

NO:ADMN(E)/2015-16/ 2751/NICH
GOVERNMENT OF SINDH
NATIONAL INSTITUTE OF CHILD HEALTH
KARACHI-75510.

TERMS AND CONDITION OF TENDER FOR REPAIR OF PLANT & MACHINERY, FOR N.I.C.H., KARARCHI, FOR THE YEAR 2015-2016

1. Cost of the Tender is : Rs.500/= (None Refundable)
2. Tender selling date is from : 07-09-2015 to 20- 09-2015
3. Tender submission date is : 21-09-2015 from 09:00 am to 10:30 am
4. Tender will be opened on : 21-09-2015 at 11:30 am
: Participants, in the **Conference Room, NICH**

NOTE: - READ THE TERMS AND CONDITIONS CAREFULLY:

Tenders are required to comply with all the clauses mentioned in the terms and conditions of the tender with all the relevant documents required and any deviation/short of document or incomplete tenders will forbid the participant from competing in the tender. The terms and conditions are as under:

1. In this tender two envelopes procedure will be followed i.e. technical bid and financial bid will be submitted in separate sealed envelopes.
 - a. In technical bid the bidder must provide original data sheet and technical brochure otherwise the bid will be rejected. The technical evaluation will be done on the basis of data sheet provided by the vendor as per technical bid.
 - b) In commercial bid the vendor should mention financial offer along with draft of 2.5 % earnest money/deposit at call in favor of the Director, NICH.
2. The Photostat copy of deposit at call if attached with the technical should be made in such a way that its value is neither disclosed nor readable. This can be done by placing a piece of paper on the place where amount is mentioned.
3. The disclosure of firm's equipment price at the time of opening of technical bid will result in to rejection of the bid.
4. First scrutiny of technical bids will be performed by the technical committee and financial bids of those firms will be opened which qualify technically.
5. The successful bidder will be required to submit bank Guarantee @ 5 % of the total value of order to NICH for warrantee period of three years, which will be required before issuing orders as security money, failing which the Earnest Money will be forfeited to Government accounts. Submission of security money will not considered as final order.
6. The vendor will give one year warrantee,
7. No tender will be entertained without call deposit conditional tender will also not be

accepted.

8. The Tenderer has to quote only one rate for each item as per tender specifications no alternate model will be accepted only first offer will be considered. Hand written tenders or any over writing, cutting, erasing shall be liable to disqualification firm.
9. The bids will be valid till **30TH June 2016 from** the date of the opening of the tenders.
10. Price escalation will not be allowed.
11. Tender which does not fulfill the prescribed condition in the tender is liable to be rejected.
12. No tender will be issued on the opening day.
13. The payment will be made after satisfactory report of the user department and electro-medical department countersigned by authorized person.
14. Income Tax will be deducted according to government rules.
32. The attached question Performa must be properly filled and should be complete in all respect in case questioner Performa is incomplete or information provided are incorrect the tender shall be rejected.
33. The Director, NICH reserves the right to reject or accept any tender without assigning reason thereof and the decision of the competent Authority will be final.
34. The rate quoted must be in Pak. Rupees.
35. If any firm refuses to deposit the security money his pay order will be forfeited to the Government Account, after depositing the security money, if he fails to supply the required item or provides the substandard quality or refuses to follow the orders, his security money also will be forfeited to the Government Accounts.. No excuse will be accepted in case of inferior and substandard quality supplies and it will be replaced forthwith.
36. The Director NICH, Karachi reserves the rights to reject or accept all or any of the tender articles on technical grounds.
37. The Director NICH, Karachi reserves the right to take all the quantities mentioned in the schedule or part thereof.
38. The Decision of the Director, NICH, Karachi, will be final.

AN UNDERTAKING CONTAINING THE MATTER MENTIONED BELOW ON STAMP PAPER

We hereby confirm to have read carefully the description of stores/items and all the terms and conditions of your tender enquiry due for opening on **21-09-2015 for the repair of Plant & Machinery for NICH, Karachi.**

We also hereby categorically confirm that the stores offered by us are exactly to the particulars and specification as laid down in your tender enquiry in all respects.

We hereby confirm that the supply be made available within 30 days after placing the order, Otherwise NICH reserves the right to take action as per terms and conditions.

We certify that we will abide by all the terms and conditions of the tender. Infringement of any of the terms/conditions will make the tender invalid as recommended by the competent authority.

Certify that no government servant has directly or indirectly a share or interest with our firm.

The tender should be addressed to the Director, NICH, Karachi and be submitted in sealed envelope and dropped in the tender box placed in the Office of the Procurement Department on above mentioned date and time. Which will be opened before the Tenderer if present at 11:30 AM the same day.

Late tender (i.e. tenders received after opening of tender box 11:30 AM will not be accepted and will be returned by hand. Instruction on work order should also be followed.

FOLLOWING DOCUMENTS ARE TO BE SUBMITTED ALONGWITH TENDER

		yes	No
a.	<i>Tender Purchase Receipt in Original.</i>		
b.	<i>Earnest Money as per tender condition.</i>		
c.	<i>Income Tax Certificate.</i>		
d.	<i>GST Registration Certificate.</i>		
e.	<i>A Photostat copy of the national Identity Card.</i>		
f.	<i>Original Terms and Conditions of the Tender duly signed by the Supplier that they have to abide by all the above mentioned terms and conditions.</i>		
g.	<i>List of Hospitals / Institutes where these type of equipment, they have supplied.</i>		
h.	<i>List of Trained / Skilled Personnel and Equipment available in their workshop for the facility of Service / Maintenance & Complete Profile of the Company.</i>		
i.	<i>Certificate that the price quoted to NICH, Karachi, are not more than the price quoted to any Government / Semi Government / Autonomous Department.</i>		
j.	<i>A Certificate that the firm will abide all terms and conditions of the tender.</i>		

CERTIFICATE

Certified that all terms and conditions mentioned in the tender are the acceptable and will abide them strictly

Signature : _____

Name : _____

Address : _____

Telephone No : _____

WITNESS: 1) _____

WITNESS: 2) _____

NO: ADMN/2015-16/ 2751 /NICH
GOVERNMENT OF SINDH
NATIONAL INSTITUTE OF CHILD HEALTH
KARACHI-75510.

**TENDER FOR REPAIR OF PLANT & MACHINERY, FOR N.I.C.H.,
KARACHI, FOR THE YEAR 2014-2015**

No.	NAME / SPECIFICATION OF MACHINERY & EQUIPMENTS	Req. Qty	Yes / no / put desired value	Rate	Amount
1.	REPAIR / REPLACEMENT OF OXYGEN VACUUM / AIR LINE AND OUTLET POINTS / BAD HEAD UNITS:				
	Model / Cat. No.	Qty	Unit		
	Make:				
	Sole Agent Exclusively for Pakistan (Yes / No)				
1	Copper pipe De-Oxidized De-Greased & half hard solid. Drown with required fittings				
	a). 1" dia	100	prft		
	b). 3/4" dia	804	prft		
	c). 1/2" dia	850	prft		
2	Medical gas outlet point should confirm to BS EN 737-1:1998, HTM 0201 and HTM 2022 standards, each quick-connect outlet shall have a large colour coded front plate for ease of gas identification and anaesthetic appeal. The primary elements in a terminal unit are a gas-specific probe and a socket with dimension that are specific for each medical gas.				
	O2 OUTLET POINT	93	Nos		
	AIR 4 BAR OUTLET POINT	69	Nos		
	VACUUM OUTLET POINT	69	Nos		
3	Isolation ball valves should be of the lever-ball type, having brass threaded adaptor connection which open and close with a 90-degree rotation the handle should be in line with the pipeline when open. Ball valves shall be designed for working pressure 600 psig WOG. Valve body shall have Teflon material ball seat and Teflon material stem seals.				
	a). BALL VALVE 1"	1	No		
	b). BALL VALVE 3/4"	2	Nos		
	c). BALL VALVE 3/8"	3	Nos		
4	Zone valve boxes shall be constructed of aluminum finish. Frames for all valve boxes shall have uniform width for balanced appearance. The front assembly shall contain visibility. Valves inside the box shall be labeled "Caution-Medical Gas Shut-Off Valves - Close Only in Emergency." Placement of the valves within the zone valve box shall be such that the visible window contains a clear viewing space shall be provided in the window to display the gas service, the area controlled by the valve, and the pressure gauges.				
	a). OXYGEN	6	Nos		
	b). AIR 4 BAR	6	Nos		
	c). VACUUM	6	Nos		
5	Power. Internal voltage shall be stepped down to 5 VDC and 15				

VDC Area alarm panels shall be designed to meet the requirements of NFPA and CSA standard. Alarms factory wiring, transformers and circuitry requiring only 230 volt AC primary for control circuit power. Voltage to external pressure or vacuum transducers shall be 15 VDC. Each gas monitored shall have a line pressure Gauge display to continuously indicate actual line pressure. The control module shall include a Silence / Enter button, shall be used to silence the audible alarm, set up the alarm panel. An LED on the control module shall illuminate green to indicate power on.

a). OXYGEN	1	Nos			
b). A R 4 BAR	1	Nos			
c). VACUUM	1	Nos			
CONTINUOUS ALUMINUM PATIENT BED HEAD PANEL.	360	prft			
With all standard accessories as per our requirement.					
Other details:					