



SHAHEED BENAZIR BHUTTO ACCIDENT EMERGENCY &
TRAUMA CENTRE KARACHI

Ref #AMS/SBB-TC/PROC/2017-18/284
Dated: 23-01-2018

Ph: 02199216380
Fax: 02199215733

CORRIGENDUM

SUBJECT: TENDER NAME: SUPPLY, INSTALLATION, TESTING & COMMISSIONING OF MEDICAL EQUIPMENT, MACHINERY & OTHER ITEMS
TENDER REFERENCE NO. AMS/SBB-TC/PROC/(LSP-03)2017-18

It is notified for information to all the interested bidders regarding Notice Inviting Tender (NIT) against **Supply, Installation, Testing & Commissioning of Medical Equipment, Machinery & Other Items** (Reference No.: AMS/SBB-TC/PROC/(LSP-03)2017-18) hoisted at Authority's website at SPPRA Serial No. 36019. Following changes have been made after a pre-bid meeting conducted on Thursday, 18-January-2018.

1. Section A. Ophthalmology page # 14 Item # 1 Operating Microscope;

Line # 3 "handgrip" is **deleted**

Line # 14 "Integrated" is **deleted** and should be read as "HD Camera system with 3 Chips"

Line # 16 Integrated is **deleted** and should be read as "Beam Splitter"

2. Section B. Anaesthesia page # 17-19 Item # 04(a) and 04(b) Flexible Bronchoscope + Light Source

Item # 04(a)	Item # 04(b) 5.5mm diameter
All below should be read as; Line # 2 Integrated / Separate Line # 3 Superfine Fiber / Fiber less Line # 8 95 to 100° Line # 09 is deleted Line # 10 88 to 93cm Line # 14, 18 & 19 are deleted Line # 16 Irrigation adapter / soaking cap Portable Imaging Monitor should be read as Line # 15 compatible biopsy forceps Line # 16 compatible grasping forceps	All below should be read as; Line # 16 4.9 – 5.5 x 60 - 65 Line # 20 Tube holder / ETT placement capability Line # 24 & 25 are deleted Line # 34 compatible biopsy forceps Line # 35 compatible grasping forceps Please add feature of "instrument channel"

3. Section C. Operation Theater / ICU page # 19-20 item # 02 ECHO Machine for Cardiac ICU

Should be read as "**Touch command screen control 08 inches or more**" instead of "Touch command screen control 10-inches or more"

Point # 06 Standard Probes: Please add one curvilinear / convex probe 2-6mhz

4. Section D. Radiology page # 23-25 Item # 2 Ultrasound Machine

Please add "**Bidders are required to demonstrate the working of quoted items as per the convenience of the end user**"

Other Terms & Conditions will remain same as advertised in previous N.I.T hoisted at Authority's website and the tender will open on 29-January-2018 as per schedule.

All interested bidders are requested to note.

ADDL MEDICAL SUPERINTENDENT/DDO
SBB ACCIDENT EMERGENCY & TRAUMA CENTRE KARACHI