

**TECHNICAL BID EVOLUTION REPORT FOR PROCUREMENT OF MACHINERY / EQUIPMENTS FOR YEAR 2015-2016**

**TECHNICAL EVALUATION OF BID**

| S.#  | Evaluation Criteria   | M/s Global Marketing Services Rawalpindi | M/s M. J Trading Corporation Kyc: | M/s Sindh Medical Stores Karachi       | M/s Scien Tech: Karachi | M/s Popular International (Pvt)  | M/s Muslim Trading Agency Karachi | M/s Lab Link Enterprises Hyd | M/s Parras Enterprise : Kyc: | M/s RTJ Enterprises Hyd: |
|--|---|--|-----------------------------------|--|-------------------------|--|-----------------------------------|------------------------------|------------------------------|--------------------------|
| 1  | Conforming technical specifications of the product  |  |                                   |  |                         |  |                                   |                              |                              |                          |
|  | Fulfillment of the tender conditions of ITB clause 8 & 11   |  |                                   |  |                         |  |                                   |                              |                              |                          |
|  | <b>ITB CLAUSE 8</b><br>Bid Form and Price Schedule (to be submitted along with financial proposal);<br>Documentary evidence to the effect that the Bidder is eligible to bid and is qualified to perform the Contract if its bid is accepted; | Y  | Y                                 | Y                                      | Y                       | Y  | Y                                 | Y                            | Y                            | Y                        |
| Documentary evidence to the effect that the goods to be supplied by the Bidder are eligible goods as defined in clause-3 and conform to the bidding documents; and | Y   | Y  | Y                                 | Y                                      | Y                       | Y  | Y                                 | Y                            | Y                            |                          |
| Bid Security   | Y   |  | Y                                 | Y                                      | Y                       | Y  | Y                                 | Y                            | Y                            |                          |
| <b>ITB CLAUSE 11</b><br>The Supplier/Agent/Importer shall have to produce letter of authorization & copy FDA-CE-ISO Certified Manufacturer.                        | Y   | N  | Y                                 | M/s Grant Instrument authority expired | Y                       | M/s S.U.E has issued the authority in favour of Lab Link for Heitich Germany not Muslim Trading Agency Karachi | Y                                 | Y                            | Y                            | Y                        |

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

*[Handwritten signature]*

|    |   |   |                 |   |   |   |   |   |   |   |   |   |   |   |                |
|----|---|---|-----------------|---|---|---|---|---|---|---|---|---|---|---|----------------|
|    | The Bidder/Manufacturer will submit an affidavit on legal stamp paper of Rs. 100/- to the effect that their firm has not been blacklisted in the past on any ground by any Government (Federal or Provincial), a local body or a public sector organization. The Bidder will be debarred from the bidding process for submitting a false statement; | Y   | N               | Y | Y   | Y | Y   | Y | Y | Y | Y | Y | Y | Y | Y              |
|    | The Bidder is required to provide with its technical proposal the names of the goods for which it has quoted rates in the financial proposal;   | Y   | Y               | Y | Y   | Y | Y   | Y | Y | Y | Y | Y | Y | Y | Y              |
|    | The Bidder / Representative must submit the authority letter during opening of Technical and financial proposals.   | Y   | Y               | Y | Y   | Y | Y   | Y | Y | Y | Y | Y | Y | Y | Y              |
| 2  | - Manufacturer Authorization  | Y   | N               | Y | M/s Grant Instrument authority expired    | Y | M/s S.U.E has issued the authority in favour of Lab Link for Heitich Germany not Muslim Trading | Y | Y | Y | Y | Y | Y | Y | Y              |
| 3  | - Certification FDA/ IVD/ CE Mark   | Y   | N               | Y | N   | Y |   | Y | Y | Y | Y | Y | Y | Y | Y              |
| 04 | - Original Technical Broacher   | Y   | Only photo copy | Y | Only photo copy                           | Y | N   | Y | Y | Y | Y | Y | Y | Y | Y              |
| 05 | - Proof of financial soundness Certificate(s) from schedule Bank about 50 (M) each (last 3 years)   | Only Bank account maintenance certificate | N               | Y | Only Bank account maintenance certificate | Y | Y   | Y | Y | Y | Y | Y | Y | Y | Y              |
| 06 | Proven Track Record last 03 years of quoted items   | Y   | N               | Y | 2 years                                   | Y | N   | Y | Y | Y | Y | Y | Y | Y | Y              |
| 07 | Registration with SBR   | Y   | N               | Y | N   | Y | N   | Y | Y | Y | Y | Y | Y | Y | Not applicable |
| 08 | Bank Statement last 3 years   | Y   | Y               | Y | Y   | Y | Y   | Y | Y | Y | Y | Y | Y | Y | Not applicable |

*A. M. P. S.*

*S. S. S.*

*SB*

|    |  |           |              |           |              |           |              |           |           |           |           |           |
|----|--|-----------|--------------|-----------|--------------|-----------|--------------|-----------|-----------|-----------|-----------|-----------|
| 09 | Audit Reports of three years ( 2012-2013 , 2013-2014and 2014-2015) (Mandatory) | Y         | N            | Y         | N            | Y         | N            | Y         | Y         | Y         | Y         | Y         |
| 10 | Proof of General Sales Tax paid during the last three years                    | Y         | Y            | Y         | Y            | Y         | N            | Y         | Y         | Y         | Y         | Y         |
| 11 | Proof of Income Tax paid during the last three years                           | Y         | 2 Years      | Y         | Y            | Y         | N            | Y         | Y         | Y         | Y         | Y         |
| 12 | Appropriateness of supply schedule offered by the bidder                       | Y         | Y            | Y         | N            | Y         | Y            | Y         | Y         | Y         | Y         | Y         |
| 13 | Copy of Professional Tax Certificate current year                              | Y         | N            | Y         | Y            | Y         | Y            | Y         | Y         | Y         | Y         | Y         |
| 14 | Copy of Chamber & commerce Industries certificate current year                 | Y         | Y            | Y         | N            | Y         | Y            | Y         | Y         | Y         | Y         | Y         |
| 15 | Copy of Income Tax Certificate   | Y         | Y            | Y         | Y            | Y         | Y            | Y         | Y         | Y         | Y         | Y         |
| 16 | Copy of Sales Tax Certificate  | Y         | Y            | Y         | Y            | Y         | Y            | Y         | Y         | Y         | Y         | Y         |
| 17 | Original tender purchase receipt   | Y         | Y            | Y         | Y            | Y         | Y            | Y         | Y         | Y         | Y         | Y         |
|    | Qualified / Disqualified   | Qualified | Disqualified | Qualified | Disqualified | Qualified | Disqualified | Qualified | Qualified | Qualified | Qualified | Qualified |

Rep. of District Accounts Officer  
Hyderabad  
(Member)

Rep. of Commissioner  
Hyderabad Division  
(Member)

Dr. Bikha Ram Devrajani  
Prof. of Medicines/Hepatologist  
LUMHS Jamshoro  
(Member)

Dr. Ikramuddin Ujian  
Prof. of Pathologist  
LUMHS Jamshoro/Hyderabad  
(Member)

Deputy Secretary (PM&I)  
Government of Sindh  
Health Department, Karachi  
Member

Dr. Abdul Khaliq Shaikh  
Program Manager  
Prevention & Control Program Sindh  
Hyderabad  
Member

Dr. Hassan Murad Shah  
Director General  
Health Services Sindh  
Hyderabad  
Member

**Technical evaluation report of Machinery / Equipments of Hepatitis Prevention & Control Program  
(Chief Minister's Initiative) Sindh for the year 2015-2016**

| S.NO | Equipment Name Tender specification  | Brand                                       | CE/FDA Marked                                     | Supplier                                    | Origin  | Remarks  |
|------|--|---|---|---|---------|--|
| 1    | <b>Real Time PCR System / Machine:</b> FDA Approved/CE IVD Marked<br>Automatic Real Time PCR Amplification Open System provide measurement of the stages of polymerase chain reaction (PCR) for IVD , with five channels or more than five channels for both qualitative and quantitative RNA , DNA assay including Genotype and capacity of up to 96 samples per run. | Rotorgene, Qiagen                           | CE marked 5 channels 96 samples                   | M/s Lab link Enterprises Hyderabad (Biogen) | Germany | Qualified as per tender specification  |
|      |  | Analyzer                                    | FDA, CE Mark 4 channels                           | M/s Popular International Hyderabad         | Germany | Not as per tender 4 Channel Only Hence disqualified                                |
|      |  | Sacche                                      | CE marked 4 Channel 96 samples                    | M/s Global Marketing Services Rawalpindi    | Italy   | Not as per tender Specs (4 Channel Only) Hence disqualified                        |
|      |  | Bioer Technology                            | CE marked 4 Channel 96 samples                    | M/s RTJ Enterprises Hyderabad               | China   | Not as per tender Specification (4 Channel Only) hence disqualified                |
|      |  | Sacche                                      | CE marked 12 to 24 samples capacity               | M/s Global Marketing Services Rawalpindi    | Italy   | Not as per tender specification (Low sample processing capacity Hence disqualified |
|      | Automated sample (RNA and DNA) Extraction and Amplification/ Detection System Automated Extraction Open System for viral nucleic acids ( RNA , DNA) extraction, based on latest technology with precision pipetting, flexible RNA,DNA protocols efficient work flow and a capacity of up to 96 samples per run.  | QIAsymphony SP- Automated extraction system | IVD, CE marked 1 to 96 sample processing capacity | M/s Lab link Enterprises Hyderabad          | Germany | Qualified as per all tender specifications   |

*[Handwritten Signature]*

*[Handwritten Signature]*

*[Handwritten Signature]*

|          |   |   |   |   |  |   |
|----------|---|---|---|---|--|---|
|          |   | <p>NEC 32<br/>Nucleic Acid<br/>Extraction<br/>System</p>  | <p>FDA, CE Mark</p>   | <p>M/s Popular<br/>International<br/>Hyderabad</p>  | <p>Germany</p>                             | <p>Extraction system has not sufficient information to decide either it is close or open system. Hepatitis kits avail DNA/RNA kits of any origin may be used with this system or not, it is not clear, so, disqualified</p> |
| <p>2</p> | <p><b>Complete System for ELISA Tests: <u>FDA Approved/CE Marked</u></b><br/><b>Microplate Absorbance Reader, Microplate Washer, Microplate Incubator</b></p> <p>Automated Programmable open system 8 Channel for 96 well Elisa Microplate Absorbance Reader, with Latest LED Technology, Windows 7-based software, Power Data Analysis with full qualitative and quantitative EIA analysis. It should provide accurate, reproducible, fast measurements, full traceability and connectivity to Laboratory Information System (LIS).</p> <p>The Microplate Washer should be Programmable 8-channel washer for 96-well microplates, and should include bottles, tubing</p> | <p>StatFax 4200, Reader Awareness Technology</p> <p>StatFax 2600 Washer Awareness Technology</p> <p>StatFax 2200 Incubator Awareness Technology</p> <p>LMP101 Washer Labcon</p> | <p>FDA Approved, CE marked 8 Channels 405 to 700 nm 4 standard filters 96 well</p> <p>FDA approved, CE marked</p> <p>FDA approved CE marked, ambient to 40 C, 575 to 1500 rpm</p> | <p>M/s Sindh Medical Store Karachi</p> <p>M/s Sind Medical Store Karachi</p> <p>M/s Sind Medical Store Karachi</p> <p>M/s RTJ Enterprises Hyderabad</p> | <p>USA</p> <p>USA</p> <p>USA</p> <p>UK</p> | <p>Qualified as Tender specifications</p>   |

*Handwritten signature*

*Handwritten signature*

|   |   |   |   |  |                |  |
|---|---|---|---|--|----------------|--|
| 3 | <p>set, built-in vacuum , dispensing pumps, and maintenance kit<br/>Incubator for microplates, should deliver a controlled temperature environment for optimal results.</p>   | LMPR 101<br>Microplate Reader<br>Labocon        |   | M/s RTJ<br>Enterprises<br>Hyderabad            | UK             |  |
|   |   | LM1-100L<br>Labocon<br>Incubator                |   | M/s RTJ<br>Enterprises<br>Hyderabad            | UK             |  |
|   |   | Humareader<br>HS<br>(Elisa Plate Reader)        |   | M/s Popular<br>International<br>Hyderabad      | Germany        | Not as per tender specification , hence disqualified           |
|   |   | Combiwas<br>(Elisa Plate Washer with incubator) |   | M/s Popular<br>International<br>Hyderabad      | Germany        | Not as per tender specification , hence disqualified           |
| 3 | <p><b>Semi Automated Chemistry Analyzer: FDA Approved/CE Marked.</b><br/>Easy Touch Screen Operation with LCD Display, Wavelength range is 340 - 630 nm, 6 filter nt open system with option of Flow Cell and Cuvette</p> | Bas-120Ts,<br>Labomed                           | CE, FDA.  | M/s RTJ<br>Enterprises<br>Hyderabad            | USA            | Not as per tender specification , hence disqualified           |
|   |   | StatFax4500                                     | FDA, CE Mark  | M/s Sindh Medical<br>Store Karachi             | USA            | As per specification Qualified                                 |
|   |   | Mini Tecno                                      | CE<br>340 to 630nm<br>8 position filter<br>wheel<br>Six filters | M/s Global<br>Marketing Services<br>Rawalpindi | Adaltis, Italy | Qualified as per tender specification.                         |
| 4 | <p><b>Vertical Auto Clave 20 Ltr Gas / Electric</b><br/>UK, USA, Japan, Europe</p>  | Humalyzer<br>3500                               |   | M/s Popular<br>International<br>Hyderabad      | Germany        | No information about channels and filters , hence disqualified |
|   |   | TY-220, EQUUS                                   |   | M/s RTJ<br>Enterprises<br>Hyderabad            | Taiwan         | Qualified as per tender specification.                         |

*Handwritten signature*

*Handwritten signature*

|  |            |  |                                   |       |  |
|--|------------|--|-----------------------------------|-------|--|
| <p><b>5 Stabilizers 10000 Watts</b><br/> <b>Technical Specifications:</b><br/> 4 Relays 10000 watts Automatic Voltage Stabilizer<br/> Input Range: 100-260 V (AC)<br/> Output Voltage: 220 V (AC)<br/> +3.5%<br/> Phase: Single<br/> Frequency: 50 + 3 Hz<br/> Winding : 100% Copper winding<br/> Warranty: 1 year full warranty</p>   | Panasonic  |  | M/s Parras Enterprises Karachi    | Local | Qualified as per tender specification. |
| <p><b>6 Stabilizers 1000C Watts</b><br/> <b>Technical Specifications:</b><br/> Input Voltage: 145V-250V<br/> Output Voltage: 110V-220V<br/> Response time: 0.5 sec. against 10% input voltage deviation<br/> Output Voltage stability: ± 0 to 1 %<br/> Frequency: 50 HZ/60HZ<br/> Wave form: Pure sine wave<br/> Servo motor control<br/> Surge protection.<br/> Warranty: 1 year full warranty<br/> Protections: Overload, Surge over &amp; under voltage<br/> In Imported Pure Sine Wave Country Of Origin Taiwan/ U.K/Japan</p> | Stabimatic |  | M/s Parras Enterprises Karachi    | Local | Qualified as per tender specification. |
| <p><b>7 Desktop Computer with following accessories.</b><br/> <b>Technical Specifications</b><br/> Motherboard: ASUS H110M-K<br/> D3 DDR3 Intel LGA1151<br/> Platform - Intel H110 Express</p>   | Intel      |  | M/s Lab link Enterprise Hyderabad | China |  |

|  |              |  |  |              |   |
|--|--------------|--|--|--------------|---|
| <p>Chipset<br/> Processor: Intel Core i7 3820<br/> 2011 Socket 3.6 GHz (10MB CACHE).<br/> Hard Disk (HDD) 500GB 3.5"<br/> SATA HARD DRIVE 64MB CACHE<br/> Memory (RAM) 4GB 1600 BUS DDR 3<br/> Optical Drive ASUS DVD RW SATA 24X<br/> PC Casing LEGEND<br/> THERMAL HT SUPPORT (WITH NORMAL PSU)<br/> LCD Monitor LED S19A100N 19" Resolution: 13,66 x 768, Panel type: TN<br/> Color Support: 16.7M, HIGH GLOSSY BLACK<br/> Keyboard: Standard Keyboard with USB port<br/> Mouse: Optical Scroll Mouse</p> | <p>Intel</p> |  | <p>M/s Parras Enterprises Karachi</p>    | <p>China</p> | <p>Qualified as per tender specification.</p> |
| <p><b>Laser Printers</b><br/> *CAPT &amp; SCOA Technology for faster printing<br/> *150 sheets paper cassette<br/> *1 sheet legal size manual feed<br/> *First print out in less than 9.3 secs<br/> *USB 2.0 Hi-Speed<br/> *12ppm (A4) printing speed<br/> *True 600 x 600dpi resolution<br/> *2400 x 600dpi equivalent with Automatic Image Refinement Technology</p>   | <p>HP</p>    |  | <p>M/s Lab link Enterprise Hyderabad</p> | <p>China</p> | <p>Qualified as per tender specification</p>  |
| <p>8</p>   | <p>HP</p>    |  | <p>M/s Parras Enterprises Karachi</p>    | <p>China</p> | <p>Qualified as per tender specification</p>  |



|   |                   |  |  |              |  |
|---|-------------------|--|--|--------------|--|
| <p><b>9</b></p> <p><b>Uninterruptible Power Supply</b></p> <p>Technical Specifications:<br/> Imported Pure Sine Wave 650VA<br/> Input Range: 145-275 V (AC)<br/> Output Voltage: 220 V (AC)<br/> +10%<br/> Phase: Single<br/> Frequency: 50 or 60 Hz<br/> Battery Type: Dry Battery New<br/> Warning System: Audio (Buzzer),<br/> Various Warning Alarm<br/> Protections: Overload, Surge<br/> over &amp; under voltage<br/> In Imported Pure Sine Wave<br/> Country Of Origin Taiwan/<br/> U/K/Japan</p> | <p>I power</p>    |  | <p>M/s Lab link<br/> Enterprise<br/> Hyderabad</p> | <p>Local</p> | <p>Qualified as per<br/> tender<br/> specification</p> |
| <p><b>10</b></p> <p>Photocopier Machine<br/> MP-2051, 25 copies /min A4 size<br/> Zoom. 50% to 200%<br/> Copy paper MaxA3 Min A6<br/> Printer networking/scanner<br/> building</p>  | <p>High Power</p> |  | <p>M/s Parras<br/> Enterprises Karachi</p>         | <p>Local</p> | <p>Qualified as per<br/> tender<br/> specification</p> |
| <p><b>11</b></p> <p><b>Plain Paper Fax Machine</b><br/> Technical Specifications:<br/> Distinctive /ring detection<br/> 2-Line LCD readout<br/> 64-level halftone resolution<br/> Polling function<br/> Friendly fax reception<br/> Extension line transfer<br/> 10-page automatic document<br/> feeder<br/> 50-sheet paper tray<br/> 1-6 one-touch dialer<br/> Phone book 100-station<br/> Electronic volume control<br/> Handset mute</p>   | <p>Panasonic</p>  |  | <p>M/s Lab link<br/> Enterprise<br/> Hyderabad</p> | <p>China</p> | <p>Qualified as per<br/> tender<br/> specification</p> |

|  |              |  |  |              |  |
|--|--------------|--|--|--------------|--|
| <p>External telephone jack.<br/>Document Memory Max. 28 Page<br/>Memory<br/>Ink Film 70m ( 10m Included<br/>Film)</p>  |              |  |  |              |  |
| <p>12 <b>Laptop Computer with following accessories</b><br/><br/><b>Operating System:</b> Windows 8.1<br/>64 bit<br/><b>Processor:</b> Intel Core i7-5500U<br/>with Intel HD Graphics 5500 (2.4<br/>GHz, 4 MB cache, 2 cores)<br/><b>Processor family:</b> Intel Core i7<br/>processor<br/><b>Memory:</b> 4 GB DDR3L-1600<br/>SDRAM (1 x 4 GB); Transfer<br/>rates up to 1600 MT/s<br/><b>Storage:</b> 500 GB 5400 rpm<br/>SATA, Dropbox, SuperMulti DVD<br/>burner<br/><b>Webcam:</b> HP TrueVision HD<br/>Webcam (front-facing) with<br/>integrated digital microphone<br/><b>Communication:</b> Integrated<br/>10/100 BASE-T Ethernet LAN;<br/>802.11b/g/n (1x1) and Bluetooth<br/>4.0 combo (Miracast compatible)<br/><b>Display:</b> 39.6 cm (15.6inch)<br/>diagonal HD BrightView WLED-<br/>backlit (1366 x 768)<br/><b>Video:</b> AMD Radeon R5 M330<br/>Graphics (2 GB DDR3 dedicated)<br/><b>Sound:</b> DTS Studio Sound with 2</p> | <p>HP450</p> |  | <p>M/s Lab link<br/>Enterprise<br/>Hyderabad</p> | <p>China</p> | <p>Qualified as per<br/>tender<br/>specification</p> |

|    |   |                           |   |         |  |         |  |
|----|---|---------------------------|---|---------|--|---------|--|
|    | speakers<br><b>Ports:</b> 1 HDMI; 1 headphone/microphone combo; 2 USB 2.0; 1 USB 3.0; 1 RJ-45 1 multi-format SD media card reader<br><b>Battery Type :</b> 4-cell, 41 Wh Li-ion polymer<br><b>Weight:</b> 2.19 kg; Packed: 3.05 kg; Weight varies by configuration<br>Branded |                           |   |         |  |         |  |
| 13 | Generator 15 KVA with canopy petrol and Gas   | Sawan                     |   |         | M/s Parras Enterprises Karachi           | China   | Qualified as per tender specification  |
| 14 | 3 K.V Generator with gas kit & petrol and Gas   | Lancer                    |   |         | M/s Parras Enterprises Karachi           | China   | Qualified as per tender specification  |
| 15 | Centrifuge Machines: 8 x15 ml rotor, with speed of 6000rpm.   | EU Poland                 | IVD<br>Marked<br>6000rpm,<br>x15 ml               | CE<br>8 | M/s RTJ Enterprises Hyderabad            | Poland  | Qualified as per tender specification  |
|    |   | NF200, NUVE               | 5000 rpm<br>12x15<br>CE                           |         | M/s Global Marketing Services Rawalpindi | Turkey  | Disqualified not as per tender specification<br>No information about CE/FDA certificates |
|    |   | HUMAX 4K                  | 12x15 ml  |         | M/s Popular International Hyderabad      | Germany | Not as per tender specification hence disqualified                                       |
| 16 | Micro Centrifuge Machines   | Biocen 22R<br>Orto Alresa | CE marked,<br>18000 rpm,<br>Angle Fixed<br>Rotor, |         | M/s RTJ Enterprises Hyderabad            | Poland  | Qualified as per tender specification  |


*A. M. Khan*

|    |   |  |  |   |                  |  |
|----|---|--|--|---|------------------|--|
|    | <b>with refrigeration: Temperature -10 to +40°C 30x2 ml rotor, with speed of 15,000rpm.</b>                             | Z323 K labnet  | 1.5-2 ml 17000 rpm<br>24x1.5/2 ml, -20 to +40C                           | M/s Global Marketing Services Rawalpindi                                  | USA              | Not as per tender specification, hence disqualified  |
| 17 | <b>Micro Centrifuge Machine: 30x2 / 1.5 ml rotor with speed of 15,000 rpm, non refrigerated</b>                         | MPW-55<br>MPW Med<br>NF048, Nuve                           | IVC CE marked<br>14500 rpm<br>12000 rpm,<br>24/1.5                       | M/s RTJ Enterprises Hyderabad<br>M/s Global Marketing Services Rawalpindi | Poland<br>Turkey | Qualified as per tender specification<br><br>Disqualified not as per tender specification & due to unavailability of CE/FDA certificates |
| 18 | <b>Micro Centrifuge Machine: 6 to 8x2 ml rotor, along with adapters for 0.5 and 0.2 ml tubes with speed of 7000 rpm</b> | Mini Star,<br>Tomos Life<br>Sciences                       | 7200 rpm,  | M/s RTJ Enterprises Hyderabad   | Singapur/USA     | As per tender specification Hence Qualified  |
| 19 | <b>Safety Cabinet class A2 size 4 feet, with hepa filter.</b>   | BO-120PP<br><br>Biobase                                    | HEPA filter,<br>57x25.6x21.65<br><br>Class II, Type A2, H14 HEPA Filters | M/s RTJ Enterprises Hyderabad<br>M/s Global Marketing Services Rawalpindi | USA<br>China     | Qualified as per tender specification<br><br>Qualified as per tender specification   |
| 20 | <b>Heat Block, temperature range Ambient to 200°C</b>   | H203-H,<br>COYOTE BIO LLC<br>Accublock<br>D1100,<br>Labnet | 2 TO 100 C<br><br>5 to 150C  | M/s RTJ Enterprises Hyderabad<br>M/s Global Marketing Services Rawalpindi | COLUMBA<br>USA   | Qualified as per tender specification<br><br>Qualified as per tender specification   |

*Handwritten signature*

|    |  |   |   |   |                    |  |
|----|--|---|---|---|--------------------|--|
|    |  | HB120S<br>Dragon Lab  | 1.5/2ml x 40,<br>upto 120C                                    | M/s Scien Tec<br>Karachi  | China              | Disqualified due to<br>incomplete<br>submission<br>of<br>documents<br>mentioned in the<br>evaluation criteria. |
| 21 | Vortex Mixer, speed flexible rpm   | Molecule-On<br>VXX200,<br>Labnet  | Flexible speed<br>1 K to 3 K<br>Up to 2850 rpm                | M/s RTJ<br>Enterprises<br>Hyderabad<br>M/s Global<br>Marketing Services<br>Rawalpindi | New Zealand<br>USA | Qualified as per<br>tender<br>specification<br>Qualified as per<br>tender<br>specification                     |
| 22 | Ultra Low Freezer, temperature<br>-40 to -80 degree C, Upright, 600<br>or more than 600 liters, with<br>temperature control Display, and<br>temperature sensitive alarm<br>system. | Gsple 6040<br>ewn, National<br>lab GMBH<br>Ultra freezer<br>600 Fiochetti         | 40 to -85,<br>585 Lit<br>-60C upto -86<br>C 600 Lit           | M/s RTJ<br>Enterprises<br>Hyderabad<br>M/s Global<br>Marketing Services<br>Rawalpindi | Germany<br>Italy   | Disqualified not as<br>per tender<br>specification<br>Qualified as per<br>tender<br>specification              |
| 23 | Pharmaceutical Refrigerator ,<br>temperature 2 to 8 degree C,<br>Upright, 300 or more than 300<br>liters, with large LED temperature<br>control Display, and temperature           | YC-300L,<br>Zhongke<br>meiling<br>Cryogenics<br>Medika 400<br>ECT F,<br>Fiochetti | CE marked,<br>300 Liter<br>2 to 10 C<br>347 Liter, 2 to<br>15 | M/s RTJ<br>Enterprises<br>Hyderabad<br>M/s Global<br>Marketing Services<br>Rawalpindi | China<br>Italy     | Qualified as per<br>tender<br>specification<br>Qualified as per<br>tender<br>specification                     |

|    |  |                     |                   |                                      |          |  |
|----|--|---------------------|-------------------|--------------------------------------|----------|--|
|    | alarm system.  | MEC-450<br>Caravell | 0 to 8 C, 450 Lit | M/s Parras<br>Enterprises Karachi    | Pakistan | Qualified as per<br>tender specification |
| 24 | UPS 3 KW Imported  | N.Q                 | N.Q               | N.Q                                  | N.Q      |  |
| 25 | Stabilizer 10 KV Imported  | N.Q                 | N.Q               | N.Q                                  | N.Q      |  |
| 26 | <b>PCR Lab Partitioning of Laboratory</b> <ul style="list-style-type: none"> <li>Glass Aluminum Partitioning for two Rooms with Fan</li> <li>Exhaust / duct for the safety cabinet.</li> </ul> Any other civil work. |                     |                   | M/s Lab link<br>Enterprise Hyderabad |          | Qualified as per<br>tender specification |



DR. ABU TALIB  
GASTROENTEROLOGIST  
DOW UNIVERSITY OF MEDICAL SCIENCES CIVIL  
HOSPITAL  
KARACHI  
(MEMBER)



DR. BIKHRA RAM DEVRAJANI  
PROFESSOR OF MEDICINES /  
HEPATOLOGIST LUMHS  
JAMSHORO  
(MEMBER)



DR. ALI M. WARYAH  
INCHARGE MOLECULAR BIOLOGY MEDICAL  
RESEARCH CENTRE LUMHS  
JAMSHORO  
(MEMBER)



DR. IKRAMUDDIN UJAN  
PROF. OF PATHOLOGY,  
LUMHS JAMSHORO /HYDERABAD  
(MEMBER)



DR. ABDUL KHALIQUE SHAIKH  
PROGRAM MANAGER  
HEPATITIS PREVENTION CONTROL  
PROGRAM SINDH HYDERABAD  
(MEMBER /SECRETARY)



DR. HASSAN MERKAD SHAH  
DIRECTOR GENERAL  
HEALTH SERVICES SINDH  
HYDERABAD.  
(CHAIRMAN)

**MINUTES OF MEETING OF PROCUREMENT COMMITTEE FOR OPENING OF FINANCIAL BID /FINANCIAL PROPOSAL FOR PROCUREMENT OF MACHINERY & EQUIPMENTS, HEPATITIS PREVENTION & CONTROL PROGRAM SINDH HYDERABAD HELD ON 25.5.2016 AT 2.00 P.M IN THE OFFICE OF THE PROGRAM MANAGER HEPATITIS PREVENTION & CONTROL PROGRAM SINDH HYDERABAD FOR THE YEAR 2015-16.**

The meeting of Procurement Committee for opening of Financial Bid/Financial Proposals for Procurement of Machinery & Equipments, Hepatitis Prevention & Control Program Sindh Hyderabad for the year 2015-16 held on 25.5.2016 at 2.00 P.M, headed by the Chairman of the committee, in the office of the Program Manager, Hepatitis Prevention & Control Program Sindh Hyderabad.

The Procurement Committee constituted by the Government vide Government Notification No;:HD(P&D)6(101)/2015-16 dated 9.5.2016.

|   |   |                  |
|---|---|------------------|
| 1 | Director General<br>Health Services Sindh Hyderabad                               | Chairman         |
| 2 | Program Manager<br>Hepatitis Prevention & Control Program Sindh<br>Hyderabad.     | Member/Secretary |
| 3 | Dr.Bikha Ram Devrajani,<br>Professor of Medicines/Hepatologist LUMHS<br>Jamshoro. | Member           |
| 4 | Professor Dr.Ikramuddin Ujjan<br>Prof: of Patholgy, LUMHS Jamshoro/Hyderabad      | Member           |
| 5 | Deputy Secretary PM&I<br>Health Department Govt: of Sindh                         | Member           |
| 6 | Rep: of Commissioner<br>Hyderabad Division  | Member           |
| 7 | Rep: of District Accounts Officer,<br>Hyderabad.                                  | Member           |

As per evaluation report, it was found that all below mentioned firms have been fulfilled the requirements mentioned in the criteria of tender for Procurement of Machinery & Equipments is declared as qualified by the committee, Evaluation Report(Annex-A).

The letters has been issued to the qualified firms as per SPPRA rule No.45.

The Procurement Committee opened the financial Proposals of following qualified firms in the meeting held on 25.5.2016, headed by the Chairman of Procurement committee, in the office of the Program Manager Program Manager Hepatitis Prevention & Control Program Sindh Hyderabad in the presence of Bidders & their representative. During opening of Financial Proposal the rate of items publicly announced. Comparative Statement was prepared at the time of opening of financial proposals/financial bid(Annex-B).

| Sr.# | Name of firm                             |
|------|--|
| 01   | M/s Global Marketing Services Rawalpindi |
| 02   | M/s Sindh Medical Stores Karachi         |
| 03   | M/s Popular International Pvt Hyd        |

|    |                                    |
|----|------------------------------------|
| 04 | M/s Lab Link Enterprises Hyderabad |
| 05 | M/s Paras Enterprises Karachi      |
| 06 | M/s RTJ Enterprises Karachi        |

The rates offered by the M/s Global Marketing for item No.2, 20 & 21 was found lowest and approved by the Procurement committee.

The rates offered by the M/s Sindh Medical Stores for item No. 1, was found lowest and approved the committee.

The rates offered by the M/s Lab Enterprises Hyderabad for item No.6,7,10,11,12,15 a & b, was found lowest and approved the committee. The rate of item no 15a & b, were found lowest then the market price the procurement Committee unanimously decided and accepted/ approved single rate as per SPPRA Rule 48.


The rates offered by the M/s Paras Enterprises Karachi for item No.5,8,9,13,14,22 & 23. was found lowest and approved by the Procurement committee.

The rates offered by the M/s RTJ Enterprises Hyderabad, for item No.3,4,16,17,18 & 19, was found lowest and approved by the procurement committee.

The rates offered by the M/s Popular International Hyderabad, for item No.21, was found lowest and approved by the Procurement committee.



Rep: of  
District Account  
Officer Hyderabad  
Member



Rep: of Commissioner  
Hyderabad Division  
Member



Dr. Ikramuddin Ujjan  
Prof: of Pathology  
LUMHS Jamshoro /  
Hyderabad  
Member



Dr. Bikha Ram Daverjani  
Prof: of Medicines/  
Hepatologist LUMHS  
Jamshoro  
Member



Deputy Secretary PM&I  
Government of Sindh  
Health Department  
Member



Dr. Abdul Khaliq Shaikh  
Program Manager  
Hepatitis Prevention & Control  
Program Sindh  
Hyderabad  
Secretary / Member



Dr. Hassan Murad Shah  
Director General  
Health Services Sindh  
Hyderabad  
Chairman

**COMPARATIVE STATEMENT FOR PROCUREMENT OF MACHINERY/ EQUIPMENTS FOR THE YEAR 2015-16.**

(Financial)

| S.# | Description / Name of items   | M/s Global Marketing Rawalpindi | M/s Sindh Medical stores Karachi | M/s Lab Link Enterprises Karachi | M/s Parras Enterprises Karachi | M/s RTI Enterprises Hyderabad | M/s Popular Intr: Hyderabad |
|-----|---|---------------------------------|----------------------------------|----------------------------------|--------------------------------|-------------------------------|-----------------------------|
| 1   | Complete System for Elisa Test: FDA approved / CE Marked Microplate Absorbance Reader, Micro Plate Washer, Microplate Incubator   | NQ                              | 1144000/-                        | NQ                               | NQ                             | 1295000/-                     | DQ                          |
| 2.  | Semi Automated Chemistry Analyzer FDA approved/ CE Marked   | 294000/-                        | 379000/-                         | NQ                               | NQ                             | 449500/-                      | DQ                          |
| 3.  | Centrifuge Machine  | DQ                              | NQ                               | NQ                               | NQ                             | 362500/-                      | DQ                          |
| 4   | Vertical Auto clave 20 Lt: Gas /Electric  | NQ                              | NQ                               | NQ                               | NQ                             | 108000/-                      | DQ                          |
| 5   | Stapilizer 10000 watts  | NQ                              | NQ                               | 18000/-                          | 17596/-                        | NQ                            | DQ                          |
| 6   | Stapilizer 1000 C watts   | NQ                              | NQ                               | 4400/-                           | 5460/-                         | NQ                            | DQ                          |
| 7   | Desktop Computer with accessories   | NQ                              | NQ                               | 88000/-                          | 89000/-                        | NQ                            | DQ                          |
| 8   | Laser Printer   | NQ                              | NQ                               | 27000/-                          | 14900/-                        | NQ                            | DQ                          |
| 9   | Uninterruptible Power supply(UPS)   | NQ                              | NQ                               | NQ                               | 5700/-                         | NQ                            | DQ                          |
| 10  | Photocopier Machine   | NQ                              | NQ                               | 190000/-                         | NQ                             | NQ                            | DQ                          |
| 11  | Plain Paper Fax Machine   | NQ                              | NQ                               | 24000/-                          | NQ                             | NQ                            | DQ                          |
| 12  | Laptop Computer with accessories  | NQ                              | NQ                               | 87000/-                          | NQ                             | NQ                            | DQ                          |
| 13  | Generator 15 KVA  | NQ                              | NQ                               | NQ                               | 490000/-                       | NQ                            | DQ                          |
| 14  | 3 KVA Generator with Gas Kit & Petrol   | NQ                              | NQ                               | NQ                               | 41990/-                        | NQ                            | DQ                          |
| 15  | Real time PCR System/Machine FDA approved/CE Marked Automated sample (RNA and DNA. Extraction and amplification /detection system | DQ                              | NQ                               | 2625000/-<br>8875000/-           | NQ                             | NQ                            | DQ                          |
|     | PCR Accessories   |                                 |                                  |                                  |                                |                               | DQ                          |
| 16  | Micro Centrifuge Machine with refrigeration   | DQ                              | NQ                               | NQ                               | NQ                             | 845000/-                      | DQ                          |
| 17  | Micro Centrifuge Machine 30x2 /1.5 ml rotor with speed of 15000rpm nonrefrigerated  | DQ                              | NQ                               | NQ                               | NQ                             | 375000/-                      | DQ                          |
| 18  | Micro Centrifuge Machine 6ti 8x2 ml rotor, alongwith adaptors   | DQ                              | NQ                               | NQ                               | NQ                             | 23000/-                       | DQ                          |

*Handwritten signature/initials*

*Handwritten signature/initials*

*Handwritten signature/initials*

|    |                                    |           |    |    |          |           |         |    |
|----|------------------------------------|-----------|----|----|----------|-----------|---------|----|
| 19 | Safety Cabinet                     | 340000/-  | NQ | NQ | NQ       | NQ        | 78000/- | DQ |
| 20 | Heat Block                         | 39500/-   | NQ | NQ | NQ       | 188000/-  | DQ      |    |
| 21 | Vortex Mixture                     | 29800/-   | NQ | NQ | NQ       | 75000/-   | 14240/- |    |
| 22 | Ultila Low Freezer                 | 1690000/- | NQ | NQ | 119000/- | 1295000/- | DQ      |    |
| 23 | Pharmaceuticals Refrigerator       | 394000/-  | NQ | NQ | 169000   | 275000/-  | DQ      |    |
| 24 | UPS 3 KW Imported                  | NQ        | NQ | NQ | NQ       | NQ        | DQ      |    |
| 25 | Stabilizer 10 KV imported          | NQ        | NQ | NQ | NQ       | NQ        | DQ      |    |
| 26 | PCR Lab Partitioning of Laboratory | NQ        | NQ | NQ | NQ       | NQ        | DQ      |    |


  
 Representative of  
 District Accounts Officer  
 Hyderabad  
 Member

  
 Representative of  
 Commissioner Hyderabad Division  
 Member

  
 Dr. Ikramuddin Ujian  
 Prof. of Pathology  
 LUMHS Jamshoro/Hyd  
 Member

  
 Dr. Bikha Ram Daverajani  
 Professor of  
 Medicine/Hepatologist  
 LUMHS, Jamshoro  
 Member

  
 Deputy Secretary PM&I  
 Health Department  
 Government of Sindh Karachi  
 Member

  
 Dr. Abdul Khaliq Shaikh  
 Program Manager  
 Prevention & Control of Hepatitis (Chief  
 Minister Initiative) Sindh Hyderabad  
 Secretary / Member

  
 Dr. Hassan Murad Shah  
 Director General  
 Health Services Sindh  
 Hyderabad  
 Chairman


## Bid Evaluation Report

1. Name of Procuring Agency: Program Manager, Hepatitis Prevention & Control Program Sindh (Chief Misnister's Imitative) Hyderabad.
2. Tender Reference No: PM/HPCPS/HYD/(Tender 15-16)/-814/22 Dated:31.3.2016.
3. Tender Description/Name of work/item: **Machinery & Equipments.**
4. Method of Procurement: Single stage Two Envelops Bidding Procedure.
5. Tender Published: leading Daily News papers Daily Express dated: 2.4.2016, Daily Dawn dated: 2.4.2016  
Daily Kawish dated: 6.4.2016, SPPRA ID NO. S. No.27213.  
*Print & Electronic Media (SPPRA ID No. & News papers names with dates)*
6. Total Bid documents Sold; 17 Nos.
7. Total Bids Received: 09 Nos.
8. Technical Bid Opening date: (if applicable) 19.4.2016 (*Provide details in separate form*)
9. No. of Bid technically qualified (if applicable): 06 Nos.
10. Bid(s) Rejected: Nos. disqualified 03.
11. Financial Bid Opening date: 25.5.2016.


### 12. Bid Evaluation Report:

| <i>S No</i> | <i>Name of Firm or Bidder</i>            | <i>Cost offered by the Bidder</i> | <i>Ranking in terms of cost</i> | <i>Comparison with Estimated cost</i> | <i>Reasons for acceptance/rejection</i> | <i>Remarks</i>   |
|-------------|--|-----------------------------------|---------------------------------|---------------------------------------|---|--|
| 0           | 1  | 2                                 | 3                               | 4                                     | 5                                       | 6  |
| 1           | M/s Global Marketing Services Rawalpindi |                                   |                                 |                                       |   | Technical Evaluation report, Comparative statement duly signed by all members of Technical Committee & Procurement committee members respectively is Attached. |
| 2           | M/s Sindh Medical Stores Karachi         |                                   |                                 |                                       |   |  |
| 3           | M/s MJ Trading Corp Karachi              |                                   |                                 |                                       |   |  |
| 4           | M/s Scien Tech Karachi                   |                                   |                                 |                                       |   |  |

|    |                                    |   |
|----|------------------------------------|---|
| 5. | M/s Popular International Pvt Hyd  | <p>Technical Evaluation report, Comparative statement duly signed by all members of Technical Committee &amp; Procurement committee members respectively is Attached.</p> |
| 6. | M/s Muslim Trading Agency Karachi  |   |
| 7  | M/s Lab Link Enterprises Hyderabad |   |
| 8  | M/s Paras Interprises Karachi      |   |
| 9  | M/s RTJ Enterprises Karachi        |   |



Representative from District Accounts Officer  
Hyderabad.  
Member



Deputy Secretary PM&I  
Govt: of Sindh, Health Department Karachi  
Member




Dr. Bikha Ram Daverjani  
Prof: of Medicine/Hepatologist  
LUMHS Jamshoro  
Member



Representative from Commissioner  
Hyderabad Division  
Member



Dr. Ikramuddin Ujjan  
Prof: of Pathology  
LUMHS Jamshoro /Hyderabad  
Member



Dr. Abdul Khalique Shaikh  
Program Manager, Hepatitis Prevention & Control  
Program  
(Chief Minister's Initiative ) Sindh Hyderabad  
Secretary /Member



Dr. Hassan Murad Shah  
Director General  
Health Services Sindh Hyderabad  
Chairman