




**MINUTES OF MEETING, BID EVALUATION REPORT AND
COMPARATIVE STATEMENTS FOR
PURCHASE OF MACHINERY / EQUIPMENT FOR
“UP-GRADATION OF NEONATAL & PEADIATRIC SERVICES”
LIAQUAT UNIVERSITY HOSPITAL, HYDERABAD**

FOR THE FINANCIAL YEAR 2017-18

TENDER OPENED ON 07.02.2018

**LIAQUAT UNIVERSITY HOSPITAL
HYDERABAD/AMSHORC**



**MINUTES OF MEETING, BID EVALUATION REPORT AND
COMPARATIVE STATEMENTS FOR
PURCHASE OF MACHINERY / EQUIPMENT FOR
“UP-GRADATION OF NEONATAL & PEADIATRIC SERVICES”
LIAQUAT UNIVERSITY HOSPITAL, HYDERABAD**

FOR THE FINANCIAL YEAR 2017-18

TENDER OPENED ON 07.02.2018

**LIAQUAT UNIVERSITY HOSPITAL
HYDERABAD - MISHORO**




**LIAQUAT UNIVERSITY HOSPITAL
HYDERABAD / JAMSHORO**

**BID
EVALUATION REPORT**

**PURCHASE OF MACHINERY /EQUIPMENT
UP-GRADATION OF NEONATAL & PEADIATRIC SERVICES
FOR THE FINANCIAL YEAR 2017-18**

TENDER INQUIRY NO.33/2017-18

OPENED ON 07.02.2018



**SINDH PUBLIC
PROCUREMENT
REGULATORY AUTHORITY
GOVERNMENT OF SINDH**

HEAD - SPPRA HEAD - PROCUREMENT - S.M.S. SPPRA QUERA SPPRA CT 2009 SPPRA CT S NOTIFICATION - TRAINING POLICY LETTERS



Serial No.	SPPRA ID	Department	Advertisement Date	Closing Date	Upload Date	City	Tender Notice	BER Date	BERS Date	CER Date	CEN
36313		LAQUATU									

Records per page 50

Columns (SAVE) Export Import Filter

36313 2147483647 2018-01-19 00:00:00 2015-02-07 19-01-2016 Jamshoro

Quaquet University Hospital

23 59 58

هڪ ئي وقت ڪراچي، حيدرآباد ۽ سکر مان شايع ٿيندڙ پهرين سنڌي اخبار

Ph: 022-9210207 SAY NO TO CORRUPTION Fax: 022-9210208

ٽيندر نوٽيس

1. مالي سال 2017-18ء لياقت يونيورسٽي هاسپيٽل، حيدرآباد/چامشورو پر نيونيٽل ۽ پيداوارڪ سروسز جي اهرنگيشن واسطي ايشيو مينٽ/مشنري جي خريداري لاءِ پاڪستان ۾ سٺي شهرت رکندڙ فرمز، دلچسپي رکندڙ ايجنٽ/سول ايجنٽ/فارين ڪمپنيز ڊسٽريبيوٽرز هتي تجويز ڪيل پروفارماتي SPPRA رولز 2010 (ترميم ٿيل 2013) هيٺ مهربند آيون ڪهرائجن ٿيون. ٽيندر دستاويزات امانت جي آخري تاريخ 7 فيبروري 2018 صبح 11:30 وڳي تائين ۽ ساڳئي ڏينهن منجهند 12:00 وڳي شرڪت ڪندڙن جي موجودگي پر ڪوليا ويندا.

2.	ٽيندر انگوائري نمبر.	33/2017-18
	ٽيندر ۽ پيداوارڪ سروسز جي اپ گريڊيشن لاءِ مشنري/ايشيو مينٽ جي خريداري	
	واڪ جو طريقو/ڪار	سٺل اسٽيج نو اينوئيٽ
	واڪ سيڪيورٽي	ڪل واڪ رقم جو 2.5%
	واڪ جو ڪارگر هنڌ	90 ڏينهن
	ٽيندر في	2000 روپيا هر هڪ (ناقابل واپسي)
	ٽيندر ڪاغذن جو اجراء	20-01-2018 کان 06-02-2018 تائين ايت LUH، حيدرآباد
	ٽيندرن جو امانت	07-02-2018 صبح 11:30 وڳي تائين ايت LUH، حيدرآباد
	ٽيندر جو ڪولڻ	07-02-2018 منجهند 12:00 وڳي ايت LUH، حيدرآباد

- ٽيندر ڪاغذن جو مڪمل سيٽ گهرجن جي شيڊيول تي مشتمل، نيڪيٽل/اسيسينفڪيشنز شرط ۽ ضابطن سان گڏ هيٺ صحيح ڪندڙ لياقت يونيورسٽي اسپتال (LUH)، حيدرآباد جي آفيس مان 2000 روپين جي ناقابل واپسي في ڏين تي خريد ڪري سگهجي ٿو.
- ميڊيڪل سپرنٽينڊنٽ، LUH حيدرآباد/چامشورو SPPRA رولز مطابق شيڊيولس جي ڪنهن به آئٽم جي مقدارن پر جڏهن به ۽ جيئن به مناسب سمجهين، اضافو يا گهٽتائي ڪرڻ جو حق محفوظ رکي ٿو. پروفارماتي ايجنسي SPPRA رولز جي واسطيدار فٽرن جي شرط سان سمورا يا ڪي به واڪ رد ڪري سگهي ٿي.
- ٽيندر ڪاغذ SPPRA ويب سائٽ تان ڊائون لوڊ ڪري سگهجن ٿا.
- واڪ پر سمورا سرڪاري ٽيڪس شامل هوندا (جيڪڏهن لاڳو ڪرڻ جو ڳا هوندا)
- بلن جي ڪٽوٽي سرڪاري قاعدن پٽائڻ ڪئي ويندي.
- واڪ ڏيندڙ پاران ڄاڻايل آئٽمن جي جملي واڪ قيمت جو 2.5% واڪ سيڪيورٽي پي آرڊر/ڊمانڊ ڊرافٽ جي صورت ۾ بحق ميڊيڪل سپرنٽينڊنٽ، لياقت يونيورسٽي اسپتال، حيدرآباد/چامشورو پيش ڪئي ويندي، جڏهن ته SPPRA رولز مطابق نيڪيٽل ڪان نيڪي جي رقم 5 سيڪڙو پروفارماتي سيڪيورٽي ورتي ويندي.
- ٽيندر نوٽيس ۽ ٽيندر ڪاغذن جي وچ ۾ مونجهارن جي صورت ۾ ٽيندر ڪاغذ سٺ حاصل ڪندا.
- عامر موٽل جي اعلان يا ڪنهن تائيدي حالت جي صورت ۾ واڪ ورنڌڙ ڪم واري ڏينهن تي ڏنل پروگرام مطابق امانتيا ۽ ڪوليا ويندا.
- پيڙاڳاڳا حنين نمبرن تان يا وڊيو آفيس وقت دوران LUH حيدرآباد مان ڪري سگهجي ٿي.

فون: 022-9210207، فيڪس: 022-9210208.

ميڊيڪل سپرنٽينڊنٽ
لياقت يونيورسٽي هاسپيٽل
حيدرآباد / چامشورو



ٹینڈرنوٹس

(1) ایس بی بی آر اے روز 2010 (ترمیم شدہ 2013) کے مطابق مالی سال 2017-18 کیلئے "ایپ گریڈیشن آف نیونائل اینڈ پیڈیاٹریک سروسز بمقام لیاقت یونیورسٹی اسپتال حیدرآباد/جامشورو" کی خریداری کیلئے خوب مشہور و معروف فرموں/خواہشمند ایجنٹ/سول ایجنٹ/غیر ملکی کمپنیوں/پاکستان میں ان کے ڈسٹری بیوٹرز سے مقررہ پروفارمہ پر ذریعہ بذاسر ممبر ٹینڈر مطلوب ہیں۔ ٹینڈر دستاویزات جمع کرانے کی آخری تاریخ 07 فروری 2018ء دن کے 11-30 بجے تک ہے اور یہ اسی روز شرکت کنندگان کی موجودگی میں دوپہر 12-00 بجے کھولے جائیں گے۔

(2)

33/2017-18	ٹینڈر انکوائری نمبر
	پیشکش کی
	صراحت
	بڈنگ پر دست
	بڈ سیکورٹی
	پیشکش کے کارآمد
	رہنے کی مدت
	ٹینڈر فیس
	ٹینڈر دستاویزات
	کا اجراء
	ٹینڈر جمع کرانے کی تاریخ
	ٹینڈر کھلنے کی تاریخ

(3) شرائط کا شیڈول، ٹیکنیکل تصدیقات مع شرائط و ضوابط پر مشتمل ٹینڈر دستاویزات کا سیت مبلغ 2000/- روپے ناقابل واپسی فیس کے عوض دفتر زیر دستگی لیاقت یونیورسٹی اسپتال (L.U.H) حیدرآباد سے خریدا جاسکتا ہے۔

(4) میڈیکل سپرنٹنڈنٹ L.U.H حیدرآباد/جامشورو کا یہ حق محفوظ ہے کہ وہ SPPRA روز کے مطابق جب بھی اور جہاں بھی ضروری سمجھے شیڈول کے مطابق کسی بھی اسٹیم کی مقدار/تعداد میں کمی/تعمیر کر سکتے ہیں۔ پروڈیورنگ ایجنسی SPPRA روز کے متعلقہ مندرجات سے شروط کسی یا جملہ پیشکشوں کو مسترد کر سکتی ہے۔

(5) ٹینڈر دستاویزات SPPRA کی ویب سائٹ سے ڈاؤن لوڈ کی جاسکتی ہیں۔

(6) پیشکشوں میں جملہ سرکاری ٹیکسز (اگر اطلاق ہوتا ہو) شامل ہونے چاہئیں۔

(7) بلوں سے کوئی سرکاری قواعد کے مطابق کی جائے گی۔

(8) بڈ سیکورٹی درج کردہ اسٹیم کی کل مالیت کا 2.5% بشکل پی آر ڈر/ڈیمانڈ ڈرافٹ بحق میڈیکل سپرنٹنڈنٹ لیاقت یونیورسٹی اسپتال حیدرآباد/جامشورو ہونی چاہئے۔ جبکہ کنٹریکٹ کی رقم کا 5% بطور بڈ سیکورٹی بمطابق ایس بی بی آر اے روز کنٹریکٹ سے حاصل کی جائے گی۔

(9) ٹینڈر نوٹس اور ٹینڈر دستاویزات کے درمیان کسی تضاد کی صورت میں ٹینڈر دستاویزات قابل عمل سمجھی جائیں گی۔

(10) کسی عام تعطیل کے اعلان یا دیگر نامساعد حالات کی صورت میں یہ پیشکشیں شیڈول کے مطابق کام کے آئندہ روز جمع کرانی اور کھولی جائیں گی۔

(11) استفسارات نیچے دیئے گئے پتہ پر درج ذیل نمبروں پر اوقات کار کے دوران بمقام L.U.H حیدرآباد کے جاسکتے ہیں۔ فون: 022-9210207، 022-9210208۔

(دستخط) ... میڈیکل سپرنٹنڈنٹ

لیاقت یونیورسٹی اسپتال حیدرآباد/جامشورو (INF/KRY-353/18)

Ph: 022-9210207

TENDER NOTICE

Fax: 022-9210208

1. According to SPPRA Rules 2010 (Amendment-2013) sealed tenders on prescribed proforma are hereby invited from well reputed firms, interested Agent / Sole Agent / Foreign Companies Distributors in Pakistan for the purchase of Machinery / Equipment for "Up-graduation of Neonatal & Paediatric Services at Liaquat University Hospital, Hyderabad / Jamshoro" for the financial year 2017-18. The last date for submission of tender documents is 7th February, 2018 upto 11:30 AM and will be opened on same day at 12:00 noon in the presence of participants.

Tender Enquiry No.	33/2017-18
	Purchase of Machinery/Equipment for "Up-Graduation of Neonatal & Paediatric Services"
Bidding Procedure	Single Stage Two Envelope
Bid Security	2.5% of the total bid value
Bid Validity	90 days
Tender Fee	Rs. 2,000/- each (non-refundable)
Issuance of Tender Documents	20.01.2018 to 06.02.2018 @ LUH, Hyderabad
Submission of Tender	07.02.2018 at 11:30 A.M. @ LUH, Hyderabad
Opening of Tender	07.02.2018 at 12:00 Noon @ LUH, Hyderabad

- 3) Complete set of tender documents containing Schedule of Requirements, Technical Specification with Terms & Conditions can be purchased with a non-refundable fee of Rs. 2,000/- cash from the office of undersigned Liaquat University Hospital (LUH), Hyderabad.
- 4) The Medical superintendent LUH, Hyderabad / Jamshoro reserves the right to increase or decrease the quantity of any item of schedules as and when it is deemed necessary according to SPPRA Rules. The Procuring agency may reject all or any bid subject to the relevant provision of SPPRA Rules.
- 5) Tender documents can be downloaded from SPPRA website.
- 6) Bid shall include all government taxes (if applicable).
- 7) Deduction in the bills will be done as per government rules.
- 8) Bid security @2.5% of the total bid value of quoted items is to be submitted by the bidders in shape of pay order / demand draft in favour of Medical Superintendent Liaquat University Hospital Hyderabad / Jamshoro whereas an amount of the contract 5% performance security is to be obtained from contractor as per SPPRA Rules.
- 9) In case of discrepancies between the Tender Notice and the Tender Documents, the Tender Documents shall take precedence.
- 10) In case of announcement of public holiday or any unfavourable circumstances, the bids shall be submitted and opened as per given schedule on the next working day.
- 11) Queries can be addressed at the following numbers or in person during office hours @ LUH, Hyderabad.

INF-KRY: No. 353/18

Say No to Corruption

ہم شہزادی کے خلاف تھہریں۔

Medical Superintendent
Liaquat University Hospital
Hyderabad/Jamshoro



NOTIFICATION

NO.HD (P&E) 1-2 (61) / 2017. In supersession of this department's notification No.HD (P&E) 3-5 (77) / 2017, dated 16.03.2017, NO.HD (P&E) 3-5 (74) / 2013, dated 03.08.2017, NO.HD (P&E) 3-5 (85) / 2017, dated 03.08.2017, NO.HD (P&E) 3-5 (78) / 2013, dated 11.09.2017, NO.HD (P&E) 3-5 (55) / 2015, dated 11.09.2017 and SO (PM&I) 2-1 / 2017-18 / PM-05 (LUH), dated 07.08.2017, the Government of Sindh, Health Department is pleased to re constitute a Procurement Committee under Clause-7 of Sindh Public Procurement Rules 2010 (Amended 2017), comprising of the following members for purchase of machinery / equipment, vehicles and furniture / fixture to be purchased under revenue component of "ADP schemes & SNE" in respect of Liaquat University Hospital Hyderabad;

- | | |
|---|-----------------|
| 1. Additional Medical Superintendent (General),
Liaquat University Hospital, Hyderabad. | Chairman |
| 2. Additional Medical Superintendent (Technical)
Liaquat University Hospital, Hyderabad. | Member |
| 3. Professor / Associate Professor / Assistant
Professor / Expert / Technical Person of Relevant field,
Liaquat University of Medical & Health Sciences (LUM&HS)
Jamshoro. | Member |

The committee will responsible to observe all the codal formalities as per Sindh Public Procurement Rules 2010 (Amended 2017), the TORS / Functions / Responsibilities of procurement committee in accordance with SPP Rule-8 (Amended 2017) are as under;

- Preparing bidding documents;
- Carrying out technical as well as financial evaluation of the bids
- Preparing evaluation report as provided in rule-45,
- Making recommendations for the award of contract to the competent authority;
- Perform any other function ancillary and incidental to the above.

(DR.FAZLULLAH PECHUHO)
SECRETARY HEALTH


NO.HD (P&E) 1-2 (61) / 2017

Karachi, dated 07-11-2017.

C.c.to:

- The Vice Chancellor, Liaquat University of Medical and Health Sciences Hyderabad / Jamshoro with request to advice the member to attend the meetings for above said committee.
- The Managing Director, Sindh Public Procurement Regularity Authority, Karachi.
- The Medical Superintendent, Liaquat University Hospital Hyderabad / Jamshoro.
- The District Accounts Officer, Hyderabad.
- All members of the committee.
- P.S. to Secretary, Health Department, Govt. of Sindh, Karachi.
- P.S to Additional Secretary (Dev), Health Department, Govt. of Sindh, Karachi.


Deputy Director (Dev)



C.814

Gen. Sec.



NO.HD(P&E) 3-2 (427)/2014
GOVERNMENT OF SINDH
HEALTH DEPARTMENT
(Procurement, Monitoring and Inspection Cell)
Karachi dated the 10th APRIL, 2017.

NOTIFICATION

No.HD(P&E)3-2(427)/2014. In supersession to this department's notification of even number dated 9-03-2017 and in pursuance of Rule 31 of the Sindh Public Procurement Rules, 2010, a Departmental Complaint Redressal Committee, comprising of following officers is hereby re-constituted as under to resolve complaints of aggrieved bidders:

- | | |
|---|--------------------------|
| 1. Special Secretary,
Health Department. | 1914 25-4-17
Chairman |
| 2. Representative from Accountant General Sindh, | Member |
| 3. Independent expert from relevant field concerning
(to be nominated by the Head of Procuring Agency) | Member |
| 4. Deputy Secretary (PM&I) | Member |
| 5. Deputy Secretary (General) | Member/Secretary |

ToRs.

- To perform according to Rule 31 of SPPRA, 2010.
- To Perform any other function ancillary and incidental to the above.

- DR. FAZLULLAH PECHUHO -
SECRETARY TO GOVERNMENT OF Sindh

No.HD(P&E)3-2(427)/2014.

Karachi dated the 10th APRIL, 2017

1. The D.G.Health Services, Sindh, Hyderabad.
2. The District Health Officers (All)
3. Medical Superintendents (All) Liaquat University Hyderabad.
4. All the Member of the Committee.
5. The PS to Secretary Health Department, Govt. of Sindh, Karachi.
6. Office order file.

SECTION OFFICER (PM&I)


Handwritten signature and initials

Handwritten number 274




REVISED ANNUAL PROCUREMENT PLAN
LIAQUAT UNIVERSITY HOSPITAL, HYDERABAD / JAMSHORO
FINANCIAL YEAR 2017-18

S.No.	Description of Procurement	Qty	Estimated Total cost	Funds Allocated	Source of funds (ADP / Non ADP)	Proposed Procurement method	Timing of Procurement				Remarks
							1st Qtr	2nd Qtr	3rd Qtr	4th Qtr	
1	Drugs / Medicines & Surgical Items (L.P)		87.75 M	87.75 M	Non ADP	Single Stage Two Envelope	In all quarters as per Requirement				
2	Dietary Items.		40.95 M	40.95 M	do	do	do				
3	Uniform & Liveries		3.33 M	3.33 M	do	do	do				
4	Misc / Petty / Electric Items.		4.389 M	4.389 M	do	do	do				
5	Medical Gases		29.23 M	29.23 M	do	do	do				
6	Janitorial Services & Items.		9.45 M	9.45 M	do	do	do				
7	Printing Articles.		1.016 M	1.016 M	do	do	do				
8	Stationary Articles		0.604 M	0.604 M	do	do	do				
9	Instruments		35.464 M	35.464 M	do	do	do				
10	Consumable Items / Pathology Kits & Chemicals.		12.183 M	12.183 M	do	do	do				
11	Deployment of Security Guards		13.2 M	13.2 M			do				
12	Repair of Machinery / Equipment		37.077 M	37.077 M	do	do	do				
13	Repair of Furniture / Fixture.		2.04 M	2.04 M	do	do	do				
14	Drugs/Medicine under Zakat fund (Bulk)		42.0 M	42.0 M	do	do	do				
15	Drugs/Medicine under Zakat fund (LP)		28.0 M		do	do	do				
16	Repair of Transport		4.692 M	4.692 M	do	do	do				
17	Machinery Equipment under SNE		374.14 M	374.14 M	do	do	do				
I	Purchase of C.T Scan Machine		70.0 M		do	do	do				
II	Purchase of Dialysis Machine		6.00 M		do	do	do				
III	Purchase of Lithotripsy Machine		70.0 M	374.14 M	do	do	do				
IV	Purchase of M.R.I Machine		125.0 M		do	do	do				
V	Purchase of Laundry Machine		4.00 M		do	do	do				
18	Purchase of Machiner / Equipment for Paeds		99.140 M		do	do	do				
19	Purchase of Machinery from Welfare fund		5.00 M	5.00 M	do	do	do				
20	Furniture / Fixture under SNE		58.925 M	58.925 M	do	do	do				
21	Purchase of Transport		39.0 M	39.0 M	do	do	do				
22	Purchase of Machinery /Equipment under various ADP/Non ADP Scheme & Rescue Center.		400.00 M		ADP	do	do				

List Enclosed


 Medical Superintendent
 Liaquat University Hospital

**MEETING OF PROCUREMENT COMMITTEE FOR OPENING OF TECHNICAL BIDS
FOR PURCHASE OF MACHINERY / EQUIPMENT (UP-GRADATION OF NEONATAL &
PAEDIATRIC SERVICES) UNDER ADP SCHEME FOR THE FINANCIAL YEAR 2017-18
IN THE COMMITTEE ROOM LUH, HYDERABAD ON 07.02.2018.**

S.NO	NAME & DESIGNATION		SIGNATURE
01	Dr. Naeem Zia Memon Addl: Medical Superintendent, LUH, Hyderabad.	Chairman	
02	Dr. Yasmeen Memon Chairperson Paediatric Department LUMHS, Jamshoro	Member	
03	Dr. Anjum Rehman Deputy Medical Superintendent, L.U. Hospital, Hyderabad.	Member	

Opening of Technical Bids for Purchase of Machinery / Equipment under ADP Scheme "Upgradation of Neonatal & Paediatric Services" for the Financial year 2017-18 at Liaquat University Hospital, Hyderabad/Jamshoro

Dated:- 07.02.2018

Tender Opened on 07.02.2018

ATTENDANCE SLIP

S.No.	Name of Firm	Phone Number	Postal Address	Name of Participant	C.N.I.C No. of participants	Signature
01	CARE HEALTH SOLUTIONS	0380-2233 062	303 3rd Floor, Mafiqe City Mall, Karimnagar	SARFRAZ AHMED	42161-1728261-3	
02	Bio Medical Technologies	0333-1333 402	407, 4th Floor, Mafiqe City Mall, Karimnagar	Haseem Ahsan	42561-6145569-9	
03	Mediquest SMC LLC	0334-262291	Banglow D-49 Mansarovar Step LLC	Jamir Rajput	41303-5534881-9	
04	BICOS	0333-2724054	Building No 2C-12 Commercial OHA Phase II KHAS	Muhammed Farooq	41303-8949622-9	
05	U.S. Enterprises	0333-2222222	Phase II OHA KHAS	Abdul Wahid	41303-8949622-9	
06	E-rotec (PVT) Ltd,	03219235662	Koraceli	Ahsanul Baryq	42901-4151007-9	
07	Khalid Corporation	0300-2507696	Koraceli	Ishaq	0300-2507696	
08	Universal Enterprises	033192665510	Koraceli	Abulhasan Vaidh	41303-8949622-9	

(Dr. Naeem Zia Memon)

Addl: Medical Superintendent (General)
LUH, Hyderabad
Chairman Procurement Committee

(Dr. Anjum Rehman)

Addl: Medical Superintendent (Technical)
LUH, Hyderabad
Member Procurement Committee

(Dr. Yasmeen Memon)

Chairperson Paediatric Department
LUMHS, Jamshoro
Member Procurement Committee

MINUTES OF BID OPENING MEETING


NIT No.	Inf No:353/18
SPPRA Number	S.No: 36313
Method of Procurement	Single Stage Two Envelope

Name of Work: **Purchase of Machinery / Equipment Up-Gradation of Neonatal & Paediatric Services for the Financial Year 2017-18**


Meeting of the Procurement Committee (PC) was held on Wednesday, 07 February 2018 for opening of above mentioned NIT. Meeting was attended by members of the Procurement Committee and the representatives of bidders (attendance sheets are enclosed).

The following companies participated in the bidding process with their respective bids by submission deadline of 11:30am on Wednesday 07.02.2018.


- 1) M/s BIOS
- 2) M/s Medequips SMC
- 3) M/s Elate C.C
- 4) M/s Bio Medical Technology
- 5) M/s Care Health Solution
- 6) M/s Universal Enterprise
- 7) M/s Al-Farooq Enterprises
- 8) M/s Universal Systematic Solution
- 9) M/s Khalid Corporation



(Chairman)
Dr. Naeem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad



(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Jamshoro






(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

**PROCUREMENT COMMITTEE
ATTENDANCE SHEET**

Tender No. 33/2017-18

Name of Work: Purchase of Machinery / Equipment Up-Gradation of Neonatal & Paediatric Services for the Financial Year 2017-18

The following members of the Procurement Committee (PC) were present on Wednesday, 07 February 2018 for opening of above mentioned NIT.

1.	Dr. Naeem Zia Memon	Addl: Medical Superintendent (Gen) Liaquat University Hospital Hyderabad	Chairman	
2.	Dr. Yasmin Memon	Chairperson of Paediatric Department LUMHS. Jamshoro	Member	
3.	Dr. Anjum Rehman	Addl: Medical Superintendent (I) Liaquat University Hospital Hyderabad	Member	

S. No.	Parameter	Max Points	M/S BIOS	M/S Medequips SMC	M/S Elite C.C	M/S Bio Medical Technology	M/S Care Health Solution	M/S Universal Enterprise	M/S Al-Farooq Enterprises	M/S Universal Systematic Solution	M/S Kha Corporate
6	b) Qualified & Trained Personnel (BE/BSc/DAB) with Bio-Data	3	3	3	3	3	3	3	0	3	0
	Original Equipment Manufacturer (OEM) or Brand Quoted	30									
	a) Manufacturer's Authorization	5	5	5	5	5	5	5	5	5	5
	b) Manufacturer's Website	2	2	2	2	2	2	2	2	2	2
	c) Brochure or Catalogue (original)	2	2	2	2	2	2	2	2	2	2
	d) ISO 9001 Certificate	2	2	2	2	2	2	2	2	2	2
	e) ISO 13485 Certificate	2	2	2	2	2	2	2	0	0	0
	f) CE	2	2	2	2	2	2	2	2	0	0
	g) FDA	4	-	-	-	-	-	-	-	-	-
	h) JIS	2	-	-	-	-	-	-	-	-	-
i) References of OEM brand / model											
(i) Government / Semi-Government Teaching Hospital / Institute	3	3	3	3	3	3	3	3	3	3	
(ii) Armed Forces / NGOs Hospital / Institute	2	2	2	2	2	-	-	-	-	-	
(iii) Private Renowned Teaching Hospital / Institute	2	2	2	2	0	2	2	2	-	-	
(iv) International Hospital / Institute	2	-	-	-	-	-	-	-	-	-	
		100	85	85	77	73	78	77	74	70	71

(Chairman)
Dr. Naem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

Tender No. 33/2017-18

Purchase of Machinery / Equipment under SNE (New) 2017-18 for PAEDS Department

Opened on 07.02.2018

BIDDERS

S. No.	Item Description	Qty	M/s BIOS	M/s Medequips SMC	M/s Elate C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universal Enterprise	M/s Al-Farooq Enterprises	M/s Universal Systematic Solution	M/s KTI Corpor
1	Installation of Central Oxygen System (for 60 Bed Ward)	1	-	-	Yes	Yes	Yes	Yes	-	-	-
2	Installation of Central Oxygen System (for 46 Bed Ward)	1	-	-	Yes	Yes	Yes	Yes	-	-	-
3	Paediatric Resuscitation Trolley	4	Yes	Yes	-	-	-	-	-	-	Yes
4	Sphygmomanometer	6	-	-	-	-	-	-	-	-	Yes
5	Heavy Duty Nebulizer	8	-	-	-	Yes	-	-	Yes	-	Yes
6	Split Air Conditioner 1.5 Ton with Installation	14	-	-	-	-	-	-	Yes	-	Yes
7	Digital Photostat Machine	2	-	-	-	-	-	-	Yes	-	Yes
8	Paediatric Ambu Bag	6	-	-	-	-	-	-	-	-	Yes
9	Digital Baby Weight Scale	2	-	-	-	-	-	-	Yes	-	Yes
10	Dual Core Computer with LED Monitor	2	-	-	-	-	-	-	-	Yes	Yes
11	Laser Printer Heavy Duty	2	-	-	-	-	-	-	-	Yes	Yes
12	Electronic Patient Numbering System	2	-	-	-	-	-	-	Yes	-	Yes
13	Sound System with Mic	1	-	-	-	-	-	-	Yes	-	Yes
14	CCTV Security System (PAEDS-I)	1	-	-	-	-	-	-	Yes	Yes	Yes
15	CCTV Security System (PAEDS-II)	1	-	-	-	-	-	-	Yes	Yes	Yes
16	BP Apparatus Mercury Stand	18	-	-	-	-	-	-	Yes	-	Yes

(Chairman)

Dr. Naeem Zia Memon

Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad

(Member)

Dr. Yasmin Memon

Chairperson of Paediatric Department
LUMHS
Jamshoro

(Member)

Dr. Anjum Rehman

Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

Tender No. 33/2017-18
 Purchase of Machinery / Equipment under SNE (New) 2017-18 for PALEDS Department
 Opened on 06.01.2018

BIDDERS TECHNICAL EVALUATION

S. No.	Item Description	Qty	M/s Elite C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universal Enterprise
01	INSTALLATION OF CENTRAL OXYGEN SYSTEM FOR 60 BEDED PEADS WARD, LUH, JAMSHORO	01 No	Yes	Yes	Yes	Yes
	a) COPPER PIPE Suitable for use with medical gases and cleaned degreased with and caps along with required fittings. Completely fitting with brazing and painting Size 1/2" - 1200 Rft. 3/4" - 1000 Rft.		Yes	Yes	Yes	Yes
	b) OUTLET POINTS Medical Gases Outlet Points. Double lock. Quick connect and gas specific. Surface Type: wall - mounted In accordance with the Healthcare Technical Memorandum HTM 2022 ● Oxygen Wall Outlet Point 60 Nos. ● Air Wall Outlet Point 80 Nos. (U.K/U.S.A/JAPAN)		Yes Yes Yes	Yes Yes Yes	Yes Yes Yes	Yes Yes Yes
	c) ISOLATION VALVES Isolation Valves with adaptor. Size 1/2" - 06 Nos. 3/4" - 04 Nos. (U.K/U.S.A/JAPAN)		GCE, UK Yes Yes	Pneumattech, UK Yes Yes	Pneumattech, UK Yes Yes	Novair, France Yes Yes
	d) ZONE SERVICE UNIT Zone Service Unit with Adaptor 04 Nos		(No brand), Italy Yes Yes	Kitz, Japan Yes Yes	Kitz, Japan Yes Yes	(No brand), Japan Yes Yes

Sum

Up

Am

S. No.	Item Description	Qty	M/s Elite C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universe Enterprise	
e)	(Local) OXYGEN FLOWMETER Oxygen Flowmeter0 – 15 L.PM60 Nos. Complete with Humidifier Bottle(U.K) and probe (UK/U.S./JAPAN)		Yes Local	Yes Local	Yes Local	Yes Local	
f)	MEDICAL GAS ALARM Supplying and installation of alarm system of different gases with mute / test facility. gives sound and flash light whenever the pressure in line decreases the set pressure for individual gas i.e. <ul style="list-style-type: none"> • Oxygen Failure Warning03 No. • Air Failure Warning03 No. (Local)		Yes Yes Local	Yes Yes Local	Yes Yes Local	Yes Yes Local	
g)	AIR COMPRESSOR SYSTEM (DUPLX)01 No. <ul style="list-style-type: none"> • Duplex System • 100% Oil free Medical Air Compressor complying with HTM 02-01 standards • 05HP Air Compressor skid mounted • One unit for regular use and another unit for standby Working. • Pre Filter (2 Pcs) • Post Filter (2 Pcs) • Compressed Air Tank 1000 Liter capacity. • Refrigerated Dryers (2 units) • Electrical Panel for the above Compressor system • By Pass system for the above Compressor System (U.K/U.S./JAPAN)		Yes No details By manufacturer	Yes Yes Not oil free	Yes Yes Yes	Yes Yes Yes	Yes No details By manufacturer
h)	OXYGEN SUPPLY MANIFOLD 01 No. The Manifold is a to switch from bank in use to reserve Bank without fluctuation in delivery supply line pressure and without the need for external power. After the switch-over the "reserve" bank becomes the "bank in use" and the bank in use becomes the reserve bank. The system is able maintain continuous supply 1500LPM Capacity		Yes Hitachi, Japan	Yes Apec, Taiwan	Yes Hitachi, Japan	Yes No brand mention	

Sum

Mosam

Mosam

S. No.	Item Description	Qty	M/s Elate C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universe Enterprise
	(2 x 10) with capacity of 180 M3/H set of connection for cylinders/Hp. Valves, discharging valves, tell pipe Make: Local with Imported parts		Yes	Yes	Yes	Yes
j)	BED HEAD PANEL HORIZONTAL FOR 60 BEDS Aluminum powder coated Bed Head Panel Vertical (Heavy Duty) with 6 electric sockets, each bed For 60 beds complete with Oxygen (UK) 02 Vacuum (UK) 01 Air 4 bar(UK) 01 On each bed drip stand, Infusion Stand Accessories Rail Accessories Basket Spot Light (Locally Made with Extended Aluminum)		Yes	Yes	Yes	Yes
k)	Mobile Suction Units Wall mounted with 2 Liter Jar 20 Nos. (U.K/U.S.A/JAPAN/ECC) Installation Testing, commissioning purging, painting inclusive of gases, welding rods etc. OR equivalent		Yes GCE, UK	Yes Pneumatech, UK	Yes Pneumatech, UK	Yes Novair, France
<p>Note: Samples of Bedhead Panel/Flowmeter to be provided on demand.</p>						
	Remarks		- Technical offer and submitted technical documents comply with the required technical specifications.	- Air compressor not as per required specification - Offered compressor is of Taiwan origin. - Technical offer says 'APEC' and	- Technical offer and submitted technical documents comply with the required technical specifications.	- Technical offer submitted technical documents comply with the required technical specifications.

Signature

Signature

Signature

S. No.	Item Description	Qty	M/s Elite C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universal Enterprise
			Responsive	M/s Bio Medical Technology brochure submitted are of -PUMA - Authorization letter from Pneumattech. UK is general and not as per required format of the tender documents.	Responsive	Responsive

Seni

Yasmin

Anjum

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Elare C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universal Enterprise
02	INSTALLATION OF CENTRAL OXYGEN SYSTEM FOR 46 BEDED PEADS WARD, LUH, HYDERABAD)	01 No	Yes	Yes	Yes	Yes
	a) COPPER PIPE Suitable for use with Medical Gases and cleaned Degreased with and caps along with required fittings Completely fitting with brazing and painting Size 1/2" - 1500 Rft. 3/4" - 700 Rft.		Yes	Yes	Yes	Yes
	b) OUTLET POINTS Medical Gases Outlet Points, Double lock. Quick connect and gas specific. Surface Type, wall mounted In accordance with the Healthcare Technical Memorandum HTM 2022 ● Oxygen Wall Outlet Point 46 Nos. ● Air Wall Outlet Point 92 Nos. (U.K/U.S.A/JAPAN)		Yes Yes Yes	Yes Yes Yes	Yes Yes Yes	Yes Yes Yes
	c) ISOLATION VALVES Isolation Valves with adaptor. Size 1/2" - 12 Nos. 3/4" - 04 Nos. (U.K/U.S.A/JAPAN)		GCE, UK Yes Yes	Pneumatech, UK Yes Yes	Pneumatech, UK Yes Yes	Novair, France Yes Yes
	d) ZONE SERVICE UNIT Zone Service Unit with Adaptor (Local) 06 Nos		(No brand), Italy Yes Yes	Kitz, Japan Yes Yes	Kitz, Japan Yes Yes	(No brand), Japan Yes Yes
	e) OXYGEN FLOWMETER Oxygen Flowmeter O - 15 LPM 36 Nos Complete with Humidifier Bottle(U.K) and probe (UK/U.S.A/JAPAN)		GCE, UK Yes Yes	Pneumatech, UK Yes Yes	Pneumatech, UK Yes Yes	Novair, France Yes Yes

Handwritten signature/initials

Handwritten signature/initials

Handwritten signature/initials

S. No.	Item Description	Qty	M/s/Elate C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Unvers Enterprise
f)	<p>MEDICAL GAS ALARM Supplying and installation of alarm system of different gases with mute / test facility. gives sound and flash light whenever the pressure in line decreases the set pressure for individual gas i.e.</p> <ul style="list-style-type: none"> ● Oxygen Failure Warning 01 No. ● Air Failure Warning 01 No. <p>(Local)</p>		Yes Yes Local	Yes Yes Local	Yes Yes Local	Yes Yes Local
g)	<p>AIR COMPRESSOR SYSTEM (DUPLIX) 01 No.</p> <ul style="list-style-type: none"> ● Duplex System ● 100% Oil free Medical Air Compressor complying with HTTM 02-01 standards ● 05 HP Air Compressor skid mounted <p>One unit for regular use and another unit for standby Working.</p> <ul style="list-style-type: none"> ● Pre Filter (2 Pcs) ● Post Filte (2 Pcs) ● Compressed Air Tank 1000 Liter capacity. ● Refrigerated Dryers (2 units) ● Electrical Panel for the above Compressor system ● By Pass system for the above Compressor System (U.K/U.S.A/JAPAN) 		Yes No details By manufacturer	Yes Yes Not oil free	Yes Yes Yes	Yes No details By manufacturer
h)	<p>OXYGEN SUPPLY MANIFOLD 01 No. The Manifold is a to switch from bank in use to reserve Bank without fluctuation in delivery supply line pressure and without the need for external power. After the switch-over the "reserve" bank becomes the "bank in use" and the bank in use becomes the reserve bank. The system is able maintain continuous supply 1500LPM Capacity (6 x 6) with capacity of 180 M3/H set of connection for cylinders/1lp. Valves, discharging valves, tell pipe Make: Local with Imported parts</p>		Hitachi, Japan Yes	Apec, Taiwan Yes	Hitachi, Japan Yes	No brand mention Yes
i)	<p>BED HEAD PANEL CONTINUOUS HORIZONTAL Aluminum powder coated Bed Head Panel Continuous</p>		Yes Yes Yes	Yes Yes Yes	Yes Yes Yes	Yes Yes Yes

Swam

Hagan

Amur

S. No.	Item Description	Qty	M/s Elate C.C	M/s Bio Medical Technology	M/s Care Health Solution	M/s Universe Enterprise
j) Mobile Suction Units with 2 Liter Jar 1.5 Nos (U.K./U.S.A/JAPAN/ECC) k) Installation Testing, commissioning, purging, painting inclusive of gases, welding rods etc. OR equivalent Note: Samples of Bedhead Panel/Flowmeter to be provided on demand.	400ft (Heavy Duty) With 6 electric sockets, each bed For 46 beds complete with Oxygen (UK) 02 Vacuum (UK)01 Air 4 bar(UK) 01 On each bed drip stand, Infusion Stand Accessories Rail Accessories Basket Spot Light (Locally Made with Extended Aluminum)	Not Continuous GCE, UK Yes GCE, UK Yes	Continuous Yes Yes Yes Yes Yes Yes Yes Yes Local Yes Pneumatech, UK	Continuous Yes Yes Yes Yes Yes Yes Yes Local Yes Pneumatech, UK	Not Continuous - - - No No No No No Local Yes Novair, France	
Remarks	- Technical offer and submitted technical documents comply with the required technical specifications.	- Air compressor not as per required specification - Offered compressor is of Taiwan origin. - Technical offer says 'APEC' and brochure submitted are of 'PUMA'. - Authorization letter from Pneumatech, UK is general and not as per required	- Technical offer and submitted technical documents comply with the required technical specifications.	- Technical offer and submitted technical documents comply with the required technical specifications.		

02/11/14

Vijay

[Signature]

S. No.	Item Description	Qty	M/S Elite C.C	M/S Bio Medical Technology	M/S Care Health Solution	M/S University Enterprise
			Responsive	format of the tender documents.	Responsive	Responsive

QAW

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

Yasmin

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Janshoro

Anjum

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s BIOS	M/s Medequips SMC	M/s Khalid Corporation
03	PAEDIATRIC RESUSCITATION TROLLEY Certification: ISO 9001, ISO 13485, CE Country of Origin: Brazil/Turkey/Korea Specification: • Control panel with large LED display. • Skin probe • Modes: pre-warm, Manual and servo • Controlled mattress heating with delayed alarms during PreWarm Mode • Remote Alarm silence • Bed movements: Trendelenburg & Reverse Trendelenburg • Drop-Down Side panels • Ceramic Heating element for rapid warming • Digital Apgar Timer of 1.5, 10, 20 minutes • Intelligent alarms for patient safety • Examination lamp • X-rays cassette tray • Air Probe • Castor with brackets & brakes • Push Handle for y movement control • Integrated Infant Resuscitator • Medical Suction Unit (Pediatric) • Air&O2 Flowmeter System • Bracket For Cylinders OR equivalent	04 Nos	OKM 730-333 Okuman Yes Turkey	Sunflower Atom Yes Japan	MX37EMG Baifida Taiwan
Remarks					
			Responsive	Responsive	Not Responsive

Sem
 (Chairman)
Dr. Naem Zia Memon
 Addl: Medical Superintendent (Gen)
 L.U.H. Hyderabad

Hassan
 (Member)
Dr. Yasmeen Memon
 Chairperson of Paediatric Department
 L.U.M.H.S. Jamshoro

Manir Akh
 (Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 L.U.H. Hyderabad

S. No.	Item Description	Qty	M/s Khalid Corporation	Responsive
04	<p>SPHYGMOMANOMETERS</p> <p>Certification: ISO 9001</p> <p>Country of Origin: USA/EU/Japan</p> <p>Specification: <ul style="list-style-type: none"> • Cuff Size For Newborn, Infant and Children • Accurate reading. • Emulsoid Bladder & Soft emulsoid Bulb • Mercury lock facility • Measure Scope: 0-300mmHg • Accuracy: 3mmHg • Glass Tube: 3.5 to 4.0mm OR equivalent </p>	06 Nos	600 Yamasu	Yes
			Japan	Yes

Sem

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

Yasmin

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

Anjum

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	QTY	M/s Bio Medical Technology	M/s Khalid Corporation	M/s Al-Farooque Enterprises
05	HEAVY DUTY NEBULIZER Certification: ISO 9001 Country of Origin: USA/EU/Japan Specification: <ul style="list-style-type: none"> Compressor: Piston, Fuse: T1.6 A/ 250V. Operating Pressure: 0.90 bar. Air flow: 14l/min. Noise Level: 50db. Temporary Use: 30/30 min. Accessories: a) Face Mask (5 piece) b) Faster Jet Bulb (1 piece) c) Mouthpiece (1 piece) d) Nasal Prong (1 piece) e) Connection Tubing (1 piece) Or Equivalent	08 Nos	Nivec III MG Electric Yes UK	Air Mist Flame Yes Italy	Ideal Plus Flame Italy Yes Italy
	Remarks		Responsive	Responsive	Responsive

S Zia

(Chairman)
Dr. Naeem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad

Memon

(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Janshoro

Anjum

(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
06	<p>SPLIT AIR CONDITIONER 1.5 TON WITH INSTALLATION</p> <p>Make: Gree / Haier / Dawlance</p> <p>Specification:</p> <ul style="list-style-type: none"> • Ton: 1.5 • Frequency: Fixed • Cooling Capacity: 18000 BTU • Air Circulation: 750 m³/h • Power Input: 2000 W • Running Current: 9.6 A • Power Supply: 1,220.50 Ph/V/Hz • EER: 2.45 <p>Includes: Installation with standard warranty. Or Equivalent</p>	14 Nos	LZA Haier or Green Air	 Haier/ Dawlance / Kenwood
	Remarks		Responsive	Responsive

(Chairman)
Dr. Naem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
07	DIGITAL PHOTOSTAT MACHINE Make: Branded Specification: <ul style="list-style-type: none"> • Floor mounted heavy duty Photostat machine • Laser dry imaging system • Warm up time: 30 seconds maximum after powering ON 1 second maximum from the sleep mode • Multiple Copies: 1 to 999 sheets • Maximum Original Size: Max. A3 / 11" x 17" • Copy / Print Speed: 20ppm OR equivalent	02 Nos	5021 Xerox Yes Yes Yes Yes Yes Yes Yes	250ISP NRG Services Yes Yes Yes Yes Yes Yes
Remarks				

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S Zia

Yasmin

Anjum

S. No.	Item Description	Qty Nos	M/s Khalid Corporation
08	PAEDIATRIC AMBU BAG <u>Certification:</u> ISO 9001 CE FDA <u>Country of Origin:</u> Imported <u>Specification:</u> <ul style="list-style-type: none"> • 100% latex free material. • Silicone rubber bag. • Autoclaved repeatedly at 134°C. • Built-in pressure limiting system that ensures a maximum inflation pressure of 70cm H2O. • Max. tidal volume 200prox.: 1300ml • Size (Length x Diameter): 275 x 135mm • Volume O2 reservoir: 200prox.: 1500ml • Expiratory connector: 30mm OR equivalent 	06 Nos	Basmed Taiwan
Remarks			
Responsive			

[Handwritten Signature]

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

[Handwritten Signature]

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

[Handwritten Signature]

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
09	DIGITAL BABY WEIGHT SCALE (Heavy Duty) Certification: ISO 9001 Country of Origin: China Specification: • Digital Weight Scale for Infants and Neonates. • Alkaline Battery • Date Time digital option • Storage of previous weight on LCD screen. • Weight capacity: up to 20-Kgs OR equivalent	02 Nos	Secca	PS3001 Laica
	Remarks		Responsive	Responsive

Zia

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

Yasmin

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

Anjum

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Universal Systematic Solution	M/s Khalid Corpora
10	DUAL CORE COMPUTER WITH LED MONITOR Specification: a) Tower CPU Core i5 3.2GHz (4th Generation) 4GB DDR3 RAM 500GB SATA Hard Drive Optical super combo b) 19" LED Display c) Key Board & Mouse d) UPS (30 minute back-up) OR equivalent	02 Nos	Prodesk 400G4 HP Yes Yes Yes	Unbranded - - -
Remarks				

Scanned

(Chairman)
Dr. Naeem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad

(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Jamshoro

Naeem

(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

Anjum

S. No.	Item Description	Qty	M/s Universal Systematic Solution	M/s Khalid Corpora
11	<p>LASER PRINTER HEAVY DUTY</p> <p>Specification:</p> <ul style="list-style-type: none"> ● Print speed black: Normal: Up to 30 ppm ● First page out (ready) Black: Up to 9.2 sec ● Print quality black (best) Up to 600 x 600 x 2 dpi ● Resolution: HP FastRes 600, HP FastRes 1200 ● Duty cycle (monthly, A4): Up to 5000 pages ● Processor speed: 266 MHz ● Print languages: Host-based printing ● Connectivity: standard 1 Hi-Speed USB 2.0 ● Paper handling input, standard 150-sheet input tray ● Paper handling output, standard 100-sheet output bin Hewlett-Packard <p>Or equivalent</p>	02 Nos	402D HP	402DN6 HP
Remarks			Responsive	Responsive

SZM

Yasmin

Anjum Rehman

(Chairman)
Dr. Naeem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
12	<p>ELECTRONIC PATIENT NUMBERING SYSTEM</p> <p>Queueing Server (Quantity: 1)</p> <ul style="list-style-type: none"> ● Intel Core i5 2 nd generation, 8GB DDR3, 250GB SATA, 15 inch monitor and accessories ● Wireless B/G/N supported wireless connectivity router ● Integrated queue management software system package ● Other required accessories, cabling and connectivity suite <p>Queue Kiosk (Quantity: 2)</p> <ul style="list-style-type: none"> ● Self-service kiosk integrated hardware ● 10 inch digital display capability ● Android 4+ based operating system ● Integrated queue management software system package ● Bluetooth connection capability ● Wireless connection capability ● Wireless printing capability ● Thermal receipt printing capability ● 80 meters paper roll tray ● Other required accessories, cabling and connectivity suite <p>Waiting Lounge Digital Display (Quantity: 2)</p> <ul style="list-style-type: none"> ● 32 inch diagonal smart digital display ● Android 4+ based operating system ● Integrated queue management software system package ● Wireless connection capability ● Announcement speakers ● Wall mount kit ● Other required accessories, cabling and connectivity suite <p>Attendant Console (Quantity: 9)</p> <ul style="list-style-type: none"> ● Intel Core i5 2 nd generation, 4GB DDR3, 250GB SATA, 15 inch monitor based portable hardware ● Advanced Kingston Lock for hardware security ● Wireless connection capability 	02 Nos	Wavelec	Imported

Sam

Wahid

Sam

S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
	<ul style="list-style-type: none"> ● Integrated queue management software system package ● Other required accessories, cabling and connectivity suite <p><u>Virtual Queue</u></p> <ul style="list-style-type: none"> ● 2 x booking kiosks ● 1 x infant consultant queue ● 1 x nutrition consultant queue ● 3 x general consultants queue for children under 5 years ● 1 x pharmacy queue ● 2 x general consultants queue for children over 5 years ● 1 x endocrine consultant queue <p>Or equivalent</p>		Yes	Yes
	Remarks		Responsive	Responsive

Slow

(Chairman)
Dr. Naem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad

(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
13	SOUND SYSTEM WITH MIC <ul style="list-style-type: none"> ● High quality sound ● Wireless speaker ● Wireless mike Or equivalent	01 No	Toa Japan	Toa Japan
	Remarks		Responsive	Responsive

Qaw

Yasmin

Anjum

(Chairman)
Dr. Naem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad


(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro

(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad


S. No.	Item Description	Qty	M/s Universal Systematic Solution	M/s Khadij Corporation	M/s Al-Farooque Enterprises
14	CCTV SECURITY SYSTEM	01 No	SCN2B1040 WD-1 Scout	HEW4PR3 HEN16102 HEN08102 Honeywell	1431SP Dahua
	Specification: Paeds Unit-I, LUH, Hyderabad. 24 Channel NVR 4.0 Megapixel Digital IP Camera Cat5 Cable Roll Cable Connectors HDD 2 TB Communication Rack Duck Putty Wire Clip Pack Power Coils 24 Port Switch 3K VA UPS Electric Board With Installation OR equivalent		Yes Yes	Yes Yes	Yes Yes
	Remarks Bidders are advised to visit the site of installation in advance to mention running feet of wires and other details in their bid for complete system.		Responsive	Responsive	Responsive



(Chairman)
Dr. Naem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad



(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Jamshoro



(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

S. No.	Item Description	QTY	M/s Universal Systematic Solution	M/s Khalid Corporation	M/s Al-Farooque Enterprises
15	CCTV SECURITY SYSTEM	01	SCN2B1040 WD-1	HEW4PR3 HEN32104 Honeywell	1431SP Dahua
	<u>Specification:</u> Paeds Unit-II, LUH, Hyderabad.	01	Yes	Yes	Yes
	32 Channel NVR	28	Yes	Yes	Yes
	4.0 Megapixel Digital IP Camera	05	Yes	Yes	Yes
	Cat5 Cable Roll	280	Yes	Yes	Yes
	Cable Connectors	03	Yes	Yes	Yes
	HDD 2 TB	04	Yes	Yes	Yes
	Communication Rack	60	Yes	Yes	Yes
	Duck Putty	03	Yes	Yes	Yes
	Wire Clip Pack	03	Yes	Yes	Yes
	Power Coils	03	Yes	Yes	Yes
	24 Port Switch	01	Yes	Yes	Yes
	3KVA UPS	01	Yes	Yes	Yes
	Electric Board		Yes	Yes	Yes
	With Installation		Yes	Yes	Yes
	OR equivalent		Yes	Yes	Yes
	Note				
	Bidders are advised to visit the site of installation in advance to mention running feet of wires and other details in their bid for complete system.				
	Remarks		Responsive	Responsive	Responsive

Sem

(Chairman)
Dr. Naeem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad


U. Yasin

(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Janshoro


Dr. Anjum Rehman

(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad


S. No.	Item Description	Qty	M/s Khalid Corporation	M/s Al-Farooque Enterprises
16	BP APPARATUS MERCURY STAND Country of Origin: USA/EU/Japan Specification: • Cuff Size For Newborn, Infant and Children • Accurate reading. • Emulsoid Bladder & Soft emulsoid Bulb • Mercury lock facility • The Height of pole is adjustable (with range of 105mc to 146cm) • Five wheels with brakes • Measure Scope: 0-300mmHg • Accuracy: 3mmHg • Glass Tube: 3.5 to 4.0mm OR equivalent	18 Nos	620 Yamasu Japan Yes Yes Yes Yes Yes Yes Yes Yes Yes OR equivalent	605P Yamasu Japan Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes
Remarks				
Responsive				



(Chairman)
Dr. Naem Zia Memon
 Addl: Medical Superintendent (Gen)
 Liaquat University Hospital
 Hyderabad



(Member)
Dr. Yasmin Memon
 Chairperson of Paediatric Department
 LUMHS
 Jamshoro



(Member)
Dr. Anjum Rehman
 Addl: Medical Superintendent
 Liaquat University Hospital
 Hyderabad



LIAQUAT UNIVERSITY HOSPITAL

HYDERABAD / JAMSHORO

FINANCIAL BID EVALUATION REPORT

OPENED ON 06.03.2018




**TENDER FOR PURCHASE OF MACHINERY /EQUIPMENT
(UP-GRADATION OF NEONATAL & PAEDIATRIC SERVICES)
FOR THE YEAR 2017-18**

**TENDER NO. 33/2017-18
(DATED: 07.02.2018)**

Enclosed:

- 1) Bid Evaluation Report**
- 2) Comparative Statement**
- 3) Meeting Minutes**





**MEETING OF PROCUREMENT COMMITTEE FOR OPENING OF FINANCIAL BIDS
FOR PURCHASE OF MACHINERY / EQUIPMENT (UP-GRADATION OF NEONATAL &
PAEDIATRIC SERVICES) UNDER ADP SCHEME FOR THE FINANCIAL YEAR 2017-18
IN THE COMMITTEE ROOM LUH, HYDERABAD ON 06.03.2018.**

S.NO	NAME & DESIGNATION		SIGNATURE
01	Dr. Naeem Zia Memon Addl: Medical Superintendent. LUH, Hyderabad.	Chairman	
02	Dr. Yasmeen Memon Chairperson Paediatric Department LUMHS, Jamshoro	Member	
03	Dr. Anjum Rehman Deputy Medical Superintendent. L.U. Hospital, Hyderabad.	Member	

Opening of Financial Bids for Purchase of Machinery / Equipment (Up-Gradation of Neonatal & Paediatric Services)
for the Financial year 2017-18 at Liaquat University Hospital, Hyderabad/Jamshoro


Dated:- 06.03.2018

ATTENDANCE SLIP


S.No.	Name of Firm	Phone Number	Postal Address	Name of Participant	C.N.I.C No: of participants	Signature
1	IRIOS	03332224059	Kareeki	Mohammad Farooq	4130389496229	
2	Medoqurps SMC	03344225918	Kareeki	M. JUNAID	4130555348819	
3	USS Enterprises	03350022660	Kareeki, Hydr	Kashif Ali	4120562278719	
4	CARE HEALTH SOLUTIONS	0330-2233062	Kareeki	Sanaullah Ahmad	0330-2233062	
5						
6						
7						


(Dr. Anjum Rehman)

Addl. Medical Superintendent (Technical)
LUH, Hyderabad
Member Procurement Committee


(Dr. Yasmeen Memon)


Chairperson Paediatric Department
LUMHS, Jamshoro
Member Procurement Committee


(Dr. Naeem Zia Memon)


Addl. Medical Superintendent (General)
LUH, Hyderabad
Chairman Procurement Committee

Bid Evaluation Report


- 1) Name of Procuring Agency: Liaquat University Hospital Hyderabad / Jamshoro
- 2) Tender Reference No: 33/2017-18
- 3) Tender Description: Purchase of Machinery /Equipment(Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18
- 4) Method of Procurement: Single Stage Two Envelope
- 5) Tender Published: Dawn - Jang 22.01.2018 Kawish 23.01.2018
SPPRA Website 36313
- 6) Total Bid Documents Sold: 09
- 7) Total Bids Received: 09
- 8) Technical Bid Opening Date: 07.02.2018
- 9) No. of Bid Technically Qualified: 08
- 10) Bid(s) Rejected: 01
- 11) Financial Bid Opening Date: 06.03.2018
- 12) **Bid Evaluation Report:** **Enclosed (i) Comparative Statement (ii) Minutes**



(Chairman)
Dr. Naeem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad



(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Jamshoro



(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

Tender No. 33/2017-18

Purchase of Machinery / Equipment (For Paediatric Department) for the Year 2017-18

(Dated: 07.02.2018)

Financial Bids Opened on 06.03.2018

COMPARATIVE STATEMENT

S. No.	Item Description	Qty	Bidder	Model	Make	C&F	DDP	Remark
1	Installation of Central Oxygen System (for 60 Bed Ward)	1	M/s Universal Enterprises, Karachi	Novair	France	NQ	Rs.5,798,800/-	1st Lowest
			M/s Elate cc (Pvt) Ltd	GCE Healthcare	UK	NQ	Rs.6,000,000/-	2nd Lowest
			M/s Care Health Solution, Karachi.	Pneumatich.	UK	£ 47,000/- (Rs.7,372,890)	Rs 7,277,200/-	3rd Lowest
2	Installation of Central Oxygen System (for 46 Bed Ward)	1	M/s Universal Enterprises, Karachi	Novair	France	NQ	Rs.5,399,800/-	1st Lowest
			M/s Care Health Solution, Karachi.	Pneumatich.	UK	£ 39,000/- (Rs.6,117,930/-)	Rs 5,966,700/-	2nd Lowest
			M/s Elate cc (Pvt) Ltd	GCE Healthcare	UK	NQ	Rs.6,000,000/-	3rd Lowest
3	Paediatric Resuscitation Trolley	4	M/s Bios.	OKM 730-333	Okuman Turkey	US \$ 7,700/- (Rs 851,620)	Rs.940,000/-	1st Lowest
			M/s Medequips	Sunflower	Atom, Japan	¥ 2,200,000/- (Rs.2,235,200/-)	Rs.2,600,000/-	2nd Lowest
4	Sphygmomanometer	6	No responsive or any offer received from eligible bidder for this item.					
5	Heavy Duty Nebulizer	8	M/s Al-Farooque	Ideal Plus	Flaem Italy	NQ	Rs.16,500/-	1st Lowest

S. No.	Item Description	Qty	Bidder	Model	Make	C&F	DDP	Remarks
6	Split Air Conditioner 1.5 Ton with Installation	14	M/s Al-Farooque Technologies.	MG Electric	UK	NQ	Rs.55,000/-	2 nd Lowest
7	Digital Photostat Machine	02	M/s Al-Farooque	Haler/ Dawlance/ Kenwood		NQ	Rs.74,500/-	Lowest. Be Evaluated
8	Paediatric Ambu Bag	06		250ISP	NRG Services	NQ	Rs.348,000/-	Lowest. Be Evaluated
No responsive or any offer received from eligible bidder for this item.								
9	Digital Baby Weight Scale	02	M/s Al-Farooque Enterprises, Hyd	PS3001	Laica		Rs.16,500/-	Lowest. Be Evaluated
10	Dual Core Computer with LED Monitor	02	M/s Universal Systematic Solution.	Prodesk 400 G4	HP	NQ	Rs.69,750/-	Lowest. Be Evaluated
11	Laser Printer Heavy Duty	02	M/s Universal Systematic Solution.	402-D	HP	NQ	Rs.32,000/-	Lowest. Be Evaluated
12	Electronic Patient Numbering System	02	M/s Al-Farooque Enterprises, Hyd		Imported	NQ	Rs 340,000/-	Lowest. Be Evaluated
13	Sound System with Mic	01	M/s Al-Farooque Enterprises, Hyd	TOA	Japan	NQ	Rs 86,500/-	Lowest. Be Evaluated
14	CCTV Security System (PAEDS-I)	01	M/s Al-Farooque Enterprises, Hyd	143ISP	Dahua	NQ	Rs 412,000/-	1 st Lowest
			M/s Universal Systematic Solution.	SCN2B1040 WD-1	Scout	NQ	Rs.586,650/-	2 nd Lowest
15	CCTV Security System (PAEDS-II)	01	M/s Al-Farooque Enterprises, Hyd	143ISP	Dahua	NQ	Rs.580,800/-	1 st Lowest
			M/s Universal Systematic Solution	SCN2B1040 WD-1	Scout	NQ	Rs.801,050/-	2 nd Lowest

Handwritten signature/initials

Handwritten signature

Handwritten signature

S. No.	Item Description	Qty	Bidder	Model	Make	C&F	DDP	Remarks
16	BP Apparatus Mercury Stand	18	M/s Al-Farooque Enterprises, Hyd	605P	Yamasu	NQ	RS.18,600/-	Lowest. Best Evaluated

Exchange Rate: \$: US Dollar 1 = 110.60 PKR € : Euro 1 = 137.49 PKR £ : British Pound 1 = 156.87 PKR CHF : Swiss Franc 1 = 118.46 PKR ¥ : Japanese Yen 1 = 1.016 PKR

S Zame

Dr. Naeem Zia Memon
Addl. Medical Superintendent
Liaquat University Hospital
Hyderabad / Jamshoro
(Chairman)

Naeem

Dr. Yasmeen Memon
Chairperson
Paeds Department
LUMHS, Jamshoro
(Member)

Anjum Rehman

Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad
(Member)

MEETING MINUTES OF THE PROCUREMENT COMMITTEE

Meeting of the Procurement Committee (PC) of Liaquat University Hospital Hyderabad / Jamshoro was held on 06.03.2018 at 12:30 pm at the Conference Hall. Dr. Naeem Zia Memon chaired the meeting. The financial bids were opened after finalization of technical bids by the procurement committee for Purchase of Machinery / Equipment (For Paediatric Department) for the Year 2017-18. Technical bids of Tender Enquiry No. 33/2017-18 were opened on 07.02.2018 and subsequently approved by the Procurement Committee for procurement of required items.

The following members attended the meeting:

No.	Name	Department	Designation
1.	Dr. Naeem Zia Memon	Addl. Medical Superintendent Liaquat University Hospital Hyderabad	Chairman <i>[Signature]</i>
2.	Dr. Yasmin Memon	Chairperson Paediatric Department LUMIIS. Jamshoro	Member <i>[Signature]</i>
3.	Dr. Anjum Rehman	Deputy Medical Superintendent Liaquat University Hospital Hyderabad	Member <i>[Signature]</i>

Initiating the discussions, this office presented the evaluated technical data sheets of bidders for perusal of the Procurement Committee (Copy of the Technical Evaluation Report enclosed.)

The financial bids of the eligible and qualified firms were as under:

- 1) M/s Bios. Karachi.
 - 2) M/s Medequips, Karachi.
 - 3) M/s Elate C.C (Pvt)Ltd
 - 4) M/s Bio Medical Technologies. Karachi.
 - 5) M/s Care Health Solution. Karachi.
 - 6) M/s Universal Enterprises, Karachi.
 - 7) M/s Al-Farooque Enterprises. Hyderabad.
 - 8) M/s Universal Systematic Solution. Karachi.
- Note: The Bid of M/s Khalid Corporation. Karachi was rejected due to less call deposit

S. No.	Bidder Name	Offered Bid Sum Value	Bid Security
1	M/s Bios, Karachi.	Rs 3,760,000/-	No. 50617378 Dated: 06-02-2018 Summit Bank Limited Karachi.
2	M/s Medequips, Karachi.	Rs 10,400,000/-	No.01593603 Dated: 02-02-2018 Habib Metropolitan Bank Ltd. Karachi

3	M/s Elate CC (Pvt) Ltd, Karachi.	Rs.300,000/-	Bank Guarantee Dated: 06-02-2018 Bankislani Pakistan Limited Karachi
4.	M/s Bio Medical Technologies, Karachi.	Rs.17,008,800/-	No. 00110330 Dated: 06-02-2018 MCB Islamic Bank Limited Karachi
5.	M/s Care Health Solution, Karachi.	Rs.13,243,900/-	No. 11949902 Dated: 06-02-2018 Allied Bank, Karachi
6.	M/s Universal Enterprises, Karachi.	Rs.11,198,600/-	No. 14042366 & 14214512 Dated: 05-1-18/06-02-2018 United Bank Limited Karachi
7.	M/s Al-Farooque Enterprises, Hyderabad.	Rs.3,998,100/-	No. 71701718 Dated: 07-02-2018 Askari Bank, Karachi
8.	M/s Universal Systematic Solution, Karachi.	Rs.1,591,200/-	No. 14285656 Dated: 06-02-2018 United Bank Limited Karachi
9.	M/s Khalid Corporation, Karachi	Rs.11,198,800/-	No:11949901 Dated: 06.02.2018 Allied Bank, Karachi.

Tender No : **33/2017-18**
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : **01**

Name of Item : **Installation of Central Oxygen System (for 60 Bed Ward)**

Bidder	M/s Universal Enterprises	M/s Elate cc (Pvt)Ltd	M/s Care Health Solution
Model	Novair	GCE Healthcare	Pneumatech
Make	France	UK	UK

The offer of M/s Universal Enterprises, M/s Elate cc (Pvt) Ltd and M/s Care Health Solution meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Universal Enterprises, Karachi offered the rates of Rs.5,798,800/- each on DDP basis for the said item. M/s Elate cc (Pvt) Ltd offered the rates of Rs.6,000,000/- each on DDP basis for the said item and M/s Care Health Solution offered the rates of £ 47,000/- each on C&F and Rs 7,277,200/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Universal Enterprises, Karachi on DDP basis.

Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : 02

Name of Item : **Installation of Central Oxygen System (for 46 Bed Ward)**

Bidder	M/s Universal Enterprises	M/s Elate cc (Pvt)Ltd	M/s Care Health Solution
Model	Novair	GCE Healthcare	Pneumatech
Make	France	UK	UK

The offer of M/s Universal Enterprises, M/s Elate cc (Pvt) Ltd and M/s Care Health Solution meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Universal Enterprises, Karachi offered the rates of Rs.5,399,800/- each on DDP basis for the said item. M/s Elate cc (Pvt) Ltd offered the rates of Rs.6,000,000/- each on DDP basis for the said item and M/s Care Health Solution offered the rates of £ 39,000/- each on C&F and Rs 5,966,700/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Universal Enterprises, Karachi on DDP basis.

Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : 03

Name of Item : **Paediatric Resuscitation Trolley.**

Bidder	M/s BIOS	M/s Medequips
Model	OKM 730-333	Sunflower
Make	Okuman Turkey	Atom, Japan

The offer of M/s Bios, Karachi and M/s Medequips Karachi meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Bios, Karachi offered the rates of \$ 7,700 each on C&F and Rs.940,000/- each on DDP basis for the said item. M/s Medequips offered the rates of ¥ 2,200,000 each on C&F and Rs.2,600,000/- each on DDP basis for the said item

The Procurement Committee approved and accepted the offer of M/s Bios, Karachi on C&F basis.

Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : 04

Name of Item : **Sphygmomanometer**

No responsive or any offer received from eligible bidder for this item.

The Procurement Committee decided to re-tender this item.

Tender No : **33/2017-18**
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : **05**

Name of Item : **Heavy Duty Nebulizer**

Bidder	M/s Al-Farooque Enterprises	M/s Bio Medical Technologies
Model	Ideal Plus	MG Electric
Make	Flaem Itally	UK

The offer of M/s Al-Farooque Enterprises and M/s Bio Medical Technologies meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Al-Farooque Enterprises offered the rates of Rs.16,500/- each on DDP basis for the said item. M/s Bio Medical Technologies offered the rates of Rs.55,000.- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

Tender No : **33/2017-18**
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : **06**

Name of Item : **Split Air Conditioner 1.5 Ton with Installation.**

Bidder	M/s Al-Farooque Enterprises
Model	
Make	Haier/Dawlance/Kenwood

The offer of M/s Al-Farooque Enterprises meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Al-Farooque Enterprises offered the rates of Rs.74,500/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

Tender No : **33/2017-18**
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : **07**

Name of Item : **Digital Photostat Machine**

Bidder	M/s Al-Farooque Enterprises
Model	250ISP
Make	NRG Services



The offer of M/s Al-Farooque Enterprises meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Al-Farooque Enterprises offered the rates of Rs.348,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

**Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18**

Item No : 08

Name of Item : Paediatric Ambu Bag

No responsive or any offer received from eligible bidder for this item.

The Procurement Committee decided to re-tender this item.

**Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18**

Item No : 09

Name of Item : Digital Baby Weight Scale

Bidder	M/s Al-Farooque Enterprises
Model	PS3001
Make	Laica

The offer of M/s Al-Farooque Enterprises meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Al-Farooque Enterprises offered the rates of Rs.16,500/- each on DDP basis for the said item.


The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

**Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18**

Item No : 10

Name of Item : Dual Core computer with LED Monitor.

Bidder	M/s Universal Systematic Solution
Model	Prodesk 400 G4
Make	HP



The offer of M/s Universal Systematic Solution meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Universal Systematic Solution offered the rates of Rs.69.750/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Universal Systematic Solution on DDP basis.

**Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradaation of Neonatal & Paediatric Services) for the Year 2017-18**

Item No : 11

Name of Item : Laser Printer Heavy Duty.

Bidder	M/s Universal Systematic Solution
Model	402-D
Make	HP

The offer of M/s Universal Systematic Solution meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Universal Systematic Solution offered the rates of Rs.32.000/- each on DDP basis

The Procurement Committee approved and accepted the offer of M/s Universal Systematic Solution on DDP basis.

**Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradaation of Neonatal & Paediatric Services) for the Year 2017-18**

Item No : 12

Name of Item : Electronic Patient Numbering System.

Bidder	M/s Al-Farooque Enterprises
Model	
Make	Imported

The offer of M/s Al-Farooque Enterprises meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Al-Farooque Enterprises offered the rates of Rs.340.000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

**Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradaation of Neonatal & Paediatric Services) for the Year 2017-18**

Item No : 13

Name of Item : Sound System with Mic

Bidder	M/s Al-Farooque Enterprises
Model	TOA
Make	Japan

The offer of M/s Al-Farooque Enterprises meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms was opened.

M/s Al-Farooque Enterprises offered the rates of Rs.86,500/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : 14

Name of Item : CCTV Security System (Paeds-I)

Bidder	M/s Al-Farooque Enterprises	M/s Universal Systematic Solution
Model	1431SP	SCN2B1040 WD-1
Make	Dahua	Scout

The offer of M/s Al-Farooque Enterprises and M/s Universal Systematic Solution meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Al-Farooque Enterprises offered the rates of Rs.412,000/- each on DDP basis for the said item. M/s Universal Systematic Solution offered the rates of Rs.586,650/- each on DDP basis

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.

Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18

Item No : 15


Name of Item : CCTV Security System (Paeds-II)

Bidder	M/s Al-Farooque Enterprises	M/s Universal Systematic Solution
Model	1431SP	SCN2B1040 WD-1
Make	Dahua	Scout

The offer of M/s Al-Farooque Enterprises and M/s Universal Systematic Solution meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Al-Farooque Enterprises offered the rates of Rs.580,800/- each on DDP basis for the said item. M/s Universal Systematic Solution offered the rates of Rs.801,050/- each on DDP basis

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.



Tender No : 33/2017-18
Purchase of Machinery /Equipment (Up-Gradation of Neonatal & Paediatric Services) for the Year 2017-18
Item No : 16
Name of Item : BP Apparatus Mercury Stand.


Bidder	M/s Al-Farooque Enterprises
Model	605P
Make	Yamasu

The offer of M/s Al-Farooque Enterprises meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Al-Farooque Enterprises offered the rates of Rs.18.600/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Al-Farooque Enterprises on DDP basis.


The meeting ended with vote of thanks from the chair.



(Chairman)
Dr. Naeem Zia Memon
Addl: Medical Superintendent (Gen)
Liaquat University Hospital
Hyderabad



(Member)
Dr. Yasmin Memon
Chairperson of Paediatric Department
LUMHS
Jamshoro




(Member)
Dr. Anjum Rehman
Addl: Medical Superintendent
Liaquat University Hospital
Hyderabad

SPPRA S.No:36313 SPPRA I.D 2147483647


CERTIFICATE

The rate of items quoted by single firm in the tender are as per
Market Rate / Price.




(Dr.Anjum Kehman)

Addl: Medical Superintendent
Liaquat University Hospital, Hyd
Member Procurement Committee



(Dr.Yasmeen Memon)

Chairperson of Paediatric Department
LUM&HS, Jamshoro
Member Procurement Committee



(Dr.Naeem Zia Memon)

Addl: Medical Superintendent (General)
Liaquat University Hospital, Hyderabad
Chairman Procurement Committee