

**MINUTES OF BID OPENING MEETING HELD ON 28<sup>TH</sup> APRIL 2015 AT 1:00 PM UNDER THE CHAIRMANSHIP OF MEDICAL SUPERINTENDENT / CHAIRMAN, PROCUREMENT COMMITTEE, CMC HOSPITAL, LARKANA AT ADMIN BLOCK, OFFICE OF M.S CMCH LARKANA**

The meeting of the Procurement Committee, CMC Hospital Larkana was held on 28<sup>th</sup> April 2015 under the Chairmanship of Medical Superintendent / Chairman, CMCH, Larkana at Admn. Block office of Medical Superintendent, CMC Hospital Larkana to open Technical Bids for the purchase of machinery, equipments, furniture / fixture in response to NIT No: CMCHL/(P&D):/104 dated: 30-03-2015

Following members of Procurement Committee, CMC Hospital, Larkana participated in the meeting:

1	Dr. Safiullah Abbasi Medical Superintendent, CMC Hospital, Larkana	Chairman
2	Dr. Khalid Shaikh Special Secretary (PH), Health Department Government of Sindh, Karachi.	Member
3	Prof. Malak Hussain Jalbani HOD Urology, SMBBMU, Larkana	Member
4	Prof. Shahida Shaikh HOD Obts. / Gynae, SMBBMU, Larkana	Member
5	Prof. Abdul Manan Bhutto HOD Dermatology, SMBBMU, Larkana	Member
6	Prof. Muhammad Ilyas Shaikh, Principal, Baby Asifa Dental College, SMBBMU, Larkana	Member
7	Dr. Aijaz Ahmed Shaikh Additional Medical Superintendent (Tech:), CMC Hospital Larkana	Member

**The List of Bidders who participated in the meeting is enclosed as Annexure "A":-**

In opening remarks the Chairman, Procurement Committee, CMC Hospital, Larkana welcomed members of Committee and Bidders or their Authorized representatives participated in the Bid opening meeting then he stated that in response to NIT No: CMCHL/(P&D):/104 dated: 30-03-2015 bids were purchased by interested Bidders till closing date i.e 28-04-2015. Then asked to proceed further, therefore sealed / locked box was opened before members and bidders or their authorized representatives. The Single stage two envelopes procedure was adopted as per NIT, only Technical Proposals were opened and sealed Financial Proposals un-touched and kept in the box. The Chairman & members of the Committee signed each & every Technical Proposal. Total 31 numbers of bids were found in the box.


After opening of Technical Bids, the Procurement Committee, CMC Hospital Larkana, unanimously

decided that evaluation of the Technical Bids should be carried-out by a Technical Committee members / stakeholders of the Heads of Department consisting of following members:


1	Prof. Malak Hussain Jalbani HOD Urology, SMBBMU, Larkana	Member
2	Prof. Shahida Shaikh HOD Obts. / Gynae, SMBBMU, Larkana	Member
3	Prof. Abdul Mana Bhutto HOD Dermatology, SMBBMU, Larkana	Member
4	Prof. Muhammad Illyas Shaikh, Principal, Baby Asifa Dental College, SMBBMU, Larkana	Member

It was also decided that in the light of recommendation of above Committee the Financial Bids will be opened to only technically qualified bidders on 05-05-2015 at 1.00 pm in the office of Medical Superintendent, CMC Hospital, Larkana the sealed financial bids of un-qualified bidders will be returned back.

The meeting ended with vote of thanks by the chair.



(Prof: Malik Hussain Jalbani)  
Head of Department of Urology  
SMBBMU, Larkana  
(Member)




(Prof: Shahida Shaikh)  
Head of Department of Obts/ Gynae,  
SMBBMU, Larkana  
(Member)




(Prof: Abdul Mana Bhutto)  
Head of Dermatology Department,  
SMBBMU, Larkana  
(Member)




(Prof: Muhammad Illyas Shaikh)  
Principal, Baby Asifa Dental College  
SMBBMU, Larkana  
(Member)



(Dr, Aijaz Ahmed Shaikh)  
Additional Medical Superintendent (Tech)  
CMC Hospital Larkano  
(Member)



(Dr. Khalid Shaikh)  
Special Secretary (P. H) Health Department,  
Govt, of Sindh, Karachi  
(Member)









(Dr. Saifullah Abbasi)

37

List of Procurement Committee Members, CMC Hospital Larkana for Bid opening meeting held on 28<sup>th</sup> April 2015 at 1.00 pm under the chairmanship of the Medical Superintendent CMC Hospital Larkana at Admin Block, Office of M.S, CMC Hospital Larkana.

NIT No: CMCHL/(P&D):/104 dated 30-03-2015

S.#	Name of Procurement Committee Member	Signature
01	Dr. Safiullah Abbasi Medical Superintendent CMC Hospital, Larkana Chairman	
02	Dr. Khalid Shaikh The Special Secretary (Public Health) Health Department, Government of Sindh Karachi.	
03	Professor Malik Hussain Jalbani Head of Department of Urology, SMBBMU Larkana.	
04	Dr. Shahida Shaikh Head of Department Obst & Gynae SMBBMU, Larkana	
05	Professor Abdul Manan Bhutto Head of Dermatology Department, SMBBMU Larkana.	
06	Professor Mohammad Ilyas Shaikh, Principal, Bibi Asifa Dental College, SMBBMU Larkana	
07	Dr. Aijaz Ahmed Shaikh Additional Medical Superintendent (Tech:) CMC Hospital Larkana.	 28/4/15

List of participants of Bid opening meeting held on 28<sup>th</sup> April 2015 at 1.00 pm under the chairmanship of the Medical Superintendent CMC Hospital Larkana at Admin Block, Office of M.S, CMC Hospital Larkana.

NIT No: CMCHL/(P&D):/104 dated 30-03-2015

42

S.No.	Name of Company	Name of Authorized Representative	Signature	Remarks.
1	Heuge Enterprise Head	Amer Fye		0334-026924
2	BIOS	Abdul Wahed		031222914587
3	Lot Link Enterprise	Gillem Mohamed		03489628091
4	Lot Link Enterprise	Gillem Mohamed		03489628091
5	Ala Binur	Raj		035795567061
6	Ala Binur	Raj		035795567061
7	Ala Binur	Raj		035795567061
8	Quansyn Medical	Raj		03512032583
9	Quon Sin Medical	Raj		-
10	Elite CCD Ltd	Jyoti An		03337125224
	Verizon	A. Sami Bhutto		030022168781

List of participants of Bid opening meeting held on 28<sup>th</sup> April 2015 at 1.00 pm under the chairmanship of the Medical Superintendent CMC Hospital Larkana at Admin Block, Office of M.S, CMC Hospital Larkana.

NIT No: CMCHL/(P&D):/104 dated 30-03-2015

23

S.No.	Name of Company	Name of Authorized Representative	Signature	Remarks
	Noor International	Faizan	[Signature]	0300-3084700
	Universal Enterprises	Nabeed Ali	[Signature]	03003156190
	Basim's medical international	M. Rizwan	[Signature]	0335-713220
	Aizba Corporation	Umar Iqbal	[Signature]	0821-2861489
	Allmed solutions	M. Zeeshan	[Signature]	0836-3852055
	ENDOKARE	Rashid Ahmed	[Signature]	0321-8925179
	Combined Engineers	Danish Bhatt	[Signature]	0342 8114878
	New Day	Afzal	[Signature]	0333-7142960

**MINUTES OF TECHNICAL COMMITTEE MEETING HELD ON 2<sup>nd</sup> MAY, 2015 AT 11:30 A.M AT MAIN ADMIN. BLOCK, OFFICE OF MEDICAL SUPERINTENDENT CMCH LARKANA**

The following participated in the meeting:


1	Prof. Malak Hussain Jalbani HOD Urology, SMBBMU, Larkana	Member
2	Prof. Shahida Shaikh HOD Obts. / Gynae, SMBBMU, Larkana	Member
3	Prof. Abdul Manan Bhutto HOD Dermatology, SMBBMU, Larkana	Member
4	Prof. Muhammad Illyas Shaikh, Principal, Baby Asifa Dental College, SMBBMU, Larkana	Member

As decided in the Bid Opening meeting held on 28-04-2015, the meeting of Technical Committee members held on 02-05-2015 at 11.30 A M in Main Admn. Block office of the Medical Superintendent, CMC Hospital, Larkana to Scrutinize & Evaluate the Technical Bids for their eligibility in response to CMC Hospital NIT No: CMCHL/ (P&D):/104 dated: 30-03-2015

In accordance with the procedure under the provision of SPPRA Rules, 2010 define in the NIT, the sealed bids submitted by the bidders (under single stage two envelop procedure) in response to procurement above reference NIT, the Committee Scrutinized / Evaluated the Technical Proposals / Bids in accordance with the criteria already defined in the bid documents and provided to bidders. The Committee after detailed scrutiny of bids awarded marks to the bidders in view of documents provided.

The detailed Evaluation Report duly signed by the members of Committee is @ Annexure-"A". The Bidders technically qualified by obtaining minimum qualifying marks and un-qualified bidders by not obtaining minimum qualifying marks due to reasons are shown in detailed Evaluation Report.


The Committee decided that the Financial Proposals of the technically qualified Bidders will be opened and scrutinized in the next meeting of Procurement Committee, CMC Hospital, Larkana that is fixed on 05-05-2015 at 1.00 pm in the Committee Room of the Office of the Medical Superintendent, CMC Hospital Larkana.




(Prof: Malik Hussain Jalbani)  
Head of Department of Urology  
SMBBMU, Larkana  
(Member)



(Prof: Abdul Manan Bhutto)  
Head of Dermatology Department,  
SMBBMU, Larkana  
(Member)



(Prof: Shahida Shaikh)  
Head of Department of Obts/ Gynae,  
SMBBMU, Larkana  
(Member)



(Prof: Mohammad Illyas Shaikh)  
Principal, Baby Asifa Dental College  
SMBBMU, Larkana








## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015

Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015
3	Tender Description/Name of work/item:	Purchase of Furniture/Fixture Skin Disease, Dental Dept., Hepatology Dept. & SZW Hospital under ADP Scheme No. 623,625 (2014-2015)
4	Method of Procurement:	Single Stage - two Envelope Procedure
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. <b>1575219283</b> <b>Upload on 15-4-2015</b> ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015
6	Total Bid documents Sold:	02
7	Total Bids Received:	02
8	Technical Bid Opening date: (if applicable) <i>(Provide details in separate form)</i>	28-04-2015
9	No. of Bid technically qualified (if applicable):	02
10	Bid(s) Rejected:	Nil
11	Financial Bid Opening date:	05-05-2015
12	<b>Bid Evaluation Report:</b>	

S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
1	2	3	4	5	6	7
		Details are @ Annexure-F				

(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt, of Sindh, Karachi (Member)	
(Prof: Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof: Shahida Shaikh) HOD Obsts/ Gynae, SMBBMU, Larkana (Member)	
(Prof: Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof: Mohammad Illyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr, Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkano	


Annexure A.

Supply, Installation, Testing, Commissioning & maintenance (During maintenance period) of Skin Disease Complex and Hepatology Department Equipments / Instruments under ADP Scheme viz:


Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623 (2014-15)

Bid Evaluation

Sr	Name Of Bidder	Name of The Firm and its Establishment	Original Tender Purchase Receipt	Sale Tax / NTN Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit /e. Firm Has Never Been Black Listed	Registration with Various Depart.	Annual Turnover Report	Bid is signed & Stamped by Authorize Person Along with Authorization Letter	Authentic /Verifiable Manufacture & Authority Letter	Bid Security Attached	REMARKS
1	M/s Quantum Medical	7	YES	Yes (5)	Yes (30)	Yes (7)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	81%
2	M/S Business International	8	YES	Yes (5)	Yes (35)	Yes (7)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	90%
3														
5														

  
 Prof: Maki Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Shahida Shaikh  
 Head of Department of Oth / Gynee,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Mohammad Ilyas Shaikh,  
 Principal, Iqbal Aista Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Kazi Jahangir Bhutto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmad Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt. of Sindh, Karachi  
 (Member)

  
 Dr: Saifullah Abbasi  
 Medical Superintendent,  
 Shaheed Mohtarma Medical College Hospital,  
 Larkana (Chairman)



**Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of Skin Disease Complex and  
Hepatology Department Equipments / Instruments under ADP Scheme viz;**

**Establishment of Shaheed Benazir Bhutto Skin Diseases Complex &  
Construction of Dental and Hepatology Department at CMC Hospital Larkana  
ADP No. 623 (2014-15)**

**Technical evaluation**

Technical Scrutiny and evaluation of offers received from different firms

S.No.	Name of the Items	Comment Of Technical Committee
1.	Trinocular Video Microscopy Systems with Digital Camera	One Firm M/s Quantum Medical has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm.
2. —	Trinocular Video Microscopy Systems with Digital Camera Brand Name: AmScope Model No: T120B-P	.
3.	IPL&Diode Laser	One Firm M/s Quantum Medical has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm.
4.	Top Advanced CO2 Fractional Laser – COFRAX	Two M/s Quantum Medical & M/S Business International has participated in the tender after technical scrutiny it is found that Both Firms are technically qualified and may be purchased from lowest bidder.
5.	New Q-Switched Nd:YAG Laser - LUCAS PLUS	One Firm M/s Quantum Medical has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm. -
6.	Aesthi Peel	One Firm M/s Quantum Medical has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm. -
7.	Ultra Peel System h	Two M/s Quantum Medical & M/S Business International has participated in the tender after technical scrutiny it is found that Both Firms are technically qualified and may be purchased from


		lowest bidder.
8.	PUVA Unit	Two M/s Quantum Medical & M/S Business International has participated in the tender after technical scrutiny it is found that Both Firms are technically qualified and may be purchased from lowest bidder.
9.	Dermalight 80, Phototherapy Device	Two M/s Quantum Medical & M/S Business International has participated in the tender after technical scrutiny it is found that Both Firms are technically qualified and may be purchased from lowest bidder.
10.	AFT Laser System	One Firm M/s Quantum Medical has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be purchased from above said firm.



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obts / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Abdul Manan Bhutto,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Mohammad Illyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aijaz Ahmed Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larakana  
(Member)



Special Secretary (Public Health)  
Health Department, Govt, of Sindh, Karachi  
(Member)



Dr: Sajidullah Abbasi  
Medical Superintendent,  
Chandka Medical College Hospital,  
Larkana (Chairman)

Annexure - B.

Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of Dental Equipments / Instruments under ADP Schemes viz:

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital  
Larkana ADP No. 623 (2014-15)

COMPARATIVE STATEMENT FINANCIAL BID DATED: 05/05/2015


S #	Quoted Item	Bidder			
		Bidder	Bidder	Bidder	Bidder
1	OPG X-Ray Unit	Pak Traders	Platinum	BIOS	
		Tech: Rej	6,000,000	4,900,000	
2	Dental X-Ray Unit	Platinum	Pak Traders	BIOS	Combined Engineering
		465,000	Tech: Rej	Tech:Rej	352,000
3	OT Table	Platinum	Pak Traders		
		260,000	645,000		
4	Surgical Halogen Light	Platinum	Pak Traders		
		1,235,000	Tech: Rej		
5	Chair Mounted (Dental)	Combined Engineering	Platinum	Pak Traders	
		1,840,000	380,000	420,000	
6	Auto Clave for Dental Instrument Medium Size	Platinum	Pak Traders		
		215,000	660,000		
7	Dental Implant Unit	Platinum	Combined Engineering	Pak Traders	
		286,000	475,000	Tech:Rej	
8	Dental Surgical Instrument	Platinum	Pak Traders		
		215,000	Tech: Rej		
9	OPG Auto Laser Machine	Platinum			
		1,428,000			
10	Intra Oral Digital Imaging X-Ray System	Platinum	Pak Traders	Combined Engineering	
		537,000	465,000	582,000	










S #	Quoted Item	Bidder	Bidder	Bidder	Bidder
11	Phantom Head Simulator	Platinum	Pak Traders	Combined Engineering	
		465,000	Tech: Rej	490,000	
12	Diathermy	Platinum	BIOS	Pak Traders	
		475,000	Tech: Rej	635,000	
13	Suction Machine	Platinum	BIOS	Combined Engineering	Pak Traders
		195,000	136,000	147,200	Tech: Rej


  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Shahida Shaikh  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Moharramad Ilyas Shakh,  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Abdul Manan Bhatto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)








  
 Special Secretary (Public Health)  
 Health Department, Govt. of Sindh, Karachi  
 (Member)

  
 Dr Sarfraz Abbasi  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)

## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015  
Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana				
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015				
3	Tender Description/Name of work/item:	Purchase of Other Items for Skin Disease & Hepatology Dept. & SZW Hospital under ADP Scheme No. 623, 625 (2014-2015)				
4	Method of Procurement:	Single Stage - two Envelope Procedure				
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. <b>1575219285</b> <i>upload on 15-4-2015</i> ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015				
6	Total Bid documents Sold:	03				
7	Total Bids Received:	03				
8	Technical Bid Opening date: (if applicable) ( <i>Provide de in separate form</i> )	28-04-2015				
9	No. of Bid technically qualified (if applicable):	03				
10	Bid(s) Rejected:	Nil				
11	Financial Bid Opening date:	05-05-2015				
12	<b>Bid Evaluation Report:</b>					
S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
1	2	3	4	5	6	7
Details are @ Annexure-G						

(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt, of Sindh, Karachi (Member)	
(Prof. Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof. Shahida Shaikh) HOD Obts./ Gynae, SMBBMU, Larkana (Member)	
(Prof. Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof. Mohammad Ilyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr. Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkano	

Supply of other items under ADP Schemes viz:

Annexure - A.

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital

Larkana ADP No. 623 (2014-15)

Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

Technical evaluation

Slr	Name Of Bidder	Name of The Firm an its Establishment	Original Tender Purchase Receipt	Sale Tax /NTN Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit i.e. Firm has Never Been Black Listed	Registration with Various Depart.	Annual Turnover Report	Bid is signed & Stamped by Authorize Person Along with Authorization Letter	Authentic 'Verifiable Manufacturer' Authority Letter	Bid Security Attached	REMARKS
1	M/s Platinum corporation	10	Yes	Yes (5)	Yes (35)	Yes (8)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	93%
2	M/S Pak Traders	5	Yes	Yes (5)	No (0)	No (0)	No (0)	No (0)	Yes (5)	No (0)	No (0)	No (0)	Yes (5)	20%
3														
5	M/s Tooba Electronics	2	Yes	Yes (5)	No (35)	No (10)	No (8)	Yes (5)	Yes (5)	No (0)	Yes (5)	Yes	Yes (5)	78%

Prof. Malik Hussain Jabani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)

Prof. Shahida Shaikh  
Head of Department of Obs / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)

Prof. Mohammad Ilyas Shaikh,  
Principal, Baby Afsa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)

Prof. Abdul Majeed Akhtar,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)

Dr. Aliza Amir Shahri,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larkana  
(Member)

Special Secretary (Public Health)  
Health Department, Govt. of Sindh, Karachi  
(Member)

Dr. Saadullah Abbasi  
Medical Superintendent,  
Chunab Medical College Hospital,  
Larkana (Chairman)

Note: M/S Pak Traders provided relevant documents before opening of Financial Bids, hence met evaluation criteria

Supply of other items under ADP Schemes viz;

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC  
Hospital Larkana ADP No. 623 (2014-15)

Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

Technical evaluation

Technical Scrutiny and evaluation of offers received from different firms

S.No.	Name of the Items	Comment Of Technical Committee
1.	Computer (HP/Dell or equivalent with <u>Laser Printer</u> alongwith <u>LED</u>	<p>Two Firms M/s Platinum and M/s Pak Traders has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
2.	Baby Cots (Export quality)	<p>Two Firms M/s Platinum and M/s Pak Traders has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
3.	Radiant Heater for Neonates (Export quality)	<p>One Firm M/S Pak Traders has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p>
4.	Central Oxygen system <i>h</i>	<p>Two Firms M/s Platinum and M/s Pak Traders has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
5.	Air Conditioners 1.5 Ton with stabilizer	<p>Three Firms M/s Platinum and M/s Pak Traders and Tooba Electronics has participated in the tender after technical scrutiny</p>

		<p>It is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum and Tooba Electronics are as per tender specifications.</p> <p>The Technical Committee has decided that the item may be purchased from lowest bidder.</p>
6.	Air-conditioner windows type 1.5 tons with stabilizer	<p>Three Firms M/s Platinum and M/s Pak Traders and Tooba Electronics has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum and Tooba Electronics are as per tender specifications.</p> <p>The Technical Committee has decided that the item may be purchased from lowest bidder.</p>
7.	Electric water cooler (Export quality)	<p>Three Firms M/s Platinum and M/s Pak Traders and Tooba Electronics has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum and Tooba Electronics are as per tender specifications.</p> <p>The Technical Committee has decided that the item may be purchased from lowest bidder.</p>
8.	Refrigerator (Export quality)	<p>Three Firms M/s Platinum and M/s Pak Traders and Tooba Electronics has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum and Tooba Electronics are as per tender specifications.</p> <p>The Technical Committee has decided that the item may be purchased from lowest bidder.</p>
9.	T.V sets (Export quality)	<p>Three Firms M/s Platinum and M/s Pak Traders and Tooba Electronics has participated in the tender after technical scrutiny it is found that</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/Platinum and Tooba Electronics are as per tender specifications.</p> <p>The Technical Committee has decided that the item may be purchased from lowest bidder.</p>
10.	Multimedia with Screen (Latest)	<p>Two Firms M/s Platinum and M/s Pak Traders has participated in the tender after technical scrutiny it is found that</p>




	(Export quality)	M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
11	Lap Top	Two Firms M/s Platinum and M/s Pak Traders has participated in the tender after technical scrutiny it is found that  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obts / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Abdul Manan Bhutto,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Mohammad Illyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aijaz Ahmed Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larkana  
(Member)



Special Secretary (Public Health)  
Health Department, Govt, of Sindh, Karachi  
(Member)










Dr Safiullah Abbasi  
Medical Superintendent,  
Chandka Medical College Hospital,  
Larkana (Chairman)

## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015  
Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana				
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015				
3	Tender Description/Name of work/item:	Purchase of Machinery & Equipments / Instrument for Skin Disease & Hepatology Dept. CMC HOSPITAL LARKANA ADP Scheme No.623 (2014-2015).				
4	Method of Procurement:	Single Stage - two Envelope Procedure				
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. <b>1575219283</b> upload on <b>15-4-2015</b> ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015				
6	Total Bid documents Sold:	02				
7	Total Bids Received:	02				
8	Technical Bid Opening date: (if applicable) ( <i>Provide details in separate form</i> )	28-04-2015				
9	No. of Bid technically qualified (if applicable):	02				
10	Bid(s) Rejected:	Nil				
11	Financial Bid Opening date:	05-05-2015				
12	<b>Bid Evaluation Report:</b>					
S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
1	2	3	4	5	6	7
1.		Details are @ Annexure-A				
2.						

(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt, of Sindh, Karachi (Member)	
(Prof: Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof: Shahida Shaikh) HOD Obts/ Gynae, SMBBMU, Larkana (Member)	
(Prof: Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof: Mohammad Illyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr, Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkana	

Technical evaluation

S#	Name Of Bidder	Name of The Firm an its Establishment	Original Tender Purchase Receipt	Sale Tax /NTN Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit I.e. Firm has Never Been Black Listed	Registration with Various Depat.	Annual Turnover Report	Bid is signed & Stamped by Authorize Person Along with Authorization Letter	Authentic /Verifiable 'Manufacture's Authority Letter	Bid Security Attached	REMARKS
1	M/S Platinum corporation	10	YES	Yes (5)	Yes (35)	Yes (8)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	93%
2	M/S Pak Traders	5	YES	Yes (5)	No	No	No	No	Yes (5)	No	No	No	Yes (5)	20%
3														



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obs / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Mohammad Ilyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Abdul Munir Bhutto  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aljaz Ahmad Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larkana  
(Member)



Dr: Shaukat Abbasi  
Medical Superintendent,  
Chandoka Medical College Hospital,  
Larkana (Chairman)



Special Secretary (Public Health)  
Health Department, Govt. of Sindh, Karachi  
(Member)

Note: M/S Pak Traders provided relevant documents before opening of Financial Bids, hence met evaluation criteria

*Supply of Furniture / Fixture under ADP Schemes viz;*

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623 (2014-15)  
Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

**Technical evaluation**

Technical Scrutiny and evaluation of offers received from different firms

S.No.	Name of the Items	Comment Of Technical Committee
1.	Fowler Beds (Surgical)	<p>Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
2	Foam Mattress with Regzine Cover (Master / Diamond)	<p>Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
3	Bed Side Lockers	<p>Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
4	Bed Side Benches (Talhi Wood ) Size 4'x1 ½	<p>Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.</p> <p>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</p> <p>Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>

5	Office Chairs (Talhi Wood)	One Firm M/s Platinum has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be purchased from above said firm.
6	Office Tables (local) (Talhi Wood)	One Firm M/s Platinum has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be purchased from above said firm.
7	Nursing Counter	One Firm M/s Platinum has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be purchased from above said firm.
8	Iron safe / Cup Board	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
9	Patients Examination Couch (Local)	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
10	Fowler Beds with Mattress (36+4 Recovery)	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
11	Blankets (Red) 60X90)	
12	Officer Table (6X4)	One Firm M/s Platinum has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be purchased from above said firm.
13	Revolving Chair	One Firm M/s Platinum has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be purchased from above said firm.
14	Chairs Seminar Hall, Library & Other Office	One Firm M/s Platinum has participated in the tender and as per the tender requirement.  Hence, Technical Committee has decided that the item may be

W

		purchased from above said firm.
15	Books Rack for Library	One Firm M/s Platinum has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm.
16	Iron Cupboards	
17	Nursing Counter	One Firm M/s Platinum has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm.
18	Stools (Steel)	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
19	I/V Stands (Steel)	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
20	Wheel Chair	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.  Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
21	Stretchers for Patients	Two Firms M/s Platinum and M/S Pak Traders has participated in the tender and as per the tender requirement.  M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.  Where as M/s Platinum is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
22	Office Chairs	One Firm M/s Platinum has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm.
23	Plastic Chairs for Waiting Area (3+1)	One Firm M/s Platinum has participated in the tender and as per the tender requirement.

h


		Hence, Technical Committee has decided that the item may be purchased from above said firm.
24	Table Side Rack	One Firm M/s Platinum has participated in the tender and as per the tender requirement. Hence, Technical Committee has decided that the item may be purchased from above said firm.



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obs / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Abdul Manan Bhutto,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Mohammad Ilyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aljaz Ahmed Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larakana  
(Member)



Special Secretary (Public Health)  
Health Department, Govt, of Sindh, Karachi  
(Member)










Pt. Safiqulah Abbasi  
Medical Superintendent,  
Chandka Medical College Hospital,  
Larkana (Chairman)

## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015  
Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana				
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015				
3	Tender Description/Name of work/item:	Purchase of Machinery, Equipment / Instrument Under Non-ADP Scheme up gradation of Urological Facilities CMCH Larkana (2014-2015)				
4	Method of Procurement:	Single Stage - two Envelope Procedure				
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. 1575219283 uplo ad on 15-4-2015 ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015				
6	Total Bid documents Sold:	07				
7	Total Bids Received:	07				
8	Technical Bid Opening date: (if applicable) ( <i>Provide date in separate form</i> )	28-04-2015				
9	No. of Bid technically qualified (if applicable):	07				
10	Bid(s) Rejected:	Nil				
11	Financial Bid Opening date:	05-05-2015				
12	<b>Bid Evaluation Report:</b>					
S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
1	2	3	4	5	6	7
		Details are @ Annexure-E				

(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt, of Sindh, Karachi (Member)	
(Prof: Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof: Shahida Shaikh) HOD Obs/ Gynae, SMBBMU, Larkana (Member)	
(Prof: Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof: Mohammad Ilyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr, Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkano	





**TENDER FOR THE SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENTS/INSTRUMENTS UNDER REVENUE COMPONENTS OF THE SCHEME "UPGRADATION OF UROLOGICAL FACILITIES. CMC HOSPITAL LARKANA**

NON ADP (2014-15) TENDER DUE DATE 28-APRIL-2015

**BID EVALUATION**

Sl	Name Of Bidder	Name of The Firm an its Establishment	Original Tender Purchase Receipt	Sale Tax /NTN Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit i.e. Firm has Never Been Black Listed	Registration with Various Deput.	Annual Turnover Report	Bid is signed & Stamped by Authorize Person Along with Authorization Letter	Authentic /Verifiable Manufacture's Authority Letter	Bid Security Attached	REMARKS
1	M/S. ALLMED SOLUTIONS	6	Yes	Yes (5)	Yes (40)	Yes (10)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	96%
2	M/S BIOS	10	Yes	Yes (5)	Yes (10)	NO (0)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	No (0)	Yes	Yes (5)	55%
3	M/S ENDO-KARE	10	Yes	Yes (5)	Yes (40)	Yes (10)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	No (0)	Yes	Yes (5)	95%
4	M/S UNIVERSAL ENTERPRISES	10	Yes	Yes (5)	Yes (20)	Yes (10)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	80%
5	M/S QUANTUM MEDICAL TECHNOLOGIES	5	Yes	Yes (5)	Yes (0)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	50%
6	M/S VERIZON	4	Yes	Yes (5)	Yes (25)	Yes (10)	Yes (10)	No(0)	Yes (5)	Yes (5)	No (0)	Yes	Yes (5)	69%
7	M/S ONTECH CORPORATION	10	Yes	Yes (5)	Yes (40)	Yes (10)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	No (0)	Yes	Yes (5)	95%


  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana

  
 Prof: Shahida Shaikh  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Muhammad Ilyas Shaikh,  
 Principal, Baby Aisla Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Abdul Masood Bhatti,  
 Head of Urology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Dr: Saad Abbasi  
 Medical Superintendent,  
 Chandak Medical College Hospital,  
 Larkana (Chairman)

**Note: M/S BIOS, M/S Quantum & M/S Veritone**  
 provided relevant documents before opening of  
 Financial Bids, hence met evaluation criteria

Annexure-A

**TENDER FOR THE SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENTS/INSTRUMENTS UNDER REVENUE COMPONENTS OF THE SCHEME "UPGRADATION OF UROLOGICAL FACILITIES, CMC HOSPITAL LARKANA NON ADP (2014-15) TENDER DUE DATE 28-APRIL-2015**

**TECHNICAL EVALUATION**

Technical Scrutiny and evaluation of offers received from different firms


S.No.	Name of the Items	Comment Of Technical Committee
1.	High Intensity Focus Ultrasound	<p>One Firm M/s Allmed Solutions has participated in the tender and as per the tender requirement.</p> <p>Hence, Technical Committee has decided that the item may be purchased from above said firm.</p>
2.	Cystoscope & Resectoscope Set Adult	<p>Two Firms M/s Allmed Solutions and M/s EndoKare has participated in the tender. After technical scrutiny it is found that</p> <p>M/s Endokare has not quoted following items - CMOS (3 in 1 system) Monitor TFT 7", Camera , Light source</p> <ul style="list-style-type: none"> <li>- Stand</li> <li>- Plastic container</li> </ul> <p>Hence technically disqualified.</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
3.	Optical Urethrotome Set	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that</p> <p>M/s Endokare has following deficiencies</p> <ul style="list-style-type: none"> <li>- Urethrotome telescope is of 12 degree which is wrong.</li> <li>- No Plasma Effect in electrosurgical unit</li> <li>- No Nerve Stimulator</li> <li>- No Vaporization Effect</li> </ul> <p>Hence technically disqualified.</p> <p>Also M/s Verizon has following deficiencies</p> <ul style="list-style-type: none"> <li>- No External hardware for Bipolar Resection</li> <li>- Resectroscope sheath is not as per tender requirement</li> <li>- Obturator is not as per Tender requirement</li> </ul> <p>Hence technically disqualified.</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>


4.	Light Source 175 w. or more	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that M/s Endokare has following deficiencies</p> <ul style="list-style-type: none"> <li>- Insertion aid is not quoted</li> <li>- instrument port with sealing system with quick release lock not quoted</li> <li>- Luer lock tube connector is not quoted</li> <li>- flow control stop cock is not quoted</li> <li>- tray for sterilization is not quoted</li> <li>- grasping forceps is not quoted</li> <li>- Biopsy forceps is not quoted</li> </ul> <p>Hence technically disqualified.</p> <p>Where has M/s Allmed Solutions and M/s Verizon are as per tender requirement committee has decided that the item may be purchase from the Lowest Bidder</p>
5.	Stone Punch	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that M/s Endokare has following deficiencies</p> <ul style="list-style-type: none"> <li>- Punch working element not quoted</li> <li>- Obturator, silicon, suction tube &amp; irrigation tube are not quoted</li> <li>- grasping and biopsy forceps are not quoted</li> </ul> <p>Hence Considered as technically disqualified</p> <p>Also M/s Verizon has following deficiencies</p> <ul style="list-style-type: none"> <li>- Punch sheath with obturator and silicon suction and irrigation tubes are not quoted.</li> <li>- Instrument port with sealing system and quick release lock are not quoted</li> <li>- minimal sheath diameter is not quoted</li> <li>- seal package of 10 is not quoted</li> </ul> <p>Hence Considered as technically disqualified</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
6.	<p>Percutaneous Nephroscope (PCNL &amp; Micro PCNL)</p> <p style="text-align: center;"><i>h</i></p>	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that M/s Endokare and M/s Verizon has 50% items are not quoted or wrongly quoted Hence both firms are technically disqualified.</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>


7.	Peads Cystoscope & Resectoscope & Urethrotome Set	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that</p> <p>M/s Endokare has following deficiencies  - miniature straight forward telescope is wrongly quoted  - Cystoscopic urethrotome sheaths are not quoted  - other multiple sheaths are not quoted  Hence Considered as technically disqualified</p> <p>M/s Verizon has following deficiencies  - miniature straight forward telescope is not as per tender specifications  - working element set is not quoted.  - Maintenance kits are not quoted  Hence Considered as technically disqualified</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
8.	Holmium Laser	<p>M/s Quantum, M/s Endokare, M/s Allmed Solutions and M/s Verizon has participated in the tender, after technical scrutiny it is found that</p> <p>All firms are technically qualified and may be purchased from the lowest bidder.</p>
9.	Urodynamic System With ENG	<p>M/s Quantum, M/s Endokare, M/s universal enterprises has participated in the tender, after technical scrutiny it is found that</p> <p>M/s Endokare has following deficiencies  - Printer &amp; optional Items are not quoted  Hence Considered as technically disqualified</p> <p>M/s Quantum has following deficiencies  - Printer &amp; optional Items are not quoted  Hence Considered as technically disqualified</p> <p>Where M/s Universal Enterprises is technically qualified hence technical committee has decided that item may be purchased from technically qualified bidder.</p>
10.	Full High Definition Video Camera System.  <i>h</i>	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that</p> <p>M/s Endokare has following deficiencies  - Quoted 26" LCD where as tender requirement is LED 27" screen  - hand instrument's lengths are short as per tender requirement  - Screen is not from same manufacturer  Hence Considered as technically disqualified</p>


		<p>M/s Verizon has following deficiencies</p> <ul style="list-style-type: none"> <li>- Quoted 26" LCD where as tender requirement is LED 27" screen</li> <li>- hand instrument's lengths are short as per tender requirement</li> </ul> <p>Hence Considered as technically disqualified</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
11.	Full High Definition Camera System	<p>Three Firms M/s Allmed Solutions, M/s Endokare and M/s Verizon has participated in the tender after technical scrutiny it is found that</p> <p>M/s Endokare has following deficiencies</p> <ul style="list-style-type: none"> <li>- Camera head is not pendulum type.</li> <li>- Hand pieces and probes are not quoted</li> <li>- air compressor is not quoted.</li> </ul> <p>Quoted 26" LCD where as tender requirement is LED 27" screen</p> <ul style="list-style-type: none"> <li>- Screen is not from same manufacturer</li> </ul> <p>Hence Considered as technically disqualified</p> <p>M/s Verizon has following deficiencies</p> <ul style="list-style-type: none"> <li>- Camera head is not pendulum type.</li> <li>- Quoted 26" LCD where as tender requirement is LED 27" screen</li> <li>- Trolley is not quoted.</li> <li>- Probes are not as per tender specifications.</li> <li>- Air compressor and roller clamps set is not quoted.</li> </ul> <p>Hence Considered as technically disqualified</p> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
12.	Combined Unit Vessel Sealing System with Argon Beamer	<p>Three Firms M/s Ontech, M/s Endokare and M/s BIOS has participated in the tender after technical scrutiny it is found that Only M/S BIOS is technically qualified and technical committee has decided that item may be purchase form technically approved bidder.</p>
13.	Combine Unit Hand Instrument	<p>Two Firms M/s Ontech &amp; M/s Endokare has participated in the tender after technical scrutiny it is found that</p> <p>Only M/S Endo-Kare is technically qualified The Technical Committee has decided that the item may be purchased from above said technically approved</p>
14.	Carm/Fluroscope	<p>Two Firms M/s Endokare &amp; M/s BIOS has participated in the tender after technical scrutiny it is found that</p> <p>M/s BIOS has following deficiency.</p> <ul style="list-style-type: none"> <li>- Items quoted is from India Origin</li> </ul> <p>Where as M/s Endokare is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>


15.	Urology Diode Laser System 180 W. or more.	<p>Three Firms M/s Ontech &amp; M/s Quantum &amp; M/S Endo-Kare have participated in the tender after technical scrutiny it is found that</p> <p>M/s Quantum has following deficiencies.</p> <ul style="list-style-type: none"> <li>- Wavelength 1470 nm is not available.</li> <li>- Agency Agreement is not Embassy Attested.</li> </ul> <p>Hence Considered as technically disqualified</p> <p>M/s Endo-Kare has following deficiencies.</p> <ul style="list-style-type: none"> <li>- Wavelength 1470 nm is not available.</li> <li>- Required fibers are not quoted</li> <li>- Pulse mod is not available.</li> </ul> <p>Hence Considered as technically disqualified</p> <p>Where as M/s Ontech is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
-----	--	---

  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)


  
 Prof: Shahida Shaikh  
 Head of Department of Obts / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Abdul Manan Bhutto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Mohammad Ilyas Shaikh,  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larakana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt, of Sindh, Karachi  
 (Member)

  
 Dr: Safiullah Abbasi  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)








## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015

Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015
3	Tender Description/Name of work/item:	Purchase of Machinery, Equipment / Instruments SZW Hospital under ADP Scheme No. 625 (2014-2015)
4	Method of Procurement:	Single Stage - two Envelope Procedure
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. 1575219283 Upload on 15-4-2015 ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015
6	Total Bid documents Sold:	12
7	Total Bids Received:	12
8	Technical Bid Opening date: (if applicable) (Provide date in separate form)	28-04-2015
9	No. of Bid technically qualified (if applicable):	12
10	Bid(s) Rejected:	Nil
11	Financial Bid Opening date:	05-05-2015
12	<b>Bid Evaluation Report:</b>	

S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
1	2	3	4	5	6	7
		Details are @ Annexure-D				


(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt, of Sindh, Karachi (Member)	
(Prof. Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof. Shahida Shaikh) HOD Obts/ Gynae, SMBBMU, Larkana (Member)	
(Prof. Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof. Mohammad Ilyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr. Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkana	

Annexure - A.

Supply, Installation, Testing, Commissioning & maintenance (During maintenance period) of Medical Equipments / Instruments under ADP Scheme viz: Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

Bid Evaluation

S#	Name Of Bidder	Name of The Firm and its Establishment	Original Tender Purchase Receipt	Sale Tax (NTN) Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit i.e. Firm has never been Black Listed	Registration with Various Dept.	Annual Turnover Report	Bid is signed & Stamped by Authorize Person Along with Authorization Letter	Authentic / Verifiably Manufacturer's Authority Letter	Bid Security Attached	REMARKS
1	M/S. Lab Link	8	YEs	Yes (5)	Yes (35)	Yes (7)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	91%
2	M/S BIOS	10	YEs	Yes (5)	Yes (30)	Yes (5)	Yes (10)	Yes (5)	No (0)	Yes (5)	Yes (5)	Yes	Yes (5)	80%
3	M/S ENDO-GAME	10	YEs	Yes (5)	Yes (35)	Yes (10)	Yes (8)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	93%
4	M/S Hospicare	10	YEs	Yes (5)	Yes (40)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	95%
5	M/S VERIZON													
6	M/S Pak Traders	5	YEs	Yes (5)	No (0)	No (0)	No(0)	No(0)	Yes (5)	No(0)	No (0)	No (0)	Yes (5)	20%
7	M/S Noor International	7	YEs	Yes (5)	No (35)	No (6)	No(7)	Yes (5)	Yes (5)	No(0)	Yes (5)	Yes	Yes (5)	80%
8	M/S Humza Enterprises	8	YEs	Yes (5)	No (30)	No (7)	No(8)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	85
9	M/S Elite	8	YEs	Yes (5)	No (35)	No (8)	No(8)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	91%
10	M/S Onitech	10	Yes	Yes (5)	Yes (40)	Yes (10)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	No (0)	Yes	Yes(5)	
11	M/S Neuson	10	Yes	Yes (5)	?	?	?	Yes (5)	Yes (5)	?	Yes (5)	Yes	Yes (5)	35%
12	M/S Allied solutions	10	Yes	Yes (5)	Yes (40)	Yes (5)	Yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	95%


  
 Prof. Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana


  
 Prof. Shahida Shaukh  
 Head of Department of OBG / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Mohammad Iftikhar  
 Principal, Babu Ahsan Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Abdul Munir Memon  
 Head of Urology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr. Agha Hameed  
 Additional Medical Superintendent (Tech)  
 CIVIC Hospital, Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt. of Sindh, Karachi  
 (Member)

  
 Dr. Subhan Ullah  
 Medical Superintendent,  
 Shaheed Mohtarma Medical University Hospital,  
 Larkana (Chairman)

**Note:** M/S Pak Traders & M/S Neuson provided relevant documents before opening of Financial Bids, hence met evaluation criteria



Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of Medical Equipments / Instruments under ADP  
Scheme viz;  
Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

TECHNICAL EVALUATION

Technical Scrutiny and evaluation of offers received from different firms

S.No.	Name of the Items	Comment Of Technical Committee
1.	HYSTEROSCOPY	<p>Five Firms M/s Lab Link, M/s Endocare, M/s Verizon, M/s Noor International, M/s Neucon &amp; M/s Allmed Solutions have participated in the tender. After technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Lab Link has the following deficiencies, Following items are not quoted. <ul style="list-style-type: none"> <li>- Operating Sheath Adapter.</li> <li>- CO2 Bottle with pin index connection.</li> </ul> </li> </ul> <p>Following items are wrongly quoted</p> <ul style="list-style-type: none"> <li>- Operating Sheath's length is short and not as per tender.</li> <li>- Biopsy &amp; Grasping Forceps are required whereas Alligator Forceps is quoted.</li> <li>- Required Vacuum Cannula, whereas Probes are quoted.</li> <li>- Required Optical Scissor, whereas Hook scissor is quoted.</li> <li>- Hysteroflator is required, whereas Endoflator is quoted.</li> </ul> <p>Hence technically disqualified.</p> <ul style="list-style-type: none"> <li>• M/s Endocare has the following deficiencies, Following items are not quoted. <ul style="list-style-type: none"> <li>- Operating Sheath Adapter.</li> <li>- Suction Catheter Pack of 06</li> <li>- Hysteroflator's filters pack of 10 are not quoted.</li> <li>- Co2 Bottle with Pin Index is not quoted.</li> </ul> </li> </ul> <p>Following items are wrongly quoted.</p> <ul style="list-style-type: none"> <li>- 5Fr Instrument Channel is required, whereas 3Fr Channel is quoted</li> <li>- Micro Hysteroflator is required, Quoted Simple Hyteroflator.</li> <li>- 100 ml/min is required, whereas 50 ml/min is quoted.</li> <li>- Integrated SCB Module is required, whereas Simple is quoted.</li> </ul> <p>Hence technically disqualified.</p> <ul style="list-style-type: none"> <li>• M/s Verizon has the following deficiencies Following items are not quoted</li> </ul>

W

		<p>- Operating Sheath Adapter is not quoted.  - Suction Catheter pack of 06 is not quoted.</p> <p>Following items are wrongly quoted.  - Blunt Scissor is required, whereas Hook scissor is quoted  - Micro Hysteroflator is required, whereas simple Hystero pump is quoted.  - Integrated SCB Module is required, whereas simple us quoted.  Hence technically disqualified.</p> <ul style="list-style-type: none"> <li>• M/s Noor International has the following deficiencies  Following items are not quoted  - Co2 Bottle with Pin Index is not quoted.</li> </ul> <p>Following items are wrongly quoted.  - Biopsy &amp; Grasping Forceps are required whereas Uretro-  renoscope is quoted.  - optical Scissor is required, whereas Hook scissor is quoted  - Hysteroflator is required, whereas suction &amp; irrigation pump is  quoted.  Hence technically disqualified.</p> <ul style="list-style-type: none"> <li>• M/s Neucon has the following deficiencies  - No Brochure, No Agency agreement attached,  Hence technically disqualified.</li> </ul> <p>Where as M/s Allmed Solutions is as per tender specifications.  The Technical Committee has decided that the item may be  purchased from above said technically approved firm.</p>
2.	Laparoscope Basic Laparoscope set with single chip Camera System	<p>Eight Firms M/s Lab Link, M/s Endocare, M/s Verizon, M/s Noor  International, M/s Neucon, M/s Hamza, M/s Pak Traders &amp; M/s  Allmed Solutions and M/s EndoKare have participated in the tender.  After technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Lab Link has the following deficiencies,  Following items are wrongly quoted.  - Length of Trocars and Forceps are not as per tender specifications.  - Electronic Endoflator with Intergrated SCB Module is not quoted.</li> </ul> <p>Hence technically disqualified</p> <p style="text-align: center;">w</p>

		<ul style="list-style-type: none"> <li>• M/s Endocare has the following deficiencies, Following items are wrongly quoted. <ul style="list-style-type: none"> <li>- Length of Trocars and Forceps are not as per tender specifications.</li> <li>- Electronic Endoflator with Intergrated SCB Module is not quoted.</li> </ul> Hence technically disqualified. </li>   <li>• M/s Verizon has the following deficiencies Following items are not quoted <ul style="list-style-type: none"> <li>- Filters pack of 10 of Endoflator is not quoted.</li> </ul> Following items are wrongly quoted. <ul style="list-style-type: none"> <li>- Length of Trocars and Forceps are not as per tender specifications.</li> <li>- Electronic Endoflator with Intergrated SCB Module is not quoted.</li> </ul> Hence technically disqualified. </li>   <li>• M/s Noor International has the following deficiencies Following items are not quoted <ul style="list-style-type: none"> <li>- Filters pack of 10 of Endoflator is not quoted.</li> </ul> Following items are wrongly quoted. <ul style="list-style-type: none"> <li>- Length of Trocars and Forceps are not as per tender specifications.</li> <li>- Electronic Endoflator with Intergrated SCB Module is not quoted.</li> </ul> Hence technically disqualified. </li>   <li>• M/s Hamza Enterprises has the following deficiencies Following items are wrongly quoted. <ul style="list-style-type: none"> <li>- 0 degree telescope is required, whereas 30 degree is quoted</li> <li>- Length of Trocars and Forceps are not as per tender specifications</li> <li>- Electronic Endoflator with Intergrated SCB Module is not quoted.</li> </ul> Hence technically disqualified. </li>   <li>• M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li>   <li>• M/s Neucon is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li> </ul> <p>Where as M/s Allmed Solutions is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
--	--	--

W


3.	O.T Light	<p>Three Firms M/s Ontech, M/s Elate and M/s Lab Link has participated in the tender after technical scrutiny it is found that</p> <p>M/s Ontech has following deficiencies -Sole Agency letter not attached.</p> <p>Where as M/s Elate &amp; M/s Lab Link are as per tender specifications. The Technical Committee has decided that the item may be purchased from the lowest Bidder.</p>
4.	Pap Smear Kit	No Firm has participated in item # 4
5.	Pulse Oximeter	<p>Three Firms M/s Hospicare, M/s Lab Link and M/s Hamza Enterprises has participated in the tender after technical scrutiny it is found that</p> <p>M/s Hospicare has quoted Chinese Product which is not acceptable as per tender</p> <p>Hence Considered as technically disqualified</p> <p>M/s Hamza Enterprises has quoted Korean Product which is not acceptable as per tender</p> <p>Hence Considered as technically disqualified</p> <p>Where as M/s Lab Link is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.</p>
6.	SUCTION MACHINE	<p>Four Firms M/s Lab Link, M/s BIOS and M/s Pak Traders &amp; M/s Hamza Enterprises has participated in the tender after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li> </ul> <p>Where as M/s Lab Link, M/s BIOS &amp; M/s Hamza Enterprises are as per tender specifications. The Technical Committee has decided that the item may be purchased from the lowest Bidder.</p>
7.	Anaesthesia Machine	Four Firms M/s Lab Link, M/s BIOS, M/s Hamza Enterprises & M/s Elate participated in the tender after technical scrutiny it is found that

		All Firms are technically qualified, The Technical Committee has decided that the item may be purchased from the lowest Bidder.
8.	Diathermy Machine	<p>Five Firms M/s Lab Link, M/s BIOS, M/s Hospicare, M/s Pak Traders &amp; M/s Ontech has participated in the tender, after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Hospicare has quoted Vessel Sealer, Where electrosurgical unit is required Hence Technically disqualified</li> <li>• M/s BIOS has quoted Korean origin. Hence Technically disqualified.</li> <li>• M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li> <li>• M/s Ontech has quoted Vessel Sealer, Where electrosurgical unit is required Hence Technically disqualified</li> </ul> <p>Where as M/s Lab Link is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
9.	Autoclave  h	<p>Four firms M/s Lab Link, M/s BIOS, M/s Hospicare, &amp; M/s Pak Traders has participated in the tender, after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li> <li>• M/s BIOS has quoted Taiwan origin. Hence Technically disqualified.</li> <li>• M/s lab Link has not quoted Chamber size as per tender specifications.</li> </ul> <p>Where as M/s Hospicare is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
10.	PV Light	One Firm M/s Pak Traders has participated in the tender and


		<p>technically approved.</p> <p>The Technical Committee has decided that the item may be purchased from above said technically approved</p>
11.	Sonicate	<p>Two Firms M/s Lab Link &amp; M/s Hamza Enterprises has participated in the tender after technical scrutiny it is found that</p> <p>Both Firms has quoted same brands. So, The Technical Committee has decided that the item may be purchased from the firm who will produce Agency Agreement.</p>
12.	Ultrasound Machine	<p>Three Firms M/s Lab Link, M/s BIOS and M/s Hamza Enterprises has participated in the tender after technical scrutiny it is found that</p> <p>M/s BIOS and M/s Hamza Enterprises has quoted items from China, Hence Technically disqualified.</p> <p>Whereas M/s Lab Link is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
13.	VACUM EXTRACTOR	<p>Four Firms M/s Lab Link, M/s BIOS, M/s Hamza Enterprises &amp; M/s Pak Traders has participated in the tender after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li> </ul> <p>Where as M/s Lab Link, M/s BIOS, M/s Hamza Enterprises are as per tender specifications. The Technical Committee has decided that the item may be purchased from Lowest bidder.</p>
14.	Delivery Table	<p>Two Firms M/s Lab Link &amp; M/s Pak Traders has participated in the tender after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified.</li> </ul> <p>Where as M/s Lab Link is as per tender specifications. The Technical Committee has decided that the item may be</p>

		purchased from above said technically approved
15.	Neonatal Resuscitation Trolley	<p>Two Firms M/s Hospicare &amp; M/s Hamza Enterprises has participated in the tender after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Hospicare has quoted China origin. Hence Technically disqualified.</li> </ul> <p>Where as M/s Hamza Enterprises is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved</p>
16.	Multipurpose Monitor	<p>Three Firms M/s Lab Link, M/s BIOS &amp; M/s Hospicare has participated after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Hospicare has quoted China origin. Hence Technically disqualified.</li> </ul> <p>M/s Lab Link &amp; M/s BIOS both are technically approved The Technical Committee has decided that the item may be purchased from Lowest bidder</p>
17.	Ventilators	<p>Four Firms Firm M/s Lab Link, M/s BIOS, M/s Hamza Enterprises &amp; M/s Elate has participated in the tender after technical scrutiny it is found that M/S Elate has quoted ICU Ventilator where as Anesthesia Ventilator is required Hence technically disqualified</p> <p>Three firms are technically approved. The Technical Committee has decided that the item may be purchased from Lowest bidder.</p>
18.	DEFIBRILLATOR	<p>Four Firms M/s Lab Link, M/s BIOS, M/s Hamza Enterprises &amp; M/s Elate has participated in the tender after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s Lab Link is not as per technical requirements Hence Technically disqualify</li> <li>• M/s Hamza Enterprises has China Origin and also its energy range is not meeting tender requirements.</li> </ul> <p>M/s BIOS and M/s Elate are technically approved. The Technical Committee has decided that the item may be purchased from Lowest bidder.</p>


19.	C.T.G MACHINE	<p>Three Firms M/s Lab Link, M/s BIOS, M/s Hamza Enterprises has participated in the tender after technical scrutiny it is found that</p> <ul style="list-style-type: none"> <li>• M/s BIOS has quoted China origin. Hence Technically disqualified.</li> </ul> <p>M/s Lab Link &amp; M/s Hamza Enterprises has technically approved. The Technical Committee has decided that the item may be purchased from Lowest bidder.</p>
-----	---------------	--



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obts / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Abdul Manan Bhutto,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Mohammad Illyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aijaz Ahmed Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larkana  
(Member)



Special Secretary (Public Health)  
Health Department, Govt, of Sindh, Karachi  
(Member)










Dr. Saifullah Abbasi  
Medical Superintendent,  
Chandka Medical College Hospital,  
Larkana (Chairman)



## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015  
Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015
3	Tender Description/Name of work/item:	Purchase of Machinery & Equipments / Instrument for Dental Dept. ADP Scheme No.623 (2014-2015)
4	Method of Procurement:	Single Stage - two Envelope Procedure
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. <b>1575219283</b> <i>upload on 15-4-2015</i> ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015
6	Total Bid documents Sold:	03
7	Total Bids Received:	03
8	Technical Bid Opening date: (if applicable) <i>(Provide date in separate form)</i>	28-04-2015
9	No. of Bid technically qualified (if applicable):	03
10	Bid(s) Rejected:	Nil
11	Financial Bid Opening date:	05-05-2015
12	<b>Bid Evaluation Report:</b>	
S No	Name of Firm or Bidder	Cost offered by the Bidder
1	2	3
4	Ranking in terms of cost	Comparison with Estimated cost
5	Reasons for acceptance/rejection	Remarks
Details are @ Annexure-B		

(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt. of Sindh, Karachi (Member)	
(Prof. Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof. Shahida Shaikh) HOD Obs/ Gynae, SMBBMU, Larkana (Member)	
(Prof. Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof. Mohammad Ilyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr. Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkana	

Supply, Installation, Testing, Commissioning & maintenance (During maintenance period) of Dental Equipments / Instruments under ADP Schemes viz:

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623 (2014-15)

Annexure - A.


Bid Evaluation

Sl#	Name Of Bidder	Name of The Firm an Its Establishment	Original Tender Purchase Receipt	Sale Tax /NTN Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit I.e. Firm has Never Been Black Listed	Registration with Various Depas.	Annual Turnover Report	Bid is signed & stamped by Authorize Person Along with Authorization Letter	Authentic /Verifiable Manufacturer's Authority Letter	Bid Security Attached	REMARKS
1	M/s Platinum corporation	10	Yes	Yes (5)	Yes (35)	Yes (8)	yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	93%
2	M/S Pak Traders	5	YES	Yes (5)	No (0)	No (0)	No(0)	No(0)	Yes (5)	No(0)	No (0)	No (0)	Yes (5)	20%
3														
5	M/s BIOS	10	Yes	Yes (5)	No (35)	No (10)	No(8)	Yes (5)	Yes (5)	No(0)	Yes (5)	Yes	Yes (5)	88%


  
 Prof. Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Shahida Shaikh  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Mohammad Ilyas Shaikh,  
 Principal, Babu Aista Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Abdul Wahid Bhatti  
 Head of Dermatology Department  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr. Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt. of Sindh, Karachi  
 (Member)

  
 Dr. Shafiqul Abbas  
 Medical Superintendent,  
 CMC Medical College Hospital,  
 Larkana (Chairman)

**Note:** M/S Pak Traders provided relevant documents before opening of Financial Bids, hence met evaluation criteria

*Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of Dental Equipments /  
Instruments under ADP Schemes viz;*

**Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction  
of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623  
(2014-15)**


## Technical evaluation


Technical Scrutiny and evaluation of offers received from different firms


S.No.	Name of the Items	Comment Of Technical Committee
1.	OPG X-Ray Unit	Three Firms M/s Platinum corporation, M/S Pak Trader & M/s BIOS has participated in the tender after technical scrutiny it is found that All Firms are technically qualified and may be purchased from lowest bidder.
2.	Dental X-Ray Unit	Three M/s Platinum corporation, M/S Pak Trader & M/s BIOS has participated in the tender after technical scrutiny it is found that M/s Pak Traders does not Comply technically Hence, technically disqualify  M/s Platinum corporatio & M/s BIOS are technically qualified and may be purchased from lowest bidder.
3.	OT Table	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that  M/s Platinum corporation does not Comply technically Hence, technically disqualify.  Where as M/s Pak Traders is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
4.	Surgical Halogen Light  h	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that  M/s Pak Traders does not Comply technically Hence, technically disqualify.  Where as M/s Platinum corporation is as per tender specifications.


		The Technical Committee has decided that the item may be purchased from above said technically approved firm.
5.	New Q-Switched Nd:YAG Laser - LUCAS PLUS	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that M/s Platinum corporation & M/s BIOS are technically qualified and may be purchased from lowest bidder.
6.	Auto Clave for Dental Instrument Medium Size	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that M/s Pak Traders does not Comply technically Hence, technically disqualify.  Where as M/s Platinum corporation is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
7.	Dental Implant Unit	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that M/s Pak Traders does not Comply technically Hence, technically disqualify.  Where as M/s Platinum corporation is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.
8.	Dental Surgical Instrument	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that  Both Firms are technically qualified and may be purchased from lowest bidder.
9.	OPG Auto Laser Machine	One firm M/s Platinum corporation has participated in the tender after technical scrutiny it is found that The Firms is technically qualified and may be purchased
10.	Intra Oral Digital Imaging X-Ray System	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that  Both Firms are technically qualified and may be purchased from lowest bidder.
11.	Phantom Head Simulator	Two Firms M/s Platinum corporation & M/s Pak Traders has participated in the tender after technical scrutiny it is found that  Both Firms are technically qualified and may be purchased from


		lowest bidder.
12.	Diathermy	Three M/s Platinum corporation, M/S Pak Trader & M/s BIOS has participated in the tender after technical scrutiny it is found that All Firms are technically qualified and may be purchased from lowest bidder.
13.	Suction Machine	Three M/s Platinum corporation, M/S Pak Trader & M/s BIOS has participated in the tender after technical scrutiny it is found that All Firms are technically qualified and may be purchased from lowest bidder.

  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)


  
 Prof: Shahida Shaikh  
 Head of Department of Obts / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Abdul Manan Bhutto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Mohammad Ilyas Shaikh,  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt, of Sindh, Karachi  
 (Member)








  
 Dr Saifullah Abbasi  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)

## Bid Evaluation Report

NIT No: CMCHL/(P&D):/104 dated: 30-03-2015

Chandka Medical College Hospital, Larkana

1	Name of Procuring Agency:	Medical Superintendent, Chandka Medical College Hospital, Larkana				
2	Tender Reference No:	NIT No: CMCHL/(P&D):/104 dated: 30-03-2015				
3	Tender Description/Name of work/Item:	Purchase of 100 KV Generator under ADP Scheme No.623 (2014-2015)				
4	Method of Procurement:	Single Stage - two Envelope Procedure				
5	Tender Published: <i>Print &amp; Electronic Media (SPPRA ID No. &amp; News papers names with dates)</i>	i) SPPRA ID No. <b>1575219283</b> upload on <b>15-4-2015</b> ii) Daily Express Tribune, dated 11-04-2015 iii) Daily Sobh, dated 11-04-2015 iv) Daily Kaleem, Sukkur, dated 11-04-2015				
6	Total Bid documents Sold:	02				
7	Total Bids Received:	02				
8	Technical Bid Opening date: (if applicable) <i>(Provide details in separate form)</i>	28-04-2015				
9	No. of Bid technically qualified (if applicable):	02				
10	Bid(s) Rejected:	Nil				
11	Financial Bid Opening date:	05-05-2015				
12	<b>Bid Evaluation Report:</b>					
S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
1	2	3	4	5	6	7
		Details are @ Annexure-C				

(Dr. Safiullah Abbasi) Medical Superintendent/Chairman Procurement Committee C. M. C Hospital, Larkana	
(Dr. Khalid Shaikh) Special Secretary (P. H), Health Department, Govt. of Sindh, Karachi (Member)	
(Prof. Malik Hussain Jalbani), HOD Urology SMBBMU, Larkana (Member)	
(Prof. Shahida Shaikh) HOD Obs/ Gynae, SMBBMU, Larkana (Member)	
(Prof. Abdul Manan Bhutto) HOD Dermatology Department, SMBBMU, Larkana (Member)	
(Prof. Mohammad Ilyas Shaikh) Principal, Baby Asifa Dental College, SMBBMU, Larkana (Member)	
(Dr. Aijaz Ahmed Shaikh) Additional Medical Superintendent (Tech), CMC Hospital Larkana	


Answer - C.

Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of 100 KV Generator with Auto Start facility under ADP Schemes viz:


Establishment of Shahzeed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana  
ADP No. 623 (2014-15)

COMPARATIVE STATEMENT FINANCIAL BID DATED: 05/05/2015

S #	Quoted Item	Qty	Bidder	Bidder
1	Generator 100 KV	1	Pak Traders	Lab Link
			3,224,200	3,300,000

  
Prof: Malik Hussain Jahani  
Head of Department of Urology,  
Shahzeed Mohiarnma Medical University, Larkana  
(Member)

  
Prof: Shahida Shaikh  
Head of Department of Obs / Gyne,  
Shahzeed Mohiarnma Medical University, Larkana  
(Member)

  
Prof: Mohtajummad Ilyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shahzeed Mohiarnma Medical University,  
Larkana  
(Member)

  
Prof: Sohail Adnan Bhutto,  
Head of Permatology Department,  
Shahzeed Mohiarnma Medical University, Larkana  
(Member)

  
Dr: Aijaz Ahmed Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larkana  
(Member)

  
Special Secretary (Public Health)  
Health Department, Govt. of Sindh, Karachi  
(Member)


  
Dr: Saad Malik Abbasi  
Medical Superintendent,  
Charudai Medical College Hospital,  
Larkana (Chairman)


Supply, Installation, Testing, Commissioning & maintenance (During maintenance period) of 100 KV Generator with Auto Start facility under ADP Schemes viz:

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623 (2014-15)

Technical evaluation

Sl	Name Of Bidder	Name of The Firm an Its Establishment	Original Tender Purchase Receipt	Sale Tax /NTN Certificate	Past Experience of Firm	Tools & Parts list Availability	Technical Staff / Trained Engineer list	Affidavit I.e. Firm has Never been Black Listed	Registration with Various Dept.	Annual Turnover Report	Bid is signed & Stamped by Authorize Person Along with Authorization Letter	Authentic /Verifiable Manufacturer's Authority Letter	Bid Security Attached	REMARKS
1	M/s. Lab Link	8	YES	Yes (5)	Yes (35)	Yes (7)	yes (10)	Yes (5)	Yes (5)	Yes (5)	Yes (5)	Yes	Yes (5)	91%
2	M/S Pak Traders	5	YES	Yes (5)	No (0)	No (0)	No(0)	No(0)	Yes (5)	No(0)	No (0)	No (0)	Yes (5)	20%

  
 Prof. Mubik Kussain Jabbari  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Shahida Shaikh  
 Head of Department of Otic / Gyne,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Mohammad Ilyas Shaikh,  
 Principal, Baby Aista Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof. Akbar Hussain Bhutto  
 Head of Otolaryngology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr. Aijaz Rehman Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt. of Sindh, Karachi  
 (Member)

  
 Dr. Sarfaraz Khobai  
 Medical Superintendent,  
 CMC Medical College Hospital,  
 Larkana (Chairman)

**Note: M/S Pak Traders provided relevant documents before opening of Financial Bids, hence met evaluation criteria**



Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of 100 KV Generator with Auto Start facility under ADP Schemes viz;

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology  
Department at CMC Hospital Larkana ADP No. 623 (2014-15)

### Technical evaluation

Technical Scrutiny and evaluation of offers received from different firms

S.No.	Name of the Items	Comment Of Technical Committee
1.	<b>Generator 100 KV</b>	Two Firms M/s Lab Link Enterprises and M/s Pak Traders has participated in the tender and as per the tender requirement. M/s Pak Traders is failed to comply Bid Evaluation Criteria Hence technically disqualified. Where as M/S Lab Link Enterprises is as per tender specifications. The Technical Committee has decided that the item may be purchased from above said technically approved firm.



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obts / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)




Prof: Abdul Manan Bhutto,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Mohammad Ilyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aijaz Ahmed Shaikh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larakana  
(Member)



Special Secretary (Public Health)  
Health Department, Govt, of Sindh, Karachi  
(Member)



Dr: Saifullah Abbasi  
Medical Superintendent,  
Chandka Medical College Hospital,  
Larkana (Chairman)

Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of Skin Disease Complex and Hepatology Department Equipments / Instruments under ADP Scheme  
viz;

Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana  
ADP No. 623 (2014-15)

COMPARATIVE STATEMENT FINANCIAL BID DATED: 05/05/2015

S#	Quoted Item	Qty	Bidder	Bidder
	<b>Skin Diseases Complex</b>			
1	Trinocular Video Microscopy Systems with Digital Camera	2	Business Medical International N.Q	Quantum 1,417,808
3	IPL&Diode Laser	2	Business Medical International N.Q	Quantum 7,065,410
4	Top Advanced CO2 Fractional Laser – COFRAX	2	Business Medical International 2,782,000	Quantum 3,972,602
5	New Q-Switched Nd:YAG Laser - LUCAS PLUS	2	Business Medical International N.Q	Quantum 5,326,232
6	Aesthipeel	2	Business Medical International N.Q	Quantum 835,616
7	Ultra Peel System	1	Business Medical International 397,646	Quantum 1,562,301
8	PUVA Unit	2	Business Medical International 1,516,287	Quantum 3,958,904
9	Dermalight 80, Phototherapy Device	2	Business Medical International 588,000	Quantum 301,369
10	AFT Laser System	1	Business Medical International N.Q	Quantum 6,716,712




S #	Quoted Item	Qty	Bidder	Bidder
	<b>Hepatology Department</b>			
1	Flow Meter	6		
2	Oxygen Cylinder	6		
3	Respirator	1		
4	Blood Band Refrigerator with Stabilizer	2		
7	Centrifuge Machine (Digital)	2		
Not Quoted				


  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Shahida Shaikh  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Mohammad Illyas Shaikh,  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University,  
 Larkana  
 (Member)

  
 Prof: Abdul Mannan Bhutto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt, of Sindh, Karachi  
 (Member)

  
 Dr: Saqibullah Abbasi  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)

Supply, Installation, Testing, Commissioning  
& maintenance (During maintenance period) of Medical Equipments / Instruments under ADP Scheme viz:  
Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

Annexure - D.

COMPARATIVE STATEMENT FINANCIAL BID DATED: 05/05/2015

S #	Quoted Item	Qty	Bidder				
			Bidder	Bidder	Bidder	Bidder	Bidder
1	HYSTEROSCOPY	1	Allmed Solution	Noor	Neocon	Lab Link	Endo-Kare
			1,799,190	Tech: Rej	Tech: Rej	Tech: Rej	Tech: Rej
2	Laparoscope Basic Laparoscope set with single chip Camera System	1	Allmed Solution	Hannza Ent:	Verizone	Pak Traders	
			3,768,605	Tech: Rej	Tech: Rej	Tech: Rej	5,385,000
			Ontech	Elate	Lab Link		
3	O.T Light LED	1	Tech: Rej	1,430,000	1,364,000		
4	Pap Smear Kit	15	Not Quoted from Any One Bidder				
5	Pulse Oximeter	8	Hospi Care	Lab Link	Hannza Ent:		
			Tech: Rej	85,000	Tech: Rej		
6	SUCTION MACHINE	4	Lab Link	BIOS	Pak Traders		
			92,000	142,000	135000		
7	Anaesthesia Machine	1	Lab Link	BIOS	Hannza Ent:	Elate	
			1,150,000	1,430,000	2,114,000	2,075,000	






S #	Quoted Item	Qty	Bidder	Bidder	Bidder	Bidder	
8	Diathermy Machine	4	Lab Link	BIOS	Hospi Care	Pak Traders	Ontech
			575,000	Tech: Rej	Tech: Rej	635,000	Tech: Rej
9	Autoclave	2	Lab Link	BIOS	Hospi Care	Pak Traders	
			Tech: Rej	Tech: Rej	900,000	Tech: Rej	
10	PV Light	2	Pak Traders				
			185,000				
11	Sonicade	4	Lab Link	Hamza Ent:			
			160,000	145,000			
12	Ultrasound Machine	2	Lab Link	BIOS	Hamza Ent:		
			1,110,000	Tech: Rej	Tech: Rej		
13	VACUM EXTRACTOR	2	Lab Link	Pak Traders	BIOS	Hamza Ent:	
			92,000	155,000	165,000	160,000	
14	Delivery Table	12	Lab Link	Pak Traders			
			226,000	135,000			
15	Neonatal Resuscitation Trolley	8	Hospi Care	Hamza Ent:			
			Tech: Rej	1,560,000			
16	Multipurpose Monitor	6	Lab Link	BIOS	Hospi Care		
			195,000	345,000	Tech: Rej		
17	Ventilators	2	Elare	Lab Link	Hamza Ent:	BIOS	
			Tech: Rej	950,000	1,650,000	499,000	


*Handwritten signature*


*Handwritten signature*


*Handwritten signature*

S #	Quoted Item	Qty	Bidder	Bidder	Bidder	Bidder
18	DEFIBRILLATOR	2	Lab Link Tech: Rej	Hamza Ent: Tech: Rej	BIOS 550,000	Elate 675,000
19	C.T.G MACHINE	4	BIOS Tech: Rej	Lab Link 140,000	Hamza Ent: Ent: 900,000	


  
**Prof: Maalik Hussain Jalbani**  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
**Prof: Shahida Shaikh**  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
**Prof: Mohammad Ilyas Shaikh,**  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
**Prof: Abdul Manaq-Bhutto,**  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
**Dr: Aijaz Ahmed Shaikh,**  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larakana  
 (Member)

  
**Special Secretary (Public Health)**  
 Health Department, Govt, of Sindh, Karachi  
 (Member)

  
**Dr: Saqatullah Abbasi**  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)

**Annexure - E**

**TENDER FOR THE SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENTS/INSTRUMENTS UNDER REVENUE COMPONENTS OF THE SCHEME "UPGRADATION OF UROLOGICAL FACILITIES, CMC HOSPITAL LARKANA NON ADP (2014-15) TENDER DUE**

DATE 28-APRIL-2015

Comparative Statement Financial Bid Dated: 05/05/2015

S #	Quoted Item	Qty	Bidder	Bidder	Bidder	Bidder
1	High Intensity Focus Ultrasound	1	Allmed Solution			
			175,450,000			
2	Cystoscope & Ressectoscope Set Adult	1	Allmed Solution	Endo-Kare		
			2,373,804	Tec:Rej		
3	Optical Urethrotome Set	1	Allmed Solution	Verizone	Endo-Kare	
			3,318,626	Tech: Rej	Tech: Rej	
4	Light Source 175 w. or more	1	Allmed Solution	Verizone	Endo-Kare	
			1,200,000	Tech: Rej	Tech: Rej	
5	Stone Punch	1	Allmed Solution	Verizone	Endo-Kare	
			960,084	Tech: Rej	Tech: Rej	




S #	Quoted Item	Qty	Bidder	Bidder	Bidder	Bidder
6	Percutaneous Nephroscope (PCNL & Micro PCNL)	1	Allmed Solution	Verizone	Endo-Kare	
			34,101,190	Tech: Rej	Tech: Rej	
7	Peads Cystoscope & Resectoscope & Urethrotome Set	1	Allmed Solution	Verizone	Endo-Kare	
			2,100,000	Tech: Rej	Tech: Rej	
			Allmed Solution	Verizone	Endo-Kare	Quantum
8	Holmium Laser	1	3,900,000	5,456,375	6,287,000	6,047,945
			Quantum	Endo-Kare	Universal	
9	Urodynamic System With ENG	1	Tech: Rej	Tech: Rej	2,526,800	
			Allmed Solution	Verizone	Endo-Kare	
10	Full High Definition Video Camera System.	1	48,048,000	Tech: Rej	Tech: Rej	
			Allmed Solution	Verizone	Endo-Kare	
11	Full High Definition Camera System	1	12,541,815	Tech: Rej	Tech: Rej	
			Allmed Solution	Verizone	Endo-Kare	
12	Combined Unit Vessel Sealing System with Argon Beamer	1	Ontech Corp:	Endo-Kare	Bios	
			3,657,486	3,937,000	3,200,000	
13	Combine Unit Hand Instrument	1	Ontech Corp:	Endo-Kare		
			Tech: Rej	1,383,000		





S #	Quoted Item	Qty	Bidder	Bidder	Bidder	Bidder
14	C. arm/Fluroscope	1	Endo-Kare 8,200,000	BIOS Tech: Rej		
15	Urology Diode Laser System 180 W. or more.	1	Ontech Corp: 16,280,000	Quantum Tech: Rej	Endo-Kare Tech: Rej	

  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Shahida Shaikh  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Mohammad Ilyas Shaikh,  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Abdul Merman Bhutto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larakana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt, of Sindh, Karachi  
 (Member)

  
 Dr: Safiullah Abbasi  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)

Annexure - G.


Supply of Other Items under ADP Schemes viz;  
 Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623 (2014-15)  
 Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)


COMPARATIVE STATEMENT FINANCIAL BID DATED: 05/05/2015


S #	Quoted Item	Qty	Bidder	Bidder	Bidder
1	Computer (HP/Dell or equivalent with Laser Printer alongwith LED	2	Platinum 124,000	Pak Traders 127,000	
2	Baby Cots	60	Platinum 82,000	Pak Traders 18,400	
3	Radiant Heater for Neonates	8		Pak Traders 354,000	
4	Central Oxygen system	Bed complete Job	Platinum 35,000	Pak Traders 7,700	
5	Air Conditioners 1.5 Ton with stabilizer	4	Platinum 75,000	Pak Traders 74,000	Tooba 96,311
6	Air-conditioner windows type 1.5 tons with stabilizer	8	Platinum 93,000	Pak Traders 98,000	Tooba 109,980
7	Electric water cooler (Export quality)	2	Platinum 38,000	Pak Traders 48,000	Tooba 58,000
8	Refrigerator (Export quality)	1	Platinum 52,000	Pak Traders 58,000	Tooba 64,428

Qin Am [Signature] Si 29- [Signature] k


S #	Quoted Item	Qty	Bidder	Bidder	Bidder
9	T.V sets (Export quality)	2	Platinum	Pak Traders	Tooba
			38,000	48,000	21,496
10	Multimedia with Screen (Latest)	1	Platinum	Pak Traders	
			98,000	110,000	
11	Lap Top	1	Platinum	Pak Traders	
			110,000	113,000	


  
 Prof: Malik Hussain Jalbani  
 Head of Department of Urology,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)


  
 Prof: Abdet Mannan Bhutto,  
 Head of Dermatology Department,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Prof: Shahida Shaikh  
 Head of Department of Obs / Gynae,  
 Shaheed Mohtarma Medical University, Larkana  
 (Member)

  
 Dr: Aijaz Ahmed Shaikh,  
 Additional Medical Superintendent (Tech)  
 CMC Hospital, Larkana  
 (Member)

  
 Prof: Mohammad Ilyas Shaikh,  
 Principal, Baby Aisfa Dental College,  
 Shaheed Mohtarma Medical University,  
 Larkana  
 (Member)

  
 Special Secretary (Public Health)  
 Health Department, Govt. of Sindh, Karachi  
 (Member)

  
 Dr: Safiullah Abbasi  
 Medical Superintendent,  
 Chandka Medical College Hospital,  
 Larkana (Chairman)

Answer - F

Supply of Furniture / Fixture under ADP Schemes viz;  
Establishment of Shaheed Benazir Bhutto Skin Diseases Complex & Construction of Dental and Hepatology Department at CMC Hospital Larkana ADP No. 623  
(2014-15)  
Revamping / Revitalization of Shaikh Zaid Hospital for Women Larkana ADP No. 625 (2014-15)

**COMPARATIVE STATEMENT FINANCIAL BID DATED: 05/05/2015**

S#	Quoted Item	Qty	Bidder	Bidder
1	Fowler Beds (Surgical)	96	Platinum	Pak Traders
			43,200	32,000
2	Foam Mattress with Regzine Cover (Master / Diamond)	100	Platinum	Pak Traders
			7,850	8,500
3	Bed Side Lockers	60	Platinum	Pak Traders
			9,164	9,900
4	Bed Side Benches (Talhi Wood ) Size 4'x1 1/2	60	Platinum	Pak Traders
			13,164	13,200
5	Office Chairs (Talhi Wood)	80	Platinum	Pak Traders
			5,300	N.Q
6	Office Tables (local) (Talhi Wood)	12	Platinum	Pak Traders
			17,534	N.Q
7	Nursing Counter	4	Platinum	Pak Traders
			54,657	N.Q

Signature:    

S #	Quoted Item	Qty	Bidder	Bidder
8	Iron safe / Cup Board	10	Platinum	Pak Traders
			20,280	24,000
9	Patients Examination Couch (Local)	6	Platinum	Pak Traders
			13,856	17,000
10	Fowler Beds with Mattress (36+4 Recovery)	136	Platinum	Pak Traders
			38,180	39,000
11	Blankets (Red) 60X90)	80	Platinum	Pak Traders
			2,200	1,150
12	Officer Table (6X4)	44	Platinum	Pak Traders
			29,178	N.Q
13	Revolving Chair	25	Platinum	Pak Traders
			14,794	N.Q
14	Chairs Seminar Hall, Library & Other Office	230	Platinum	Pak Traders
			8,424	N.Q
15	Books Rack for Library	6	Platinum	Pak Traders
			15,615	N.Q

*Handwritten signature*

*Handwritten signature*

*Handwritten signature*

*Handwritten signature*

S#	Quoted Item	Qty	Bidder	Bidder
16	Iron Cupboards	44	Platinum	Pak Traders
			20,273	24,000
17	Nursing Counter	12	Platinum	Pak Traders
			34,657	N.Q
18	Stools (Steel)	66	Platinum	Pak Traders
			4,800	5,400
19	I/V Stands (Steel)	95	Platinum	Pak Traders
			2,650	8,500
20	Wheel Chair	8	Platinum	Pak Traders
			11,200	9,800
21	Stretchers for Patients	8	Platinum	Pak Traders
			26,760	28,000
22	Office Chairs	138	Platinum	Pak Traders
			2,300	N.Q
23	Plastic Chairs for Waiting Area (3+1)	20	Platinum	Pak Traders
			6,850	N.Q






S#	Quoted Item	Qty	Bidder	Bidder
24	Table Side Rack	10	Platinum 9745	Pak Traders N.Q



Prof: Malik Hussain Jalbani  
Head of Department of Urology,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Shahida Shaikh  
Head of Department of Obs / Gynae,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Prof: Mohammad Illyas Shaikh,  
Principal, Baby Aisfa Dental College,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Special Secretary (Public Health)  
Health Department, Govt. of Sindh, Karachi  
(Member)



Prof: Abdat Manan Bhutto,  
Head of Dermatology Department,  
Shaheed Mohtarma Medical University, Larkana  
(Member)



Dr: Aijaz Ahmed Sharkh,  
Additional Medical Superintendent (Tech)  
CMC Hospital, Larkana  
(Member)



Dr: Safiullah Abbasi  
Medical Superintendent,  
Chandka Medical College Hospital,  
Larkana (Chairman)