

SINDH PUBLIC PROCUREMENT REGULATORY AUTHORITY

CONTRACT EVALUATION FORM

TO BE FILLED IN BY ALL PROCURING AGENCIES FOR PUBLIC CONTRACTS OF WORKS, SERVICES & GOODS

- 1) NAME OF THE ORGANIZATION / DEPTT. WELFARE SOCIETY FOR PATIENTS CARE THROUGH CIVIL HOSPITAL KARACH
- 2) PROVINCIAL / LOCAL GOVT./ OTHER Provincial
- 3) TITLE OF CONTRACT MEDICAL EQUIPMENT FOR WSPC UNDER ADP # 787
- 4) TENDER NUMBER No. MS(CHK)/2014-2015/- 3163
- 5) BRIEF DESCRIPTION OF CONTRACT MEDICAL EQUIPMENT FOR WSPC ADP # 787
- 6) FORUM THAT APPROVED THE SCHEME Planning Commission
- 7) TENDER ESTIMATED VALUE Rs. 56.358 Million
- 8) ENGINEER'S ESTIMATE (For civil works only) NOT APPLICABLE
- 9) ESTIMATED COMPLETION PERIOD (AS PER CONTRACT) 2014-15
- 10) TENDER OPENED ON (DATE & TIME) 25-04-2015 at 12.00 Noon
- 11) NUMBER OF TENDER DOCUMENTS SOLD (Attach list of buyers) Seventeen
- 12) NUMBER OF BIDS RECEIVED Seventeen
- 13) NUMBER OF BIDDERS PRESENT AT THE TIME OF OPENING OF BIDS Seventeen
- 14) BID EVALUATION REPORT (Enclose a copy) B.E.R already hoisted on authority's Website ID # 1485/2015
- 15) NAME AND ADDRESS OF THE SUCCESSFUL BIDDER Already Hoisted
- 16) CONTRACT AWARD PRICE Rs. 56.358 Million
- 17) RANKING OF SUCCESSFUL BIDDER IN EVALUATION REPORT (i.e. 1st, 2nd, 3rd EVALUATION BID). Ranking of the bidder as per Evaluation report already hoisted on Authority's Website @ ID # 1485/2015

18) METHOD OF PROCUREMENT USED :- (Tick one)

- a) SINGLE STAGE – ONE ENVELOPE PROCEDURE N/A Domestic/ Local
- b) SINGLE STAGE – TWO ENVELOPE PROCEDURE Yes
- c) TWO STAGE BIDDING PROCEDURE N/A
- d) TWO STAGE – TWO ENVELOPE BIDDING PROCEDURE N/A

PLEASE SPECIFY IF ANY OTHER METHOD OF PROCUREMENT WAS ADOPTED i.e. EMERGENCY, DIRECT CONTRACTING ETC. WITH BRIEF REASONS:

19) APPROVING AUTHORITY FOR AWARD OF CONTRACT _____

20) WHETHER THE PROCUREMENT WAS INCLUDED IN ANNUAL PROCUREMENT PLAN?

Yes	<input type="checkbox"/>	No	<input checked="" type="checkbox"/>
-----	--------------------------	----	-------------------------------------

21) ADVERTISEMENT :

i) SPPRA Website
(If yes, give date and SPPRA Identification No.)

Yes	SPPRA ID 1485/2015
No	

ii) News Papers
(If yes, give names of newspapers and dates)

Yes	Daily Kawish 09.5.15, Daily Dawn 10.5.15, Daily Jang 11.5.15 & Others
No	

22) NATURE OF CONTRACT

Domestic/ Local	<input checked="" type="checkbox"/>	Int.	<input type="checkbox"/>
--------------------	-------------------------------------	------	--------------------------

23) WHETHER QUALIFICATION CRITERIA
WAS INCLUDED IN BIDDING / TENDER DOCUMENTS?
(If yes, enclose a copy)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

24) WHETHER BID EVALUATION CRITERIA
WAS INCLUDED IN BIDDING / TENDER DOCUMENTS?
(If yes, enclose a copy)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

25) WHETHER APPROVAL OF COMPETENT AUTHORITY WAS OBTAINED FOR USING A
METHOD OTHER THAN OPEN COMPETITIVE BIDDING?

N/A

Yes	<input type="checkbox"/>	No	<input checked="" type="checkbox"/>
-----	--------------------------	----	-------------------------------------

26) WAS BID SECURITY OBTAINED FROM ALL THE BIDDERS?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

27) WHETHER THE SUCCESSFUL BID WAS LOWEST EVALUATED
BID / BEST EVALUATED BID (in case of Consultancies)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

28) WHETHER THE SUCCESSFUL BIDDER WAS TECHNICALLY
COMPLIANT?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

29) WHETHER NAMES OF THE BIDDERS AND THEIR QUOTED PRICES WERE READ OUT AT
THE TIME OF OPENING OF BIDS?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

30) WHETHER EVALUATION REPORT GIVEN TO BIDDERS BEFORE THE AWARD OF
CONTRACT?

(Attach copy of the bid evaluation report)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

31) ANY COMPLAINTS RECEIVED
(If yes, result thereof)

Yes	
No	No

32) ANY DEVIATION FROM SPECIFICATIONS GIVEN IN THE TENDER NOTICE / DOCUMENTS
(If yes, give details)

Yes	
No	No

33) WAS THE EXTENSION MADE IN RESPONSE TIME?
(If yes, give reasons)

Yes	
No	No

34) DEVIATION FROM QUALIFICATION CRITERIA
(If yes, give detailed reasons.)

Yes	
No	No

35) WAS IT ASSURED BY THE PROCURING AGENCY THAT THE SELECTED FIRM IS NOT BLACK LISTED?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

36) WAS A VISIT MADE BY ANY OFFICER/OFFICIAL OF THE PROCURING AGENCY TO THE SUPPLIER'S PREMISES IN CONNECTION WITH THE PROCUREMENT? IF SO, DETAILS TO BE ASCERTAINED REGARDING FINANCING OF VISIT, IF ABROAD:
(If yes, enclose a copy)

Yes	<input type="checkbox"/>	No	<input checked="" type="checkbox"/>
-----	--------------------------	----	-------------------------------------


37) WERE PROPER SAFEGUARDS PROVIDED ON MOBILIZATION ADVANCE PAYMENT IN THE CONTRACT (BANK GUARANTEE ETC.)?

Yes	<input type="checkbox"/>	No	<input checked="" type="checkbox"/>
-----	--------------------------	----	-------------------------------------

38) SPECIAL CONDITIONS, IF ANY
(If yes, give Brief Description)

Yes	No.
No	No.

Signature & Official Stamp of
Authorized Officer


Add. Medical Superintendent
(Procurement)
Civil Hospital, Karachi

FOR OFFICE USE ONLY

SPPRA, Block. No.8, Sindh Secretariat No.4-A, Court Road, Karachi
Tele: 021-9205356; 021-9205369 & Fax: 021-9206291

Print

Save

Reset



No. MS/CHK/14-15/ 2278
Civil Hospital,
Baba-e-Urdu Road, Karachi
Dated: 7-06-2015

M/s. Universal Enterprises
Karachi

Sub: **LETTER OF INTENT FOR SUPPLY OF MEDICAL EQUIPMENTS**

- Notice Inviting Tender issued in respect of ADP scheme # 787/2014-15 Viz: " Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi " to be supplied during the Financial Year 2014-2015.
- The Technical and Financial Bid submitted on 25th May, 2015 by your firm.
 - The meetings of Procurement Committee of the Civil Hospital, Karachi held on 25-05-2015 to open Technical proposal and the Financial proposal opened on 04-06-2015 respectively for supply of the following Medical Equipments as amount mentioned against each.

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Amount
1	Hemodialysis Machine with all Standard Accessories Model: Surdial 55, Make: Nipro Corp. - Japan	10	1,163,000.00	11,630,000.00
2	Reprocessing Unit with all Standard Accessories Model: Renatron Make: Medivator - USA	1	1,188,000.00	1,188,000.00
Grand Total Rs.				12,818,000.00

- The Procurement Committee has evaluated bids submitted by firms in detail and is pleased to inform you that you have been selected as the "Preferred Bidder" for supply of the above said Medical Equipments. The Civil Hospital, Karachi now wishes to offer you for supply of these equipments on the terms & conditions as per the bid documents of the said NIT.
- "That this letter of intent does not form a contractual or legal relationship under the relevant laws of Pakistan between Civil Hospital Karachi and the addressee of this "Letter of Intent"
- You are therefore, requested to confirm, within seven (07) days from the date of receipt of this Letter of Intent, regarding your willingness to accept this offer for supply of these equipments of the said scheme during current Financial Year 2014 – 15.
- If agreed form of Contract on Non-Judicial stamp along with Pay Order of Performance Security Bond of 5% of the contract value may be submitted to this Hospital within stipulated time

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to the:

- The Accountant General Sindh, Karachi
- The Director, SPPRA Karachi
- The Accounts Officer, Civil Hospital, Karachi


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2271 dated the 11/06/2015

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. Universal Enterprises, Karachi		
3.	Contractor's Tender No. & Date	Ref#. UBM-0138/05/15 Dated: 21-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indenter's Name and Address	President Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
1	Hemodialysis Machine with all Standard Accessories Model: Surdial 55, Make: Nipro Corp. - Japan	10 Nos.	Rs. 1,163,000/- Each	11,630,000/-
TOTAL AMOUNT				Rs. 11,630,000/-

TOTAL VALUE IN WORDS: Eleven Million Six Hundred Thirty Thousand Only.

8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologist, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 640,900/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 09977783 dated 12-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department(Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2292 dated the 15/06 2015

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. Universal Enterprises, Karachi		
3.	Contractor's Tender No. & Date	Ref#. UBM-0138/05/15 Dated: 21-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
2	Reprocessing Unit with all Standard Accessories Model: Renatron Make: Medivator - USA	01 Nos.	Rs. 1,188,000/- Each	1,188,000/-
TOTAL AMOUNT				Rs. 1,188,000/-

TOTAL VALUE IN WORDS: One Million One Hundred Eighty Eight Thousand Only.


8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 640,900/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 09977783 dated 12-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

**MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI**

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department(Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT



ARIF BAIG STAMP VENDOR
Licence No. 09, Shop No. 16, Block-14,
Karim Plaza, Gulshan-e-Iqbal, Karachi.

093456

S. No. Date 09 MAY 2015

Issued to With Address

Through With Address

Purpose ZAHID MEHMOOD

Value Rs. Attached 2 A Advocate

Stamps Vendor's Signature

Not For Used Bank Guaranty/will/divorce

Vendor Not Responsible Any Fake Documentation

CONTRACT AGREEMENT

Letter of Intent No. MS/CHK/2014-2015/2278 Dated 09-06-2015.

This agreement made this 09th June, 2015, between the Medical Superintendent, Civil Hospital, Karachi "The Purchaser" herein after and **M/s. Universal Enterprise**: Address: 29- Block-3, Overseas Co-operative Housing Society Stadium Road, Karachi, herein the supplier".


Whereas the purchaser is desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quote the rate which has accepted by purchase for supply of the following goods in the sum of **Rs. 395,000/-** herein after " The Contract Price"

Item #	Description of Stores	Rate On DDP Basis	Quantity Ordered	Total value in figures.
01	Hemodialysis Machine with all Standard Accessories Model: Surdial 55, Make: Nipro Corp. - Japan	Rs. 1,1630,00/- Each	10 No.	11,630,000.00
02	Reprocessing Unit with all Standard Accessories Model: Renatron Make: Medivator - USA	Rs. 1,188,000/- Each	01 No.	1,188,000.00
Total Amount				Rs. 12,818,000/-

The delivery period will commence from the date of acceptance of the contract by the bidder.

Now this agreement witnessed as following:

1. In this agreement words and expression shall have the same meanings as are respectively assigned to them in the Terms & Conditions of Tender Enquiry referred to.
2. Following document shall be seemed top form and be read and constructed as per of this agreement, viz:
 - a. Contract Agreement.
 - b. Price Schedule.
 - c. Technical Specifications.


Prof. M. Saad Quraishy
Medical Superintendent
Civil Hospital, Karachi.

- d. Terms & Conditions of Tender Enquiry.
- e. Terms & Conditions of the Contract.
- f. Schedule of requirement.
- g. Bidder offer including the relevant correspondence with all annexes made before signing the contract.

3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.

4. The purchase hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects therein, the contracts price payable under the provision of the Contract at the time and in the manner prescribed in the contract.

In witness whereof the parties here to here caused this agreement in accordance with their respective hands and seals, the day, the month and the year first above written.

SUPPLIER



SIGNATURE & SEAL

Abdul Munir

NAME : Abdul Munir

Designation : Customer Care Manager

Address : XIII A, 93, Gulsara-e-Maqbal Karachi

PURCHASER

SIGNATURE & SEAL

Prof. M. Saeed Quraishy

NAME : _____

Designation : Prof. M. Saeed Quraishy

Address : Medical Superintendent Civil Hospital, Karachi.

WITNESS (SUPPLIER)



SIGNATURE & SEAL

Muhammael Khelich

NAME : Muhammael Khelich

Designation : Feild Manager

Address : 29/3 Overseas cooperative Housing Society Karachi

WITNESS (PURCHASER)

SIGNATURE & SEAL

Dr. Sohail Mir Momon

NAME : Dr. Sohail Mir Momon

Designation : Adl. Medical Superintendent (Procurement) Civil Hospital, Karachi.

Address : _____



No. MS/CHK/14-15/ 22781
Civil Hospital,
Baba-e-Urdu Road, Karachi
Dated: 04-06-2015

M/s. Life One
Karachi

Sub: LETTER OF INTENT FOR SUPPLY OF MEDICAL EQUIPMENTS

- Notice Inviting Tender issued in respect of ADP scheme # 787/2014-15 Viz: " Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi " to be supplied during the Financial Year 2014-2015.
- The Technical and Financial Bid submitted on 25th May, 2015 by your firm.
 - The meetings of Procurement Committee of the Civil Hospital, Karachi held on 25-05-2015 to open Technical proposal and the Financial proposal opened on 04-06-2015 respectively for supply of the following Medical Equipments as amount mentioned against each.

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Amount
3	OT Light with all Standard Accessories Model: ALB-100C Make: AFA - Taiwan	1	605,000.00	605,000.00
4	OT Table with all Standard Accessories Model: OT-125 A Make: St-Frances - Taiwan	1	585,000.00	585,000.00
9	Autoclave with all Standard Accessories Model: AS-450 Make: AFA - Taiwan	1	1,065,000.00	1,065,000.00
13	Syringe Pump with all Standard Accessories Model: Medifusion-MS1200, Make: Daiwah - Korea	10	91,000.00	910,000.00
15	Patient Monitor with all Standard Accessories Model: G3L, Make: General Meditech - China	11	146,000.00	1,606,000.00
17	Diathermy with all Standard Accessories Model: DT-400P, Make: Daiwha - Korea	1	369,000.00	369,000.00
Grand Total Rs.				5,140,000.00

- The Procurement Committee has evaluated bids submitted by firms in detail and is pleased to inform you that you have been selected as the "Preferred Bidder" for supply of the above said Medical Equipments. The Civil Hospital, Karachi now wishes to offer you for supply of these equipments on the terms & conditions as per the bid documents of the said NIT.
- "That this letter of intent does not form a contractual or legal relationship under the relevant laws of Pakistan between Civil Hospital Karachi and the addressee of this "Letter of Intent"
 - You are therefore, requested to confirm, within seven (07) days from the date of receipt of this Letter of Intent, regarding your willingness to accept this offer for supply of these equipments of the said scheme during current Financial Year 2014 - 15.
 - If agreed form of Contract on Non-Judicial stamp along with Pay Order of Performance Security Bond of 5% of the contract value may be submitted to this Hospital within stipulated time

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to the:

- The Accountant General Sindh, Karachi
- The Director, SPPRA Karachi
- The Accounts Officer, Civil Hospital, Karachi


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2293 dated the 15/06/2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. Life One, Karachi		
3.	Contractor's Tender No. & Date	Ref #. NIL Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
3.	OT Light with all Standard Accessories Model: ALB-100C Make: AFA - Taiwan	01 No.	Rs. 605,000/- Each	605,000/-
4.	OT Table with all Standard Accessories Model: OT-125 A Make: St-Frances - Taiwan	01 No.	Rs. 585,000/- Each	585,000/-
9.	Autoclave with all Standard Accessories Model: AS-450 Make: AFA - Taiwan	01 No.	Rs. 1,065,000/- Each	1,065,000/-
13.	Syringe Pump with all Standard Accessories Model: Medifusion-MS1200, Make: Daiwah - Korea	10 Nos.	Rs. 91,000/- Each	910,000/-
15.	Patient Monitor with all Standard Accessories Model: G3L, Make: General Meditech - China	11 Nos.	Rs. 146,000/- Each	1,606,000/-
17.	Diathermy with all Standard Accessories Model: DT-400P, Make: Daiwha - Korea	01 No.	Rs. 369,000/- Each	369,000/-
Grand Total Rs.				Rs. 5,140,000/-

TOTAL VALUE IN WORDS: Five Million One Hundred Forty Thousand Only.


8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 170,000/- & 87,000/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 70500511 & 70500536 dated 22-05-2015 & 11-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

**MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI**

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	iii) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care iv) Representative of Health Department (Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT



JAMIL AKHTER SHAIKH STAMP VENDORS

Licence # 17, Room No: 302,
Tahir Plaza, Near City Court Karachi.

SR. NO. 13383 DATE 19/12/14 19 DEC 2014
ISSUED TO WITH ADDRESS MR. Atiqam Sarwar Advocate;
THROUGH WITH ADDRESS MR. Ledger No: 177
PURPOSE _____
VALUE RS. 100/- ATTACHED _____
STAMP VENDOR'S SIGNATURE _____

CONTRACT AGREEMENT

Letter of Intent No. MS/CHK/2014-2015/2281 Dated 09-06-2015.

This agreement made this 10th June, 2015, between the Medical Superintendent, Civil Hospital, Karachi "The Purchaser" herein after and **M/s. Life One**. Address: Suite D-1, 4th Floor West land, Trade Centre, C-5, Central Commercial Area, Block 7 & 8, KCHUS-Ltd. Karachi, herein the supplier".

Whereas the purchaser is desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quote the rate which has accepted by purchase for supply of the following goods in the sum of **Rs. 5,140,000/-** herein after " The Contract Price"

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Total Amount
3	OT Light with all Standard Accessories Model: ALB-100C Make: AFA - Taiwan	01	Rs. 605,000/- Each	605,000/-
4	OT Table with all Standard Accessories Model: OT-125 A Make: St-Frances - Taiwan	01	Rs. 585,000/- Each	585,000/-
9	Autoclave with all Standard Accessories Model: AS-450 Make: AFA - Taiwan	01	Rs. 1,065,000/- Each	1,065,000/-
13	Syringe Pump with all Standard Accessories Model: Medifusion-MS1200, Make: Daiwah - Korea	10	Rs. 91,000/- Each	910,000/-
15	Patient Monitor with all Standard Accessories Model: G3L, Make: General Meditech - China	11	Rs. 146,000/- Each	1,606,000/-
17	Diathermy with all Standard Accessories Model: DT-400P, Make: Daiwha - Korea	01	Rs. 369,000/- Each	369,000/-
Grand Total Rs.				Rs. 5,140,000/-

The delivery period will commence from the date of acceptance of the contract by the bidder.

Now this agreement witnessed as following:

1. In this agreement words and expression shall have the same meanings as are respectively assigned to them in the Terms & Conditions of Tender Enquiry referred to.
2. Following document shall be seemed top form and be read and constructed as per of this agreement, viz:
 - a. Contract Agreement.
 - b. Price Schedule.

Prof. M. Saeed Quraishy
Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.



- c. Technical Specifications.
- d. Terms & Conditions of Tender Enquiry.
- e. Terms & Conditions of the Contract.
- f. Schedule of requirement.
- g. Bidder offer including the relevant correspondence with all annexes made before signing the contract.

3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.

4. The purchase hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects therein, the contracts price payable under the provision of the Contract at the time and in the manner prescribed in the contract.

In witness whereof the parties here to here caused this agreement in accordance with their respective hands and seals, the day, the month and the year first above written.

SUPPLIER

PURCHASER

SIGNATURE & SEAL

SIGNATURE & SEAL

NAME: Muhammad Arshad Jani Brig.

NAME: _____

Designation: General Manager

Designation: Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.

Address: PU 17/18 Block 13-A
Gulshan Ehsan, Karachi

Address: _____

WITNESS (SUPPLIER)

WITNESS (PURCHASER)

SIGNATURE & SEAL

SIGNATURE & SEAL

NAME: KHURAM KHAN

NAME: Dr. Sohail Mir Memon

Designation: _____

Designation: Addl. Medical Superintendent
(Procurement)
Civil Hospital, Karachi.

Address: _____

Address: _____



No. MS/CHK/14-15/ 2082
Civil Hospital,
Baba-e-Urdu Road, Karachi
Dated: 07-06-2015

M/s. Zam Trading
Karachi

Sub: **LETTER OF INTENT FOR SUPPLY OF MEDICAL EQUIPMENTS**

- Notice Inviting Tender issued in respect of ADP scheme # 787/2014-15 Viz: "Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi" to be supplied during the Financial Year 2014-2015.
- The Technical and Financial Bid submitted on 25th May, 2015 by your firm.
 - The meetings of Procurement Committee of the Civil Hospital, Karachi held on 25-05-2015 to open Technical proposal and the Financial proposal opened on 04-06-2015 respectively for supply of the following Medical Equipments as amount mentioned against each.

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Amount
5	Surgical Beds (Dialysis Chair) with all Standard Accessories, Model: WM3000, Make: Winmed - Local	3	164,800.00	494,400.00
6	Surgical Beds (Dialysis Recliners) with all Standard Accessories, Model: WM5001, Make: Winmed - Local	3	165,000.00	495,000.00
8	Stretcher with all Standard Accessories Model: ER1000, Make: Winmed - Local	13	49,900.00	648,700.00
Grand Total Rs.				1,638,100.00

- The Procurement Committee has evaluated bids submitted by firms in detail and is pleased to inform you that you have been selected as the "Preferred Bidder" for supply of the above said Medical Equipments. The Civil Hospital, Karachi now wishes to offer you for supply of these equipments on the terms & conditions as per the bid documents of the said NIT.
- "That this letter of intent does not form a contractual or legal relationship under the relevant laws of Pakistan between Civil Hospital Karachi and the addressee of this "Letter of Intent"
- You are therefore, requested to confirm, within seven (07) days from the date of receipt of this Letter of Intent, regarding your willingness to accept this offer for supply of these equipments of the said scheme during current Financial Year 2014 - 15.
- If agreed form of Contract on Non-Judicial stamp along with Pay Order of Performance Security Bond of 5% of the contract value may be submitted to this Hospital within stipulated time

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to the:

- The Accountant General Sindh, Karachi
- The Director, SPPRA Karachi
- The Accounts Officer, Civil Hospital, Karachi


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2094 dated the 18/11/2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. Zam Trading Co., Karachi		
3.	Contractor's Tender No. & Date	Ref#. NIL Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
5	Surgical Beds (Dialysis Chair) with all Standard Accessories, Model: WM3000, Make: Winmed - Local	03 Nos.	Rs. 164,800/- Each	494,400/-
6	Surgical Beds (Dialysis Recliners) with all Standard Accessories, Model: WM5001, Make: Winmed - Local	03 Nos.	Rs. 165,000/- Each	495,000/-
8	Stretcher with all Standard Accessories Model: ER1000, Make: Winmed - Local	13 Nos.	Rs. 49,900/- Each	648,700/-
Grand Total Rs.				Rs. 1,638,100/-

TOTAL VALUE IN WORDS: One Million Six Hundred Thirty Eight Thousand One Hundred Only.


8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 81,900/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. _____ dated _____, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

**MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI**

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department (Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT



JAMIL AKHTER SHAIKH STAMP VENDOR

License # 12, Room No: 302,

Tahir Plaza, Near City Court, Karachi.

SRL NO. 27196 DATE 11/06/15

ISSUED TO WITH ADDRESS

THROUGH WITH ADDRESS MP

PURPOSE

VALUE RS. 100/- ATTACHED

STAMP VENDOR SIGNATURE

NOT ISSUED FOR

Muzaffaruddin A. Siddiqui
Advocate High Court
Lodger No: 643

CONTRACT AGREEMENT

Letter of Intent No. MS/CHK/2014-2015/2294 Dated 15-06-2015.

This agreement made this 15th June, 2015, between the Medical Superintendent, Civil Hospital, Karachi "The Purchaser" herein after and **M/s. Zam Trading Co.** Address: Room No. 82, 3rd Floor, National Chamber Near Hamdard Dawakhana, Arambagh, Karachi, herein the supplier".

Whereas the purchaser is desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quote the rate which has accepted by purchase for supply of the following goods in the sum of **Rs. 1,638,100/-** herein after " The Contract Price"

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Total Amount
5	Surgical Beds (Dialysis Chair) with all Standard Accessories, Model: WM3000, Make: Winmed - Local	3	Rs. 164,800/- Each	494,400.00
6	Surgical Beds (Dialysis Recliners) with all Standard Accessories, Model: WM5001, Make: Winmed - Local	3	Rs. 165,000/- Each	495,000.00
8	Stretcher with all Standard Accessories Model: ER1000, Make: Winmed - Local	13	Rs. 49,900/- Each	648,700.00
Grand Total Rs.				Rs. 1,638,100.00

The delivery period will commence from the date of acceptance of the contract by the bidder.

Now this agreement witnessed as following:

- In this agreement words and expression shall have the same meanings as are respectively assigned to them in the Terms & Conditions of Tender Enquiry referred to.
- Following document shall be seemed top form and be read and constructed as per of this agreement, viz:
 - Contract Agreement.
 - Price Schedule.
 - Technical Specifications.
 - Terms & Conditions of Tender Enquiry.
 - Terms & Conditions of the Contract.
 - Schedule of requirement.

[Signature]
Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.

[Signature]

g. Bidder offer including the relevant correspondence with all annexes made before signing the contract.

3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.

4. The purchase hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects therein, the contracts price payable under the provision of the Contract at the time and in the manner prescribed in the contract.

In witness whereof the parties here to here caused this agreement in accordance with their respective hands and seals, the day, the month and the year first above written.

SUPPLIER

SIGNATURE & SEAL

NAME :

Designation :

Address :

Saheem Khan

Saheem Khan

Manager

*Plot 32 3rd Floor National Chamber
Near Hamdard Dawakhana Main Bazaar Karachi*



PURCHASER

SIGNATURE & SEAL

NAME :

Designation :

Address :

Prof. M. Saeed Quraishy

**Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.**

WITNESS (SUPPLIER)

SIGNATURE & SEAL

NAME :

Designation :

Address :

Dyaz ud Din

Dyaz ud Din

Sales Officer

WITNESS (PURCHASER)

SIGNATURE & SEAL

NAME :

Designation :

Address :

Dr. Sohail Mir Memori

**Dr. Sohail Mir Memori
Addl. Medical Superintendent
(Procurement)
Civil Hospital, Karachi.**



No. MS/CHK/14-15/ 2019
Civil Hospital,
Baba-e-Urdu Road, Karachi
Dated: 07/06/2015

M/s. BIOS
Karachi

Sub: **LETTER OF INTENT FOR SUPPLY OF MEDICAL EQUIPMENTS**

- a. Notice Inviting Tender issued in respect of ADP scheme # 787/2014-15 Viz: "Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi" to be supplied during the Financial Year 2014-2015.
- b. The Technical and Financial Bid submitted on 25th May, 2015 by your firm.
- c. The meetings of Procurement Committee of the Civil Hospital, Karachi held on 25-05-2015 to open Technical proposal and the Financial proposal opened on 04-06-2015 respectively for supply of the following Medical Equipments as amount mentioned against each.

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Amount
7	Crash Trolley with all Standard Accessories Model: CAVS, Make: Hersill - Spain	3	150,000.00	450,000.00
10	Ventilator with all Standard Accessories Model: Inspiration 5i, Make: Event - California / USA	4	2,195,000.00	8,780,000.00
11	Fluoroscopy with DR System 500mA with all Standard Accessories i) Digital X-Ray: Model: N40HF, Make: Stephanix - France ii) C-Arm: Model: Omniscope Make: Stephanix - France	1	15,585,000.00 8,330,000.00	23,915,000.00
14	Echo Color Doppler with all Standard Accessories Model: S11, Make: Sonoscap - China	1	1,899,000.00	1,899,000.00
Grand Total Rs.				35,044,000.00

- 1 The Procurement Committee has evaluated bids submitted by firms in detail and is pleased to inform you that you have been selected as the "Preferred Bidder" for supply of the above said Medical Equipments. The Civil Hospital, Karachi now wishes to offer you for supply of these equipments on the terms & conditions as per the bid documents of the said NIT.
- 2 "That this letter of intent does not form a contractual or legal relationship under the relevant laws of Pakistan between Civil Hospital Karachi and the addressee of this "Letter of Intent"
- 3 You are therefore, requested to confirm, within seven (07) days from the date of receipt of this Letter of Intent, regarding your willingness to accept this offer for supply of these equipments of the said scheme during current Financial Year 2014 - 15.
- 4 If agreed form of Contract on Non-Judicial stamp along with Pay Order of Performance Security Bond of 5% of the contract value may be submitted to this Hospital within stipulated time

**MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI**

Copy to the:

- The Accountant General Sindh, Karachi
- The Director, SPPRA Karachi
- The Accounts Officer, Civil Hospital, Karachi


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2295 dated the 15/06. 2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. BIOS, Karachi		
3.	Contractor's Tender No. & Date	Ref#. BIOS/864/2015 Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
7	Crash Trolley with all Standard Accessories Model: CAVS, Make: Hersill - Spain	03 Nos.	Rs. 150,000/- Each	450,000/-
Grand Total Rs.				Rs. 450,000/-

TOTAL VALUE IN WORDS: Four Hundred Fifty Thousand Only.

8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 802,200/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 70500535 dated 11-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department(Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2296 dated the 10/07/2015. 2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. BIOS, Karachi		
3.	Contractor's Tender No. & Date	Ref#. BIOS/864/2015 Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
10	Ventilator with all Standard Accessories Model: Inspiration 5i, Make: Event - California / USA	04 Nos.	Rs. 2,195,000/- Each	8,780,000/-
Grand Total Rs.				Rs. 8,780,000/-

TOTAL VALUE IN WORDS: Eight Million Seven Hundred Eighty Thousand Only.

8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 802,200/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 70500535 dated 11-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department(Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2297 dated the 10/06. 2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. BIOS, Karachi		
3.	Contractor's Tender No. & Date	Ref #. BIOS/864/2015 Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
11	Fluoroscopy with DR System 500mA with all Standard Accessories			
	i) Digital X-Ray: Model: N40HF, Make: Stephanix – France	01 No.	Rs. 15,585,000/- Each	Rs. 15,585,000/-
	ii) C-Arm: Model: Omniscope Make: Stephanix - France	01 No.	Rs. 8,330,000/- Each	Rs. 8,330,000/-
Grand Total Rs.				Rs. 23,915,000/-

TOTAL VALUE IN WORDS: Twenty Three Million Nine Hundred Fifteen Thousand Only.

8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 950,000/- & 802,200/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 70500510 & 70500535 dated 22-05-2015 & 11-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department (Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2299 dated the 19/06 2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. BIOS, Karachi		
3.	Contractor's Tender No. & Date	Ref#. BIOS/864/2015 Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
14	Echo Color Doppler with all Standard Accessories Model: S11, Make: Sonoscap - China	01 No.	1,899,000.00	1,899,000/-
Grand Total Rs.				Rs. 1,899,000/-


TOTAL VALUE IN WORDS: One Million Eight Hundred Ninety Nine Thousand Only.


8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 802,200/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 70500535 dated 11-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

**MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI**

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department (Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT



ARI-BAG STAMP VENDOR

Licence No. 09, Shop No. 16, Block-14,
Karim Plaza, Gulshan-e-Iqbal, Karachi.

S. No. **093455** Date **09 MAY 2015**

Issued to With Address **ZAHID MEHMOOD**

Through With Address **Defence**

Purpose **A**

Value Rs. Attached

Stamps Vendor's Signature

Not For Used Bank Guaranty/will/divorce

Vendor Not Responsible Any Fake Documentation

CONTRACT AGREEMENT

Letter of Intent No. MS/CHK/2014-2015/2279 Dated 09-06-2015.

This agreement made this 09th June, 2015, between the Medical Superintendent, Civil Hospital, Karachi "The Purchaser" herein after and M/s. BIOS. Address: 2nd Floor Building No. 2C 12 Commercial, Phase II Ext. D.H.A. Karachi, herein the supplier".

Whereas the purchaser is desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quote the rate which has accepted by purchase for supply of the following goods in the sum of **Rs. 35,044,000/-** herein after " The Contract Price"

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Total Amount
7	Crash Trolley with all Standard Accessories Model: CAVS, Make: Hersill - Spain	03	Rs. 150,000/- Each	450,000/-
10	Ventilator with all Standard Accessories Model: Inspiration 5I, Make: Event - California / USA	04	Rs. 2,195,000/- Each	8,780,000/-
11	Fluoroscopy with DR System 500mA with all Standard Accessories i) Digital X-Ray: Model: N40HF, Make: Stephanix - France ii) C-Arm: Model: Omniscope Make: Stephanix - France	01	Rs. 15,585,000/- Each Rs. 8,330,000/- Each	23,915,000/-
14	Echo Color Doppler with all Standard Accessories Model: S11, Make: Sonoscap - China	01	Rs. 1,899,000/- Each	1,899,000/-
Grand Total Rs.				Rs. 35,044,000/-

The delivery period will commence from the date of acceptance of the contract by the bidder.

Now this agreement witnessed as following:

1. In this agreement words and expression shall have the same meanings as are respectively assigned to them in the Terms & Conditions of Tender Enquiry referred to.
2. Following document shall be seemed top form and be read and constructed as per of this agreement, viz.
 - a. Contract Agreement.
 - b. Price Schedule.
 - c. Technical Specifications.

Prof. M. Saeed Quraishy
Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.





- d. Terms & Conditions of Tender Enquiry.
 - e. Terms & Conditions of the Contract.
 - f. Schedule of requirement.
 - g. Bidder offer including the relevant correspondence with all annexes made before signing the contract.
3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.
4. The purchase hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects therein, the contracts price payable under the provision of the Contract at the time and in the manner prescribed in the contract.

In witness whereof the parties here to here caused this agreement in accordance with their respective hands and seals, the day, the month and the year first above written.

3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.

SUPPLIER


SIGNATURE & SEAL  

NAME : Munir Ahmed

Designation : CEO

Address: BIOS 2nd Floor Bld 2C Ph 2 DHA - Karachi

PURCHASER


SIGNATURE & SEAL 

NAME : _____

Designation : _____

Address: Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.

WITNESS (SUPPLIER)


SIGNATURE & SEAL 

NAME : A. Waqid

Designation : Tech Engineer

Address: BIOS 2nd Floor Building 2C Ph 2 - DHA - Karachi.

WITNESS (PURCHASER)

SIGNATURE & SEAL 

NAME : _____

Designation : Dr. Sohail Mir Memon
Addl. Medical Superintendent
(Procurement)

Address: Civil Hospital, Karachi.



No. MS/CHK/14-15/ 2780
 Civil Hospital,
 Baba-e-Urdu Road, Karachi
 Dated: 06/06/2015

M/s. Hospicare Systems
 Karachi

Sub: **LETTER OF INTENT FOR SUPPLY OF MEDICAL EQUIPMENTS**

- Notice Inviting Tender issued in respect of ADP scheme # 787/2014-15 Viz: " Purchase of Medical
- Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi "** to be supplied during the Financial Year 2014-2015.
 - The Technical and Financial Bid submitted on 25th May, 2015 by your firm.
 - The meetings of Procurement Committee of the Civil Hospital, Karachi held on 25-05-2015 to open Technical proposal and the Financial proposal opened on 04-06-2015 respectively for supply of the following Medical Equipments as amount mentioned against each.

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Amount
12	ABG Machine System with all Standard Accessories Model: Gastat - 600, Make: Techno Medical - Japan	1	985,000.00	985,000.00
Grand Total Rs.				985,000.00

- Sub: The Procurement Committee has evaluated bids submitted by firms in detail and is pleased to inform you that you have been selected as the "Preferred Bidder" for supply of the above said Medical Equipments. The Civil Hospital, Karachi now wishes to offer you for supply of these equipments on the terms & conditions as per the bid documents of the said NIT.
- "That this letter of intent does not form a contractual or legal relationship under the relevant laws of Pakistan between Civil Hospital Karachi and the addressee of this "Letter of Intent"
 - You are therefore, requested to confirm, within seven (07) days from the date of receipt of this Letter of Intent, regarding your willingness to accept this offer for supply of these equipments of the said scheme during current Financial Year 2014 – 15.
 - If agreed form of Contract on Non-Judicial stamp along with Pay Order of Performance Security Bond of 5% of the contract value may be submitted to this Hospital within stipulated time

**MEDICAL SUPERINTENDENT
 CIVIL HOSPITAL KARACHI**

Copy to the:

- The Accountant General Sindh, Karachi
- The Director, SPPRA Karachi
- The Accounts Officer, Civil Hospital, Karachi


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2278 dated the 11.05.2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. Hospicare Systems, Karachi		
3.	Contractor's Tender No. & Date	Ref#. NIL Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
12	ABG Machine System with all Standard Accessories Model: Gastat - 600, Make: Techno Medical - Japan	01 Nos.	Rs. 985,000/- Each	985,000/-
Grand Total Rs.				Rs. 985,000/-

TOTAL VALUE IN WORDS: Nine Hundred Eighty Five Thousand Only.


8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 49,250/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 01213086 dated 12-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department (Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT



GUNUS KHAN STAMP VENDOR

No. 8, Seal of the Compound District East, Karachi

S No: 67160 Date: _____

Made with Order _____

Through with Address _____

Purposed _____

Valued Rs. _____

Stamp Vendor Sign _____

SYED FAROOQ ALI
Advocate High Court

CONTRACT AGREEMENT

Letter of Intent No. MS/CHK/2014-2015/2280 Dated 09-06-2015.

This agreement made this 09th June, 2015, between the Medical Superintendent, Civil Hospital, Karachi "The Purchaser" herein after and **M/s. Hospicare Systems**: Address: Ground floor, Rabbiya Garden, Block-3, M.C.H.S, Shaheed-e-Millat Road, Karachi, herein the supplier".

Whereas the purchaser is desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quote the rate which has accepted by purchase for supply of the following goods in the sum of **Rs. 985,000/-** herein after " The Contract Price"

Item #	Description of Stores	Rate On C & F Basis	Quantity Ordered	Total value in figures.
12.	ABG Machine System with all Standard Accessories Model: Gastat - 600, Make: Techno Medical - Japan	Rs. 985,000/- Each	01 No.	985,000.00
Total Amount				Rs. 985,000/-

The delivery period will commence from the date of acceptance of the contract by the bidder.

Now this agreement witnessed as following:

1. In this agreement words and expression shall have the same meanings as are respectively assigned to them in the Terms & Conditions of Tender Enquiry referred to.
2. Following document shall be seemed top form and be read and constructed as per of this agreement, viz:
 - a. Contract Agreement.
 - b. Price Schedule.
 - c. Technical Specifications.
 - d. Terms & Conditions of Tender Enquiry.
 - e. Terms & Conditions of the Contract.
 - f. Schedule of requirement.

Muhammad Quraishy
 Muhammad Quraishy
 Medical Superintendent
 Civil Hospital, Karachi
 Civil Hospital, Karachi.

Muhammad Quraishy


g. Bidder offer including the relevant correspondence with all annexes made before signing the contract.

3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.
4. The purchase hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects therein, the contracts price payable under the provision of the Contract at the time and in the manner prescribed in the contract.

In witness whereof the parties here to here caused this agreement in accordance with their respective hands and seals, the day, the month and the year first above written.

SUPPLIER





SIGNATURE & SEAL

NAME : MASOOD AHMED

Designation : CEO

Address : Rabiya Garden, Karachi

PURCHASER




SIGNATURE & SEAL

NAME : Prof. M. Saeed Quraishy

Designation : Medical Superintendent
Civil Hospital, Karachi.

Address : _____

WITNESS (SUPPLIER)



SIGNATURE & SEAL 

NAME : Aman Ullah

Designation : CEO BioMedics

Address : Rabiya Garden, Karachi

WITNESS (PURCHASER)


SIGNATURE & SEAL

NAME : Dr. Sohail Mir Memon

Designation : Addl. Medical Superintendent
(Procurement)

Address : Civil Hospital, Karachi.



No. MS/CHK/14-15/ 2283
Civil Hospital,
Baba-e-Urdu Road, Karachi
Dated: 09-06-2015

M/s. Elate CC (Pvt.) Ltd.
Karachi

Sub: **LETTER OF INTENT FOR SUPPLY OF MEDICAL EQUIPMENTS**

Notice Inviting Tender issued in respect of ADP scheme # 787/2014-15 Viz: " **Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi** " to be supplied during the Financial Year 2014-2015.

b. The Technical and Financial Bid submitted on 25th May, 2015 by your firm.

The meetings of Procurement Committee of the Civil Hospital, Karachi held on 25-05-2015 to open Technical proposal and the Financial proposal opened on 04-06-2015 respectively for supply of the following Medical Equipments as amount mentioned against each.

Item #	Description (s) of Equipment	Qty	Unit price on DDP Basis	Amount
16	Defibrillator with all Standard Accessories Model: XD1, Make: Primedic - Germany	1	395,000.00	395,000.00
Grand Total Rs.				395,000.00

The Procurement Committee has evaluated bids submitted by firms in detail and is pleased to inform you that you have been selected as the "Preferred Bidder" for supply of the above said Medical Equipments. The Civil Hospital, Karachi now wishes to offer you for supply of these equipments on the terms & conditions as per the bid documents of the said NIT.

"That this letter of intent does not form a contractual or legal relationship under the relevant laws of Pakistan between Civil Hospital Karachi and the addressee of this "Letter of Intent"

You are therefore, requested to confirm, within seven (07) days from the date of receipt of this Letter of Intent, regarding your willingness to accept this offer for supply of these equipments of the said scheme during current Financial Year 2014 – 15.

If agreed form of Contract on Non-Judicial stamp along with Pay Order of Performance Security Bond of 5% of the contract value may be submitted to this Hospital within stipulated time

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to the:

- The Accountant General Sindh, Karachi
- The Director, SPPRA Karachi
- The Accounts Officer, Civil Hospital, Karachi


MEDICAL SUPERINTENDENT

OFFICE OF THE MEDICAL SUPERINTENDENT, CIVIL HOSPITAL, KARACHI
NO.MS(CHK)ADP/787/2014-15/- 2300 dated the 15/07. 2015.

1.	Contract No & Date	Purchase of Medical Equipment for Welfare Society for Patient Care Through Civil Hospital, Karachi		
2.	Name & Address of the Contractor	M/s. Elate CC (Pvt.) Ltd., Karachi		
3.	Contractor's Tender No. & Date	Ref#. NIL Dated: 25-05-2015		
4.	Tender Inquiry No. & Date	No. MS/CHK/2014-2015/3163 DATED 07-05-2015		
5.	Indentor's Name and Address	Welfare Society for Patient Care		
6.	Conditions of the Contract	As per Bid Documents & Contract Agreement on Non-Judicial stamp paper provided by the Contractor/Supplier		
7.	Particular of Stores	Details are as under :		
Item #	Description of Stores	Rate On D.D.P Basis	Quantity Ordered	Total value in figures.
16	Defibrillator with all Standard Accessories Model: XD1, Make: Primedic - Germany	01 No.	Rs. 395,000/- Each	395,000/-
Grand Total Rs.				Rs. 395,000/-

TOTAL VALUE IN WORDS: Three Hundred Ninety Five Thousand Only.


8.	Mode of Payment	As per Bid Documents Supply order (s) will be placed on D.D.P basis in currency of Pak Rupees to the Contractors/Suppliers, as the current financial year is nearing to end. To avoid lapse of funds the bills will be claimed from office of the Accountant General Sindh after getting necessary guarantee from the Contractors / Suppliers in shape of pay order, demand draft, call deposit receipt, Bank Guarantee in favor of procuring agency (PA) equivalent of the claimed amount the same will be refunded to the Contractors/Suppliers after receiving the equipment by the Welfare Society for Patient Care, as per the required specification.
9.	Place of Delivery.	Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
10.	Dispatch Instruction:	As per Agreement on Non-Judicial Stamp Paper.
11.	Name and Address of the Consignee:	Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care, Clifton Kidney & General Hospital, F-17/2, Rojhan Street, Block 5, Clifton, Karachi
12.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani, President Welfare Society for Patient Care ii) Representative of Development Wing, Health Department, Gov. of Sindh
13.	Security Deposit	Worth Rs. 19,750/- i.e 5% of A/T Value obtained in shape of Pay Order / Call Deposit No. 00860530 dated 12-06-2015, which will be released after receiving satisfactory performance / certificate from End User.
14.	Date Of Delivery	As per bid documents

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

Copy to:-

1.	The Accountant General Sindh, Karachi.	Head of Account No:- Health 073-Hospital Services 073101- General Hospital Services.
2.	Indenting Officer	End User: Prof. Dr. Iffat Yazdani, Consultant Nephrologists, President Welfare Society for Patient Care
3.	Inspection Authority	i) End User: Prof. Dr. Iffat Yazdani President Welfare Society for Patient Care ii) Representative of Health Department (Development Wing) Gov. of Sindh.


MEDICAL SUPERINTENDENT



MUHAMMAD MUSLIM STAMIR VENDOR
No. 83 Shop # 8, D.C. Central Karachi

No. 35308 Date 27 NOV 2014

Contract Agreement

Letter of Intent No. MS/CHK/2014-2015/2283 Dated 09-06-2015.

This agreement made this 09th June, 2015, between the Medical Superintendent, Civil Hospital, Karachi "The Purchaser" herein after and M/s. Elate CC (Pvt.) Ltd., Address: Suit No. 1,2,3, St plot-3 Block-3, Gulshan-e-Iqbal, Karachi, herein the supplier".

Whereas the purchaser is desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quote the rate which has accepted by purchase for supply of the following goods in the sum of Rs. 395,000/- herein after " The Contract Price"

Item #	Description of Stores	Rate On DDP Basis	Quantity Ordered	Total value in figures.
16.	Defibrillator with all Standard Accessories Model: XD1, Make: Primedic - Germany	Rs. 395,000/- Each	01 No.	395,000.00
Total Amount				Rs. 395,000/-

The delivery period will commence from the date of acceptance of the contract by the bidder.

Now this agreement witnessed as following:

1. In this agreement words and expression shall have the same meanings as are respectively assigned to them in the Terms & Conditions of Tender Enquiry referred to.
2. Following document shall be seemed top form and be read and constructed as per of this agreement, viz:
 - a. Contract Agreement.
 - b. Price Schedule.
 - c. Technical Specifications.
 - d. Terms & Conditions of Tender Enquiry.
 - e. Terms & Conditions of the Contract.
 - f. Schedule of requirement.

Muhammad Quraishy
Prof. M. Quraishy
Medical Superintendent
Civil Hospital, Karachi.

g. Bidder offer including the relevant correspondence with all annexes made before signing the contract.

3. In consideration of the payment to be made by the purchase to the, we supplier as herein after mentioned, the supplier hereby covenant with the purchase to provide the goods and services and to remedy the defects therein conformity in all respects with the provision of the goods.
4. The purchase hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects therein, the contracts price payable under the provision of the Contract at the time and In the manner prescribed in the contract.

In witness whereof the parties here to here caused this agreement in accordance with their respective hands and seals, the day, the month and the year first above written.

SUPPLIER

SIGNATURE & SEAL



NAME : ELATE CONSULTANTS PVT. LTD.

Designation : _____

Address : Suite #1, 2, 3 ST-3 Block-3
CHIKSHAN IQBAL KARACHI.

PURCHASER

SIGNATURE & SEAL

NAME : _____

Designation : Prof. M. Saeed Quraishy
Medical Superintendent
Civil Hospital, Karachi.

Address : _____

WITNESS (SUPPLIER)

SIGNATURE & SEAL

NAME : SYED MOHD HASHIM

Designation : MANAGER ACCOUNTS

Address : FLAT #408 SHARAD PRIDE SURJANT
TOWN KARACHI

WITNESS (PURCHASER)

SIGNATURE & SEAL

NAME : _____

Designation : Dr. Sohail Mir Memon
Addl. Medical Superintendent
(Procurement)

Address : Civil Hospital, Karachi.



BID EVALUATION REPORT

**MEDICAL EQUIPMENT FOR
WELFARE SOCIETY FOR
PATIENTS CARE UNDER ADP
SCHEME # 787/2014-15
THROUGH CIVIL HOSPITAL
KARACHI**

CIVIL HOSPITAL, BABA E URDU ROAD, KARACHI

Ph. +92 21 99215740-45

DAWN

DAWN SUNDAY MAY 10, 2015



Civil Hospital

Baba-e-Urdu Road, Karachi

Ph: No. 99215727 — Fax No. 99215733

NO: (CHK)/2014-2015/3163

DATED the 7th May, 2015

NOTICE INVITING TENDERS

Civil Hospital Karachi, invites sealed bids for the supply of following categories on **Single Stage-Two Envelope procedure** as per SPP Rules-2010 (Amended 2013/14) from the authorized agents, manufacturers registered with General Sales Tax / Income Tax Authority, Sindh Sales Tax (if applicable). A complete set of tender enquiry (SBD) with terms and conditions can be downloaded from Authority's website. It can also be obtained from Procurement Office of this Hospital and the same should be dropped by hand / mail on **25th May, 2015 at 11:00 am** along with pay order of the Tender Fee in favour of the undersigned.

The bids must be dropped by hand / mail on 25th May, 2015 at 11:00 am along with all the required documents in the office of the Addl. Medical Superintendent (Procurement) CHK @ 1st Floor, Admin Block, Civil Hospital, Karachi. These will be opened publicly in the presence of bidders or their authorized representatives, who choose to attend in the Board Room, 2nd Floor, Admin Block, Civil Hospital, Karachi at **12:00 noon** on the same day.

S. No.	Name of Items	Bid Security	Tender Fee	Date of Opening
1.	Supply of Medical Equipment for Welfare Society for Patient Care under ADP Scheme # 787/2014-15.	2.5%	Rs. 1000/-	25.05.2015

1. Technical and Financial proposals should be submitted in separate envelopes.
2. In case of announcement of Public Holiday or any unfavorable circumstances, the tenders / bids will be opened on next working day. Other terms and conditions, venue and time for dropping and opening will remain the same as advertised in this Notice.
3. All quotations shall include Government Taxes (if applicable) and these will be deducted from the bills, according to Government Rules / Policy.
4. D.D.P (in Pak Rupees) prices should be quoted.
5. Bid Security @ 2.5% of total value of the quoted items by firms in shape of Pay Order / Demand Draft must be attached with "Financial Bid" in favor of the undersigned.
6. The Hospital Procurement Committee reserves the right to reject the tender subject to relevant provision of SPP Rules 2010. (Amended-2013)

آئی ایم ایس اور ڈیپارٹمنٹ کی طرف سے جاری ہے۔

MEDICAL SUPERINTENDENT
CIVIL HOSPITAL KARACHI

INF-KRY No. 1797/15

MONDAY MAY 11, 2015.

روزنامہ جنگ
SS-023
11/05/2015
Monday

بنی میر ظلیل الرحمن

جنگ سب سے بڑی

21 رجب المرجب 1436ھ 11 مئی 2015ء نمبر 130

سول اسپتال

پانے اردو ڈاکر ایفون 99215727 99215733 فیکس

نمبر (CHK)/2014-2015/3163 مورخہ 07 مئی 2015

نوٹس طلبی سینڈرز

سول اسپتال کراچی کو ریجنل ڈپٹی کمشنر کی سپاہی کیلئے SPP ریٹز 2010 (ترمیم شدہ 2013) کے مطابق واحد مرحلہ۔ دو اطلاق طریق کار کے تحت جنرل سٹیٹس / آف فیکس / سندھ سٹیٹس / آف فیکس (آرگنائزیشن) کے پاس رجسٹرڈ بااختیار ایجنٹوں، میڈیکل پروفیشنلز سے سربراہ پیشکشیں مطلوب ہیں۔

نیشنل ریکروٹمنٹ (SBD) کا ایک مکمل سیٹ مع شرائط و ضوابط اتھارٹی کی ویب سائٹ سے ڈاؤن لوڈ کیا جاسکتا ہے۔ اسے پرو کیورمنٹ آفس ہڈا سے بھی حاصل کی جاسکتا ہے جسے 25 مئی 2015 دن 11:00 بجے تک مع نیشنل فیکس کے پے آرڈر ذریعہ ذریعہ پیشگی طور پر بذریعہ ایک پہنچا دیا جائے۔

پیشکشیں 25 مئی 2015 دن 11:00 بجے تک مع تمام درکار دستاویزات دفتر ایڈمنسٹریشن میڈیکل سپرٹنڈنٹ (پرو کیورمنٹ) سول اسپتال کراچی واقع پگلی منزل ایڈمن ہڈاک، سول اسپتال کراچی میں ذریعہ ایک پہنچا دیا جانا چاہیے۔ آفیس ہورڈروم واقع دوسری منزل، ایڈمن ہڈاک، سول اسپتال کراچی میں دوپہر 12:00 بجے میں حاضر رہنے کے خواہشمند پیشکش دہندگان یا ان کے مجاز نمائندوں کی موجودگی میں سرعام کھولا جائیگا۔

نمبر شمار	آئٹم کا نام	ہڈ سیکورٹی	نیشنل فیکس	کھلنے کی تاریخ
1	سپاہی ہڈا میڈیکل کیمپنٹ برائے ہڈا سہ ماہی برائے پیشکشیں	2.5%	1000/- روپے	25-05-2015
	پیشکش کیلئے اسے ڈی پی آئی نمبر 787/2014/15			

- 1۔ ٹیکنیکل اور فائنل پروپوزلز جداگانہ لفافوں میں جمع کرائی جانی چاہئیں۔
- 2۔ سرکاری تعطیل کا اعلان ہونے یا ناموافق حالات پیدا ہونے کی صورت میں نیشنل ریزرو پیشکشیں آئندہ ایم کار پر کھولی جائیں گی۔ دیگر شرائط و ضوابط نیشنل ریزرو کے لئے اور کھلنے کا مقام اور وقت برطانیہ نیشنل ریزرو سے برقرار رہیں گے۔
- 3۔ تمام کوئٹیشن میں تمام سرکاری ٹیکسز (بصورت قابل اطلاق) شامل ہوں گے اور آفیس ہڈا میں سے برطانیہ سرکاری قواعد یا کسی منہا کر لیا جائے گا۔
- 4۔ ڈی ڈی پی قیمت (پاکستانی روپے میں) پیش کی جائے۔
- 5۔ فرم کی جانب سے پیشکش کی کل مالیت کی 2.5% رقم کی ہڈا سیکورٹی ہنگل پے آرڈر ذریعہ ضمانت ڈرافٹ بحق ذریعہ پیشگی "فنانس ہڈا" کے سر اوائل نامہ تسلیم کی جائے۔
- 6۔ ہڈا سٹیٹس پرو کیورمنٹ سٹیٹس میں پیشکشیں ہڈا سے سب سے زیادہ مسترد کر دے۔ دستخط:۔ **میڈیکل سپرٹنڈنٹ**

سول اسپتال کراچی (INF/KRY-1797/15)

Daily

KAWISH

روزانه

ڪاوش

جنر 09 مئي 2015 ع

هڪ ئي وقت ڪراچي، حيدرآباد ۽ سکر مان شايع ٿيندڙ پهرين سنڌي اخبار

سول اسپتال، باب اردو روڊ، ڪراچي

No. (CHK) 2014-2015/3163

Dated: 7/05/2015



فون نمبر: 99215727، فیکس نمبر: 99215733

ٽيندر گھرائڻ جو نوٽيس

سول اسپتال ڪراچي، سبڙ ٽيڪس/آنٽي ٽيڪس اختيارين وٽ رجسٽرڊ ٿيل مجاز ايجنٽ مينوليٽي چورس کان سگنل اسٽيج-ٽو اينوليٽي طريقيڪار تحت ايس بي بي رولز-2010 (ٽرميمر ٽيبل 2013/14) تحت هيٺ ڄاڻايل ڪنٽريڪٽ جي فراھمي واسطي مھربند وائڪ گھرائڻ تي.

ٽيندر آنٽي ٽيڪس (SBD) جو هڪ مھربند ميٽ سان شرط ۽ ضابطا تيار ٿي، ويب سائيٽ تان ڏکڻ لاء ڪري سگھجن ٿا. اھي پڻ هن اسپتال جي پروفيسور ميٽ آفيس مان پڻ حاصل ڪري سگھجن ٿا. ساڳيا هيٺ صحيح ڪندڙ جي حذر ٽيندر ٿي جي بي آرڊر سميت 25 مئي، 2015 تي صبح 11:00 وڳي، شوٽ/ٿيل معرفت زما وٺين.

واحد لاء ٽي طرح شوٽ/ٿيل معرفت 25 مئي، 2015 تي صبح 11:00 وڳي سمورن گھريل ڪاغذن سميت ايڊيشنل ميڊيڪل سپرنٽينڊنٽ (پروفيسور ميٽ CHK) اھت فرسٽ فلور ايڊمن بلاڪ سول اسپتال، ڪراچي، جي آفيس پري وٺا وڃن. اھي سر عامر واحد ٽينڊرن يا سندن مجاز نمائندن جيڪي موجود رھن چاھين جي رويرو بورڊ آف، سيڪنڊ فلور ايڊمن بلاڪ، سول اسپتال ڪراچي پراڻي ٽينڊرن منعقد 12:00 وڳي ڪرايا ويندا.

نمبر شمار	آئٽم جو نالو	واڪ سيڪيورٽي	ٽينڊر في	ٽينڊر ڪولڻ جي تاريخ
1.	اي ڊي پي اسڪيم نمبر 787/2014-15 تحت فرينس جي سار سنڀال واسطي ويلفيئر سوسائٽي لاء ميڊيڪل ايجنٽي جي فراھمي.	2.5%	1000 روپيا	25-05-2015

- ٽيڪنيڪل ۽ فنانشل پروپوزس ڏاڙ ڦار لھان پر جمع ڪرائڻ گھرجن.
- عام موڪل يا ڪنهن اٿوئنڊ حالتن جي صورت ۾ ٽينڊر/واڪ ورٽڊ ڪم ڪار واري ڏينھن تي ڪرايا ويندا، ٻيا شرط ۽ ضابطا، ٽينڊر نمائڻ ۽ ڪولڻ واسطي هڪ ۽ وقت هن نوٽيس پر مشتمل ڪيل موجب ساڳيو هوندو.
- سمورا ڪنٽريڪٽ بشمول گورنمينٽ ٽيڪسز (جيڪڏهن لاڳو ڪرڻ جو چوڳو هجي) ۽ سرڪاري ڏاڪڻن/پاليسي مطابق بلن مان ڪٽائيا ويندا.
- ڊي ڊي پي (پاڪستاني رين پرا) ٽيڪسٽن ڄاڻائڻ گھرجن.
- فرم پاران ڄاڻايل آئٽم جي جملي قيمت جي 2.5% تي واڪ سيڪيورٽي پي آرڊر/مانڊ ڊرافٽ هيٺ صحيح ڪندڙ جي حق ۾ "مالياتي واڪ" سان لازمي شامل هئڻ گھرجن.
- اسپتال پروفيسور ميٽ ڪميٽي ايس بي بي رولز 2010 (ٽرميمر ٽيبل 2013) جي لاڳاپيل تشريعات جي شرط سان ٽينڊر وڌ ڪرڻ جو حق محفوظ رکي ٿي.

ميڊيڪل سپرنٽينڊنٽ
سول اسپتال ڪراچي

INF/KRY.No.1797/2015



SPPRA

Sindh Public Procurement Regulatory Authority Government of Sindh

- Home
- Public Tenders
- SPPRA Notifications
- SPPRA Board
- SPP Act 2009
- SPP Rules 2010
- Amendment 2010
- SPPRA Team
- Contact Us

- Tenders
- Committee
- Sample Documents
- Policy Instructions
- Review Panel
- Black Listed Firms
- Capacity Building
- Complaints
- Contract Evaluation Form
- Departments
- Download Statistics
- Downloads

Serial No.

Text

Tender Title

SPPRA ID

Advertise Date

Closing Date

City

submit

Advance Search

WORKS
SERVICES
GOODS

1 Page 1 Documents Viewed 772023

S.No.	SPPRA ID	Tender Title	Download	Uploaded On	Advertised On	Closing On	Contingendum	Comments
24148	SPPRA ID WTH-ELD	Civil Hospital Karachi Medical Equipments		12 May 2015	12 May 2015	25 May 2015	SBD 12 May 2015	SEE. CS. TER. AOC Required



Government of Sindh
Health Department

NOTIFICATION.

No.HD(P&E)3-2(426)/2014-15 In supersession of the Health Department Government of Sindh notification of even number dated 22-04-2015 and subsequent notification issued on 12-05-2015 Procurement Committee comprising of following members is constituted for procurement in respect of development scheme "Expansion & Strengthening of Dialysis Facilities Through WSPC(Welfare Society for Patients Care) during current financial year 2014-15 under Rule-7 of Sindh Public Procurement Rules,2010 (Amended 2013)

- | | | |
|----|--|----------|
| 01 | Dr.Suhail Mir Memon,
A.M.S (Procurement) Civil Hospital Karachi. | Chairman |
| 02 | Dr.Sikandar Rafique Qureshi,
Assistant Professor,Shaheed Mohtarm Benzîr Bhutto
Medical College, Karachi. | Member |
| 03 | Mr.Waheed Shaikh,
Deputy Director,
Excise & Taxation Department,
Government of Sindh,Karachi. | Member |
| 04 | Representative of Health Department (Development Wing)
Government of Sindh,Karachi | Member |
| 05 | Mr.Waris Khalil,
Divisional Accounts Officer,
office of the Accountant General Sindh, Karachi | Member |

The committee shall be held responsible to observe all the codal formalities as per Sindh Public Procurement Rules, 2010 (Amended 2013).

SECRETARY HEALTH
GOVERNMENT OF SINDH,KARACHI.
Karachi/Dated 24-05-2015

No.HD(P&E)3-2(426)/2014-15

C.c to:

- The Secretary Excise & Taxation Department Government of Sindh, Karachi with a request to kindly nominate a member / representative for Procurement Committee.
- The Accountant General Sindh, Karachi.
- The Managing Director Sindh Public Procurement Regularity Authority, Karachi.
- All members of the Committee.
- PS to the Secretary Health Government of Sindh, Karachi.
- PA to the Additional Secretary (Dev) Health Department Government of Sindh, Karachi.

Deputy Director (Dev-II)

Date:- 25-05-2015.

A Hospital Procurement Committee meeting held on 25-05-2015 at 12.00 noon in the Committee Room of Civil Hospital, Karachi under the Chairman of Dr. Sohail Mir Memon, Addl. Medical Superintendent (Procurement) Civil Hospital, Karachi to open the **Technical Bids** for the **Supply of Medical Equipment for Welfare Society for Patient Care** under ADP Scheme # 787/2014-15, during the financial year 2014-2015 thought C.H.K.

The following Officers have attended the meeting:

S.NO	NAME OF OFFICER(s)	SIGNATURE
1)	DR SOHAIL MIR MEMON	
2)	DR. SHANWAR RAFIQUE QURATSHI	
3)	Bader Shauk, Dy Director	
	Wahed Sheikh (Dev) Health	
	MR. Wazir Khalil JAO	

Sealed tenders were invited through the publication of Tender Notice in prominent News Papers i.e. The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 and other News Papers and hoisted on SPPRA website @ Sr. No.24148 for the Supply of Medical Equipment for Welfare Society for Patient Care under ADP Scheme # 787/2014-15, thought Civil Hospital Karachi.

The Technical Bids opened on 25-05-2015 at 12.00 noon in the Committee Room, Civil Hospital, Karachi before the Procurement Committee.

The following firms have participated in the competition, the representatives who were present at the time of opening tender their signatures were taken as a token of presence.

S. No.	Name of Firms	Signature of Representative
1 ✓	M/s S. Sjozuddin & Co.	
2 ✓	M/s Hospital Supply Corp	
3 ✓	M/s Life One	
4 ✓	M/s BIOS	Usman Ali
5	M/s Meer International	S. SHAJA UR REHMAN
6 ✓	M/s. Aspirecare Systems	Zulfi
7 ✓	M/s Zam Trading	Saeed Anjum
8 ✓	M/s Technology Links (Pvt) Ltd	Fahim 03332750819
9 ✓	M/s Ontech Corporation	huf 0345-3090020
10 ✓	M/s Siemens Pakistan Eng.	Javed 0322-2224495 Jawaid
11 ✓	M/s. Medequips SMC	Ahmed (0333-3877242)
12 ✓	M/s. Elate. cc (Pvt)	Imdad 0311-0209044

for 25/5/15

S. No.	Name of Firms	Signature of Representative
13	M/s Karachi Medical Co	H. Khan
14 ✓	M/s Universal Enterprise	[Signature]
15	M/s. Shurazi Trading.	[Signature]
16 ✓	M/s Boro track Trading	[Signature] 0346-2787775
17	M/s Oasis Medical Systems	Talhar M. Nadeem
<p>Total (17) Seventeen firms participated in the competition.</p>		

25/5/15

[Signature]
25/5/15
25/5/15

TECHNICAL STATEMENT

PURCHASE OF MEDICAL EQUIPMENT FOR
WELFARE SOCIETY FOR PATIENT CARE UNDER

ADP SCHEME # 787/2014-15

THROUGH

CIVIL HOSPITAL, KARACHI

DURING THE FINANCIAL YEAR

2014-2015



TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

1

Name of Item.

Hemodialysis Machine with all Standard Accessories

Quantity Required

10 Nos.

Estimated Price

Rs. 1,000,000.00

Country of Origin

USA / EEC / JAPAN

DK

Name of Firms	Hospital Supply Corp.	M/s. Universal Ent.
Make	Toray	Nipro Corp
Model	TQS-88	Surdial 55
Dialysis machine complete with battery backup	Yes	Yes
Dialysis Machine Capable of Acetate Dialysis, Bicarbonate Dialysis and Isolated Ultra-filtration	Yes	Yes
LCD Touch control Panel, 8" or more screen.	Yes	10.4"
Dialysis treatment parameters such as UF Goal, UF Rate, Blood Flow and Syringe Flow 2 or more Selectable Concentrate	Yes	Yes
Double Chamber System to precise UF Volumetric control	Yes	Yes
CPU System for Control		
Driving and Monitoring Functions: Sensors for Syringe check for Syringe fitting and speed	Yes	Yes
Machine should perform self check of the closed circuit, capable to monitor reverse rotation of blood pump	Yes	Yes
Supply Water Flow 600 ml/min or more	800ml/min	750ml/min
Dialysis Flow 400-600 ml / min	300-700ml/min	300-700ml/min
UF rate 0.00 – 3.00 L / h or better	Yes	0, 0.10 – 5.00 L/h
UF volume 0.00 – 35L or more Dialysate type Bicarbonate / Acetate	Yes	Yes
Dialysis Conductivity 12.0 – 16.0mS/cm or better	Yes	Yes
Dialysis Step 1 – 6 (0.3Mm/ cm per one step)	Yes	Yes
Blood Pump Flow Rate 20-600ml/min or better	Yes	Yes
Heparin Infusion rate 10ml-20ml/h.	Yes	Yes
Detector		
Air bubble detector. Ultrasonic detection system	Yes	Yes
Blood leak detector Photoelectric detection system	Yes	Yes

Wahed Shaikh
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi

WAKS KARACHI
 Divisional Accounts Officer
 Accident & Emergency Centre &
 Ancillary Services Complex A1
 Civil Hospital Karachi

(Signature)
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi

(Signature)
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi

Monitor Section		
Dialysate temperature indicator	Yes	Yes
Dialysate flow indicator	Yes	Yes
Dialysate conductivity indicator	Yes	Yes
Dialysate pressure indicator	Yes	Yes
Arterial pressure indicator	Yes	Yes
Venous pressure indicator	Yes	Yes
Transmembrane pressure indicator	Yes	Yes
UF volume indicator	Yes	Yes
Blood pump flow indicator	Yes	Yes
Heparin pump flow indicator	Yes	Yes
Disinfection and Rinsing	Yes	Yes
Chemicals Sodium hypochlorite	Yes	Yes
Audio/visual alarms for conductivity , pressures, water shortage, concentrate shortage, air bubble, blood leak etc.	Yes	Yes
Country of Origin : USA / Europe / Japan	Japan	Japan
Delivery Period:	As per TE	As per TE
Warranty:	One Year	One Year
Catalogue / Brochure should be attached duly flagged.	Yes	Yes

as









TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital Karachi

Item No. 01 HEMODIALYSIS MACHINE WITH ALL STANDARD ACCESSORIES

M/s. Hospital Supply Corporation

M/s. Hospital Supply Corporation has quoted model: TQS-88, TMQ Toray from China

M/s. Hospital Supply Corporation has quoted Country of Origin China instead of USA / Europe / Japan. Which does not meets tender specification / requirement.

Offer of M/s. Hospital Supply Corporation is declared as non-responsive.

M/s. Universal Enterprises

M/s. Universal Enterprises has quoted model: Surdial 55 from Nipro Corporation – Japan

Quoted model of M/s. Universal Enterprises meets tender specification.

Offer of M/s. Universal Enterprises is declared as responsive.



TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

2

Name of Item.

Reprocessing Unit with all Standard Accessories

Quantity Required

01 No.

Estimated Price

Rs. 1,000,000.00

Country of Origin

USA / ECC / JAPAN

Name of Firms	M/s. Hopital Supply Corp	M/s. Universal Ent.
Make	Meditop Kidney Kleen	Medivator
Model	Compact	Renatron
Self-diagnosis and user programmable dialyzer re-processing system.	Yes	Yes
Dialyzer rinse process with check connection	Yes	Yes
2.5% concentrate preparation /clearing and rinsing	Yes	Yes
Dialyzer test process with TCV test and pressure test	Yes	Yes
Dialyzer sanitization processes rinsing 3.5% concentrate preparation	Yes	Yes
Rinsing and filling 3.5% concentration	Yes	Yes
Country of Origin : USA / Europe / Japan	Thailand	USA
Delivery Period:	As per TE	As per TE
Warranty:	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes

Handwritten mark

Handwritten signature

Large handwritten signature

Handwritten signature

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

ITEM NO. 2 REPROCESSING UNIT

M/s. Hospital Supply Corp

M/s. Hospital Supply Corporation has quoted model: Compact from Meditop Kidny.

Quoted model of M/s. Hospital Supply Corp. is from Thailand which is against tender required country of origin.


Due to above mentioned reason, the offer of M/s. Hospital Supply Corp is declared as non-responsive.

M/s. Universal Enterprise

M/s. Universal Enterprise has quoted model: Renatron from Medivator-USA.

Quoted model of M/s. Universal Enterprise meets tender specification.

Offer of M/s. Universal Enterprise is complete as per tender specification hence may consider as responsive.



Iffat Yazdani
Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

3

Name of Item.

OT Light with all Standard Accessories

Quantity Required

01 No.

Estimated Price

Rs. 750,000.00

Country of Origin

Korea / Taiwan / Turkey

Name of Firms	M/s. Life One	M/s. Tehnology Links	M/s. Universal Ent.	M/s. Oasis Medical System
Make	AFA	Reunmed	Uzumcu	DomelUX 6200
Model	ALB-100C	ROL-51C	DUO X1	DomelUX 6200
Single cupola Ceiling OT Light	Yes	Yes	Yes	Yes
Halogen lamp	Yes	Yes	LED	Yes
Color temperature 4300k or better	Yes	Yes	4270K	not mentioned
Lamp dia. (Single head) : 62cm	Yes	Yes	650cm	not mentioned
Polygon Reflector	Yes	Polygon not mentioned	Yes	not mentioned
110,000LUX or more intensity	Yes	Yes	160,000Lux	140,000Lux
Cri: 93± 3	Yes	Yes	Yes	Yes
Action radius: 170cm or better	Yes	Yes	Yes	not mentioned
CE or FDA Approved	CE	CE	Yes	CE
Country of Origin : Korea / Taiwan / Turkey	Taiwan	Taiwan	Turkey	Taiwan
Delivery Period:	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes

OK

APC

[Handwritten signature]

[Handwritten signature]

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 3 OT LIGHT

M/s. Life One

M/s. Life One has quoted model: ALB 100C from M/s. AFA-Taiwan.

Quoted model of M/s. Life One meets tender specification.

Offer of M/s. Life One is declared as responsive.

M/s. Universal Enterprise

M/s. Universal Enterprise has quoted model: Due X1 from Uzumsu-Turkey

Quoted model of M/s. Universal Enterprise meets tender specification.

Offer of M/s. Universal Enterprise is declared as responsive.

M/s. Owasis Medical

M/s. Owasis Medical has quoted model: Bomelux 6200 from Benqu-Taiwan

M/s. Owasis Medical has not enclosed the technical brochure to verify the tender specification parameters such as color temp, Cri, lamp diameter, polygon reflector, radius etc.


Due to above reason, offer of M/s. Owasis Medical is declared as non- responsive.

M/s. Technology Link

M/s. Technology Link has quoted model: ROL 51C from Rexmed-Taiwan

Quoted model of M/s. Technology Link meets tender specification.

Offer of M/s. Technology Link is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

4

Name of Item.

OT Table with all Standard Accessories

Quantity Required

01 No.

Estimated Price

Rs. 650,000.00

Country of Origin

Korea / Taiwan / Turkey

Name of Firms	M/s. Life One	M/s. Tehnology Links	M/s. Universal Ent.	M/s. Oasis Medical System
Make	St-Francis	Rexmmed	Bilcakkilar	Bengq
Model	OT-125 A	ROT-160X	5505	Novel-330T
Multipurpose hydraulic Operation Table	Yes	Yes	Yes	Yes
360° Tabletop rotation	Yes	Yes	Yes	not mentioned
Stainless steel side rail, accessories and clamps	Yes	Yes	Yes	Yes
Built-in kidney lifter	Yes	Yes	Yes	Yes
Table Top : 180 cm or more	Yes	Yes	Yes	Yes
Table Top Width: 50 cm or more	Yes	Yes	Yes	Yes
Table Top Elevation: 70-100 cm or more	Yes	74-104 cm	Yes	71 ~ 95 cm
Lateral Tilt: 25°/25° or more	Yes	Yes	25°/25°	20°/20°
Trendelenburg / Reverse: 20°/25° or more	22°/28°	Yes	Yes	25°/25°
Head Section Up/Down: +30° /- 90°	Yes	Yes	Yes	+60° /- 90°
Leg Section Up/Down: +20° /- 90°	Yes	Yes	Yes	+0° /- 90°
Table should complete with:				
1 x Screen Frame	Yes	Yes	Yes	NQ
1 x Shoulder Supporter	Yes	Yes	Yes	NQ
1 x Body Support	Yes	Yes	Yes	NQ
1 x Arm Rest	Yes	Yes	Yes	NQ
1 x Knee Crutches	Yes	Yes	Yes	NQ
1 x Mattress	Yes	Yes	Yes	NQ
1 x Kidney Elevator	Yes	Yes	Yes	NQ
CE or FDA Approved	CE	CE	CE	CE
Country of Origin : Korea / Taiwan / Turkey	Taiwan	Taiwan	Turkey	Turkey
Delivery Period:	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes

Handwritten initials

Handwritten signature

Handwritten signature

Handwritten signature

Handwritten signature

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 04 OT TABLE

M/s. Life One

M/s. Life One has quoted model: OT-125A from M/s. St-Francis-Taiwan.

Quoted model of M/s. Life One meets tender specification.

Offer of M/s. Life One is declared as responsive.

M/s. Universal Enterprise

M/s. Universal Enterprise has quoted model: 550S from Bicakcilar-Turkey

Quoted model of M/s. Universal Enterprise meets tender specification.

Offer of M/s. Universal Enterprise is declared as responsive.

M/s. Owasis Medical

M/s. Owasis Medical has quoted model: Novel -330T from Benqu-Taiwan

M/s. Owasis Medical has not enclosed the technical brochure to verify the tender specification parameters .
They have not mentioned required accessories in their technical offer hence offer of M/s. Owasis Medical is
declared as non- responsive.

M/s. Technology Link

M/s. Technology Link has quoted model: ROT-160X from Rexmed-Taiwan

Quoted model of M/s. Technology Link meets tender specification.

Offer of M/s. Technology Link is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

5

Name of Item.

Surgical Beds (Dialysis Chair) with all Standard Accessories

Quantity Required

03 Nos.

Estimated Price

Rs. 1,65,000.00

Country of Origin

LOCAL (Export-Quality)

Name of Firms	M/s. Hospital Supply Crop Nanning Passon Medical	M/s. Zam Trading Winmed	M/s. Tehnology Links
Model	PY-YD 340	WM3000	ME-310
Width 600mm	Yes	Yes	width not mentioned
Length 1900mm	Yes	Yes	1680mm
Seat Width including arms 880	Yes	Yes	width not mentioned
Height Adjust 560mm ~ 820 mm	Yes	Yes	adjusted from 290mm
Frame Q235 Steel	Yes	Yes	ABS
Cushion PU Leather	Yes	Yes	Yes
Castor 100mm	Yes	Yes	Yes
Back Adjust 12° ~ 75°	Yes	Yes	-10° ~ 80°
Leg Adjust 80° ~ 12°	Yes	Yes	10° ~ 80°
Tendelenburg 12°	Yes	Yes	Yes
Covering ABS	Yes	Yes	Yes
Motors 3	Yes	Yes	Yes
Power Supply AC 110~240 V 50/60 Hz	Yes	Yes	Yes
Total Weight 90 KG	Yes	Yes	Yes
Country of Origin : LOCAL (Export Quality)	China	Local	China
Delivery Period:	As per TE	As per TE	As per TE
Warranty:	One Year	As per TE	As per TE

AP

Sharma

Yes

[Signature]

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 5 SURGICAL BEDS (DIALYSIS CHAIR)

M/s. Hospital Supply Corp.

M/s. Hospital Supply Corp has quoted model:PY-YD 340 from China.

Quoted model of M/s. Hospital Supply meets tender specification.

Offer of M/s. Hospital Supply is declared as responsive.

M/s. ZAM Trading

M/s. ZAM has quoted model: WM-3000 from Winmed-Pakistan

Quoted model of M/s. ZAM meets tender specification.

Offer of M/s. ZAM Trading is declared as responsive.

M/s. Technology Link

M/s. Technology Link has quoted model: ME-310 from China.

Quoted model of M/s. Technology Link meets tender specification.

Offer of M/s. Technology Link is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price
Country of Origin

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015
6
Surgical Beds (Dialysis Recliners) with all Standard Accessories
03 Nos.
Rs. 165,000.00
LOCAL (Export Quality)

Name of Firms	M/s. Hospital Supply Crop Nanning Passion Medical	M/s. Zam Trading Winmed	M/s. Tehnology Links
Model	PV-YD 380	WMMS001	ME-3805
Total Length 1920mm	Yes	Yes	length not mentioned
Width 700mm	Yes	Yes	width not mentioned
Height 520mm ~ 840mm	Yes	Yes	560mm ~ 800mm
Back Segment Angle $0^{\circ} \sim 80^{\circ}$ 45	Yes	Yes	$0^{\circ} \sim 70^{\circ}$
Leg Segment Angle $45^{\circ} \sim 25^{\circ}$ 45	Yes	Yes	$-45^{\circ} \sim 0^{\circ}$
Castor 125mm	Yes	Yes	Yes
Upholstery P U	Yes	Yes	Yes
NW 90 Kg	Yes	Yes	Yes
Safe maximum load 240 Kg	Yes	Yes	Yes
Motors 3	Yes	Yes	Yes
Power Supply AC 110-240 V 50/60 Hz 100 VA	Yes	Yes	Yes
Digital Weighing Range 0-240 kg	Yes	Yes	not mentioned
Country of Origin : LOCAL (Export Quality)	China	Local	China
Delivery Period:	As per TE	As per TE	As per TE
Warranty:	One Year	As per TE	As per TE

AS

Chhanna

AS

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 6 SURGICAL BEDS (DIALYSIS RECLINERS)

M/s. Hospital Supply Corp.

M/s. Hospital Supply Corp has quoted model: PY-YD 380 from China.

Quoted model of M/s. Hospital Supply meets tender specification.

Offer of M/s. Hospital Supply is declared as responsive.

M/s. ZAM Trading

M/s. ZAM has quoted model: WM-5001 from Winmed-Pakistan

Quoted model of M/s. ZAM meets tender specification.

Offer of M/s. ZAM Trading is declared as responsive.

M/s. Technology Link

M/s. Technology Link has quoted model: ME-380S from China.

Quoted model of M/s. Technology Link meets tender specification.

Offer of M/s. Technology Link is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

7

Name of Item.

Crash Trolley with all Standard Accessories

Quantity Required

03 Nos.

Estimated Price

Rs. 150,000.00

Country of Origin

UK / USA / EEC/Japan

Name of Firms	M/s. BIOS	M/s. Eliat cc	M/s. Karachi Medical Company
Make	Hersill	France Hospital	Crosmed
Model	CAVS	Perso-Life	C1122
Crash trolley with 4 swivel castors with brakes	Yes	Yes	Yes
Having 4 drawers or more	7 Drawers	Yes	Yes
Top tray for defibrillator	Yes	Yes	Yes
Side tray for instrument	Yes	Not Mentioned	Not Mentioned
Complete with adult and child resuscitator with face mask.	Yes	Yes	Yes
Venture suction with collection jar	Yes	Not Mentioned	Yes
Container for waste	Yes	Yes	Yes
Drip stand support	Yes	Yes	Yes
Cardiac board	Yes	Yes	No
Oxygen regulator	Yes	Not Mentioned	No
Flow meter with humidifier	Yes	Not Mentioned	No
Country of Origin : UK / USA / EEC/ Japan	Spain	Italy	China
Delivery Period:	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged:	Yes	Yes	Yes

AP

Adhain

FE

[Signature]

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 7 CRASH TROLLEY

M/s. BIOS

M/s. BIOS has quoted CAVS from Hersill-Spain.

Quoted model of M/s. BIOS meets tender specification.

Offer of M/s. BIOS is declared as responsive.

M/s. Elate Cc.

M/s. Elate Cc has quoted model: Persolife from France-Hopitol


From the attached brochure and technical quotation, it is observed that the mentioned model is a series of different models of combination instead of specific model. In their offer, M/s. Elate cc has not clarify which specific model, they have quoted.

With the above reason, offer of M/s. Elate Cc cannot be evaluated and compared with tender specification hence offer of M/s. Elate Cc is declared as non-responsive.

M/s. Karachi Medical Corp.

M/s. Karachi Medical has quoted model: C1122 from China.

Quoted model is of Chinese origin which is against tender specification hence offer of M/s. Karachi Medical Corp. is declared as non-responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

8

Name of Item.

Stretcher with all Standard Accessories

Quantity Required

13 Nos.

Estimated Price

Rs. 50,000.00

Country of Origin

LOCAL (Export Quality)

Name of Firms	M/s. Hospital Supply Corp	M/s. Zam Trading
Make	Nanning Medical PY-CZ	Winmed ER1000
Model	Not Mention	Yes
Stainless steel	Yes	Yes
With swivel castors	China	Local
Country of Origin : LOCAL (Export Quality)	As Per T.E	As Per T.E
Delivery Period:	One Year	As Per T.E
Warranty:		

CP

[Signature]

M. H. Khan

[Signature]

[Signature]

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 8 STRETCHER

M/s. Hospital Supply Corp.

M/s: Hospital Supply Corp has quoted model:PY-CZ from China.

Quoted model of M/s. Hospital Supply meets tender specification.

Offer of M/s. Hospital Supply is declared as responsive.

M/s. ZAM Trading

M/s. ZAM has quoted model: ER-1000 from Winmed-Pakistan

Quoted model of M/s. ZAM meets tender specification.

Offer of M/s. ZAM Trading is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

9

Name of Item.

Autoclave with all Standard Accessories

Quantity Required

01 No.

Estimated Price

Rs. 1,500,000.00

Country of Origin

Korea / Taiwan / Turkey

Name of Firms	M/s. Life One
Make	AFA
Model	AS-450
Chamber capacity: 143 liters	Yes
Sterilization temperature 121 c	Yes
Design temperature 138 C	Yes
Dry function	Yes
Re-dry function	Yes
Heater power consumption 6 kw	Yes
Frequency 50/60Hz	Yes
Over temperature protection	Yes
Over Pressure protection	Yes
Electric overload protection	Yes
Emergency exhaust switch	Yes
Low water buzzer	Yes
Over heat/temp. Indication light	Yes
Adjustable Sterilize drying time from 0-60 minutes on timer	Yes
Automatic air venting device	Yes
With air filter to prevent contaminated air entrance into the chamber	Yes
Country of Origin : Korea / Taiwan / Turkey	Taiwan
Delivery Period:	As per TE
Warranty:	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi


Item No. 9 AUTOCLAVE

M/s. Life One

M/s. Life One has quoted model:AS-450 from AFA-Taiwan.

Quoted model of M/s.Life One meets tender specification.

Offer of M/s. Life One is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

10

Name of Item.

Ventilator with all Standard Accessories

Quantity Required

04 No.

Estimated Price

Rs. 2,000,000.00

Country of Origin

USA / EEC / JAPAN

Name of Firms	M/s. BIOS	M/s. Noor International	M/s. Hospicare
Make	Event	Hamilton	Purete & Bennett
Model	Inspiration SI	CI	PB 840
Modes:			
Assisted Control Mandatory Ventilation (CMV), Synchronized Intermittent Mandatory Ventilation (SIMV), Spontaneous Ventilation (CPAP + PS, CPAP + VS, SPAP)	Yes	CPAP + VS not mentioned	Yes
Auto Control : Apnea Backup Ventilation	Yes	Yes	Yes
SPAP Philosophy Options: Time, Cycle + Time, Cycle + I:E	Yes	Yes	Not mentioned
IBW Calculator	Yes	Yes	Yes
NIV On/Off in all Modes	Yes	Yes	Yes
Active Exhalation Valve	Yes	Yes	Yes
Automatic Leak Compensation (up to 60 lpm)	Yes	Yes	Leak volume not mentioned
User defined flow sensor position	Yes	Yes	Yes
Proximally at the circuit Wye	Yes	Yes	Yes
Distally at the Expiratory Port	Yes	Yes	Yes
Breath Types			
Volume-controlled (V-CMV; V-SIMV)	Yes	Yes	Yes
Pressure-controlled (P-CMV; P-SIMV; PS)	Yes	Yes	Yes
Volume Targeted Pressure-controlled (PRVC-CMV; PRVC-SIMV; VS)	Yes	Yes	not mentioned
Dual Level PEEP (SPAP)	Yes	Yes	dual level not mentioned
Equivalent to APBV (Carefusion) & BiLevel (Covidien)	Yes	Yes	Yes
Breath Triggering			
Pressure triggering 0.5 - 20 cmH2O	Yes	0.1 - 20 cmH2O	0.1 - 20 cmH2O
Flow triggering 0.1 - 25 lpm	Yes	0.1 - 20 lpm	0.2 - 20 lpm
Respiratory rate 0 - 350 bpm	Yes	Yes	Yes
Tidal volume 2 - 2000 ml	Yes	20 - 2000 ml	5 to 2,500 mL, 2 to 315 mL with NeoMode

PEEP / CPAP 0 - 50 cmH2O	Yes	0 - 35 cmH2O	0 - 45 cmH2O
Psupport 0 - 90 cmH2O	Yes	0-60 cmH2O	0 - 70 cmH2O
I-Times 0.1 - 10 sec	Yes	Yes	0.2 to 8.0 seconds
Oxygen concentration 21 - 100%	Yes	Yes	Yes
Rise time settings 1 - 20	Yes	Yes	Yes
Exhalation sensitivity % 10 - 80% (spontaneous peak flow)	Yes	Yes	Yes
Base Flow (with leak comp off) 2.5 - 25 lpm	Yes	Adults/ped: fixed at 3 l/min Neonatal: fixed at 4 l/min	Not mentioned
Plateau (Insp. pause) 0 - 42 sec (0 - 70%)	Yes	0 - 2000ms	Yes
Manual inspiration, inspiratory and expiratory	Yes	Yes	0.0 to 2.0 seconds
Alarm silence breath holds	Yes	Yes	Yes
100% Oxygen (+20% Oxygen (Neonate) Standby	Yes	Yes	Yes
Multiple Easy View Tabs	Yes	Yes	Yes
Main, Settings, Monitoring, Alarms, Configuration	Yes	Yes	Yes
Manual & Automatic Time Based Event Markers	Yes	Yes	Yes
Graphics Display (3 - Loops and 3 - Waveforms)	Yes	Yes	Yes
Apnea Back-up	Yes	Yes	Yes
Smart Nebulizer	Yes	Yes	Not mentioned
Smart Sigh	Yes	Yes	Yes
Humidification Selection (HME, Humidifier, and None)	Yes	Yes	Yes
Heliox 0 - 80%	Yes	NO	No
User Defined Custom Default Settings	Yes	Yes	Yes
Data Monitoring			
Delivered oxygen concentrations: 21 - 100%	Yes	Yes	Yes
Trend Data (1-72 hrs.) (34 trend parameters)	Yes	Yes	Yes
Alarms			
Auto set alarm feature & 1000 Event log	Yes	Yes	Yes
Complete with Humidifier	Yes	Yes	not quoted
Country of Origin : USA / EEC / JAPAN	California / USA	Switzerland	Malaysia
Delivery Period:	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes

Handwritten signatures and initials:
 A large blue signature, possibly "K. Shankar".
 A green signature, possibly "A.S.".
 A blue signature, possibly "P.R.".
 A blue signature, possibly "M.P.".

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 10 VENTILATOR

M/s. BIOS

M/s. Bios has quoted model: Inspiration 5i from e-Vent Medical-USA

Quoted model of M/s. BIOS is from neonate to Adult.

Offered model meet tender specification such as modes, breath type, breather reading , data monitoring, alarms are as per tender specification.

Offer of M/s. BIOS is declared as responsive.

M/s. Noor International

M/s. Noor International has quoted model: C-1 from Hamilton -Switzerland

From the attached brochure, quoted model do not have mode CPAP+VS as required in the tender specification.

From the attached brochure and technical data sheet , tidal volume is from 20~2000ml instead of required tender specification of 2-2000ml i.e. the quoted model cannot support the ventilation from infant/neonatal.

Offer of M/s. Noor International is declared as non- responsive.

M/s. Hospicare

M/s. Hospicare has quoted model: PB 840 from Puret & Benat from Ireland.

Quoted model of M/s. Hospicare meet tender specification.

Offer of M/s. Hospicare is declared as responsive.

Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No. _____

Item No. _____

Name of Item. _____

Quantity Required _____

Estimated Price _____

Country of Origin _____

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

11

Fluoroscopy with DR System 500mA with all Standard Accessories

01 No.

Rs. 25,585,000.00

USA / EEC / Japan

Name of Firms	M/s. BIOS	M/s. Ontech Corp	M/s. Medequips SMC	M/s. Medequips SMC (Alternat)	M/s. Siemens Pakistan
Make	Stephanix	Hitachi + Technix SPA	Toshiba + Primax	Primax	Siemens
Model	N40HF + Omniscope	Radnext 50 + TCA 6	Mred-A50S + Cyber Bloc	Riviera + Cyber Bloc	Multixtusion + Siremobile Compact I
500mA Digital Radiography machine	Yes	Yes	Yes	Yes	Yes
500mA/125KV High Frequency X-Ray consisting of floating top table	Yes	Yes	Yes	Yes	Yes
Chest bucky stand	Yes	Yes	Yes	Yes	Yes
Floor to ceiling tube stand	Yes	Yes	Yes	floor mounted only	Ceiling mounted only
X-ray tube 200KHU	Yes	Yes	Yes	Yes	Yes
Flat panel detector for table bucky and chest bucky	Yes	Yes	Yes	Yes	Yes
Size 14 x 17, 2,560 x 3,072 pixels	Yes	Z208x2688	2540x3052	2775x2874	2552x3054
complete with work station, CPU, Thermal or laser printer	Yes	Yes	Yes	Yes	Yes
Micro processor control, Three phase , 40 KW high frequency X-Ray Generator	Yes	Yes	Yes	Yes	Yes
Automatic line compensation	Yes	Yes	Yes	Yes	Yes
Over load tube protection	Yes	Yes	Yes	Yes	Yes
Radiographic kVP 40-125KV in one KV step	Yes	Yes	Yes	Yes	Yes
Radiographic msa 10-500mA or more in 18 steps	Yes	10 - 800mA (20 or 39 steps)	18 steps not mentioned	Yes	Yes
mas range: 0.1-500mas	Yes	Yes	Yes	Yes	Yes
Exposure time: 1 ms - 10s	Yes	1 ms - 8s	1 ms - 9s	Yes	1ms with tonormak
2 point kVP and mas selection, 3 point kVP, ma & time selection, 500 or more Anatomical program.	Yes	Yes	3 point	Yes	not mentioned
Four way floating top fixed table:		Yes			
Table top dimension					
length: 200-220 cm	Yes	not mentioned	Yes	Yes	240cm
Height 80-90cm	80cm	not mentioned	68cm	81cm	51.5 - 95.5 cm

Table top weight capacity not less than 250kg	250kg	not mentioned	150kg	180 (no movement limitation) 250 (static)	300kg
Electromagnetic brakes Longitudinal and lateral travel movement	Yes	Yes	Yes	Yes	not mentioned
Super speed moving potter bucky	Yes	Yes	Yes	Yes	Yes
Longitudinal bucky travel	Yes	Yes	Yes	Yes	Yes
Cassette size : 18 x 24 to 35x43cm	Yes	Yes	Yes	Yes	Yes
Moving grid 10:1	Yes	ratio not mentioned	ratio not mentioned	Yes	Yes
Foot switch for electro magnetic brake control.	Yes	Yes	Yes	Yes	not mentioned
Floor Mounted Tube stand					
Balancing System with counterweights	Yes	Yes	Yes	Yes	not mentioned
Electromagnetic brakes	Yes	Yes	Yes	Yes	Yes
Control panel with push button to control vertical, longitudinal and transverse travel	Yes	Yes	Yes	Yes	Yes
Column rotation +/-90deg or better	Yes	Yes	deg. Not mentioned	Yes	Yes
Tube arm rotation : horizontal axis: +/-180deg	Yes	Yes	deg. Not mentioned	Yes	Yes
Goniometer for X-Ray beam incidence angulations indication	Yes	Yes	Yes	Yes	not mentioned
Collimator					
Manual collimator with 6 pairs of leafs. LED light indicator for alignment with bucky	Yes	Yes	not mentioned in attached brochure	Yes	Yes
Field light Lamp with 160LUX. lamp timing: retractable measuring tape	Yes	Yes		Yes	Lux not mentioned
Wall Bucky stand					
Cassettes format from 18X24 to 35x43 on either orientation	Yes	Yes	Yes	13x18 to 35x43cm	Yes
Moving Potter Bucky super speed	Yes	Yes	Yes	Yes	not mentioned
Vertical travel of bucky not less than 150cm	Yes	Yes	120cm	Yes	Yes
Moving grid: 10:1 ratio	Yes	ratio not mentioned	ratio not mentioned	Yes	Yes
X-Ray Tube :					
Focus sizes: Small focus 1.0 mm. Large focus 2.0 mm. Maximum Tube, 125 KVP . Anode Heat capacity 200 KHU	Yes	0.6 (Small) 1.2 (Large), 150KVP	0.6 ~ 1.2 mm	0.6 ~ 1.2 mm	0.6 ~ 1 mm
Anode rotation 2700/3200 rpm		1800/8500 rpm	rpm not mentioned		
Detector					
(1) FLAT PANEL DIGITAL X-RAY DETECTOR (FPD) 2 Nos. for Table bucky and Chest bucky stand	Yes	Yes	Yes	Yes	Yes
Easy upgrade to DR system from General Radiography	Yes	Yes	Yes	Yes	Yes
Highest image quality Detector type : FT - Amorphous selenium or CsI/Amorphous Silicon	Yes	Yes	Yes	Yes	Yes
Detector area: 14 x 17"	Yes	Yes	Yes	Yes	Yes


Pixel resolution: 2500 x 3000 pixels (7.5M) or better	3556x4320	2208x2688	2540x3052	2775x2874	2552x3054
Pixel pitch: 140 µm or less	100 µm	160 x 160 µm	Yes	148 µm	139 µm
Analog to digital conversion 14 bits or better	16 Bits	Yes	Yes	16 Bits	16 Bits
Image Acquisition time: < 10 sec or better for full image	Yes	Yes	Yes	Yes	Yes
Optimum DOE value for low contrast objects	Yes	Yes	Yes	Yes	not mentioned
(2) Workstation and software					
Integrated work station with a Medical grade LCD display of 21" or more	Yes	Yes	Yes	Yes	Yes
Software to support: patient registration, image acquisition, image viewing, image archiving and export	Yes	Yes	Yes	Yes	Yes
Image processing for noise reduction, image enhancement & adjustment, windowing, zoom IN/OUT etc	Yes	Yes	Yes	Yes	Yes
Image manipulation: digital marker, customize annotation, crop, angle and length measurement, rotate, image reject, image save	Yes	Yes	Yes	Yes	Yes
Image lay out : 1 x1, 2 x 1, 1x2 & 2x2	Yes	Yes	Yes	Yes	Yes
DKCOM 3.0 connectivity for store, media storage, print , work list etc	Yes	Yes	Yes	Yes	Yes
Hard ware based with dual core CPU, image disk storage of 500GB or more and interface capability with touch screen or mouse	Yes	Yes	Yes	Yes	Yes
Operating system: Windows XP.	Yes	Yes	Yes	Yes	Yes
C-Arm Fluoroscopy					
Image intensifier for use in urology and other surgery	Yes	Yes	Yes	Yes	Yes
Compact unit consisting of C-Arm control unit and display unit	Yes	Yes	Yes	Yes	Yes
Generator : high frequency inverter type with the frequency of 50KHz or better	Yes	40KHz	Yes	Yes	15 ~ 30 KHz
Out put 2.5Kw or more	4.2Kw	Yes	4.2Kw	4.2Kw	1.4Kw
Stationary or rotating, anode x-ray tube	Yes	Yes	Stationary	Stationary	rotating
Focal spot not more than 0.5 small and 1.5mm large	Yes	Yes	Yes	Yes	single focus 0.6mm
X-ray tube Heat capacity 40KHU or more	Yes	Yes	Yes	Yes	not mentioned
Fluoro mA : 110KV max	Yes	Yes	Yes	Yes	Yes
Fluoro mA : 4~5 mA	Yes	Yes	Yes	Yes	Yes
Boost mode 5-10ma	Yes	0.5-8mA	0.3-8mA	0.3-8mA	not mentioned
Radiography : 20ma or more	Yes	Yes	Yes	Yes	0.2 ~ 12.2mA
Radiography mA x t: 4-100mAs	Yes	Yes	Yes	Yes	not mentioned
Automatic brightness control (ABC)	Yes	Yes	Yes	Yes	Yes
Collimator beam limiter	Yes	Yes	Yes	Yes	not mentioned
Wide movement of C-Arm up and down , left and right, forward and backward	Yes	Yes	Yes	Yes	Yes
Sliding rotation, C-Arm rotation : +/- 180°	Yes	Yes	Yes	Yes	not mentioned
Image Intensifier triple field	Yes	Yes	Yes	Yes	not mentioned

Prakash

Handwritten signatures and initials in blue and green ink.

Input field : 9"/6"/4.5"	Yes	Yes	Yes	Yes	Yes	9" & 6"
Resolution (centre): 55LPM/cm	Yes	Yes	Yes	Yes	Yes	not mentioned
Camera : Type Digital CCD Camera	Yes	Yes	Yes	Yes	Yes	CCD not mentioned
Auto Gain control, Brightness	Yes	Yes	Yes	Yes	Yes	not mentioned
TV Monitor : Black 2 x 17" Monitors on trolley	2 x 19"	Yes	Yes	Yes	Yes	Yes
Image memory: 100 frame	Yes	Yes	Yes	Yes	Yes	Yes
CE or FDA Approved	CE	CE	CE	CE	CE	FDA
Country of Origin : USA /EEC/ JAPAN	France	Japan + Italy	Japan + France	France	France	Germany
Delivery Period:	As per TE	As per TE	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes	Yes	Yes



TECHNICAL EVALUATION REPORT


For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 11 FLUOROSCOPY WITH DR SYSTEM 500MA WITH ALL STANDARD ACCESSORIES

M/S. ONTECH CORPORATION:

500mA Digital Radiography Machine

- M/s. Ontech Corp. has quoted model Radnext 50 from M/s. Hitachi Corporation.
- Exposure Time: The quoted model Radnext 50 has exposure time of 1ms to 4 sec instead of required exposure time 1ms to 10s.
- Four way table top: From the attached brochure and technical quotation, the required parameters for four way floating top fix table are not available for comparison from tender specification.
- FPD Pixel Resolution : Pixel resolution for the flat panel detector is 2208x2688 (5.9Million) instead of required tender specification 2500x3000(7.5Million).
- FPD Pixel pitch: The quoted flat panel has pixel pitch of 160µm instead of required tender specification 140µm or less.
- C-Arm Fluoroscopy: The quoted model of C-Arm Fluoroscopy has inverter frequency 40KHZ instead of 50KHZ or better.
- The fluoro mA of quoted model in the boost mode is 0.5 to 8mA instead of 5-10mA.
- Due to above discrepancies from the required tender specification, the offer of M/s. Ontech Corporation is declared as non-responsive.


Dr. Sikandar Rafique Qurnishi
MBBS MChS
Assistant Professor,
Radiology
SMBS Medical College Lyari.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 11 FLUOROSCOPY WITH DR SYSTEM 500MA WITH ALL STANDARD ACCESSORIES

M/S. MEDEQUIPS:

500mA Digital Radiography Machine


1st offer:

- M/s. Medequips has quoted Mred -A50S from M/s. Toshiba-Japan.
- Exposure time has minor deficiency i.e. quoted model has 1ms -9sec instead of required 1ms - 10sec.
- The quoted model of Four way floating top fix table height is 68cm instead of required tender specification 80-90cm.
- The quoted model of table top weight capacity is 150kg with an optional attachment of 200kg instead of required 250kg.
- From the available brochure/technical data sheet, the quoted model of floor mounted tube stand does not show the required column rotation of +/-90° and tube arm rotation horizontal axis +/-180°.
- From the attached brochure of collimator, the required specification is not verifiable i.e. six pair of leaf and LED light indicator.
- The quoted Flat panel Detector type is TFT-GOS(Gadolinium Oxysulfide) instead of required Amorphous Silicon
- Due to above discrepancies from the required tender specification, the offer of M/s. Medequips is declared as non-responsive.

Alternate Offer:

- M/s. Mediequip has quoted model Riviera 500mA Digital Radiography Machine from Primax-France. It has been observed from the website of the Primax that they are not the Manufacturer rather they outsource the equipment from Europe, China etc.
- The table top weight capacity is 180kg with no movement limitation and 250kg for static only whereas the tender required 250kg weight with full movement etc.
- M/s. Mediequip has quoted Flat Panel Detector Flying Pix which is not a known brand.
- The quoted model has pixel resolution of 2775 x 2874 instead of 2500x3000.
- The quoted model has a pixel pitch of 148µm instead of 140µm or less.
- Due to above discrepancies from the required tender specification, the offer of M/s. Medequips is declared as non-responsive.

Page 2 of 4


Dr. Sikandar Rafique Qureshi
MBBS MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.

Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Diabetic Center



TECHNICAL EVALUTION REPORT


For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi


Item No. 11 FLUOROSCOPY WITH DR SYSTEM 500MA WITH ALL STANDARD ACCESSORIES

M/s. BIOS

500mA Digital Radiography Machine

- M/s. BIOS has quoted model: N-40HF from M/s. Stephanix-France.
- Quoted model of M/s. BIOS is complete as per tender specification.
- M/s. BIOS has quoted Csi type Flat Panel Detector with the pixel resolution of 3556 x 4320.
- The pixel pitch of quoted FPD is 100um.
- M/s. BIOS has quoted C-Arm Fluoroscopy , model : OMNISCOP from Stephanix-France.
- The quoted C-Arm frequency is 50KHZ.
- The offer of M/s. BIOS is complete as per tender specification and hence may be declare as responsive.


Dr. Sikandar Rafique Qurnishi
MBBS, MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 11 FLUOROSCOPY WITH DR SYSTEM 500MA WITH ALL STANDARD ACCESSORIES

M/s. Simens Pkaistan

500mA Digital Radiography Machine

- M/s. Simens has quoted model Multix fusion and Siremobile compact L.
- M/s. Simens has quoted ceiling mounted tube stand instead of floor to ceiling tube stand. Ceiling mounted tube stand needs more space for installation and etc.
- The quoted model of M/s. Simens has exposure time 1ms with iontomat, which is a lon chamber instead of 1ms -10s.
- The quoted model of C-Arm Fluoroscopy is 15-30KHZ instead of 50KHz required in the tender.
- The quoted model of C-Arm Fluoroscopy has output of 1.4kW instead of 3.5kW required in the tender.
- The quoted model of C-Arm fluoroscopy has single focus 0.6mm instead of required dual focus i.e. 0.5 small and 1.5large.
- From the attached brochure of Siremobile, the required parameter such as Radiography mAs, automatic brightness control (ABC), C-Arm rotation , image intensifier input field are not verifiable.
- Due to above discrepancies from the required tender specification, the offer of M/s. Simens Pakistan is declared as non-responsive.


Dr. Sikandar Rafique Quraishi
MBBS MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.


Prof. Jffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

Tender Enquiry No.
 Item No.
 Name of Item.
 Quantity Required
 Estimated Price
 Country of Origin

TECHNICAL DATA
 Machinery Equipment Under ADP Scheme NO 787 / 2014-2015
 12
 ABG Machine System with all Standard Accessories
 O1 No.
 Rs. 685,000.00
 USA / ECC / JAPAN

Name of Firms	M/s. Life One	M/s. Hospicare	M/s. Karachi
Make	Medica	Techno Medical	Cobas
Model	Easy Blood Gas	Gastat - 600	8121
Type	Parameter	Yes	Yes
pH	Yes	Yes	Yes
Blood Gas	pCO ₂ , pO ₂	Yes	Yes
Electrolytes	K ⁺ , Na ⁺ , Cl	not mentioned	Yes
Hematocrit	Hct	Yes	Yes
Shelve life of solution	3-4 months	not mentioned	Yes
Pouch volume	-440 ml	not mentioned	not mentioned
Fluidic cycles	230	not mentioned	not mentioned
Storage temperature	12-12 °C/54-77 °F	Yes	+15 ~ +33
Automatic: 1 & 2 point calibration every 8 hours	Time period not mentioned	Time period not mentioned	Yes
Full visual graphic array (VGA)	Yes	Yes	Yes
Full active thin film transistor (TFT)	Yes	Yes	Yes
800X600 resolutions	Yes	Yes	Yes
Resistive touch screen	Yes	Yes	Yes
Country of Origin : USA / ECC / JAPAN	USA	Japan	Switzerland
Delivery Period:	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes

Handwritten signature

Handwritten signature

Handwritten initials

Large handwritten signature

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital Karachi

Item No. 12 ABG MACHINE SYSTEM

M/s. Life One

M/s. Life One has quoted model: Easy Blood Gas from Medica-USA

Quoted model of M/s. Life One meets tender specification.

Offer of M/s. Life One is declared as responsive.

M/s. Hospicare

M/s. Hospicare has quoted model: Gastat 600 from Techno medical – Japan

Quoted model of M/s. Hospicare meets tender specification.

Offer of M/s. Hospicare is declared as responsive.

M/s. Karachi Medical Company

M/s. Karachi Medical Company has quoted model: B-121 Cobas, from Roche – Switzerland

Quoted model of M/s. Karachi Medical Company meets tender specification.

Offer of M/s. Karachi Medical Company is declared as responsive.

Prof. Iffat Yazdani
MBBS, FRCP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price
Country of Origin

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015
13
Syringe Pump with all Standard Accessories
10 Nos.
Rs. 100,000.00
Korea / Taiwan / Turkey

Name of Firms	M/s. S. Ejazuddin & Co.	M/s. Life One Daiwah	M/s. Hospicare B. Braun	M/s. Elate cc Medina	M/s. Karachi Medical Company MDX Instruments	M/s. Universal Ent. Terumo
Automatic Syringe- Size Detection	No	Yes	Yes	Yes	Yes	Yes
K.V.O. (Keep Veine open Function)	No	Yes	Not mentioned	Yes	not mentioned	not mentioned
Flow Rate Range						
Syringe 10 ml : 0.1-300 ml /h	~200	Yes	1-200 ml/h	Yes	Yes	300 ml/h
Syringe 20ml : 0.1- 400 ml /h	~350	Yes	1-700 ml/h	Yes	Yes	300 ml/h
Syringe 30ml : 0.1- 500 ml /h	~450	Yes	1-800 ml/h	Yes	Yes	300 ml/h
Syringe 50ml : 0.1- 1200 ml /h	~800	Yes	1-1200 ml/h	Yes	Yes	Yes
Volume limit : 0.1- 999.9 ml (0.1 ml steps)	Yes	Yes	Yes	Yes	Yes	Yes
Purge Rate:						
Syringe 10ml: 300ml/h	~200	Yes	1-200 ml/h	Yes	Yes	300 ml/h
Syringe 20ml: 400ml/h	~350	Yes	1-700 ml/h	Yes	Yes	400 ml/h
Syringe 30ml: 500ml/h	~450	Yes	Yes	Yes	Yes	500 ml/h
Syringe 50ml: 1200ml/h	~800	Yes	1-1200 ml/h	Yes	Yes	Yes
KVO : 0.1 ml/h	No	Yes	Not mentioned	Yes	not mentioned	not mentioned
Accuracy With Approved Syringes	Yes	Yes	Yes	Yes	Yes	Yes
Mechanical accuracy : ±1%	Yes	Yes	Yes	Yes	not mentioned	Yes
Accuracy including Syringe: ±3%	Yes	Yes	Yes	Yes	not mentioned	Yes
ALARM:						
Occlusion, Empty syringe, Low Battery, Syringe Dislocated, Plunger / Clutch Disengaged, Repeat Alarm, Start Reminder Alarm	Yes	Yes	Yes		Yes	Yes
Country of Origin : Korea / Taiwan / Turkey	Japan	Korea	Germany	Poland	USA	Japan
Delivery Period:	As per TE	As per TE	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes	Yes	Yes

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 13 SYRINGE PUMP

M/s. S.Ejaz udin

M/s. S.Ejazuddin has quoted model: SP-500 from JMS-Japan

Quoted model of M/s. S.Ejazuddin meets tender specification.

Offer of M/s. S.Ejazuddin is declared as responsive.

M/s. Life One

M/s. Life One has quoted model: MS-1200 from Daiwha-Korea

Quoted model of M/s. Life One meets tender specification.

Offer of M/s. Life One is declared as responsive.

M/s. Hospicare

M/s. Hospicare has quoted model: Perfusor compact S from B.Braun Germany

Quoted model of M/s. Hospicare meets tender specification.


Offer of M/s. Hospicare is declared as responsive.

M/s. Elate CC

M/s. Elate Cc has quoted model: Medima S from Medima-Poland

Quoted model of M/s. Elate Cc meets tender specification.

Offer of M/s. Elate Cc is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 13 SYRINGE PUMP

M/s. Karachi Medical Company

M/s. Karachi Medical Company has quoted model: CF 111 from MDS-China

From enclosed quotation country of origin and Port of shipment is China which is against tender specification.

Due to above reason, offer of M/s. Karachi Medical is declared as non- responsive.

M/s. Universal Enterprise

M/s. Universal Enterprise has quoted model: SS-700 from Terumo-Japan

Quoted model of M/s. Universal Enterprise meets tender specification.

Offer of M/s. Universal Enterprise is declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

14

Name of Item.

Echo Color Doppler with all Standard Accessories

Quantity Required

01 No.

Estimated Price

Rs. 2,000,000.00

Country of Origin

USA / EEC / JAPAN

Name of Firms	M/s. BIOS	M/s. Ontech Corp	M/s. Medequips SMC	M/s. Karachi Medical Company	M/s. Shrazil Trading
Make	Sonoscap	Hitachi	Toshiba	MDX Instruments	GE HealthCare
Model	S11	Aloka F31	Xario 100	Q7	Logic F6
Echo Cardiography System for Vascular and Cardiology application	Yes	Yes	Yes	Yes	Yes
Original Trolley mounted Fully Digital multi beam former	Yes	Yes	Yes	Yes	Yes
Fast Image processor and three active connectors	Yes	Yes	Yes	Yes	Yes
Gray shades 256 gray shades or better	Yes	Yes	Yes	Yes	Yes
Image depth: Image depth: 25 cm or more	Yes	5 - 30cm	40cm	depth not mentioned	33cm
Monitor: 15" or more high resolution, Non Flicker TFT/LCD/LED	Yes	Yes	19"	Yes	17"
Imaging modes: B mode, 2B/4B, M mode, TH1 Mode, Color Mode, DPl Mode, PW Mode, PW Mode, Trapezoidal Imaging, compound imaging, etc	Yes	Yes	TH1 not mentioned	Trapezoidal Imaging & TH1 not mentioned	Trapezoidal Imaging & TH1 not mentioned
Measurement Packages:					
Obstetric Measurement Package	Yes	Yes	Yes	Yes	Yes
Gynecology Measurement Package	Yes	Yes	Yes	Yes	Yes
Cardiology Measurement Package	Yes	Yes	Yes	Yes	Yes
Urology Measurement Package	Yes	Yes	Yes	Yes	Yes
Small parts Measurement Package	Yes	Yes	Yes	Yes	Yes
Vascular Measurement Package	Yes	Yes	Yes	Yes	Yes
Orthopedic Measurement Package	Yes	Yes	Yes	Yes	Yes
High pulse repetition frequency	Yes	Yes	Yes	Yes	Yes
Spectral Doppler	Yes	Yes	Yes	Yes	Yes
DICOM 3.0	Yes	Yes	Yes	Yes	Yes
Complete with the following:					
Conver probe Multi frequency between 2-6 MHz or better	Yes	2.5 - 5.71	Yes	frequency not mentioned	frequency not mentioned
Linear probe Multi frequency between 5-10 MHz or better	Yes	5 - 13.3	Yes	frequency not mentioned	frequency not mentioned

Sector Probe Multi frequency 2-5 MHz or better	Yes	1.82 ~ 3.64	Yes	frequency not mentioned	frequency not mentioned
Black & White digital Ultrasound printer	Yes	Yes	Yes	Yes	Yes
Sine wave on line UPS	Yes	Yes	Yes	Yes	Yes
Country of origin should be USA/ Western Europe/Japan with their respective standards MDD (CE) compliance / MH/W Japan / FDA 510k. Equipment from other countries which is FDA 510k approved can be considered.	FDA 510k China	Yes	Japan	Port of Shipping China	FDA 510k China
Country of Origin : USA / EEC / JAPAN	China	Japan	Japan	USA	China
Delivery Period:	As per TE	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes	Yes

Shankar

So

Yes

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 14 ECHO COLOR DOPPLER WITH ALL STANDARD ACCESSORIES

M/S. ONTECH CORPORATION:

- M/s. Ontech Corp. has quoted model Aloka-F31 from M/s. Hitachi Corporation.
- The quoted model has 30cm image depth.
- Offer model have all imaging modes as per tender requirement.
- Offer is complete with Multi frequency Convex, Linear and Sector Probes.
- M/s. Ontech has offered with Echo Color Doppler with B/W printer and Online UPS.
- The offer of M/s. Ontech is complete as per tender specification and hence may be declared as responsive.

M/S. MEDEQUIPS:

- M/s. Medequips has quoted model Xario 100 from M/s. Toshiba.
- The quoted model has 40cm image depth.
- Quoted model of M/s. Medequips has 19" display.
- Offered model meets required imaging modes .
- Offer is complete with Multi frequency Convex, Linear and Sector Probes.
- M/s. Medequips has offered with Echo Color Doppler with B/W printer and Online UPS.
- The offer of M/s. Mediequip meets tender specification hence may be declared as responsive.

M/s. BIOS

- M/s. BIOS has quoted model S-11 from M/s. Sonoscap.
- The image depth of quoted model is as per tender specification.
- Display of quoted model is as per tender specification.
- Offered model have all imaging modes as per tender specification.
- Offer is complete with Multi frequency Convex, Linear and Sector Probes.
- M/s. BIOS has offered with Echo Color Doppler with B/W printer and Online UPS.
- Offer model do have FDA 510K certification.
- The offer of M/s. BIOS meets tender specification and hence may be declared as responsive.

Page 1 of 2


Dr. Sikandar Rafique Quraishi
MBBS MCPS
Assistant Professor,
Nephrology
SMBB Medical College Lyari.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 14 ECHO COLOR DOPPLER WITH ALL STANDARD ACCESSORIES


M/s. Karachi Medical Company

- M/s. Karachi Medical Company has quoted model Q7 from M/s. MDX Instruments.
- The image depth of quoted model is neither mentioned in technical brochure nor in their technical offer.
- The offered model of Karachi Medical Company do not have THI mode and Trapezoidal image mode as required in tender specification.
- Frequency ranges of offered probes are not mentioned in their offer.
- As per their technical literature, the offered model is of Chinese origin and do not have FDA 510K certification.
- Due to above discrepancies from the required tender specification, the offer of M/s. Karachi Medical Company is declared as non-responsive.

M/s. Shirazi Trading

- M/s. Shirazi Trading has quoted model Logic F-6 from M/s. G.E Healthcare.
- The quoted model has 33cm image depth.
- Quoted model of M/s. Shirazi Trading has 17" display.
- Offered model meets requirement of imaging modes as per tender specification.
- Offer is complete with Multi frequency Convex, Linear and Sector Probes.
- M/s. Shirazi Trading has offered with Echo Color Doppler with B/W printer and Online UPS.
- The offer of M/s. Shirazi Trading meets tender specification hence may be declared as responsive.


Dr. Sibandar Rafique Quraishi
MBBS MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

15

Name of Item.

Patient Monitor with all Standard Accessories

Quantity Required

11 Nos.

Estimated Price

Rs. 150,000.00

Country of Origin

USA / EEC / JAPAN

Name of Firms	M/s. Life One	M/s. Medequips SMC	M/s. Karachi Medical Company	M/s. Shirazi Trading	M/s. Bio Track Trading
Make	General Meditech	Nihonkohdens	MDX Instruments	GE HealthCare	Shenzhen- N-Tech
Model	GLL	SVM-7521	Accut Sign8	B-40	9000F
ECG, NIBP, SpO2, RESP, TEMP	Yes	Yes	Resp & Temp Optional	Temp not mentioned	Yes
Standard 5 lead ECG with ST segment monitoring, respiration, dual temperature monitoring	Yes	Yes	respiration, dual temperature monitoring optional	dual temperature monitoring not mentioned	Single Temp.
SpO2, non-invasive blood pressure, Auto setting of alarm limits	Yes	Yes	Yes	Yes	Yes
Suitable for adults and pediatric	Yes	Yes	Yes	Yes	Yes
Battery backup for 90 minutes atleast	Yes	Yes	not mentioned	Yes	Yes
12" or more Touch Screen display	15"	Yes	Yes	Yes	Yes
Trace 8 Wave forms	Yes	6 waves	Yes	6 waves	6 waves
Audible And Visual Alarms	Yes	Yes	Yes	Yes	Yes
Country of origin should be UK/USA/ EEC/Japan with their respective standards MDD (CE) compliance / MILW Japan / FDA 510K. In case equipment is from other country FDA 510K approval is mandatory	China FDA 510K	Japan	CE	China FDA 510K	CE
Country of Origin : USA / EEC / JAPAN	China	Japan	USA	China	China
Delivery Period:	As per TE	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes	Yes

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 15 PATIENT MONITOR

M/S. Life One

- M/s. Life One has quoted model G3L from M/s. General Meditech-China.
- Offered model is FDA 510K approved.
- The offer of M/s. Life One is complete as per tender specification and hence may be declared as responsive.

M/S. MEDEQUIPS:


- M/s. Medequips has quoted model SVM-7521 from M/s. Nihon Kohdan-Japan
- Offered model meets tender specification.
- The offer of M/s. Medequips is complete as per tender specification and hence may be declared as responsive.

M/s. Karachi Medical Company

- M/s. Karachi Medical Company has quoted model Accuit Sign 8 from M/s. MDX Instruments.
- As per their technical literature, the offered model is of Chinese origin and do not have FDA 510K certification.
- Due to above discrepancies from the required tender specification, the offer of M/s. Karachi Medical Company is declared as non-responsive.

M/s. Shirazi Trading

- M/s. Shirazi Trading has quoted model B-40 from M/s. G.E Healthcare-China.
- Offered model is FDA 510K approved.
- The offer of M/s. Shirazi Trading meets tender specification hence may be declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL EVALUATION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 15 PATIENT MONITOR

M/s. Biotrac Trading

- M/s. Biotrach Trading has quoted model 9000 F from M/s. Shenzhen – N- Tech-China
- Offered model do not have FDA 510K Approval.
- With the above reason, offer of M/s. Biotrac Trading is declared as non-responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Tender Enquiry No.

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

Item No.

16

Name of Item.

Defibrillator with all Standard Accessories

Quantity Required

01 No.

Estimated Price

Rs. 475,000.00

Country of Origin

USA / EEC / JAPAN

Name of Firms	M/s. BIOS	M/s. Medequips SMC	M/s Elate CC	M/s. Karachi Medical Company	M/s. Shirazi Trading
Make	Osatu	Nilhonkohdens	Primedic	MOX Instruments	Zoil Medica
Model	R700	TEC-5521K	X01	Hartstart Pro	M Series
Defibrillator with Biphasic technology. Truncated Exponential waveform. Integrated pediatric paddles	Yes	Yes	Yes	Yes	Yes
Monitor:					
EKG monitoring : 3,5, or 10 lead patient cable, paddles or multifunction disposable electrodes. Heart rate : 0 to 300 ppm.	Yes	Yes	30 to 300 ppm	Yes	Yes
Defibrillator					
Biphasic truncated exponential	Yes	Truncated Exponential Not Mentioned	not mentioned	Truncated Exponential Not Mentioned	Yes
Manual : 1-2-3-5-7-9-10-15-20-30-50-70-100-125-150-200J	200J	270 J	360J	270 J	360J
Synchronized cardioversion	Yes	Yes	not mentioned	not mentioned	Yes
Defibrillator energy : 200 joules	Yes	270 J	360J	270 J	360J
Charging time: 6 sec at 200 joules	6 sec at 200 J	3 sec at 200 J	Yes	not mentioned	< 7 sec
Recorder:					
Thermal array 50mm	Yes	Yes	Yes	Yes	Yes
10, 25, and 50mm/sec speed	Yes	Yes	Yes	Yes	Yes
Manual and automatic	Yes	Yes	Yes	Yes	Yes
Complete with one hour or more battery back	Yes	Yes	Yes	Yes	Yes
Country of Origin : USA / EEC / JAPAN	Spain	Japan	Germany	USA	USA
Delivery Period:	As per TE	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes	Yes

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15 through Civil Hospital, Karachi

Item No. 16 DEFIBRILLATOR

M/s. BIOS

- M/s. BIOS has quoted model R-700 from M/s. Osatu-Spain.
- Offered model meets tender specification.
- The offer of M/s. BIOS is complete as per tender specification and hence may be declared as responsive.

M/s. Elate Cc


- M/s. Elate Cc has quoted model XD -1 from M/s. Primedic-Germany
- Offered model meets tender specification.
- The offer of M/s. Elate Cc is complete as per tender specification and hence may be declared as responsive.

M/s. Medequips

- M/s. Medquips has quoted model Tech 5521 K from M/s. Nihon Kohdan.
- Offered model meets tender specification.
- The offer of M/s. Medequips is complete as per tender specification and hence may be declared as responsive.

M/s. Karachi Medical Company

- M/s. Karachi Medical Company has quoted model Heart Start Pro from.
- Offered model do not meet tender specification as the Synchronized cardioversion is not available in quoted model.
- From the attached brochure, charging time is not verifiable hence offer of M/s. Karachi Medical Company is declared as non-responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre


TECHNICAL EVALUATION REPORT

for purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 16 DEFIBRILLATOR

M/s. Shirazi Trading

- M/s. Shirazi Trading has quoted model M series from M/s. Zoll Medical-USA
- Offered model meets tender specification.
- The offer of M/s. Shirazi Trading is complete as per tender specification and hence may be declared as responsive.


Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

TECHNICAL DATA

Machinery Equipment Under ADP Scheme NO 787 / 2014-2015

17

Diathermy with all Standard Accessories

01 No.

Rs. 375,000.00

Korea / Taiwan / Turkey

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price
Country of Origin


Name of Firms	M/s. Life One	M/s. Hospicare	M/s. Karachi Medical Company	M/s. Universal Ent.
Make	Dainiha	Vally-Lab	Beful	Tekno
Model	DT-400P	Force FX	Doctanz 400	90010-04
Electro Surgical Diathermy machine 400W to be used in area of General Surgery, Gynecology, Endoscopy, Vascular Surgical, Neuro surgery, orthopedic, urology, TURP procedure and etc	Yes	300W, TURP not mentioned	Yes	Yes
Unit should be capable having max 400W at 500ohms with a continuous wave in the pure cut mode	Yes	300W	400W at 300ohms	400W at 340ohms
Cut mode should have four different mode: pure, blend I,IIa, III with respective output power of 230W at 300ohms, 180 W at 300ohms, 120 W at 300ohms	Yes	Yes	Yes	Yes
Two modes of monopolar coagulation	Yes	Yes	Yes	Yes
Contact coagulation 100W max power and spray coagulation 80W max power	Yes	Yes	Yes	Yes
Spray coagulation to be used effectively in TURP and endoscopic procedure in urology	Yes	TURP not mentioned	Yes	Yes
The contact coagulation to be used for laparoscope. Two modes of bipolar coagulation	Yes	Yes	Yes	Yes
Standard 99W and soft 50W for large tissue coagulation and small tissue coagulation	Yes	Yes	soft 50W not mentioned	Yes
Unit must have a memory function to remember its last setting in case of power failure	not mentioned	Yes	Yes	not mentioned
The system should have ability to check safety system and automatically return to the last activated setting	Yes	Yes	Yes	Yes
For patient safety, unit should have patient return electrode monitoring system	Yes	Yes	Yes	Yes
Unit should complete with: bipolar forceps	Yes	Yes	Yes	Yes
Reusable bipolar forceps cable	Yes	Yes	Yes	Yes
Reusable mono polar handle with cable	Yes	Yes	Yes	Yes
Reusable twin button handle with cable	Yes	Yes	Yes	Yes
Electrodes double foot switch	Yes	Yes	Yes	Yes
Single foot switch	Yes	Yes	Yes	Yes
Disposable dual return plate	Yes	Yes	Yes	Yes


Reusable return plate cable	Yes	Yes	Yes	Yes
Reusable metal plate with cable	Yes	Yes	Yes	Yes
Country of Origin : Korea / Taiwan / Turkey	Korea	USA	Korea	Germany
Delivery Period:	As per TE	As per TE	As per TE	As per TE
Warranty:	As per TE	As per TE	As per TE	As per TE
Catalogue / Brochure should be attached duly flagged.	Yes	Yes	Yes	Yes


WARIS KHAILI
 Divisional Accounts Officer
 Mr. Wazir Khalil
 Accident & Disaster Accounts Office, Complex At
 Office of Accounts, Government of Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


 Deputy Director (Dev.)
 Health Department
 Government of Sindh
 KARACHI


 Representative of E.S. Department
 Government of Sindh
 Deputy Registrar
 Motor Registration
 Karachi


 Asst. Prof. Dr. Sikandar Rafique Qurashi
 Shahheed Mohtarma Benazir Bhutto Medical College, Karachi


 (Member)
Dr. Sikandar Rafique Qurashi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

TECHNICAL EVALUTION REPORT

For purchase of Medical Equipment for Welfare Society for Patient Care Under ADP Scheme no. 787/2014-15
through Civil Hospital, Karachi

Item No. 17 DIATHERMY

M/s. Life One

M/s. Life One has quoted model: DT-400P C from M/s. Daiwha-Korea.

Quoted model of M/s. Life One meets tender specification.

Offer of M/s. Life One is declared as responsive.

M/s. Universal Enterprise

M/s. Universal Enterprise has quoted model: 90010-04 from Tekno.

Quoted model of M/s. Universal Enterprise meets tender specification.

Offer of M/s. Universal Enterprise is declared as responsive.

M/s. Karachi Medical Company

M/s. Karachi Medical Company has quoted model: Beful from Doctanz 400.

Quoted model of M/s. Karachi Medical Company meets tender specification.

Offer of M/s. Karachi Medical Company is declared as responsive.

M/s. Hospicare

M/s. Hospicare has quoted model: Force-FX from Valley Lab .

Quoted model of M/s. Hospicare meets tender specification.

Offer of M/s.Hospicare is declared as responsive.



Prof. Iffat Yazdani
MD, FACP
Consultant Nephrologist
Clifton Dialysis Centre

MINUTES


Date:-04-06-2015

A Hospital Procurement Committee meeting held on 04-06-2015 at 03.00 p.m. in the Board Room under the Chairmanship of Dr. Sohail Mir Memon Additional Medical Superintendent (Procurement) Civil Hospital, Karachi to open the financial bids after finalization of technical bids by the Procurement Committee for procurement of **Machinery / Equipment for WELFARE SOCIETY FOR PATIENT CARE UNDER ADP SCHEME # 787/2014-15 THROUGH CIVIL HOSPITAL, Karachi**. The technical bids was opened on 25-05-2015 at 12.00 Noon and subsequently recommended by the Procurement Committee for procurement of the required items.

The following members attended the meeting:-

1	Dr. Sohail Mir Memon Addl. Medical Superintendent (Procurement) Civil Hospital Karachi	Chairman
2	Asstt. Prof. Dr. Sikandar Rafique Quraishi Shaheed Mohtarma Benazir Bhutto Medical College, Karachi	Member
3	Representative of Finance Department Government of Sindh, Karachi	Member
4	Representative of Health Department (Dev. Wing) Government of Sindh, Karachi	Member
5	Mr. Waris Khalil District Accounts Officer Office of Accountant General Sindh, Karachi	Member

Initiating the discussions, this office presented the evaluated Technical Data Sheets for perusal of the Hospital Procurement Committee.


Dr. Sohail Mir Memon
Addl. Medical Superintendent
(Procurement)
Civil Hospital, Karachi


Deputy Director (Devt)
Health Department
Government of Sindh
KARACHI


Waheed Shaikh
Deputy Director (Admn)
Motor Registration Wing
Karachi


WARIS KHALIL
Divisional Accounts Officer
Accident & Emergency Centre &
Ancillary Services Complex At
Civil Hospital Karachi


Dr. Sikandar Rafique Quraishi
MBBS MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.

Sealed tenders were invited through the publication of Tender Notice in prominent News Papers i.e. The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 and other News Papers and hoisted on SPPRA website @ Sr. No.24148 for the **Supply of Medical Equipment for Welfare Society for Patient Care under ADP Scheme # 787/2014-15, thought Civil Hospital Karachi.**

The **Financial Bids** opened on 04-06-2015 at 03.00 pm. in the Committee Room, Civil Hospital, Karachi before the Procurement Committee.

The following firms have participated in the competition. The representatives who were present at the time of opening of Financial Bids their signatures were taken as a token of presence.

S. No.	Name of Firms	Signature of Representative
1	M/s. S. Ejazuddin & Co.	 0300-3782957
2	M/s. Hospital Supply Corporation	 0300-2069330
3	M/s. Life One	 0545-5235017
4	M/s. BIOS	 03333831988
5	M/s. Noor International	 0331-2153212
6	M/s. HospiCare System	 03000550116
7	M/s. ZAM Trading	 0341-3105990
8	M/s. Technology Link (Pvt.) Ltd.	 03332750819
9	M/s. Ontech Corporation	 0345-3590020
10	M/s. Siemens Pakistan	
11	M/s. Medequips SMC	 0333-3899842
12	M/s. Elate CC (Pvt.) Ltd.	 0311-0209044
13	M/s. Karachi Medical Co.	 03333016546 M. JAHANGIR
14	M/s. Universal Enterprises	 0333-2264534
15	M/s. Sherazi Trading	 0312-2038616
16	M/s. Biotrack Trading	 0546-2787775
17	M/s. Oasis Medical System	 0322-296847






BID EVALUATION REPORT

1. Name of Procuring Agency: **Civil Hospital, Karachi**
 2. Tender Reference No: **MS/CHK/ADP/787/2014-15/- 3163**
 3. Tender Description/Name of item: **Hemodialysis Machine with all Standard Accessories**
 4. Method of Procurement: **Single Stage Two Envelope Procedure Basis**
 5. Tender Published: **The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148,**

Print & Electronic Media (SPPRA ID No. & News papers names with dates)

6. Total Bid documents Sold: **Seventeen**
 7. Total Bids Received: **Two**
 8. Technical Bid Opening date: (if applicable) **25-05-2015** (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): **One**
 10. Bid(s) Rejected: **Nil**
 11. Financial Bid Opening date: **04.06.2015**
 12. Bid Evaluation Report: **copy of minutes attached.**


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Universal Enterprise	Rs. 1,163,000/- each	1 st Lowest	Over all under PC-I	Lowest Basis	The Committee is recommended to award contract M/s Universal Enterprises @ Cost of Pak Rs. 1,163,000/- each on D.D.P basis

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Accounts Officer
 Divisional Accounts Office Centre &
 Office of Accountant General Sindh, Karachi
 Civil (Member) Karachi


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)

**Deputy Director (Dev.)
 Health Department
 Government of Sindh
 KARACHI**


 Representative of E. S. T. Department
Waheed Shaikh
 (Deputy Director (Admn))
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)

Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBS Medical College Lyari.


Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT

1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of item: Reprocessing Unit with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148,

Print & Electronic Media (SPPRA ID No. & News papers names with dates)


6. Total Bid documents Sold: Seventeen
 7. Total Bids Received: Two
 8. Technical Bid Opening date: (If applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Universal Enterprise	Rs. 1,188,000/- each	1st	Over all under PC-I	Lowest Basis	The Committee is recommended to award contract M/s Universal Enterprises @ Cost of Pak Rs. 1,188,000/- each on D.D.P basis

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 D/o Mr. Waris Khalil, District Accounts Officer, Centre & Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


 Representative of F&T Department
Shaheed Shaikh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT

1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of item: OT Light with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148,

Print & Electronic Media (SPPRA ID No. & News papers names with dates)


6. Total Bid documents Sold; Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Life One	Rs. 605,000/- each	1 st Lowest	Over all under PC-1	Lowest Basis	The Committee is recommended to award contract M/s Life One @ Cost of Pak Rs.605,000/- each on D.D.P Lowest basis.
2	M/s. Technology Links	Rs. 799,000/- each	3 rd Lowest			
3	M/s. Universal Enterprise	Rs. 683,000/- each	2 nd Lowest			

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Accounts Officer
 Accident & Emergency Unit &
 Ancillary Services Unit, Civil Hospital
 Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Dev)
 Health Department
 Government of Sindh
 KARACHI


 Representative of E.P.T. Department
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)
Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT

1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of item: Surgical Beds (Dialysis Chair) with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
 6. Total Bid documents Sold; Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S#	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Hospital Supply Corporation	Rs. 450,000/- each	3 rd Lowest			
2	M/s. ZAM Trading Co.	Rs. 164,800/- each	1 st Lowest	Over All under PC-I	On Lowest Basis	The Committee is recommended to award contract M/s ZAM @ Cost of Pak Rs. 164,800/- each on D.D.P Lowest basis.
3	M/s. Technology Link (Pvt.) Ltd.	Rs. 395,000/- each	2 nd Lowest			


Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Accounts Officer
 Accounts & Finance Complex At
 District Accounts Office
 Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


 Representative of E. S. T. Department
 Government of Sindh, Karachi
 (Member)
Waheed Shaikh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)


Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of Item: Surgical Beds (Dialysis Recliners) with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
 6. Total Bid documents Sold: Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Hospital Supply Corporation	Rs. 375,000/- each	2 nd Lowest			
2	M/s. ZAM Trading Co.	Rs. 165,000/- each	1 st Lowest	Over All Under PC-I	On Lowest basis	The Committee is recommended to award contract M/s ZAM @ Cost of Pak Rs. 165,000/- each on D.D.P Lowest basis
3	M/s. Technology Link (Pvt.) Ltd.	Rs. 389,000/- each	3 rd Lowest			


Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Div: Mr. Waris Khalil, Officer
 District Accounts Office Centre 3
 Office of Accountant General Sindh, Karachi
 Civil Hospital Karachi
 Member


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Dist)
 Health Department
 Government of Sindh
 KARACHI


 Representative of E.T. Department
 Government of Sindh, Karachi
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)
Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of Item: Crash Trolley with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
 6. Total Bid documents Sold: Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S#	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s.BIOS	Rs. 150,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s BIOS @ Cost of Pak Rs. 150,000/- each on D.D.P basis

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Mr. Waris Khalil's Officer
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)


 Representative of E. & T. Department
 Government of Sindh, Karachi
 (Member)
Waseed Shaikh
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi

Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT

1. Name of Procuring Agency: Civil Hospital, Karachi
2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
3. Tender Description/Name of Item: Stretcher with all Standard Accessories
4. Method of Procurement: Single Stage Two Envelope Procedure Basis
5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
6. Total Bid documents Sold: Seventeen
7. Total Bids Received: Four
8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
9. No. of Bid technically qualified (if applicable): One
10. Bid(s) Rejected: Nil
11. Financial Bid Opening date: 04.06.2015
12. Bid Evaluation Report: copy of minutes attached.

S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Hospital Supply Corporation	Rs. 325,000/- each	2 nd Lowest			
2	M/s. ZAM Trading Co.	Rs. Rs. 49,900/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s ZAM @ Cost of Pak Rs. 49,900/- each on D.D.P basis

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
Divisional Officer
District Accounts Officer
Office of Accountant General Sindh, Karachi
(Member)


Representative of Health Department (Dev. Wing)
Government of Sindh, Karachi
(Member)


Representative of Motor Registration Wing
Government of Sindh, Karachi
(Member)


Asstt. Prof. Dr. Sikandar Rafique Quraishi
Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
(Member)
Dr. Sikandar Rafique Quraishi
MBBS MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.


Dr. Sohail Mir Memon
Addl. Medical Superintendent (Procurement)
Civil Hospital Karachi
(Chairman)
Dr. Sohail Mir Memon
Addl. Medical Superintendent
(Procurement)
Civil Hospital, Karachi.

BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of item: Autoclave with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
 6. Total Bid documents Sold; Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.

S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Life One	Rs. 1,065,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s Life One @ Cost of Pak Rs. 1,065,000/- each on D.D.P basis

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Accounts Officer
 Mr. Waris Khalil
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
Deputy Director (Dev)
 Health Department
 Government of Sindh
 KARACHI


 Representative of E.S.T. Department
 Government of Sindh, Karachi
Wahed Shaikh
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
3. Tender Description/Name of item: Ventilator with all Standard Accessories
4. Method of Procurement: Single Stage Two Envelope Procedure Basis
5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
6. Total Bid documents Sold: Seventeen
7. Total Bids Received: Four
8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
9. No. of Bid technically qualified (if applicable): One
10. Bid(s) Rejected: Nil
11. Financial Bid Opening date: 04.06.2015
12. Bid Evaluation Report: copy of minutes attached.


S#	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. BIOS	Rs. 2,195,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s Life One @ Cost of Pak Rs. 2,195,000/- each on D.D.P basis
2	M/s. HospiCare System	Rs. 2,450,000/- each	2 nd Lowest			

Minutes in which reasons for acceptance / rejection have been recorded


Mr. Waris Khalil
District Accounts Officer
Office of Accountant General Sindh, Karachi
(Member)


Representative of Health Department (Dev. Wing)
Government of Sindh, Karachi
(Member)
Deputy Director (Dev)
Health Department
Government of Sindh
KARACHI


Representative of E.S.T. Department
Government of Sindh
Waneed Shaikh
(Member)
Deputy Director (Admin)
Motor Registration Wing
Karachi


Asstt. Prof. Dr. Sikandar Rafique Quraishi
Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
(Member)
Dr. Sikandar Rafique Quraishi
MBBS MCPS
Assistant Professor,
Radiology
SIMBB Medical College Lyari.


Dr. Sohail Mir Memon
Addl. Medical Superintendent (Procurement)
Civil Hospital Karachi
(Chairman)

Dr. Sohail Mir Memon
Addl. Medical Superintendent
(Procurement)
Civil Hospital, Karachi.


BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
3. Tender Description/Name of item: Fluoroscopy with DR System 500mA with all Standard Accessories
4. Method of Procurement: Single Stage Two Envelope Procedure Basis
5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
6. Total Bid documents Sold; Seventeen
7. Total Bids Received: Four
8. Technical Bid Opening date: (If applicable) 25-05-2015 (Provide details in separate form)
9. No. of Bid technically qualified (if applicable): One
10. Bid(s) Rejected: Nil
11. Financial Bid Opening date: 04.06.2015
12. Bid Evaluation Report: copy of minutes attached.


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. BIOS	Rs. 23,915,000/- each	1 st Lowest	Over All Under PC-1		The Committee is recommended to award contract M/s BIOS @ Cost of Pak Rs. 23,915,000/- each on D.D.P basis

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Officer
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


 Representative of E & T Department
 Government of Sindh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi
 (Member)


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: M5/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of item: ABG Machine System with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
 6. Total Bid documents Sold: Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Life one	Rs. 1,195,000/- each	2 nd Lowest			
2	M/s. HospiCare System	Rs. 985,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s. HospiCare System @ Cost of Pak Rs. 985,000/- each on D.D.P basis
3	M/s. Karachi Medical Co.	Rs. 2,750,000/- each	3 rd Lowest			

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)
 Ancillary Services Complex At
 Civil Hospital Karachi


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Dev.)
 Health Department
 Government of Sindh
 KARACHI


 Representative of E & T Department
 Government of Sindh, Karachi
 (Member)
Waqeem Shaikh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT

- | | |
|--|--|
| 1. Name of Procuring Agency: | <u>Civil Hospital, Karachi</u> |
| 2. Tender Reference No: | <u>MS/CHK/ADP/787/2014-15/- 3163</u> |
| 3. Tender Description/Name of Item: | <u>Syringe Pump with all Standard Accessories</u> |
| 4. Method of Procurement: | <u>Single Stage Two Envelope Procedure Basis</u> |
| 5. Tender Published: | <u>The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)</u> |
| 6. Total Bid documents Sold; | <u>Seventeen</u> |
| 7. Total Bids Received: | <u>Four</u> |
| 8. Technical Bid Opening date: (if applicable) | <u>25-05-2015</u> (Provide details in separate form) |
| 9. No. of Bid technically qualified (if applicable): | <u>One</u> |
| 10. Bid(s) Rejected: | <u>Nil</u> |
| 11. Financial Bid Opening date: | <u>04.06.2015</u> |
| 12. Bid Evaluation Report: | <u>copy of minutes attached.</u> |

S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Universal Enterprises	Rs. 213,400/- each	5 th Lowest			
2	M/s. S. Ejazuddin & Co.	Rs. 155,000/- each	4 th Lowest			
3	M/s. Life One	Rs. 91,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s Life One @ Cost of Pak Rs. 91,000/- each on D.D.P basis
4	M/s. HospiCare System	Rs. 145,000/- each	3 rd Lowest			
5	M/s. Elate CC (Pvt.) Ltd.	Rs. 98,000/- each	2 nd Lowest			

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Officer
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 Civil Member Karachi


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Dev.)
 Health Department
 Government of Sindh
KARACHI


 Representative of E.S. Department
 Government of Sindh, Karachi
Waheed Shaikh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi


Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.


BID EVALUATION REPORT


1. Name of Procuring Agency: Civil Hospital, Karachi
 2. Tender Reference No: MS/CHK/ADP/787/2014-15/- 3163
 3. Tender Description/Name of item: Echo Color Doppler with all Standard Accessories
 4. Method of Procurement: Single Stage Two Envelope Procedure Basis
 5. Tender Published: The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)
 6. Total Bid documents Sold: Seventeen
 7. Total Bids Received: Four
 8. Technical Bid Opening date: (if applicable) 25-05-2015 (Provide details in separate form)
 9. No. of Bid technically qualified (if applicable): One
 10. Bid(s) Rejected: Nil
 11. Financial Bid Opening date: 04.06.2015
 12. Bid Evaluation Report: copy of minutes attached.


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. BIOS	Rs. 1,899,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s BIOS @ Cost of Pak Rs. 1,899,000/- each on D.D.P basis
2	M/s. Ontech Corporation	Rs. 2,999,000/- each	3 rd Lowest			
3	M/s. Sherazi Trading	Rs. 3,350,000/- each	4 th Lowest			
4	M/s. Medequips SMC	Rs. 2,974,000/- each	2 nd Lowest			


Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)
 Civil Hospital Karachi


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Dev.)
 Health Department
 Government of Sindh
KARACHI


 Representative of EL & T Department
 Government of Sindh, Karachi
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT


- | | |
|--|--|
| 1. Name of Procuring Agency: | <u>Civil Hospital, Karachi</u> |
| 2. Tender Reference No: | <u>MS/CHK/ADP/787/2014-15/- 3163</u> |
| 3. Tender Description/Name of Item: | <u>Patient Monitor with all Standard Accessories</u> |
| 4. Method of Procurement: | <u>Single Stage Two Envelope Procedure Basis</u> |
| 5. Tender Published: | <u>The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)</u> |
| 6. Total Bid documents Sold: | <u>Seventeen</u> |
| 7. Total Bids Received: | <u>Four</u> |
| 8. Technical Bid Opening date: (if applicable) | <u>25-05-2015</u> (Provide details in separate form) |
| 9. No. of Bid technically qualified (if applicable): | <u>One</u> |
| 10. Bid(s) Rejected: | <u>Nil</u> |
| 11. Financial Bid Opening date: | <u>04.06.2015</u> |
| 12. Bid Evaluation Report: | <u>copy of minutes attached.</u> |


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Life One	Rs. 146,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s Life One @ Cost of Pak Rs. 146,000/- each on D.D.P basis
2	M/s. Sherazi Trading	Rs. 450,000/- each	3 rd Lowest			
4	M/s. Medequips SMC	Rs. 292,000/- each	2 nd Lowest			

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 Divisional Accounts Officer
 Mr. Waris Khalil
 District Accounts Officer
 Accident & Emergency Centre &
 Ancillary Services Complex At
 Office of Accountant General Sindh, Karachi
 Civil Hospital Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)
 Deputy Director (Dev)
 Health Department
 Government of Sindh
 KARACHI


 Representative of Motor Registration Wing
 Government of Sindh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)

Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.


BID EVALUATION REPORT


- | | |
|--|--|
| 1. Name of Procuring Agency: | <u>Civil Hospital, Karachi</u> |
| 2. Tender Reference No: | <u>MS/CHK/ADP/787/2014-15/- 3163</u> |
| 3. Tender Description/Name of item: | <u>Defibrillator with all Standard Accessories</u> |
| 4. Method of Procurement: | <u>Single Stage Two Envelope Procedure Basis</u> |
| 5. Tender Published: | <u>The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)</u> |
| 6. Total Bid documents Sold; | <u>Seventeen</u> |
| 7. Total Bids Received: | <u>Four</u> |
| 8. Technical Bid Opening date: (if applicable) | <u>25-05-2015</u> (Provide details in separate form) |
| 9. No. of Bid technically qualified (if applicable): | <u>One</u> |
| 10. Bid(s) Rejected: | <u>Nil</u> |
| 11. Financial Bid Opening date: | <u>04.06.2015</u> |
| 12. Bid Evaluation Report: | <u>copy of minutes attached.</u> |


S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. BIOS	Rs. 475,000/- each	2 nd Lowest			
2	M/s. Elate CC (Pvt.) Ltd.	Rs. 395,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s. Elate CC (Pvt.) Ltd. @ Cost of Pak Rs. 395,000/- each on D.D.P basis
4	M/s. Sherazi Trading	Rs. 650,000/- each	4 th Lowest			
5	M/s. Medequips SMC	Rs. 534,500/- each	3 rd Lowest			


Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHALIL
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)


 Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


 Representative of EST Department
 Government of Sindh, Karachi
 (Member)
Wajid Shaikh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


 Deputy Director (Dev)
 Health Department
 Government of Sindh
 KARACHI
 Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


 Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)


Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi.

BID EVALUATION REPORT

- | | |
|--|--|
| 1. Name of Procuring Agency: | <u>Civil Hospital, Karachi</u> |
| 2. Tender Reference No: | <u>MS/CHK/ADP/787/2014-15/- 3163</u> |
| 3. Tender Description/Name of Item: | <u>Diathermy with all Standard Accessories</u> |
| 4. Method of Procurement: | <u>Single Stage Two Envelope Procedure Basis</u> |
| 5. Tender Published: | <u>The Daily Kawish dated 09-05-2015, The Daily DAWN dated 10-05-2015 & The Daily JANG dated 11-05-2015 & Other Newspapers and hoisted on SPPRA Website vide Sr. # 24148, Print & Electronic Media (SPPRA ID No. & News papers names with dates)</u> |
| 6. Total Bid documents Sold: | <u>Seventeen</u> |
| 7. Total Bids Received: | <u>Four</u> |
| 8. Technical Bid Opening date: (if applicable) | <u>25-05-2015</u> (Provide details in separate form) |
| 9. No. of Bid technically qualified (if applicable): | <u>One</u> |
| 10. Bid(s) Rejected: | <u>Nil</u> |
| 11. Financial Bid Opening date: | <u>04.06.2015</u> |
| 12. Bid Evaluation Report: | <u>copy of minutes attached.</u> |

S #	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for Acceptance / Rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Universal Enterprises	Rs. 772,000/- each	3 rd Lowest			
2	M/s. Life One	Rs. 369,000/- each	1 st Lowest	Over All Under PC-I		The Committee is recommended to award contract M/s. Life One @ Cost of Pak Rs. 369,000/- each on D.D.P basis
4	M/s. HospiCare System	Rs. 895,000/- each	4 th Lowest			
5	M/s. Karachi Medical Co.	Rs. 375,000/- each	2 nd Lowest			

Minutes in which reasons for acceptance / rejection have been recorded


WARIS KHILIL
 Divisional Accounts Officer
 District Accounts Officer
 Office of Accountant General Sindh, Karachi
 (Member)


Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)


Waseed Shaikh
 Deputy Director (Admn)
 Motor Registration Wing
 Karachi


Asstt. Prof. Dr. Sikandar Rafique Quraishi
 Shaheed Mohtarma Benazir Bhutto Medical College, Karachi
 (Member)
Dr. Sikandar Rafique Quraishi
 MBBS MCPS
 Assistant Professor,
 Radiology
 SMBB Medical College Lyari.


Dr. Sohail Mir Memon
 Addl. Medical Superintendent (Procurement)
 Civil Hospital Karachi
 (Chairman)


Dr. Sohail Mir Memon
 Addl. Medical Superintendent
 (Procurement)
 Civil Hospital, Karachi

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 01
Name of Item. Hemodialysis Machine with all Standard Accessories
Quantity Required 10 Nos.
Estimated Price Rs. 1,000,000/-

Name of Firms	Hospital Supply Corp.	M/s. Universal Ent.
Make	TQM Toray	Nipro Corp
Model	TQS-88	Surdial 55
Country of Origin : USA / Europe / Japan	China	Japan

The offer of M/s. Hospital Supply Corp. for Model: TQS-88, Make: TQM Toray, China was not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi, because the offered model does not meet specifications as per tender requirement, Hence, not considered.

The offer of M/s. Universal Enterprises for Model: Surdial 55, Make: Nipro Corp. Japan meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firm was opened and offered the rates for Rs. 1,163,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Universal Enterprises may be awarded at the rate of Pak Rs. 1,163,000/- each on D.D.P.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 02
Name of Item. Reprocessing Unit with all Standard Accessories
Quantity Required 01 Nos.
Estimated Price Rs. 1,000,000/-

Name of Firms	M/s. Hospital Supply Corp	M/s. Universal Ent.
Make	Meditop Kidny Kleen	Medivator
Model	Compact	Renatron
Country of Origin : USA / Europe / Japan	Thailand	USA

The offer of M/s. Hospital Supply Corp. for Model: Compact, Make: Meditop Kidny Kleen, Thailand was not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meets required specifications as per tender requirement, Hence, not considered.

The offer of M/s. Universal Enterprises for Model: Renatron, Make: Nedicator, USA meets all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bid of qualified firm was opened and offered the rates for Rs. 1,188,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Universal Enterprises may be awarded at the rate of Pak Rs. 1,188,000/- each on D.D.P.

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price

Machinery / Equipment under ADP Scheme # 787/ 2014-2015
03
OT Light with all Standard Accessories
01 Nos.
Rs. 750,000/-

Name of Firms	M/s. Life One	M/s. Technology Links	M/s. Universal Ent.	M/s. Oasis Medical System
Make	AFA	Rexmed	Uzumcu	Benq
Model	ALB-100C	ROL-51C	DUO X1	DomeLUX 6200
Country of Origin : Korea / Taiwan / Turkey	Taiwan	Taiwan	Turkey	Taiwan

The offer of M/s. Oasis Medical System. for Model: DomeLUX 6200, Make: Benq Medical, Taiwan, was not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meets required specifications as per tender requirement, Hence, not considered.

The offer of M/s. Life One for Model: ALB-100C, Make: AFA Medical, Taiwan, M/s. Technology Link for Model: ROL-51C, Make: Rexmed Medical, Taiwan and M/s. Universal Enterprises for Model: DUO X1, Make: Uzumcu, Turkey meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee the financial bids of qualified firms were opened. M/s. Life One offered the rates for Rs. 605,000/- each, M/s. Technology Link offered the rates of Rs. 799,000/- each and M/s. Universal Enterprise offered the rates for Rs. 683,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Life One may be awarded at the rate of Rs. 605,000/- each on lowest basis.

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price

Machinery / Equipment under ADP Scheme # 787/ 2014-2015
04
OT Table with all Standard Accessories
01 Nos.
Rs. 650,000/-

Name of Firms	M/s. Life One	M/s. Technology Links	M/s. Universal Ent.	M/s. Oasis Medical System
Make	St-Frances	Rexmed	Bicakcilar	Benq
Model	OT-125 A	ROT-160X	550S	Novel-330T
Country of Origin : Korea / Taiwan / Turkey	Taiwan	Taiwan	Turkey	Turkey

The offer of M/s. Oasis Medical System for Model: Novel-330T, Make: Benq Medical, Taiwan was not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meets the required specifications, Hence, not considered.

The offer of M/s. Life One for Model: OT-25 A, Make: St-Frances, Taiwan, M/s. Technology Link for Model: ROT-160X, Make: Rexmed Medical, Taiwan and M/s. Universal Enterprises for Model: 550S, Make: Bicakcilar, Turkey meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. Life One offered the rates @ Rs. 585,000/- each, M/s. Technology Link offered the rates of Rs. 699,000/- each and M/s. Universal Enterprise offered the rates for Rs. 1,533,000/- each on D.D.P basis



Hence, the Procurement Committee is recommended that M/s. Life One may be awarded at the rate of Rs. 585,000/- each on lowest basis.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 05
Name of Item. Surgical Beds (Dialysis Chair) with all Standard Accessories
Quantity Required 03 Nos.
Estimated Price Rs. 165,000/-

Name of Firms	M/s. Hospital Supply Crop	M/s. Zam Trading	M/s. Technology Links
Make	Nanning Passion Medical	Winmed	Mermaid
Model	PY-YD 340	WM3000	ME-310
Country of Origin : LOCAL (Export Quality)	China	Local	China

The offer of M/s. Hospital Supply Corp. for Model: PY-YD 340, Make: Nanning Passion Medical, China, M/s. Zam Trading for Model: WM3000, Make: Winmed, Pakistan and M/s. Technology Link for Model: ME-310, Make: Mermaid, China meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. Hospital Supply Corp. offered the rates for Rs. 450,000 /- each, M/s. Zam Trading offered the rates of Rs. 164,800/- each and M/s. Technology Link offered the rates for Rs. 395,000 /- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Zam Trading may be awarded at the rate of Rs. 164,800/- each on lowest basis.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 06
Name of Item. Surgical Beds (Dialysis Chair) with all Standard Accessories
Quantity Required 03 Nos.
Estimated Price Rs. 165,000/-

Name of Firms	M/s. Hospital Supply Crop	M/s. Zam Trading	M/s. Tehnology Links
Make	Nanning Passion Medical	Winmed	Mermaid
Model	PY-YD 380	WM5001	ME-380S
Country of Origin : LOCAL (Export Quality)	China	Local	China

The offer of M/s. Hospital Supply Corp. for Model: PY-YD 380, Make: Nanning Passion Medical, China, M/s. Zam Trading for Model: WM5001, Make: Winmed, Pakistan and M/s. Technology Link for Model: ME-380S, Make: Mermaid, China meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. Hospital Supply Corp. offered the rates for Rs. Rs. 375,000/- each, M/s. Zam Trading offered the rates of Rs. 165,000/- each and M/s. Technology Link offered the rates for Rs. 389,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Zam Trading may be awarded at the rate of Rs. 165,000/- each on lowest basis.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 07
Name of Item. Crash Trolley with all Standard Accessories
Quantity Required 03 Nos.
Estimated Price Rs. 150,000/-

Name of Firms	M/s. BIOS	M/s. Elate CC	M/s. Karachi Medical Company
Make	Hersill	France Hospital	Crosmed
Model	CAVS	Perso-Life	C1122
Country of Origin : UK / USA / EEC / Japan	Spain	Italy	China

The offer of M/s. Elate CC (Pvt.) Ltd. for Model: Perso-Life, Make: France Hospital, Italy and M/s. Karachi Medical Company for Model: C1122, Make: Crosmed, China were not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as offered model does not meet the required specifications as per tender requirement, Hence, not considered.

The offer of M/s. BIOS for Model: CAVS, Make: Hersill, Spain meets all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firm was opened and offered the rates for Rs. 150,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. BIOS may be awarded at the rate of Pak Rs. 150,000/- each on D.D.P. basis

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 08
Name of Item. Stretcher with all Standard Accessories
Quantity Required 13 Nos.
Estimated Price Rs. 50,000/-

Name of Firms	M/s. Hospital Supply Corp	M/s. Zam Trading
Make	Nanning Medical	Winmed
Model	PY-CZ	ER1000
Country of Origin : LOCAL (Export Quality)	China	Local

The offer of M/s. Hospital Supply Corp. for Model: PY-CZ, Make: Nanning Passion Medical, China, M/s. Zam Trading for Model: ER1000, Make: Winmed, Local meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. Hospital Supply Corp. offered the rates for Rs. Rs. 325,000/- each & M/s. Zam Trading offered the rates of Rs. 49,900/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Zam Trading may be awarded at the rate of Rs. 49,900/- each on lowest basis.

Handwritten signatures in green, blue, and black ink, likely representing the Procurement Committee members.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
 Item No. 09
 Name of Item. Autoclave with all Standard Accessories
 Quantity Required 01 No.
 Estimated Price Rs. 1,500,000/-

Name of Firms	M/s. Life One
Make	AFA
Model	AS-450
Country of Origin : Korea / Taiwan / Turkey	Taiwan

The offer of M/s. Life One for Model: AS-450, Make: AFA Medical, Taiwan, meets all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bid of qualified firm was opened and offered the rates for Rs. 1,065,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. Life One may be awarded at the rate of Rs. 1,065,000/- each on D.D.P.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
 Item No. 10
 Name of Item. Ventilator with all Standard Accessories
 Quantity Required 04 Nos.
 Estimated Price Rs. 2,000,000/-

Name of Firms	M/s. BIOS	M/s. Noor International	M/s. HospiCare
Make	Event	Hamilton	Purete & Bennet
Model	Inspiration 5i	C1	PB 840
Country of Origin : USA / EEC / JAPAN	California / USA	Switzerland	Malaysia

The offer of M/s. Noor International for Model: C1, Make: Hamilton, Switzerland not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meets required specifications as per tender requirement, Hence, not considered.

The offer of M/s. BIOS for Model: Inspiration 5i, Make: Event Medical, California / USA and M/s. HospiCare Systems for Model: PB 840, Purete & Bennet meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bid of qualified firms were opened. M/s. BIOS offered the rates of 2,195,000/- each & M/s. HospiCare Systems offered the rates of Rs. 2,450,000/- each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. BIOS may be awarded at the rate of Pak Rs. 2,195,000/- each on D.D.P. lowest basis.

The image shows several handwritten signatures in green, blue, and black ink, likely representing the procurement committee members or the end user.

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price

Machinery / Equipment under ADP Scheme # 787/ 2014-2015
11
Fluoroscopy with DR System 500mA with all Standard Accessories
01 No.
Rs. 25,585,000/-

Name of Firms	M/s. BIOS	M/s. Ontech Corp	M/s. Medequips SMC	M/s. Medequips SMC (Alternat)	M/s. Siemens Pakistan
Make	Stephanix	Hitachi + Technix SPA	Toshiba + Primax	Primax	Siemens
Model	N40HF + Omniscope	Radnext 50 + TCA 6	Mred-A50S + Cyber Bloc	Riviera + Cyber Bloc	Multifusion + Siremobile Compact L
Country of Origin : USA / EEC / JAPAN	France	Japan + Italy	Japan + France	France	Germany

The offer of M/s. Ontech Corp. for Model: Radnext 50 + TCA 6, Make: Hitachi + Technix SPA, Japan / Italy, M/s. Medequips SMS for Model: Mred-A50S + Cyber Bloc, Make: Toshiba + Primax, Japan+ France, M/s. Medequips SMC Alternate offered Model: Riviera + Cyber Bloc, Primax, France and M/s. Siemens Pakistan for Model: Multifusion + Siremobile Compact L Make: Siemens, Germany not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meet specifications as per tender requirement, Hence, not considered.

The offer of M/s. BIOS for Model: N40HF + Omniscope, Make: Stephanix Medical, France meets all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bid of qualified firm was opened and offered the rates for Rs. 15,585,000.00 (Digital X-Ray) Model: N40HF Rs. 8,330,000.00 (C-Arm) Model: Omniscope Make: Stephanix – France Total: Rs. 23,915,000.00 each on D.D.P basis

Hence, the Procurement Committee is recommended that M/s. BIOS may be awarded at the rate of Pak Rs. 15,585,000.00 (Digital X-Ray) Model: N40HF Rs. 8,330,000.00 (C-Arm) Model: Omniscope Make: Stephanix – France Total: Rs. 23,915,000.00 each on D.D.P basis

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price

Machinery / Equipment under ADP Scheme # 787/ 2014-2015
12
ABG Machine System with all Standard Accessories
01 No.
Rs. 685,000/-

Name of Firms	M/s. Life One	M/s. HospiCare	M/s. Karachi Medical Company
Make	Medica	Techno Medical	Cobas
Model	Easy Blood Gas	Gastat - 600	B121
Country of Origin : USA / EEC / JAPAN	USA	Japan	Switzerland

The offer of M/s. Life One for Model: Easy Blood Gas, Make: Medica, USA, M/s. HospiCare Systems for Model: Gastat-600, Make: Techno Medical, Japan and M/s. Karachi Medical Company for Model: B121, Make: Cobas, Switzerland meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of



qualified firms were opened. M/s. Life One offered the rates for Rs. 1,195,000/- each, M/s. HospiCare Systems offered the rates for Rs. 985,000/- each and M/s. Karachi Medical Company offered the rates of Rs. 2,750,000/- each.

Hence, the Procurement Committee is recommended that M/s. HospiCare Systems may be awarded at the rate of Pak Rs. 985,000/- each on Lowest basis.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
Item No. 13
Name of Item. Syringe Pump with all Standard Accessories
Quantity Required 10 Nos.
Estimated Price Rs. 100,000/-

Name of Firms	M/s. S. Ejazuddin & Co.	M/s. Life One	M/s. HospiCare Systems	M/s. Elate CC (Pvt.) Ltd	M/s. Karachi Medical Co.	M/s. Universal Ent.
Make	JMS	Daiwah	B. Braun	Medima	MDX Instruments	Terumo
Model	SP-500	Medifusion-MS1200	Perfusor Compact S	Medima S	CF111	SS-700
Country of Origin : Korea / Taiwan / Turkey	Japan	Korea	Germany	Poland	USA	Japan

The offer of M/s. Karachi Medical Company for Model: CF111, Make: MDX Instruments, USA not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meets required specifications as per tender requirement, Hence, not considered.

The offer of M/s. S. Ejazuddin & Co for Model: SP-500, Make: JMS, Japan, M/s. Life One for Model: Medifusion-MS1200 , Make: Daiwah, Korea, M/s. HospiCare Systems for Model: Perfusor Compact S, Make: B. Braun, Germany, M/s. Elate CC (Pvt.) Ltd. For Model: Medima S, Make: Medima, Poland and M/s. Universal Enterprises for Model: SS-700, Make: Terumo, Japan meets all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. S. Ejazuddin & Co. offered the rates for Rs. 155,000/- each, M/s. Life One offered the rates for Rs. 91,000/- each, M/s. HospiCare System offered the rates of Rs. 145,000/- each, M/s. Elate CC (Pvt.) Ltd. offered the rates for Rs. Rs. 98,000/- each and M/s. Universal Enterprises offered the rates of Rs. 213,400/- each.

Hence, the Procurement Committee is recommended that M/s. Life One may be awarded at the rate of Pak Rs. 91,000/- each on Lowest basis.

The image shows several handwritten signatures in blue and green ink, likely representing the members of the Procurement Committee or the End User. There are approximately six distinct signatures scattered across the lower half of the page.

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price

Machinery / Equipment under ADP Scheme # 787/ 2014-2015
14
Echo Color Doppler with all Standard Accessories
01 No.
Rs. 2,000,000/-

Name of Firms	M/s. BIOS	M/s. Ontech Corp	M/s. Medequips SMC	M/s. Karachi Medical Company	M/s. Shirazi Trading
Make	Sonoscap	Hitachi	Toshiba	MDX Instruments	GE HealthCare
Model	S11	Aloka F31	Xario 100	Q7	Logic F6
Country of Origin : USA / EEC / JAPAN	China	Japan	Japan	USA	China

The offer of M/s. Karachi Medical Company for Model: Q7, Make: MDX Instruments, USA not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meets required specification as per tender requirement, Hence, not considered.

The offer of M/s. BIOS for Model: S11, Make: Sonoscap, China, M/s. Ontech Corp for Model: Aloka F31, Make: Hitachi, Japan, M/s. Medequips SMC for Model: Xario 100, Make: Toshiba, Japan and M/s. Shirazi Trading for Model: Logic F6, Make: GE HealthCare, China meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. BIOS offered the rates for Rs. 1,899,000/- each, M/s. Ontech Corp offered the rates for Rs. 2,999,000 /- each, M/s. Medequips SMC offered the rates of Rs. 2,974,000 /- each and M/s. Shirazi Trading Offered the rates for Rs. 3,350,000 /- each.

Hence, the Procurement Committee is recommended that M/s. BIOS may be awarded at the rate of Pak Rs. 1,899,000/- each on Lowest basis.

Tender Enquiry No.
Item No.
Name of Item.
Quantity Required
Estimated Price

Machinery / Equipment under ADP Scheme # 787/ 2014-2015
15
Patient Monitor with all Standard Accessories
11 Nos.
Rs. 150,000/-

Name of Firms	M/s. Life One	M/s. Medequips SMC	M/s. Karachi Medical Company	M/s. Shirazi Trading	M/s. Bio Track Trading
Make	General Meditech	Nihonkohdens	MDX Instruments	GE HealthCare	Shenzhen- N-Tech
Model	G3L	SVM-7521	Accuit Sign8	B-40	9000F
Country of Origin : USA / EEC / JAPAN	China	Japan	USA	China	China

The offer of M/s. Karachi Medical Company for Model: Accuit Sign8, Make: MDX Instruments, USA and M/s. Bio Track Trading are not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meet required specifications as per tender requirement, Hence, not considered.



The offer of M/s. Life One for Model: G3L, Make: General Meditech, China, M/s. Medequips SMC for Model: SVM-7521, Make: Nihon Kohdens, Japan and M/s. Shirazi Trading for Model: B-40, Make: GE Healthcare, China meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. Life One offered the rates for Rs. 146,000/- each, M/s. Medequips SMC offered the rates for Rs. 292,000/- each and M/s. Shirazi Trading Offered the rates for Rs. 450,000/- each.

Hence, the Procurement Committee is recommended that M/s. Life One may be awarded at the rate of Pak Rs. 146,000/- each on Lowest basis.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
 Item No. 16
 Name of Item. Defibrillator with all Standard Accessories
 Quantity Required 01 No.
 Estimated Price Rs. 475,000/-

Name of Firms	M/s. BIOS	M/s. Medequips SMC	M/s Elate CC	M/s. Karachi Medical Company	M/s. Shirazi Trading
Make	Osatu	Nihonkohdens	Primedic	MDX Instruments	Zoll Medica
Model	R700	TEC-5521K	XD1	Hartstart Pro	M Series
Country of Origin : USA / EEC / JAPAN	Spain	Japan	Germany	USA	USA

The offer of M/s. Karachi Medical Company for Model: Harstart Pro, Make: MDX Instruments, USA not considered on the technical ground basis by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as the offered model does not meet required specifications as per tender requirement, Hence, not considered.

The offer of M/s. BIOS for Model: R700, Make: Osatu Medical, Spain, M/s. Medequips SMC or Model: TEC-5521K, Make: Nihon Kohen, Japan, M/s. Elate CC (Pvt.) Ltd. and M/s. Shirazi Trading for Model: M Series, Make: Zoll Medica, USA meet all the required specifications as per tender requirement and also recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. BIOS offered the rates for Rs. 475,000/- each, M/s. Medequips SMC offered the rates for Rs. 534,500/- each M/s. Elate CC (Pvt.) Ltd offered the rates of Rs. 395,000/- each and M/s. Shirazi Trading Offered the rates for Rs. 650,000/- each.

Hence, the Procurement Committee is recommended that M/s. Elate CC (Pvt.) Ltd. may be awarded at the rate of Pak Rs. 395,000/- each on Lowest basis.

Tender Enquiry No. Machinery / Equipment under ADP Scheme # 787/ 2014-2015
 Item No. 17
 Name of Item. Diathermy with all Standard Accessories
 Quantity Required 01 No.
 Estimated Price Rs. 375,000/-


Name of Firms	M/s. Life One	M/s. HospiCare Systems	M/s. Karachi Medical Company	M/s. Universal Ent.
Make	Daiwha	Vally-Lab	Beful	Tekno
Model	DT-400P	Force FX	Doctanz 400	90010-04
Country of Origin : Korea / Taiwan / Turkey	Korea	USA	Korea	Germany

The offer of M/s. Life One for Model: DT-400P, Make: Daiwha, Korea, M/s. Hospicare Systems for Model: Force FX, Make: Vally Lab, USA, M/s. Karachi Medical Company for Model: Doctanz 400, Make: Beful, Korea and M/s. Universal Enterprise for Model: 90010-04, Make: Tekno, Germany meet all the required specifications as per tender requirement and recommended by the End User Prof. Iffat Yazdani Consultant Nephrologist, Clifton Dialysis Centre Karachi as approved by the Procurement Committee, the financial bids of qualified firms were opened. M/s. Life One offered the rates for Rs. 369,000/- each, M/s. Hospicare Systems offered the rates for Rs. 895,000/- each, M/s. Karachi Medical Company offered the rates of Rs. 375,000/- each and M/s. Universal Enterprises Offered the rates for Rs. 772,000/- each.


Hence, the Procurement Committee is recommended that M/s. Life One may be awarded at the rate of Pak Rs. 369,000/- each on Lowest basis.


WARIS KHALIL
Divisional Officer
Mr. Waris Khalil
District Accounts Officer
Office of Accountant General Sindh, Karachi
(Member)


Representative of Health Department (Dev. Wing)
Government of Sindh, Karachi
(Member)


Representative of E & T Department
Government of Sindh, Karachi
(Member)
Waleed Shaikh
Deputy Director (Admn)
Motor Registration Wing
Karachi

Deputy Director (E&T)
Health Department
Government of Sindh
KARACHI


Asstt. Prof. Dr. Sikandar Rafique Quraishi
Shaheed Mohtarma Benazir Bhutto Medical College,
Karachi
Dr. Sikandar Rafique Quraishi
(Member)
MBBS MCPS
Assistant Professor,
Radiology
SMBB Medical College Lyari.


Dr. Sohail Mir Memon
Addl. Medical Superintendent (Procurement)
Civil Hospital Karachi
(Chairman)


Dr. Sohail Mir Memon
Addl. Medical Superintendent
(Procurement)
Civil Hospital, Karachi.

THE COMPARATIVE STATEMENT FOR THE SUPPLY OF MACHINERY EQUIPMENT UNDER ADP SCHEME NO 787/2014-2015 QUOTED BY THE VARIOUS FIRMS VIDE TENDER ENQUIRY NO.MS/CHK/2014-2015/3163 DATED 07-05-2015 PUBLISHED IN NEWS PAPERS I.E. THE DAILY KAWISH DATED 09-05-2015, THE DAILY DAWN DATED 10-05-2015 & THE DAILY JANG DATED 11-05-2015 AND OTHER NEWS PAPERS AND HOISTED ON SPRA WEBSITE @ SR. NO.24148

TECHNICAL BIDS OPENED ON 25-05-2015 AND FINANCIAL BIDS OPENED ON 04-06-2015 AT 03.00 P.M.

S.#	Name of Items	Quantity	M/s. Universal Enterprises	M/s. S. Ejazuddin & Co.	M/s. Hospital Supply Corporation	M/s. Life One	M/s. BIOS
1	Hemodialysis Machine with all Standard Accessories	10	Rs. 1,163,000.00 Model: Sundial 55 Make: Nipro Corp. - Japan				
2	Reprocessing Unit with all Standard Accessories	1	Rs. 1,188,000.00 Model: Renatron Make: Medivator - USA				
3	OT Light with all Standard Accessories	1	Rs. 683,000.00 Model: DUO X1 Make: Uzunca - Turkey			Rs. 605,000.00 Model: ALB-100C Make: AFA - Taiwan	
4	OT Table with all Standard Accessories	1	Rs. 1,533,000.00 Model: 5505 Make: Bircaklar - Turkey			Rs. 585,000.00 Model: OT-115 A Make: St-Francis - Taiwan	
5	Surgical Beds (Dialysis Chair) with all Standard Accessories	3		Rs. 450,000.00 Model: PV-YD 340 Make: Nanning Passion Medical - China			
6	Surgical Beds (Dialysis Recliners) with all Standard Accessories	3		Rs. 375,000.00 Model: PV-YD 380 Make: Nanning Passion Medical - China			
7	Crash Trolley with all Standard Accessories	3		Rs. 325,000.00 Model: PV-CZ Make: Nanning Passion Medical - China			Rs. 150,000.00 Model: CAVS Make: Herstill - Spain
8	Stretcher with all Standard Accessories	13					
9	Autoclave with all Standard Accessories	1				Rs. 1,065,000.00 Model: AS-450 Make: AFA - Taiwan	


Dr. Sikandar Rafique Qurashi
 Assistant Professor,
 MBBS MCPS
 Radiology
 SMBB Medical College Lyari.


Waheed Shaikh
 Deputy Director (Admin)
 Motor Registration Wing
 Karachi

S. #	Name of Items	Quantity	M/s. Universal Enterprises	M/s. S. Ejazuddin & Co.	M/s. Hospital Supply Corporation	M/s. Life One	M/s. BIOS
10	Ventilator with all Standard Accessories	4					Rs. 2,195,000.00 Model: Inspiration S1 Make: Event - California / USA
11	Fluoroscopy with DR System 500ma with all Standard Accessories	1					Rs. 15,585,000.00 (Digital X-Ray) Model: N40HF Rs. 8,330,000.00 (C-Arm) Model: Omniscope Make: Stephanik - France Total: Rs. 23,915,000.00
12	ABG Machine System with all Standard Accessories	1				Rs. 1,195,000.00 Model: Easy Blood Gas Make: Medica - USA	
13	Syringe Pump with all Standard Accessories	10	Rs. 213,400.00 Model: SS-700 Make: Terumo - Japan	Rs. 155,000.00 Model: SP-500 Make: JMS - Japan		Rs. 91,000.00 Model: Medifusion-MS1200 Make: Daiwah - Korea	
14	Echo Color Doppler with all Standard Accessories	1				Rs. 1,899,000.00 Model: S11 Make: Sonoscap - China	
15	Patient Monitor with all Standard Accessories	11				Rs. 146,000.00 Model: G1L Make: General Meditech - China	
16	Defibrillator with all Standard Accessories	1				Rs. 369,000.00 Model: DT-400P Make: Daiwha - Korea	Rs. 475,000.00 Model: R700 Make: Osatu - Spain
17	Dialtherapy with all Standard Accessories	1	Rs. 772,000.00 Model: 90010-04 Make: Tekno - Germany				

WARRINGTON
 Mr. Waqar Khalid
 Divisional Accounts Officer
 Office of Accountant General Sindh, Sector-8
 Ancillary Services Complex At
 Civil Hospital Karachi

Representative of Health Department (Dev. Wing)
 Government of Sindh, Karachi
 (Member)

Wahid Shaikh
 Deputy Registration Officer
 Motor Registration
 Karachi

