



# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**
3. Tender Description/Name of item: **(1) Repair and Service of Mobile X-Ray Machine Model:- 325 R  
Make:- Dongmun at Medical ICU.**
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA  
website at S.No.26000**
6. Total Bid documents Sold; **One**
7. Total Bids Received: **One**
8. Tender Opening date:- **31-12-2015**
9. Bid(s) Rejected: **Nil**
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. BIOS	Rs. 112,000	1 <sup>st</sup> Lowest	As per the Estimated cost	Authorized Agent	The Committee is recommended to award the contract to M/s. BIOS at Rs. 112,000


**The Committee is recommended to award the contract on Authorized Agent basis.**

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(2) Repair and Service of Color Doppler Model: - Imageic Maestro S. No. 0980472 Make:- Kontron / France at Radiology Department.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**  
6. Total Bid documents Sold; **One**  
7. Total Bids Received: **One**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. BIOS	Rs. 315,000	1 <sup>st</sup> Lowest	As per the Estimated cost	Authorized Agent	The Committee is recommended to award the contract to M/s. BIOS at Rs. 315,000


**The Committee is recommended to award the contract on Authorized Agent basis.**

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(3) Repair of Multipara Monitors Model:- OMNI & UT4000  
Make:- Infinium Medical & Gold Way at E.O.T.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA  
website at S.No.26000**  
6. Total Bid documents Sold; **Three**  
7. Total Bids Received: **Three**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s Zam Trading	Rs. 222,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 222,000
2	M/s BIOS	Rs. 228,000	2 <sup>nd</sup> Lowest			
3	M/s Life One	Rs. 230,000	3 <sup>rd</sup> Lowest			


The Committee is recommended to award the contract on lowest basis.

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(4) Repair of Multipara Monitors Model: - OMNI-III Make:-  
Infinium Medical at Surgical ICU.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA  
website at S.No.26000**  
6. Total Bid documents Sold; **Three**  
7. Total Bids Received: **Three**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/ rejection	Remarks
0	1	2	3	4	5	6
1	M/s Zam Trading	Rs. 516,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 516,000
2	M/s Life One	Rs. 518,000	2 <sup>nd</sup> Lowest			
3	M/s BIOS	Rs. 522,000	3 <sup>rd</sup> Lowest			


The Committee is recommended to award the contract on lowest basis.

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(5) Repair and Service of Ventilator Model:- C2 Make:- Hamilton S # 2783 at Medical ICU.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**  
6. Total Bid documents Sold; **One**  
7. Total Bids Received: **One**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Noor International	Rs. 220,000	1 <sup>st</sup> Lowest	As per the Estimated cost	Authorized Agent	The Committee is recommended to award the contract to M/s. Noor International at Rs. 220,000


The Committee is recommended to award the contract on Authorized Agent basis.

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(6) Repair of Ultrasound Machine Model:- HS2000 Make:- Honda at Medical ICU.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**  
6. Total Bid documents Sold: **Three**  
7. Total Bids Received: **Three**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Zam Trading	Rs. 158,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 158,000
2	M/s. Health Tech	Rs. 162,000	2 <sup>nd</sup> Lowest			
3	M/s Biomatrix Pak	Rs. 165,000	3 <sup>rd</sup> Lowest			


The Committee is recommended to award the contract on lowest basis.

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi

# **BID EVALUATION REPORT**


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(7) Repair of Multipara Monitors Model:- DST5000 Make:- DST / USA at Medical Unit-III.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**  
6. Total Bid documents Sold: **Three**  
7. Total Bids Received: **Three**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Zam Trading	Rs. 165,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 165,000
2	M/s. Health Tech	Rs. 171,000	2 <sup>nd</sup> Lowest			
3	M/s Biomatrix Pak	Rs. 174,900	3 <sup>rd</sup> Lowest			


**The Committee is recommended to award the contract on lowest basis.**

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(10) Repair of Multipara Monitors Model:- UT4000 Make:- Gold Way at Peads Unit-II.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**  
6. Total Bid documents Sold: **Three**  
7. Total Bids Received: **Three**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Zam Trading	Rs. 186,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 186,000
2	M/s. Health Tech	Rs. 189,000	2 <sup>nd</sup> Lowest			
3	M/s Biomatrix Pak	Rs. 195,000	3 <sup>rd</sup> Lowest			


**The Committee is recommended to award the contract on lowest basis.**

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi

  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi





# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**
3. Tender Description/Name of item: **(11) Repair of Baby Incubator Model:- YP-90A Make:- Ningbo David at Peads Unit-II.**
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**
6. Total Bid documents Sold: **Three**
7. Total Bids Received: **Three**
8. Tender Opening date:- **31-12-2015**
9. Bid(s) Rejected: **Nil**
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Zam Trading	Rs. 300,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 300,000
2	M/s. Biomatrix Pak	Rs. 304,800	2 <sup>nd</sup> Lowest			
3	M/s. Health Tech	Rs. 315,000	3 <sup>rd</sup> Lowest			


**The Committee is recommended to award the contract on lowest basis.**

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# **BID EVALUATION REPORT**


1. Name of Procuring Agency: **Civil Hospital, Karachi**  
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**  
3. Tender Description/Name of item: **(14) Repair of Multipara Monitors Model:- OMNI-III Make:- Infinium at Medical ICU.**  
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**  
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**  
6. Total Bid documents Sold; **Three**  
7. Total Bids Received: **Three**  
8. Tender Opening date:- **31-12-2015**  
9. Bid(s) Rejected: **Nil**  
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Zam Trading	Rs. 580,000	1 <sup>st</sup> Lowest	As per the Estimated cost	On Lowest Basis	The Committee is recommended to award the contract to M/s. Zam Trading at Rs. 580,000
2	M/s. Biomatrix Pak	Rs. 590,000	2 <sup>nd</sup> Lowest			
3	M/s. Health Tech	Rs. 595,000	3 <sup>rd</sup> Lowest			


**The Committee is recommended to award the contract on lowest basis.**

  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi


  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi


# BID EVALUATION REPORT


1. Name of Procuring Agency: **Civil Hospital, Karachi**
2. Tender Reference No: **No. AMS(Tech) /-9988 dated 07-12-2015.**
3. Tender Description/Name of item: **(15) Installation of Air Compressor System for Neurology Unit.**
4. Method of Procurement: **Single Stage One Envelop Procedure Basis**
5. Tender Published: **The Daily "Dawn" dated 10-12-2015 and hoisted on SPPRA website at S.No.26000**
6. Total Bid documents Sold; **One**
7. Total Bids Received: **One**
8. Tender Opening date:- **31-12-2015**
9. Bid(s) Rejected: **Nil**
10. Bid Evaluation Report:


S. No.	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1	M/s. Standard Suppliers	Rs. 660,000	1 <sup>st</sup> Lowest	As per the Estimated cost	Authorized Agent	The Committee is recommended to award the contract to M/s. Standard Suppliers at Rs. 660,000


**The Committee is recommended to award the contract on Authorized Agent basis.**


  
**Mr. Shahid Ali Khan**  
Accounts Officer  
Civil Hospital, Karachi

  
**Dr. Abdul Qadir Siddiqui**  
Addl. Medical Superintendent (General)  
Civil Hospital Karachi

  
**Dr. M. Nawaz Lashari**  
Asstt. Prof. of Cardiology  
Dow University of Health Sciences Karachi

  
**Dr. Siraj Haider**  
Asstt. Prof. of Surgery  
Dow University of Health Sciences Karachi

  
**Dr. Neelam Jan**  
Addl. Medical Superintendent (Technical)  
Civil Hospital Karachi



**N.I.T NO. 10 FOR REPAIR OF MACHINERY EQUIPMENTS**  
**CRITERIA FOR EVALUATION OF THE BID**


**NIT NO. INF/KRY-4085/15 SPPRA ID @ S. No. 26000**

		M/s Standard Suppliers		M/s Noor Int.		M/s BIOS		M/s Life One		M/s Biomatrix		M/s Health Tech	
		Yes	No	Yes	No	Yes	No	Yes	No	Yes	No	Yes	No
1.	Compliance of Terms & Conditions / Instructions mentioned in the Bid Form / NIT.	✓	-	✓	-	✓	-	✓	-	✓	-	✓	-
2.	Relevant Experience <i>(Three years)</i>	✓	-	✓	-	✓	-	✓	-	✓	-	✓	-
3.	Registration with Income Tax / Sales Tax. (Mandatory) / Sindh Sales Tax (SST) If applicable.	✓	-	✓	-	✓	-	✓	-	✓	-	✓	-
4.	Copy of Bank Certificate regarding financial soundness and Turn over for the last three years. (5.000 (M) Turn over in each year).	✓	-	✓	-	✓	-	✓	-	✓	-	✓	-

*[Handwritten signature]*

*[Handwritten mark]*

*Nadeem*



**BIOS**

SERVING HUMANITY WITH TECHNOLOGY

December 28, 2015

Medical Superintendent  
Civil Hospital  
Karachi.

Sub: Repair and Service Estimate of Mobile X-ray Machine

Dear Sir,

With reference to your call & thorough check up of the subject equipment by our engineers. We are pleased to submit as follows.

Medical I.C.U

01. Mobile X-ray Machine

Mod: 325R

Make: Dongmun

WORK REQUIRE AS FOLLOWS:


- Repair require of collimator circuit.
- Repair require of filament.
- Repair require of mechanical part of column.
- Repair require of lock circuit.
- Complete Service & calibration.

Rs: 112,000/-

(Rupees: One Lac Twelve Thousand Only)

Amir  
31/12/15

*[Handwritten signature]*  
31/12/15



**Head Office:**  
2nd, Floor Building No. 2C  
12 Commercial, Phase II Ext.  
DHA, Karachi-Pakistan  
Ph: 021-35891031 / 021-35891032  
021-35891033  
E-mail : bios@cyber.net.pk  
Web.: www.bios.com.pk

**Lahore Office:**  
1st Floor, Munawar Apartment,  
Behind Queens Centre,  
33/1 Queens Road, Lahore.  
Ph : 042-36310519  
Fax : 042-36311697  
E-mail : bios@cyber.net.pk

**Islamabad Office:**  
Room No. 3-A,  
Flat No. 12, 3rd Floor,  
Building # 82 / West Muzaffar Chamber,  
Fazal-e-Haq Road,  
Blue Area, Islamabad  
Phone : 051-2802555 Fax : 051-2802555

**ANNEXURE-II**




Tender Fee Rs. 500/-  
(Non-Refundable)

**STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE &  
REPAIR OF THE MACHINERY EQUIPMENT AT CIVIL HOSPITAL  
KARACHI DURING THE FINANCIAL YEAR 2015-16**

Date of Submission & Time: - 31-12-2015 @ 11:30 a.m  
Date of Opening & Time: - 31-12-2015 @ 12:00 noon


S. No.	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
1	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325R Make:- Dongmun	01 No.	0.112	112,000/-
2	Repair and Service of Color Doppler Machine at Radiology Department. Model:- Imageic Maestro Make:- Kontron / France S. No. 0980472	01 No.	0.315	315,000/-
3	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold way	04 Nos.	0.222	228,000/-
4	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	0.516	522,000/-
5	Repair and Service of Ventilator at Medical ICU. Model: - C2 Make:- Hamilton Sr. No. 2783 Mixer Block Assembly Service and Calibration	01 No.	0.220	
6	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	0.158	
7	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	0.165	
8	Repair and Service of Convex Probe of Toshiba NEMIO/XG at Radiology Department. Convex Probe PVM-375AT	01 No.	0.714	
9	Repair and Service of SYSC-PWB of Toshiba NEMIO/17 at Radiology Department. SYSC-PWB	01 No.	0.684	
10	Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold way	03 Nos.	0.186	
11	Repair of Baby Incubator at Peads Unit-II Model:- YP-90A Make:- Ningbo David	06 Nos.	0.300	
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	0.735	



	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
3	Repair, Maintenance and Installation of Liquid Chiller at M.R.I Department.	01 No.	0.715	
4	Repair of Multipara Monitors at Medical ICU. Model:- OMNI-III Make:- Infinium Medical	10 Nos.	0.580	
15	Installation of Air Compressor System for Neurology Unit.  Supply of Air Compressor System for Neurology Ward  Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.  Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.  Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.  After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.  Electric Panel for above system  Water Separator	01 No.  02 Nos.  02 Nos.  02 Nos.  01 No.  01 No.	0.665	
<b>Total Estimated Cost</b>			<b>6.287</b>	<b>11,77,000/-</b>

**Terms & Conditions:**

1. Warranty period for repair work must be given with offer if the repaired equipment is out of order during the warranty period the same will be repaired without charging any cost by the concerned firm.
2. If the offer / quoted amount value is more than Rs. 25,000/- 2% Security deposit in shape of pay order in favor of Medical Superintendent Civil Hospital, Karachi, should be attached with offer the same will be refunded after 03 Months. If not received any complain of the repaired equipment.
3. The bid shall be opened on the above said date & time in the presence of the Hospital Procurement Committee (For Repair). The offer will be evaluated on best evaluated basis.





**B I O S**

SERVING HUMANITY WITH TECHNOLOGY

December 28, 2015

Medical Superintendent  
Civil Hospital  
Karachi.

Sub: Repair and Service Estimate of Color Doppler Machine

Dear Sir,

With reference to your call & thorough check up of the subject equipment by our engineers. We are pleased to submit as follows.

Radiology Department

02. Color Doppler

Mod: Imageic Maestro

S.No: 0980472

Make: Kontron/France

WORK REQUIRE AS FOLLOWS:

- Replacement require of linear probe.
- Complete Service & calibration.

Rs:315,000/-

(Rupees: Three Lac Fifteen Thousand Only)

*Handwritten signatures and notes:*  
1. A signature with the date 21/12/15.  
2. A signature with the amount 315,000/- written below it.  
3. A large, stylized signature.

**Head Office:**

2nd, Floor Building No. 2C

12 Commercial, Phase II Ext.

D.H.A. Karachi-Pakistan

Ph: 021-35891031 / 021-35891032

021-35891033

E-mail : bios@cyber.net.pk

Web.: www.bios.com.pk

**Lahore Office:**

1st Floor, Munawar Apartment,

Behind Queens Centre,

33/1 Queens Road, Lahore.

Ph : 042-36310519

Fax : 042-36311697

E-mail : bios@cyber.net.pk

**Islamabad Office:**

Room No. 3-A,

Flat No. 12, 3rd Floor,

Building # 82 / West Muzaffar Chamber,

Fazal-e-Haq Road,

Blue Area, Islamabad

Phone : 051-2802555 Fax : 051-2802555





# ZAM TRADING

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Multipara Monitor

Dear Sir,

We here by submit our estimated price for the repair item.

E.O.T

S.No.03. Multipara monitors (04Nos)  
Model: OMNI & UT4000  
Make: Infinium Medical & Goldway

Work Require as follows:

- ❖ Replacement need of Complete accessories.
- ❖ Repair need of NIBP module.
- ❖ Replacement need of battery pack.
- ❖ Complete Servicing & Calibration.

1No. Rs. 55,500/=

Total Rs. 222,000/- (04Nos)

Rupees: Two Lac Twenty Two Thousand Only.

TERMS & CONDITIONS:

VALIDITY : 3 Weeks

WARRANTY : Six Months only for those faults which are mentioned in our quotation.

REPAIR PERIOD : 3-5 weeks or earlier.

PAY ORDER : Rs. 44400/- Dated. 26/12/15 No. 50611632

If the work awarded to us, we assure you our best cooperation and service.

Your Truly

ZAM TRADING  
ZAM Trading

*[Handwritten signature and date 31/12/15]*

**ANNEXURE-II**

Tender Fee Rs. 500/-  
(Non-Refundable)

**STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE &  
REPAIR OF THE MACHINERY EQUIPMENT AT CIVIL HOSPITAL  
KARACHI DURING THE FINANCIAL YEAR 2015-16**

Date of Submission & Time: - 31-12-2015 @ 11:30 a.m

Date of Opening & Time: - 31-12-2015 @ 12:00 noon

S. No.	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
1	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325R Make:- Dongmun	01 No.	0.112	/
2	Repair and Service of Color Doppler Machine at Radiology Department. Model:- Imageic Maestro Make:- Kontron / France S. No. 0980472	01 No.	0.315	/
3	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold way	04 Nos.	0.222	222,000/=
4	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	0.516	516,000/=
5	Repair and Service of Ventilator at Medical ICU. Model: - C2 Make:- Hamilton Sr. No. 2783 Mixer Block Assembly Service and Calibration	01 No.	0.220	
6	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	0.158	158,000/=
7	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	0.165	165,000/=
8	Repair and Service of Convex Probe of Toshiba NEMIO/XG at Radiology Department. Convex Probe PVM-375AT	01 No.	0.714	
9	Repair and Service of SYSC-PWB of Toshiba NEMIO/17 at Radiology Department. SYSC-PWB	01 No.	0.684	
10	Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold way	03 Nos.	0.186	186,000/=
11	Repair of Baby Incubator at Peads Unit-II Model:- YP-90A Make:- Ningbo David	06 Nos.	0.300	300,000/=
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	0.735	

	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
	Maintenance and Installation of Liquid ... at M.R.I Department.	01 No.	0.715	
	Supply of Multipara Monitors at Medical ICU. OMNI-III Infinium Medical	10 Nos.	0.580	580,000/-
	Installation of Air Compressor System for Neurology Unit.			
	Supply of Air Compressor System for Neurology Ward	01 No.		
	Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.	02 Nos.		
	Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.	02 Nos.	0.665	
15	Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.		
	After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.		
	Electric Panel for above system	01 No.		
	Water Separator	01 No.		
	<b>Total Estimated Cost</b>		<b>6.287</b>	

**Terms & Conditions:**

1. Warranty period for repair work must be given with offer if the repaired equipment is out of order during the warranty period the same will be repaired without charging any cost by the concerned firm.
2. If the offer / quoted amount value is more than Rs. 25,000/- 2% Security deposit in shape of pay order in favor of Medical Superintendent Civil Hospital, Karachi, should be attached with offer the same will be refunded after 03 Months. If not received any complain of the repaired equipment.
3. The bid shall be opened on the above said date & time in the presence of the Hospital Procurement Committee (For Repair). The offer will be evaluated on best evaluated basis.

  
ZAM TRADING



**B I O S**

SERVING HUMANITY WITH TECHNOLOGY

December 28, 2015

Medical Superintendent  
Civil Hospital  
Karachi.

Sub: Repair and Service Estimate of Multipara Monitor

Dear Sir,

With reference to your call & thorough check up of the subject equipment by our engineers. We are pleased to submit as follows.

E.O.T

03. Multipara Monitor (04 Nos)

Mod: OMNI & UT4000

Make: Infinium/Goldway

WORK REQUIRE AS FOLLOWS:

- Replacement require of complete accessories.
- Repair require of NIBP module.
- Replacement require of battery pack.
- Complete Servicing & Calibration.

RS:57,000/- (01Nos)

Total Rs:228,000/- (04Nos.)

(Rupees: Two Lac Twenty Eight Thousand Only)

**Head Office:**

1st Floor Building No. 2C  
Commercial, Phase II Ext.  
A. Karachi-Pakistan  
021-35891031 / 021-35891032  
021-35891033  
E-mail: bios@cyber.net.pk  
www.bios.com.pk

**Lahore Office:**

1st Floor, Munawar Apartment,  
Behind Queens Centre,  
33/1 Queens Road, Lahore.  
Ph : 042-36310519  
Fax : 042-36311697  
E-mail : bios@cyber.net.pk

**Islamabad Office:**

Room No. 3-A,  
Flat No. 12, 3rd Floor,  
Building # 82 / West Muzaffar Chamber,  
Fazal-e-Haq Road,  
Blue Area, Islamabad  
Phone : 051-2802555 Fax : 051-2802555

# LIFE ONE

28 December 2015  
Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation Regarding Repair for Multipara Monitor

Dear Sir,  
We here by submit our estimated price for the repair item.

## E.O.T

Item # 03.  
**Multipara Monitor (04NOS)**  
Model: OMNI&UT4000  
Make: Infinium/Goldway

### Work Require as follows:

- \* Replacement need of Accessories.
- \* Repair need of main pcb.
- \* Replacement need of NIBP pumps.
- \* Replacement need of battery pack.
- \* Need Complete Service & Calibration.

01 No. Rs.57,500/-  
Total 04 Nos. Rs.230,000/-  
Rs: Two Lac Thirty Thousand Only.

### TERMS & CONDITIONS:

VALIDITY : 3 Weeks.  
WARRANTY : One year only for those faults which are mentioned  
in our quotation and service report.  
REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 50611670 Date: 27/12/15 Amount Rs: 46007

If the work awarded to us, we assure you our best cooperation and service.

Your Truly,



Suit D-1, 1<sup>st</sup> floor, West Land Trade Center, C-5, Central Commercial Area, Block 7&8, KCHSU Ltd, Karachi  
Ph : 021-34306050, 36067758 – 36067757 , Fax: 021-34802091



ANNEXURE-II

Tender Fee Rs. 500/-  
(Non-Refundable)

**STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE &  
REPAIR OF THE MACHINERY EQUIPMENT AT CIVIL HOSPITAL  
KARACHI DURING THE FINANCIAL YEAR 2015-16**

Date of Submission & Time: - 31-12-2015 @ 11:30 a.m  
Date of Opening & Time: - 31-12-2015 @ 12:00 noon

S. No.	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
1	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325R Make:- Dongmun	01 No.	0.112	/
2	Repair and Service of Color Doppler Machine at Radiology Department. Model:- Imageic Maestro Make:- Kontron / France S. No. 0980472	01 No.	0.315	/
3	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold way	04 Nos.	0.222	230,000/-
4	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	0.516	518,000/-
5	Repair and Service of Ventilator at Medical ICU. Model: - C2 Make:- Hamilton Sr. No. 2783 Mixer Block Assembly Service and Calibration	01 No.	0.220	/
6	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	0.158	/
7	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	0.165	/
8	Repair and Service of Convex Probe of Toshiba NEMIO/XG at Radiology Department. Convex Probe PVM-375AT	01 No.	0.714	/
9	Repair and Service of SYSC-PWB of Toshiba NEMIO/17 at Radiology Department. SYSC-PWB	01 No.	0.684	/
10	Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold way	03 Nos.	0.186	/
11	Repair of Baby Incubator at Peads Unit-II Model:- YP-90A Make:- Ningbo David	06 Nos.	0.300	/
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	0.735	/



S. No.	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered	
13	Repair, Maintenance and Installation of Liquid Chiller at M.R.I Department.	01 No.	0.715	748,000/-	
14	Repair of Multipara Monitors at Medical ICU. Model:- OMNI-III Make:- Infinium Medical	10 Nos.	0.580		
15	Installation of Air Compressor System for Neurology Unit.  Supply of Air Compressor System for Neurology Ward	01 No.	0.665		
	Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.	02 Nos.			
	Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.	02 Nos.			
	Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.			
	After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.			
	Electric Panel for above system	01 No.			
	Water Separator	01 No.			
<b>Total Estimated Cost</b>			<b>6.287</b>		


**Terms & Conditions:**

1. Warranty period for repair work must be given with offer if the repaired equipment is out of order during the warranty period the same will be repaired without charging any cost by the concerned firm.
2. If the offer / quoted amount value is more than Rs. 25,000/- 2% Security deposit in shape of pay order in favor of Medical Superintendent Civil Hospital, Karachi, should be attached with offer the same will be refunded after 03 Months. If not received any complain of the repaired equipment.
3. The bid shall be opened on the above said date & time in the presence of the Hospital Procurement Committee (For Repair). The offer will be evaluated on best evaluated basis.

*[Handwritten Signature]*  
21/12/2017



*[Handwritten Signature]*  
21/12/17



# ZAM TRADING

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Multipara Monitor

Dear Sir,  
We here by submit our estimated price for the repair item.

### SICU

S.No.04. Multipara monitors (08Nos)  
Model: OMNI III  
Make: Infinium Medical

### Work Require as follows:

- ❖ Replacement need of accessories.
- ❖ Repair need of main board.
- ❖ Repair need of NIBP module.
- ❖ Replacement need of battery pack.
- ❖ Complete Servicing & Calibration.

1No. Rs. 64,500/=

Total Rs. 516,000/- (08Nos)

Rupees: Five Lac Sixteen Thousand Only.

### TERMS & CONDITIONS:

VALIDITY : 3 Weeks  
WARRANTY : Six Months only for those faults which are mentioned in our quotation.  
REPAIR PERIOD : 3-5 weeks or earlier.  
PAY ORDER : Rs. 10,320/- Dated. 28/12/15 NO: 56611673

If the work awarded to us, we assure you our best cooperation and service.

Your Truly

ZAM TRADING  
ZAM Trading

*[Handwritten signatures and dates]*  
31/12/15  
31/12/15



# LIFE ONE

28 December 2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation Regarding Repair for Multipara Monitor

Dear Sir,

We here by submit our estimated price for the repair item.

S.I.C.U

Item # 04.

**Multipara Monitor (08NOS)**

Model: OMNI III

Make: Infinium Medical

Work Require as follows:

- \* Replacement need of Accessories.
- \* Repair need of main pcb.
- \* Repair need of ecg amplifier pcb.
- \* Replacement need of NIBP pumps.
- \* Replacement need of battery pack.
- \* Need Complete Service & Calibration.

01 No. Rs 64,750/-

08 Nos. Rs.518,000/=

Rs: Five Lac Eighteen Thousand Only.

TERMS & CONDITIONS:

VALIDITY : 3 Weeks.


WARRANTY : One year only for those faults which are mentioned  
in our quotation and service report.

REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 56611631 Date: 28/12/15 Amount Rs: 10,360/-

If the work awarded to us, we assure you our best cooperation and service.

Your Truly,



*Handwritten signature and date: 28/12/15*

*Handwritten signature and date: 31/12/15*



**B I O S**  
SERVING HUMANITY WITH TECHNOLOGY

December 28, 2015

Medical Superintendent  
Civil Hospital  
Karachi.

Sub: Repair and Service Estimate of Multipara Monitor

Dear Sir,

With reference to your call & thorough check up of the subject equipment by our engineers. We are pleased to submit as follows.

S.I.C.U

04. Multipara Monitor (08Nos)  
Mod: OMNI III  
Make: Infinium

WORK REQUIRE AS FOLLOWS:

- Replacement require of complete accessories.
- Repair require of mother board.
- Repair need of NIBP pack module.
- Replacement require of battery pack.
- Complete Servicing & Calibration.


RS:65,250/- (01Nos)  
Total Rs:522,000/- (08Nos.)  
(Rupees: Five Lac Twenty Thousand Only)

*Handwritten signature and date: 31/12/2015*

Head Office:  
1st Floor Building No. 2C  
Commercial, Phase II Ext.  
Karachi-Pakistan  
021-35891031 / 021-35891032  
021-35891033  
Email: bios@cyber.net.pk  
www.bios.com.pk

Lahore Office:  
1st Floor, Munawar Apartment,  
Behind Queens Centre,  
33/1 Queens Road, Lahore.  
Ph : 042-36310519  
Fax : 042-36311697  
E-mail : bios@cyber.net.pk

Islamabad Office:  
Room No. 3-A,  
Flat No. 12, 3rd Floor,  
Building # 82 / West Muzaffar Chamber,  
Fazal-e-Haq Road,  
Blue Area, Islamabad  
Phone : 051-2802555 Fax : 051-2802555



# NOOR INTERNATIONAL

Head Office : Noor House, 29-D, Block-6, P.E.C.H. Society, Karachi-75400, Pakistan.  
Tel # 92-21-34527001-3, 34322311-3, Fax # 92-21-3452-7000, E-mail : niho@noor-international.com  
Website : www.noor-international.com

## QUOTATION

Our Ref#NI/WS/DEC-0370/15

December 26, 2015

Attn: The Medical Superintendent  
Civil Hospital  
MICU  
Karachi- Pakistan

Subject: Quotation for Repairing of ventilator.

Make: Hamilton  
Model: C2  
Serial No: 2783

Respected Sir,

We are pleased to offer our best prices for the below mentioned items as under:


S.R.#	Description	Qty.	Unit Price/ Rs.
1	Mixer Block Assembly	1	210,000.00
2	Service and Calibration		10,000.00
<b>Total Charges</b>			<b>220,000.00</b>

## TERMS AND CONDITIONS


VALIDITY: 30 Days  
PAYMENT: At the time of delivery.  
Delivery: After receiving the work order  
Earnest Money: Pay Order Enclosed.  
Amount: 4400.00  
Pay Order No: 10831525  
Warranty: 1 year

Please feel free to contact us for any information / assistance that you may require in respect of subject offer.

FOR NOOR INTERNATIONAL

  
Service Department



  
31/12/15



# ZAM TRADING

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Ultrasound Machine

Dear Sir,  
We here by submit our estimated price for the repair item.

### Medical ICU

S.No.06. Ultrasound Machine (01No)  
Model: HS2000  
Make: Honda

### Work Require as follows:

- ❖ Replacement need of convex probe.
- ❖ Complete Servicing & Calibration.

1No. Rs. 158,000/=

Rupees: One Lac Fiftv Eight Thousand Only.

### TERMS & CONDITIONS:

VALIDITY : 3 Weeks  
WARRANTY : Six Months only for those faults which are mentioned in our quotation.  
REPAIR PERIOD : 3-5 weeks or earlier.  
PAY ORDER : Rs. 316000/- Dated. 28 Dec 2015

101-10 799685

If the work awarded to us, we assure you our best cooperation and service.

Your Truly

ZAM TRADING

ZAM Trading

*Handwritten signature and date: 31/12/15*

# HEALTH TECH PAKISTAN

Sale & Service of Medical Equipment

Superintendent

Date: 30-12-15

Hospital

Subject: Repair Estimate of Ultrasound Machine installed at Medical ICU.

Description	Qty	Unit Price	Total Price
Repair of Ultrasound Machine Mod: HS2000 Make: Honda	01	162,000/-	162,000/-
<u>Work Details:</u> <ul style="list-style-type: none"><li>Replacement need of convex probe.</li><li>Repair need of filter board.</li></ul>			
Total Rs: One Lac Sixty Two Thousand Rupees Only.			162,000/-

## TERMS & CONDITIONS:

VALIDITY : 02 Weeks.

WARRANTY : Six months only for those faults which are mentioned  
in our quotation and service report.

REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 10799680 Date: 28-12-15 Amount Rs: 3240/-

Your Faithfully

Health Tech Pakistan





ANNEXURE-II

Tender Fee Rs. 500/-  
(Non-Refundable)

STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE &  
REPAIR OF THE MACHINERY EQUIPMENT AT CIVIL HOSPITAL  
KARACHI DURING THE FINANCIAL YEAR 2015-16

Date of Submission & Time: - 31-12-2015 @ 11:30 a.m

Date of Opening & Time: - 31-12-2015 @ 12:00 noon

	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325R Make:- Dongmun	01 No.	0.112	
2	Repair and Service of Color Doppler Machine at Radiology Department. Model:- Imageic Maestro Make:- Kontron / France S. No. 0980472	01 No.	0.315	
3	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold way	04 Nos.	0.222	
4	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	0.516	
5	Repair and Service of Ventilator at Medical ICU. Model:- C2 Make:- Hamilton Sr. No. 2783 Mixer Block Assembly Service and Calibration	01 No.	0.220	
6	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	0.158	162,000/-
7	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	0.165	171,000/-
8	Repair and Service of Convex Probe of Toshiba NEMIO/ XG at Radiology Department. Convex Probe PVM-375AT	01 No.	0.714	
9	Repair and Service of SYSC-PWB of Toshiba NEMIO/ 17 at Radiology Department. SYSC-PWB	01 No.	0.684	
10	Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold way	03 Nos.	0.186	189,000/-
11	Repair of Baby Incubator at Peads Unit-II Model:- YP-90A Make:- Ningbo David	06 Nos.	0.300	315,000/-
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	0.735	




M.F.  
Handwritten signatures and initials.

Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
Maintenance and Installation of Liquid at M.R.I Department.	01 No.	0.715	
Supply of Multipara Monitors at Medical ICU. - OMNI-III - Infinium Medical	10 Nos.	0.580	595,000/-
Installation of Air Compressor System for Neurology Unit.			
Supply of Air Compressor System for Neurology Ward	01 No.		
Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.	02 Nos.		
Refrigerated Air Dryers Model:- LD-10'HA (PUMED / TAIWAN) or equivalent.	02 Nos.	0.665	
Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.		
After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.		
Electric Panel for above system	01 No.		
Water Separator	01 No.		
Total Estimated Cost		6.287	

Terms & Conditions:

1. Warranty period for repair work must be given with offer if the repaired equipment is out of order during the warranty period the same will be repaired without charging any cost by the concerned firm.
2. If the offer / quoted amount value is more than Rs. 25,000/- 2% Security deposit in shape of pay order in favor of Medical Superintendent Civil Hospital, Karachi, should be attached with offer the same will be refunded after 03 Months. If not received any complain of the repaired equipment.
3. The bid shall be opened on the above said date & time in the presence of the Hospital Procurement Committee (For Repair). The offer will be evaluated on best evaluated basis.



Handwritten signatures and initials, including 'AMS' and '3/12/18'.



# BIOMATF


Manufacturer, Sales & Service o

## ESTIMATE

**CUSTOMER:**  
MEDICAL SUPERINTENDENT  
CIVIL HOSPITAL  
KARACHI.



**DATE:** 26 December 2015

Item No.	Qty	Description	Unit Price	Total Price
		<b><u>Medical ICU</u></b>		
6	01	Ultrasound Machine Model: HS5000 Make: Honda  <b><u>Work Require:</u></b> Replacement of Convex probe. Repair of power supply. Over hauling and service.	Rs.165,000/=	Rs.165,000/=
TOTAL AMOUNT:-One Lac Sixty Five Thousand Only.				Rs.165,000/=

  
 For BIOMATRIX PAKISTAN  
**BIO-MATRIX PAKISTAN**

### TERMS & CONDITIONS

- Payment 100% after Completion of work.
- Estimate valid up to 30 days
- Warranty period 06 moths.
- Pay Order : 10799674 Date: 26/12/15 Rs: 3310/-

  
  
 31/12/15



CARDIOVIEW



Eternity Electronic Technology Co. Ltd



ST. FRANCIS TAIPEI TAIWAN R.O.C

BMP SPO2 & ECG

EDHNB

Analytical Instrument Inc. Advanced Instruments Inc. USA



**ANNEXURE-II**

Tender Fee Rs. 500/-  
(Non-Refundable)

**STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE &  
REPAIR OF THE MACHINERY EQUIPMENT AT CIVIL HOSPITAL  
KARACHI DURING THE FINANCIAL YEAR 2015-16**

Date of Submission & Time: - 31-12-2015 @ 11:30 a.m

Date of Opening & Time: - 31-12-2015 @ 12:00 noon

S. No.	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
1	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325R Make:- Dongmun	01 No.	0.112	
2	Repair and Service of Color Doppler Machine at Radiology Department. Model:- Imageic Maestro Make:- Kontron / France S. No. 0980472	01 No.	0.315	
3	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold way	04 Nos.	0.222	
4	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	0.516	
5	Repair and Service of Ventilator at Medical ICU. Model: - C2 Make:- Hamilton Sr. No. 2783 Mixer Block Assembly Service and Calibration	01 No.	0.220	
6	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	0.158	165,000/-
7	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	0.165	174,900/-
8	Repair and Service of Convex Probe of Toshiba NEMIO/XG at Radiology Department. Convex Probe PVM-375AT	01 No.	0.714	/
9	Repair and Service of SYSC-PWB of Toshiba NEMIO/17 at Radiology Department. SYSC-PWB	01 No.	0.684	/
10	Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold way	03 Nos.	0.186	195,000/-
11	Repair of Baby Incubator at Peads Unit-II Model:- YP-90A Make:- Ningbo David	06 Nos.	0.300	304,800/-
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	0.735	

BIO-MATRIX PAKISTAN

S. No.	Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
13	Repair, Maintenance and Installation of Liquid Chiller at M.R.I Department.	01 No.	0.715	
14	Repair of Multipara Monitors at Medical ICU. Model:- OMNI-III Make:- Infinium Medical	10 Nos.	0.580	500,000/-
15	<p>Installation of Air Compressor System for Neurology Unit.</p> <p>Supply of Air Compressor System for Neurology Ward</p> <p>Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.</p> <p>Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.</p> <p>Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.</p> <p>After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.</p> <p>Electric Panel for above system</p> <p>Water Separator</p>	<p>01 No.</p> <p>02 Nos.</p> <p>02 Nos.</p> <p>02 Nos.</p> <p>02 Nos.</p> <p>01 No.</p> <p>01 No.</p>	0.665	
<b>Total Estimated Cost</b>			<b>6.287</b>	

**Terms & Conditions:**

1. Warranty period for repair work must be given with offer if the repaired equipment is out of order during the warranty period the same will be repaired without charging any cost by the concerned firm.
2. If the offer / quoted amount value is more than Rs. 25,000/- 2% Security deposit in shape of pay order in favor of Medical Superintendent Civil Hospital, Karachi, should be attached with offer the same will be refunded after 03 Months. If not received any complain of the repaired equipment.
3. The bid shall be opened on the above said date & time in the presence of the Hospital Procurement Committee (For Repair). The offer will be evaluated on best evaluated basis.

**BIO-MATRIX PAKISTAN**



# ZAM TRADING

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Multipara Monitors

Dear Sir,  
We here by submit our estimated price for the repair item.

Medical Ward III

S.No.07. Multipara Monitors (3Nos)  
Model: DST5000  
Make: DST/USA

Work Require as follows:

- ❖ Replacement need of complete accessories.
- ❖ Repair need of NIBP module.
- ❖ Complete Servicing & Calibration.

Each Rs. 55,000/=

Total Rs: 165,000/= (3Nos)

Rupees: One Lac Sixty Five Thousand Only.

TERMS & CONDITIONS:

VALIDITY : 3 Weeks  
WARRANTY : Six Months only for those faults which are mentioned in our quotation.  
REPAIR PERIOD : 3-5 weeks or earlier.  
PAY ORDER : Rs. 3,300/- Dated. 28-12-15 No:- 10799681

If the work awarded to us, we assure you our best cooperation and service.

Yours Truly

ZAM TRADING  
ZAM Trading

*Handwritten signature and date: 31/12/15*

# HEALTH TECH PAKISTAN

Sale & Service of Medical Equipment

Superintendent

Date: 30-12-15

Repair Estimate of Multipara Monitor installed at Medical Unit III.

Description	Qty	Unit Price	Total Price
Repair of Multipara Monitor Mod: DST5000 Make: DST / USA			
07 <u>Work Details:</u> <ul style="list-style-type: none"><li>• Replacement need of NIBP board.</li><li>• Replacement need of accessories.</li><li>• Repair need of filter board.</li></ul>	03	57,000/-	171,000/-
Total Rs: One Lac Seventy One Thousand Rupees Only.			171,000/-

## TERMS & CONDITIONS:

VALIDITY : 02 Weeks.

WARRANTY : Six months only for those faults which are mentioned  
in our quotation and service report.

REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 10799679 Date: 28-12-15 Amount Rs: 3420/-

*[Handwritten signature and date: 31/12/15]*

Your Faithfully





# BIOMATF

Manufacturer, Sales & Service of Medical Equipments, Consumables & Disposables

NTN # 2815734-6

## ESTIMATE

**CUSTOMER:**  
MEDICAL SUPERINTENDENT  
CIVIL HOSPITAL  
KARACHI.

**DATE:** 26 December 2015

Item No.	Qty	Description	Unit Price	Total Price
		<b><u>Medical Unit III</u></b>		
7	03	Multipara Monitor Model: UT4000 Make: Goldway  <b><u>Work Require:</u></b> Replacement of accessories. Repair of power supply. Replacement of NIBP pump. Over hauling and service.	Rs.58,300/=	Rs.174,900/=
TOTAL AMOUNT:-One Lac Seventy Four Thousand Nine Hundred Only.				Rs.174,900/=

*Handwritten signature and date: 31/12/15*

*Handwritten signature and date: 31/12/15*  
For BIOMATRIX PAKISTAN  
BIO-MATRIX PAKISTAN

### TERMS & CONDITIONS

- Payment 100% after Completion of work.
- Estimate valid up to 30 days
- Warranty period 06 moths.
- Pay Order : 107 99673 Date: 3, 5007 Rs: 28-12-2015



CARDIOVIEW



Eternity Electronic Technology Co. Ltd




ST. FRANCIS  
TAIPEI TAIWAN R.O.C

BMP  
SP02 & ECG



Analytical Industries Inc.  
Advanced Instruments Inc. USA



# ZAM TRADING

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Multipara Monitor

Dear Sir,  
We here by submit our estimated price for the repair item.

Paeds Ward II

S.No.10. Multipara monitors (03Nos)  
Model: UT4000  
Make: Goldway

Work Require as follows:

- ❖ Replacement need of accessories.
- ❖ Repair need of main board.
- ❖ Replacement need of battery pack.
- ❖ Complete Servicing & Calibration.

1No. Rs. 62,000/=

Total Rs. 186,000/- (3Nos)

Rupees: One Lac Eighty Six Thousand Only.

TERMS & CONDITIONS:

VALIDITY : 3 Weeks  
WARRANTY : Six Months only for those faults which are mentioned in our quotation.

REPAIR PERIOD : 3-5 weeks or earlier.

PAY ORDER : Rs. 317201/- Dated. 28-12-15 No: 10799684

If the work awarded to us, we assure you our best cooperation and service.

Your Truly

ZAM Trading

*[Handwritten signatures and dates]*  
31/12/15  
31/12/15  
31/12/15

# HEALTH TECH PAKISTAN

Sale & Service of Medical Equipment  
tendent

Date: 30-12-15

## Repair Estimate of Multipara Monitor installed at Paeds Unit II.

Description*	Qty	Unit Price	Total Price
Repair of Multipara Monitor Mod: UT4000 Make: Goldway			
<u>Work Details:</u> <ul style="list-style-type: none"><li>• Repair need of main processor board.</li><li>• Replacement need of battery pack.</li><li>• Replacement need of accessories.</li><li>• Repair need of filter board.</li></ul>	03	63,000/-	189,000/-
Total Rs: One Lac Eighty Nine Thousand Rupees Only.			189,000/-

### TERMS & CONDITIONS:

VALIDITY : 02 Weeks.

WARRANTY : Six months only for those faults which are mentioned  
in our quotation and service report.

REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 10799678 Date: 28-12-15 Amount Rs: 37801/-

*Handwritten signature and date: 31/12/15*

Your Faithfully





# BIOMATRIX

Manufacturer, Sales & Service o

NIN 11 -

## ESTIMATE

**CUSTOMER:**  
MEDICAL SUPERINTENDENT  
CIVIL HOSPITAL  
KARACHI.

**DATE:** 26 December 2015

Item No.	Qty	Description	Unit Price	Total Price
		<b><u>Peads Unit II</u></b>		
10	03	Multipara Monitor Model: UT4000 Make: Goldway  <b><u>Work Require:</u></b> Repair of mother board. Repair of power supply. Replacement of accessories. Replacement of battery pack. Over hauling and service.	Rs.65,000/=	Rs.,195,000/=
TOTAL AMOUNT:-One Lac Ninety Five Thousand Only.				Rs.195,000/=

AS  
21/12/15

For BIOMATRIX PAKISTAN  
**BIO-MATRIX PAKISTAN**

### TERMS & CONDITIONS

- Payment 100% after Completion of work.
- Estimate valid up to 30 days
- Warranty period 06 moths.

• Pay Order: 10799670 Date: 28/12/15 Rs: 3,900/=



CARDIOVIEW



Eternity Electronic Technology Co. Ltd



ST. FRANCIS  
TAIPEI TAIWAN R.O.C

BMP  
SPO2 & ECG

TECHNO

Analytical Industries Inc.  
Advanced Instruments Inc. USA





# ZAM

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Baby Incubator

Dear Sir,  
We here by submit our estimated price for the repair item.

Paeds Ward II

S.No.11. Baby Incubtor (06Nos)  
Model: YP-90A  
Make: Ningbo David

Work Require as follows:

- ❖ Replacement need of Heater.
- ❖ Repair need of main board.
- ❖ Complete Servicing & Calibration.

1No. Rs. 50,000/=

Total Rs. 300,000/- (6Nos)

Rupees: Three Lac Only.

TERMS & CONDITIONS:

VALIDITY : 3 Weeks

WARRANTY : Six Months only for those faults which are mentioned in our quotation.

REPAIR PERIOD : 3-5 weeks or earlier.

PAY ORDER : Rs. 60000/= Dated. 28-12-15 No. 16799683

If the work awarded to us, we assure you our best cooperation and service.

Your Truly

**ZAM TRADING**  
ZAM Trading

*Handwritten signatures and dates:*  
31/12/15  
31/12/15



# BIOMATRIX

Manufacturer, Sales & Service

## ESTIMATE

**CUSTOMER:**  
MEDICAL SUPERINTENDENT  
CIVIL HOSPITAL  
KARACHI.

**DATE:** 26 December 2015

Item No.	Qty	Description	Unit Price	Total Price
		<b><u>Peads Unit II</u></b>		
11	06	Baby Incubator Model: YP90-A Make: Ningbo David  <b><u>Work Require:</u></b> Repair of power supply. Replacement of heater. Repair of power supply. Over hauling and service.	Rs.50,800/=	Rs.,304,800/=
TOTAL AMOUNT:-Three Lac Four Thousand Eight Hundred Only.				Rs.304,800/=

*Handwritten signature and date: 21/12/15*

*Handwritten signature*

*Handwritten signature and date: 31/12/15*

For BIOMATRIX PAKISTAN  
**BIO-MATRIX PAKISTAN**

### TERMS & CONDITIONS

- Payment 100% after Completion of work.
- Estimate valid up to 30 days
- Warranty period 06 moths.
- Pay Order : 10799669 Date: 28-12-15 Rs: 6:100/=



CARDIOVIEW



Eternity Electronic Technology Co. Ltd



ST. FRANCIS  
TAIPEI TAIWAN R.O.C

BMP  
SPO2 & ECG

TECHNO

Analytical Instrument Inc.  
Advanced Instrument's Inc. USA

# HEALTH TECH PAKISTAN

Sale & Service of Medical Equipment  
tendent

Date:30-12-15

## Repair Estimate of baby Incubator installed at Paeds Unit II.

Description	Qty	Unit Price	Total Price
Repair of baby Incubator Mod: YO-90A Make: Ningbo David			
<u>Work Details:</u> <ul style="list-style-type: none"><li>• Repair need of heater assembly.</li><li>• Repair need of controller pcb.</li><li>• Replacement need of temperature controller.</li></ul>	06	52,500/-	315,000/-
Total Rs: Three Lac Fifteen Thousand Rupees Only.			315,000/-

### TERMS & CONDITIONS:

VALIDITY : 02 Weeks.

WARRANTY : Six months only for those faults which are mentioned  
in our quotation and service report.

REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 10799677 Date: 28-12-15 Amount Rs: 6300/-

*Handwritten signature and date: 31/12/15*

Your Faithfully

*Handwritten signature*  
Health Tech Pakistan





# ZAM TRADING

Room No.82 3rd Floor, National Chamber Near Hamdard Dawakhana Arambagh Karachi. Mob:0312-2914587

Date: 28<sup>th</sup> DEC-2015

Medical Superintendent,  
Civil Hospital,  
Karachi.

SUB: Quotation For Repair of Multipara Monitor

Dear Sir,  
We here by submit our estimated price for the repair item.

Medical ICU

S.No.14. Multipara monitors (10Nos)  
Model: OMNI III  
Make: Infinium Medical

Work Require as follows:

- ❖ Replacement need of accessories.
- ❖ Repair need of Nibp module.
- ❖ Replacement need of battery pack.
- ❖ Complete Servicing & Calibration.

1No. Rs. 58,000/=

Total Rs. 580,000/- (10Nos)

Rupees: Five Lac Eighty Thousand Only.

TERMS & CONDITIONS:


VALIDITY : 3 Weeks  
WARRANTY : Six Months only for those faults which are mentioned in our quotation.  
REPAIR PERIOD : 3-5 weeks or earlier.  
PAY ORDER : Rs. 11,600/- Dated. 28-12-15 NO:- 107 99682


If the work awarded to us, we assure you our best cooperation and service.

Your Truly

  
~~ZAM TRADING~~  
ZAM Trading







# BIOMATRIX

Manufacturer, Sales & Service

NTN # 2815734-6

## ESTIMATE

**CUSTOMER:**  
**MEDICAL SUPERINTENDENT**  
**CIVIL HOSPITAL**  
**KARACHI.**

**DATE: 26 December 2015**

Item No.	Qty	Description	Unit Price	Total Price
		<b><u>Medical ICU</u></b>		
14	10	Multipara Monitor Model: OMNI III Make: Infinium Medical  <b><u>Work Require:</u></b> Repair of power supply. Replacement NIBP boards. Repair of power supply. Replacement of accessories. Over hauling and service.	Rs.59,000/=	Rs.590,000/=
TOTAL AMOUNT:- Five Lac Ninety Thousand Only.				Rs.590,000/=

*Handwritten signature*

*Handwritten signature*

For BIOMATRIX PAKISTAN  
 BIO-MATRIX PAKISTAN

### TERMS & CONDITIONS

- Payment 100% after Completion of work.
- Estimate valid up to 30 days
- Warranty period 06 moths.
- Pay Order: 10799671 Date: 28-12-2015

RS: 11,200/-



CARDIOVIEW



Eternity Electronic Technology Co. Ltd



ST. FRANCIS  
TAIPEI TAIWAN R.O.C



SPO2 & ECG



# HEALTH TECH PAKISTAN

Sale & Service of Medical Equipment

Superintendent

Date: 30-12-15

## Repair Estimate of Multipara Monitor installed at Medical ICU.

Description	Qty	Unit Price	Total Price
Repair of Multipara Monitors Mod: OMNI III Make: Infinium Medical			
14 <u>Work Details:</u> <ul style="list-style-type: none"><li>• Repair need of ERT board.</li><li>• Repair need of NIBP modules.</li><li>• Replacement need of accessories,</li></ul>	10	59,500/-	595,000/-
Total Rs: Five Lac Ninety Five Thousand Rupees Only. <u>IC</u>			595,000/-

### TERMS & CONDITIONS:


VALIDITY : 02 Weeks.

WARRANTY : Six months only for those faults which are mentioned  
in our quotation and service report.

REPAIR PERIOD : 3-5 weeks or earlier.

Pay Order : No. 10799676 Date: 28-12-15 Amount Rs: 11900/-

*[Handwritten signature and date 31/12/15]*





# STANDARD SUPPLIERS

Supplier of Medical Equipments/Medical Pipeline System/Hospital Furniture/Medicines



OUR REF: - SS/QT - 7403/2015  
DATED: - 28/12/2015

Medical Superintendent,  
Civil hospital  
Karachi

Dear Sir,

STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE & REPAIR OF THE  
MACHINERY EQUIPMENTS AT CIVIL HOSPITAL KARACHI DURING THE FINANCIAL  
YEAR 2015 - 2016  
DUE ON: - 31 - 12 - 2015

Reference your above tender, we are pleased to quote as under: -

<u>S.NO.</u>	<u>DESCRIPTION</u>	<u>QTY</u>	<u>AMOUNT</u>
16.	Installation of Air Compressor System for Neurology Ward.	01 No.	Rs. 660.000/=

AIR COMPRESSOR SYSTEM

Comprising of: -

- |    |   |                  |
|----|---|------------------|
| a) | Air Compressor System 5 HP<br><u>Model - PK 50160</u><br>Repairing and complete overhauling<br>(PUMED / TAIWAN) | 02 No.<br>01 No. |
| b) | Refrigerated Air Dryers<br><u>Model - LD - 1011A</u><br>(PUMED / TAIWAN)  | 02 Nos.          |
| c) | Pre - Air Filter<br><u>Model - TLQ - 00065</u><br>(PUMED / TAIWAN)  | 02 Nos.          |
| d) | After Air Filter<br><u>Model - TLP - 00065</u><br>(PUMED / TAIWAN)  | 02 Nos.          |



**ANNEXURE-II**

**Tender Fee Rs. 500/-**  
(Non-Refundable)


**STANDARD BIDDING DOCUMENTS (SBD) FOR MAINTENANCE & REPAIR OF THE MACHINERY EQUIPMENT AT CIVIL HOSPITAL KARACHI DURING THE FINANCIAL YEAR 2015-16**

Date of Submission & Time: - 31-12-2015 @ 11:30 a.m  
Date of Opening & Time: - 31-12-2015 @ 12:00 noon

Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325R Make:- Dongmun	01 No.	0.11	N/A
Repair and Service of Color Doppler Machine at Radiology Department. Model:- Imageic Maestro Make:- Kontron / France S. No. 0980472	01 No.	0.31	N/A
Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold way	04 Nos.	0.24	N/A
Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	0.51	N/A
Repair and Service of Ventilator at Medical ICU. Model: - C2 Make:- Hamilton Sr. No. 2783 Mixer Block Assembly Service and Calibration	01 No.	0.22	N/A
Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	0.158	N/A
Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	0.165	N/A
Repair and Service of Convex Probe of Toshiba NEMIO/XG at Radiology Department. Convex Probe PVM-375AT	01 No.	0.714	N/A
Repair and Service of SYSC-PWB of Toshiba NEMIO/17 at Radiology Department. SYSC-PWB	01 No.	0.684	N/A
Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold way	03 Nos.	0.186	N/A
Repair of Baby Incubator at Peads Unit-II Model:- YP-90A Make:- Ningbo David	06 Nos.	0.300	N/A
2 Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	0.735	N/A





Description of Work	Qty.	Estimated Cost In (Million)	Amount Offered
Repair, Maintenance and Installation of Liquid Chiller at M.R.I Department.	01 No.	0.715	N/A
Repair of Multipara Monitors at Medical ICU. Model:- OMNI-III Make:- Infinium Medical	10 Nos.	0.580	N/A
Installation of Air Compressor System for Neurology Unit.			Rs. 660,000/-
Supply of Air Compressor System for Neurology Ward	01 No.		(Rupees Six Hundred & Sixty thousand Only). 
Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.	02 Nos.		
Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.	02 Nos.	0.665	
Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.		
After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.	02 Nos.		
Electric Panel for above system	01 No.		
Water Separator:	01 No.		
<b>Total Estimated Cost</b>		<b>6.287</b>	

**Terms & Conditions:**

1. Warranty period for repair work must be given with offer if the repaired equipment is out of order during the warranty period the same will be repaired without charging any cost by the concerned firm.
2. If the offer / quoted amount value is more than Rs. 25,000/- 2% Security deposit in shape of pay order in favor of Medical Superintendent Civil Hospital, Karachi, should be attached with offer the same will be refunded after 03 Months. If not received any complain of the repaired equipment.
3. The bid shall be opened on the above said date & time in the presence of the Hospital Procurement Committee (For Repair). The offer will be evaluated on best evaluated basis.






*Handwritten signature and text:*  
 31/12/2015  
 [Signature]



Date: 31-12-2015

Tender opening meeting held on 31-12-2015 at 12:00 p.m in the Committee Room of Civil Hospital Karachi under the Chairmanship of Addl: Medical Superintendent, Civil Hospital Karachi for "Repair of Machinery Equipments" at Civil Hospital Karachi.

The following members have attended the meeting:-

S. No.	Name of Officer(s)	Signature
1	Addl: Medical Superintendent (Technical), Civil Hospital Karachi	
2	Assistant Professor of Cardiology, Dow University of Health Science, Karachi	
3	Assistant Professor of Surgery, Dow University of Health Science, Karachi	
4	Addl: Medical Superintendent (General), Civil Hospital Karachi	
5	Accounts Officer, Civil Hospital Karachi	

Date: 31-12-2015

Scaled tenders were invited through the publication of Tender Notice published in the Daily "Dawn" dated 10<sup>th</sup> December' 2015 and hoisted on SPPRA website vide Sr. No. 26000 for the "Repair of Machinery Equipments" at Civil Hospital Karachi.

The Bids opened on 31-12-2015 at 12:00 p.m. in the Committee Room, Civil Hospital Karachi before the Procurement Committee of this Hospital.

The following firms have participated in the competition, the representative who were present at the time of opening tender their signatures were taken as a token of presence.

S. No.	Name of Firms	Signature of Representative
01.	M/S. STANDARD SUPPLIERS.	
✓ 02.	M/S. NOOR INTERNATIONAL	
✓ 03.	M/S. BIOS.	
04.	M/S. LIFE ONE.	
✓ 05.	M/S. ZANI TRADING.	
06.	M/S. BIOMATRIX PAK.	
07	M/S. HEALTH TEC PR	

Serial 02, 03, 05, 07, 08, 09, 10, 11, 12, 13, 14, 15, 16, 17, 18, 19, 20, 21, 22, 23, 24, 25, 26, 27, 28, 29, 30, 31, 32, 33, 34, 35, 36, 37, 38, 39, 40, 41, 42, 43, 44, 45, 46, 47, 48, 49, 50, 51, 52, 53, 54, 55, 56, 57, 58, 59, 60, 61, 62, 63, 64, 65, 66, 67, 68, 69, 70, 71, 72, 73, 74, 75, 76, 77, 78, 79, 80, 81, 82, 83, 84, 85, 86, 87, 88, 89, 90, 91, 92, 93, 94, 95, 96, 97, 98, 99, 100

31/12/2015  
Procurement



## CIVIL HOSPITAL KARACHI

NIT NO. INF/KRY-4085/15 SPPRA ID @ S. No. 26000

**Method and procedure of procurement: (Single Stage – One Envelope)**

### MINUTES OF BID OPENING MEETING

A meeting of the procurement committee of this department was held on 31-12-2015 for opening the bids received in respect of subject NIT for **Maintenance & Repair of Machinery Equipment at Civil Hospital Karachi 2015-16** till the deadlines of submission. The meeting was attended by following members of the procurement committee and the representative of bidders. (**Attendance sheet is enclosed**).

The following bidders submitted their bids till the deadline of submission:

- 1) M/s Standard Suppliers
- 2) M/s Noor International
- 3) M/s BIOS
- 4) M/s Life One
- 5) M/s Zam Trading Co.
- 6) M/s Biomatrix Pak
- 7) M/s Health Tech Pakistan

The bids were opened at 12:00 noon in the presence of the above mentioned participants and the rates quoted by bidders were read aloud and encircled by Chairman of the Procurement Committee. All the members of the procurement committee signed each and every page of financial proposal / bids. The bids do not contain any over-writing or cutting.

Following is the details of bids announced:

S. No.	Item name	Name of Bidders	Offered Price	Amount of Bid Security	Pay Order No. / Dated
1	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325 R Make:- Dongmun	M/s BIOS	Rs. 112,000/-	Rs. 2,240/-	1084000146 28-12-2015
		-	-	-	-
		-	-	-	-
2	Repair and Service of Color Doppler at Radiology Department. Model:- Imageic Maestro S. No. 0980472 Make:- Kontron / France	M/s BIOS	Rs. 315,000/-	Rs. 6,300/-	1084000147 28-12-2015
		-	-	-	-
		-	-	-	-

  
Neelam

S. No.	Item name	Name of Bidders	Offered Price	Amount of Bid Security	Pay Order No. / Dated
3	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold Way	M/s Zam Trading	Rs. 222,000/-	Rs. 4,440/-	50611632 28-12-2015
		M/s BIOS	Rs. 228,000/-	Rs. 4,560/-	1084000148 28-12-2015
		M/s Life One	Rs. 230,000/-	Rs. 4,600/-	50611630 28-12-2015
4	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	M/s Zam Trading	Rs. 516,000/-	Rs. 10,320/-	50611633 28-12-2015
		M/s Life One	Rs. 518,000/-	Rs. 10,360/-	50611631 28-12-2015
		M/s BIOS	Rs. 522,000/-	Rs. 10,440/-	1084000149 28-12-2015
5	Repair and Service of Ventilator at Medical ICU. Model:- C2 Make:- Hamilton S # 2783 Mixer Block Assembly Service and Calibration	M/s Noor International	Rs. 220,000/-	Rs. 4,400/-	10831525 23-12-2015
		-	-	-	-
		-	-	-	-
6	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	M/s Zam Trading	Rs. 158,000/-	Rs. 3,160/-	10799685 28-12-2015
		M/s Health Tech	Rs. 162,000/-	Rs. 3,240/-	10799680 28-12-2015
		M/s Biomatrix Pak	Rs. 165,000/-	Rs. 3,310/-	10799674 28-12-2015
7	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	M/s Zam Trading	Rs. 165,000/-	Rs. 3,300/-	10799681 28-12-2015
		M/s Health Tech	Rs. 171,000/-	Rs. 3,420/-	10799679 28-12-2015
		M/s Biomatrix Pak	Rs. 174,900/-	Rs. 3,500/-	10799673 28-12-2015
8	Repair and Service of Convex Probe of Toshiba NEIMO / XG at Radiology Department. Convex Probe PVM-375AT	NQ	NQ	NQ	NQ
9	Repair and Service of SYSC-PWB of Toshiba NEIMO / XG at Radiology Department. SYSC-PWB	NQ	NQ	NQ	NQ
10	Repair of Multipara Monitors at Peads Unit-II. Model:- UT4000 Make:- Gold Way	M/s Zam Trading	Rs. 186,000/-	Rs. 3,720/-	10799684 28-12-2015
		M/s Health Tech	Rs. 189,000/-	Rs. 3,780/-	10799678 28-12-2015
		M/s Biomatrix Pak	Rs. 195,000/-	Rs. 3,900/-	10799670 28-12-2015
11	Repair of Baby Incubator at Peads Unit-II. Model:- YP-90A Make:- Ningbo David	M/s Zam Trading	Rs. 300,000/-	Rs. 6,060/-	10799683 28-12-2015
		M/s Biomatrix Pak	Rs. 304,800/-	Rs. 6,100/-	10799669 28-12-2015
		M/s Health Tech	Rs. 315,000/-	Rs. 6,300/-	10799677 28-12-2015
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	NQ	NQ	NQ	NQ
13	Repair, Maintenance and Installation of Liquid Chiller at M.R.I Department.	NQ	NQ	NQ	NQ

Attesting


S. No.	Item name	Name of Bidders	Offered Price	Amount of Bid Security	Pay Order No. / Dated
14	<b>Repair of Multipara Monitors at Medical ICU.</b> Model:- OMNI-III Make:- Infinium Medical	M/s Zam Trading	Rs. 580,000/-	Rs. 11,600/-	10799682 28-12-2015
		M/s Biomatrix Pak	Rs. 590,000/-	Rs. 11,800/-	10799671 28-12-2015
		M/s Health Tech	Rs. 595,000/-	Rs. 11,900/-	10799676 28-12-2015
15	<b>Installation of Air Compressor System for Neurology Unit.</b>  Supply of Air Compressor System for Neurology Ward  Air Compressor System Comprising of: - Air Compressor System 5HP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.  Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.  Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.  After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.  Electric Panel for above system  Water Seprator	M/s Standard Suppliers	Rs. 660,000/-	Rs. 13,500/-	13021625 28-12-2015
		-	-	-	-
		-	-	-	-

The committee shall examine all the bids as per the Qualification/ eligibility criteria provided in the bidding documents, arithmetical checks and verify the documents and bid security submitted by the bidders.

The meeting ended with the note of thanks to and from the chair.



Neelam






## CIVIL HOSPITAL KARACHI

Statement showing the comparative rates for the "Maintenance & Repair of the Machinery Equipment at Civil Hospital Karachi during the financial year 2015-16 quoted by the various firms vides Tender Enquiry No. AMS (Tech) /-9988 dated 07-12-2015, published in the Daily "Dawn" dated 10-12-2015 and hoisted on SPRA website at S. No. 26000 technical / financial opened on 31-12-2015.

Item No.	Descriptions of Works	Qty	Name of Firm & Rate Quoted	Name of Firm & Rate Quoted	Name of Firm & Rate Quoted	Remarks
01	Repair and Service of Mobile X-Ray Machine at Medical ICU. Model:- 325 R Make:- Dongmun	01 No.	Rs. 112,000/- M/s BIOS	-	-	Qualified on Authorized Agent basis
02	Repair and Service of Color Doppler at Radiology Department. Model:- Imageic Maestro S. No. 0980472 Make:- Kontron / France	01 No.	Rs. 315,000/- M/s BIOS	-	-	Qualified on Authorized Agent basis
03	Repair of Multipara Monitors at E.O.T. Model:- OMNI & UT4000 Make:- Infinium Medical & Gold Way	04 Nos.	Rs. 222,000/- M/s Zam Trading	Rs. 228,000/- M/s BIOS	Rs. 230,000/- M/s Life One	On Lowest Basis
04	Repair of Multipara Monitors at Surgical ICU. Model:- OMNI-III Make:- Infinium Medical	08 Nos.	Rs. 516,000/- M/s Zam Trading	Rs. 518,000/- M/s Life One	Rs. 522,000/- M/s BIOS	On Lowest Basis
05	Repair and Service of Ventilator at Medical ICU. Model:- C2 Make:- Hamilton S # 2783 Mixer Block Assembly Service and Calibration	01 No.	Rs. 220,000/- M/s Noor International	-	-	Qualified on Authorized Agent basis
06	Repair of Ultrasound Machine at Medical ICU. Model:- HS2000 Make:- Honda	01 No.	Rs. 158,000/- M/s Zam Trading	Rs. 162,000/- M/s Health Tech	Rs. 165,000/- M/s Biomatrix Pak	On Lowest Basis
07	Repair of Multipara Monitors at Medical Unit-III. Model:- DST5000 Make:- DST / USA	03 Nos.	Rs. 165,000/- M/s Zam Trading	Rs. 171,000/- M/s Health Tech	Rs. 174,900/- M/s Biomatrix Pak	On Lowest Basis
08	Repair and Service of Convex Probe of Toshiba NEIMO / XG at Radiology Department. Convex Probe PVM-375AT	01 No.	NQ	NQ	NQ	NQ

Nedra

Item No.	Descriptions of Works	Qty	Name of Firm & Rate Quoted	Name of Firm & Rate Quoted	Name of Firm & Rate Quoted	Remarks
09	Repair and Service of SYSC-PWB of Toshiba NEIMO / XG at Radiology Department. SYSC-PWB	01 No.	NQ	NQ	NQ	NQ
10	Repair of Multipara Monitors at Peds Unit-II. Model:- UT4000 Make:- Gold Way	03 Nos.	Rs. 186,000/- M/s Zam Trading	Rs. 189,000/- M/s Health Tech	Rs. 195,000/- M/s Biomatrix Pak	On Lowest Basis
11	Repair of Baby Incubator at Peds Unit-II. Model:- YP-90A Make:- Ningbo David	06 Nos.	Rs. 300,000/- M/s Zam Trading	Rs. 304,800/- M/s Biomatrix Pak	Rs. 315,000/- M/s Health Tech	On Lowest Basis
12	Repair, Maintenance and Installation of the Chilled Water Pipes with related Accessories at M.R.I. Department.	01 No.	NQ	NQ	NQ	NQ
13	Repair, Maintenance and Installation of Liquid Chiller at M.R.I Department.	01 No.	NQ	NQ	NQ	NQ
14	Repair of Multipara Monitors at Medical ICU. Model:- OMNI-III Make:- Infinium Medical	10 Nos.	Rs. 580,000/- M/s Zam Trading	Rs. 590,000/- M/s Biomatrix Pak	Rs. 595,000/- M/s Health Tech	On Lowest Basis








Nedra





Item No.	Descriptions of Works	Qty	Name of Firm & Rate Quoted	Name of Firm & Rate Quoted	Name of Firm & Rate Quoted	Remarks
15	<p>Installation of Air Compressor System for Neurology Unit.</p> <p>Supply of Air Compressor System for Neurology Ward</p> <p>Air Compressor System Comprising of: - Air Compressor System SHP Model: - PK-50160 Repairing and Complete with overhauling (PUMED / TAIWAN) or equivalent.</p> <p>Refrigerated Air Dryers Model:- LD-10 HA (PUMED / TAIWAN) or equivalent.</p> <p>Pre-Air Filters Model:- TLQ-00065 (PUMED / TAIWAN) or equivalent.</p> <p>After Air Filters Model:- TLP-00065 (PUMED / TAIWAN) or equivalent.</p> <p>Electric Panel for above system</p> <p>Water Separator</p>	<p>01 No.</p> <p>02 Nos.</p> <p>02 Nos.</p> <p>01 No.</p> <p>01 No.</p>	<p>Rs. 660,000</p> <p>M/s Standard Suppliers</p>	-	-	<p>Qualified on Authorized Agent basis</p>

  
Accounts Officer  
Civil Hospital, Karachi  
Sethi D Ali Khattar

  
Addl: Medical Superintendent (General)  
Civil Hospital Karachi  
Mr. Ahmad Ali Khan

  
Asst. Prof. of Cardiology  
Dow University of Health  
Sciences Karachi  
M. Nadeem

  
Asst. Prof. of Surgery  
Dow University of Health  
Sciences Karachi  
Dr. Sidra Hashmi

  
Addl: Medical Superintendent  
(Technical)  
Civil Hospital Karachi  
Dr. Azeem Khan