

No.LUH/Gen-Br/- 22571/72

Office of the
Medical Superintendent
Liaquat University Hospital
Hyderabad/Jamshoro
Dated the 16-5-2018


To,

The Managing Director,
Government of Sindh,
Sindh Public Procurement Regulatory Authority,
Karachi.

SUBJECT:- COMPARATIVE STATEMENT AND MINUTES OF MEETING.

Reference:- SPPRA No.36633, 36634, 37150 & I.D No. 2147483647


I have the honour to enclose herewith the Technical Evaluation Report, Bid Evaluation Report, Comparative Statement and Minutes of the meeting for purchase of Machinery / Equipment of Tender No. 34, 35 and 41 for the financial year 2017-18 for Liaquat University Hospital, Hyderabad / Jamshoro to paste on www.pprasindh.gov.pk.


Medical Superintendent
Liaquat University Hospital
Hyderabad/Jamshoro
Dated the 17-5-2018.


No.LUH/Gen-Br/-

Copy forwarded for information to :-

1. The P.S to Secretary, Government of Sindh, Health Department, Karachi.
2. General File.


Medical Superintendent
Liaquat University Hospital
Hyderabad/Jamshoro

SPPRA INWARD DIARY
NO : 9169
DATED 17-05-2018



36653

**LIAQUAT UNIVERSITY HOSPITAL
HYDERABAD / JAMSHORO**

**BID
EVALUATION REPORT**

**PURCHASE OF MACHINERY /EQUIPMENT UNDER SNE /
ADP FOR THE YEAR 2017-18
(RESCUE CENTRE) (UROLOGY DEPARTMENT) (PART-1)**

TENDER INQUIRY No.34/2017-18

OPENED ON 27.02.2018



**LIAQUAT UNIVERSITY HOSPITAL
HYDERABAD / JAMSHORO**

**BID
EVALUATION REPORT**

**PURCHASE OF MACHINERY /EQUIPMENT UNDER SNE /
ADP FOR THE YEAR 2017-18
(RESCUE CENTRE) (UROLOGY DEPARTMENT) (PART-1)**

TENDER INQUIRY No.34/2017-18

OPENED ON 27.02.2018

S.No.	Description of Procurement.	Qty	Estimated Total cost	Funds Allocated	Source of funds (ADP / Non ADP)	Proposed Procurement method	Timing of Procurement				Remarks	
							1st Qtr	2nd Qtr	3rd Qtr	4th Qtr		
1	Drugs / Medicines & Surgical Items (L.P)		87.75 M	87.75 M	Non ADP	Single Stage Envelope	In all quarters as per Requirement					
2	Dietary Items.		40.95 M	40.95 M	---do---	---do---						
3	Uniform & Liveries		3.33 M	3.33 M	---do---	---do---						
4	Misc / Petty / Electric Items.		4.389 M	4.389 M	---do---	---do---						
5	Medical Gases		29.23 M	29.23 M	---do---	---do---						
6	Janitorial Services & Items.		9.45 M	9.45 M	---do---	---do---						
7	Printing Articles.		1.016 M	1.016 M	---do---	---do---						
8	Stationary Articles		0.604 M	0.604 M	---do---	---do---						
9	Instruments		35.464 M	35.464 M	---do---	---do---						
10	Consumable Items / Pathology Kits & Chemicals.		12.183 M	12.183 M	---do---	---do---						
11	Deployment of Security Guards		13.2 M	13.2 M								
12	Repair of Machinery / Equipment		37.077 M	37.077 M	---do---	---do---						
13	Repair of Furniture / Fixture.		2.04 M	2.04 M	---do---	---do---						
14	Drugs/Medicine under Zakat fund (Bulk)		42.0 M	42.0 M	---do---	---do---						
15	Drugs/Medicine under Zakat fund (LP)		28.0 M		---do---	---do---						
16	Repair of Transport		4.692 M	4.692 M	---do---	---do---						
17	Machinery Equipment under SNE		374.14 M	374.14 M	---do---	---do---						
I	Purchase of C. T Scan Machine		70.0 M		---do---	---do---						
II	Purchase of Dialysis Machine		6.00 M		---do---	---do---						
III	Purchase of Lithotripsy Machine		70.0 M		---do---	---do---						
IV	Purchase of M.R.I Machine		125.0 M		---do---	---do---						
V	Purchase of Laundry Machine		4.00 M		---do---	---do---						
18	Purchase of Machiner / Equipment for Paeds		99.140 M		---do---	---do---						
19	Purchase of Machinery from Welfare fund		5.00 M	5.00 M	---do---	---do---						
20	Furniture / Fixture under SNE		58.925 M	58.925 M	---do---	---do---						
21	Purchase of Transport		39.0 M	39.0 M	---do---	---do---						
22	Purchase of Machinery /Equipment under various ADP/Non ADP Scheme & Rescue Center.		400.00 M		ADP	---do---						

List Enclosed



NOTIFICATION

NO.HD (P&E) 1-2 (61) / 2017. In supersession of this department's notification No.HD (P&E) 3-5 (77) / 2017, dated 16.03.2017, NO.HD (P&E) 3-5 (74) / 2013, dated 03.08.2017, NO.HD (P&E) 3-5 (85) / 2017, dated 03.08.2017, NO.HD (P&E) 3-5 (78) / 2013, dated 11.09.2017, NO.HD (P&E) 3-5 (55) / 2015, dated 11.09.2017 and SO (PM&I) 2-1 / 2017-18 / PM-05 (LUH), dated 07.08.2017, the Government of Sindh, Health Department is pleased to re constitute a Procurement Committee under Clause-7 of Sindh Public Procurement Rules 2010 (Amended 2017), comprising of the following members for purchase of machinery / equipment, vehicles and furniture / fixture to be purchased under revenue component of "ADP schemes & SNE" in respect of Liaquat University Hospital Hyderabad;

- | | |
|---|-----------------|
| 1. Additional Medical Superintendent (General),
Liaquat University Hospital, Hyderabad. | Chairman |
| 2. Additional Medical Superintendent (Technical)
Liaquat University Hospital, Hyderabad. | Member |
| 3. Professor / Associate Professor / Assistant
Professor / Expert / Technical Person of Relevant field,
Liaquat University of Medical & Health Sciences (LUM&HS)
Jamshoro. | Member |

The committee will responsible to observe all the codal formalities as per Sindh Public Procurement Rules 2010 (Amended 2017), the TORS / Functions / Responsibilities of procurement committee in accordance with SPP Rule-8 (Amended 2017) are as under;

- Preparing bidding documents;
- Carrying out technical as well as financial evaluation of the bids
- Preparing evaluation report as provided in rule-45,
- Making recommendations for the award of contract to the competent authority;
- Perform any other function ancillary and incidental to the above.

(DR.FAZLULLAH PECHUHO)
SECRETARY HEALTH

NO.HD (P&E) 1-2 (61) / 2017

Karachi, dated 07-11-2017.

C.c.to:

- The Vice Chancellor, Liaquat University of Medical and Health Sciences Hyderabad / Jamshoro with request to advice the member to attend the meetings for above said committee.
- The Managing Director, Sindh Public Procurement Regularity Authority, Karachi.
- The Medical Superintendent, Liaquat University Hospital Hyderabad / Jamshoro.
- The District Accounts Officer, Hyderabad.
- All members of the committee.
- P.S. to Secretary, Health Department, Govt. of Sindh, Karachi.
- P.S to Additional Secretary (Dev), Health Department, Govt. of Sindh, Karachi.


Deputy Director (Dev)



C. 814

Clean Br.



NO.HD(P&E) 3-2 (427)/2014
GOVERNMENT OF SINDH
HEALTH DEPARTMENT
 (Procurement, Monitoring and Inspection Cell)
 Karachi dated the 10th APRIL, 2017.

NOTIFICATION

No.HD(P&E)3-2(427)/2014. In supersession to this department's notification of even number dated 9-03-2017 and in pursuance of Rule 31 of the Sindh Public Procurement Rules, 2010, a Departmental Complaint Redressal Committee, comprising of following officers is hereby re-constituted as under to resolve complaints of aggrieved bidders:

- | | | |
|---|--------------|------------------|
| 1. Special Secretary,
Health Department. | 1914 25-4-17 | Chairman |
| 2. Representative from Accountant General Sindh, | | Member |
| 3. Independent expert from relevant field concerning
(to be nominated by the Head of Procuring Agency) | | Member |
| 4. Deputy Secretary (PM&I) | | Member |
| 5. Deputy Secretary (General) | | Member/Secretary |

ToRs.

- To perform according to Rule 31 of SPPRA, 2010.
- To Perform any other function ancillary and incidental to the above.

- DR. FAZLULLAH PECHUHO -
 SECRETARY TO GOVERNMENT OF Sindh

No.HD(P&E)3-2(427)/2014.

Karachi dated the 10th APRIL, 2017

1. The D.G.Health Services, Sindh, Hyderabad.
2. The District Health Officers (All)
3. Medical Superintendents (All) Liaquat University Hyderabad.
4. All the Member of the Committee.
5. The PS to Secretary Health Department, Govt. of Sindh, Karachi.
6. Office order file.

SECTION OFFICER (PM&I)

Handwritten signature/initials

Handwritten date: 27/4
Handwritten text: Gen B

SAY NO TO CORRUPTION

ٽيندر نوٽيس

1- مالي سال 2017-18 لاءِ ADP/SNE (رسڪيو سينٽر) (يورولوجي ڊپارٽمينٽ) لياقت يونيورسٽي اسپتال، حيدرآباد/ڄامشورو هيٺ. هيٺين مشينري/ايڪوئيمينٽ جي خريداري لاءِ سٺي ساڪ رکندڙ دلچسپي رکندڙ ايجنٽس/سول ايجنٽ/پاڪستان ۾ دسترقيوتز پرڏيهي ڪمپنين کان. ايس پي پي آري رولز 2010 (ترميم ٿيل 2013) مطابق پرسڪرائيڊ پروفارما تي مهربند ٽيندر طلب ڪجن ٿا. واک دستاويز جمع ڪرائڻ جي آخري تاريخ 2018-02-27 صبح 11:30 وڳي آهي. جيڪي ساڳي ڏينهن تي منجهند 12:00 وڳي شرڪت ڪندڙن جي زيرو ڪوليا ويندا.

سيڪورٽي نمبر	ٽيندر نمبر	ڊسڪريشن	ٽيندر في نا قابل واپسي	سيڪورٽي ڊپازٽ
1	34/2017-18	SNE/ADP سال 18-2017 لاءِ (رسڪيو سينٽر) (يورولوجي ڊپارٽمينٽ) پارت-1، تحت مشينري/ايڪوئيمينٽ جي خريداري	2000 رپيا	2.5 سيڪڙو

پروسيجر اينڊ شيڊول

بڊنگ پروسيجر	سنگل اسٽيج ٽوانولپ
واڪ جي موٽريت	90 ڏينهن
ٽيندر دستاويزن جواجر	2018-02-08 کان 2018-02-26 واقع LUH، حيدرآباد
ٽيندر جمع ڪرائڻ	2018-02-27 بوقت صبح 11:30 وڳي، واقع LUH حيدرآباد
ٽيندر ڪوليا ويندا	2018-02-27 بوقت منجهند 12:00 وڳي واقع LUH حيدرآباد

- شرط ۽ ضابطن سان، ٽيڪنيڪل صراحتن ۽ گهرجن جي شيڊول تي مشتمل ٽيندر دستاويزن جو هڪ سيٽ، هيٺ صحيح ڪندڙ جي آفيس، لياقت يونيورسٽي اسپتال (LUH) حيدرآباد مان هر هڪ سيٽ جي لاءِ 2000 رپيا في نا قابل واپسي ادا ڪري خريد ڪري سگهجي ٿو.
- ايس پي پي آري رولز جي مطابق، ميڊيڪل سپرنٽينڊنٽ LUH حيدرآباد/ڄامشورو شيڊول جي ڪنهن به ائٽر جي تعداد کي جڏهن ۽ جتي ضروري سمجهندو ان مطابق گهٽائڻ يا وڌائڻ جو حق محفوظ رکي ٿو. پروڪيورنگ ايجنسي ايس پي پي آري رولز جي لاڳاپيل شق مطابق ڪنهن به يا سمورن واکن کي رد ڪري سگهي ٿي.
- ٽيندر دستاويزن ايس پي پي آري جي ويب سائيٽ تان ڏاڻوڻ لود به ڪري سگهجن ٿا.
- واڪ ۾ (لاڳو ٿيڻ جي صورت ۾) سمورا سرڪاري ٽيڪس شامل هئڻ گهرجن.
- سرڪاري قانون مطابق بلن مان ڪوتاهي ڪئي ويندي.
- واڪ ڏيندڙ کي ميڊيڪل سپرنٽينڊنٽ لياقت يونيورسٽي اسپتال حيدرآباد/ڄامشورو جي حق ۾، ڪوٽ ڪيل آئٽمز جي ڪل بڊ واري ماليت جي 2.5 سيڪڙو بڊ سيڪورٽي طور، پي آرڊر/ڊمانڊ ڊرافٽ شامل هئڻ گهرجي. جڏهن ته ايس پي پي آري رولز مطابق نيڪيدار کان نيڪي جي ڪل ماليت تي 5 سيڪڙو پرفارمنس سيڪورٽي ورتي ويندي.
- ٽيندر نوٽيس ۽ ٽيندر دستاويزن ۾ تفاوت جي صورت ۾ ٽيندر دستاويزن کي اوليت ڏني ويندي.
- سرڪاري موڪل جي اعلان يا ڪنهن اوجھتي صورتحال پيدا ٿيڻ جي صورت ۾ واک ايندڙ ڪم واري ڏينهن تي ڄاڻايل شيڊول مطابق جمع ڪيا ۽ ڪوليا ويندا.
- وڌيڪ معلومات، LUH حيدرآباد ۾ آفيس وقت دوران هيٺ ڄاڻايل نمبرن تان يا ذاتي طور اچي وٺي سگهجي ٿي.
فون نمبر: 021-9210207
فيڪس نمبر: 022-9210208

ميڊيڪل سپرنٽينڊنٽ

لياقت يونيورسٽي اسپتال

حيدرآباد/ڄامشورو

INF/KRY.NO.694/2018

ڪريشن کان انڪار ڪريو، اسان دهشتگردي خلاف متحد آهيون

سنڌ و تعليم جي بهتري لاءِ Text علمي ۽ پنهنجو پيغام لکي 8398 تي ايس ايم ايس ڪريو

لیاقت یونیورسٹی اسپتال حیدرآباد/جامشورو

فون نمبر 022-9210207 بد عنوانی سے انکار فیکس نمبر 022-9210208

ٹینڈر نوٹس

سپر اتوائمن-2010 (ترمیم شدہ-2013) کے تحت رواں مالی سال 18-2017 کے دوران لیاقت یونیورسٹی اسپتال، حیدرآباد/جامشورو کیلئے ایس این ای/سالانہ ترقیاتی پروگرام کے تحت مندرجہ ذیل مشینری/ایکیپمنٹ (ریسکے سینٹر) (یورولوجی ڈیپارٹمنٹ)، کیلئے درج ذیل ایشیا کی خریداری کیلئے خواہشمند ایجنٹ/سول ایجنٹ/پاکستان میں غیر ملکی کمپنیوں کے تیسرے نمبر کنندگان سے مقررہ پروفارما (خانکے) پر سرسہر ٹینڈرز مطلوب ہیں۔ ٹینڈرز دستاویزات جمع کرانے کی آخری تاریخ 27-02-2018 دن کے ساڑھے گیارہ بجے تک ہے اور ٹینڈرز حاضر رہنے کے خواہشمند بولی دہندگان/شرکت کنندگان کے سامنے اسی دن بارہ بجے کھولے جائیں گے۔

نمبر شار	ٹینڈر نمبر	تفصیلات	ٹینڈر کیس (ناقابل واپس)	زر مضامنت
1	34/2017-18	یورولوجی ڈیپارٹمنٹ اور ریسکے سینٹر کیلئے مشینری/ایکیپمنٹ کی خریداری ایس این ای/سالانہ ترقیاتی پروگرام کے تحت برائے سال 18-2017	2,000	2.5 لاکھ

شیڈول اور طریقہ کار

پیشکشوں کا طریقہ کار	سینٹر ایچ۔ دو لگانے
بولی کی مستند میعاد	90 یام
ٹینڈرز دستاویزات کا اجرا	2018-02-08 تا 2018-02-26 واقع لیاقت یونیورسٹی اسپتال، حیدرآباد
ٹینڈرز کا جمع کرنا	2018-02-27 کو دن کے ساڑھے گیارہ بجے ایل یو ایچ، حیدرآباد پے
ٹینڈرز کا کھلنا	2018-02-27 کو دن کے بارہ بجے ایل یو ایچ، حیدرآباد پر

- 1- ضروریات کے شیڈول، تکنیکی تفصیلات مع شرائط اور ضوابط پر مشتمل ٹینڈرز دستاویزات کا مکمل سیٹ ٹی کس=2000 روپے کی ناقابل واپسی ادائیگی پر زیر تحفظ لیاقت یونیورسٹی اسپتال (ایل یو ایچ) حیدرآباد کے آفس سے خریدی جاسکتی ہیں۔
- 2- میڈیکل سپرنٹنڈنٹ، لیاقت یونیورسٹی اسپتال، حیدرآباد/جامشورو کو یہ حق حاصل ہے کہ وہ سپر اتوائمن کے تحت کسی بھی آن لائن کی مقدار میں کسی یا اضافہ کر دے۔ پروکیورنگ ایجنسی کو یہ حق حاصل ہے کہ وہ سپر اتوائمن کی متعلقہ سٹیشن کے تحت کسی ایک یا تمام پیشکشوں کو مسترد کر دے۔
- 3- ٹینڈرز دستاویزات سپر ایکی ویب سائٹ سے بھی ڈاؤن لوڈ کی جاسکتی ہیں۔
- 4- پیشکش میں تمام سرکاری ٹیکسز شامل ہونے چاہئیں (اگر قابل عمل ہوں)
- 5- ٹیکسز سرکاری اتوائمن کے تحت بلوں سے منہا کر لئے جائیں گے۔
- 6- بولی کی مجموعی مالیت کے ذمہ داری فیصد کے مساوی رقم کا بے آر ڈر/ڈیمانڈ ڈرافٹ نام میڈیکل سپرنٹنڈنٹ لیاقت یونیورسٹی اسپتال حیدرآباد/جامشورو جمع کرنا ہوگا اور ایک پانچ لاکھ کا کارڈنگی سکیورٹی ٹیکس یا سپر اتوائمن کے تحت جمع کرنا ہوگا۔
- 7- اگر کسی وجہ سے ٹینڈرز نوٹس اور ٹینڈرز دستاویزات میں کوئی فرق پایا گیا تو ٹینڈرز دستاویزات کو ترجیح دی جائے گی۔
- 8- اگر کسی وجہ سے عام تعطیل کا اعلان ہو گیا یا کسی دیگر وجہ سے سب ٹینڈرز مکمل نہ کیے تو یہی ٹینڈرز دئے گئے شیڈول کے تحت اگلے کام کے دن جمع کرانے اور کھولے جائیں گے۔
- 9- کوٹیز درج ذیل نمبر یا فرو کو لیاقت یونیورسٹی اسپتال واقع حیدرآباد پر اوقات کار کے دوران ارسال کریں۔

فون 022-9210207
فیکس 022-9210208



INF/KRY No.694/18

بد عنوانی سے انکار

8 3 9 8

TENDER NOTICE

1) According SPPRA Rules 2010 (Amendment-2013) sealed tenders on prescribed proforma are hereby invited from well reputed firms, interested Agents / Sole Agents / Foreign Companies Distributors in Pakistan for the purchase of following Machinery / Equipment under ADP/SNE (Rescue Centre) (Urology Department) Liaquat University Hospital, Hyderabad / Jamsoro for the financial year 2017-18. The last date for submission of tender documents is 27-02-2018 up to 11.30 a.m. and will be opened on same day at 12.00 Noon in the presence of participants.

S. No.	Tender #	Description	Tender-Fee (Non-Refundable)	Security Deposit
1.	34/2017-18	PURCHASE OF MACHINERY / EQUIPMENT UNDER SNE / ADP FOR THE YEAR 2017-18 (RESCUE CENTRE) (UROLOGY DEPARTMENT) (PART-1)	Rs.2000/-	2.5%

PROCEDURE & SCHEDULE

Bidding Procedure	Single-Stage — Two Envelope System
Bid Validity	90 days
Issuance of Tender Documents	08-02-2018 to 26-02-2018 at LUH, Hyderabad
Submission of Tender	27-02-2018 at 11:30 AM at LUH, Hyderabad
Opening of Tender	27-02-2018 at 12:00 Noon at LUH, Hyderabad

2) Complete set of tender documents containing Schedule of Requirements, Technical Specification with Terms & Conditions can be purchased with a non-refundable fee of Rs.2,000/- each from the office of undersigned Liaquat University Hospital (LUH), Hyderabad.

3) The Medical Superintendent LUH, Hyderabad / Jamsoro reserves the right to increase or decrease the quantity of any item of schedules as and when it is deemed necessary according to SPPRA Rules. The Procuring Agency may reject all or any bid subject to the relevant provision of SPPRA Rules.

4) Tender documents can be downloaded from SPPRA website.

5) Bids shall include all government taxes (if applicable).

6) Deduction in the bills will be done as per government rules.

7) Bid Security @ 2.5% of the total bid value of quoted items by the bidder in shape of pay order / demand draft in favour of Medical Superintendent Liaquat University Hospital Hyderabad / Jamsoro whereas an amount of the contract 5% performance security is to be obtained from contractor as per SPPRA Rules.

8) In case of discrepancies between the Tender Notice and the Tender Documents, the Tender Documents shall take precedence.

9) In case of announcement of public holiday or any unfavourable circumstance, the bids shall be submitted and opened as per given schedule on the next working day.

10) Queries can be addressed at the following numbers or in person during office hours at LUH, Hyderabad.

Ph: 022-9210207

Fax: 022-9210208




Medical Superintendent
Liaquat University Hospital
Hyderabad / Jamsoro

INF-KPRY: 694/18

لا تبيعوا
معلوماتكم
بمقابل
مال

Say No to Corruption

**MEETING OF PROCUREMENT COMMITTEE FOR OPENING OF TECHNICAL BIDS
FOR PURCHASE OF MACHINERY / EQUIPMENT FOR UROLOGY DEPARTMENT
(RESCUE CENTRE PART-I) UNDER ADP / SNE FOR THE FINANCIAL YEAR 2017-18
IN THE COMMITTEE ROOM LUH, HYDERABAD ON 27.02.2018.**

S.NO	NAME & DESIGNATION		SIGNATURE,
01	Dr. Naeem Zia Memon Addl: Medical Superintendent, LUH, Hyderabad.	Chairman	
02	Dr. Jawed Altaf Jat Head of Urology Department, LUMHS, Jamshoro	Member	
03	Dr. Anjum Rehman Deputy Medical Superintendent, L.U. Hospital, Hyderabad.	Member	

Opening of Tender Documents for Purchase of Machinery / Equipment for Urology Department (Rescue Centre Part-I) under ADP /SNE
for the Financial year 2017-18 at Liaquat University Hospital, Hyderabad/Jamshoro

Dated:- 27.02.2018

ATTENDANCE SLIP

S.No.	Name of Firm	Phone Number	Postal Address	Name of Participant	C.N.I.C No: of participants	Signature
1	Hospital Supply Corporation	0344-3546831	Hyderabad.	Akram Khan	44202-9888398-5	A. M. Khan.
2	Nasir International	03313551771	Karachi	SADIK HUSSAIN	42501-4911940-9	S. I.
3	Butech Corporation	0342-2600250	Karachi	M. Hassam Khan	42101-2922753-3	M. Hassam Khan.
4	Verizon Verizon					
5	Allmed Solutions	0333-2753119	Karachi	Arif	413036967844-5	A. M.
6	Verizon	0336-3988062 0342-2438328	Karachi	Faiz ullah	42201-1789218-3	F. U.
7	PoFular International	03018380765	Hyderabad	Uzayaz Ali	54400-5254237-1	U. A.
8	Universal Enterprises	03312663310	Hyd	Obaidullah Sheikh	41304-4423797-5	O. S.
9						
10						

(Dr. Anjum Rehman)
Addl: Medical Superintendent (Technical)
LUH, Hyderabad
Member Procurement Committee

(Dr. Javed Altaf Jat)
Head of Urology Department
LUMHS, Jamshoro
Member Procurement Committee

(Dr. Naeem Zia Memon)
Addl: Medical Superintendent (General)
LUH, Hyderabad
Chairman Procurement Committee

MINUTES OF BID OPENING MEETING


NIT No.	Inf No:694/2018
SPPRA Number	S.No:36633
Method of Procurement	Single Stage Two Envelope

Name of Work: **Purchase of Machinery / Equipment Under SNE / ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-1)**

Meeting of the Procurement Committee (PC) was held on Tuesday, 27 February 2018 for opening of above mentioned NIT. Meeting was attended by members of the Procurement Committee and the representatives of bidders (attendance sheets are enclosed).

The following companies participated in the bidding process with their respective bids by submission deadline of 12:30pm on Tuesday 27.02.2018.


- 1) M/s Hospital Supply Corporation
- 2) M/s Ontech Corporation
- 3) M/s Allmed Solution
- 4) M/s Universal Enterprise
- 5) M/s Verizon
- 6) M/s Noor International
- 7) M/s Popular International



(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad



(Member)
Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro






(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

**PROCUREMENT COMMITTEE
ATTENDANCE SHEET**

Tender No. **34/2017-18**

Name of Work: **Purchase of Machinery / Equipment Under SNE / ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-1)**


The following members of the Procurement Committee (PC) were present on Tuesday, 27 February 2018 for opening of above mentioned NIT.

1.	Dr. Naeem Zia Memon	Additional Medical Superintendent Liaquat University Hospital Hyderabad	Chairman	
2.	Dr. Jawed Altaf Jat	Head of Urology Department LUMHS Jamshoro	Member	
3.	Dr. Anjum Rehman	Deputy Medical Superintendent Liaquat University Hospital Hyderabad	Member	

BIDDERS ELIGIBILITY EVALUATION

S. No.	Parameter	Max Points	M/s Hospital Supply Corporation	M/s Ontech Corporation	M/s Allmed Solution	M/s Universal Enterprise	M/s Verizon	M/s Noor International	M/s Popular International
1	Conformity to the Technical Specifications	20							
(a)	Fully compliant with the required specifications	20	-	-	20	-	20	-	-
(b)	Compliant with minor deviation $\leq 10\%$ (subject to main function is not affected)	15	15	15	-	15	-	15	15
2	Warranty	20							
(a)	Warranty (period more than required)	20	-	-	-	-	-	-	-
(b)	Warranty (period as required)	15	15	15	15	15	15	15	15
3	Bidder's Legal Status, Certification & Experience	15							
(a)	NTN & GST Certificate	2	2	2	2	2	2	2	2
(b)	Professional Tax Certificate	2	2	2	2	2	2	2	2
(c)	SECP Incorporation Certificate	2	-	-	-	-	-	-	-
(d)	ISO 9001 Certificate	1	-	-	1	1	-	-	-
(e)	Purchase Orders of last year worth ≥ 1 million rupees (2 point/each order copy)	8	-	6	8	2	-	8	6
4	Bidder's Financial Status	9							
(a)	Bank Certificate	3	3	3	3	3	3	3	3
(b)	Income Tax Return for annual business of ≥ 20 million rupees (last three years)	3	3	2	3	3	2	3	2
(c)	Audited Financial Statement (last three years)	3	3	2	3	3	2	3	2
5	Bidder's After Sales Services in Sindh	6							
(a)	Service Workshop	3	3	3	3	3	3	3	3
(b)	Qualified & Trained Personnel (BE/BSc/DAE) with Bio-Data	3	3	3	3	3	3	3	3

S. No.	Parameter	Max Points	M/s Hospital Supply Corporation	M/s Ontech Corporation	M/s Allmed Solution	M/s Universal Enterprise	M/s Verizon	M/s Noor International	M/s Popular International
6	Original Equipment Manufacturer (OEM) or Brand Quoted	30	1	2	3	4	5	6	7
(a)	Manufacturer's Authorization	5	-	5	5	5	5	5	-
(b)	Manufacturer's Website	2	-	2	2	2	2	2	2
(c)	Brochure or Catalogue (original)	2	-	2	2	2	2	2	2
(d)	ISO 9001 Certificate	2	-	2	2	-	2	-	-
(e)	ISO 13485 Certificate	2	-	2	2	2	2	2	2
(f)	CE	2	-	2	2	2	2	2	2
(g)	FDA	4	-	-	-	-	4	-	4
(h)	JIS	2	-	-	-	-	-	-	-
(i)	References of OEM brand / model								
	(i) Government / Semi-Government Teaching Hospital / Institute	3	-	3	3	3	3	3	3
	(ii) Armed Forces / NGOs Hospital / Institute	2	-	2	2	2	-	-	-
	(iii) Private Renowned Teaching Hospital / Institute	2	-	2	2	-	-	2	-
	(iv) International Hospital / Institute	2	-	-	-	-	-	-	-
	Total Points	100	49	73	85	70	74	75	68
			Ineligible	Eligible	Eligible	Eligible	Eligible	Eligible	Ineligible



(Chairman)

Dr. Naeem Zia Memon

Additional Medical Superintendent
Liaquat University Hospital
Hyderabad



(Member)

Dr. Jawed Altaf Jat

Head of Urology Department
LUMHS
Jamshoro



(Member)

Dr. Anjum Rehman

Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

DOCUMENTS CHECKLIST


S. No.	Document Description	M/s Hospital Supply Corporation	M/s Ontech Corporation	M/s Allmed Solution	M/s Universal Enterprise	M/s Verizon	M/s Noor International	M/s Popular International
		1	2	3	4	5	6	7
General Bid Documents:								
1	Tender Purchase Receipt (Original)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
2	Bid Security (Pay Order / Bank Draft)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
3	General & Special Conditions of Contract	Yes	Yes	Yes	Yes	Yes	Yes	Yes
4	Schedule of Requirements	Yes	Yes	Yes	Yes	Yes	Yes	Yes
5	Technical Specifications	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Bidder's Documents:								
6	Manufacturer's Authorization (as per sample form)	No	Yes	Yes	Yes	Yes	Yes	No
7	Undertaking (as per sample form)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
8	Certificate (as per sample form)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
9	Income Tax & GST Registration Certificate	Yes	Yes	Yes	Yes	Yes	Yes	Yes
10	Professional Tax Certificate (Sindh)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
11	SECP Incorporation Certificate (if applicable)	-	-	-	-	-	-	-
12	Company Profile	No	Yes	Yes	Yes	Yes	Yes	Yes
13	Bank Certificate	Yes	Yes	Yes	Yes	Yes	Yes	Yes
14	Income Tax Return (last year)	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Original Equipment Manufacturer (OEM) or Brand Quoted:								
15	Catalogue / Brochures (original)	Yes	Yes	Yes	Yes	Yes	Yes	Yes

Swain

Order

[Signature]

S. No.	Document Description	M/s Hospital Supply Corporation	M/s Ontech Corporation	M/s Allmed Solution	M/s Universal Enterprise	M/s Verizon	M/s Noor International	M/s Popular International
16	ISO 9001 / ISO 13485	No	Yes	Yes	Yes	Yes	Yes	Yes
17	CE / FDA / JIS	No	Yes	Yes	Yes	Yes	Yes	Yes
18	References of the quoted brand & model (<i>local</i>)	No	Yes	Yes	Yes	Yes	Yes	Yes
19	References of the quoted brand & model (<i>international</i>)	-	-	-	-	-	-	-
Additional Documents:								
-	-	Ineligible	Eligible	Eligible	Eligible	Eligible	Eligible	Ineligible



(Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad



(Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro



(Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

Tender No. 34/2017-18


Purchase of Machinery / Equipment under SNE / ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)

Opened on 27.02.2018

SCHEDULE OF REQUIREMENTS

S. No.	Item Description	Qty	M/s Hospital Supply Corporation	M/s Ontech Corporation	M/s Allmed Solution	M/s Universal Enterprise	M/s Verizon	M/s Noor International	M/s Popular International
1.	Minimal Invasive Percutaneous Nephroscope (XS/S)	01	-	-	Yes	-	Yes	-	-
2.	Minimal Invasive Percutaneous Nephroscope (M)	01	-	-	Yes	-	Yes	-	-
3.	Minimal Invasive Percutaneous Nephroscope (L)	01	-	-	Yes	-	Yes	-	-
4.	Nephroscope Adult	01	-	-	Yes	-	Yes	-	-
5.	URS Adult	01	-	-	Yes	Yes	Yes	-	-
6.	Cystoscope Adult	01	-	-	Yes	Yes	Yes	Yes	-
7.	Stone Punch	01	-	-	Yes	Yes	Yes	-	-
8.	Lithorite	01	-	-	Yes	-	Yes	Yes	-
9.	Urethrotome Adult	01	-	-	Yes	Yes	Yes	Yes	-
10.	Resectoscope Adult (Monopolar)	01	-	-	Yes	Yes	Yes	Yes	-
11.	Bipolar Resectoscope Adult	01	-	-	Yes	Yes	Yes	-	-
12.	Electro Surgery Unit 400 Watts	01	-	-	Yes	Yes	Yes	-	Yes
13.	Ultrasonic Intra-Corporeal Lithotripter	01	-	Yes	-	-	Yes	-	-
14.	Fiber Optic CystoUrethro Fiberscope	01	-	-	Yes	-	Yes	-	-
15.	Video Cyst Scope	01	-	-	Yes	-	Yes	-	-
16.	Video URS	01	-	-	Yes	-	Yes	-	-

S. No.	Item Description	Qty	M/s Hospital Supply Corporation	M/s Ontech Corporation	M/s Allmed Solution	M/s Universal Enterprise	M/s Verizon	M/s Noor International	M/s Popular International
17.	Full HD Camera System	01	-	-	Yes	Yes	Yes	Yes	Yes
18.	Holmium Laser System For Urology Applications	01	Yes	Yes	Yes	-	Yes	-	-
19.	PeadsNephroscope	01	-	-	Yes	Yes	Yes	-	-
20.	URS Paeds	01	-	-	Yes	Yes	Yes	-	-
21.	PeadsCystoscope	01	-	-	Yes	Yes	Yes	Yes	-
22.	UrethrotomePeads	01	-	-	Yes	Yes	Yes	Yes	-
23.	PeadsResectoscope	01	-	-	Yes	Yes	Yes	Yes	-



(Chairman)

Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad



(Member)

Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro



(Member)

Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

Tender No. 34/2017-18


Purchase of Machinery / Equipment under SNE / ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-1)

Opened on 27.02.2018

TECHNICAL SPECIFICATION EVALUATION & COMPLIANCE

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
1.	<p>Minimal Invasive Percutaneous Nephroscope (XS/S)</p> <p>1- Nephroscope, with angled eyepiece, 1 7.5 Fr., length 24 cm, autoclavable, 1 working channel 2 Fr., 1 irrigation channel 3 Fr., fiber optic light transmission incorporated, for use with Operating Sheaths</p> <p>Following accessories are included: Instrument Port with Sealing System and Quick Release Lock LUER-Lock Tube Connector, male LUER-Lock Tube Connector, with stopcock, dismantling Insertion Aid, for guide wires Wire Tray Multiport Bridge Cleaning Adaptor</p> <p>2- One Step Dilator, with central 1 channel for guide wires, for use with 8.5/9.5 Fr. Operating Sheaths</p> <p>3- Operating Sheath, 8.5/9.5 Fr., 1 working length 15 cm, for use with Nephroscope for MIP XS/S One Step Dilator and Applicator</p> <p>4- One Step Dilator, with central 1 channel for guide wires, for use with Operating Sheath 11/12 Fr.</p> <p>5- Operating Sheath, 11/12 Fr. working length 15 cm, for continuous irrigation and suction, for use with Nephroscope for One Step Dilator and Applicator 1</p> <p>6- Operating Sheath, for the supine position, 11/12 Fr., working length 18 cm, for continuous irrigation and suction, for use with Nephroscope One Step Dilator and Applicator 1</p> <p>7- Applicator for Sealing, including sheath and pusher for use with Operating Sheaths 1</p>	01	Karl Storz Yes	Richard Wolf Yes

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizo
8-	Applicator, for supine position, with rod for sealing with sheath, for use with Operating Sheaths	1	Yes	Yes
9-	Suction System, power supply 100 - 240 VAC 50/60 Hz, Release 45 Consisting of: Suction Unit Connecting Cable, length 100 cm Basic Tubing Set, for single use Control Cable	1	Yes	Yes
10-	One-Pedal Footswitch, one-stage, digital FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes CE	Yes CE
Remarks			Germany Complies with the required items, its specifications and criteria.	Germany Complies with the required items, its specifications and criteria.
			Responsive	Responsive


(Chairman)

Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad


(Member)

Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro


(Member)


Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
2.	<p>Minimal Invasive Percutaneous Nephroscope (M)</p> <p>1- Nephroscope , with angled eyepiece, 12 Fr., 12°, length 22 cm, autoclavable, 1 working channel 6.7 Fr. for instruments up to 5 Fr., fiber optic light transmission incorporated, for use with Operating Sheaths Following accessories are included in delivery: Insertion Aid, for guide wires Instrument Port with Sealing System and Quick Release Lock Seal, package of 10 LUER-Lock Tube Connector, male LUER-Lock Tube Connector, with stopcock, dismantling Cleaning Adaptor 1</p> <p>2- One Step Dilator, with central channel for guide wires, for use with 15/16 Fr. Operating Sheaths 1</p> <p>3- One Step Dilator, with central channel and a second eccentric channel for guide wires, for use with 16.5/17.5 Fr. Operating Sheaths 1</p> <p>4- One Step Dilator, with central channel for guide wires and distal curved channel for deflection of guide wires, for use with 21/22 Fr. Operating Sheaths 1</p> <p>5- Operating Sheath, for the supine position, 15/16 Fr., working length 18 cm, for continuous irrigation and suction, for use with Nephroscope for MIP M One Step Dilator and Applicator 1</p> <p>6- Operating Sheath, for the supine position, 16.5/17.5 Fr., working length 18 cm, for continuous irrigation and suction, for use with Nephroscope for MIP M One Step Dilator and Applicator 1</p> <p>7- Operating Sheath, for the supine position, 21/22 Fr., working length 18 cm, for continuous irrigation and suction, for use with Nephroscope for MIP M One Step Dilator and</p>	01	<p>Karl Storz</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>	<p>Richard Wolf</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>


[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
	<p>Applicator 1</p> <p>8- Grasping Forceps for Foreign Bodies, 5 Fr., double action jaws, flexible, length 40 cm,</p> <p>9- Biopsy Forceps, 5 Fr., double action jaws, flexible, length 40 cm, 1</p> <p>10- Scissors, single action jaws, 5 Fr., flexible, length 40 cm, for use with Nephroscope 1</p> <p>11- Grasping Forceps, rigid, for large stones and stone fragments, 3 expanding jaws and small fixation spikes, with spring handle, length 36 cm, for use with Nephroscope for and instrument port 1</p> <p>FDA/CE/JIS Marked OR Equivalent</p> <p>USA/UK/Europe /Japan/Germany</p>		<p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>	<p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>
Remarks			<p>Complies with the required items, its specifications and criteria.</p> <p>Responsive</p>	<p>Complies with the required items, its specifications and criteria.</p> <p>Responsive</p>


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad


 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro



 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
3.	<p>Minimal Invasive Percutaneous Nephroscope (L)</p> <p>1- Nephroscope L, with angled eyepiece, 19.5 Fr., rod lens system 12 °, length 22 cm, autoclavable, 12.4 Fr. Working channel for use with instruments up to 11.5 Fr., fiber optic light transmission incorporated, for use with Operating Sheaths and Instrument Port</p> <p>Following accessories are included in delivery: Instrument Port LUER-Lock Tube Connector, male LUER-Lock Tube Connector, with stopcock, dismantling Insertion Aid, for guide wires Silicone Leaflet Washer, package of 10 Seal, package of 10 Cleaning Adaptor</p> <p>2- Dilator with central channel and a second eccentric channel for guide wires, for use with 23/24 Fr. Operating Sheaths.</p> <p>3- Operating Sheath, 23/24 Fr., working length 15 cm, for continuous irrigation and suction, for use with Nephroscope Dilator and Applicator</p> <p>4- Operating Sheath, for the supine position, 23/24 Fr., working length 18 cm, for continuous irrigation and suction, for use with Nephroscope One Step Dilator and Applicator</p> <p>5- Applicator for Sealant, including sheath and rod, for use with Operating Sheaths</p> <p>6- Forceps, for grasping stone fragments and coagula, with fenestrated jaws and U-spring handle, 11.5 Fr., length 38 cm, color code: red-black</p>	01	Karl Storz Yes Yes Yes Yes Yes Yes	Richard Wolf Yes Yes Yes Yes Yes Yes

[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
	FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes CE Germany	Yes CE Germany
Remarks			Responsive	Responsive
			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.


 (Chairman)

Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad



(Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro



(Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
4.	Nephroscope Adult	01	Karl Storz	Richard Wolf
	1- Telescope 6°, with parallel eyepiece, autoclavable, with LUER-Lock connection for inflow, withinstrument channel, fiber optic light transmission incorporated, Color code: yellow-red		Yes	Yes
	2- Operating Sheath, 26 Fr., for continuous irrigation and suction, with LUER-Lock stopcock, rotatable, Color-code: black-red		Yes	Yes
	3- Hollow Obturator and Fascial Dilator, Color code: black-red		Yes	Yes
	4- Telescoping Dilation Set, consisting of: set of 6 dilators, sizes 9, 12, 15, 18, 21 and 24 Fr., with 2 rigid and 2 flexible guide rods		Yes	Yes
	5- Dilator, 27 Fr.,		Yes	Yes
	6- Grasping Forceps for large stones and stone fragments, 3 expanding jaws and small fixation spikes, with spring handle, length 38 cm, for use with Telescopes and Operating Sheaths Color code: red-black		Yes	Yes
	7- Grasping Forceps for large stones and stone fragments, serrated, double, action jaws, with ring handle, length 38 cm, for use with Telescopes and Operating Sheaths Color code: red-black		Yes	Yes
	8- Grasping Forceps for large stones and stone fragments with fenestrated jaws and ring handle, double action jaws, length 38 cm, for use with Telescopes and Operating Sheaths Color code: red-black		Yes	Yes
	9- Knife, straight, with 3-ring- handle, length 38 cm, for use with Telescopes and Operating Sheaths Color code: red-black		Yes	Yes
	10- Puncture Cannula, for localization of renal calculi including inner and outer cannulas, package of 5		Yes	Yes


[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
	11- Guide wire, for introduction through the cannula package of 2		Yes	Yes
	12- Dilation Cannula, ø 3 mm, for introduction of a second safety guide wire, consisting of an inner and outer cannula		Yes	Yes
	13- Grasping Forceps for large stone fragments and coagula, fenestrated jaws and spring handle, length 38 cm, for use with percutaneous nephroscopes		Yes	Yes
	14- Biopsy Forceps, single action jaws, with ring handle, length 38 cm, for use with percutaneous nephroscopes		Yes	Yes
	15- Scissors, single action jaws, length 38 cm, for use with percutaneous nephroscopes		Yes	Yes
	16- Knife, sickle-shaped, with 3-ring-handle, length 38 cm, for use with percutaneous nephroscopes		Yes	Yes
	17- Suction Tube, length 38 cm, for use with percutaneous nephroscopes		Yes	Yes
	FDA/CE/JIS Marked OR Equivalent		CE	CE
	USA/UK/Europe /Japan/Germany		Germany	Germany
	Remarks		Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.


(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad


(Member)
Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro


(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

Responsive


Responsive


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
5.	<p>URS Adult</p> <p>1- Uretero-Renoscope, 8 Fr., 6°, Length Distal tip instrument sheath autoclavable, with angled eyepiece, fiber optic light transmission incorporated, 2 lateral irrigation ports and 1 working channel 5 Fr., for use with instruments up to 4 Fr.</p> <p>Following accessories are included: Insertion Aid Instrument Port with Sealing System and Quick Release Lock LUER-Lock Tube Connector LUER-Lock Tube Connector Flow Control Stopcock Wire Tray</p> <p>2- Forceps, long jaws for Steinstraße, rigid, double action jaws, 4 Fr., length 60 cm, color code: blue</p> <p>3- Grasping Forceps for stone fragments, double action jaws, 4 Fr., rigid, length 60 cm, Color code: blue</p> <p>4- Biopsy Forceps, rigid, double action jaws, 4 Fr., length 60 cm, color code: blue</p> <p>5- Grasping Forceps for large stone fragments, double action jaws, 4 Fr., rigid, length 60 cm, Color code: blue</p>	01	Karl Storz Yes Yes Yes Yes Yes Yes	Richard Wolf Yes Yes Yes Yes Yes Yes	Tekno Yes Yes Yes Yes Yes Yes


Dr. [Signature]

[Signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Univer Enterprise
6-	Stone Basket, sterile, disposable 2.5 Fr., length 120 cm FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes CE Germany	Yes CE Germany	Yes CE Germany
Remarks			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Limited references, Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non-Responsive


(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad


(Member)
Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro



(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s Noor International
6.	Cystoscope Adult 1- Telescope 30°, enlarged view, diameter 4 mm, length 30 cm, autoclavable, fiber optic light transmission incorporated 2- Cystoscope-Urethroscope-Sheath, 22 Fr., with obturator and 2 LUER-Lock adaptors. 3- Cystoscope-Urethroscope Sheath, 17 Fr., with obturator and 2 LUER-Lock adaptor. 4- Telescope Bridge with 2 lockable instrument channels 5- Catheter Deflecting Mechanism, with 2 instrument channels, with ratchet 6- Grasping Forceps for removal of foreign bodies, 7 Fr. double action jaws, flexible, length 40 cm 7- Biopsy Forceps, 7 Fr., double action jaws, flexible, length 40 cm 8- Coagulating Electrode, 4 Fr., unipolar, length 53 cm 9- Unipolar High Frequency Cord, with 4 mm plug, length 300 cm, for models, Erbe type T, older models and Elliman	01	Karl Storz Yes Yes Yes Yes Yes Yes Yes Yes Yes	Richard Wolf Yes Yes Yes Yes Yes Yes Yes Yes Yes	Tekno Yes Yes Yes Yes Yes Yes Yes Yes Yes	Stema Yes Yes Yes Yes Yes Yes Yes Yes Yes






S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s M/F International
	tray(same manufacturer) FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		CE Germany	CE Germany	CE Germany	CE Germany
Remarks			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.	No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non-Responsive	Non-Responsive


(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad


(Member)
Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro


(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
7.	<p>Stone Punch</p> <p>1- Telescope 30°, enlarged view, diameter 4 mm, length 30 cm, autoclavable, fiber optic light transmission incorporated.</p> <p>2- Punch-Working Element</p> <p>3- Punch Sheath, with Central Valve, including connecting tubes for in- and outflow, 25 Fr., straight beak, with obturator</p> <p>4- Insert Tube, with channel for flexible instruments, 7 Fr., with atraumatic beak for urethroscopy tray (same manufacturer)</p> <p>FDA/CE/JIS Marked OR Equivalent</p> <p>USA/UK/Europe /Japan/Germany</p>	01	<p>Karl Storz</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>	<p>Richard Wolf</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>	<p>Tekno</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>
			<p>Complies with the required items, its specifications and criteria.</p> <p>Responsive</p>	<p>Complies with the required items, its specifications and criteria.</p> <p>Responsive</p>	<p>Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.</p> <p>Non-Responsive</p>

Remarks

(Signature)

(Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital

(Signature)

(Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS

(Signature)


(Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Noor International
8.	<p>Lithorite</p> <p>1- Lateral Telescope 70Å°, enlarged view, Å, 4 mm, autoclavable, fiber optic light transmission incorporated, Color code: yellow</p> <p>2- Lithotrite, incorporating handle and turning screw action, 24 Fr., including 2 LUER-lock adaptors</p> <p>FDA/CE/JIS Marked OR Equivalent</p> <p>USA/UK/Europe /Japan/Germany</p>	01	<p>Karl Storz</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>	<p>Richard Wolf</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>	<p>Stema</p> <p>Yes</p> <p>Yes</p> <p>CE</p> <p>Germany</p>
Remarks					
			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non-Responsive



(Chairman)

Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad



(Member)

Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro



(Member)


Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s Noor International
9.	Urethrotome Adult	01	Karl Storz	Richard Wolf	Tekno	Stema
	1- Telescope 30°, enlarged view, diameter 4 mm, length 30 cm, autoclavable, fiber optic lighttransmission incorporated.		Yes	Yes	Yes	Yes
	2- SACHSE, Urethrotome Sheath, 21 Fr., with channel for Filiiform-Bougies and 2 LUER-lock adaptors		Yes	Yes	Yes	Yes
	3- Obturator, for urethrotome sheath, 21 Fr.,		Yes	Yes	Yes	Yes
	4- Telescope bridge, with channel for instruments 5 Fr.		Yes	Yes	Yes	Yes
	5- Supplementary Sheath, half-round, to insert a balloon catheter, to slip on urethrotome sheath		Yes	Yes	Yes	Yes
	6- Supplementary Sheath, for continuous irrigation and suction, to slip on urethrotome sheath		Yes	Yes	Yes	Yes
	7- Working Element (also for use with optical Urethrotome) Motion by means of a spring. The thumb support is movable. In rest position the electrode is inside the sheath.		Yes	Yes	Yes	Yes
	8- SACHSE Cold Knife straight, not to be used with HF current tray (same manufacturer)		Yes	Yes	Yes	Yes

[Handwritten signature]

[Handwritten signature]


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s N International
	FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		CE Germany Complies with the required items, its specifications and criteria.	CE Germany Complies with the required items, its specifications and criteria.	CE Germany Limited references, Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.	CE Germany No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied.
	Remarks		Responsive	Responsive	Non-Responsive	Non-Responsive


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad


 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro


 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s Noor International
10.	Resectoscope Adult (Monoploar) 1- Telescope 30°, enlarged view, diameter 4 mm, length 30 cm, autoclavable, fiber optic light transmission incorporated 2- Working Element Motion by means of a finger grip. In rest position the electrode is outside the sheath. 3- Resectoscope Sheath, including connecting tubes for in- and outflow, 26 Fr., oblique beak, rotatable inner tube with ceramic insulation, color code: yellow 4- Standard-Obturator for use with sheaths 24/26 Fr., 5- ELLIK Evacuator 6- Cutting Loop angled, 24/26 Fr., color code: yellow 7- Unipolar High Frequency Cord, with 4 mm plug, length 300 cm, tray (same manufacturer) FDA/CE/IIIS Marked OR Equivalent USA/UK/Europe /Japan/Germany	01	Karl Storz Yes Yes Yes Yes Yes Yes Yes CE Germany	Richard Wolf Yes Yes Yes Yes Yes Yes Yes CE Germany	Tekno Yes Yes Yes Yes Yes Yes Yes CE Germany	Stema Yes Yes Yes Yes Yes Yes Yes CE Germany
Remarks			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.	No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non-Responsive	Non-Responsive



(Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital, Hyderabad



(Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS, Jamshoro




(Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital, Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
11.	Bipolar Resectoscope Adult 1- Telescope 30°, diameter 4 mm, length 30 cm, autoclavable, fiber optic light transmission incorporated. 2- Electrotome, Including: Working Element Cutting Loop, bipolar Coagulation Electrode, bipolar High Frequency Cord Protection Tube 3- Resectoscope Sheath, including connecting tubes for in- and outflow, 26 Fr., oblique beak, rotatable inner tube with ceramic insulation. 4- SCHMIEDT Visual Obturator, for use with sheaths 24/26 Fr., Color code: yellow 5- Cutting Loop bipolar, 24/26 Fr., for use with Telescope color code: yellow 6- Bipolar High Frequency Cord 7- 400 High End-Set, SCB, power supply 220 - 240 VAC, 50/60 Hz, HF connecting sockets: Bipolar combination Multifunction Unipolar 3-pin + Erbe Neutral electrode combination 6.3 mm jack and 2-pin, System requirements: software based diathermy Consisting of: Autocon 400, Mains Cord Connecting Cable, length 100 cm 8- External Hardware Upgrade for "bipolar resection" function. Consists of: External Auxiliary Module Bipolar HF-Cabel Assembly, Holder Auxiliary Module Valid for devices with Software Generator	01	Karl Storz Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	Richard Wolf Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes	Tekno Yes Yes Yes Yes Yes Yes Yes Yes Yes Yes






S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
	control 10.00 and above. Devices with software below 10.00 need an additional software and possible hardware update, made in the factory				
9-	Coagulating Electrode bipolar, pointed, 24 Fr., for use with Telescope color code: yellow		Yes	Yes	Yes
10-	Vaporization Electrode half-moon, bipolar, ball-shaped, 24/26 Fr., for use with HOPKINS® II Telescope color		Yes	Yes	Yes
11-	Protection Tube, for sterilization and storage of electrodes, loops, curettes and knives		Yes	Yes	Yes
12-	Three-pedal footswitch for HF Generators.		Yes	Yes	Yes
13-	Neutral Electrode made of conductive silicone, with 2 rubber straps for fastening, A = 500 cm ² surface area. Connecting cord is required		Yes	Yes	Yes
14-	Connecting Cord, for connecting the neutral electrodes length 400 cm		Yes	Yes	Yes
	FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		CE Germany	CE Germany	CE Germany
Remarks			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non-Responsive



Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital, Hyderabad



Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS, Jamshoro



Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital, Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
12.	<p>Electrosurgery Unit 400 watts</p> <p>DIATHERMY Special Features</p> <ul style="list-style-type: none"> - Functional Test - Progressive power control - Endo Mode" Time Control cutting - Bipolar coagulation with foot switch - Connector for Monopolar hand switches - Multifunctional connector for Bipolar instruments - Monopolar cutting & coagulation by foot Switch <p>Technical Specification of the Unit</p> <ul style="list-style-type: none"> - Supply voltage: <ul style="list-style-type: none"> - 220-240 V; 50-60 Hz or 100-127 V; 50-60 Hz - Selectable via dummy plug located inside the unit by Technical Service - Power input: with no HF power output: approx.16 VA, at max. power output: approx. 800 VA - Class of protection : I - Classified acc. to MDD: II b - Leakage currents LF and HF : in acc. with EN IEC 60601, Part 2-2 - Type of equipment: CF; defibrillator-proof - Nominal frequency: 450 kHz - Pulse frequency: 30 kHz - HF output power - Type of current - Weight: 8.6 kg - Interference suppression : Limits in conformity with EN 55011, - Interference immunity in conformity with IEC 801 - Dimensions: 405 mm x 155 mm x 380 mm (W x H x D) - CE marking conform with 93/42/EEC - Unit ENDO, 220-240 V <p>Consisting of: •</p>	01	<p>Karl Storz</p> <p>Yes</p> <p>Yes</p> <p>-</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>	<p>Richard Wolf</p> <p>Yes</p> <p>-</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>	<p>Tekno</p> <p>Yes</p> <p>Not mentioned</p> <p>Not mentioned</p> <p>Yes</p> <p>Yes</p> <p>Not mentioned</p> <p>Not mentioned</p> <p>Yes</p>

[Handwritten signature]


[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
	<ul style="list-style-type: none"> - Double-pedal foot switch - Rubber neutral electrode for adults single 15X26 cm - Non-stick Red forceps, Straight, Blunt. 0.6 mm, 20 cm - Bipolar Cable - Reusable Electrode Handle with two push buttons - Electrode (Set of 5 pcs) - Trolley (Local - Best Quality) <p>FDA/CE/JIS Marked OR Equivalent</p> <p>USA/UK/Europe /Japan/Germany</p>		<p>Yes Yes Not mentioned Yes Not mentioned Yes Yes</p> <p>CE</p> <p>Germany</p>	<p>Yes Yes Not mentioned Yes Not mentioned Yes Yes</p> <p>CE</p> <p>Germany</p>	<p>Yes Not mentioned Not mentioned Yes Not mentioned Not mentioned Yes</p> <p>CE</p> <p>Germany</p>
Remarks					
			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Does not comply with the required specifications and criteria.
			Responsive	Responsive	Non-Responsive



(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad



(Member)
Dr. Jawed Altraf Jat
Head of Urology Department
LUMHS
Jamshoro



(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

S. No.	Item Description	Qty	M/s Verizon	M/s Onctech Corporation
	<p>Pneumatic probe, Ø 1.0 mm, length 605 mm Pneumatic probe, Ø 0.8 mm, length 605 mm Pneumatic probe Ø 1.0 mm, length 570 mm for combination with the ultrasound probes Ø 3.3 and 3.8 mm, length 403 mm (FR-084 I FR-083) Pneumatic probe, Ø 1.0 mm length 497 mm for combination with the ultrasound probes Ø 3.3 mm and Ø 3.8 mm, length 330 mm (FR-105/FR-106)</p> <p>Ultrasound Probes (2Probes at Choice) Consisting of: Probe Ø 3.8 mm, Length 403 mm, for ultrasound handpiece OR Probe Ø 3.3 mm, length 403 mm for ultrasound handpiece OR Probe Ø 3.3 mm, length 330 mm for ultrasound handpiece OR Probe Ø 3.8 mm, length 330 mm for ultrasound handpiece</p> <p>Compressed Air Tubing Compressed Air tube 3.0 m for the compressor</p> <p>Main Cord Main Cord Europe</p> <p>Litho Vac LV3 Kit Consisting of: Litho Vac suction tube Ø 4.0mm (12FR) Litho Vac suction tube Ø 4.0mm (12FR) Litho Vac suction tube Ø 4.0 x 335 mm (compatible with EL-044) Litho Vac suction tube Ø 3.5 (10.5FR) Litho Vac suction tube 3.5 x 380 mm (compatible with EL-081) Probe Ø 1.6 x 453 mm Litho Vac suction tube Ø 1.6 (4.8FR) Litho Vac suction tube 1.6 x 595 mm (compatible with EL-080) Probe Ø 0.8 x 380 mm</p> <p>Litho Pump</p>		<p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>	<p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>

[Handwritten signature]

[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Verizon	M/s Onctech Corporation
	Litho Pump control unit, Consisting of: Compressed air connection Connection set Set of connectors for exhaust Overflow protection filters at choice With Medela connection (Pack of 10) OR With conical connection Double reusable suction jar set Compressor 220V All-In-One Cart Instruction Manual FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes Yes Yes Yes Yes Yes Yes CE	Yes Yes Yes Yes Yes Yes Yes CE
Remarks			Germany Complies with the required items, its specifications and criteria.	Switzerland Complies with the required items, its specifications and criteria.

Dr. Naeem Zia Memon

(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad

Dr. Jawed Altaf Jat

(Member)
Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro

Dr. Anjum Rehman


(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
14.	<p>Fiber Optic CystoUrethro Fiberscope</p> <p>1- Cysto-Urethro-Fiberscope, Deflection of distal tip 210 °/140 °, Direction of view 0 °, Angle of view 110 °, Working channel inner diameter 7 fr., Sheath size 15.5 fr., Working length 37 cm Consisting of: Cysto-Urethro-Fiberscope Case Grasping Forceps Biopsy Forceps Leakage Tester Pressure Compensation Cap Cleaning Brush LUER-Adaptor</p> <p>2- Stone Basket, 5 Fr., length 60 cm Consisting of: 3-Ring Handle Basket Coil</p> <p>3- Coagulating Electrode, unipolar, 4 Fr., length 73 cm</p> <p>4- Seal, for Instrument Ports</p> <p>5- Plastic Container for Sterilization, specially suited for gas and hydrogen peroxide (Sterrad ®) sterilization and storage, perforated, with lid, for use with flexible endoscopes up to max. 39 cm working length, external dimensions (w x d x h): 690 x 170 x 92 mm</p> <p>FDA/CE/JIS Marked</p>	01	Karl Storz Yes Yes Yes Yes Yes Yes Yes Yes	Richard Wolf Yes Yes Yes Yes Yes Yes Yes Yes


[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
	OR Equivalent USA/UK/Europe /Japan/Germany		CE Germany	CE UK
Remarks				
			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.
			Responsive	Responsive


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad


 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro



 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
15.	<p>VIDEO CYSTOSCOPE</p> <p>1- Video Cysto-Urethroscope, steerable, working channel inner diameter 6.5 Fr., direction of view 0°, angle of view 120°, sheath size 16 Fr., working length 37 cm Following accessories are included: Video Cysto-Urethroscope Case Grasping Forceps Biopsy Forceps Pressure Compensation Cap Leakage Tester Cleaning Brush LUER-Adaptor</p> <p>2- Cleaning Adaptor, for Instrument Ports</p> <p>3- Monitor for Endoscopes, Kit, screen size 7" with 1280 x 800 pixel resolution, two camera inputs, a USB and a HDMI port, optimized user interface, video and image capturing in real time on SD card, playback of recorded video clips and J Still images, data transfer from SD card to USB flash drive possible, splash-proof according to IP54, suitable for wipe disinfection, shock-resistant ABS plastic housing, intelligent power management with rechargeable Li-Ion batteries, VESA 75 mounting option, power adaptor for EU, UK, USA and Australia, power supply 110 - 240 VAC, 50/60 Hz Consisting of: Monitor Protection Cap VESA 75 Quick Clip Power Supply Set</p> <p>4- Stand For Mounting Monitor stainless steel made (Locally Supply)</p>	01	Karl Storz Yes Yes Yes	Richard Wolf Yes Yes Yes

[Handwritten signature]

[Handwritten signature]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
5-	Plastic Container for Flexible Endoscopes, suitable for gas and hydrogen peroxide (Sterrad®) sterilization and storage, external dimensions (w x d x h): 550 x 260 x 90 mm, for use with a flexible endoscope FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes Yes CE Germany	Yes Yes CE Germany
Remarks			Complies with the required items, its specifications and criteria. Responsive	Complies with the required items, its specifications and criteria. Responsive


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad


 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro



 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
16.	<p>Video URS</p> <p>1- Video Uretero-Renoscope Kit steerable, compatible with HD, Working channel 3.6 Fr., Direction of view: 0 °, Angle of view: 90 °, Working length: 70 cm, Sheath size: 8.5 Fr., Following accessories are included in delivery: Case Pressure Compensation Cap Leakage Tester Cleaning Brush LUER-Adaptor, with seal</p> <p>2- HD module, for use with up to 3 link modules, resolution 1920 x 1080 pixels, with integrated and digital Image Processing Module, power supply 100 - 120 VAC/200 - 240 VAC, 50/60 Hz, Consisting of: Mains Cord, length 300 cm DVI-D Connecting Cable, length 300 cm Connecting Cable, length 100 cm USB Flash Drive, USB Silicone Keyboard, with touchpad, DE</p> <p>3- HD link module, for use with flexible video endoscopes and one-chip camera heads (up to FULL HD), power supply 100 - 120 VAC/200 - 240 VAC, 50/60 Hz, Consisting of: Mains Cord, length 300 cm Link Cable, length 20 cm</p> <p>4- Video Endoscope Adaptor, color systems PAL/NTSC, length 60 cm</p> <p>5- 27" FULL HD Monitor, color systems PAL/NTSC, max. screen resolution 1920 x 1080,</p>	01	Karl Storz Yes	Richard Wolf Yes


Handwritten signature

Handwritten signature

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon
	image format 16:9, Interface: RS 232, power supply 85 - 264 VAC, 50/60 Hz, wall mount with VESA 100 adaptor Including: External 24 VDC Power Supply Mains Cord		Yes	Yes
	6- Equipment Cart (Locally Supply) FDA/CE/JIS Marked OR Equivalent USA/UK/EEC/Japan/Germany		Yes CE Germany	Yes CE Germany
Remarks			Responsive	Responsive


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad



 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro



 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s Noor International
17.	FULL HD Camera System	01	Karl Storz	Richard Wolf	Tekno	Stema
1-	HD module, for use with up to 3 link modules, resolution 1920 x 1080 pixels, with integrated and digital Image Processing Module, power supply 100 - 120 VAC/200 - 240 VAC, 50/60 Hz, Consisting of: Mains Cord, length 300 cm DVI-D Connecting Cable, length 300 cm Connecting Cable, length 100 cm USB Flash Drive, 32 GB USB Silicone Keyboard, with touchpad, US		Yes	Yes	Yes	Yes
2-	HD module, for use with FULL HD threechip camera heads, power supply 100 - 120 VAC/200 - 240 VAC, 50/60 Hz, Consisting of: Mains Cord, length 300 cm Link Cable, length 20 cm		Yes	Yes	Yes	Yes
3-	Three-Chip FULL HD Camera Head, progressive scan, soakable, gas- and plasma-sterilizable, with integrated Parfocal Zoom Lens, focal length $f = 15 - 31$ mm (2x), 2 freely programmable camera head buttons		Yes	Yes	Yes	Yes
4-	LED 175 Watt light source high-performance LED and one light outlet, power supply 110 - 240 VAC, 50/60 Hz Including: Mains Cord Connecting Cable		Yes	Yes	Yes	Yes
5-	Fiber Optic Light Cable, with straight connector, extremely heat-resistant, enhanced light transmission, diameter 4.8 mm, length 250 cm		Yes	Yes	Yes	Yes
6-	27" FULL HD Monitor, color systems PAL/NTSC, max. screen		Yes	Yes	Not Full HD Fiber Optic	Not Full HD Fiber Optic

[Handwritten signatures and initials]

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s For International
	resolution 1920 x 1080, image format 16:9, Interface: RS 232, power supply 85 - 264 V.A.C., 50/60 Hz, wall mount with VESA 100 adaptor Including: External 24 VDC Power Supply Mains Cord 7- Equipment Cart (Locally Supply) FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes Yes CE Germany	Yes Yes CE Germany	Yes Yes CE Germany	26" Yes CE Germany
	Remarks		Complies with the required items, its specifications and criteria. Responsive	Complies with the required items, its specifications and criteria. Responsive	Does not comply with the required specifications and criteria. Non-Responsive	No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied. Non-Responsive


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad


 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro


 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Ontech Corporation
18.	<p>HOLMIUM Laser System for Urology Applications</p> <p>1- System, with power supply 230 VAC, 50/60 Hz Special Features Laser power should be 20 watt (portable system) High Mobility & Flexibility Due To The Desktop Design Easy To Operate Easy To Maintain Small Diameter Fibers Low Stress For Patient Short Treatment Time Least Possible Tissue Damage High Flexibility Fibers For Use With Rigid & Flexible Power consumption should not exceed by 1800VA Dimension of the equipment 610x250x430 with all necessary accessories</p> <p>2- Fiber Kit Basic Set Including: 3 pcs. CALCULASE II Fiber 230 µm, reusable 3 pcs. CALCULASE II Fiber 365 µm, reusable 3 PCS. CALCULASE II Fiber 600 µm, reusable</p> <p>Optional Accessories 3- Fiber Cutter, for use with fibers 4- Fiber Stripper, for use with fibers 5- Fiber Stripper Set, sterilizable, including Silicone Pad, Ceramic Knife, Fiber Stripper 230 µm, Fiber Stripper 365 µm, Fiber Stripper 600 µm.</p>	01	<p>Karl Storz</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>	<p>Quanta</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>	<p>EMS</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p> <p>Yes</p>

[Handwritten signature]

[Handwritten signature]


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Czech Corporation
	FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		CE Germany Complies with the required items, its specifications and criteria.	CE Germany Complies with the required items, its specifications and criteria.	CE Switzerland Complies with the required items, its specifications and criteria.
Remarks					
Responsive					


(Chairman)

Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital
Hyderabad


(Member)

Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS
Jamshoro



(Member)


Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital
Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
19.	Peads Nephroscope	01	Karl Storz	Richard Wolf	Tekno
	1- Straight Forward Telescope 6 °, with angled eyepiece, autoclavable, with instrument channel 5 Fr., fiber optic light transmission incorporated, Color code: green		Yes	Yes	Yes
	2- Telescoping Dilation Set, consisting of: set of 3 dilators, 9, 12 and 15 Fr., with 2 rigid and 2 flexible guide rods		Yes	Yes	Yes
	3- Operating Sheath, 17 Fr., including connecting tubes for in- and outflow and 2 LUER-Lock adaptors		Yes	Yes	Yes
	4- Hollow Obdurate and Fascial Dilator		Yes	Yes	Yes
	5- Grasping Forceps for stone fragments, double action jaws, 5 Fr., length 30 cm		Yes	Yes	Yes
	6- Grasping Forceps for larger stones and fragments, double action jaws, 5 Fr., length 30 cm		Yes	Yes	Yes
	7- Biopsy Forceps. double action jaws, 5 Fr., length 30 cm		Yes	Yes	Yes
	8- LED Nova 150, High-Performance LED Cold Light Fountain with one light outlet, power supply 100 - 240 VAC, 50/60 Hz Including: Main cords		Yes	Yes	Yes
	9- Fiber Optic Light Cable, with straight connector, diameter 3.5 mm, length 180 cm		Yes	Yes	Yes




S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
	FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		CE Germany	CE Germany	CE Germany
Remarks					
			Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non Responsive


 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad


 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro


 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise
20.	<p>URS PAEDS</p> <p>1- Pediatric Uretero-Renoscope 7.3 Fr., 6Å°, angled eyepiece, conical, one-step, 7.3 - 7.6 Fr., working length 25 cm, autoclavable, fiber optic light transmission incorporated, 2 lateral irrigation ports and 1 working channel 3.6 Fr. for instruments up to 3 Fr., Following accessories are included: Instrument Port with Sealing System and Quick Release Lock Seal LUER-LockTube Connector Flow Control Stopcock Cleaning Adaptor, Insertion Aid Tray</p> <p>2- Grasping Forceps for stone fragments, double action jaws, 3 Fr., length 60 cm</p> <p>3- LED Nova 150, High-Performance LED Cold Light Fountain with one light outlet, power supply 100 - 240 VAC, 50/60 Hz Including: Mains cord</p> <p>4- Fiber Optic Light Cable, with straight connector, diameter 3.5mm, length 180cm</p> <p>FDA/CE/JIS MarkedOR Equivalent USA/UK/Europe /Japan/Germany</p>	01	Karl Storz Yes	Richard Wolf Yes Yes Yes	Tekno Yes Yes Yes
			Complies with the required items, its specifications and criteria. Responsive	Complies with the required items, its specifications and criteria. Responsive	Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied. Non-Responsive
			Responsive	Responsive	Non-Responsive

[Signature]

Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital, Hyderabad


[Signature]

Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS, Jamshoro

[Signature]

Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital, Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/oor International
21.	Peads Cystoscope 1- Straight Forward Tele- scope 0°, ø 1.9/2.1 mm, autoclavable, fiber optic light transmission incorporated, 2- Cystoscope-Urethroscope Sheath, 9.5 Fr., working length 14 cm, with 4 Fr. working channel, with Obturator and 2 LUER-Lock 3- Cystoscope-Urethroscope Sheath 11 Fr., working length 14 cm, with working channel 5 Fr., with Obturator and 2 LUER-Lock connectors. 4- Grasping Forceps double action jaws, flexible, 3 Fr., length 28 cm 5- Biopsy Forceps, 3 Fr., double action jaws, flexible, length 28 cm 6- Coagulating Electrode, 3 Fr., unipolar, length 53 cm 7- Needle Electrode, 3 Fr., unipolar, length 53 cm FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany	01	Karl Storz Yes Yes Yes Yes Yes Yes Yes CE Germany	Richard Wolf Yes Yes Yes Yes Yes Yes Yes CE Germany	Tekno Yes Yes Yes Yes Yes Yes Yes CE Germany	Stema Yes Yes Yes Yes Yes Yes Yes CE Germany
	Remarks		Complies with the required items, its specifications and criteria.	Complies with the required items, its specifications and criteria.	Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.	No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied.
			Responsive	Responsive	Non Responsive	Non-Responsive



(Chairman)

Dr. Naeem Zia Memon

Additional Medical Superintendent

Liaquat University Hospital, Hyderabad



(Member)

Dr. Jawed Altaf Jat

Head of Urology Department

LUMHS, Jamshoro



(Member)

Dr. Anjum Rehman

Deputy Medical Superintendent


Liaquat University Hospital, Hyderabad


S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s Moor International
22.	Urethrotome Peads	01	Karl Storz	Richard Wolf	Tekno	Stema
1-	Straight Forward Telescope 0°, enlarged view, ø 4 mm, autoclavable, fiber optic light transmission		Yes	Yes	Yes	Yes
2-	Working Element, motion by means of a spring. The thumb support is movable. In rest position the electrode is inside the sheath.		Yes	Yes	Yes	Yes
3-	Resectoscope Sheath with LUER-Lock stopcock, including connecting tube for inflow, 11 Fr. and obturator		Yes	Yes	Yes	Yes
4-	Urethrotome Sheath with LUER-Lock stopcock, 10 Fr., with obturator and 2 LUER- lock adaptors		Yes	Yes	Yes	Yes
5-	Telescope Bridge with 1 lockable instrument channel		Yes	Yes	Yes	Yes
6-	Cold Knife straight This article is only available in units of 6		Yes	Yes	Yes	Yes
7-	Cold Knife round This article is only available in units of 6		Yes	Yes	Yes	Yes
8-	Cold Knife sickle-shaped This article is only available in units of 6		Yes	Yes	Yes	Yes
9-	Cold Knife hook-shaped This article is only available in units of 6		Yes	Yes	Yes	Yes
10-	Cutting Loop angled, color code: green This article is only available		Yes	Yes	Yes	Yes




S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s For International
	in units of 6 FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany		Yes CE Germany Complies with the required items, its specifications and criteria.	Yes CE Germany Complies with the required items, its specifications and criteria.	Yes CE Germany Limited references. Not much feedback on material quality from urology users. Neither complies required criteria nor is the committee satisfied.	Yes CE Germany No trained person / workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied.
Remarks			Responsive	Responsive	Non-Responsive	Non-Responsive



 (Chairman)
Dr. Naeem Zia Memon
 Additional Medical Superintendent
 Liaquat University Hospital
 Hyderabad



 (Member)
Dr. Jawed Altaf Jat
 Head of Urology Department
 LUMHS
 Jamshoro


 (Member)
Dr. Anjum Rehman
 Deputy Medical Superintendent
 Liaquat University Hospital
 Hyderabad

S. No.	Item Description	Qty	M/s Allmed Solution	M/s Verizon	M/s Universal Enterprise	M/s Noor International
23.	Peads Resectoscope 1- Straight Forward Tele- scope 0°, ø 1.9/2.1 mm. autoclavable, fiber optic light transmission incorporated, 2- Resectoscope Sheath with LUER-Lock stopcock, including connecting tube for inflow, 11 Fr. and obturator. 3- Working Element, motion by means of a spring. The thumb support is movable. In rest position the electrode is inside the sheath. 4- Cutting Loop angled, color code: black This article is only available in units of 6 5- Telescope Bridge with 1 lockable instrument channel FDA/CE/JIS Marked OR Equivalent USA/UK/Europe /Japan/Germany	01	Karl Storz Yes Yes Yes Yes Yes CE Germany	Richard Wolf Yes Yes Yes Yes Yes CE Germany	Tekno Yes Yes Yes Yes Yes CE Germany	Stema Yes Yes Yes Yes Yes CE Germany
	Remarks		Complies with the required items, its specifications and criteria. Responsive	Complies with the required items, its specifications and criteria. Responsive	No factory trained person / workshop for surgical instrument. Does not comply with the required criteria and satisfaction of the committee. Non-Responsive	No trained person/ workshop for surgical instrument. Limited references. Neither complies required criteria nor is the committee satisfied. Non-Responsive


(Chairman)
Dr. Naeem Zia Memon
Additional Medical Superintendent
Liaquat University Hospital, Hyderabad


(Member)
Dr. Jawed Altaf Jat
Head of Urology Department
LUMHS, Jamshoro


(Member)
Dr. Anjum Rehman
Deputy Medical Superintendent
Liaquat University Hospital, Hyderabad

LIAQUAT UNIVERSITY HOSPITAL

HYDERABAD / JAMSHORO

BID FINANCIAL EVALUATION REPORT

OPENED ON 24.04.2018

PURCHASE OF MACHINERY/EQUIPMENT UNDER SNE/ADP FOR THE YEAR 2017-18 (RESCUE CENTRE) (UROLOGY DEPARTMENT) (PART-I)

TENDER NO. 34/2017-18

(DATED: 27.02.2018)

Enclosed:

- 1) Bid Evaluation Report**
- 2) Comparative Statement**
- 3) Meeting Minutes**

LIAQUAT UNIVERSITY HOSPITAL

HYDERABAD / JAMSHORO

BID FINANCIAL EVALUATION REPORT

OPENED ON 24.04.2018

PURCHASE OF MACHINERY/EQUIPMENT UNDER SNE/ADP FOR THE YEAR 2017-18 (RESCUE CENTRE) (UROLOGY DEPARTMENT) (PART-I)

TENDER NO. 34/2017-18









(DATED: 27.02.2018)

Enclosed:

- 1) Bid Evaluation Report**
- 2) Comparative Statement**
- 3) Meeting Minutes**

MEETING OF CENTRAL PROCUREMENT COMMITTEE FOR OPENING OF FOLLOWING FINANCIAL BIDS FOR PURCHASE OF MACHINERY / EQUIPMENT UNDER ADP / SNE FOR THE FINANCIAL YEAR 2017-18 FOR LIAQUAT UNIVERSITY HOSPITAL, HYDERABAD/ JAMSHORO HELD ON 24.04.2018 AT THE OFFICE OF VICE CHANCELLOR DOW UNIVERSITY OF HEALTH SCIENCES OJHA CAMPUS KARACHI.




1. Purchase of Machinery / Equipment for Establishment of Cardiothoracic Surgery Department at LUH, Hyderabad.
2. Purchase of Machinery / Equipment for Rescue Center (Part-I), Urology Department.
3. Purchase of Machinery / Equipment for Rescue Center (Part-II), Neuro Surgery Deptt.
4. Purchase of Laundry Machine.


S.NO	NAME & DESIGNATION		SIGNATURE
01	Professor Dr.M.Saeed Qureshi Vice Chancellor, Dow University of Health Sciences, Karachi.	Chairman	
02	Professor Dr. Tariq Rafi, Vice Chancellor, Jinnah Sindh Medical University, Karachi.	Member	
03	Professor Dr. Muneer Junejo Pro Vice Chancellor, LUM&HS, Jamshoro.	Member	
04	Professor Aneela Ataur Rehman Pro Vice Chancellor, Dean of Faculty Community Medicine, LUMHS, Jamshoro	Member	
05	Vice Chancellor, Chandka Medical University, Larkana.	Member	
06	Mr. Shafi Mohammad Qureshi Additional Director, Directorate of Pharmacy, Health Department, Government of Sindh, Karachi.	Member	
07	DR. JALAL UDDIN JALALANI Additional Secretary (PM&I), Health Department, Government of Sindh, Karachi.	Member	
08	AT H A R M A N Z O O T M e e s a n i Additional Finance Secretary, Finance Department, Government of Sindh, Karachi.	Member	
09	MR. TAHIR. Additional Director NAB NAB Islamabad.	Member	

Dated:- 24.04.2018

Opening of Financial Bids for Purchase of Machinery / Equipment
For Rescue Center (Part-I) Urology Department
Financial year 2017-18

ATTENDANCE SLIP

S.No.	Name of Firm	Phone Number	Postal Address	Name of Participant	C.N.I.C No: of participants	Signature
1	ALLMED SOLUTIONS	03312153212 03002545977		M.OMAR EJAZ	42101-85553767	
2	Onitech Corporation	0342-2606250		M. Hassam Khan	42101-2922753-3	
3	Verizon	03422438328		Faiz ullah	42201-1789203	


Member

Member


Member

Member

Member


Member

Prof Anwar Atta-ur-Rahman


Member


Member


Member


Chairman


Member

Member


Member

BID EVALUATION REPORT


- 1) Name of Procuring Agency: Liaquat University Hospital Hyderabad / Jamshoro
- 2) Tender Reference No: 34/2017-18
- 3) Tender Description: Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)
- 4) Method of Procurement: Single Stage Two Envelope
- 5) Tender Published: Dawn – Jang – Awami Awaz , Nai Bat
SPPRA Website S.No:36633
- 6) Total Bid Documents Sold: 07
- 7) Total Bids Received: 07
- 8) Technical Bid Opening Date: 27.02.2018
- 9) No. of Bid Technically Qualified: 03
- 10) Bid(s) Rejected: 04
- 11) Financial Bid Opening Date: 24.04.2018
- 12) Bid Evaluation Report: **Enclosed (i) Comparative Statement (ii) Minutes**



Prof. Dr. M. Saeed Qureshi
Vice Chancellor
DUHS, Karachi




Prof. Dr. Fatiq Rafi
Vice Chancellor
JSMU, Karachi




Prof. Dr. Muneer Junejo
Pro Vice Chancellor
LUMHS, Jamshoro

Vice Chancellor
CMU,
Larkana



Prof. Anceela Ataur Rehman
Pro Vice Chancellor
LUMHS, Jamshoro




Mr. Shafi Mohammad Qureshi
Addl: Director, Directorate of Pharmacy
Health Department, GoS, Karachi



Additional Secretary (PM&I)
Health Department,
GoS, Karachi



Additional Finance Secretary
Finance Department, GoS,
Karachi



Additional Director NAB
National Accountability Bureau
Islamabad

Tender No. 34/2017-18


Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)
(Dated: 27.02.2018)


Financial Bids Opened on 24.04.2018

COMPARATIVE STATEMENT

S. No.	Item Description	Qty	Bidders	Make	C&F (Foreign Currency)	C&F (Equivalent to Pak Rupees)	DDP (Pak Rupees)	Remarks
1	Minimal Invasive Percutaneous Nephroscope (XS/S)	01	M/s Verizon	Richard Wolf	€ 12,500	Rs 1,715,875	Rs 1,920,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 28,225	Rs 3,874,446	Rs 4,233,750	2nd Lowest
2	Minimal Invasive Percutaneous Nephroscope (M)	01	M/s Verizon	Richard Wolf	€ 7,600	Rs 1,043,252	Rs 1,180,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 13,735	Rs 1,885,403	Rs 2,060,250	2nd Lowest
3	Minimal Invasive Percutaneous Nephroscope (L)	01	M/s Verizon	Richard Wolf	€ 4,600	Rs 631,442	Rs 720,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 10,428	Rs 1,431,452	Rs 1,564,200	2nd Lowest
4	Nephroscope Adult	01	M/s Verizon	Richard Wolf	€ 6,500	Rs 892,255	Rs 985,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 12,070	Rs 1,656,849	Rs 1,810,500	2nd Lowest
5	URS Adult	01	M/s Verizon	Richard Wolf	€ 5,700	Rs 782,439	Rs 798,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 11,893	Rs 1,632,552	Rs 1,783,950	2nd Lowest
6	Cystoscope Adult	01	M/s Verizon	Richard Wolf	€ 4,500	Rs 617,715	Rs 700,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 9,219	Rs 1,265,492	Rs 1,382,850	2nd Lowest
7	Stone Punch	01	M/s Verizon	Richard Wolf	€ 4,700	Rs 645,169	Rs 726,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 9,412	Rs 1,291,985	Rs 1,411,800	2nd Lowest
8	Lithorite	01	M/s Verizon	Richard Wolf	€ 6,500	Rs 892,255	Rs 990,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 11,631	Rs 1,596,587	Rs 1,744,650	2nd Lowest












S. No.	Item Description	Qty	Bidders	Make	C&F (Foreign Currency)	C&F (Equivalent to Pak Rupees)	DDP (Pak Rupees)	Remarks
9	Urethrotome Adult	01	M/s Verizon	Richard Wolf	€ 5,600	Rs 768,712	Rs 630,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 9,051	Rs 1,242,431	Rs 1,357,650	2nd Lowest
10	Resectoscope Adult (Monoploar)	01	M/s Verizon	Richard Wolf	€ 4,100	Rs 562,807	Rs 645,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 8,471	Rs 1,162,814	Rs 1,270,650	2nd Lowest
11	Bipolar Resectoscope Adult	01	M/s Verizon	Richard Wolf	€ 5,350	Rs 734,395	Rs 835,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 36,699	Rs 5,037,672	Rs 5,504,850	2nd Lowest
12	Electro Surgery Unit 400 Watts	01	M/s Allmed Solution	Karl Storz SE & CO KG	€ 19,569	Rs 2,686,237	Rs 2,935,350	Lowest & best evaluated
13	Ultrasonic Intra-Corporeal Lithotripter	01	M/s Ontech Corporation	EMS	€ 42,225	Rs 5,796,226	Rs 7,458,530	Lowest & best evaluated
14	Fiber Optic CystoUrethro Fiberscope	01	M/s Verizon	Richard Wolf	€ 10,100	Rs 1,386,427	Rs 1,570,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 16,198	Rs 2,223,499	Rs 2,429,700	2nd Lowest
15	Video Cyst Scope	01	M/s Allmed Solution	Karl Storz SE & CO KG	€ 20,147	Rs 2,765,579	Rs 3,022,050	1st Lowest
			M/s Verizon	Richard Wolf	€ 20,200	Rs 2,772,854	Rs 3,160,000	2nd Lowest
16	Video URS	01	M/s Verizon	Richard Wolf	€ 47,875	Rs 6,571,801	Rs 7,500,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 59,994	Rs 8,235,376	Rs 8,999,100	2nd Lowest
17	Full HD Camera System	01	M/s Verizon	Richard Wolf	€ 39,500	Rs 5,422,165	Rs 6,183,000	1st Lowest
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 56,942	Rs 7,816,428	Rs 8,541,300	2nd Lowest
18	Holmium Laser System For Urology Applications	01	M/s Verizon	Quanta	€ 29,000	Rs 3,980,830	Rs 4,480,000	1st Lowest
			M/s Ontech Corporation	EMS	€ 53,000	Rs 7,275,310	Rs 9,000,000	
			M/s Allmed Solution	Karl Storz SE & CO KG	€ 81,571	Rs 11,197,251	Rs 12,235,650	2nd Lowest
19	PeadsNephroscope	01	M/s Verizon	Richard Wolf	€ 8,200	Rs 1,125,614	Rs 1,276,000	1st Lowest
			M/s Allmed Solution	Karl Storz	€ 10,371	Rs 1,423,627	Rs 1,555,650	2nd Lowest

S. No.	Item Description	Qty	Bidders	Make	C&F (Foreign Currency)	C&F (Equivalent to Pak Rupees)	DDP (Pak Rupees)	Remarks
20	URS Paeds	01	M/s Verizon M/s Allmed Solution	SE & CO KG Richard Wolf Karl Storz SE & CO KG	€ 6,800 € 11,018	Rs 933,436 Rs 1,512,441	Rs 1,055,000 Rs 1,652,700	1st Lowest 2nd Lowest
21	PeadsCystoscope	01	M/s Verizon M/s Allmed Solution	Richard Wolf Karl Storz SE & CO KG	€ 4,500 € 7,158	Rs 617,715 Rs 982,579	Rs 705,000 Rs 1,073,700	1st Lowest 2nd Lowest
22	UrethrotomePeads	01	M/s Verizon M/s Allmed Solution	Richard Wolf Karl Storz SE & CO KG	€ 5,600 € 7,191	Rs 768,712 Rs 987,109	Rs 872,000 Rs 1,078,650	1st Lowest 2nd Lowest
23	PeadsResectoscope	01	M/s Verizon M/s Allmed Solution	Richard Wolf Karl Storz SE & CO KG	€ 4,400 € 6,217	Rs 603,988 Rs 853,408	Rs 691,000 Rs 932,550	1st Lowest 2nd Lowest

Exchange Rate: € : Euro 1= 137.27 PKR


Prof. Dr. M. Saeed Qureshi
 Vice Chancellor
 DUHS, Karachi



Prof. Dr. Farid Rafi
 Vice Chancellor
 JSMU, Karachi


Prof. Dr. Muneer Junejo
 Pro Vice Chancellor
 LUMHS, Jamshoro


Prof. Anceela Ataur Rehman
 Pro Vice Chancellor
 LUMHS, Jamshoro


Vice Chancellor
 Chandka Medical University
 Larkana


Mr. Shafi Mohammad Qureshi
 Addl: Director, Directorate of Pharmacy
 Health Department, GoS, Karachi


Additional Secretary (PM&I)
 Health Department,
 GoS, Karachi


Additional Finance Secretary
 Finance Department, GoS,
 Karachi










Additional Director NAB
 National Accountability Bureau
 Islamabad

MEETING MINUTES OF THE PROCUREMENT COMMITTEE

Meeting of the Central Procurement Committee of Health Department was held on 24.04.2018 at the office of Vice Chancellor Dow University of Health Sciences Ojha Campus Karachi. Prof. Dr. M.Saeed Qureshi chaired the meeting. The financial bids were opened after finalization of technical bids by the central procurement committee for Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I).

Technical bids of Tender Enquiry No. 34/2017-18 were opened on 27.02.2018 and subsequently approved by the Procurement Committee for procurement of required items.

The following members attended the meeting:

S. No.	Name & Designation		Signature
01	Prof. Dr. M.Saeed Qureshi Vice Chancellor Dow University of Health Sciences Karachi	Chairman	
02	Prof. Dr. Tariq Rafi Vice Chancellor Jinnah Sindh Medical University Karachi	Member	
03	Prof. Dr. Muneer Junejo Pro Vice Chancellor Liaquat University of Medical and Health Sciences Jamshoro	Member	
04	Prof. Aneela Ataur Rehman Pro Vice Chancellor Dean of Faculty Community Medicine Liaquat University of Medical and Health Sciences Jamshoro	Member	
05	Vice Chancellor Chandka Medical University Larkana	Member	
06	Mr. Shafi Mohammad Qureshi Additional Director Directorate of Pharmacy Health Department Government of Sindh Karachi	Member	
07	Additional Secretary (PM&I) Health Department, Government of Sindh Karachi	Member	
08	Additional Finance Secretary Finance Department, Government of Sindh, Karachi	Member	
09	Additional Director NAB National Accountability Bureau Islamabad	Member	

Initiating the discussions, the evaluated technical data sheets of bidder was presented for perusal of the Central Procurement Committee (Copy of the Technical Evaluation Report enclosed.)

While financial evaluation, the requirement of Rule 48 of SPPRA was also observed. The financial bid of eligible and qualified firms with responsive technical offers are as under:

No.	Bidder Name	Bid Security
1	M/s Allmed Solutions	No. 01873726 Dated: 21-02-2018 Habib Metropolitan Bank Limited
2	M/s Verizon	No. 00001141 Dated: 23-02-2018 Summit Bank Limited
3	M/s Ontech Corporation	No. 04226147 Dated: 26-02-2018 Meezan Bank Limited

Tender No : **34/2017-18**
Purchase of Machinery / Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)

Item No : **01**

Name of Item : **Minimal Invasive Percutaneous Nephroscope (XS/S)**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 28,225/- each on C&F and Rs 4,233,750/- each on DDP basis for the said item. M/s Verizon offered the rates € 12,500/- each on C&F and Rs 1,920,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)

Item No : **02**

Name of Item : **Minimal Invasive Percutaneous Nephroscope (M)**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 13,735/- each on C&F and Rs 2,060,250/- each on DDP basis for the said item. M/s Verizon offered the rates of € 7,600/- each on C&F and Rs 1,180,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)

Item No : **03**

Name of Item : **Minimal Invasive Percutaneous Nephroscope (L)**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 10,428/- each on C&F and Rs 1,564,200/- each on DDP basis for the said item. M/s Verizon offered the rates of € 4,600/- each on C&F and Rs 720,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : 04

Name of Item : Nephroscope Adult

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 12,070/- each on C&F and Rs 1,810,500/- each on DDP basis for the said item. M/s Verizon offered the rates of € 6,500/- each on C&F and Rs 985,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : 05

Name of Item : URS Adult

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 11,893/- each on C&F and Rs 1,783,950/- each on DDP basis for the said item. M/s Verizon offered the rates of € 5,700/- each on C&F and Rs 798,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.


Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)


Item No : 06


Name of Item : Cystoscope Adult

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.







M/s Allmed Solutions offered the rates of € 9,219/- each on C&F and Rs 1,382,850/- each on DDP basis for the said item. M/s Verizon offered the rates of € 4,500/- each on C&F and Rs 700,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

**Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)**

Item No : 07

Name of Item : Stone Punch

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 9,412/- each on C&F and Rs 1,411,800/- each on DDP basis for the said item. M/s Verizon offered the rates of € 4,700/- each on C&F and Rs 726,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

**Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)**

Item No : 08

Name of Item : Lithorite

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 11,631/- each on C&F and Rs 1,744,650/- each on DDP basis for the said item. M/s Verizon offered the rates of € 6,500/- each on C&F and Rs 990,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

**Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)**

Item No : 09

Name of Item : Urethrotome Adult

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 9,051/- each on C&F and Rs 1,357,650/- each on DDP basis for the said item. M/s Verizon offered the rates of € 5,600/- each on C&F and Rs 630,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.



Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : 10

Name of Item : Resectoscope Adult (Monoploar)

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 8,471/- each on C&F and Rs 1,270,650/- each on DDP basis for the said item. M/s Verizon offered the rates of € 4,100/- each on C&F and Rs 645,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : 11

Name of Item : Bipolar Resectoscope Adult

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 36,699/- each on C&F and Rs 5,504,850/- each on DDP basis for the said item. M/s Verizon offered the rates of € 5,350/- each on C&F and Rs 835,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : 12

Name of Item : Electro Surgery Unit 400 Watts

Bidder	M/s Allmed Solution
Make	Karl Storz SE & CO KG

The offer of M/s Allmed Solutions meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 19,569/- each on C&F and Rs 2,935,350/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Allmed Solutions on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **13**

Name of Item : **Ultrasonic Intra-Corporeal Lithotripter**

Bidder	M/s Ontech Corporation
Make	EMS

The offer of M/s Ontech Corporation meets all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Ontech Corporation offered the rates of € 42,225/- each on C&F and Rs 7,458,530 /- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Ontech Corporation on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **14**

Name of Item : **Fiber Optic CystoUrethro Fiberscope**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 16,198/- each on C&F and Rs 2,429,700/- each on DDP basis for the said item. M/s Verizon offered the rates of € 10,100/- each on C&F and Rs 1,570,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **15**

Name of Item : **Video Cyst Scope**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 20,147/- each on C&F and Rs 3,022,050/- each on DDP basis for the said item. M/s Verizon offered the rates of € 20,200/- each on C&F and Rs 3,160,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Allmed Solutions on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **16**
Name of Item : **Video URS**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 59,994/- each on C&F and Rs 8,999,100/- each on DDP basis for the said item. M/s Verizon offered the rates of € 47,875/- each on C&F and Rs 7,500,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **17**
Name of Item : **Full HD Camera System**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 56,942/- each on C&F and Rs 8,541,300/- each on DDP basis for the said item. M/s Verizon offered the rates of € 39,500/- each on C&F and Rs 6,183,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)


Item No : **18**
Name of Item : **Holmium Laser System For Urology Applications**

Bidder	M/s Allmed Solution	M/s Verizon	M/s Ontech Corporation
Make	Karl Storz SE & CO KG	Quonta	EMS

The offers of M/s Allmed Solutions, M/s Verizon and M/s Ontech Corporation meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 81,571/- each on C&F and Rs 12,235,650/- each on DDP basis for the said item. M/s Verizon offered the rates of € 29,000/- each on C&F and Rs 4,480,000/- each on DDP basis for the said item. M/s Ontech Corporation offered the rates of € 35,150/- each on C&F and Rs 5,510,500/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.







Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **19**

Name of Item : **PeadsNephroscope**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 10,371/- each on C&F and Rs 1,555,650/- each on DDP basis for the said item. M/s Verizon offered the rates of € 8,200/- each on C&F and Rs 1,276,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **20**

Name of Item : **URS Paeds**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 11,018/- each on C&F and Rs 1,652,700/- each on DDP basis for the said item. M/s Verizon offered the rates of € 6,800/- each on C&F and Rs 1,055,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : **34/2017-18**
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18
(Rescue Centre) (Urology Department) (Part-I)

Item No : **21**


Name of Item : **PeadsCystoscope**

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 7,158/- each on C&F and Rs 1,073,700/- each on DDP basis for the said item. M/s Verizon offered the rates of € 4,500/- each on C&F and Rs 705,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.





Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)

Item No : 22

Name of Item : UrethrotomePeads

Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf

The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.

M/s Allmed Solutions offered the rates of € 7,191/- each on C&F and Rs 1,078,650/- each on DDP basis for the said item. M/s Verizon offered the rates of € 5,600/- each on C&F and Rs 872,000/- each on DDP basis for the said item.

The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.

Tender No : 34/2017-18
Purchase of Machinery/Equipment under SNE/ADP for the Year 2017-18 (Rescue Centre) (Urology Department) (Part-I)

Item No : 23

Name of Item : Peads Resectoscope


Bidder	M/s Allmed Solution	M/s Verizon
Make	Karl Storz SE & CO KG	Richard Wolf


The offers of M/s Allmed Solutions and M/s Verizon meet all the required specification and quality as per tender requisite. As approved by the Procurement Committee, the financial offers of said firms were opened.


M/s Allmed Solutions offered the rates of € 6,217/- each on C&F and Rs 932,550/- each on DDP basis for the said item. M/s Verizon offered the rates of € 4,400/- each on C&F and Rs 691,000/- each on DDP basis for the said item.


The Procurement Committee approved and accepted the offer of M/s Verizon on C&F basis.


The meeting ended with vote of thanks from the chair.



Prof. Dr. M. Saeed Qureshi
 Vice Chancellor
 DUHS, Karachi



Prof. Dr. Tariq Rafi
 Vice Chancellor
 JSMU, Karachi



Prof. Dr. Muneer Junejo
 Pro Vice Chancellor
 LUMHS, Jamshoro



Vice Chancellor
 CMU,
 Larkana


Prof. Anela Ataur Rehman
 Pro Vice Chancellor
 LUMHS, Jamshoro


Mr. Shafi Mohammad Qureshi
 Addl: Director, Directorate of Pharmacy
 Health Department, GoS, Karachi


Additional Secretary (PM&I)
 Health Department,
 GoS, Karachi


Additional Finance Secretary
 Finance Department, GoS,
 Karachi


Additional Director NAB
 National Accountability Bureau
 Islamabad

Additional Finance Secretary
 Finance Department, GoS,
 Karachi

Additional Director NAB
 National Accountability Bureau
 Islamabad