

**PHONE NO: 071-9310213 FAX 9310119.**

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR  
MEDICAL COLLEGE HOSPITAL SUKKUR**

NO: MS/GMCHS/ SUKKUR/

DATED

2018

To,

✓  
The Managing Director,  
SPPRA Karachi.

**SUBJECT: SUBMISSION OF BID EVALUATION REPORT OF NIT AT SERIAL  
NO.37244 OF GHULAM MUHAMMAD MAHAR MEDICAL  
COLLEGE HOSPITAL SUKKUR.**

It is stated that the complete Bid Evaluation Report along with following documents is submitted herewith for early hoisting on SPPRA Website.

7. BER Original.
8. Minutes of Meeting with attendance sheets.
9. Notifications of Committees.
10. Newspaper Cuttings.
11. Certificate of Single Bids.
12. Screenshot of SPPRA ID/Serial No.37243

This is for your kind information and for early hoisting.

*31/5/18*  
**MEDICAL SUPERINTENDENT  
GMMC HOSPITAL SUKKUR**

Copy Submitted with compliment:

1. The Secretary Government of Sindh Health Department Karachi for kind information.
2. The Additional Secretary (PM&I) Cell Health Department Government of Sindh Karachi for information.
3. The Additional Secretary (Development) Health Department Government of Sindh Karachi for information.
4. The District Accounts Officer Sukkur.

**MEDICAL SUPERINTENDENT  
GMMC HOSPITAL SUKKUR**

SPPRA INWARD DIARY

NO : 9604

DATED : 04 Feb 2018

**RING AGENCIES FACING TECHNICAL ISSUES WHEN POSTING BERS IN PPMS MAY SEND HARD COPIES COPIES TO SPPRA FOR POSTING OFFICIAL WEBSITE. PROCURING AGENCIES & CONTRACTORS ARE ADVISED TO VISIT BOTH PPMS & SPPRA OFFICIAL WEBSITE TO VIEW BERS OF PPMS NITs. For PPMS click here :: Procurement Performance Management System ::**

37244	2147483647	Ghulam	2018-03-12	2018-03-	13-03
		Muhammad	00:00:00	29	2018
		Mahar Medical College Hospital		23:59:59	
37243	2147483647	Ghulam	2018-03-12	2018-03-	13-03
		Muhammad	00:00:00	29	2018
		Mahar Medical College Hospital		23:59:59	
35828	2147483647	Ghulam	2017-12-22	2018-01-	22-12
		Muhammad	00:00:00	09	2017

**MINUTES OF MEETING REGARDING THE OPENING OF FINANCIAL PROPOSALS OF MACHINERY, EQUIPMENTS AND FURNITURE ADP SCHEME NAMEDLY 50 BEDDED MEDICAL & SURGICAL ICU WITH EXPANSION OF OTHER UNITS AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR FY 2017-18 HELD ON 30-5-2018 AT 9:00 AM AT OJHA CAMPUS DOW UNIVERSITY OF MEDICAL & HEALTH SCIENCES KARACHI.**

The Meeting started in the Holy name of Allah in the office of the Vice Chancellor Dow University of Medical & Health Sciences at Ojha Campus Karachi. The Chair welcomed all the participants.

It was briefed by the Medical Superintendent GMMMC Hospital Sukkur that the NIT was floated and published in the Newspapers such as Daily Dawn, Jang and Kawish and others on 23-2-2018 and 24-2-2018 and hoisted on SPPRA Website at Serial NO:37244. The Technical Proposals were finalized by the Technical Committee of GMMMC Hospital Sukkur. The same were submitted before Central Procurement Committee for the opening of Financial Proposals.

The Financial Proposals were opened in the presence of the Chairman and Members of Central Procurement Committee at Ojha Campus of Dow University of Medical & Health Sciences Karachi before all bidders. Attendance sheets are attached herewith.

The meeting proceeded with one by one opening of bids by signing and encircling rates which were also read aloud before committee and bidders. The Committee decided the verification of furniture and fixture on next date.

  
**MEMBER**

Prof. Dr. Tarique Rafi,  
Vice Chancellor, Jinnah Sindh  
Medical University Karachi

**MEMBER**


Prof. Dr. Muneer Junejo,  
Pro. Vice Chancellor Liaquat University of  
Medical & Health Science Jamshoro

  
**MEMBER**


Prof. Dr. Aneela Ataur Rehman  
Pro. Vice Chancellor, Dean of Faculty  
Community Medicines, Liaquat University  
of Medical & Health Science Jamshoro

  
**MEMBER**

Dr. Shafi Muhammad Qureshi  
Additional Director, Director of Pharmacy  
Government of Sindh Karachi

  
**MEMBER** (Mr. Jamaluddin Jaleelani)

Additional Secretary, (PM & I)  
Health Department Government of Sindh  
Karachi

  
**MEMBER** (Prof. Dr. Ghulam  
Asghar Channa)

Vice Chancellor,  
Chandka Medical University Larkana

  
**MEMBER** (Muhammad Tabir)

Additional Director NAB,  
NAB Islamabad

  
**MEMBER** (Mrs. Athar Mirani)

Additional Finance, Secretary (I.F)  
Finance Department Government of Sindh,  
Karachi

  
**CHAIRMAN**

Prof. Dr. Muhammad Saeed Qureshi  
Vice Chancellor

Dow University of Medical & Health Sciences, Karachi

Note:- Prof. Dr. Muneer Junejo, Pro Vice Chancellor, Liaquat University of Medical & Health Sciences, Jamshoro did not attend meeting. He is on Ex-Pakistan Leave.

Medical Superintendent  
GMMMC Hospital Sukkur  
M.S. GMMMC, Sukkur.


**BID EVALUATION REPORT**

- |                                      |   |
|--------------------------------------|---|
| 1. Name of Procuring Agency          | GMMC Hospital Sukkur.   |
| 2. Tender Reference No.              | Tender inquiry NO:INF/KRY.NO. 1253/2018.  |
| 3. Tender Description/Name of item   | Purchase of Machinery, Equipments And Furniture Under ADP Scheme (50 Bedded Medical & Surgical ICU With Expansion Of Other Units At Ghulam Muhammad Mahar Medical College Hospital Sukkur |
| 4. Method of Procurement             | Single Stage Two Envelope Procedure.  |
| 5. Tender Published                  | Published in Daily Dawn, Daily Moomal & Daily Kawish on 24.02.2018 , Daily Jang on 25.02.2018 and Corrigendum on 08.03.2018 on SPPRA SR.NO. 37244   |
| 6. Bid Documents Sold                | 29 (Twenty Nine)  |
| 7. Total Bids Received               | 18 (Eighteen)   |
| 8. Tender Opening Date (Technical)   | 29.03.2018  |
| 9. No. of Bids Technically Qualified | 13 (Thirteen)   |
| 10. Bids Rejected                    | 05 (Five)   |
| 11. Financial Bid Opening Date       | 30.05.2018  |


**12. Bid Evaluation Report**


S.No	Name of Firm/Bidder /Detail of Items	Cost offered by the Bidder	Ranking in terms of the cost	Comparison with Estimated Cost	Reasons for acceptance / Rejection	Remarks
00	01	02	03	04	05	
Comparative Statement Attached						


**The Committee has recommended to award the contract to the successful bidder**


  
**MEMBER**  
Prof. Dr. Tarique Rafi,  
Vice Chancellor, Jinnah Sindh Medical University  
Karachi


**MEMBER**  
Prof. Dr. Muneer Junejo,  
Pro. Vice Chancellor Liaquat University of Medical & Health Science Jamshoro

  
**MEMBER**  
Prof. Dr. Aneela Ataur Rehman  
Pro. Vice Chancellor, Dean of Faculty Community Medicines, Liaquat University of Medical & Health Science Jamshoro

  
**MEMBER**  
Dr. Shafi Muhammad Qureshi  
Additional Director, Director of Pharmacy Government of Sindh Karachi

  
**MEMBER** (Mr. Jamatuddin Jalalani)  
Additional Secretary, (PM & I)  
Health Department Govt. of Sindh Karachi

  
**MEMBER** (Prof. Dr. Ghulam Asghar Channas)  
Vice Chancellor  
Chandka Medical University Larkana

  
**MEMBER** (Muhammad Tahir)  
Additional Director NAB,  
NAB Islamabad

  
**MEMBER** (Mr. Athar Mirani)  
Additional Finance, Secretary (I.F)  
Finance Department Government of Sindh, Karachi

  
**CHAIRMAN**  
Prof. Dr. Muhammad Saeed Qureshi  
Vice Chancellor  
Dow University of Medical & Health Sciences, Karachi

Note:- Prof. Dr. Muneer Junejo, Pro Vice Chancellors, Liaquat University of Medical & Health Sciences did not attend meeting. He is on Ex-Pakistan Leave.

M.S. GMMC Hospital Sukkur




**COMPARATIVE STATEMENT OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU (UNDER ADP SCHEME) AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR THE FINANCIAL YEAR 2017-18 OPENED ON 30.05.2018 AT 09:00 A.M AT OJHA CAMPUS @ DOW UNIVERSITY OF MEDICAL & HEALTH SCIENCES KARACHI ANNEXURE (A)**

Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
1	ICU VENTILATOR	M/S BIOS Karachi	Evolution 3e Essential USA	2,472,000	1 <sup>st</sup> Lowest	M/s Bios Karachi Approved as 1st Lowest
2	MULTIPARAMETER MONITOR	M/S Platinum Corporation	PG M9000 / Progetti /Italy	449,500	1 <sup>st</sup> Lowest	M/s Platinum Corporation Approved as 1st Lowest
3	DEFIBRILLATOR	M/S Elate CC Karachi	EVO 1/Germany	680,000	1 <sup>st</sup> Lowest	M/s Elate CC approved as 1st Lowest
		M/S Zahid Sons	Defibgard 5000/Swiss	1,500,000	3rd Lowest	
		M/s Radiant Medical	M-Series/USA	694,000	2 <sup>nd</sup> Lowest	
4	INFUSION PUMP	M/S Platinum Corporation	PG 9071/Progetti/ Italy	199,500	1 <sup>st</sup> Lowest	M/s Platinum Corporation Approved as 1st Lowest
		M/S Medi-Urge	Agila VP/ France	211,000	2 <sup>nd</sup> Lowest	
5	ABG MACHINE	M/S S Ejazuddin	ABL80 FLEX CO-OX Analyzer	1,975,000	2 <sup>nd</sup> Lowest	M/s Platinum Corporation Approved as 1st Lowest
		M/S Platinum Corporation	USA	1,150,000	1 <sup>st</sup> Lowest	
6	ULTRASOUND MACHINE WITH PROBE	M/s Medequips SMC	USA. Aplio 300	8,500,000	1 <sup>st</sup> Lowest	M/s Medequips SMC approved as 1st Lowest
7	DOUBLE VACCUM SYSTEM	M/S Elate CC Karachi	UK	4,303,180	1st Lowest	M/s Elate CC Approved as 1st Lowest for both Items
		M/S Elate CC Karachi	UK			
9	CRASH TROLLEY	M/S Zahid Sons	MX37EMG/ Taiwan	220,000	1 <sup>st</sup> Lowest	M/s Zahid Sons Approved as 1st Lowest
10	PORTABLE X RAY MACHINE 100MA	M/S BIOS Karachi	Movix-4 E+ , France	2,590,000	2 <sup>nd</sup> Lowest	M/s Radiant Medical Approved as 1st Lowest
		M/s Radiant Medical	Roller 4, Italy	1,688,000	1 <sup>st</sup> Lowest	
11	LARYNGO SCOPE	M/s Hi-Tech Engineering	Local	17,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest
12	AMBO BAG SILICON	M/s Hi-Tech Engineering	Taiwan	20,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest






Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
13	B.P APPRATUS	M/s Hi-Tech Engineering	Japan	30,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest
14	FOOD TROLLEY	M/s Hi-Tech Engineering	Local	250,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest
15	BIPAP VENTILATORS	M/S Elate CC Karachi	DreamStar Due ST/France	300,000	1 <sup>st</sup> Lowest	M/s Elate CC Approved as 1st Lowest
		M/S Zahid Sons	Sleepcube Bilevel ST/Devilbiss/USA	350,000	2 <sup>nd</sup> Lowest	
		M/S Jafferi Medical	Seteler 150/Resmed/ Australia	795,000	3 <sup>rd</sup> Lowest	
16	WARMER	M/S Zahid Sons	AM-301 Japan	220,000	1 <sup>st</sup> Lowest	M/s Zahid Sons Approved as 1st Lowest
		M/S Radiant	Barkey S Line/ Germany	225,000	2 <sup>nd</sup> Lowest	
17	BLOOD WARMER	M/S Elate CC Karachi	Nouva 5/ Germany	300,000	3 <sup>rd</sup> Lowest	M/s Zahid Sons Approved as 1st Lowest
		M/S Zahid Sons	AM-301 Japan	220,000	1 <sup>st</sup> Lowest	
		M/s Radiant Medical	Barkey S Line/ Germany	225,000	2 <sup>nd</sup> Lowest	
		M/S Zahid Sons	KU-200/ Japan	120,000	2 <sup>nd</sup> Lowest	
18	NEUBILIZER MACHINE	M/s Hi-Tech Engineering	CA - MI , Italy	10,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest
19	ECG MACHINE 6 CHANNEL	M/S Elate CC Karachi	KES601T/ Germany	188,000	1 <sup>st</sup> Lowest	M/s Elate CC Approved as 1st Lowest
		M/S Zahid Sons	Cardiovit FT-1/ Swiss	300,000	2 <sup>nd</sup> Lowest	
		M/s Radiant Medical	ELI-150c/ USA	274,000	4 <sup>th</sup> Lowest	
		M/S Platinum Corporation	EPG 6 View Plus/ Progetti/ Italy	299,000	3 <sup>rd</sup> Lowest	
		M/s Radiant Medical	Conex Sport Monitor/USA	264,000	1 <sup>st</sup> Lowest	
20	VITAL SIGN MONITORS	M/S Platinum Corporation	Compactsign 300/Humtleigh/ UK	355,000	2 <sup>nd</sup> Lowest	M/s Radiant Medical Approved as
21	AIR MATTRESS	Not Quoted				
22	UV LIGHTS	M/s Hi-Tech Engineering	Local	22,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest
23	X-RAY ILLMINATOR	Not Quoted				

Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
24	ELECTROLYTE ANALYZER			Not Quoted		
25	STEAMER	M/s Hi-Tech Engineering China		7,000 6000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Approved as 1st Lowest

**ORDERS OF KIND:**

*[Signature]*  
**MEMBER**  
 Prof. Dr. Tarique Rafi,  
 Vice Chancellor, Jinnah Sindh Medical University  
 Karachi

**MEMBER**  
 Prof. Dr. Muneer Junjo,  
 Pro. Vice Chancellor Liaquat University of  
 Medical & Health Science Jamshoro

**MEMBER**  
 Prof. Dr. Aneela Ataur Rehman  
 Pro. Vice Chancellor, Dean of Faculty  
 Community Medicines, Liaquat University  
 of Medical & Health Science Jamshoro

*[Signature]*  
**MEMBER**  
 Dr. Shafi Muhammad Qureshi  
 Additional Director, Director of Pharmacy  
 Government Of Sindh Karachi

*[Signature]*  
**MEMBER**  
 (Mr. Jamaluddin Talib)  
 Additional Secretary, (PM & I)  
 Health Department Government Of Sindh  
 Karachi

*[Signature]*  
**MEMBER**  
 Vice Chancellor,  
 Chandka Medical University Larkana

*[Signature]*  
**MEMBER**  
 (Prof: Dr. Ghulam Asghar  
 Channa)

*[Signature]*  
**MEMBER** (Muhammad Talib)  
 Additional Director NAB,  
 NAB Islamabad

as per remarks  
 Price may be compared with P.C.L  
 For single bidder please  
 best value for money

**CHAIRMAN**  
 Prof. Dr. Muhammad Saeed Qureshi, Vice Chancellor,  
 Dow University of Health Sciences, Karachi

*[Signature]*  
**MEMBER** (Ms. Athar Mirani)  
 Additional Finance, Secretary (I.F)  
 Finance Department Government Of Sindh,  
 Karachi

Note:-> Prof: Dr. Muneer Junjo, Pro-vice Chancellor, Liaquat University of Medical & Health Sciences, Jamshoro, did not attend Meeting. He is on Ex-Pakistan leave.

Medical Superintendent  
 Jinnah Hospital Sukkur  
 Ms. Ghammat Sultkur

**COMPARATIVE STATEMENT OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU (UNDER ADP SCHEME) AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR THE FINANCIAL YEAR 2017-18 OPENED ON 30.05.2018 AT 09:00 A.M AT OJHA CAMPUS @ DOW UNIVERSITY OF MEDICAL & HEALTH SCIENCES KARACHI ANNEXURE (B)**

Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
1	ICU VENTILATOR	M/S BIOS Karachi	Evolution 3e Essential USA	2,472,000	1 <sup>st</sup> Lowest	M/s Bios Karachi approved as 1st Lowest
2	MULTIPARAMETER MONITOR	M/S Platinum Corporation	PG M9000/ Progetti/ Italy	449,500	1 <sup>st</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest
3	DEFIBRILLATOR	M/S Elate CC Karachi	EVO 1/Germany	680,000	1 <sup>st</sup> Lowest	M/s Elate CC Approved as 1st Lowest
		Zahid Sons	Defibgard 5000/Swiss	1,500,000	3 <sup>rd</sup> Lowest	
4	INFUSION PUMP	M/s Radiant Medical	M-Series/USA	694,000	2 <sup>nd</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest
		M/S Platinum Corporation	PG 9071/Progetti/ Italy	199,500	1 <sup>st</sup> Lowest	
5	ABG MACHINE	M/s Medi-Urge	Agila VP France	211,000	2 <sup>nd</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest
		S Ejazuddin	ABL80 FLEX CO-OX Analyzer	1,975,000	2 <sup>nd</sup> Lowest	
6	DOUBLE OXYGEN LINE	M/S Platinum Corporation	USA	1,150,000	1 <sup>st</sup> Lowest	M/s Elate CC Approved as 1st Lowest for both Items
		M/S Elate CC Karachi	UK	4,303,180	1 <sup>st</sup> Lowest	
7	DOUBLE VACCUM SYSTEM	M/S Elate CC Karachi	UK			
8	CRASH TROLLEY	M/S Zahid Sons	MX37EMG/ Taiwan	220,000	1 <sup>st</sup> Lowest	M/s Zahid Sons Approved as 1st Lowest
9	LARYNGO SCOPE	M/s Hi-Tech Engineering	Local	17,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
10	AMBO BAG (SILICON)	M/s Hi-Tech Engineering	Taiwan	20,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest




Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
11	B.P APPRATUS	M/s Hi-Tech Engineering	Japan	30,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
12	BIPEP VENTILATORS	M/S Elate CC Karachi	DreamStar Due ST/France	300,000	1 <sup>st</sup> Lowest	M/s Elate CC Approved as 1st Lowest
		M/S Zahid Sons	Sleepcube Bilevel ST/Devilbiss/USA	350,000	2 <sup>nd</sup> Lowest	
		M/S Jafferi Medical	Seteler 150/Ressmed/ Australia	795,000	3 <sup>rd</sup> Lowest	
13	BLOOD WARMER	M/S Elate CC Karachi	Nouva 5/ Germany	300,000	3 <sup>rd</sup> Lowest	M/s Zahid Sons Approved as 1st Lowest
		M/S Zahid Sons	AM-301 Japan	220,000	1 <sup>st</sup> Lowest	
		M/s Radiant Medical	Barkey S Line/ Germany	225,000	2 <sup>nd</sup> Lowest	
14	NEUBILIZER MACHINE	M/S Zahid Sons	KU-200/ Japan	120,000	2 <sup>nd</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
15	ECG MACHINE 6 CHANNEL	M/S Elate CC Karachi	KES601T/ Germany	188,000	1 <sup>st</sup> Lowest	M/s Elate CC Approved as 1st Lowest
		M/S Zahid Sons	Cardiovit FT-1/ Swiss	300,000	4 <sup>th</sup> Lowest	
		M/s Radiant Medical	ELJ-150e/ USA	274,000	2 <sup>nd</sup> Lowest	
		M/S Platinum Corporation	EPG 6 View Plus/ Progetti/ Italy	299,000	3 <sup>rd</sup> Lowest	
16	VITAL SIGN MONITORS	M/s Radiant Medical	Conex Sport Monitor/USA	264,000	1 <sup>st</sup> Lowest	M/s Radiant Medical Approved as 1st Lowest
17	AIR MATTRESS	M/S Platinum Corporation	Compactsign 300/ Huntleigh/UK	355,000	2 <sup>nd</sup> Lowest	
Not Quoted						
18	UV LIGHTS	M/s Hi-Tech Engineering	Taiwan	22,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
19	X-RAY ILLUMINATOR	M/s Hi-Tech Engineering	Japan	18,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest


Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
20	STEAMER	M/s Hi-Tech Engineering	Local	6,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
21	CENTRAL MONITOR	M/s Radiant Medical M/S Platinum Corporation	Surveyors/ Mortara/ USA M6000/Progetti/ Italy	1,296,000 1,600,000	1 <sup>st</sup> Lowest 2 <sup>nd</sup> Lowest	M/s Radiant Medical Approved as 1st Lowest
22	AMBULATORY HOLTHER WITH BLOOD PRESURE MONITORING SYSTEM	M/s Radiant Medical	ABPM 7100/USA	788,000	1 <sup>st</sup> Lowest	M/s Radiant Medical Approved as 1st Lowest
23	SUCTION MACHINE	M/s Radiant Medical M/S Platinum Corporation	Master/ Switzerland Askir C30/CAMI/Italy	466,000 162,500	2 <sup>nd</sup> Lowest 1 <sup>st</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest
24	AUTOCLAVE BENCH TOP	Not Quoted				
25	BLANKET WARMER	M/s Hi-Tech Engineering	Local	695,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
26	PULSE OXIMETER	M/s Radiant Medical M/S Platinum Corporation	7500/USA Smartsign/ Huntleigh/UK	98,000 150,000	1 <sup>st</sup> Lowest 2 <sup>nd</sup> Lowest	M/s Radiant Medical Approved as 1st Lowest
27	OESOPHAGAL STETHOSCOPE	M/s Hi-Tech Engineering	ST SB UK	25,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
28	MINI TRACHEOSTOMY SET	M/s Hi-Tech Engineering	ST SB UK	50,000	1 <sup>st</sup> Lowest	M/s Hi-Tech Engineering Service Approved as 1st Lowest
29	OPERATION TABLE WITH BATTERY BACKUP	M/S Platinum Corporation	Surgitable 4/ATA Surgical/ France	3,988,000	1 <sup>st</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest
30	O.T LIGHT- DOUBLE DOOM SURGICAL LIGHT WITH BATTERY + CAMERA + MONITOR & CAMERA SYSTEM WITH COMPLETE INSTALLATION	M/S Platinum Corporation	7W/5W/ATA Surgical/ France	DDP Light: 3,630,000 DDP Camera : 350,000	1 <sup>st</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest

*[Handwritten signatures and initials]*

Items #	ITEM NAME	NAME OF BIDDER	MAKE MODEL	UNIT PRICE QUOTED IN P.K.R	RANK	REMARKS
31	ANESTHESIA MACHINE WITH INTEGRATED VENTILATOR AND VAPORIZER	M/S Platinum Corporation	Mistral III/ Norvap/ UK	DDP Anesthesia: 1,865,000 DDP Gas Agent 980,000 DDP Monitoring: 385,000 DDP Vaporizer : 220,000	1 <sup>st</sup> Lowest	M/s Platinum Corporation approved as 1st Lowest
32	PORTABLE X RAY MACHINE 100MA	M/S Radiant Medical	Roller 4, Italy	1,688,000	1 <sup>st</sup> Lowest	M/s Radiant Medical Approved as 1st Lowest
		M/S BIOS Karachi	Movix 4 E+ /France	2,590,000	2 <sup>nd</sup> Lowest	

**ORDERS OF KIND:**

*onyalla*

**MEMBER**

Prof. Dr. Tarique Rafi,

Vice Chancellor, Jinnah Sindh Medical University Karachi

**MEMBER**

Prof. Dr. Muneer Junejo,

Pro. Vice Chancellor Liaquat University of Medical & Health Science, Jamshoro

**MEMBER**

Prof. Dr. Anceela Ataur Rehman

Pro. Vice Chancellor, Dean of Faculty Community Medicines, Liaquat University of Medical & Health Science, Jamshoro

**MEMBER**

Dr. Shafi Muhammad Qureshi

Additional Director, Director of Pharmacy Government Of Sindh Karachi

**MEMBER**

(Ma. Jamaluddin jalalani)

Additional Secretary, (PM & I) Health Department Government Of Sindh Karachi

**MEMBER**

(Prof. Dr. Ghulam Asghar Channa)

Vice Chancellor Chandka Medical University, Larkana

**MEMBER**

(Muhammad Talib)

Additional Director NAB, NAB Islamabad

AS Pro sendor

**MEMBER**

(Mr. Abbas Mianji)

Additional Finance, Secretary (L.F) Finance Department Government Of Sindh, Karachi

**CHAIRMAN**

Prof. Dr. Muhammad Saeed Qureshi, Vice Chancellor, Dow University of Medical & Health Sciences, Karachi

*Note:- Prof: Dr. Muneer Junejo, Pro-vice chancellor, Liaquat University of Medical & Health Sciences Jamshoro did not attend meeting. He is on Ex-Pakistan leave.*

Medical Superintendent  
General Hospital  
MS. GIMMETH SANSUR

ATTENDANCE SHEET OF BIDDERS AT OJHA CAMPUS.

DATED 30-5-2018 AT 9:00 AM

FINANCIAL OPENING

ADP SCHEME

SNO:	COMPANY	NAME OF REPRESENTATIVE:	CONTACT NO:	SIGNATURE
01	ZAHID AND SONS	AAMIR ALAM	0333-3345581	
02	Perksten Furniture Tradex	Abdul Sattar	0315-8154803	
03	BIDS	Karimullah Shoukat	0333-7531950	
04	Mediurge	OSAMIS	03366828480	
05	Radiant medical	STEED ISRAE	0321-8558360	
06	Jafri Medical	NOMAN	0313 2651960	
07	Medline Technologies Pvt Ltd	M. Imdad	0321-8279258	
08	Medeqwips CSMC	Jahangir	0321-2792508	
09	Biotech services	M. Rakeed	03008961993	
10	S Etaz walid & Co	ZEHAN	0345-2441775	
11	Khadija Corporation	M. Ghay	03002807686	
12	Abdullah Platinum Corp	Abdullah	03337110483	
13	Lab Lucis	Prof. Dr. Saeed Qureshi	0304083357	

Medical Superintendent  
GMMHC Hospital Sukkur  
(Dr. Muhammad Amroz Baloch)

(Dr. Tariqina Rafiq)

(Prof. Dr. Ghulam Asghar Channa)

(Dr. Shafi Muhammad Qureshi)

(Prof. Dr. Anwar-ul-Haq)

(Additional Secretary)

**MINUTES OF MEETING REGARDING THE TENDER OPENING (TECHNICAL PROPOSALS) OF ITEMS OF MACHINERY AND EQUIPMENTS ADP SCHEME NAMELY 50 BEDDED MEDICAL & SURGICAL ICU WITH EXPANSION OF OTHER UNITS FOR THE FINANCIAL YEAR 2017-18 HELD ON 29-03-2018 AT 3.00 PM IN THE OFFICE OF THE MEDICAL SUPERINTENDENT GMMMC HOSPITAL SUKKUR.**

The meeting started with the recitation of the Holy Quran. All the participants of procurement committee and the bidders of the Tender were welcomed by the Chairman Procurement committee.


All the participants were briefed by the Chairman Procurement Committee about the process of tender of ADP items of machinery and equipments for the FY-2017-18. It was reiterated that the Technical Proposals of all the bidders are hereby to be opened in the presence of all the bidders publically. Everyone showed satisfaction and the process was started. All the bidders as well as the chairman signed the attendance sheets. The attendance sheets are attached herewith.


The meeting proceeded with the opening of Technical Proposals one by as per serial nos of submission of bids AT 2.00 PM and the bidders who were present and had submitted their bids on the due time.

Subsequently technical proposals were seen and signed. All the submitted bids were read out loudly before the committee and the bidders. The unopened Financial Proposals of the bidders were kept under lock and key in the possession of Chairman Procurement Committee.

The meeting ended with the note of thanks to and from the chair.

  
**(MR. AWAIS MUSHTAQ)**  
ASSISTANT COMMISSIONER  
(REPRESENTATIVE FROM COMMISSIONER SUKKUR DIVISION SUKKUR)  
**(MEMBER)**

  
**(DR. ABDUL AZIZ MEMON)**  
ADDITIONAL MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR  
**(MEMBER)**

  
**(DR. MUHAMMAD ANWAR PALARI)**  
MEDICAL SUPERINTENDENT  
CHAIRMAN PROCUREMENT COMMITTEE  
GMMMC HOSPITAL SUKKUR  
**(CHAIRMAN)**  
MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR


**MINUTES OF MEETING REGARDING THE TECHNICAL EVALUATION OF ITEMS OF MACHINERY AND EQUIPMENTS ADP SCHEME NAMELY 50 BEDDED MEDICAL & SURGICAL ICU WITH EXPANSION OF OTHER UNITS FOR THE FINANCIAL YEAR 2017-18 AT GMMMC HOSPITAL SUKKUR.**

The meeting started with the recitation of the Holy Quran. All the participants of Technical Committee were welcomed by the Chairman Technical Committee.

The meeting proceeded as per agenda. All the participants were briefed by the Chairman who is also member of Procurement Committee about the process of tender of ADP items of machinery and equipments for the FY-2017-18. It was reiterated that the Technical Proposals of all the bidders were evaluated as per check list and Technical Specifications. The attendance sheets are attached herewith.

Subsequently the result of technical evaluation was signed. All the members unanimously agreed that the items are as per tender specification and in the best interest of the institute. The Technical Evaluation was further sent to Central Procurement Committee for endorsement.

The meeting ended with the note of thanks to and from the chair.

  
**( PROF. ABRAR SHAIKH )**  
HEAD OF DEPARTMENT OF MEDICINE  
GMMM COLLEGE SUKKUR


**(MEMBER)**

  
**(PROF. INAYATULLAH MEMON )**  
HEAD OF DEPARTMENT OF RADIOLOGY  
GMMM COLLEGE SUKKUR


**(MEMBER)**

Assistant Professor Radiology  
G.M.M.C Hospital Sukkur


PROF. ABRAR SHAIKH  
F. R. C. P  
Head Dept of Medicine  
G. M. C. SUKKUR

  
**( PROF. JAN MUHAMMAD MEMON )**  
HEAD OF DEPARTMENT OF SURGERY  
GMMM COLLEGE SUKKUR

**(MEMBER)**

  
**( DR. MAQSOOD AHMED SIDDIQUI )**  
HEAD OF DEPARTMENT OF ANAESTHESIA  
GMMM COLLEGE SUKKUR

**(MEMBER)**

  
**(DR. ABDUL AZIZ MEMON)**  
ADDITIONAL MEDICAL SUPERINTENDENT  
CHAIRMAN PROCUREMENT COMMITTEE  
GMMMC HOSPITAL SUKKUR  
**(CHAIRMAN)**

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 01	ICU VENTILATOR	M/S Elate CC Karachi	4000 ICU/Siare/ Italy	Non Responsive	The required Turbine life from 7 to 8 years is not mentioned by bidder on his letterhead in Technical specification and brochure of M/S Elate CC attached in bid. The Battery backup support for 90 minutes instead of required 120 minutes or more. The Bidder has mentioned on his letter head automode and manual mode present on Brochure but they are not automatic mode as requirements and not as per Tender specification.
Item No : 02	MULTIPARAMETER MONITOR	M/S BIOS Karachi	Evolution 3e Essential USA	Responsive	As Per Tender Specification.
		M/S Elate CC Karachi	KMO 121/Kalamed/ Germany	Non Responsive	* Option of IBP recorder not available in this model and also no upgradation option. * NIBP Measuring range for adult 25-265 instead of 10-270. * NIBP Measuring range for Neonate 25-135 instead of 10-135. * Others deviations also found in specification and not as per tender specification.
		M/S Zahid Sons Karachi	BT 770/ Bistos/ Korea	Non Responsive	Country of Origin Korea. Bidders mislead by mentioning USA in his bidding documents and not as per tender specification.

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S Radiant	Surveyors S 12/Mortara/USA	Non Responsive	Display size less than 12.1" Inches. Trend Data Recording time is only upto 24 hours, instead of 168 hours as required. NIBP Measurement range 30~250 instead on 10~270. Others deviations also found in specification and not as per tender specification.
		M/S Platinum Corporation	PG M9000 / Progetti /Italy	Responsive	As Per Tender Specification.
		M/S Elate CC Karachi	EVO 1/Germany	Responsive	As Per Tender Specification.
		M/S Zahid Sons	Defibgard 5000/Swiss	Responsive	As Per Tender Specification.
		M/S Radiant	M-Series/USA	Responsive	As Per Tender Specification.
		M/S Platinum Corporation	PG 9071/Progetti/ Italy	Responsive	As Per Tender Specification.

Item No  
03  
DEFIBRILLATOR



**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 04	INFUSION PUMP	M/S Medi Urge	Agila VP/ France	Responsive	As Per Tender Specification.
		M/S S Ejazuddin	OT-701, JMS, Japan	Non Responsive	Quoted Item is very old model. No TFT Touch screen. KVO rate 1ml/hr instead of 0.01 to 5ml/hr KVO rate. Purge rate offered 450ml/hr but required is 1200ml/hr Battery backup 4 hours instead of 9 hours after full charge and not as per tender specification.
		M/S Zahid Sons	SK-6001, China	Non Responsive	Country of Origin is China and not as per tender specification.
Item No : 05	ABG MACHINE	M/S S Ejazuddin	ABL80 FLEX CO-OX Analyzer	Responsive	As Per Tender Specification.
		M/S Platinum Corporation	USA	Responsive	As Per Tender Specification.
Item No : 06	ULTRASOUND MACHINE WITH PROBE	M/S MedEquips	USA, Aplio 300	Responsive	As Per Tender Specification.
Item No	DOUBLE VACCUM SYSTEM	M/S Elate CC Karachi	UK	Responsive	As Per Tender Specification.

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR**  
**LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION**  
**OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME**  
**ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
:07		M/S Platinum Corporation	Japan	Non Responsive	Authority not attached and Not as per tender specification
Item No :08	DOUBLE OXYGEN LINE	M/S Elate CC Karachi	UK	Responsive	As Per Tender Specification.
Item No :09	CRASH TROLLEY	M/S Platinum Corporation	Japan	Non Responsive	Authority not attached and Not as per tender specification
Item No :10	PORTABLE X RAY MACHINE 100MA	M/S BIOS Karachi	China	Non Responsive	Country of Origin is China and not as per tender specification.
		M/S Zahid Sons	MX37EMG/ Taiwan	Responsive	As Per Tender Specification.
		M/S Radiant Medical	Movix-4 E+, France	Responsive	As Per Tender Specification.
Item No :11	LARYNGO SCOPE	M/S High Tech	Roller 4, Italy	Responsive	As Per Tender Specification.

*[Handwritten signatures and initials]*

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 12	AMBO BAG SILICON	M/S High Tech	Taiwan	Responsive	As Per Tender Specification.
Item No : 13	B.P APPRATUS	M/S High Tech	Japan	Responsive	As Per Tender Specification.
Item No : 14	FOOD TROLLEY	M/S High Tech	Local	Responsive	As Per Tender Specification.
Item No : 15	BIPAP VENTILATORS	M/S Elate CC Karachi	DreamStar Due ST/France	Responsive	As Per Tender Specification.
		M/S Zahid Sons	Sleepcube Bilevel ST/Devilbiss/USA	Responsive	As Per Tender Specification.
		M/S Jafferi Medical	Seteler 150/Resmed/ Australia	Responsive	As Per Tender Specification.
Item No : 16	WARMER	M/S Platinum Corporation	Steler 150/Resmed/ Australia	Non Responsive	Authority Letter of Manufacturer not found and not as per tender specification.
		M/S Zahid Sons	AM-301 Japan	Responsive	As Per Tender Specification.

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S Radiant	Barkey S Line/ Germany	Responsive	As Per Tender Specification.
		M/S Elate CC Karachi	Nouva 5/ Germany	Responsive	As Per Tender Specification.
Item No : 17	BLOOD WARMER	M/S Zahid Sons	AM-301 Japan	Responsive	As Per Tender Specification.
		M/S Radiant	Barkey S Line/ Germany	Responsive	As Per Tender Specification.
		M/S Zahid Sons	KU-200/ Japan	Responsive	As Per Tender Specification.
Item No : 18	NEUBILIZER MACHINE	M/S High Tech	CA - MI , Italy	Responsive	As Per Tender Specification.
		M/S Platinum Corporation	CA-MI/ Italy	Non Responsive	Authority Letter of Manufacturer not found and not as per tender specification.
Item No : 19	ECG MACHINE 6 CHANNEL	M/S Elate CC Karachi	KES601T/ Germany	Responsive	As Per Tender Specification.


**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR - EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S Zahid Sons	Cardiovit FT-1/ Swiss	Responsive	As Per Tender Specification.
		M/S Radiant	ELI-150c/ USA	Responsive	As Per Tender Specification.
		M/S Platinum Corporation	EPG 6 View Plus/ Progetti/ Italy	Responsive	As Per Tender Specification.
		M/S Radiant	Conex Sport Monitor/USA	Responsive	As Per Tender Specification.
Item No : 20	VITAL SIGN MONITORS	M/S Platinum Corporation	Compactsign 300/Huntleigh/ UK	Responsive	As Per Tender Specification.
Item No : 21	AIR MATTRESS	M/S High Tech	China	Responsive	As Per Tender Specification.
Item No : 22	UV LIGHTS	M/S High Tech	Local	Responsive	As Per Tender Specification.
Item No : 23	X-RAY ILLIMINATOR	M/S High Tech	Local	Responsive	As Per Tender Specification.
Item No : 24	ELECTROLYTE ANALYZER	M/S Lab Link Enterprises	Humalyte Plus/Human/ Germany	Responsive	As Per Tender Specification.


**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION  
OF TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (A)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 25	STEAMER	M/S High Tech	China	Responsive	As Per Tender Specification.

  
**Prof. Iqbal Muhammad Memon**  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member


  
**Prof: Abrar Shaikh**  
 Head Of Department of Medicine,  
 GMMM College Sukkur  
 Member

**PROF ABRAR SHAIKH**  
**F. R. C. P**  
 Head Deptt of Medicine  
 G. M. C SUKKUR

  
**Dr. Maqsood Ahmed Siddiqui**  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

  
**Dr. Inayatullah Memon**  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

**Assistant Professor Radiology**  
**G. M. M. C Hospital Sukkur**

  
**Dr. Abdul Aziz Memon**  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION OF  
TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 01	ICU VENTILATOR	Elate CC Karachi	4000 ICU/Siare/ Italy	Non Responsive	The required Turbine life from 7 to 8 years is not mentioned by bidder on his letterhead in Technical specification and brochure of M/S Elate CC attached in bid. The Battery backup support for 90 minutes instead of required 120 minutes or more. The Bidder has mentioned on his letter head automode and manual mode present on Brochure but they are not automatic mode as requirements and not as per tender specification.
Item No : 02	MULTIPARAMETER MONITOR	BIOS Karachi	Evolution 3e Essential USA	Responsive	As per Tender Specification.
		Elate CC Karachi	KMO 121/Kalamed/ Germany	Non Responsive	Option of IBP recorder not available in this model and also no upgradation option. NIBP Measuring range for adult 25-265 instead of 10-270. NIBP Measuring range for Neonate 25-135 instead of 10-135. Others deviations also found in specification and not as per tender specification.
		Zahid Sons	BT 770/ Bistos/ Korea	Non Responsive	Option of IBP recorder not available in this model and also no upgradation option. NIBP Measuring range for adult 25-265 instead of 10-270. NIBP Measuring range for Neonate 25-135 instead of 10-135. Others deviations also found in specification and not as per tender specification.

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION OF  
TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		Radiant	Surveyors S 12/Mortara/USA	?	Display size less than 12.1 Inches. Trend Data Recording time is only upto 24 hours, instead of 168 hours as required. NIBP Measurement range 30~250 instead on 10~270. Others deviations also found in specification and not as per tender specification.
		Platinum Corporation	PG M9000/ Progetti/ Italy	Responsive	As per Tender Specification.
	DEFIBRILLATOR	Elate CC Karachi	EVO 1/Germany	Responsive	As per Tender Specification.
		Zahid Sons	Defibgard 5000/Swiss	Responsive	As per Tender Specification.
		Radiant	M-Series/USA	Responsive	As per Tender Specification.
	INFUSION PUMP	Platinum Corporation	PG 9071/Progetti/ Italy	Responsive	As per Tender Specification.
		Medi Urge	Agila VP France	Responsive	As per Tender Specification.
Item No : 04		S Ejazuddin	OT-701, JMS, Japan	NON-Responsive	M/S S Ejazuddin Quoted very old model. No TFT Touch screen. KVO rate 1ml/hr instead of 0.01 to 5ml/hr KVO rate. Purge rate offered 450ml/hr but required is 1200ml/hr Battery backup 4 hours instead of 9 hours after full charge and not as per tender specification.
		Zahid Sons	SK-6001, China	Non Responsive	Country of Origin is China and not as per tender specification.
	ABG MACHINE	S Ejazuddin	ABL80 FLEX CO-OX Analyzer	Responsive	As per Tender Specification.
Item No : 05		M/S Platinum Corporation	USA	Responsive	As per Tender Specification.



**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION OF  
TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 06	DOUBLE OXYGEN LINE	M/S Elate CC Karachi	UK	Responsive	As per Tender Specification.
Item No : 07	DOUBLE VACCUM SYSTEM	M/S Platinum Corporation M/S Elate CC Karachi	Japan UK	Non Responsive Responsive	Not as per required specification As per Tender Specification.
Item No : 08	CRASH TROLLEY	M/S Platinum Corporation M/S Platinum Corporation	Japan China	Non Responsive Non Responsive	Not as per required specification Country of Origin is China and not as per tender specification
Item No : 09	LARYNGO SCOPE	M/S Zahid Sons	MX37EMG/ Taiwan	Responsive	As per Tender Specification.
Item No : 10	AMBO BAG (SILICON)	M/S High Tech	Local	Responsive	As per Tender Specification.
Item No : 11	B.P APPRATUS	M/S High Tech	Taiwan	Responsive	As per Tender Specification.
Item No : 12	BIPEP VENTILATORS	M/S Elate CC Karachi M/S Zahid Sons M/S Jafferi Medical	DreamStar Due ST/France Sleepcube Bilevel ST/Devilbiss/USA Seteler 150/Resmed/Australia	Responsive Responsive Responsive	As per Tender Specification. As per Tender Specification. As per Tender Specification.
Item No : 13	BLOOD WARMER	M/S Elate CC Karachi M/S Zahid Sons M/S Radiant	Steler 150/Resmed/Australia Nouva 5/ Germany AM-301 Japan Barkey S Line/Germany	Non Responsive Responsive Responsive	Authority Letter of Manufacturer not found As per Tender Specification. As per Tender Specification.
Item No :	NEUBILIZER MACHINE	M/S Zahid Sons M/S High Tech	KU-200/ Japan	Responsive Responsive	As per Tender Specification. As per Tender Specification.

*[Handwritten signatures and marks]*

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION OF  
TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
14		M/S Platinum Corporation	CA-MI/ Italy	Non Responsive	Not as per required specification
	ECG MACHINE 6 CHANNEL	M/S Elate CC Karachi	KES601T/ Germany	Responsive	As per Tender Specification.
Item No : 15		M/S Zahid Sons	Cardiovit FT-1/ Swiss	Responsive	As per Tender Specification.
		M/S Radiant	ELI-150c/ USA	Responsive	As per Tender Specification.
		M/S Platinum Corporation	EPG 6 View Plus/ Progetti/ Italy	Responsive	As per Tender Specification.
	VITAL SIGN MONITORS	M/S Radiant	Conex Sport Monitor/USA	Responsive	As per Tender Specification.
Item No : 16		M/S Platinum Corporation	Compactsign 300/ Huntleigh/UK	Responsive	As per Tender Specification.
Item No : 17	AIR MATTRESS	M/S High Tech	Local	Responsive	As per Tender Specification.
Item No : 18	UV LIGHTS	M/S High Tech	Taiwan	Responsive	As per Tender Specification.
Item No : 19	X-RAY ILLUMINATOR	M/S High Tech	Japan	Responsive	As per Tender Specification.
Item No : 20	STEAMER	M/S High Tech	Local	Responsive	As per Tender Specification.
	CENTRAL MONITOR	M/S Zahid Sons	Bistos/Korea	Non Responsive	Country of Origin Korea. Bidders mislead by mentioning USA in his bidding documents and not as per tender specification.
Item No : 21		M/S Radiant	Surveyors/ Mortara/ USA	Responsive	As per Tender Specification.
		M/S Platinum Corporation	M6000/Progetti/ Italy	Responsive	As per Tender Specification.
Item No : 22	AMBULATORY HOLTER WITH BLOOD PRESSURE MONITORING SYSTEM	M/S Radiant	ABPM 7100/USA	Responsive	As per Tender Specification.
	SUCTION MACHINE	M/S Zahid Sons	SU-305/Taiwan	Non Responsive	Country of origin is Taiwan and not as per tender specification.
Item No : 23		M/S Radiant	Master/ Switzerland	Responsive	As per Tender Specification.

*[Handwritten Signature]*


**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION OF  
TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME  
ANNEXURE (B)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S Platinum Corporation	Askir C30/CAMI/Italy	Responsive	As per Tender Specification.
Item# 24	<i>Autoclave Benchtop</i>	M/S Platinum Corporation	PG Sterile/Progetti/Italy	Responsive	As per Tender Specification.
Item No : 25	BLANKET WARMER	M/S High Tech	Local	Responsive	As per Tender Specification.
	PULSE OXIMETER				
Item No : 26		M/S Zahid Sons	BT-720/Bistos/Korea	Non Responsive	Country of Origin Korea. Bidders mislead by mentioning USA in his bidding documents and not as per tender specification and not as per tender specification.
		M/S Radiant	7500/USA	Responsive	As per Tender Specification.
		M/S Platinum Corporation	Smartsign/Huntleigh/UK	Responsive	As per Tender Specification.
Item No : 27	OESOPHAGAL STETHOSCOPE	M/S High Tech	ST SB UK	Responsive	As per Tender Specification.
Item No : 28	MINI TRACHEOSTOMY SET	M/S High Tech	ST SB UK	Responsive	As per Tender Specification.
	OPERATION TABLE WITH BATTERY BACKUP				
Item No : 29		M/S Elate CC Karachi	MO460/Spain	Non Responsive	. Given table length is 1960mm instead of 2100mm and lateral tilt given -15 to + 15 instead of -20 / + 20 degree. Manual Head plate adjustment -90, +45 instead of -90 to + 90 degree, Kidney bridge 100 mm instead of 120 mm height required, the battery backup is not given which is required upto 4 hours and not as per tender specification.
		M/S Zahid Sons	OT-2300 PLUS/Taiwan	Non Responsive	Country of origin is Taiwan and not as per tender specification.
		M/S Platinum Corporation	Surgitable 4/ATA Surgical/ France	Responsive	As per Tender Specification.

*Handwritten signatures and marks at the bottom right of the page.*


**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR**  
**LIST OF MACHINERY AND EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMC HOSPITAL SUKKUR EVALUATION OF**  
**TECHNICAL PROPOSALS / BID FOR MACHINERY, EQUIPMENTS OF ADP SCHEME**  
**ANNEXURE (B)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 30	O.T LIGHT- DOUBLE DOOM SURGICAL LIGHT WITH BATTERY + CAMERA + MONITOR & CAMERA SYSTEM WITH COMPLETE INSTALLATION	M/S Elate CC Karachi	CLVM 700/500 LED Spain	Non Responsive	* Quoted Color temprature 3800 to 5000K instead of 3500 to 5000K. * Illumination depth 1000mm instead of 1500 mm * Service life of LED 20000 hours instead of 50000 hours which is too short and decrease life of light below then half of required and not as per tender specification.
Item No : 31	ANESTHESIA MACHINE WITH INTEGRATED VENTILATOR AND VAPORIZER	M/S Platinum Corporation M/S Elate CC Karachi M/S Platinum Corporation	7W/5W/ATA Surgical/ France Morpheus/Siare/ Italy Mistral III/ Norvap/ UK	Responsive Non Responsive Responsive	As per Tender Specification. Do not comply with the specification such as TFT touch screen not available and not as per tender specification.
Item No : 32	PORTABLE X RAY MACHINE 100MA	M/S BIOS Karachi M/S Radiant M/S BIOS Karachi	Roller 4 /Italy Movix 4 E+ /France	Responsive Responsive Responsive	As per Tender Specification. As per Tender Specification. As per Tender Specification.


  
**Prof: Jan Muhammad Memon**  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

  
**Prof: Maqsood Ahmed Siddiqui**  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

  
**Prof: Abrar Shaikh**  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

  
**Prof: Inayatullah Memon**  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

  
**Dr. Abdul Aziz Memon**  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

  
**Assistant Professor Radiology**  
**G.M.M.C Hospital Sukkur**

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 01	Beds for Surgical ICU three folders	M/S Hi Tech Engineering	103-K-00, Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
		M/S Medline	Local	Responsive	As per tender Specification
Item No: 02	Foam Mattress	M/S The Dcare	HCB 340, China	Non Responsive	Sample not provided
		M/S The Dcare	Local	Non Responsive	Sample not provided
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 03	Blanket Red Size 60 X 90	M/S Pakistan Furniture	Local	Responsive	As per tender Specification
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
		M/S Medline	Local	Responsive	As per tender Specification

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S The Dcare	CB-9000, China	Non Responsive	Sample not provided
Item No: 04	Chairs for Seminar Hall	M/S Hi Tech Engineering M/S Pakistan Furniture	Local Local	Non Responsive Responsive	Sample not provided As per tender Specification
Item No: 05	Tables for Seminar Hall	M/S Hi Tech Engineering M/S Pakistan Furniture	Local Local	Non Responsive Responsive	Sample not provided As per tender Specification
Item No: 06	Chairs for Offices	M/S Hi Tech Engineering M/S Pakistan Furniture	Local Local	Non Responsive Responsive	Sample not provided As per tender Specification
Item No: 07	Book Rack for Library	M/S Hi Tech Engineering M/S Pakistan Furniture	Local Local	Non Responsive Responsive	Sample not provided As per tender Specification
Item No: 08	Chairs for Laboratory	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided

*(Handwritten signatures and marks)*

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
 24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
 ANNEXURE (A)  
 SCHEME**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification

Item No: 09	Tables for Laboratory	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
Item No: 10	Office Table	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		Pakistan Furniture	Local	Responsive	As per tender Specification
Item No: 11	Revolving Chairs	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		Pakistan Furniture	Local	Responsive	As per tender Specification

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
SCHEME  
ANNEXURE (A)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No: 12	Bedside Lockers	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
Item No: 13	Iron Cupboard	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
	Nursing Counter	Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 14		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
	Stool Steel	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 15		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
Item No: 16	IV Stand Steel	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided

*[Handwritten signatures and marks]*



**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
ANNEXURE (A)  
SCHEME**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
Item No: 17	Wheel Chairs	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
Item NO: 18	Steture with Side arms	Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per tender Specification
<b>OTHERS ITEMS OF OFFICE EQUIPMENTS FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMMMC HOSPITAL SUKKUR</b>					
Item No: 01	Air Conditioner Split type 1.5 Tonn with Tablizer 10,000 watts	M/S Hi Tech Engineering	Dawlance	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Orient	Responsive	As per tender Specification
	Electric Water Cooler with stabilizer 5000 watts	M/S Hi Tech Engineering	Fischer	Non Responsive	Sample not provided

Handwritten signatures and marks are present at the bottom of the page, including a large signature on the right and a checkmark on the left.

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
ANNEXURE (A)  
SCHEME**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No: 02		M/S Pakistan Furniture	Orient	Responsive	As per tender Specification
	Refrigerator with stabilizer 3000 Watts	M/S Hi Tech Engineering	Dawlance	Non Responsive	Sample not provided
Item No: 03		M/S Pakistan Furniture	Orient	Responsive	As per tender Specification
	Branded Desktop Computer with Monitor	M/S Hi Tech Engineering	HP 1102W	Non Responsive	Sample not provided
Item No: 04		M/S Pakistan Furniture	Dell 3669	Responsive	As per tender Specification
	Multimedia Screen	M/S Hi Tech Engineering	X402, Optama, USA	Non Responsive	Sample not provided
Item No:05		M/S Pakistan Furniture	-	Responsive	As per tender Specification


Handwritten signatures and marks are present at the bottom of the page, including a large signature on the right and several scribbles on the left.


**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF  
 24 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP  
 SCHEME  
 ANNEXURE (A)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No: 06	Laptop	M/S Hi Tech Engineering Pakistan Furniture	HP Dell 5570	Non Responsive Responsive	Sample not provided As per tender Specification

  
 Prof. Jan Muhammad Memon  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

  
 Dr. Maqsood Ahmed Siddiqui  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

  
 Prof: Abrar Shaikh  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

  
 Dr. Inayatullah Memon  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member  
 Assistant Professor Radiology  
 G.M.M.M.C Hospital Sukkur

  
 Dr. Abdul Aziz Memon  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR TECHNICAL  
EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU  
TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR**

**ADP SCHEME  
ANEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 01	Beds for Surgical ICU three folders	M/S Hi Tech Engineering	103-K-00, Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
		M/S Medline	Local	Responsive	As per Tender Specification
Item No: 02	Foam Mattress	M/S The Dcare	HCB 340, China	Non Responsive	Sample not provided
		M/S The Dcare	Local	Non Responsive	Sample not provided
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 03	Blanket Red Size 60 X 90	M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Medline	Local	Responsive	As per Tender Specification
Item No: 04	Chairs for Seminar Hall	M/S The Dcare	CB-9000, China	Non Responsive	Sample not provided
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No:		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided

Tahlee for Examining Hall

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR      TECHNICAL**  
**EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU**  
**TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR**

**ADP SCHEME**  
**ANEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
05	TABLES FOR SEMINAR HALL	M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No; 06	Chairs for Offices	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No; 07	Book Rack for Library	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No; 08	Chairs for Laboratory	Hi Tech Engineering	Local	Non Responsive	Sample not provided
		Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No; 09	Tables for Laboratory	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
		M/SPakistan Furniture	Local	Responsive	As per Tender Specification
Item No; 10	Office Table	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
		M/SPakistan Furniture	Local	Responsive	As per Tender Specification
Item No;	Revolving Chairs	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR TECHNICAL  
EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU  
TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR**

**ADP SCHEME  
ANEXURE (B)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
11		M/SPakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 12	Bedside Lockers	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
		M/SPakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 13	Iron Cupboard	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
		M/SPakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 14	Nursing Counter	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
		M/SPakistan Furniture	Local	Responsive	As per Tender Specification

Handwritten signatures and initials are present at the bottom of the page, including a large signature on the left and several other marks on the right.

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR TECHNICAL  
EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU  
TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HSOPITAL SUKKUR FOR**


**ADP SCHEME  
ANEXURE (B)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No: 15	Stool Steel	M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
Item No:16	IV Stand Steel	M/SPakistan Furniture	Local	Responsive	As pert Tender Specification
		M/SHi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 01	Air Conditioner Split type 1.5 Tonn with Tablizer 10,000 watts	M/S Pakistan	Local	Responsive	As pert Tender Specification
		M/S Hi Tech Engineering	Dawlance	Non Responsive	Sample not provided
Item No: 02	Electric Water Cooler with stabilizer 5000 watts	M/S Pakistan Furniture	Orient	Responsive	As pert Tender Specification
		M/S Hi Tech Engineering	Fischer	Non Responsive	Sample not provided
Item No: 03	Refrigerator with stabilizer 3000 Watts	M/S Pakistan Furniture	Orient	Responsive	As pert Tender Specification
		M/S Hi Tech Engineering	Dawlance	Non Responsive	Sample not provided
Item No: 04	Branded Desktop Computer with Monitor	M/S Pakistan Furniture	Orient	Responsive	As pert Tender Specification
		M/S Hi Tech Engineering	HP 1102W	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Dell 3669	Responsive	As pert Tender Specification

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR      TECHNICAL  
EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24 BEDDED MEDICAL ICU  
TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR**


**ADP SCHEME  
ANEXURE (B)**


Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No:05	Multimedia Screen	M/S Hi Tech Engineering M/S Pakistan Furniture	X402, Optama, USA -	Non Responsive  Responsive	Sample not provided  As per Tender Specification

  
**Prof: Jan Muhammad Memon**  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member


  
**Prof: Abrar Shaikh**  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

**PROF ABRAR SHAIKH**  
 F.R.C.P  
 Head Dept: of Medicine  
 G. M. M. C. Hospital Sukkur

  
**Dr. Maqsood Ahmed Siddiqui**  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

  
**Dr. Inayatullah Memon**  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

**Assistant Professor Radiology**  
**G.M.M.C Hospital Sukkur**

  
**Dr. Abdul Aziz Memon**  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman




**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24  
BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP SCHEME  
ANNEXURE (C)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No : 01	Beds for Surgical ICU three folders	M/S Hi Tech Engineering	103-K-00, Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
		M/S Medline	Local	Responsive	As per Tender Specification
Item No; 03	Chairs for Offices	M/S The Dcare	HCB 340, China	Non Responsive	Sample not provided
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 04	Office Table	M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 05	Revolving Chairs	M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
Item No: 6	Bedside Lockers	M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification

*[Handwritten signatures and initials are present at the bottom of the page, including a large signature across the bottom right and several initials on the left.]*

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24  
BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP SCHEME  
ANNEXURE (C)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No:07	Iron Cupboard	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 08	Nursing Counter	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No:09	Stool Steel	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No:10	IV Stand Steel	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 11	Wheel Chairs	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item NO: 12	Steture with Side arms	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No:		M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided

  
 Ghulam Muhammad  
 Director, Establishment Branch

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24  
BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP SCHEME  
ANNEXURE (C)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
13	Patient Examination Couch	M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 14	Dressing Table	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item NO: 15	Patient Waiting Chairs Stainless steel	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 16 A	Benches Steel Side Arms	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 16 B	Cardiac Trolley	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
Item No: 16 C	ECG Trolley	M/S Hi Tech Engineering	Local	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Local	Responsive	As per Tender Specification
<b>OTHERS ITEMS OF OFFICE EQUIPMENTS FOR ESTABLISHMENT OF 24 BEDDED SURGICAL ICU AT GMMMC HOSPITAL SUKKUR</b>					
Item No:	Air Conditioner Split type 1.5 Tonn with Tablizer	M/S Hi Tech Engineering	Dawlance	Non Responsive	Sample not provided

*[Handwritten signatures and marks]*

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24  
 BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP SCHEME  
 ANNEXURE (C)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
01	10,000 watts	M/S Pakistan Furniture	Orient	Responsive	As per Tender Specification
Item No: 02	Electric Water Cooler wth stabilizer 5000 watts	M/S Hi Tech Engineering	Fischer	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Orient	Responsive	As per Tender Specification
Item No: 03	Refrigerator with stabilizer 3000 Watts	M/S Hi Tech Engineering	Dawlance	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Orient	Responsive	As per Tender Specification
Item No: 04	Branded Desktop Computer with Monitor	M/S Hi Tech Engineering	HP 1102W	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Dell 3669	Responsive	As per Tender Specification

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
TECHNICAL EVALUATION TECHNICAL PROPOSALS / BID FOR FURNITURE / FIXTURE / OFFICE EQUIPMENT FOR ESTABLISHMENT OF 24  
BEDDED SURGICAL ICU TO BE INSTALLED AT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR FOR ADP SCHEME  
ANNEXURE (C)**

Items #	Items	Bidder Name	Make & Model	Technical Evaluation	Remarks
Item No:05	Multimedia Screen	M/S Hi Tech Engineering	X402, Optama, USA	Non Responsive	Sample not provided
		M/S Pakistan Furniture	-	Responsive	As per Tender Specification
Item No: 06	Laptop	M/S Hi Tech Engineering	HP	Non Responsive	Sample not provided
		M/S Pakistan Furniture	Dell 5570	Responsive	As per Tender Specification

*[Signature]*  
**Prof: Jan Muhammad Memon**  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

*[Signature]*  
**Dr. Maqsood Ahmed Siddiqui**  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

*[Signature]*  
**Dr. Abdul Aziz Memon**  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

*[Signature]*  
**Prof: Abrar Shaikh**  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

*[Signature]*  
**Dr. Inayatullah Memon**  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

*[Signature]*  
**Assistant Professor Radiology**  
 G.M.M.C Hospital Sukkur

*[Signature]*  
**PROF ABBAR SHAKH**  
 F.R.C.S  
 Head Dept: of Medicine  
 G. M. C. Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME

OPENED ON 29 MARCH 2018

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5	6	7	8	9	10
		M/S Medequips SMC	M/S Elite CC	M/S BIOS	M/S s. Ejazuddin & Co.	M/S Jafari Medical	M/S Zahid & Sons	M/S Platinum Corporation	M/S Radiant Medical	M/S Medi Urge	M/S Hi Tech Engineering
1	Registration with income tax (active income tax certificat)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
2	Registration with sales tax (active sales tax certificate)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
3	Relevant Experience minimum 5 years (relevant experience certificate)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
4	Fresh Bank Certificate / Bank Statement Showing End Transaction of 60 M	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
5	Signed and Company Sealed Bidding Documents & Contract Agreement	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
6	Delivery Time	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
7	WEBOC ID of Firm , Copy Certificate	Y	Y	Y	Y	Y	Y	Y	Y	Y	N/A
8	Copy of Earnest money without Amount	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
9	Original Tender Purchase Receipt	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y


OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF


ADP SCHEME


OPENED ON 29 MARCH 2018


S.#	1	2	3	4	5	6	7	8	9	10
ELEGIBILITY PARAMETERS	M/S Medequips SMC	M/S Elate CC	M/S BIOS	M/S S. Ejazuddin & Co.	M/S Jafari Medical	M/S Zahid & Sons	M/S Platinum Corporation	M/S Radiant Medical	M/S Medi Urge	M/S Hi Tech Engineering
E = ELIGIBLE / N = NOT ELIGIBLE	E	E	E	E	E	E	E	E	E	E

  
 Prof. Javed Muhammad Memon  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

Prof: Abrar Shaikh  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

  
 Dr. Maqsood Ahmed Siddiqui  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

  
 Dr. Inayatullah Memon  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

  
 Dr. Abdul Aziz Memon  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

  
 Assistant Professor Radiology  
 G.M.M.C Hospital Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME

OPENED ON 29 MARCH 2018

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5	6	7	8	9	10
		M/S Medequips SMC	M/S Elite CC	M/S BIOS	M/S S. Ejazuddin & Co.	M/S Jaferi Medical	M/S Zahid & Sons	M/S Platinum Corporation	M/S Radiant Medical	M/S Medi Urge	M/S Hi Tech Engineering
1.	Tender Purchase Receipt (Original)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
5.	Technical Specifications (dully filled, Signed with Stamp)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
6.	Technical Proposal on Bidder's Letterhead	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
<b>Bidders Documents</b>											
7.	Manufacturer's Authorization (as per sample form)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
8.	Undertaking (as per sample form)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
9.	Certificate (as per sample form)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y



OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME

OPENED ON 29 MARCH 2018

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5	6	7	8	9	10
		M/S Medequips SMC	M/S Elite CC	M/S BIOS	M/S S. Ejazuddin & Co.	M/S Jaferi Medical	M/S Zahid & Sons -	M/S Platinum Corporation	M/S Radiant Medical	M/S Medi Urge	M/S Hi Tech Engineering
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory. Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
11.	Valid PNRA registration certificate where applicable	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
12.	Company Profile	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
13.	Bank certificate/Statement with last three years turnover.	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
14.	Income Tax Return (last two years)	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
15.	Workshop for after sales services	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
16.	Technical Team detail	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
17.	The Bidder will ensure provide WEOC ID of Bidder must be active for Imported items.	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
<b>Original Equipment Manufacturer (OEM or brand quoted)</b>											
18.	CE / FDS / JIS	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y




OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME

OPENED ON 29 MARCH 2018

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5	6	7	8	9	10
19.	References of offered model or brand (in Pakistan preferable in Sindh)	M/S Medequips SMC	M/S Elite CC	M/S BIOS	M/S S. Ejazuddin & Co.	M/S Jafari Medical	M/S Zahid & Sons	M/S Platinum Corporation	M/S Radiant Medical	M/S Medi Urge	M/S Hi Tech Engineering
	<b>E = ELIGIBLE / N = NOT ELIGIBLE</b>	Y	Y	Y	Y	Y	Y	Y	Y	Y	Y
		E	E	E	E	E	E	E	E	E	E

Prof: Jan Muhammad Memon  
Head Of Department of Surgery  
GMMM College Sukkur  
Member

Prof: Abrar Shaikh  
Head Of Department of Medicine  
GMMM College Sukkur  
Member

Dr. Maqsood Ahmed Siddiqui  
Head Of Department Anesthesia  
GMMM College Sukkur  
Member

Dr. Inayatullah Memon  
Head Of Department of Radiology  
GMMM College Sukkur

Dr. Abdul Aziz Memon  
Additional Medical Superintendent  
GMMMC Hospital Sukkur  
Chairman

Member  
Assistant Professor Radiology  
G.M.M.C Hospital Sukkur

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME  
OPENED ON 29 MARCH 2018**

S.#	1	2	3	4	5
	<b>ELEGIBILITY PARAMETERS</b>				
	M/S PAKISTAN FURNITURE	M/S HI-TECH ENGINEERING	M/S MEDLINE TECHNOLOGIES	M/S Khalid Corporation	M/S THE 'D' CARE
1.	Tender Purchase Receipt (Original)	Y	Y	Y	Y
2.	Bid Security (Pay Order/Bank Draft) / Bank Guarantee	Y	Y	Earnest is visible i.e. Rs. 200,000/-	Y
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Y	Y	Samples not provided	
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Y	Y	Samples not provided	
5.	Technical Specifications (dully filled, Signed with Stamp)	Y	Y	Samples not provided	
6.	Technical Proposal on Bidder's Letterhead	Y	Y	N	Y
	<b>Bidders Documents</b>				
7.	Manufacturer's Authorization (as per sample form)	Y	Y	Samples not provided	
8.	Undertaking (as per sample form)	Y	Y	Samples not provided	
9.	Certificate (as per sample form)	Y	Y	Samples not provided	
10.	Income Tax & GST Registration Professional	Y	Y	Samples not provided	
11.	Professional Tax Certificate (Sindh)	Y	Y	Samples not provided	

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME  
OPENED ON 29 MARCH 2018**

S.#	ELIGIBILITY PARAMETERS				
	1	2	3	4	5
	M/S PAKISTAN FURNITURE	M/S HI-TECH ENGINEERING	M/S MEDLINE TECHNOLOGIES	M/S Khalid Corporation	M/S THE 'D' CARE
12	SECP Incorporation Certificate (if applicable)	Y	Y		
13	Company Profile	Y	Y		
14	Bank certificate	Y	Y		Y
15	Bank Statement	Y	Y		Y
16	Audited Financial Satatement (last year)	Y	Y		Y
17	Income Tax Return (last years)	Y	Y		Y
<b>Original Equipment Manufacturer (OEM or brand of Hospital Furniture:)</b>					
18.	Catalogue / Brochures (Original)	Y	Y		
19.	ISO 9001 / CE Certificate	Y	Y		
20.	Chamber of Commerce & Industry Membership Certificate (Local manufacturer)	Y	Y		
21.	EOBI Registration Certificate (local Manufacturer)	N/A	N/A		Samples not provided
22.	Social Security Registration Certificate (local manufacturer)	N/A	N/A		Samples not provided
23.	References of the quoted brand / manufacturer (local)	Y	Y		
24.	references of the quoted brand / manufacturer (international)	Y	Y		

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR**  
**OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME**  
**OPENED ON 29 MARCH 2018**

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5
E = ELIGIBLE / N= NOT ELIGIBLE		M/S PAKISTAN FURNITURE E	M/S HI-TECH ENGINEERING N	M/S MEDLINE TECHNOLOGIES E	M/S Khalid Corporation N	M/S THE 'D' CARE N

*Jan Muhammad Memon*  
**Prof: Jan Muhammad Memon**  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

**Dr. Maqsood Ahmed Siddiqui**  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

**Prof: Abrar Shaikh**  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

**Dr. Inayatullah Memon**  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

*Dr. Abdul Aziz Memon*  
**Dr. Abdul Aziz Memon**  
 Assistant Professor Radiology  
 Additional Medical Superintendent  
 G.M.M.C Hospital Sukkur  
 Chairman

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME  
OPENED ON 29 MARCH 2018**

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5	6	7
		M/S Iqbal & Company	M/S Med Elec Corporation	M/S Medline Technologies	M/S Human Health Care	M/S Khalid Corporation	M/S BIO Tech Services	M/S The D Care
1	Registration with income tax (active income tax certificate)	Y	Y	Y	Y	N	Y	Y
2	Registration with sales tax (active sales tax certificate)	Y	Y	Y	Y	Y	Y	Y
3	Relevant Experience minimum 5 years (relevant experience certificate)	N	Y	1. Relevant Experience does not match with STRN which is w.e.f. 13 March 2015 2. From supply orders relevant experience of 5 years is not found	Y	N	Y	Y
	Fresh Bank Certificate /	N	N	Y	Y	Y	Y	Y
4	Bank Statement Showing End Transaction of 60 M	N	N	Bank Gaurantee 26.06.2018	Y	Y, But the detailed estimate is not of 60 M	Y	N
5	Signed and Company Sealed Bidding Documents & Contract Agreement	Y	Y	Y	Y	Y	Y	Y
6	Delivery Time	Y	Y	Y	Y	N	Y	N
7	WEOB ID of Firm , Copy Certificate	N	N	N	N	N	N	N

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
OBLIGATORY CRITERIA FOR ELIGIBILITY OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME**

OPENED ON 29 MARCH 2018

S.#	ELEGIBILITY PARAMETERS	1	2	3	4	5	6	7
		M/S Iqbal & Company	M/S Med Elec Corporation	M/S Medline Technologies	M/S Human Health Care	M/S Khalid Corporation	M/S BIO Tech Services	M/s The D Care
8	Copy of Earnest money without Amount	Y	visible 54000/-	Y	Visible 828,000/-	N	Y	Y
9	Original Tender Purchase Receipt	Y	Y	Y	Y	Y	Y	Y
	<b>E = ELIGIBLE / N= NOT ELIGIBLE</b>	<b>N</b>	<b>N</b>	<b>N</b>	<b>N</b>	<b>N</b>	<b>N</b>	<b>N</b>

~~Prof. Jan Muhammad Memon~~  
Head Of Department of Surgery  
GMMM College Sukkur  
Member

Prof: Abrar Shaikh  
Head Of Department of Medicine  
GMMM College Sukkur  
Member

Dr. Maqsood Ahmed Siddiqui  
Head Of Department Anesthesia  
GMMM College Sukkur  
Member

Dr. Inayatullah Memon  
Head Of Department of Radiology  
GMMM College Sukkur  
Member

Dr. Abdul Aziz Memon  
Additional Medical Superintendent  
GMMMC Hospital Sukkur  
Chairman

Assistant Professor Radiology  
G.M.M.C Hospital Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S THE D CARE

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	Yes
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer).	Yes
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Yes
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Not attached
5.	Technical Specifications (dully filled, Signed with Stamp)	Yes
6.	Technical Proposal on Bidder's Letterhead	Yes
	<b>Bidders Documents</b>	
7.	Manufacturer's Authorization (as per sample form)	ARJO Hurtleigh's Expired Authority attached and all authorities are in the name of "A Care"
8.	Undertaking (as per sample form)	Yes
9.	Certificate (as per sample form)	Yes
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Yes
11.	Valid PNRA registration certificate where applicable	N/A
12.	Company Profile	Yes
13.	Bank certificate/Statement with last three years turnover.	Bank Certificate Scanned without duly signed and stamped by bidder and Bank Statement doesnot showing the 60 M turnover for the last year
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Yes
16.	Technical Team detail	Yes
17.	The Bidder will ensure provide WBOC ID of Bidder must be active for Imported items.	Not attached



OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S THE D CARE

ELEGIBILITY PARAMETERS		DETAILS / REMARKS
S.#	Original Equipment Manufacturer (OEM or brand quoted)	Yes
18.	CE / FDS / JIS	Yes
19.	References of offered model or brand (in Pakistan preferable in Sindh)	Not Eligible
<b>E = ELIGIBLE / N = NOT ELIGIBLE</b>		

*[Signature]*  
 Prof: Jan Muhammad Memon  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

Prof: Abrar Shaikh  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

Dr. Maqsood Ahmed Siddiqui  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

*[Signature]*  
 Dr. Inayatullah Memon  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

*[Signature]*  
 Dr. Abdul Aziz Memon  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

*[Signature]*  
 Assistant Professor Radiology  
 G.M.M.C Hospital Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S RIO TECH

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	Yes
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Yes
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Yes
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Yes
5.	Technical Specifications (dully filled, Signed with Stamp)	Yes
6.	Technical Proposal on Bidder's Letterhead	Yes
	<b>Bidders Documents</b>	
7.	Manufacturer's Authorization (as per sample form)	Authority of Medlab has no date of issue, the authority of Esolate photocopied on company letter head
8.	Undertaking (as per sample form)	Yes
9.	Certificate (as per sample form)	Yes
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Yes
11.	Valid PNRA registration certificate where applicable	Not attached
12.	Company Profile	Photocopy instead of original
13.	Bank certificate/Statement with last three years turnover.	Photocopy instead of original
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Photocopy instead of original
16.	Technical Team detail	Attached, on simple paper, instead of letter head
17.	The Bidder will ensure provide WEBOC ID of Bidder must be active for Imported items.	Not attached

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S BIO TECH

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
	<b>Original Equipment Manufacturer (OEM or brand quoted)</b>	
18.	CE / FDS / JIS	Brochures photocopy
19.	References of offered model or brand (in Pakistani preferable in Sindh)	Yes
	<b>E = ELIGIBLE / N = NOT ELIGIBLE</b>	<b>Not Eligible</b>

Prof: Jan Muhammad Memon  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

Prof: Abrar Shaikh  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

Dr. Maqsood Ahmed Siddiqui  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

Dr. Abdul Aziz Memon  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

Dr. Inayatullah Memon  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member  
 Assistant Professor Radiology  
 G.M.M.M.C Hospital Sukkur

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR**  
**DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME (PART-II)**

OPENED ON 29 MARCH 2018

M/S MEDLINE CORPORATION

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	Yes
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Bank Guarantee Expires on 26-6-2018
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Yes
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Yes
5.	Technical Specifications (dully filled, Signed with Stamp)	Yes
6.	Technical Proposal on Bidder's Letterhead	Yes
<b>Bidders Documents</b>		
7.	Manufacturer's Authorization (as per sample form)	Yes
8.	Undertaking (as per sample form)	Yes
9.	Certificate (as per sample form)	Yes
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Yes
11.	Valid PNRA registration certificate where applicable	Yes
12.	Company Profile	Yes
13.	Bank certificate/Statement with last three years turnover.	Bank Guarantee Expires on 26-6-2018
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Yes
16.	Technical Team detail	Yes
17.	The Bidder will ensure provide WEBOC ID of Bidder must be active for Imported items.	Not attached
<b>Original Equipment Manufacturer (OEM or brand quoted)</b>		
18.	CE / FDS / JIS	Yes

*(Handwritten signature)*

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S MEDLINE CORPORATION

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
19.	References of offered model or brand (in Pakistan preferable in Sindh)	1. sRelevant experience does not match with STRN which is w.e.f. 13 March 2015 2. fromsupply orders relevant experince of 5 years is not found
<b>E = ELIGIBLE / N= NOT ELIGIBLE</b>		<b>Not Eligible</b>

Prof: Jan Muhammad Memon  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

Prof: Abrar Shaikh  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

Dr. Maqsood Ahmed Siddiqui  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

Dr. Inayatullah Memon  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

Dr. Abdul Aziz Memon  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

Assistant Professor Radiology  
 G.M.M.M.C Hospital Sukkur

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME (PART-II)**

OPENED ON 29 MARCH 2018  
M/S KHALID CORPORATION

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	CDR NO 11949914 amount not hidden i.e. 200,000/-
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Yes but detailed estimate is not of 60 (M)
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Yes
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Yes
5.	Technical Specifications (dully filled, Signed with Stamp)	Not signed and stamped
6.	Technical Proposal on Bidder's Letterhead	Not attached
<b>Bidders Documents</b>		
7.	Manufacturer's Authorization (as per sample form)	Authority of Noor Internaational and Pak Traders
8.	Undertaking (as per sample form)	Yes
9.	Certificate (as per sample form)	Not as per sample form
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Registration with Income Tax not attached
11.	Valid PNRA registration certificate where applicable	N/A
12.	Company Profile	Yes
13.	Bank certificate/Statement with last three years turnover.	Bank Certificate is not fresh and 1 month Bank Statement is short
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Yes
16.	Technical Team detail	Yes
17.	The Bidder will ensure provide WEBOC ID of Bidder must be active for Imported items.	Not attached
<b>Original Equipment Manufacturer (OEM or brand quoted)</b>		

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)**

OPENED ON 29 MARCH 2018

M/S KHALID CORPORATION

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
18.	CE / FDS / JIS	Yes
19.	References of offered model or brand (in Pakistan preferable in Sindh)	Yes
<b>E = ELIGIBLE / N = NOT ELIGIBLE</b>		<b>Not Eligible</b>

Prof: Jan Muhammad Memon  
Head Of Department of Surgery  
GMMM College Sukkur  
Member

Prof: Abrar Shaikh  
Head Of Department of Medicine  
GMMM College Sukkur  
Member

Dr. Maqsood Ahmed Siddiqui  
Head Of Department Anesthesia  
GMMM College Sukkur  
Member

Dr. Inayatullah Memon  
Head Of Department of Radiology  
GMMM College Sukkur  
Member

Dr. Abdul Aziz Memon  
Additional Medical Superintendent  
GMMM Hospital Sukkur  
Chairman

Assistant Professor Radiology  
G.M.M.C Hospital Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S HUMAN HEALTH

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	Yes
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Yes
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Not Signed
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Yes
5.	Technical Specifications (dully filled, Signed with Stamp)	Yes
6.	Technical Proposal on Bidder's Letterhead	without sign
<b>Bidders Documents</b>		
7.	Manufacturer's Authorization (as per sample form)	Manufacturer's Signature, Stamp, date of issue validity period is not mentioned on the Authority
8.	Undertaking (as per sample form)	Yes
9.	Certificate (as per sample form)	Yes
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Yes
11.	Valid PNRA registration certificate where applicable	Yes
12.	Company Profile	Yes
13.	Bank certificate/Statement with last three years turnover.	Bank Statement 17 months short instead of 36 required 36 months
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Yes
16.	Technical Team detail	Yes
17.	The Bidder will ensure provide WEBOC ID of Bidder must be active for Imported items.	Not attached



OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S HUMAN HEALTH

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
	<b>Original Equipment Manufacturer (OEM or brand quoted)</b>	
18.	CE / FDS / JIS	Yes
19.	References of offered model or brand (in Pakistan preferable in Sindh)	Yes
<b>E = ELIGIBLE / N= NOT ELIGIBLE</b>		<b>NOT ELIGIBLE</b>

Prof: Jap-Muhammad Memon  
 Head Of Department of Surgery  
 GMMM College Sukkur  
 Member

Prof: Abrar Shaikh  
 Head Of Department of Medicine  
 GMMM College Sukkur  
 Member

Dr. Maqsood Ahmed Siddiqui  
 Head Of Department Anesthesia  
 GMMM College Sukkur  
 Member

Dr. Inayatullah Memon  
 Head Of Department of Radiology  
 GMMM College Sukkur  
 Member

Dr. Abdul Aziz Memon  
 Additional Medical Superintendent  
 GMMMC Hospital Sukkur  
 Chairman

Assistant Professor Radiology  
 G.M.M.-M.C Hospital Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL  
 DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018

M/S MED ELEC CORPORATION

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	Yes
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Earnest money is visible (54000)
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Yes
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Yes
5.	Technical Specifications (dully filled, Signed with Stamp)	Yes
6.	Technical Proposal on Bidder's Letterhead	Yes
<b>Bidders Documents</b>		
7.	Manufacturer's Authorization (as per sample form)	Yes
8.	Undertaking (as per sample form)	Not as per sample form
9.	Certificate (as per sample form)	Not as per sample form
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh) Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Yes
11.	Valid PNRA registration certificate where applicable	N/A
12.	Company Profile	
13.	Bank certificate/Statement with last three years turnover.	Yes
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Yes
16.	Technical Team detail	Yes

**OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL  
DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)**

OPENED ON 29 MARCH 2018

M/S MED ELEC CORPORATION

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
17.	The Bidder will ensure provide WEBOC ID of Bidder must be active for Imported items.	WEBOC ID is not showing (a) NTN/FTN (b) STRN © Business Address (d) CNIC NO: & Contact #
	<b>Original Equipment Manufacturer (OEM or brand quoted)</b>	
18.	CE / FDS / JIS	Yes
19.	References of offered model or brand (in Pakistan preferable in Sindh)	Not attached
	<b>E = ELIGIBLE / N= NOT ELIGIBLE</b>	<b>Not Eligible</b>

Prof. Jan Muhammad Memon  
Head Of Department of Surgery  
GMMM College Sukkur  
Member

Prof: Abrar Shaikh  
Head Of Department of Medicine  
GMMM College Sukkur  
Member

Dr. Magsood Ahmed Siddiqui  
Head Of Department Anesthesia  
GMMM College Sukkur  
Member

Dr. Inayatullah Memon  
Head Of Department of Radiology  
GMMM College Sukkur  
Member

Dr. Abdul Aziz Memon  
Additional Medical Superintendent  
GMMMC Hospital Sukkur  
Chairman

Assistant Professor Radiology  
GMMMC Hospital Sukkur

OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018  
M/S IQBAL TRADERS

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
1.	Tender Purchase Receipt (Original)	Yes
2.	Bid Security (Pay Order/Bank Draft) (Original in Financial offer)	Yes
3.	General & Special Conditions of Contract (Duly filled, Signed & Stamped by bidder each & every page)	Yes
4.	Schedule of Requirements (dully filled, Signed with Stamp)	Not as per Sample form
5.	Technical Specifications (dully filled, Signed with Stamp)	Yes
6.	Technical Proposal on Bidder's Letterhead	Yes
<b>Bidders Documents</b>		
7.	Manufacturer's Authorization (as per sample form)	Yes
8.	Undertaking (as per sample form)	Yes
9.	Certificate (as per sample form)	Yes
10.	Income Tax & GST Registration Professional Tax Certificate (Sindh)/Certificates are mandatory, Bidder's FBR Status should be ACTIVE (For NTN and Sales Tax)	Yes
11.	Valid PNRA registration certificate where applicable	N/A
12.	Company Profile	Yes
13.	Bank certificate/Statement with last three years turnover.	Bank certificates of both banks May 2015 and June 2013 and Bank Statement 6 months short.
14.	Income Tax Return (last two years)	Yes
15.	Workshop for after sales services	Yes
16.	Technical Team detail	On Simple Paper without sign & Stamp



OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

DETAILED CRITERIA OF TECHNICAL BIDS FOR MACHINERY, EQUIPEMENTS OF ADP SCHEME (PART-II)

OPENED ON 29 MARCH 2018  
M/S IQBAL TRADERS

S.#	ELEGIBILITY PARAMETERS	DETAILS / REMARKS
17.	The Bidder will ensure provide WEBOC ID of Bidder must be active for Imported items.	WEBOC ID is not showing (a) NTN/FTN (b) STRN © Business Address (d) CNIC NO: & Contact #
18.	<b>Original Equipment Manufacturer (OEM or brand quoted)</b> CE / FDS / JIS	Yes
19.	References of offered model or brand (in Pakistan preferable in Sindh)	References are very old i.e 2005 to 2013
<b>E = ELIGIBLE / N= NOT ELIGIBLE</b>		<b>Not Eligible</b>

Prof: Jan Muhammad Memon  
Head Of Department of Surgery  
GMMM College Sukkur  
Member

Dr. Maqsood Ahmed Siddiqui  
Head Of Department Anesthesia  
GMMM College Sukkur  
Member

Dr. Abdul Aziz Memon  
Additional Medical Superintendent  
GMMMC Hospital Sukkur  
Chairman






Prof: Abrar Shaikh  
Head Of Department of Medicine  
GMMM College Sukkur  
Member

Dr. Inayatullah Memon  
Head Of Department of Radiology  
GMMM College Sukkur  
Member

Assistant Professor Radiology  
G.M.M.C Hospital Sukkur

NO. 07-03-2018 AT 01:00PM

PARTICIPANTS

S.NO	NAME	DESIGNATION	DEPARTMENT	CONTACT	SIGNATURE
1.	Dr. Abdul Aziz (Genim)		AMS		
2.	Dr. Abeyasinghe				
3.	Dr. Inayatullah	Assistant Professor		0333-7565346	
4.	Dr. Magsasa-Abad.	Assistant Professor		03337025717	
5.	Prof. Safia Wahid	Prof. Gynecologist		03332027719	

  
MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR

**SCHEME NAMELY 50 BEDDED MEDICAL & SURGICAL ICU MACHINERY EQUIPMENTS & ITEMS FOR FINANCIAL YEAR 2017-18 AT GIMMC HOSPITAL SUKKUR E HELD ON 29-03-2018 UPTO 2:00 PM.**




**FOR BIDDERS**

S.NO	COMPANY	NAME OF REPRESENTATIVE	CONTACT/CELL	SIGNATURE
01	The D Care A & B	Abdul Jabbar	0333 711 3463	[Signature]
02	Medequps SMC	Ahmed Zafar	0300-2120609	[Signature]
03	Jybal and Company	HASAN AHYUB	0334-3091531	[Signature]
04	Elite c.c Pvt	Hussain Ahmed	0332-2726623	[Signature]
05	M/S Rfos	Abdul Wahed	0336-3725873	[Signature]
06	Siejanudin & Co (Pvt)	Spd Saad Gilani	0333-8164865	[Signature]
07	MEDLINE TECHNOLOGIES	SYED FARRUKH	0302-8553608	[Signature]
08	JARRI MEDICAL	ASIF ALI	0313 3434444	[Signature]
09	Human Health Care	M. Kaeed	03322-842729	[Signature]
10	ZAHID AND SONS	AAMIR ALAM	0333-3345581	[Signature]
11	Labdar z.	Prof. Dr. Awi	0304-0083357	[Signature]
12	Platinium of	Abdul sb z.	03337118433	[Signature]
13	Khalid Corporation	M. Shauqee	03002807696	[Signature]
14	BIO TECH SERVICES	Sywan Ahmed	0300-826199	[Signature]
15	M/S M-Tech Engineers (Pvt. A, B)	IMRAN	0334-8906186	[Signature]
16	Pakistan Furniture (Pvt. B)	Faisal	03453886313	[Signature]
17	Radiant Interiors	ISRAEL AHMED	0321-8550360	[Signature]
18	Medi. store	FARID AHMED	0333-7102400	[Signature]

*Q. J. J. J.*  
 MEDICAL SUPERINTENDENT  
 GIMMC HOSPITAL SUKKUR  
 29/3/18

ATTENDANCE SHEET REGARDING TENDER FOR  
PROCUREMENT OF ADP SCHEME NAMELY ESTABLISHMENT OF 50 BEDDED MEDICAL & SURGICAL ICU  
ITEMS OF MACHINERY, EQUIPMENTS & OTHER ITEMS FOR FINANCIAL YEAR 2017-18 AT GMMMC  
HOSPITAL SUKKUR E HELD ON 29-03-2018 AT 3:00 PM.

PROCUREMENT COMMITTEE

S.NO	NAME	DESIGNATION	CONTACT/ CELL	SIGNATURE
1.	DR. M. Anwar Pasha	Medical Supr.	0300-9373939	
2.	AWAIS MUSHTAQ	AC CITY	0300 9439995	
3.	DR. A AZIZ MBBS	AMS	0333762366	





  
MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR





ITEMS OF MACHINERY, EQUIPMENTS & OTHER ITEMS FOR FINANCIAL YEAR 2017-18 AT GMMMC HOSPITAL SUKKUR HELD ON 29-03-2018 AT 3:00 PM.

TECHNICAL COMMITTEE

S.NO	NAME	DESIGNATION	CONTACT/ CELL	SIGNATURE
1.	DR. ABDUL AZIZ	AMS	03337102366	
2.	DR. Abbas	ANESTHESIA	03337028860	
3.	Dr. Inayatullah	ASSI prof. Radiology	0333-7665346	
4.	Mr. Mansoor Ahmed	Assistant Prof Anesthesia	03337027717	
5.				

  
MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR





**OFFICE OF THE MEDICAL SUPERINTENDENT  
GHULAM MUHAMMAD MAHAR MEDICAL  
COLLEGE HOSPITAL SUKKUR**  
PHONE NO. 071-9310213 — FAX NO. 071-9310119  
NO. GMCH/Sukkur/4259/69 DATED: 03 MARCH 2018

**CORRIGENDUM**

In continuation of NIT NO: MS/GMCHS/3112/15 dated 15-02-2018, already published in various Newspapers Dawn, Jung and Kawish.

Description	1. Purchase of Plant, Machinery & Equipments	Cost of Tender Rs. 3000/- for each component
Dated of Sale of Tender Document	From the publication of Tender in Newspapers	
Last date of Sale of Tender	27-03-2018 upto 02:00 PM	
Date of Submission of Tender	29-03-2018 upto 02:00 PM	
Date of Opening	29-03-2018 at 03:00 PM	
Tender Opening Venue	Office of the Medical Superintendent GMMMC Hospital Sukkur	

Whereas all the other terms & conditions are same.

**MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR**

INF-KRY No. 1252/18 Say No to Corruption

**SAY NOT TO CORRUPTION**

Ph: 022-9210207 Fax: 022-9210208

**NOTICE**

The advertisement published in various newspapers under No. Inf/Kry: 1235/2018 for Purchase of Laundry Machine is hereby cancelled. The new date will be announced shortly.

**Medical Superintendent  
Liaquat University Hospital  
Hyderabad / Jamshoro**

INF-KRY No. 1251/18 Say No to Corruption

**KARACHI POLICE**

**JOB OPPORTUNITY**

**SCHEDULE OF LADY SEARCHER /  
POLICE CONSTABLE (BPS-05)**

It is inform to all candidates who submitted their applications for Lady Searcher / Police Constable of 16 remaining vacancies in Training and Recruitment (T&R) Branch, PHQ, Garden Karachi Range published in daily newspapers (Dawn, Jang and Sindh Express, dated 24/25-01-2018) are hereby bound to appear for Physical / Written Test and Interview on 15-03-2018.

All candidates who have applied are hereby requested to please appear at *Police Headquarters Garden, South Zone, Karachi* at 0900 hours as per above schedule.

**Note:**

- No TA/DA will be paid.
- All candidates are advised to bring with them original documents; (Matriculation Certificate, Domicile / PRC and CNIC) in the above test and interview.
- The candidate who fails or absents at any of above tests will not be entitled for the next test.
- In case any candidate is found deficient in any of required standards as per recruitment policy or submission of any fake documents or wrong information will entail her to disqualification at any stage.

**DIGP Administration  
For Addl: Inspector General of Police  
Karachi Range**

INF-KRY No. 1254/18 Say No to Corruption

**OFFICE OF THE MEDICAL SUPERINTENDENT  
GHULAM MUHAMMAD MAHAR MEDICAL  
COLLEGE HOSPITAL SUKKUR**  
PHONE NO. 071-9310213 — FAX NO. 071-9310119  
NO. GMCH/Sukkur/4247/52 DATED: 03 March 2018

**CORRIGENDUM**

In continuation of NIT NO: MS/GMCHS/3116/19 dated 15-02-2018, already published in various Newspapers Dawn, Jung and Kawish.

Description	1. Purchase of Revenue Component of ADP Scheme Namely Establishment of 50 Bedded and Surgical ICU with Expansion of Other Units at GMMMC Hospital Sukkur.	Cost of Tender Rs. 3000/- for each component
Dated of Sale of Tender Document	From the publication of Tender in Newspapers	
Last date of Sale of Tender	27-03-2018 upto 02:00 PM	
Date of Submission of Tender	29-03-2018 upto 02:00 PM	
Date of Opening	29-03-2018 at 03:00 PM	
Tender Opening Venue	Office of the Medical Superintendent GMMMC Hospital Sukkur	

Whereas all the other terms & conditions are same.

**MEDICAL SUPERINTENDENT  
GMMMC HOSPITAL SUKKUR**

INF-KRY No. 1253/18 Say No to Corruption

**NOTICE INVITING TENDER**

The Balochistan Development Authority invites sealed percentage rates tenders based on Composite Schedule of Rates, 1998 Government of Balochistan (Amended upto date) from interested contractors / firms for the works mentioned below:-

**Bid Security 4% and Time Limit 30-06-2018 for each work:**

S.#.	Name of Scheme	E/Cost
1	Construction of black Topping Road Tuka Zareen Road Baboan District Harnai (PSDP No.1230) <b>TS138631E</b>	14.709 M
2	Construction of Road Gochina to Harnai via Talib Leas Short Cut to Harnai District Harnai. (PSDP No.1231) <b>TS138632E</b>	9.802 M
3	Construction of Black Topping Road Shahrage Area Sheikh Mussa, Alizai, Khair Abad, etc District Harnai. (PSDP 2017-18, S.No.1249) <b>TS138633E</b>	19.609 M
4	Const: of B/T Road Zarmana Pari Distt: Harnai (PSDP 2017-18, S.No.1233) <b>TS138634E</b>	9.799 M
5	Construction of Black Topping Road Zarmana Road Warikha District Harnai (PSDP 2017-18, S.No.1250) <b>TS138635E</b>	4.905 M
6	Construction of Black Topping Road Gudda Ghuzza to Sar Kan Road District Harnai (PSDP 2017-18, S.No.1234) <b>TS138636E</b>	9.807 M
7	Construction of Black Topping Road & Cutting Road in Spin Tangi to Wari Khan	14.707 M

**ITY INSTITUTION**

DATED: 28-02-2018

**TICE**

**ULES, 1973**

the following act of misconduct



OFFICE OF THE MEDICAL SUPERINTENDENT GHULAM MUHAMMAD MAHAR MEDICAL COLLEGE HOSPITAL SUKKUR

PHONE NO: 071-9310713 FAX: 9315115

NO:MSGMCHS / SUKKUR / 3115/10

DATE: 24-2-2018

NOTICE INVITING TENDER

The Medical Superintendent, GMMMC Hospital Sukkur hereby invites the sealed bids from interested and eligible bidders who fulfill the eligibility criteria for the Purchase of Revenue Component ( RCP ) Scheme namely Establishment of 50 Bedded Medical & Surgical ICU with Expansion of other units of GMMMC Hospital Sukkur, Plant, Machinery and Fixture for GMMMC Hospital Sukkur under the relevant possession of Sindh Public Procurement Rules 2010 (Amended 2017).

The bids must be submitted on opening date of tenders on or before 16.00 PM upto 2.00 PM which will be opened publicly in the presence of all bidders or their authorized representatives who choose to attend at 2.00 PM in the Office of the Medical Superintendent Ghulam Muhammad Mahar Medical College Hospital Sukkur. All bids must be accompanied by a bid security @ 2% of the total bid cost in shape of pay order/bank guarantee.

The chairman procurement committee reserves the right to post-poned / accept / reject any or all bids under the relevant provision of Rule 10 of Sindh Public Procurement Rules 2010 (Amended 2017).

The complete set of tender enquiry / bidding documents may be purchased from Office of the Medical Superintendent GMMMC Hospital Sukkur on submission of written application on the letterhead of registered firm/company upon cash payment of non-refundable amount mentioned below from the date of publication of advertisement in newspaper.

Table with 2 columns: Description of Tender, and Date of Submission of Tender. Rows include: Bid to be submitted on, Date of Submission of Tender, Date of Opening of Tender, and Date of Submission of Tender.

In case the Chairman Procurement Committee is out of Headquarters or any of its members is not available and if the Govt. employees Public Holiday their tenders will be submitted and opened on next working day following all the same terms and conditions. All Bids shall be issued Government Tenders such as Professional Fee, S&T and GST and others, if applicable. Information regarding this MIT may also be downloaded from GPPRA website www.gppra.gov.pk & Sindh Government Website www.sindh.gov.pk and the website of GMMMC Hospital.

MEDICAL SUPERINTENDENT GMMMC HOSPITAL SUKKUR. Say No to Corruption

Handwritten signature and date: 24-2-2018 DAWAN