






BID EVALUATION REPORT

1.	Name of Procuring Agency:	Dow University of Health Sciences, Karachi.
2.	Tender Reference No:	No. DUHS/DP/2017/24 Dated 12 July 2017
3.	Tender Description:	Procurement of Hospital and Lab. Equipment / Instruments
4.	Method of Procurement:	Single Stage – Two Envelope
5.	Tender Published:	a. SPPRA websites at Sr. No. 33972 / SPPRA ID No. 2147483647; b. Daily Dawn (English Newspaper) on 14 July 2017; c. Daily Jang (Urdu Newspaper) on 14 July 2017; d. Daily Ibrat (Sindhi Newspaper) on 14 July 2017. (PID(K)139/18) (Copy attached as Appendix 1)
6.	Total Bid documents Sold:	Twenty Eight (28) Nos.
7.	Total Bids Received:	Twenty Four (24) Nos.
8.	Technical Bid Opening date:	7 th August 2017 at 11:30 a.m. (Appendix 2)
9.	No. of Bid technically qualified:	Sixteen (16) Nos. (Appendix 3)
10.	Bid(s) Disqualified / Rejected:	Eight (08) Nos. (Appendix 3)
11.	Financial Bid Opening date:	30 October 2017 at 12:00 Noon (Appendix 4)
12.	Bid Evaluation Report:	Attached as Appendix 5

1.	Prof. Dr. Muhammad Masroor Pro-Vice Chancellor (Chairman) Dow University of Health Sciences, Karachi	 Prof. Dr. Muhammad Masroor MBBS, MCPS, FCPS, FRCP (Glasgow) Pro-Vice Chancellor Dow University of Health Sciences Karachi
2.	Professor Amanullah Registrar (Member / Coordinator) Dow University of Health Sciences, Karachi	 Prof. Amanullah Abbasi MBBS, FCPS, MRCP, FRCP Professor of Medicine Registrar Dow University of Health Sciences Karachi.
3.	Mr. Munaf Advani Director Finance (Member) Benazir Bhutto Shaheed University, Lyari, Karachi.	 DIRECTOR FINANCE BENAZIR BHUTTO SHAHEED UNIVERSITY LYARI KARACHI
4.	Mr. Nadeem Shakoor Javeri Director Finance (Member) Dow University of Health Sciences, Karachi	 Nadeem Shakoor Javeri Director Finance Dow University of Health Sciences Karachi
5.	Professor M. Shuja Farrukh Medical Superintendent Dow University Hospital, Ojha Campus, Karachi	 Prof. Dr. M. Shuja Farrukh MBBS, DLO, FCPS Medical Superintendent Dow University Hospital, Ojha Campus Dow University of Health Sciences

**HOISTING ON SPPRA & DUHS WEBSITES
PROCUREMENT OF HOSPITAL AND LAB. EQUIPMENT / INSTRUMENTS
No. DUHS/DP/2017/24 Dated 12 July 2017**

The screenshot shows the SPPRA website interface. At the top, there is a navigation menu with links for HOME, SPPRA, PUBLIC PROCUREMENT, NEWS, SEND QUERY, SPP ACT 2009, SPP RULES, and NOTIFICATIONS. Below the navigation, there is an 'Upcoming Events' section with the text 'C.B. Workshop at Karachi on: 2017 / May / 10-11'. The main content area is titled 'Search Tenders List' and contains a table of search results. The table has columns for Serial, SPPRA ID, City, Department, Advertisement Date, Closing Date, Upload Date, Tender Notice, BER Date, BERs, Corrigendum Date, Corrigendums, and Corr 2 Da. A single record is visible in the table.

Serial	SPPRA ID	City	Department	Advertisement Date	Closing Date	Upload Date	Tender Notice	BER Date	BERs	Corrigendum Date	Corrigendums	Corr 2 Da
33972	2147483647	KARACHI	DOW University of Health Science	2017-07-18 00:00:00	2017-09-07 23:59:59	18-07-2017						

The screenshot shows the DUHS website 'Tenders' section. It contains a list of various procurement notices and tenders, each with a brief description and a link to download bidding documents. The list includes items such as '1. RFP - REQUEST FOR PROPOSALS (APPOINTMENT OF CONSULTANT)', 'CORRIGENDUM (Technical Specification) PROCUREMENT OF HOSPITAL EQUIPMENT/ INSTRUMENTS', and 'PROCUREMENT OF DRUGS / MEDICINES / SURGICAL CONSUMABLES / DISPOSABLES / CONTRAST MEDIA & ALLIED ITEMS'.

- 1. RFP - REQUEST FOR PROPOSALS (APPOINTMENT OF CONSULTANT) (Click here to download the Bidding Documents)
- (CORRIGENDUM) (Technical Specification) PROCUREMENT OF HOSPITAL EQUIPMENT/ INSTRUMENTS (DUHS/DP/2017/26 Dated 21 Sep. 2017)
- NOTICE FOR CANCELLATION OF TENDER FOR DEPLOYMENT OF SECURITY SERVICES (DUHS/P&D/2017/8767 Dated 09-07-2017)
- (CORRIGENDUM) PROCUREMENT OF HOSPITAL EQUIPMENT/ INSTRUMENTS (DUHS/DP/2017/26 Dated 21 Sep. 2017)
- PROCUREMENT OF DRUGS / MEDICINES / SURGICAL CONSUMABLES / DISPOSABLES / CONTRAST MEDIA & ALLIED ITEMS (Ref. No. DUHS/DP/2017/-27 Dated: 02/10/2017) (Click here to download the Bidding Documents)
- PROCUREMENT OF HOSPITAL EQUIPMENT / INSTRUMENT (Ref. No. DUHS/DP/2017/- 26 Dated: 21/09/2017) (Click here to download the Bidding Documents)
- PROCUREMENT OF TWO MODULES OF ACCOUNTING SOFTWARE (GENERAL LEDGER AND PAYROLL) AND HARDWARE (SERVER) (Ref. No. DUHS/P&D/2017/8922 Dated: 08/09/2017) (Click here to download the Bidding Documents)
- NOTICE INVITING TENDER (NIT) PRINTING OF PROSPECTUS MBBS & BDS FOR SESSION 2017-2018 (Ref. No. DUHS/AD/2017/08-339 Dated: 17th August, 2017) (Click here to download the Bidding Documents)
- EXPRESSION OF INTEREST HIRING THE SERVICES OF TESTING AGENCY (Click here to download the Bidding Documents)
- Procurement of Services for Convocation 2017 (Ref. No. DUHS/DP/2017/- 25) (Click here to download the Bidding Documents)
- Procurement of Lab. & Hospital Equipment / Instruments (Ref. No. DUHS/DP/2017/- 24) (Click here to download the Advertisement) (Click here to download the Bidding Documents)



Friday
July 14, 2017
Volume 111, Issue
KARACHI
Page No. 20
No. of Pages
Price Rs. 100/-
www.dawn.com

Published on 14th July, 2017

DOW UNIVERSITY OF HEALTH SCIENCES

NOTICE INVITING TENDERS (NIT)

No. DUHS/DP/2017/24

Dated 12 July 2017

Bids are invited under sealed cover for the Procurement of Hospital and Lab. Equipment/ Instruments on DDP / C&F basis from authorized Dealers / Distributors / Manufacturers, having registration with Federal Board of Revenue (FBR) for Income Tax and Sales Tax.

Tender Fee	Rs. 2,000/- (Rupees two thousand only) Non-Refundable in shape of Pay Order / Demand Draft in favor of Dow University of Health Sciences, Karachi
Bidding Procedure	Single Stage - Two Envelope
Bid Security	2% of the total bid value
Purchasing Date	21 July 2017 to 05 August 2017
Bids Delivery Date & Time	07 August 2017 at 11:00 a.m.
Bid Opening Date & Time	07 August 2017 at 11:30 a.m.

In case of any unforeseen situation or government holiday resulting in closure of office on the date of opening, bids shall be submitted / opened on next working day at the given time.

Bidding document may be purchased by interested bidders on the submission of a written application, NTN, GST Certificates and Pay Order / Demand Draft of tender fee during office hours. Bidding Documents are also available at DUHS and SPPRA websites. Interested Bidders may obtain further information personally from 11:00 A.M. to 02:00 P.M.

Conditional Bids, Telegraphic Bids, Bids not accompanied by Bid Security of required amount and form, bids received after specific date and time and bids of Black Listed firms will be rejected.

The Dow University of Health Sciences, Karachi (DUHS) reserves the right to reject any or all the bids subject to the relevant provisions of SPP Rules 2010 (Amended 2013).

Director Procurement
Procurement Directorate
Prof. Masood Hameed Khan Library Building
Dow University of Health Sciences (Ojha Campus)
SUPARCO Road, off Main University Road,
Gulzar-e-Hijri, Scheme No. 33, Karachi.
Phone No. + 92-21-99261497
Email: director.procurement@duhs.edu.pk

www.duhs.edu.pk
Baba-e-Urdu Road, Karachi - 74200 Pakistan
Tel: 021-99216065 Fax: 021-99216065



Published on 14th July, 2017

DOW UNIVERSITY OF HEALTH SCIENCES

طبی ٹینڈرنوٹس

نمبر: DUHS/DP/2017/24 تاریخ: 12 جولائی 2017

ہسپتال اور ایب ایکو ہسٹ کی فراہمی کیلئے ڈی ڈی پی ای اینڈ ایف کی بنیاد پر انڈیکس اور سٹیکس کیلئے فیڈرل بورڈ آف ریو لنڈ (ایف بی آر) سے ریٹنڈر ایکو ہسٹ چارج کنڈکٹنگ ایگنٹ کے طور پر ڈی ڈی پی ای اینڈ ایف کی فراہمی کیلئے طلب ہے۔

2,000/- (دو ہزار روپے)؛ قابل واپسی	ٹینڈر فیس
پائل سپر ڈارڈر ڈیٹا ڈرافٹ، تمام ڈیٹا ڈرافٹس اور ڈیٹا ڈرافٹس کے ساتھ ساتھ کراچی	بڈنگ طریقہ کار
سنگل اینج، واپس	بولی کی سیکورٹی
کس بولی کا 2%	بولی دستاویزات خریدنے کی تاریخ
21 جولائی 2017 تا 05 اگست 2017	بولیاں پہنچانے کی آخری تاریخ اور وقت
07 اگست 2017، وقت صبح 11:00 بجے	بولی کھنڈنے کی تاریخ اور وقت
07 اگست 2017، وقت صبح 11:30 بجے	

ٹینڈر کھنڈنے کی تاریخ پر آؤٹ ریسٹروکٹس اور ہسپتال کے ساتھ ساتھ ڈی ڈی پی ای اینڈ ایف کی فراہمی کیلئے طلب ہے۔

پہلی رات 10 بجے بولی دیکھ کر بولی دستاویزات اور بولی پاپ سے تحریری درخواست کے ساتھ بولی ڈی ڈی پی ای اینڈ ایف کی فراہمی کیلئے طلب ہے۔ بولی دیکھ کر بولی دستاویزات اور بولی پاپ سے تحریری درخواست کے ساتھ بولی ڈی ڈی پی ای اینڈ ایف کی فراہمی کیلئے طلب ہے۔ بولی دیکھ کر بولی دستاویزات اور بولی پاپ سے تحریری درخواست کے ساتھ بولی ڈی ڈی پی ای اینڈ ایف کی فراہمی کیلئے طلب ہے۔

مشروطاً بولیاں، نئی گراف بڈ، مطلوب بڈ سیکورٹی کے بغیر، مقررہ تاریخ اور وقت کے بعد وصول ہانے والی بولیاں، بلیک لسٹ فرم کی جانب سے دی گئیں بولیاں قابل قبول نہیں ہوگی۔

ڈی ڈی پی ای اینڈ ایف ہسپتال اور ایب ایکو ہسٹ کراچی (ڈی ڈی پی ای اینڈ ایف) کی فراہمی کیلئے ڈی ڈی پی ای اینڈ ایف کی فراہمی کیلئے طلب ہے۔

ڈائریکٹر پروکیورمنٹ
پروکیورمنٹ ڈائریکٹر
پروفیسر مسعود سعید خان ایگزیکٹو ڈائریکٹر
ڈی ڈی پی ای اینڈ ایف ہسپتال اور ایب ایکو ہسٹ کراچی
سپارڈ ڈی ڈی پی ای اینڈ ایف ہسپتال اور ایب ایکو ہسٹ کراچی
فون: +92-21-99261497
ای میل: director_procurement@duhs.edu.pk

www.duhs.edu.pk

Babae-Undu Road, Karachi - 74200 Pakistan
Tel: 021-99216065 Fax: 021-99216065

PIQ/2017/159/15



Published on 14th July, 2017

DOW UNIVERSITY OF HEALTH SCIENCES

ٽينڊر گھرائڻ لاءِ نوٽيس

تاريخ: 12 جولاءِ، 2017

نمبر: DUHS/DPI/2017/24

اسپتال ۽ لیب ایڪویمنٽ جي فراهمي لاءِ ڏي ڏي بي/سي اينڊ ايف جي بنياد تي انڪر ٽيڪس ۽ سيلز ٽيڪس لاءِ فيڊرل بورڊ آف روينيو (ايف بي آر) سان رجسٽرڊ ايڪویمنٽ تيار ڪندڙن/مجاز ڊيلرن/ڊسٽري بيوٽرن کان مھربند واک گھريل آھن.

2000 (ٻه هزار روپيا) نه وڻڻ جوڳا بي آرڊر/ڊمانڊ ڊرافٽ جي صورت ۾ نالي ڊاٽو يونيورسٽي آف هيلٿ سائنسز، ڪراچي	ٽينڊر في
سنگل اسٽيج بہ لافا	بڊنگ جو طريقو
ڪل واک جو 2 سيڪڙو	واڪ جي سيڪيورٽي
21 جولائي 2017 کان 5 آگسٽ 2017	واڪ دستاويز خريد ڪرڻ جي تاريخ
7 آگسٽ 2017 وقت صبح 11:00 وڳي	واڪ پهچائڻ جي آخري تاريخ ۽ وقت
7 آگسٽ 2017 وقت صبح 11:30 وڳي	واڪ کولڻ جي تاريخ ۽ وقت

ٽينڊر کٽڻ جي تاريخ تي جيڪڏهن ٽيڊر اڪائي صرتاً يا حڪومت ٻارڻ مڙسڻ جي پيش نظر آفيس بند ٿيندي ته واک ايندڙ ڏينهن تي ڏنل وقت تي جمع / کٽيا ويندا.

دلچسپي رکندڙ واک ڏيندڙ، واک دستاويز هيٺ ڄاڻايل پتي تان لکت ۾ درخواست سان اين تي اين، جي ايس تي سرٽيفڪيٽس ۽ ٽينڊر في جو بي آرڊر/ڊمانڊ ڊرافٽ آفيس وقت وچ ۾ جمع ڪرائي حاصل ڪري سگهجن ٿا. واک دستاويز ڏي يو ايڇ ايس / ايس بي بي آر اي جي ويب سائٽس تي به دستياب آهن. واک ڏيندڙ وڌيڪ معلومات پنهنجي سر پاڻ صبح 11:00 وڳي کان منجهند 2:00 وڳي تائين حاصل ڪري سگهن ٿا.

شرطي واک، ٽيلي گراف بڊ، گھريل بڊ سيڪيورٽي جي رقم ۽ فارم بنا، مقرر تاريخ ۽ وقت کان پوءِ وصول ٿيندڙ واک، بليڪ لسٽ فرم پاران ڏنل واک رد ڪيا ويندا.

ڊاٽو يونيورسٽي آف هيلٿ سائنسز، ڪراچي (ڊي يو ايڇ ايس) ايس بي بي قانونن 2010 (ترميم ٿيل 2013) جي لاڳاپيل شق تحت ڪنهن هڪ يا سورن واکن کي رد ڪرڻ جو حق محفوظ رکي ٿي.

ڊائريڪٽر پروڪيورمنٽ
پروڪيورمنٽ ڊائريڪٽوريٽ
پروفيسر مسعود حميد خان لائبريري بلڊنگ
ڊاٽو يونيورسٽي آف هيلٿ سائنسز (اوجها ڪيمپس)
سپارڪو روڊ آف مين يونيورسٽي روڊ، گلزار هجري، اسڪيم نمبر 33، ڪراچي

فون: +92-21-99261497
اي ميل: director.procurement@duhs.edu.pk

www.duhs.edu.pk

Baba-e-Urdu Road, Karachi - 74200 Pakistan.
Tel: 021-99216065 Fax: 021-99216065

PID(K) 139/18



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.
ATTENDANCE SHEET

Procurement of Lab. & Hospital Equipment / Instruments
Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF TECHNICAL BIDS / PROPOSALS ON 07 AUGUST 2017 AT 11:30 A.M.

S#	BIDDER (Company / Firm / Contractor)	REPRESENTATIVE (Name)	DESIGNATION	CONTACT NUMBER	SIGNATURE
1-	Qualitron Corporation	Nadeem ul-Haque	Manager Sales	0336-2824143	<i>[Signature]</i>
2-	Technodes Link	Muzaffer Husain	Manager Sales	0333-3135975	<i>[Signature]</i>
3,	Diagnostic System	Bilal Basit	Manager Sales	0333-2024249	<i>[Signature]</i>
④	Bal Chohan & Company	Bilal Khan	Sales Executive	0303-2184504	<i>[Signature]</i>
⑤	Seeed Supply Pvt. Ltd. Muhammad Shafiqe	SHAFIQ ALI QURANI	Sen. Application Specialist	0300-2439810	<i>[Signature]</i>
⑥	Hospital Supply Corporation	Muhammad Shafiqe	Sales & Service Engineer	0333-4989673	<i>[Signature]</i>
7	Popular & Revision	M. ASHRAF	Sales Rep. U	03018279272	<i>[Signature]</i>
8	BMI-Business Medical Int'l	M. Nadeem	Manager	0300-8244148	<i>[Signature]</i>
9	Global Marketing Service	Zaeem Bagci	Sales Engineer	0331-2318513	<i>[Signature]</i>

[Signature]
Prof. Dr. M. Shuja Haque
MBBS, DLO, FCPS, FRCP (Glasgow)
Medical Superintendent,
Dow University Hospital, Ojha Campus
Dow University of Health Sciences

[Signature]
Prof. Dr. Muhammad Masroor
MBBS, DLO, FCPS, FRCP (Glasgow)
Vice-Chancellor
Dow University of Health Sciences
Karachi

[Signature]
DIRECTOR FINANCE
ZAKIR BHUTTO SHAHEED
Director Finance
DOW UNIVERSITY LYARI KARACHI
Dow University of Health Sciences

[Signature]
Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.
ATTENDANCE SHEET

Procurement of Lab. & Hospital Equipment / Instruments
Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF TECHNICAL BIDS / PROPOSALS ON 07 AUGUST 2017 AT 11:30 A.M.

S#	BIDDER (Company / Firm / Contractor)	REPRESENTATIVE (Name)	DESIGNATION	CONTACT NUMBER	SIGNATURE
10	Suze Bio Diagnostic & Pharm.	M. Aamir Saeed	Sales & Service Engr	0333-3659632	
11	Mughal Trader	MUSTAJEEM	Sales + Marketing	0300-3655709	
12	SAJID BHATTI RECH INK	SAJID BHATTI	NATIONAL MANAGER	0300-0997122	
13	Waqas Chemical House	WAQAS KHAN	Sales / Application Specialist	0312-2151143	
14	H.A. Shah & Sons	Fahad Qamar	Asst Mgr. Sales	0334-3305201	
15	Faizur Hussain Lucky Labs Gulistan	Faizur Hussain	MANAGER	0345-2628269	
16	UNIBRAND INK QUAZI	M. OVAIS	Sales & Service Engr.	0333-2714388	
17	Chem Tech Pakistan	Atiqur Mirza	Manager	0300-9201742	
18	Scientec	Mr. SYED ADNAN	Manager	030-0202420	

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent

Dow University Hospital, Ojha Campus
Dow University of Health Sciences

Prof. Dr. Muhammad Nadeem
MBBS, FCPS, FRCGS, FRCR
Director Finance

Dow University of Health Sciences

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.
ATTENDANCE SHEET

Procurement of Lab. & Hospital Equipment / Instruments
Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF TECHNICAL BIDS / PROPOSALS ON 07 AUGUST 2017 AT 11:30 A.M.

S#	BIDDER (Company / Firm / Contractor)	REPRESENTATIVE (Name)	DESIGNATION	CONTACT NUMBER	SIGNATURE
19	INSITEC II	SIAHID IMAN	Executive	03002318938	
20	AMS (Pvt) Ltd	M. Zain Najmeddin	Executive	0332-3080207	
21	CWTC	Muhammad Ali	Executive	03335333198	
22	STG	Kashif	Sales	0300-2287865	
23	Latif Brothers	Farazuddin Shaikh	Sales and Service Expt	0334-2638495	
24	Ray Technologies	Fabish	Sales &	0345-2727064	

Nadeem Shakoor Javeri
Director Finance

Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent

Dow University Hospital, Ojha Campus
Dow University of Health Sciences

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Science
Karachi.



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.

ATTENDANCE SHEET (Procurement Committee)

Procurement of Lab. & Hospital Equipment / Instruments

Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF TECHNICAL BIDS / PROPOSALS ON 07 AUGUST 2017 AT 11:30 A.M.

S#	NAME	DESIGNATION	SIGNATURE
1	Prof. Muhammad Masroor	Pro-Vice Chancellor/ Chairman P.C.	
2	Munaf Advari	D.F. Benazir Bhutto Subject University Lyari	
3	Nadeem Shikoh Tariq	DF. DUHS	
4	Prof. M. Shuja Farooq	Prof of ENT / MS-DNH	
5	Prof. Amrutha	Professor / Neurology	

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT
PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
NIT No. DUHS/DP/2017/-24 Dated 12 July 2017
TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017

APPENDIX 3

GROUP 1

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
1	Small Vertical Electrophoresis Assembly with Power Pack	Chemical House	165-8001 + 170-3935 + 164-5050	Bio-Rad Laboratories,	China	Offer meets the tender requirement, hence acceptable	76	Qualified
1	Small Vertical Electrophoresis Assembly with Power Pack	Rays Technologies	Elite 300	Wealtec	USA	Offered model meets the technical specs with minor deviation (Run up to 2 gels only. Reference from 3 reputed organization is not provided).	73	Qualified
1	Small Vertical Electrophoresis Assembly with Power Pack	Analytical Measuring Systems Pvt. Ltd.	846-017-104	Analytik Jena AG	Germany	Rejected. Run upto 2 gels only. Performance from 3 reputed organizations are not provided. Bid not signed	33	Disqualified
1	Small Vertical Electrophoresis Assembly with Power Pack	Scientific Supplies (Pvt.) Ltd.	V16 (Cat # 21070010 + CBEPS-300X)	Biometra / VWR	Germany	Rejected. Run up to 1 gel only. Reference from 3 reputed organization is not provided	39	Disqualified
1	Small Vertical Electrophoresis Assembly with Power Pack	Scientific Supplies (Pvt.) Ltd.	V16-2 (Cat # 31071010)	Biometra / VWR	Germany	Offered model meets the technical specs with minor deviation (Run up to 2 gels only. Reference from 3 reputed organization is not provided). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification	59	Disqualified
1	Small Vertical Electrophoresis Assembly with Power Pack	Sure Bio-Diagnostics & Pharmaceuticals	P8DS-1 + EC300XL 2	Thermo Scientific	USA	Rejected. Run up to 2 gels only. Blotting module not quoted. Reference from 3 reputed organization is not provided	44	Disqualified
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Chemical House	170-4481 + 170-4416 + 170-4415 + 170-4446 + 170-444/8 + MD-20/25	Bio-Rad Laboratories	China	Offer meets the tender requirement, hence acceptable	76	Qualified
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Rays Technologies	Elite 300	Wealtec	USA	Offered model meets the technical specs with minor deviation (Casting tray is 15 x 15cm. Compatible power pac is upto 300V & 400 mA).	73	Qualified
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Analytical Measuring Systems Pvt. Ltd.	846-025-200	Analytik Jena AG	Germany	Rejected. Tray is 12 x 14 cm. 2 combs of 18 well. Power Pac is max. upto 400V Bid not signed	33	Disqualified
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Scientific Supplies (Pvt.) Ltd.	Horizon 58 (Cat # 41060039)	Biometra	Germany	Rejected. Casting tray is 5.7 x 8.3cm. Comb of 8 and 14 wells.	39	Disqualified
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Scientific Supplies (Pvt.) Ltd.	Horizon 11-14 (Cat # 11068020)	Biometra	Germany	Offered model meets the technical specs with minor deviation (Casting tray is 11 x 14cm with Comb of 1 well). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	59	Disqualified
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Scientific Technical Corporation (Pvt) Ltd.	ME 15-7-10-15	Major Science	USA	Offered model meets the technical specs with minor deviation (Standard form of UV transillumination is not provided with instrument). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification	57	Disqualified

Prof. Dr. Muhammad Masroor Director Finance
 MBBS, MCPS, FCPS, FRCP (D) 1 of 5
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences
 Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi.

GROUP 1

APPENDIX 3

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
3	Thermal Cyclers	H. A. Shah & Sons	Mastercycler Nexus Gradient (6331000017)	Eppendorf AG	Germany	Offer meets the tender requirement, hence acceptable	72	Qualified
3	Thermal Cyclers	H. A. Shah & Sons	Veriti 96 Well (Cat # 4375786)	Life Technology	USA	Offer meets the tender requirement, hence acceptable	76	Qualified
3	Thermal Cyclers	Chemical House	T100 Cat #: 1861096	Bio-Rad Laboratories	China	Offered model meets the technical specs with minor deviation (Block material and lid temperature range is not mentioned. Temp accuracy is 0.5 °C. Evaporation protection lid not present).	71	Qualified
3	Thermal Cyclers	Analytical Measuring Systems Pvt. Ltd.	846-2-070-301	Analytik Jena AG	Germany	Rejected. Block and lid temp. range. Max. number of cycles are not mentioned. Evaporation protection lid is not present. Bid not signed	33	Disqualified
3	Thermal Cyclers	Chem Tech Pakistan	TG1000G	Diab Scientific Inc.	China	Offered model meets the technical specs with minor deviation (Temp accuracy is 0.3 °C. USB / Ethernet connectivity not mentioned. Evaporation protection lid not present). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	43	Disqualified
3	Thermal Cyclers	Diagnostic Systems	PrimeG/02	Cole Parmer/Techn	UK	Rejected. Not compatible with PCR plate. Evaporation protection lid is not present.	43	Disqualified
3	Thermal Cyclers	Lucky Lab. International	96G Gradient	QLS	UK / EU	Rejected. Max. no of cycles not mentioned. Evaporation protection lid is not present	45	Disqualified
3	Thermal Cyclers	Rays Technologies	PEQSTAR, ?????	?????, Germany	Germany	Rejected. Specific model of PEQSTAR is not mentioned. Min temp gradient is 35 °C. Lid temp is 70 - 120 °C. Evaporation protection lid not present.	61	Disqualified
3	Thermal Cyclers	ScienTec	AERIS-BG096	ESCO	Singapore	Offered model meets the technical specs with minor deviation (Block material and maximum no. of cycles are not mentioned. Evaporation protection lid is not present). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	44	Disqualified
3	Thermal Cyclers	Scientific Supplies (Pvt.) Ltd.	TCST-9620	Blue Ray	Taiwan	Rejected. Taiwan made. Block material, max. no of cycles, lid temp range are not mentioned. Temp accuracy is 0.3 °C.	34	Disqualified
3	Thermal Cyclers	Scientific Supplies (Pvt.) Ltd.	TAdvanced 96 G (Cat # 846-070-201)	Biometra	Germany	Rejected. Max no of cycles are not mentioned. Evaporation protection lid not present.	39	Disqualified
3	Thermal Cyclers	Scientific Technical Corporation (Pvt) Ltd.	Alpha Cyclers AC-1	PCRmax Limited (A Bibby Scientific company)	UK	Rejected. Block temp range is 10-100 °C. Block material not mentioned. Evaporation protection lid is not present	32	Disqualified
4	Shaking Refrigerated Incubator	Ins-Tech	Innova 42R (Cat. No. M1335-0016)	Eppendorf Inc.	USA	Offer meets the tender requirement, hence acceptable	76	Qualified
4	Shaking Refrigerated Incubator	Mughal Traders	SI900R	Robus	UK	Rejected. Unit dimensions and temperature range is not as per tender specs	7	Disqualified
4	Shaking Refrigerated Incubator	Scientific Technical Corporation (Pvt) Ltd.	Thermostable 1S-20R	Daihan Scientific, North America Inc.	USA	Rejected. Temp range is 10-60 °C and 20 - 250 RPM	40	Disqualified

Nadeem Shakoor Javeri
Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

Prof. Dr. M. Shuja Farrukh
 MBBS, D.O, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University of Health Sciences
 Karachi

[Signature]
DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Dr. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi.

GROUP 1

APPENDIX 3

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
5	Spectrophotometer	Scientific Technical Corporation (Pvt) Ltd.	EPOCH2TC	Biotech Instruments Inc.	USA	Offer meets the tender requirement, hence acceptable	73	Qualified
5	Spectrophotometer	Sure Bio-Diagnostics & Pharmaceuticals	51119300DP	Thermo Fisher Scientific	USA / Finland	Offered model meets the technical specs with minor deviation (Not for 6 well plate, only for 96 & 384 well microplate. Linear shaking option is available only). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	64	Disqualified
5	Spectrophotometer	Chem Tech Pakistan	SP-UV1100	Dlab Scientific Inc.	China	Rejected. offered model is Not for Miscrowell Plate	23	Disqualified
5	Spectrophotometer	H. A. Shah & Sons	BioSpectometer Basic (Cat # 6135000904)	Eppendorf AG	Germany	Rejected. Wavelength rage is 200 - 830 nm, not for microwell plate	47	Disqualified
5	Spectrophotometer	Qualitron Corporation	V-730 Bio	Jasco	Japan	Rejected. Provision of micro-well plate options along with Linear or orbital shaking were not available.	47	Disqualified
5	Spectrophotometer	Rays Technologies	Colibri	Titertek Berthold	Germany	Rjected. Miscrowell Plate & cuvette option is not available	61	Disqualified
6	- 40 Freezer	Rays Technologies	MDF-U5412	Panasonic	Japan	Offer meets the tender requirement, hence acceptable	86	Qualified
6	- 40 Freezer	Sure Bio-Diagnostics & Pharmaceuticals	7240V	Thermo Scientific	USA	Offer meets the tender requirement, hence acceptable	77	Qualified
6	- 40 Freezer	Global Marketing Services	Plasma Super Artic 400 ECT-F Touch	Fiocchetti	Italy	Rejected. Temp: range -30 to -40 °C	43	Disqualified
6	- 40 Freezer	Lucky Lab. International	OVEVF-510-45	Ovan	Spain / EU	Rejected. Temperature range -35 to -45 °C	45	Disqualified
6	- 40 Freezer	Mughal Traders	UPLTF 275	Arctiko	Denmark	Rejected. 246 Lit. Capacity is not as per tender specs	7	Disqualified
6	- 40 Freezer	ScienTec	HF3-400S-1	ESCO	Singapore	Offered model meets the technical specs with minor deviation. However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification	49	Disqualified
6	- 40 Freezer	Scientific Supplies (Pvt.) Ltd.	AL07-12-230	Advantage	Germany	No document attached in support of product certification and failed to score the 70 marks, which was mandatory requirement for technical qualification.	64	Disqualified
7	Refrigerated Centrifuge	H. A. Shah & Sons	Centrifuge 5810 R (Cat # 5811000010)	Eppendorf AG	Germany	Offer meets the tender requirement, hence acceptable	72	Qualified
7	Refrigerated Centrifuge	Lucky Lab. International	UNI-CEN MR (21 14 100 + 21 14 401 + 21 14 550 + 25 53 80)	Hero Lab.	Germany / EU	Offer meets the tender requirement, hence acceptable	70	Qualified
7	Refrigerated Centrifuge	Rays Technologies	Z366 K	Hermle	Germany	Offer meets the tender requirement, hence acceptable	81	Qualified
7	Refrigerated Centrifuge	Scientific Supplies (Pvt.) Ltd.	ALLEGRA x-30R + FX301.5 + S2096+ C0650 + Adapter Cat #: A99471 + 392274 + 361111 + 364670 + 393088	Beckman Coulter	USA / EU	Offer meets the tender requirement, hence acceptable	72	Qualified
7	Refrigerated Centrifuge	Sure Bio-Diagnostics & Pharmaceuticals	Heraeus Centrifuge Megafuge 16R	Thermo Fisher Scientific / Thermo Heraeus	USA / Germany	Offer meets the tender requirement, hence acceptable	77	Qualified

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

Nadeem Shakir Javeri
Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University of Health Sciences
Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAFEEQ
UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi

GROUP 1

APPENDIX 3

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
7	Refrigerated Centrifuge	Chem Tech Pakistan	D3024R	Diab Scientific Inc.	China	Rejected. Rotor for PCR plate and 15/50ml falcon tube is not quoted.	23	Disqualified
7	Refrigerated Centrifuge	Diagnostic Systems	D3024R	Diab	China	Rejected. Chinese origin. Rotor for PCR plate & 15/50 ml falcon tube is not quoted only 24 place for 2ml tubes.	35	Disqualified
7	Refrigerated Centrifuge	Scientific Technical Corporation (Pvt) Ltd.	K2015R	Centurion Scientific Ltd.	UK	Rejected. Only 24 place rotor for 2ml tubes	37	Disqualified
8	Rotary Evaporator	Continental Worldwide Trading Co.	Rotavapor R-300 with manual lift & 1 liter water bath (95 oC) with Diagnoal Glass assembly (11R300111A001) AND B-301 Water Bath	Buchi	Switzerland	Offer meets the tender requirement, hence acceptable	70	Qualified
8	Rotary Evaporator	Rays Technologies	Hei-VAP HL	Heidolph	Germany	Offer meets the tender requirement, hence acceptable	86	Qualified
8	Rotary Evaporator	Chem Tech Pakistan	RE100	Diab Scientific Inc.	China	Rejected. Max. temp is 180 °C. Distillation assembly not quoted.	23	Disqualified
8	Rotary Evaporator	Continental Worldwide Trading Co.	Rotavapor R-100 with glass assembly (11100V101) and B-100 Water Bath	Buchi	Switzerland	Offered model meets the technical specs, however, no document attached in support of product certification thus failed to score the 70 marks, which was mandatory requirement for technical qualification.	63	Disqualified
8	Rotary Evaporator	Qualitron Corporation	N-1200 AS-W	Eyela	Japan / China	Rejected. Less rpm (20-180). Temperature range not as per requirement (5+RT to 180 °C)	39	Disqualified
8	Rotary Evaporator	ScienTec	RE100 Pro	Dragon Lab.	China	Rejected. Maximum temp. is 180 °C. Distillation assembly not quoted	24	Disqualified
8	Rotary Evaporator	Scientific Technical Corporation (Pvt) Ltd.	WEV1001 V/L	Daihan Scientific, North America Inc.	USA	Rejected. Rotary speed 10-180 RPM. Tem range 5-99 °C.	40	Disqualified
9	Micropipette Set	H. A. Shah & Sons	3120000216 + 3120000232 + 3120000259 + 3120000267	Eppendorf AG	Germany	Offer meets the tender requirement, hence acceptable	72	Qualified
9	Micropipette Set	Rays Technologies	?????	Gilson	France	Offer meets the technical specs with minor deviaiton (Lowest range 0.2ml only. Only HP holder & HP ejection are fully autocolavable).	81	Qualified
9	Micropipette Set	Sure Bio-Diagnostics & Pharmaceuticals	4642010 + 46400304640050 + 4640060	Thermo Fisher Scientific	Finland	Offer meets the technical specs with minor deviaiton (Lowest range 0.1 - 2ul only. Spring loaded tip cone ejection not available).	72	Qualified
9	Micropipette Set	Chem Tech Pakistan	????	VWR	UK	Rejected. Catalogue number not provided to evaluate the offer	23	Disqualified
9	Micropipette Set	Chohan & Company	?????	Merux	France	Rejected. Quoted range of Set is not as per tender specs	25	Disqualified
9	Micropipette Set	Lucky Lab. International	B02-1 + B20-1 + B200-1 +B1000-1	Capp	Dnmark / EU	Offer meets the technical specs with minor deviaiton (Lowest range 0.1 - 2ul only. No spring loaded tip cone ejection not available). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification	65	Disqualified

Prof. Dr. Muhammad Maksoom Shakoor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Director Finance
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University of Health Sciences
 Karachi

DIRECTOR FINANCE
GENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi.

GROUP 1

APPENDIX 3

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
9	Micropipette Set	Popular International (Pvt) Ltd.	19110P + 19120P + 19125P + 19130P + 19140P	Human	Germany	Offer meets the technical specs with minor deviation (Lowest range is 0.5ml, spring loaded HP cone ejection property is not available). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	57	Disqualified
9	Micropipette Set	ScienTec	????	Dragon Lab.	China	Rejected. Chinese origin. Spring loaded tip cone ejection property is not available. Micropipette plus is autoclavable either you have quoted micropipette or micropipette plus. Quoted model is not clear	32	Disqualified
9	Micropipette Set	Scientific Supplies (Pvt.) Ltd.	AP-2 + AP-20 + AP-200 + AP-1000	Axygen	USA	Offer meets the technical specs with minor deviation (Lowest range 0.1 - 2ul only. Spring loaded tip cone ejection not available). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	59	Disqualified
9	Micropipette Set	Scientific Technical Corporation (Pvt) Ltd.	825.0002 + 825.0020 + 825.0200 + 825.1000 + (Serial Acura 825)	Socorex	Switzerland	Offer meets the technical specs with minor deviation (Spring load tip cone ejection of tip property is not available). However, bidder had failed to score the 70 marks, which was mandatory requirement for technical qualification.	57	Disqualified
10	White Light / UV Trans	Lucky Lab. International	UVT-20 (Cat # 2975100 +29 87 300 + 29 85 200 + 29 85 300)	Hero Lab.	Germany / EU	Offer meets the tender requirement, hence acceptable	70	Qualified
10	White Light / UV Trans	Scientific Supplies (Pvt.) Ltd.	846-057-550 + 846-20510-0	Biometra	Germany	Meets EQ, IQ, OQ Standards. Hence acceptable	72	Qualified
10	White Light / UV Trans	Rays Technologies	MD20	Wealtec	USA	Offered model meets the technical specs with minor deviation (White light option is not available).	73	Qualified
10	White Light / UV Trans	Scientific Technical Corporation (Pvt) Ltd.	KA. TN.4LC	Korea ACE Scientific Co. Ltd.	Korea	Rejected. Simple UV lamp not a trans-illuminator. Wavelength range 254 / 365. white light option no available	32	Disqualified

Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

Nadeem Shakoor Javeri
 Director Finance
 Dow University of Health Sciences
 Karachi

Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences
 Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi.

TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT
PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
NIT No. DUHS/DP/2017/-24 Dated 12 July 2017
TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017

APPENDIX 3

GROUP 2

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
1	Fluorescent Microscope	Global Marketing Services	Leica DM1000 + dmc 2900	Leica Microsystems	Germany	Offer meets the tender requirement, hence acceptable	76	Qualified
1	Fluorescent Microscope	Qualitron Corporation	Ci-L	Nikon Corporation	Japan / China	Offer meets the tender requirement, hence acceptable	83	Qualified
1	Fluorescent Microscope	Diagnostic Systems	IS.3153-EPLI/3	Euromex	Netherlands	Rejected. 20x not quoted.	43	Disqualified
1	Fluorescent Microscope	Latif Brothers	Axio Scope A1, Carl Zeiss Microscopy	Carl Zeiss	China	Rejected. Chinese origin. 4X not quoted. Binocular Tube	56	Disqualified
1	Fluorescent Microscope	Lucky Lab. International	LB-264	Labomed Inc.	USA	Rejected. 20x not quoted.	45	Disqualified
1	Fluorescent Microscope	Popular International (Pvt) Ltd.	14800	Human	Germany	Rejected. 4x not quoted. Binocular tube V/S Trinocular tube	37	Disqualified
1	Fluorescent Microscope	Rays Technologies	MT6300	Meiji	Japan	Rejected. 4x not quoted. White LED not available	56	Disqualified
2	Embedding Centre	Technology Links	Histostar (Item Code: A81000001)	ThermoShandon Limited	UK	Offered product is technically acceptable with minor deviation.	70	Qualified
2	Embedding Centre	Scientific Technical Corporation (Pvt) Ltd.	MPS/P2 + MPS/W + MPS/C	Slee Medical GmbH Mainz	Germany	Offered product is technically acceptable with minor deviation.	70	Qualified
2	Embedding Centre	Global Marketing Services	Leica Histocore Arcadia H+C	Leica Microsystems	Germany	Rejected. 1.4 Ltr. V/S 1.5 Ltr. Cool Plate 0 to -6 °C V/S 0 to -15 °C No LED Display	53	Disqualified
2	Embedding Centre	ScienTec	YD-6L	Yidi	China	Rejected. Chinese origin	29	Disqualified
2	Embedding Centre	Uni Brand International	5232-Tissue-Tek TEC-5	Sakura Finetek Japan Co., Ltd.	Japan	Rejected. Parafin Tank 4 Ltr. V/S 5 Ltr. Cool plate temp 0 to -10°C V/S 0 to -15°C	38	Disqualified
3	Dual Head Microscope	Global Marketing Services	Leica DM750	Leica Microsystems	Germany	Offer meets the tender requirement, hence acceptable	71	Qualified
3	Dual Head Microscope	Qualitron Corporation	E-200 with TH-2F	Nikon Corporation	Japan / China	Acceptable with minor deviation (Cross travel vernier calibrated stage not available).	78	Qualified
3	Dual Head Microscope	Diagnostic Systems	IS.1156-PLI2	Euromex	Netherlands	Rejected. 2x & 20x not quoted. Cross travel vernier calibrated stage not available. Trinocular Tube	43	Disqualified
3	Dual Head Microscope	Latif Brothers	Axio Lab A1, Carl Zeiss Microscopy	Carl Zeiss	China	Rejected. Chinese origin. Cross Travel vernier calibrated stage not available	53	Disqualified
3	Dual Head Microscope	Lucky Lab. International	LB-931	Labomed Inc.	USA	Rejected. 2x & 20x not quoted. Cross travel vernier calibrated stage not available	45	Disqualified

Nadeem Shakoor
 Director Finance
 Dow University of Health Sciences

Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Superintendent
 Dow University Hospital, Ojha Campus
 Dow University of Health Sciences


Prof. Dr. Muhammad Anwar
 MBBS, MCPS, FCPS
 Director Finance
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY (VADIKAPACHI)


Prof. Amanullah Abbas
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 University of Health Sciences
 Karachi.


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT
PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
NIT No. DUHS/DP/2017/-24 Dated 12 July 2017
TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017

GROUP 3


Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Remarks	Marks Obtained	Remarks
1	Phototherapy System	Business Medical International (Bmi)	Houva 4*	National Biological Corporation	USA	Offer meets the tender requirement, hence acceptable	77	Qualified
1	Phototherapy System	Latif Brothers	3-Series, X Model 220 – 240 VAC (311/350-24/24)	Daavlin	USA	Offer meets the tender requirement, hence acceptable	87	Qualified
2	IPL & PIP Toning System	Business Medical International (Bmi)	HS-660	Med Apolo	China	Offer meets the tender requirement, hence acceptable	75	Qualified
3	Centrifuge Machine with Rotors	Global Marketing Services	CM-6MT	ELMI Tech.	Lithuania (Europe)	Offer meets the tender requirement, hence acceptable	78	Qualified
3	Centrifuge Machine with Rotors	Hospital Supply Corporation	Genesis Centrifuge 416	Gyrozen Co. Ltd.	Korea	Offer meets the tender requirement, hence acceptable	79	Qualified
3	Centrifuge Machine with Rotors	Rech International	Rotofix 32A (1206- ART)	Arthrex	Germany	Offer meets the tender requirement, hence acceptable	70	Qualified
3	Centrifuge Machine with Rotors	Business Medical International (Bmi)	LC-04P	Genesis Centrifuge	Korea	Rejected. Manufacturer authorization not attached	37	Disqualified
4	Platelet Rich Plasma (PRP) Kits	Global Marketing Services		DR. PRP	Korea	Offer meets the tender requirement, hence acceptable	78	Qualified
4	Platelet Rich Plasma (PRP) Kits	Hospital Supply Corporation	Cure PRP	Good Morning Bio Co. Ltd.	Korea	Offer meets the tender requirement, hence acceptable	79	Qualified
4	Platelet Rich Plasma (PRP) Kits	Rech International	ABS-10011	Arthrex	Germany	Offer meets the tender requirement, hence acceptable	70	Qualified
4	Platelet Rich Plasma (PRP) Kits	Business Medical International (Bmi)	Neo Genesis PRP Series	Neo Genesis PRP Series	Korea	Rejected. Manufacturer authorization not attached	37	Disqualified


Prof. Dr. Muhammad Macroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences


Nadeem Shakoor
 Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi.



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.
ATTENDANCE SHEET

(Procurement of Lab. & Hospital Equipment / Instruments)

Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF FINANCIAL BIDS / PROPOSALS ON 30 OCTOBER, 2017 AT 12:00 NOON

S#	BIDDER (Company / Firm / Contractor)	REPRESENTATIVE (Name)	DESIGNATION	CONTACT NUMBER	SIGNATURE
1	Hospital Supply Corporation	Muhammad Shafiqe	Sales & Service Engineer	0333-4989673	M. J. * Lue
2	Chemical House	wajaz Khan	Sales/ Application	0312-2151143	M. J. * Lue
3	Cure Bio Diagnostic + Pharm	M. Aamir Sayed	Sales / Service Eng.	0333-3659632	Umi
4	Lucky Lab International	M. Samiullah Khan	Application & Service Engineer	0341-2015730	M. J. * Lue
5	BMI, Bawon Medical	M. Nadeem	Manager	0300-8244148	M. J. * Lue
6	Continental World wide Pvt.	S. Zai ulobedi	Ast. Manager Sales	034822821691	M. J. * Lue
7	H.A. Shah & Sons	Fahad Qamar	Ast. Manager Sales	03343305201	M. J. * Lue
8	Qualitron Corp.	Nadeem Akbar Khan	Manager Sales	0336-2824143	M. J. * Lue
9	INSITECH	Shahid Anwar	Executive	0300-2318938	M. J. * Lue
10	Global Marketing Service	Zeem Bagai	sales & service Eng	0331-2318513	M. J. * Lue

Prof. Dr. Muhammad Macroob
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

Nadeem Shakoor Javed
Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University of Health Sciences
Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Dr. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.

ATTENDANCE SHEET

(Procurement of Lab. & Hospital Equipment / Instruments)

Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF FINANCIAL BIDS / PROPOSALS ON 30 OCTOBER, 2017 AT 12:00 NOON

S#	BIDDER (Company / Firm / Contractor)	REPRESENTATIVE (Name)	DESIGNATION	CONTACT NUMBER	SIGNATURE
11	Latif Brothers (Instruments)	Farazuddin Shaikh	Sales & Service Engnr	0334-2638495	
12	RAYS Technology.	Imtiaz	Sales Manager	0300-0550119	
13	UNIBRAND INTL	M. Ovais	Sales & Service Engnr	0345-8568241	
14	Sci'artefi-Supplier.	SUHAN ALI Qureshi	Owner	0300-2439810	
15	Alchem Technology Links Pvt. Ltd.	S Rehan Ahmed	Sales Engineer	0322-2214755	
16	SCIENTIFIC TECHNICAL.	Md. Hashif	Sales	0300-2287865	

Prof. Dr. Muhammad Masood
MBBS, MCPS, FCPS, FRCP (Glasgow) Director Finance
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.



DOW UNIVERSITY OF HEALTH SCIENCES, KARACHI.

ATTENDANCE SHEET (Procurement Committee)

Procurement of Lab. & Hospital Equipment / Instruments

Invitation for Bid (NIT) No: DUHS/DP/2017/-24 Dated 12 July, 2017

OPENING OF FINANCIAL BIDS / PROPOSALS ON 30 OCTOBER, 2017 AT 12:00 NOON.

S#	NAME	DESIGNATION	SIGNATURE
1	Prof. M. Masroor	Prov. / Chairman	
2	Prof. M. Shuja Farooq	MS DUH	
3	Nadeem Shekhar	DF DUHS	
4	Abdul Muneem Adnan	DF BBSUL	
5	Mr. Amal Khan	Professor	

**TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT
PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
NIT No. DUHS/DP/2017/-24 Dated 12 July 2017
TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017 AND FINANCIAL BIDS OPENED ON 30 OCTOBER 2017
GROUP 1**

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Qty.	Currency	Unit Price C&F Basis	Total Price C&F Basis	Unit Price C&F Basis Equal to PRs.	Total Price C&F Basis Equal to PRs.	Unit Price DDP Basis Pak Rs.	Total Price DDP Basis Pak Rs.	Ranking	Recommendations / Remarks
1	Small Vertical Electrophoresis Assembly with Power Pack	Chemical House	165-8001 + 170-3935 + 164-5050	Bio-Rad Laboratories,	China	2	USD	3,068	6,136	323,674	647,348	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
1	Small Vertical Electrophoresis Assembly with Power Pack	Chemical House	165-8001 + 170-3935 + 164-5050	Bio-Rad Laboratories,	China	2	PRs.	-	-	-	-	382,590	765,180	2	Conditional Financial offer, as price will be changed if foreign currency value and taxes will be increased. hence rejected. Higher Bid
1	Small Vertical Electrophoresis Assembly with Power Pack	Rays Technologies	Elite 300	Wealtec	USA	2	PRs.	-	-	-	-	465,000	930,000	3	Highest Bid
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Rays Technologies	Elite 300	Wealtec	USA	3	PRs.	-	-	-	-	120,000	360,000	1	Recommended to purchase being the lowest evaluated responsive bid.
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Chemical House	170-4481 + 170-4416 + 170-4415 + 170-4446 + 170-4448 + MD-20/25+ 164-5070	Bio-Rad Laboratories	China	3	USD	4,288	12,864	452,384	1,357,152	204,750	614,250	2	Rs. 614,250/- (3 units) quoted separately for UV Transilluminator, which will be supplied locally. Higher Bid
2	Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack	Chemical House	170-4481 + 170-4416 + 170-4415 + 170-4446 + 170-4448 + MD-20/25+ 164-5070	Bio-Rad Laboratories	China	3	PRs.	-	-	-	-	757,458	2,272,374	3	Conditional Financial offer, as price will be changed if foreign currency value and taxes will be increased. hence rejected. Highest Bid
3	Thermal Cycler	Chemical House	T100 Cat #: 1861096	Bio-Rad Laboratories	China	2	USD	4,812	9,624	507,666	1,015,332	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
3	Thermal Cycler	Chemical House	T100 Cat #: 1861096	Bio-Rad Laboratories	China	2	PRs.	-	-	-	-	563,940	1,127,880	2	Conditional Financial offer, as price will be changed if foreign currency value and taxes will be increased. hence rejected. Higher Bid
3	Thermal Cycler	H. A. Shah & Sons	Veriti 96 Well (Cat # 4375786)	Life Technology	USA	2	USD	6,200	12,400	654,100	1,308,200	-	-	3	Highest Bid
3	Thermal Cycler	H. A. Shah & Sons	Mastercycler Nexus Gradient (6331000017)	Eppendorf AG	Germany	2	Euro	8,100	16,200	1,002,699	2,005,398	-	-	4	Highest Bid
4	Shaking Refrigerated Incubator	Ins-Tech	Innova 42R (Cat. No. M1335-0016)	Eppendorf Inc.	USA	2	USD	20,700	41,400	2,183,850	4,367,700	-	-	1	Recommended to purchase being the lowest evaluated responsive bid. Cost including 12 Gassing Manifold, Photosynthesis light bank, UV Germicidal and Culture Drawer (2 units each)
5	Spectrophotometer	Scientific Technical Corporation (Pvt) Ltd.	EPOCH2TC	Biotek Instruments Inc.	USA	2	USD	14,900	29,800	1,571,950	3,143,900	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
6	-40 Freezer	Sure Bio-Diagnostics & Pharmaceuticals	7240V	Thermo Scientific	USA	1	USD	7,500	7,500	791,250	791,250	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
6	-40 Freezer	Rays Technologies	MDF-U5412	Panasonic	Japan	1	PRs.	-	-	-	-	925,000	925,000	2	Highest Bid
7	Refrigerated Centrifuge	Rays Technologies	Z366 K	Hermle	Germany	1	Euro	6,625	6,625	820,109	820,109	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
7	Refrigerated Centrifuge	Sure Bio-Diagnostics & Pharmaceuticals	Heraeus Centrifuge Megafuge 16R	Thermo Fisher Scientific / Thermo Heraeus	USA / Germany	1	Euro	8,430	8,430	1,043,550	1,043,550	-	-	2	Highest Bid
7	Refrigerated Centrifuge	Lucky Lab. International	UNI-CEN MR (21 14 100 + 21 14 401 + 21 14 550 + 25 53 80)	Hero Lab.	Germany / EU	1	Euro	9,047	9,047	1,119,928	1,119,928	-	-	3	Highest Bid
7	Refrigerated Centrifuge	H. A. Shah & Sons	Centrifuge 5810 R (Cat # 5811000010)	Eppendorf AG	Germany	1	Euro	11,000	11,000	1,361,690	1,361,690	-	-	4	Highest Bid

Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Registrar
University of Health Sciences
Karachi.

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Karachi

Prof. Dr. Muhammad Maqsood
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

Director Finance
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYAR PAK
Karachi


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT
 PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
 NIT No. DUHS/DP/2017/-24 Dated 12 July 2017

TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017 AND FINANCIAL BIDS OPENED ON 30 OCTOBER 2017

GROUP 1

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Qty.	Currency	Unit Price C&F Basis	Total Price C&F Basis	Unit Price C&F Basis Equal to PRs.	Total Price C&F Basis Equal to PRs.	Unit Price DDP Basis Pak Rs.	Total Price DDP Basis Pak Rs.	Ranking	Recommendations / Remarks
7	Refrigerated Centrifuge	Scientific Supplies (Pvt.) Ltd.	ALLEGRA x-30R + FX301.5 + S2096+ C0650 + Adapter Cat #: A99471 + 392274 + 361111 + 364670 + 393088	Beckman Coulter	USA / EU	1	USD	13,111	13,111	1,383,211	1,383,211	-	-	5	Higest Bid
8	Rotary Evaporator	Rays Technologies	Hei-VAP HL	Heidolph	Germany	1	Euro	10,000	10,000	1,237,900	1,237,900	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
8	Rotary Evaporator	Continental Worldwide Trading Co.	Rotavapor R-300 (11R300111A001) + B-301 + V-300 (11V300000) + FL801 (Julabo, Germany) Distillation Assembly 5 Lit. + 3 0029 58 + 5 1365 04	Buchi	Switzerland	1	PRs.	-	-	-	-	1,573,650	1,573,650	2	Higher Bid
9	Micropipette Set	Sure Bio-Diagnostics & Pharmaceuticals	4642010 + 46400304640050 + 4640060	Thermo Fisher Scientific	Finland	3	USD	340	1,020	35,870	107,610	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
9	Micropipette Set	H. A. Shah & Sons	3120000216 + 3120000232 + 3120000259 + 3120000267	Eppendorf AG	Germany	3	PRs.	-	-	-	-	93,600	280,800	2	Higher Bid
9	Micropipette Set	Rays Technologies	?????	Gilson	France	3	PRs.	-	-	-	-	114,000	342,000	3	Higest Bid
10	White Light / UV Trans	Scientific Supplies (Pvt.) Ltd.	846-057-550 + 846-20510-0	Biometra	Germany	2	PRs.	-	-	-	-	183,000	366,000	1	Recommended to purchase being the lowest evaluated responsive bid.
10	White Light / UV Trans	Rays Technologies	MD20	Wealtec	USA	2	PRs.	-	-	-	-	195,000	390,000	2	Higher Bid


Nadeem Shah
 Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi


DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Prof. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Health Sciences


TECHNICAL EVALUATION / BIDDER'S QUALIFICATION REPORT
PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
NIT No. DUHS/DP/2017/-24 Dated 12 July 2017


TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017 AND FINANCIAL BIDS OPENED ON 30 OCTOBER 2017

GROUP 2


Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Qty.	Currency	Unit Price C&F Basis	Total Price C&F Basis	Unit Price C&F Basis Equal to PRs.	Total Price C&F Basis Equal to PRs.	Unit Price DDP Basis Pak Rs.	Total Price DDP Basis Pak Rs.	Ranking	Recommendations / Remarks
1	Fluorescent Microscope	Qualitron Corporation	Ci-L	Nikon Corporation	Japan / China	1	JPY	2,052,600	2,052,600	1,958,796	1,958,796	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
1	Fluorescent Microscope	Global Marketing Services	Leica DM1000 + dmc 2900	Leica Microsystems	Germany	1	Euro	22,700	22,700	2,810,033	2,810,033	-	-	2	Higher Bid
1	Fluorescent Microscope	Global Marketing Services	Leica DM1000 + dmc 2900	Leica Microsystems	Germany	1	PRs.	-	-	-	-	4,102,410	4,102,410	3	Highest Bid
2	Embedding Centre	Scientific Technical Corporation (Pvt) Ltd.	MPS/P2 + MPS/W + MPS/C	Slee Medical GmbH Mainz	Germany	1	Euro	8,900	8,900	1,101,731	1,101,731	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
2	Embedding Centre	Technology Links	Histostar (Item Code: A81G00001)	ThermoShandon Limited	UK	1	GBP	13,560	13,560	1,877,382	1,877,382	-	-	2	Higher Bid
3	Dual Head Microscope	Qualitron Corporation	E-200 with TH-2F	Nikon Corporation	Japan / China	1	JPY	394,860	394,860	376,815	376,815	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
3	Dual Head Microscope	Global Marketing Services	Leica DM750	Leica Microsystems	Germany	1	Euro	11,900	11,900	1,473,101	1,473,101	-	-	2	Higher Bid
3	Dual Head Microscope	Global Marketing Services	Leica DM750	Leica Microsystems	Germany	1	PRs.	-	-	-	-	2,150,602	2,150,602	3	Highest Bid


Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi


Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences


Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi


DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI



Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.


COMPARATIVE STATEMENT AND PURCHASE RECOMMENDATION
PROCUREMENT OF HOSPITAL & LAB. EQUIPMENT / INSTRUMENTS
 NIT No. DUHS/DP/2017-24 Dated 12 July 2017
TECHNICAL BIDS / PROPOSALS OPENED ON 07 AUGUST 2017 AND FINANCIAL BIDS OPENED ON 30 OCTOBER 2017

GROUP 3

Item #	Name of Equipment	Name of Bidder	Model	Make	Country of Origin	Qty.	Currency	Unit Price C&F Basis	Total Price C&F Basis	Unit Price C&F Basis Equal to PRs.	Total Price C&F Basis Equal to PRs.	Unit Price DDP Basis Pak Rs.	Total Price DDP Basis Pak Rs.	Ranking	Recommendations / Remarks
1	Phototherapy System	Latif Brothers	3-Series, X Model 220 – 240 VAC (311/350-24/24)	Daavlin	USA	1	USD	26,755	26,755	2,822,653	2,822,653	-	-	1	Recommended to purchase being the lowest evaluated responsive bid.
1	Phototherapy System	Business Medical International (Bmi)	Houva 4*	National Biological Corporation	USA	1	USD	37,240	37,240	3,928,820	3,928,820	-	-	2	Higher Bid
1	Phototherapy System	Latif Brothers	3-Series, X Model 220 – 240 VAC (311/350-24/24)	Daavlin	USA	1	PRs.	-	-	-	-	3,998,000	3,998,000	3	Highest Bid
2	IPL & PIP Toning System	Business Medical International (Bmi)	HS-660	Med Apolo	China	1	PRs.	-	-	-	-	1,125,000	1,125,000	1	Recommended to purchase being the lowest evaluated responsive bid.
3	Centrifuge Machine with Rotors	Global Marketing Services	CM-6MT	ELMI Tech.	Lithuania (Europe)	1	Euro	970	970	120,076	120,076	174,699	174,699		Recommended to purchase being the lowest evaluated responsive bid.
4	Platelet Rich Plasma (PRP) Kits	Global Marketing Services		DR. PRP	Korea	50	Euro	29	1,450	3,590	179,496	4,518	225,904		
									2,420		299,572		400,602	1	
3	Centrifuge Machine with Rotors	Hospital Supply Corporation	Genesis Centrifuge 416	Gyrozen Co. Ltd.	Korea	1	PRs.	-	-	-	-	350,000	350,000		Higher Bid
4	Platelet Rich Plasma (PRP) Kits	Hospital Supply Corporation	Cure PRP	Good Morning Bio Co. Ltd.	Korea	50	PRs.	-	-	-	-	7,000	350,000		
													700,000	2	
3	Centrifuge Machine with Rotors	Rech International	Rotofix 32A (1206-ART)	Arthrex	Germany	1	Euro	2,500	2,500	309,475	309,475	-	-		Highest Bid
4	Platelet Rich Plasma (PRP) Kits	Rech International	ABS-10011	Arthrex	Germany	50	Euro	90	4,500	11,141	557,055	-	-		
									7,000		866,530			3	


Nadeem Shakoor Javeri
 Director Finance
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS. DLO. FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University Of Health Sciences


Prof. Dr. Muhammad Masroor
 MBBS. MCPS, FCPS, FRCP (Glasgow)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi


DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


Prof. Amanullah Abbasi
 MBBS. FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi.

**MINUTES OF THE TECHNICAL BID OPENING MEETING
HELD AT PROCUREMENT DIRECTORATE
DOW UNIVERSITY OF HEALTH SCIENCES (OJHA CAMPUS), KARACHI
ON 7 AUGUST 2017 AT 11:30 A.M.**

A meeting was held under the chairmanship of Prof. Dr. Muhammad Masroor, Pro-Vice Chancellor, DUHS / Chairman Procurement Committee on 7th August 2017 at 11:30 a.m. to open and evaluate the bids for the "Procurement of Hospital & Lab. Equipment / Instruments".


The participants include:


PROCUREMENT COMMITTEE


1. Professor Dr. Muhammad Masroor (Chairman)
Pro- Vice Chancellor,
Dow University of Health Sciences, Karachi.
2. Professor Amanullah (Member / Coordinator)
Registrar
Dow University of Health Sciences, Karachi
3. Mr. Munaf Advani (Member)
Director Finance
Benazir Bhutto Shaheed University, Lyari, Karachi.
4. Mr. Nadeem Shakoor Javeri (Co-opted Member)
Director Finance
Dow University of Health Sciences, Karachi
5. Professor M. Shuja Farrukh (Co-opted Member)
Medical Superintendent
Dow University Hospital
Ojha Campus, Karachi

BIDDERS WITH NAME OF REPRESENTATIVE

1. Mr. Shahid Ali Quraishi M/s. Scientific Supplies (Pvt.) Ltd.
2. Mr. Farazuddin Shaikh M/s. Latif Brothers
3. Mr. Syed Adnan Ali M/s. ScienTec
4. Mr. Ather Mirza M/s. Chem Tech
5. Mr. Mustaqeem M/s. Mughal Traders
6. Mr. M. Nadeem M/s. BMi-Business Medical International
7. Mr. Tabish M/s. Rays Technologies
8. Mr. Shahid Imam M/s. Ins-Tech
9. Mr. Muhammad Ali M/s. Continental Worldwide Trading Company
10. Mr. Faizan M/s. Lucky Lab. International


Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.


Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences


DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI


Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi

11.	Mr. Bilal Khan	M/s. Chohan & Company
12.	Mr. M. Ovais	M/s. Unibrand International
13.	Mr. Fahad Qamar	M/s. H. A. Shah & Sons
14.	Mr. Bilal Basit	M/s. Diagnostic Systems
15.	Mr. Nadeemullah Khan	M/s. Qualitron Corporation
16.	Mr. Zaeem Baqai	M/s. Global Marketing Services
17.	Mr. Sajid Bhatti	M/s. Rech International
18.	Mr. Waqas Khan	M/s. Chemical House
19.	Mr. Muzaffar Hussain	M/s. Technology Links (Pvt) Ltd.
20.	Mr. Muhammad Shafique	M/s. Hospital Supply Corporation
21.	Mr. Kashif	M/s. Scientific Technical Corporation (Pvt) Ltd.
22.	Mr. M. Zain Najmuddin	M/s. Analytical Measuring Systems Pvt. Ltd.
23.	Mr. M. Ashraf	M/s. Popular International (Pvt) Ltd.
24.	Mr. M. Aamir Sayeed	M/s. Sure-Bio Diagnostics & Pharmaceuticals

Meeting started with recitation from the Holy QURAN. Chair welcomed the participants and then commenced the proceedings. Forum was apprise that the Notice Inviting Tender (NIT) No. DUHS/DP/2017/-24 Dated 12-07-2017 was hoisted on the DUHS and SPPRA websites at Sr. No. 33972 / SPPRA ID No. 2147483647 and the same was published on 14 July 2017 in 3 widely circulated leading dailies of English, Urdu and Sindhi languages viz. Daily Dawn, Daily Jang and Daily Ibrat.

The participants were informed that:

Following 28 bidders purchased the bidding documents:

1. M/s. Scientific Supplies (Pvt.) Ltd.
2. M/s. Latif Brothers
3. M/s. ScienTec
4. M/s. Chem Tech
5. M/s. Mughal Traders
6. M/s. BMi-Business Medical International
7. M/s. Rays Technologies
8. M/s. Hospicare Systems
9. M/s. Ins-Tech
10. M/s. Continental Worldwide Trading Company
11. M/s. Lucky Lab. International
12. M/s. Chohan & Company
13. M/s. Unibrand International
14. M/s. H. A. Shah & Sons
15. M/s. Faraz Associates
16. M/s. Diagnostic Systems
17. M/s. Qualitron Corporation
18. M/s. Global Marketing Services
19. M/s. Rech International


Prof. Amanullah Abbasi
 MBBS, FCPS, MRCP, FRCP
 Professor of Medicine
 Registrar
 Dow University of Health Sciences
 Karachi


Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Qjha Campus
 Dow University of Health Sciences


DIRECTOR FINANCE
 BENAZIR BHUTTO SHAHEED
 UNIVERSITY LYARI KARACHI


Nadeem Shahkhor Javeri
 Director Finance
 Dow University of Health Sciences
 Karachi






20. M/s. Chemical House
21. M/s. Technology Links (Pvt) Ltd.
22. M/s. Hospital Supply Corporation
23. M/s. Scientific Technical Corporation (Pvt) Ltd.
24. M/s. Analytical Measuring Systems Pvt. Ltd.
25. M/s. Agfa Pakistan (Private) Ltd.
26. M/s. Popular International (Pvt) Ltd.
27. M/s. Sure-Bio Diagnostics & Pharmaceuticals
28. M/s. Karachi Medical Company

The receipt of bids was closed at 11:00 a.m. local time on 07 August 2017 as scheduled. Out of twenty eight (28) firms which purchased the bidding documents, twenty four (24) bidders submitted their bids.

It was briefed to the participants that initially, only the Technical Bids / Proposals will be opened and evaluated in a manner prescribed in advance, without reference to the price and reject any proposal which does not conform to the specified requirements. Financial Bids / Proposals of technically qualified bids shall be opened publicly at a time, date and venue announced and communicated to the bidders in advance. Financial Bids / Proposals found technically disqualified / non-responsive shall be returned un-opened to the respective bidder(s). Accordingly, prior to evaluate bid in depth, preliminary examination of the bids was made by the procurement committee.

Bidders were informed that the detailed technical evaluation will be carried-out and communicate to all the bidders, and thereafter Financial Bids will be opened in presence of the bidder's representatives.

Meeting ended with the vote of thanks to and from the chair.

1.	Prof. Dr. Muhammad Masroor Pro-Vice Chancellor (Chairman) Dow University of Health Sciences, Karachi	 Prof. Dr. Muhammad Masroor MBBS, MCPS, FCPS, FRCP (Glasgow) Pro-Vice Chancellor Dow University of Health Sciences Karachi
2.	Professor Amanullah Registrar (Member / Coordinator) Dow University of Health Sciences, Karachi	 Prof. Amanullah Abbasi MBBS, FCPS, MRCP, FRCP Professor of Medicine Registrar Dow University of Health Sciences Karachi.
3.	Mr. Munaf Advani Director Finance (Member) Benazir Bhutto Shaheed University, Lyari, Karachi.	 DIRECTOR FINANCE BENAZIR BHUTTO SHAHEED UNIVERSITY LYARI KARACHI
4.	Mr. Nadeem Shakoor Javeri Director Finance (Member) Dow University of Health Sciences, Karachi	 Nadeem Shakoor Javeri Director Finance Dow University of Health Sciences Karachi
5.	Professor M. Shuja Farrukh Medical Superintendent Dow University Hospital, Ojha Campus, Karachi	 Prof. Dr. M. Shuja Farrukh MBBS, DLO, FCPS Medical Superintendent Dow University Hospital, Ojha Campus Dow University Of Health Sciences


**MINUTES OF THE FINANCIAL BID OPENING MEETING
HELD AT PROCUREMENT DIRECTORATE
DOW UNIVERSITY OF HEALTH SCIENCES (OJHA CAMPUS), KARACHI
ON 30 OCTOBER 2017 AT 12:00 NOON**

A meeting was held under the chairmanship of Prof. Dr. Muhammad Masroor, Pro-Vice Chancellor, DUHS / Chairman Procurement Committee, on 30 October 2017 at 12:00 Noon to open and evaluate the financial bids for the "Procurement of Hospital & Lab. Equipment / Instruments".

The participants include:

PROCUREMENT COMMITTEE

1. Professor Dr. Muhammad Masroor (Chairman)
Pro- Vice Chancellor,
Dow University of Health Sciences, Karachi.
2. Professor Amanullah (Member / Coordinator)
Registrar
Dow University of Health Sciences, Karachi
3. Mr. Munaf Advani (Member)
Director Finance
Benazir Bhutto Shaheed University, Lyari, Karachi.
4. Mr. Nadeem Shakoor Javeri (Co-opted Member)
Director Finance
Dow University of Health Sciences, Karachi
5. Professor M. Shuja Farrukh (Co-opted Member)
Medical Superintendent
Dow University Hospital
Ojha Campus, Karachi


Prof. Amanullah Abbasi
MBBS, FCPS, MRCP, FRCP
Professor of Medicine
Registrar
Dow University of Health Sciences
Karachi.

BIDDERS WITH NAME OF REPRESENTATIVE

1. Mr. Shahid Ali Quraishi M/s. Scientific Supplies (Pvt.) Ltd.
2. Mr. Farazuddin Shaikh M/s. Latif Brothers
3. Mr. M. Nadeem M/s. BMi-Business Medical International
4. Mr. Imtiaz M/s. Rays Technologies
5. Mr. Shahid Imam M/s. Ins-Tech
6. Mr. S. Zain ul abedin M/s. Continental Worldwide Trading Company
7. Mr. M. Samiullah Khan M/s. Lucky Lab. International
8. Mr. M. Ovais M/s. Unibrand International
9. Mr. Fahad Qamar M/s. H. A. Shah & Sons


Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCF
Medical Superintendent
Dow University Hospital, Ojha Cam
Dow University Of Health Scienc


DIRECTOR FINANCE


Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi

10. Mr. Nadeemullah Khan M/s. Qualitron Corporation
11. Mr. Zaeem Baqai M/s. Global Marketing Services
12. Mr. Waqas Khan M/s. Chemical House
13. Mr. S. Rehan Ahmed M/s. Technology Links (Pvt) Ltd.
14. Mr. Muhammad Shafique M/s. Hospital Supply Corporation
15. Mr. Kashif M/s. Scientific Technical Corporation (Pvt) Ltd.
16. Mr. M. Aamir Sayeed M/s. Sure-Bio Diagnostics & Pharmaceuticals

Meeting started with recitation from the Holy QURAN. Chair welcomed the participants and then commenced the proceedings. It was informed to the forum that Technical Bids were opened on 7th August 2017 for the above said tender, where out of twenty eight (28) firms which purchased the bidding document, twenty four (24) bidders submitted their bids.

The details of qualified / disqualified bids for supply of Hospital and Lab. Equipment / Instruments are as under;

GROUP - 1

Item #	Name of Bidder	Model	Make	Country of Origin	Marks Obtained	Remarks
1 Small Vertical Electrophoresis Assembly with Power Pack						
1	Chemical House	165-8001 + 170-3935 + 164-5050	Bio-Rad Laboratories,	China	76	Qualified
2	Rays Technologies	Elite 300	Wealtec	USA	73	Qualified
3	Analytical Measuring Systems Pvt. Ltd.	846-017-104	Analytik Jena AG	Germany	33	Disqualified
4	Scientific Supplies (Pvt.) Ltd.	V16 (Cat # 21070010 + CBEPS-300X)	Biometra / VWR	Germany	39	Disqualified
5	Scientific Supplies (Pvt.) Ltd.	V16-2 (Cat # 31071010)	Biometra / VWR	Germany	59	Disqualified
6	Sure Bio-Diagnostics & Pharmaceuticals	P8DS-1 + EC300XL 2	Thermo Scientific	USA	44	Disqualified
2 Medium (10 x 15 cm) Horizontal Electrophoresis Assembly with Power Pack						
2	Rays Technologies	Elite 300	Wealtec	USA	73	Qualified
1	Chemical House	170-4446 + 170-4446 + 170-4415 + 170-4446 + MD-20/25	Bio-Rad Laboratories	China	76	Qualified
3	Analytical Measuring Systems Pvt. Ltd.	846-025-200	Analytik Jena AG	Germany	33	Disqualified
4	Scientific Supplies (Pvt.) Ltd.	Horizon 58 (Cat # 41060039)	Biometra	Germany	39	Disqualified
5	Scientific Supplies (Pvt.) Ltd.	Horizon 11-14 (Cat # 11068020)	Biometra	Germany	59	Disqualified

Prof. Dr. Muhammad Macroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCP
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University of Health Sciences

#	Name of Bidder	Model	Make	Country of Origin	Marks Obtained	Remarks
6	Scientific Technical Corporation (Pvt) Ltd.	ME 15-7-10-15	Major Science	USA	57	Disqualified
3 Thermal Cyclers						
1	H. A. Shah & Sons	Mastercycler Nexus Gradient (6331000017)	Eppendorf AG	Germany	72	Qualified
2	H. A. Shah & Sons	Veriti 96 Well (Cat # 4375786)	Life Technology	USA	76	Qualified
3	Chemical House	T100 Cat #: 1861096	Bio-Rad Laboratories	China	71	Qualified
4	Analytical Measuring Systems Pvt. Ltd.	846-2-070-301	Analytik Jena AG	Germany	33	Disqualified
5	Chem Tech Pakistan	TG1000G	Dlab Scientific Inc.	China	43	Disqualified
6	Diagnostic Systems	PrimeG/02	Cole Parmer/Techne	UK	43	Disqualified
7	Lucky Lab. International	96G Gradient	QLS	UK / EU	45	Disqualified
8	Rays Technologies	PEQSTAR, ?????	??????, Germany	Germany	61	Disqualified
9	ScienTec	AERIS-BG096	ESCO	Singapore	44	Disqualified
10	Scientific Supplies (Pvt.) Ltd.	TCST-9620	Blue Ray	Taiwan	34	Disqualified
11	Scientific Supplies (Pvt.) Ltd.	TAdvanced 96 G (Cat # 846-070-201)	Biometra	Germany	39	Disqualified
12	Scientific Technical Corporation (Pvt) Ltd.	Alpha Cyclers AC-1	PCRmax Limited (A Bibby Scientific company)	UK	32	Disqualified
4 Shaking Refrigerated Incubator						
1	Ins-Tech	Innova 42R (Cat. No. M1335-0016)	Eppendorf Inc.	USA	76	Qualified
2	Mughal Traders	SI900R	Robus	UK	7	Disqualified
3	Scientific Technical Corporation (Pvt) Ltd.	Thermostable 1S-20R	Daihan Scientific, North America Inc.	USA	40	Disqualified
5 Spectrophotometer						
1	Scientific Technical Corporation (Pvt) Ltd.	EPOCH2TC	Biotek Instruments Inc.	USA	73	Qualified
2	Sure Bio-Diagnostics & Pharmaceuticals	51119300DP	Thermo Fisher Scientific	USA / Finland	64	Disqualified
3	Chem Tech Pakistan	SP-UV1100	Dlab Scientific Inc.	China	23	Disqualified
4	H. A. Shah & Sons	BioSpectrophotometer	Eppendorf AG	Germany	47	Disqualified
5	Qualitron Corporation	30-Bio	Jasco	Japan	47	Disqualified
6	Rays Technologies	Colibri	Titertek Berthold	Germany	61	Disqualified
6 - 40 Freezer						
1	Rays Technologies	MDF-U5412	Panasonic	Japan	86	Qualified
2	Sure Bio-Diagnostics & Pharmaceuticals	7240V	Thermo Scientific	USA	77	Qualified

Prof. Dr. Muhammad Masroor
 MBBS, MCPS, FCPS, FRCP (Glas. W)
 Pro-Vice Chancellor
 Dow University of Health Sciences
 Karachi

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Nadeem Phaloor Javeri
 Director Finance
 Dow University of Health Sciences
 Karachi

Prof. Dr. M. Shuja Farrukh
 MBBS, DLO, FCPS
 Medical Superintendent
 Dow University Hospital, Ojha Campus
 Dow University of Health Sciences

Item #	Name of Bidder	Model	Make	Country of Origin	Marks Obtained	Remarks
3	Global Marketing Services	Plasma Super Artic 400 ECT-F Touch	Fiocchetti	Italy	43	Disqualified
4	Lucky Lab. International	OVEVF-510-45	Ovan	Spain / EU	45	Disqualified
5	Mughal Traders	UPLTF 275	Arctiko	Denmark	7	Disqualified
6	ScienTec	HF3-400S-1	ESCO	Singapore	49	Disqualified
7	Scientific Supplies (Pvt.) Ltd.	AL07-12-230	Advantage	Germany	64	Disqualified
7 Refrigerated Centrifuge						
1	H. A. Shah & Sons	Centrifuge 5810 R (Cat # 5811000010)	Eppendorf AG	Germany	72	Qualified
2	Lucky Lab. International	UNI-CEN MR (21 14 100 + 21 14 401 + 21 14 550 + 25 53 80)	Hero Lab.	Germany / EU	70	Qualified
3	Rays Technologies	Z366 K	Hermle	Germany	81	Qualified
4	Scientific Supplies (Pvt.) Ltd.	ALLEGRA x-30R + FX301.5 + S2096+ C0650 + Adapter Cat #: A99471 + 392274 + 361111 + 364670 + 393088	Beckman Coulter	USA / EU	72	Qualified
5	Sure Bio-Diagnostics & Pharmaceuticals	Heraeus Centrifuge Megafuge 16R	Thermo Fisher Scientific / Thermo Heraeus	USA / Germany	77	Qualified
6	Chem Tech Pakistan	D3024R	Dlab Scientific Inc.	China	23	Disqualified
7	Diagnostic Systems	D3024R	Dlab	China	35	Disqualified
8	Scientific Technical Corporation (Pvt) Ltd.	K2015R	Centurion Scientific Ltd.	UK	37	Disqualified
8 Rotary Evaporator						
1	Continental Worldwide Trading Co.	Rotavapor R-300 with manual lift & 1 liter water bath (95 oC) with Diagonal Glass assembly (11R300111A001) AND B-301 Water Bath	Buchi	Switzerland	70	Qualified
2	Rays Technologies	Heidolph	Heidolph	Germany	86	Qualified
3	Chem Tech Pakistan	D3024R	Dlab Scientific Inc.	China	23	Disqualified
4	Continental Worldwide Trading Co.	Rotavapor R-100 with glass assembly (11100V101) and B-100 Water Bath	Buchi	Switzerland	63	Disqualified
5	Qualitron Corporation	N-1200 AS-W	Eyela	Japan / China	39	Disqualified
6	ScienTec	RE100 Pro	Dragon Lab.	China	24	Disqualified
7	Scientific Technical Corporation (Pvt) Ltd.	WEV1001 V/L	Daihan Scientific, North America Inc.	USA	40	Disqualified

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (G), FRCR
Pro-Vice-Chancellor
Dow University of Health Sciences
Karachi

DIRECTOR FINANCE

BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University of Health Sciences

Item #	Name of Bidder	Model	Make	Country of Origin	Marks Obtained	Remarks
9	Micropipette Set					
1	H. A. Shah & Sons	3120000216 + 3120000232 + 3120000259 + 3120000267	Eppendorf AG	Germany	72	Qualified
2	Rays Technologies	?????	Gilson	France	81	Qualified
3	Sure Bio-Diagnostics & Pharmaceuticals	4642010 + 46400304640050 + 4640060	Thermo Fisher Scientific	Finland	72	Qualified
4	Chem Tech Pakistan	?????	VWR	UK	23	Disqualified
5	Chohan & Company	?????	Merux	France	25	Disqualified
6	Lucky Lab. International	B02-1 + B20-1 + B200-1 + B1000-1	Capp	Dnmark / EU	65	Disqualified
7	Popular International (Pvt) Ltd.	19110P + 19120P + 19125P + 19130P + 19140P	Human	Germany	57	Disqualified
8	ScienTec	?????	Dragon Lab.	China	32	Disqualified
9	Scientific Supplies (Pvt.) Ltd.	AP-2 + AP-20 + AP-200 + AP- 1000	Axygen	USA	59	Disqualified
10	Scientific Technical Corporation (Pvt) Ltd.	825.0002 + 825.0020 + 825.0200 + 825.1000 + (Serial Acura 825)	Socorex	Switzerland	57	Disqualified
10	White Light / UV Trans					
1	Lucky Lab. International	UVT-20 (Cat # 2975100 +29 87 300 + 29 85 200 + 29 85 300)	Hero Lab.	Germany / EU	70	Qualified
2	Scientific Supplies (Pvt.) Ltd.	846-057-550 + 846-20510-0	Biometra	Germany	72	Qualified
3	Rays Technologies	MD20	Wealtec	USA	73	Qualified
4	Scientific Technical Corporation (Pvt) Ltd.	KA. TN.4LC	Korea ACE Scientific Co. Ltd.	Korea	32	Disqualified

GROUP 2

1	Fluorescent Microscope					
1	Global Marketing Services	Leica DM1000 + DMC 2900	Leica Microsystems	Germany	76	Qualified
2	Qualitron Corporation	Ci-L	Nikon Corporation	Japan / China	83	Qualified
3	Diagnostic Systems	IS.3153-FRLi/3	Euromex	Netherlands	43	Disqualified
4	Latif Brothers	Leica Scope A1, Carl Zeiss Microscopy	Carl Zeiss	China	56	Disqualified
5	Lucky Lab. International	LB-264	Labomed Inc.	USA	45	Disqualified
6	Popular International (Pvt) Ltd.	14800	Human	Germany	37	Disqualified
7	Rays Technologies	MT6300	Meiji	Japan	56	Disqualified

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Gastro.)
Pro-Vice Chancellor
Dow University of Health Sciences

DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farru
MBBS, DLO, FC
Medical Superintendent
Dow University Hospital, Ojha Can

Item #	Name of Bidder	Model	Make	Country of Origin	Marks Obtained	Remarks
2 Embedding Centre						
1	Technology Links	Histostar (Item Code: A81000001)	ThermoShandon Limited	UK	70	Qualified
2	Scientific Technical Corporation (Pvt) Ltd.	MPS/P2 + MPS/W + MPS/C	Slee Medical GmbH Mainz	Germany	70	Qualified
3	Global Marketing Services	Leica Histocore Arcadia H+C	Leica Microsystems	Germany	53	Disqualified
4	ScienTec	YD-6L	Yidi	China	29	Disqualified
5	Uni Brand International	5232-Tissue-Tek TEC-5	Sakura Finetek Japan Co., Ltd.	Japan	38	Disqualified
3 Dual Head Microscope						
1	Global Marketing Services	Leica DM750	Leica Microsystems	Germany	71	Qualified
2	Qualitron Corporation	E-200 with TH-2F	Nikon Corporation	Japan / China	78	Qualified
3	Diagnostic Systems	IS.1156-PLi2	Euromex	Netherlands	43	Disqualified
4	Latif Brothers	Axio Lab A1, Carl Zeiss Microscopy	Carl Zeiss	China	53	Disqualified
5	Lucky Lab. International	LB-931	Labomed Inc.	USA	45	Disqualified

GROUP 3

1 Phototherapy System						
1	Business Medical International (BMi)	Houva 4*	National Biological Corporation	USA	77	Qualified
2	Latif Brothers	3-Series, X Model 220 – 240 VAC (311/350-24/24)	Daavlin	USA	87	Qualified
IPL & PIP Toning System						
1	Business Medical International (BMi)	HS-660	Med Apolo	China	75	Qualified
Centrifuge Machine with Rotors						
1	Global Marketing Services	CM-6MT	ELMI Tech.	Lithuania (Europe)	78	Qualified
2	Hospital Supply Corporation	Genesis Centrifuge 416	Gyrozen Co. Ltd.	Korea	79	Qualified
3	Rech International	Rotofix 32A (1206-ART)	Arthrex	Germany	70	Qualified
4	Business Medical International (BMi)	LC-04P	Genesis Centrifuge	Korea	37	Disqualified
Platelet Rich Plasma (PRP) Kits						
1	Global Marketing Services		Neo Genesis PRP Series	Korea	78	Qualified
2	Hospital Supply Corporation		Good Morning Bio Co. Ltd.	Korea	79	Qualified
3	Rech International		Arthrex	Germany	70	Qualified
4	Business Medical International (BMi)		Neo Genesis PRP Series	Korea	37	Disqualified

Prof. Dr. Muhammad Masroor
MBBS, MCPS, FCPS, FRCP (Glasgow)
Pro-Vice Chancellor
Dow University of Health Sciences
Karachi






DIRECTOR FINANCE
BENAZIR BHUTTO SHAHEED
UNIVERSITY LYARI KARACHI

Nadeem Shakoor Javeri
Director Finance
Dow University of Health Sciences
Karachi

Prof. Dr. M. Shuja Farrukh
MBBS, DLO, FCPS
Medical Superintendent
Dow University Hospital, Ojha Campus
Dow University Of Health Sciences






As per the agenda, Technically Qualified offers were taken into consideration by the committee exclusively. The rates quoted by the bidders were read aloud along with amount of 2% bid security attached by the bidder. Each rate in the comparative statement was confirmed from the original bid by the committee and recommendations were made, which were documented in the Bid Evaluation Report (BER).

The meeting ended with the vote of thanks to the chair.

1.	Prof. Dr. Muhammad Masroor Pro-Vice Chancellor (Chairman) Dow University of Health Sciences, Karachi	 Prof. Dr. Muhammad Masroor MBBS, MCPS, FCPS, FRCP (Glasgow) Pro-Vice Chancellor Dow University of Health Sciences Karachi
2.	Professor Amanullah Registrar (Member / Coordinator) Dow University of Health Sciences, Karachi	 Prof. Amanullah Abbas MBBS, FCPS, MRCP, FRCP Professor of Medicine Registrar Dow University of Health Sciences Karachi
3.	Mr. Munaf Advani Director Finance (Member) Benazir Bhutto Shaheed University, Lyari, Karachi.	 DIRECTOR FINANCE BENAZIR BHUTTO SHAHEED UNIVERSITY LYARI KARACHI
4.	Mr. Nadeem Shakoor Javeri Director Finance (Member) Dow University of Health Sciences, Karachi	 Nadeem Shakoor Javeri Director Finance Dow University of Health Sciences Karachi
5.	Professor M. Shuja Farrukh Medical Superintendent Dow University Hospital, Ojha Campus, Karachi	 Prof. Dr. M. Shuja Farrukh MBBS, DLO, FCPS Medical Superintendent Dow University Hospital, Ojha Campus Dow University Of Health Sciences

CERTIFICATE

This is to certify that the prices are comparable to the market rates in respect of single bid received for the Procurement of Hospital and Lab. Equipment / Instruments, in response to NIT No. DUHS/DP/2017/- 24 Dated 12 July 2017, for which Technical Bids were opened on 07 August 2017 and Financial Bids were opened on 30 October 2017.

1.	Prof. Dr. Muhammad Masroor Pro-Vice Chancellor (Chairman) Dow University of Health Sciences, Karachi	 Prof. Dr. Muhammad Masroor MBBS, MCPS, FCPS, FRCP (Glasgow) Pro-Vice Chancellor Dow University of Health Sciences Karachi
2.	Professor Amanullah Registrar (Member / Coordinator) Dow University of Health Sciences, Karachi	 Prof. Amanullah Abbasi MBBS, FCPS, MRCP, FRCP Professor of Medicine Registrar Dow University of Health Sciences Karachi.
3.	Mr. Munaf Advani Director Finance (Member) Benazir Bhutto Shaheed University, Lyari, Karachi.	 DIRECTOR FINANCE BENAZIR BHUTTO SHAHEED UNIVERSITY LYARI KARACHI
4.	Mr. Nadeem Shakoor Javeri Director Finance (Member) Dow University of Health Sciences, Karachi	 Nadeem Shakoor Javeri Director Finance Dow University of Health Sciences Karachi
5.	Professor M. Shuja Farrukh Medical Superintendent Dow University Hospital, Ojha Campus, Karachi	 Prof. Dr. M. Shuja Farrukh MBBS, DLO, FCPS Medical Superintendent Dow University Hospital, Ojha Campus Dow University Of Health Sciences



Office of the Registrar

Dow University of Health Sciences, Karachi.

Baba-e-Urdu Road
Karachi-74200
Pakistan
Phone: 021-99216901
Facsimile: 021-99215763
registrar@duhs.edu.pk

Ref. No. DUHS/Reg./2017/-08-42

Dated: 15th August, 2017.

MEMORANDUM

Subject: PROCUREMENT COMMITTEE, DUHS.

With the approval of the Competent Authority, a Procurement Committee is hereby re-constituted comprising of the following composition with immediate effect and until further orders:-

- | | | |
|----|---|--------------------|
| 1. | Pro-Vice Chancellor,
Dow University of Health Sciences,
Karachi. | Chairman |
| 2. | Registrar,
Dow University of Health Sciences,
Karachi. | Member/Coordinator |
| 3. | Director Finance,
Benazir Bhutto Shaheed University
Lyari, Karachi. | Member |
| 4. | Director Finance,
Dow University of Health Sciences,
Karachi. | Co-opted Member |
| 5. | Concerned Dean/Principal/Director
Dow University of Health Sciences,
Karachi. | Co-opted Member |

2. The Committee shall perform its functions in accordance with the SPP Rules 2010 (Amended 2013).

(Professor Amanullah)
Registrar,
Dow University of Health Sciences
Karachi.

Copy to:

1. The Chairman of the Committee.
2. All Members of the Committee.
3. PA to the Vice Chancellor, DUHS.
4. PA to the Pro-Vice Chancellor, DUHS.



OFFICE OF THE VICE CHANCELLOR
DOW UNIVERSITY OF HEALTH SCIENCES

Baba-e-Urdu Road, Karachi-74200 Pakistan. **Direct No.** 021-32745824 **Fax:** 99215763
Tel: 99215754-57 **Ext:** 3002 **Website:** www.duhs.edu.pk, **E-mail:** vc@duhs.edu.pk

No. DUHS/VC/2017/- 11- 01

Dated 02 November 2017

The Pro-Vice Chancellor /
Chairman Procurement Committee
Dow University of Health Sciences
Karachi.


Subject: Extension in bid validity period for Procurement of Hospital and Lab. Equipment / Instruments N.I.T No. DUHS/DP/2017/24 (SPPRA Sr. No. 33972).

Permission is hereby accorded for extension in bid validity period for further 30 days in respect of NIT No. DUHS/DP/2017/24 Dated 12 July 2017 (SPPRA Sr. No. 33972) for the **Procurement of Hospital and Lab. Equipment / Instruments.**

With best regards.

Yours sincerely,




Prof. Mohammed Saeed Quraishy
Vice Chancellor

cc. The Registrar, DUHS.
cc. The Director Finance, DUHS.
cc. The Director Procurement, DUHS.

Latif Brothers (Instruments)

Approved Government Contractors
And Manufacturers Representatives

G.P.O. Box 394, 14-Commercial Buildings Shahrah-e-Quaid-e-Azam, Lahore, Pakistan.

Tel: (042) 35406972, 35407027, 35407028, 37225950, Fax: (042) 35407024, 35407025, E-mail: pakistani@pakistani.org.pk



LBK/036/1760

November 02, 2017

The Director Procurement
Dow University of Health Sciences
Karachi

Dear Sir,

Sub: Extension of Bid Validity

Ref. Tender No. DUHS/DP/2017/24 dated 12 July 2017 Procurement of
LAB/Hospital Equipment

This is in reference to the above mentioned subject. We do hereby extend the
validity of our Quot # LB/852-1728 dated 04-08-17 till November 30, 2017.

Thanking you, very sincerely yours

Yours faithfully

For **LATIF BROTHERS (Instruments)**

BRANCHES

Suite No. 202, Marine Pride, Plot No. BC-2,
Block No. 7, Clifton, **KARACHI**
Tel: (021) 35379589
Fax: (021) 35379590

H. No. 113/5, Saadi Road,
RAWALPINDI CANTT
Tel: (051) 5130267-68
Fax: (051) 5120556

Apt. # B-1, First Floor,
Town Centre, Main Abdara Road,
University Town, **PESHAWAR**
Tel/Fax: (091) 5842877



Global Marketing Services

111 Hali Road Westridge 1, UAN: +92 51 111 145 236 | email: info@gms-world.com
Rawalpindi Cantt, Pakistan. Fax: +92 51 8314413 | website: www.gms-world.com



November 2, 2017

Director Procurement
Procurement Directorate
Dow University of Health Sciences (Ojha Campus)
Karachi.

Subject: Extension of Bid validity.

Dear Sir,

Reference to tender No.N.I.T No.DUHS/DP/2017/-24 for the procurement of Lab & Hospital Equipment/Instruments opening date 07.08.2017 ,we M/s Global Marketing Services extend the bid validity for further period of 60 days from the date of this letter.

Regards.

For Global Marketing Services

Tanveer Subhani
Manager Institutions



Recd: 2-11-2017.

Directin Procurement
DUHS, Karachi.

Subject: Ref. DUHS/DP 2017/-24
Extension of Validity

Dear Sir,

With Ref: DUHS/DP 2017/-24. Submitted
our quotation No SSR-1284-08/2017.
are
Price Validated till 31st Dec, 2017.

(Signature)
(For Scientific Supplies)



RAYS TECHNOLOGIES

205-B, 2nd Floor, Anum Classic,
Opp. Duty Free Shop,
Shahrah-e-Faisal, Karachi-Pakistan.
Tel: 0092-21-34391383-34388739
Fax: 0092-21-34535882
Email: info@raystechno.com
Website: www.raystechno.com

To,

Procurement Director

Dow University of Health Sciences,

Karachi

Subject: Price Validity Letter


Sir,

*We are pleased to inform you that our quoted prices in **Tender No. DUHS/DP/2017/- 24 Dated 12 July 2017** are valid further one month. (Nov 30th 2017)*

As we remain

Thanking you

For


Rays Technologies



QUALITRON CORPORATION

Tel: 021-3453-9991 to 94 (4 Lines)
Fax: 021-34547747
E-mail: qualtra@super.net.pk
URL: www.qualitron.com.pk

172-E, Block-2, Tariq Road,
PECHS, Karachi-75400 Pakistan.



REF. : VofD/DUHS-Khi/2017

October 02 '2017

Dow University of Health Sciences
Karachi.

REF : TENDER ENQUIRY # DUHS/DP/2017/24 DATED AUGUST 07 '2017 FOR
NIKON MICROSCOPE.
SUB : EXTENSION OF BID VALIDTY OF DATE.

Dear Sir,

With reference to the above tender, we are pleased to inform you that we have extended our date of validity of bid for **One Month**.

While thanking you, we remain,

Very truly yours

QUALITRON CORPORATION



M. TAZAM SHAIKH
(Proprietor.)

DUHS-Khi/17



...empowering diagnostics

Sure Bio-Diagnostics & Pharmaceuticals

November 2, 2017

Director Procurement,
Dow University of Health Sciences,
Karachi - Pakistan.

**Subject: EXTENSION IN TENDER QUOTATION VALIDITY PERIOD AGAINST
BID NO. DUHS/DP/2017/24 DATED 12 JULY 2017.**

This is with reference to Bid No. DUHS/DP/2017/24 Dated 12th July 2017 & Sure Bio Quote No. 133-2017-DUHS dated August 01, 2017 for the "PROCUREMENT OF LAB. & HOSPITAL EQUIPMENT / INSTRUMENTS", as requested to extend the validity period of tender quotation for further 30 days extension with effective from 04-11-2017. So we hereby confirm that we are pleased to extend validity period of above quotation for further 30 days with effective from 04-11-2017 till 04-12-2017.

You are requested to please consider our above quotation with validity till 04-12-2017.

Assuring you of our best services.

Thanks and best regards,

For
Shakeel Ahmed
Chief Executive



Muhammad Irfan Akbar
Manager Customer Relations



To,
Procurement Directorate,
Dow University of Health Sciences,
Karachi, Pakistan.

Date: 02nd November, 2017
Ref No: 0031/CHS/2018

Subject: Extension of Bid Validity Period against your Tender Ref. DUHS/DP/2017/24 Dated 12-July 2017


Dear Sir,

This is with reference to extending of bid validity period of our offer against your tender number subject cited above. It is hereby inform you that we are submitting the letter for the extend bid validity period of our offer No. T-CHS-5912-04-08-2017, up to thirty (30) days further, i.e. 05th December 2017 (as per SPP Rules)

We sincerely hope that it may fulfill your desire. If you need any further assistance please let us know.

Thanking you.

Sincerely,


For: Chemical House

Shoaib Rana, ★
Business Manager (CEO).

HEAD OFFICE:
6-C, Skandar Malhi Road, Canal Park,
Gulberg-E, Lahore - 54660 Pakistan
G.P.O. Box 1138 Lahore - Pakistan
Tel: +92-42-35755146-47, 35752919-20,
35763112-3 Fax: +92-42-35762675
E-mail: info@chemical-house.com

KARACHI OFFICE:
Flat # 1, Mezzanine Floor, 1-C Sunset, Lane # 8,
Phase # 2 (Ext.) Main Commercial Avenue,
D.H.A. Karachi - 75500 Pakistan
Tel: +92-21-35407892, 35396848
Fax: +92-21-35396848

RAWALPINDI OFFICE:
House No. Old NA/258, New No. 389,
New Mill Pur, Salafita Town,
Rawalpindi-Pakistan
Tel: +92-51-4453627, 5891371

PESHAWAR OFFICE:
Flat # 1, 1st Floor, Qamar Mansion,
Opposite UBL Bank, Saddar Road,
Peshawar (Cantt) - 25000 Pakistan
Tel: +92-91-5277858
Fax: +92-91-5277858

Scientific Technical Corporation (Pvt.) Ltd.

Deals in: Scientific, Analytical, Medical, Industrial, Research Equipment



Ref: STC-DUHSK(24)/2017-18(MW)

Dated: 02-11-2017

Director Procurement,
Dow University of Health Sciences (Ojha Campus)
SUPARCO Road, off Main University Road,
Gulzar-e-Hijri, Scheme No. 33,
Karachi.
Tel: 021-99261497

Subject: Bid Extension Letter

Dear Sir,

With the reference of your email, Dated: 02-11-2017 to your Tender No. DUHS/DP/2017/-24,
Due On: 07-08-2017, for the "Procurement of Hospital and Lab. Equipments/Instruments" For
which Technical Proposals were opened on 07th August 2017 and Financial Bid were opened on 30th October,
2017.

We confirm you Bid validity extension till 05-12-2017.

Thanking you and waiting for your valued order.

Yours truly,

MUHAMMAD WASEEM
SALES MANAGER
SCIENTIFIC TECHNICAL CORPORATION (PVT.) LTD.