

PHONE NO.0233-9290097.  
Fax # 0233-9290098

**OFFICE OF THE**  
**EXECUTIVE DISTRICT OFFICER HEALTH,**  
**DISTRICT GOVERNMENT MIRPURKHAS.**

Dated, 13/06 2011.

NO:EDOH/Store/MPS/ 6298/102


To,

The 1. M/S Muslim Medical Services. 2 M/S Sindh Medical Store. 3 M/S Cosmos Enterprises. 4 M/S Global Marketing.

Subject:-


ACCEPTANCE / REJECTION OF BIDS OFFERED BY YOUR FIRM IN TENDER (INSTRUMENTS / EQUIPMENT MACHINERY FOR THE UP GRADATION OF BLOOD BANK DHQ HOSPITAL MIRPURKHAS FOR F.Y.2010.11.

Please find enclosed herewith detail of acceptance / rejection of bids technical & Financial offered by your Firm as per evaluation report copy enclosed herewith.

  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

CC to:

The Deputy Secretary, (PM&I Cell) Barrack # 24 Sindh Secretariat @ Karachi.  
The Administrator / District Coordination Officer District Mirpurkhas.  
The Director (CB) Sindh Public Procurement Regulatory Authority,  
Block #.08 Sindh Secretariat No. 4-A Court Road @ Karachi.

  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

PHONE NO.0233-9290097.  
Fax # 0233-9290098


**OFFICE OF THE**  
**EXECUTIVE DISTRICT OFFICER HEALTH,**  
**DISTRICT GOVERNMENT MIRPURKHAS.**  
Dated, 13/06/2011.

NO:EDOII/ Store /MPS/ 62/03/106  
To,


The 1. M/S SM Trader. 2 M/S Platinum . 3 M/S Oassis Medical System. 4 Ms Lab Link Enterprises. 5 M/S Hospi Care Systems. 6 SK Trading. 7 M/S Shah Enterprises.

Subject:-  
REJECTION OF ALL THE TECHNICAL BIDS OFFERED BY YOUR FIRM  
(TENDER (INSTRUMENTS / EQUIPMENT MACHINERY FOR THE UP  
GRADATION OF BLOOD BANK DHQ HOSPITAL MIRPURKHAS) FOR  
F.Y.2010.11.

As all the technical bids offered by your firm have been rejected by the technical evaluation committee, therefore you are requested to receive back your financial bid against receipt from the office of the undersigned at any time of your convenience.

  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

CC to:  
The Deputy Secretary,(PM&I Cell) Barrack # 24 Sindh Secretariat @ Karachi.  
The Administrator /District Coordination Officer District Mirpurkhas.  
The Director (CB) Sindh Public Procurement Regulatory Authority,  
Block #.08 Sindh Secretariat No, 4-A Court Road @ Karachi.

  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.



S.K. TRADING CORPORATION

H-1, 1st Floor, Mayfair Centre,  
Dr. Dawood Pota Road,  
Saddar, Karachi-74400, Pakistan.  
Tel : (92-21) 32254042, 32220265  
Fax : (92-21) 32232446  
E-mail: info@sk-tc.com  
Website: www.sk-tc.com

To,  
E.D.O. (HEALTH),  
MIRPURKHAS.

Date: 24-06-2011

Tender No. EDOH/Procurement/MPS/4248/-

SUBJECT: REFUND OF EARNEST MONEY

Dear Sir

We have participated in the tender enquiry opened on 19-04-11 Poh#: 4511071 and deposited Rs. 264,620/- as earnest money in favour of E.D.O. (HEALTH) MIRPURKHAS

So please take necessary action in our favor and release earnest money as soon as possible, thanking you in anticipation, we remain.

Yours Sincerely  
S. K. Trading Corporation



*Mr. Abdul Rehman  
CMB of accounts  
27/28/6/11*

*MC  
Relief  
27/6/11*

*27/6/11*



*K*

PHONE NO.0233-9290097.  
Fax # 0233-9290098

**OFFICE OF THE**  
**EXECUTIVE DISTRICT OFFICER**  
**DISTRICT GOVERNMENT**  
Dated, 29/6-2011.

NO:EDOH/ Store /MPS/ 6813/16

To,

The Director (CB)  
Sindh Public Procurement Regulatory Authority,  
Block #.08 Sindh Secretariat No, 4-A Court Road  
@ Karachi.

Subject: -

ILLEGAL / MALAFIDE REJECTION OF OUR TECHNICAL  
PROPOSAL IN THE TENDER OF UP GRADATION BLOOD BANK @  
CIVIL HOSPITAL MIRPURKHAS.

Reference:

Your kind Office Letter No. Dir (CB-0675)SPPRA 3-5( MPS)/ 10.11/3125 dated 25.06.2011 on  
the Subject captioned above.

I have the honour to submit herewith photocopies of detail of reasons for the rejection of  
Technical bids offered by all the participating firms including the complainant.

*efe*

29/6

EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

CC to:

The Secretary Local Government Department Gos Karqachi.  
The Administrator /District Coordination Officer District Mirpurkhas.  
Additional Secretary (M&I cell) Health Department Karachi.

*efe*

29/6

EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

PHONE NO.0233-9290097.  
Fax # 0233-9290098

**OFFICE OF THE**  
**EXECUTIVE DISTRICT OFFICER**  
**DISTRICT GOVERNMENT MIRPURKHA.**  
Dated, \_\_\_\_\_ 2011.

NO:EDOII/ Store /MPS/

6918/22

Dated, 28-6-2011

To,

The M/S SK Trading Corporation,  
H- 1<sup>st</sup> floor Mayfair Center  
Dr Dawood pota Road Saddar  
Karachi.

Subject: -


REJECTION OF ALL THE TECHNICAL BIDS OFFERED BY YOUR FIRM ( TENDER (INSTRUMENTS / EQUIPMENT MACHINERY FOR THE UP GRADATION OF BLOOD BANK DHQ HOSPITAL MIRPURKHAS) FOR F.Y.2010.11.

Reference:

Your Letter No. Nil dated 16.06.2011 on the Subject captioned above.


by your firm.

Please find enclosed herewith detail of reasons for the rejection of technical bids offered

  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

CC to:

The Secretary Local Government Department Gos Karachi,  
The Administrator /District Coordination Officer District Mirpurkhas.  
Additional Secretary (M&I cell) Health Department Karachi.

  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

PHONE NO.9290097. Fax # 9290098  
NO:EDO/Procurement  
MPS \_\_\_\_\_  
Dated, \_\_\_\_\_ 2011.

OFFICE OF THE,  
EXECUTIVE DISTRICT OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.

NOTING PROCESS.

Subject:- REQUEST FOR APPROVAL OF LOWEST EVALUATED BIDS TENDER (  
INSTRUMENTS/ QUIPMENT / MACHINERY) FOR THE UP-GRADATION OF  
BLOOD BANK DHQ HOSPITAL MIRPURKHAS.

*I have the honour to submit herewith:*

1. *That evaluation report on technical bids is attached herewith as annexure "A".*
2. *That comparative statement on lowest evaluated bids is attached herewith as annexure "B".*


Therefore it is requested to approve lowest evaluated bids for purchase,  
as per SPPRA rules.

*Submitted for approval*

  
Dr Shafquat Hussain Daheri  
Executive District Officer Health  
District Mirpurkhas

  
Mr Syed Zakir Ali Shah  
Executive District Officer  
(Finance & Planning)  
District Mirpurkhas.

  
Mr Ghulam Mustafa Abro  
Executive District Officer  
(Works & Services)  
District Mirpurkhas.

  
District Administrator /  
District Coordination Officer  
District Mirpurkhas

PHONE NO.9290097. Fax # 9290098

NO:EDOH/Procurement /MPS

Dated, 04/6/2011.

Comactor

OFFICE

EXECUTIVE DISTRICT COORDINATION OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.

To,

1. Mr Syed Zakir Ali Shah,  
Executive District Officer  
Finance & Planning Mirpurkhas.
2. Mr Ghulam Mustafa Abro  
Executive District Officer  
Works & Services Mirpurkhas.

Subject:-

MINUTES OF MEETING HELD ON 30.05.2011 AT 01:PM IN THE OFFICE OF THE  
DISTRICT COORDINATION OFFICER MIRPURKHAS REGARDING OPENING  
OF TENDER (EQUIPMENT / INSTRUMENTS / MACHINERY) FOR THE UP  
GRADATION OF BLOOD BANK DHQ HOSPITAL MIRPURKHAS.


1. A total of 12 technical bids were received.
2. Representatives of 12 participating firms were present.
3. M/S Seico Scientific Karachi submitted financial bid instead of technical bid hence was rejected on the spot and returned to the relevant bidder against receipt.
4. Every bid envelop was opened by the hand of relevant bidder and thereafter bid papers were signed and fool proofed by all the members of the district purchase committee.
5. 6<sup>th</sup> of June 2011 was fixed up for the opening of financial bids and intimated to all the bidders. TIME 01:PM

Meeting ended with vote of thanks to and from the chair.

  
Dr. Shafquat Hussain Daheri  
Executive District Office Health  
District Mirpurkhas.

CC:to:

The Administrator / District Coordination Officer District Mirpurkhas.

  
Dr. Shafquat Hussain Daheri  
Executive District Office Health  
District Mirpurkhas.

PHONE NO.9290097. Fax # 9290098

NO:EDOIL/Procurement /MPS *6208/10*

Dated, *11-6-* 2011.

**OFFICE OF THE**  
**EXECUTIVE DISTRICT OFFICER HEALTH,**  
**DISTRICT GOVERNMENT MIRPURKHAS.**

To,

1. The Executive District Officer,  
(F&P) District Mirpurkhas.
2. The Executive District Officer,  
( W&S) District Mirpurkhas.

Subject:- MINUTES OF THE MEETING HELD ON 06.06.2011 IN THE OFFICE THE DISTRICT COORDINATION OFFICER REGARDING OPENING OF FINANCIAL BIDS (TENDER EQUIPMENT / INSTRUMENTS & MACHINERY FOR THE FY 2010.11.

1	<i>Tender Box containing was opened at 1:00pm in presence of the attending bidders after verification of the seal by the bidders.</i>
2	<i>A total No, of 11envelops were taken out of the tender box,</i>
3	<i>Seven Financial bids against which technical bids had already been rejected were not opened and kept aside for returning back un-opened to the relevant bidders.</i>
4	<i>Remaining four financial bids were opened by the hands of the relevant bidders.</i>
5	<i>Quoted rates and call deposits were announced.</i>
6	<i>Meeting ended with vote of thanks to and from the chair</i>

*11/6/11*  
*J*  
DR SHAFQUAT HUSSAIN DAHERI  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.

CC to:

The District Administrator / District Coordination Officer District Mirpurkhas.

*11/6/11*  
*J*  
DR SHAFQUAT HUSSAIN DAHERI  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT MIRPURKHAS.




Report evaluation committee regarding technical bids / tender Equipment Instrument and Machinery for the up gradation of Blood Bank DHQ Hospital Mirpurkhas and Laboratory Dialysis Center At Civil Hospital Mirpurkhas.


Tender #	Name of Items	M/S Muslim Medical Services	M/S Sindh Medical Store	M/S SM Trader	M/S Platinum	M/S Cosmos Enterprises (Offer No.1)	M/S Cosmos Offer No.2	M/S Oasis Medical System	M/S Global Marketing	M/s Lab Link Enterprises	M/S Hospi Care Systems	SK Trading	M/S Shah Enterprises
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Hecht-Germany Against tender specification / qualification criteria	1 Make Heimer USA 2 against tender specification	NO	HEIMLE-GERMANY'S GMA-GERMANY Against tender specification / qualification criteria	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1 Make Drexel against tender specification/ qualification criteria	Make Sorval/ ThermoFisher Germany as per tender specification / qualification criteria	Against tender specification / qualification criteria	Make Heimer against specification	Hecht-Korea Against tender specification	Against tender specification / qualification criteria
2	Platelet Agitator	NO	1 Make Heimer USA 2 as per tender specification	NO	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1 Make Heimer Against qualification criteria	Make Heimer USA BMS Canada as per tender specification / qualification criteria	Against tender specification / qualification criteria	Boehr USA against qualification criteria	Heimer USA qualification criteria	Against tender specification / qualification criteria
3	Platelet Agitator Incubator	NO	1 Make Heimer USA 2 as per tender specification	NO	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1 Make Heimer Against qualification criteria	Make Heimer USA / BMS Canada as per tender specification / qualification criteria	NO	Boehr USA against qualification criteria	Heimer USA qualification criteria	Against tender specification / qualification criteria
4	Blood Bag Expresser	NO	1 Make _____ 2 against tender specification	NO	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1 Make Corcon Germany against qualification criteria	NO	Against tender specification / qualification criteria	Boehr USA Against tender specification / qualification criteria	Quattri - Italy Against tender specification	Against tender specification / qualification criteria
5	Tube Sealer	NO	1 Make Corcon Italy 2 against tender specification	NO	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1 Make Dilon Germany against qualification criteria	1 Make Corcon Korea as per tender specification qualification criteria	Against tender specification / qualification criteria	Make Genesis Bunker against tender specification / qualification criteria	Decon-Italy against qualification criteria	Against tender specification / qualification criteria
6	Weighing Scale	NO	NO	NO	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1 Make M. Radweg Germany against tender specification	NO	NO	NO	NO	Against tender specification / qualification criteria
7	Table Top Centrifuge	Dahan Korea Against tender specification / qualification criteria	1 Make Sigma Germany As Per specification	BOECO-GERMANY Against tender specification / qualification criteria	SIGNA GERMANY Against tender specification / qualification criteria	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make Centaur Scientific UK against tender qualification	NO	Against tender specification / qualification criteria	Make Sigma against qualification criteria	NPWH Poland Against tender specification	Against tender specification / qualification criteria


8	Table Top Centrifuge	Hecht-Germany Against tender specification / qualification criteria	1. Make Sigma Germany 2. against tender specification	BOECO-GERMANY Against tender specification / qualification criteria	SIGMA-GERMANY Against tender specification / qualification criteria	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make Centurion scientific UK, against tender qualification	NQ	Against tender specification / qualification criteria	Make Seax against qualification criteria	MPW- Poland Against tender specification	Against tender specification / qualification criteria
9	Water Bath-Large	Haver-China as per tender specification / qualification criteria	1. Make GS research Korea 2. against tender specification	MEDLINE-UK Against tender specification / qualification criteria	Mettler Germany Against tender specification	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make Memmert Germany, against qualification criteria.	NQ	Against tender specification / qualification criteria	NQ	Biosan-Lava Against tender specification / qualification criteria	Against tender specification / qualification criteria
10	Water Bath-Small	Haver-China as per tender specification / qualification criteria	1. Make GS research Korea 2. against tender specification	MEDLINE-UK Against tender specification / qualification criteria	Mettler Germany Against tender specification	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make Memmert Germany, against qualification criteria.	NQ	Against tender specification / qualification criteria	NQ	Biosan-Lava Against tender specification / qualification criteria	Against tender specification / qualification criteria
11	Microscope	Motic Against tender specification / qualification criteria	NQ	BOECO-GERMANY Against tender specification / qualification criteria	Mett Japan Against tender specification	1. Make BX Optical UK against tender specification	Against tender specification / qualification criteria	Make Otoka Italy, against tender specification	Make Leica Germany as per tender specification / qualification criteria	Against tender specification / qualification criteria	NQ	Olympus Against tender specification	Against tender specification / qualification criteria
12	Blood Bag Shaker	NQ	1. Henstock / Dakone Italy as per tender specification	NQ	NQ	NQ	Against tender specification / qualification criteria	1. Make Decon Italy, against tender specification / qualification criteria	Make Centron Korea as per tender specification / qualification criteria	NQ	Genesis Against tender specification / qualification criteria	Dalton-Italy against qualification criteria	Against tender specification / qualification criteria
13	Plasma Freezer with temp chart recorder	Haver-China Against tender specification / qualification criteria	1. Make Italy Zagari Against tender specification	LABLOGIC-FRANCE Against tender specification / qualification criteria	Sanyo Japan Against tender specification	1. Make Fouchelet Italy 2. as per tender specification / qualification criteria	Against tender specification / qualification criteria	1. Make Argemont Italy, against tender specification	NQ	Against tender specification / qualification criteria	NQ	Bisim - Korea Against tender specification / qualification criteria	Against tender specification / qualification criteria
14	Plasma Freezer	Haver-China Against tender specification / qualification criteria	Make Italy as per tender specification / qualification criteria	LABLOGIC-FRANCE Against tender specification / qualification criteria	Sanyo Japan Against tender specification	1. Make Fouchelet Italy 2. as per tender specification / qualification criteria	Against tender specification / qualification criteria	1. Make Argemont Italy, against tender specification	NQ	Against tender specification / qualification criteria	NQ	Melting-China against qualification criteria	Against tender specification / qualification criteria
15	Blood Bank Refrigerator	Haver-China Against tender specification / qualification criteria	1. Make Finned Italy BMS Canada as per tender specification / qualification criteria	NQ	Sanyo Japan, Against tender specification.	1. Make Vison scientific south Korea against tender specification	Against tender specification / qualification criteria	1. Make Argemont Italy, against tender specification	NQ	Against tender specification / qualification criteria	Make Finned against qualification criteria	Melting-China against qualification criteria	Against tender specification / qualification criteria


16	Blood Bank Refrigerator	Heier-China Against tender specification / qualification criteria	Make: Homer USA BMS Canada as per tender specification / qualification	NO		Sony Japan Against tender specification	1. Make: Vison specific south korea as per tender specification / qualification criteria	Against tender specification / qualification criteria	1. Make: Angeon Italy against tender Specification	NO	Against tender specification / qualification criteria	Make Finnet against qualification criteria	Meling-China against qualification criteria	Against tender specification / qualification criteria
17	Pharmacy Refrigerator	Heier-China Against tender specification / qualification criteria	1. Make: CCI USA 2. against qualification criteria	NO		Sony Japan Against tender specification	1. Make: Vison specific south korea against tender specification / qualification criteria	Against tender specification / qualification criteria	1. Make: Angeon Italy against tender Specification	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Meling Against tender specification / qualification criteria	Against tender specification / qualification criteria
18	Blood Donation Chair/Venation couches	NO	NO	NO	NO	NO	NO	Against tender specification / qualification criteria	Against tender specification / qualification criteria	NO	NO	NO	NO	Against tender specification / qualification criteria
19	Computer	NO	NO	NO	NO	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria
20	Printer	NO	NO	NO	NO	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria
21	Printer	NO	NO	NO	NO	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria
22	Generator	NO	NO	NO	NO	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria
23	Split Air conditioner	NO	NO	NO	NO	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria	NO	NO	Against tender specification / qualification criteria

24	Gri Card System	N-Q	1. Diemed / Swiss As per tender specification / qualification criteria	N-Q	N-Q	N-Q	Against tender specification / qualification criteria	N-Q	N-Q	N-Q	N-Q	N-Q	Against tender specification / qualification criteria
25	Weighing Machine	N-Q	N-Q	N-Q	N-Q	N-Q	Against tender specification / qualification criteria	N-Q	N-Q	N-Q	N-Q	N-Q	Against tender specification / qualification criteria
26	Stabilizer	N-Q	N-Q	N-Q	N-Q	N-Q	Against tender specification / qualification criteria	N-Q	N-Q	N-Q	N-Q	N-Q	Against tender specification / qualification criteria
27	Haematology Analyzer	N-Q	1. Make: Med France As per tender specification / qualification criteria	N-Q	N-Q	1. Make: Nikon Koncon Japan against tender specification	Make BGT Germany against tender specification / qualification criteria	1. Make: Cos-Tec / Nikon Japan against tender specification / qualification criteria.	N-Q	Against tender specification / qualification criteria	MAKE : DIAG CHINA again-spec & Q/criteria	N-Q	Against tender specification / qualification criteria
28	Semi Automated Bio Chemistry Analyzer	N-Q	1. Make: Awareness Technology USA As per tender specification / qualification criteria	N-Q	N-Q	1. Make: ISE Italy against tender specification	Make Ondag Belgium against tender specification / qualification criteria	1. Make: Analyzer UK against tender specification / qualification criteria	Make Vcal Scientific, Netherlands as per tender specification / qualification criteria	Against tender specification / qualification criteria	MAKE : DIAG CHINA again-spec & Q/criteria	Beiba-France Against tender specification	Against tender specification / qualification criteria
29	Electrolyte Analyzer	N-Q	1. Make: SFR France Against tender specification	N-Q	N-Q	1. Make: Electrical Medical USA against per tender specification	Make: K-Lite Electrolyte Aus. against tender specification / qualification criteria	Make Diamond USA against tender specification / qualification criteria	Make: Opt Medical USA as per tender specification / qualification criteria	Against tender specification / qualification criteria	Diamond USA Against tender specification	Mehou - China Against tender specification	Against tender specification / qualification criteria

  
Dr. N. Baloch  
(DDO) Mirpurkhas

  
Dr. Ahmed Khan  
(DDO) Procurement Logistics

  
Mr. Aziz Ahmad  
(DDO) F&P Mirpurkhas

  
Dr. Hussain Khan Mari  
Focal Person (WFP)  
District Mirpurkhas

  
Dr. Waheed Bin Azhar  
(Chief Operating Officer)  
Fatimid Foundation Karachi  
Technical Advisor to Evaluation Committee


  
Dr. Sharique Hussain Daberi  
Executive District Officer Health  
District Mirpurkhas


4/6/19


Comparative statement on financial bids regarding tender Equipment Instrument and Machinery for the up gradation Hospital Mirpurkhas and Laboratory Dialysis Center At Civil Hospital Mirpurkhas.


Tender #	Name of Items	M/S Muslim Medical Services	M/S Sindh Medical Store	M/S Cosmos Enterprises	M/S Global Marketing
1	Cryofuge Automated Refrigerated centrifuge floor standing.				Make: Sorval USA RS. 289500000
2	Platelet Agitator		1. Make Helmer USA RS. 279000.00		Make: Helmer USA / BMS RS. 275000.00
3	Platelet Agitator Incubator		1. Make Helmer USA RS. 873000.00		Make: Helmer USA / BMS RS. 450000.00
4	Blood Bag Expresser				
5	Tube Sealer				1. Make Centron korea Rs. 235000.00
6	Weighing Scale				
7	Table Top Centrifuge		1. Make Sigma Germany: RS. 450000.00		
8	Table Top Centrifuge				
9	Water Bath- Large	Haier-China: RS. 35000.00			
10	Water Bath- Small	Haier-China: RS. 25000.00			
11	Microscope				Make: Leica Germany RS. 145000.00
12	Blood Bag Shaker		1. Hemotech / Dalcone Italy: RS. 468000.00		Make: Centron korea RS. 235000.00
13	Plasma Freezer with temp chart recorder			1. Make : Foucchett Italy. Rs. 3082000.00	
14	Plasma Freezer		Make Italy: RS. 554913.00	1. Make : Foucchett Italy: Rs. 4156000.00	
15	Blood Bank Refrigerator		1. Make Fridm Italy/ BMS Canada: RS. 786609.00		
16	Blood Bank Refrigerator		Make: Helmer USA / BMS Canada. RS. 585000.00	1. Make : Vision scientific south korea. RS. 715000.00	

17	Pharmacy Refrigerator			
18	Blood Donation Chair/Venation couches			
19	Computer			
20	Printer			
21	Printer			
22	Generator			
23	Split Air conditioner			
24	Gel Cord System		1. Diamed / Swiss' RS 783000.00	
25	Weighing Machine			
26	Stabilizer			
27	Haematology Analyzer		1. Make Melet France: RS.495000.00	
28	Semi Automated Bio Chemistry Analyzer		1. Make Awareness technology USA:RS 333000.00	Make Vital Scientific, Neatherland; RS,345000.00
29	Electrolyte Analyzer			Make:Opti Medical USA; RS,290000..00

  
 DR M Hanif Baloch  
 (DOH) Admin/Account

  
 Dr Ahmed Khan  
 (DDO Procurement/ Logistics)

  
 Mr Aziz Ahmed Qaudpoto  
 (DDO F&P) Mirpurkhas

  
 Dr Hussain Khan Marri  
 Focal Person (WFP)  
 Mirpurkhas

  
 Dr. Waleed Bin Azhar  
 C.O.O Fatimid Foundation

  
 Dr Shafquat Hussain Daheri  
 Executive District Officer

Comparative statement tender Equipment / Instruments and Machinery for the up gradation of Blood Bank DHQ Hospital Mirpurkhas and Laboratory Dialysis Center At Civil Hospital Mirpurkhas.

Tender #	Name of Items	M/S Muslim Medical Services	M/S Sindh Medical Store	M/S Cosmos Enterprises	M/S Global Marketing
1	Cryofuge Automated Refrigerated centrifuge floor	-	-	-	Make: Sorval USA: Rs. 2895000.00
2	Platelet Agitator	-	Make: Helmer USA / BMS Canada: Rs. 279000.00	-	Make: Helmer USA / BMS Canada: Rs. 275000.00
3	Platelet Agitator Incubator	-	Make: Helmer USA RS. 873000.00	-	Make: Helmer USA RS. 450000.00
4	Blood Bag Expresser	Not Quoted	Not Quoted	Not Quoted	Not Quoted
5	Tube Sealer	-	-	-	1.Make Centroan korea: RS. 235000.00
6	Weighing Scale	Not Quoted	Not Quoted	Not Quoted	Not Quoted
7	Table Top Centrifuge	-	1.Make Sigma Germany :RS.450000.00	-	-
8	Table Top Centrifuge	Not Quoted	Not Quoted	Not Quoted	Not Quoted
9	Water Bath- Large	Haier-China: RS.35000.00	-	-	-
10	Water Bath- Small	Haier-China: RS.25000.00	-	-	-
11	Microscope	-	-	-	Make: Leica Germany: Rs. 145000.00

Tender #	Name of Items	M/S Muslim Medical Services	M/S Sindh Medical Store	M/S Cosmos Enterprises	M/S Global Marketing
12	Blood Bag Shaker	-	Make: Delcon Italy. Rs. 468000.00	-	Make: Centron Korea. Rs. 235000.00
13	Plasma Freezer with temp chart recorder	-	-	1. Make Fouchelet Italy Rs. 3082000.00	-
14	Plasma Freezer	-	Make Italy : Rs. 54913.00	1. Make Fouchelet Italy Rs. 1156000.00	-
15	Blood Bank Refrigerator	-	1. Make Frimed Italy/ Rs. 786609.00	-	-
16	Blood Bank Refrigerator	-	Make: Helmer USA Canada: Rs. 585000.00	1. Make Vision scendf south korea Rs. 715000.00	-
17	Pharmacy Refrigerator	Not Quoted	Not Quoted	Not Quoted	Not Quoted
18	Blood Donation	Not Quoted	Not Quoted	Not Quoted	Not Quoted
19	Computer	Not Quoted	Not Quoted	Not Quoted	Not Quoted
20	Printer	Not Quoted	Not Quoted	Not Quoted	Not Quoted
21	Printer	Not Quoted	Not Quoted	Not Quoted	Not Quoted
22	Generator	Not Quoted	Not Quoted	Not Quoted	Not Quoted
23	Split Air conditioner	Not Quoted	Not Quoted	Not Quoted	Not Quoted
24	Gel Cord System	-	1. Diamed / Swiss: Rs. 783000.00	-	-
25	Weighing Machine	Not Quoted	Not Quoted	Not Quoted	Not Quoted
26	Stabilizer	Not Quoted	Not Quoted	Not Quoted	Not Quoted
27	Haematology Analyzer	-	1. Make Melet France: Rs. 495000.00	-	-
28	Semi Automated Bio Chemistry Analyzer	-	1. Make Awareness technology USA: Rs. 333000.00	-	Make Vital Scientific, Neatherland: Rs. 345000.00
29	Electrolyte Analyzer	-	-	-	Make Opti Medical USA. Rs. 890,000

Dr. Shaquiat Hussain Daheri  
Executive District Officer Health  
District Mirpurkhas.

Mr. Syed Zakir Ali Shah  
Executive District Officer Finance & Planning  
District Mirpurkhas.

Administrator / District Coordination Officer  
District Mirpurkhas

Mr. Ghulam Mustafa Akbar  
Executive District Officer Works & Services  
District Mirpurkhas




**Detail of deviation from tender specification / qualification criteria in respect of**

**M/S Cosmos Enterprises ( Offer #.02)**

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ml) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers - perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .	CE/ FDA Certificate missing	Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 °C Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	CE/ FDA Certificate missing	Authority letter missing
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	CE/ FDA Certificate missing	Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	CE/ FDA Certificate missing	Authority letter missing
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE/ FDA Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing

12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature - 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
18	Blood Donation Chair /Venation couches	Standard Size Locally Manufactured	CE/ FDA Certificate missing	Authority letter missing
19	Computer	P4 FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	CE/ FDA Certificate missing	Authority letter missing
21	Printer	Laser FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
22	Generator	30 KV Local/ ISO Approved	CE/ FDA Certificate missing	Authority letter missing

23	Split Air conditioner	1 ½ ton ISO Certified	CE/ FDA Certificate missing	Authority letter missing
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
25	Weighing Machine	Mechanical / Locally Manufactured.	CE/ FDA Certificate missing	Authority letter missing
26	Stabilizer	UPS 5 kv Local/ ISO Approved	CE/ FDA Certificate missing	Authority letter missing
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing


  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS

Detail of deviation from tender specification / qualification criteria in respect of  
M/S Oasis Medical systems

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500MI) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	4x500ml single bucket	Authority letter missing
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .		Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 °Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved		Authority letter missing
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.		Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	Hand held	Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	CE/ FDA Certificate missing	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE/ FDA Certificate missing	
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE/ FDA Certificate missing	
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .		Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .		Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE/ FDA Certificate missing	

12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	CE/ FDA Certificate missing	
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	temp:chart recorder missing	
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature – 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	temp:chart recorder missing	
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature – 2 0C to-6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE Certificate missing / Lesser capacity	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE certificate missing	
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	Interface for PC Missing	
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	Manufacturer not mentioned	Authority letter missing
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	Authority letter missing
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	Authority letter missing
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	Authority letter missing
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	Authority letter missing

26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	Authority letter missing
27	Haematology Analyzer	18 Parameters 3 part Diff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing

  
**EXECUTIVE DISTRICT OFFICER HEALTH**  
**DISTRICT GOVERNMENT MIRPURKHAS**

Detail of deviation from tender specification / qualification criteria in respect of  
M/S Lab Link Enterprises.

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ml) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .	CE/ FDA Certificate missing.	Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 ° Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	Item quoted <i>not</i>	
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.		Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Item quoted <i>not</i>	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE/ FDA Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	RMP 4000	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing

12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	Item quoted <i>not</i>	
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature - 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Capacity 659 ltrs tem: chart recorder missing	Authority letter missing
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	- Interface for PC missing - CE/FDA Certificate missing	Authority letter missing
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	Item quoted <i>not</i>	
19	Computer	P4 FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved		Authority letter missing
21	Printer	Laser FDA/ CE Approved .		Authority letter missing
22	Generator	30 KV Local/ ISO Approved		Authority letter missing
23	Split Air conditioner	1 ½ ton ISO Certified		Authority letter missing



**Detail of deviation from tender specification / qualification criteria in respect of  
M/S Hospicare**

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500MI) with Accessories Temperature range -9- +40 °C.Control Microprocessor, FDA/ CE Approved .	Item not quoted	
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .		Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 °Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	CE/ FDA Certificate missing	Authority letter missing
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.		Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Item not quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.		Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .		Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	Item not quoted	
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	Item not quoted	
11	Microscope	Binocular FDA/ CE Approved .	Item not quoted	
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing

13	Plasma Freezer with temp:chart recorder	Capacity 450-500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	Item not quoted	
14	Plasma Freezer	Capacity 250- 300 liter upright ultra low temperature - 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	Item not quoted	
15	Blood Bank Refrigerator	Capacity 250- 300 liter low temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .		Authority letter missing
16	Blood Bank Refrigerator	Capacity 100- 150 liter temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .		Authority letter missing
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2- 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	CE/FDA Certificate missing	Authority letter missing
18	Blood Donation Chair /Venation couches	Standard Size Locally Manufactured	Item not quoted	
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH,-RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system: Minimum reagent volume 250ul FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved.		Authority letter missing

120  
EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT GOVERNMENT MIRPURKHAS

**Detail of deviation from tender specification / qualification criteria in respect of  
M/S SK Trading**

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500Ml) with Accessories Temperature range -9~ +40 oC Control Microprocessor, FDA/ CE Approved .	CE/ FDA Certificate missing	
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .		Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 oC Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved		Authority letter missing
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	ISO / CE Certificate missing	Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .		Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Item not quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE Certificate missing	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 oC to 99 oC FDA/ CE Approved .	CE Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE Certificate missing	
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .		Authority letter missing

13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE Certificate missing	Authority letter missing
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature – 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .		Authority letter missing
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE Certificate missing	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for-PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE Certificate missing	
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	CE/FDA Certificate missing	Authority letter missing
18	Blood Donation Chair /Venation couches	Standard Size Locally Manufactured	Item not quoted	
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	Item not quoted	
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system: Minimum reagent volume 250ul FDA/ CE Approved .	CE/ FDA Certificate missing	
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved.	CE/ FDA Certificate missing	

  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS

Detail of deviation from tender specification / qualification criteria in respect of  
M/S Shah Enterprises.

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500MI) with Accessories Temperature range -9~ +40 0C Control Microprocessor, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers -perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .	CE/ FDA Certificate missing	Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 0 Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	CE/ FDA Certificate missing	Authority letter missing
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	CE/ FDA Certificate missing	Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	CE/ FDA Certificate missing	Authority letter missing
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE/ FDA Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 0 C to 99 0C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature - 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing


15	Blood Bank Refrigerator	Capacity 250- 300 liter low temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
16	Blood Bank Refrigerator	Capacity 100- 150 liter temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2- 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	CE/ FDA Certificate missing	Authority letter missing
19	Computer	P4 FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	CE/ FDA Certificate missing	Authority letter missing
21	Printer	Laser FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
22	Generator	30 KV Local/ ISO Approved	CE/ FDA Certificate missing	Authority letter missing
23	Split Air conditioner	1 ½ ton ISO Certified	CE/ FDA Certificate missing	Authority letter missing
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
25	Weighing Machine	Mechanical / Locally Manufactured.	CE/ FDA Certificate missing	Authority letter missing
26	Stabilizer	UPS 5 kv Local/ ISO Approved	CE/ FDA Certificate missing	Authority letter missing
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system: Minimum reagent volume 250ul FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved.	CE/ FDA Certificate missing	Authority letter missing

  
**EXECUTIVE DISTRICT OFFICER HEALTH**  
**DISTRICT GOVERNMENT MIRPURKHAS**

Detail of deviation from tender specification / qualification criteria in respect of  
M/S Muslim Medical Services.

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ml) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	Single bucket 4x500 ml	
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved	CE/FDA Certificate missing	
3	Platelet Agitator Incubator	Digital Temp 20~24 °C Capacity to hold one unit of platelet agitator Locally Manufactured, / ISO Certified or CE/FDA Approved	CE/FDA Certificate missing	
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	Imported	
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .		Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Not Quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.		Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	3900 RPM	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	No Deviation	No Deviation
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	No Deviation	No Deviation
11	Microscope	Binocular FDA/ CE Approved .	Not quoted	
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	Not Quoted	
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE	temp:chart recorder missing	

		Approved .		
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature – 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	temp:chart recorder missing	
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	temp:chart recorder missing	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control temp chart recorder FDA/ CE Approved .	temp:chart recorder missing	
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	Interface for PC missing	
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	Not Quoted	
19	Computer	P4 FDA/ CE Approved .	Not Quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Not Quoted	
21	Printer	Laser FDA/ CE Approved .	Not Quoted	
22	Generator	30 KV Local/ ISO Approved .	Not Quoted	
23	Split Air conditioner	1 ½ ton -ISO Certified	Not Quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Not Quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Not Quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Not Quoted	
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	Not Quoted	
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	Not Quoted	
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	Not Quoted	


  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS



Detail of deviation from tender specification / qualification criteria in respect of  
M/S Sindh Medical Store

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ml) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	Single bucket 4x500 ml	
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved	No Deviation	No Deviation
3	Platelet Agitator Incubator	Digital Temp 20~24 ° Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	No Deviation	No Deviation
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	Imported	
5	Tube Sealer	Table top heat sealer FDA/ CE/ Approved .		Authority letter missing
6	Weighing Scale	Weighing-capacity 10kg Locally Manufactured	Not Quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	No Deviation	No Deviation
8	Table Top Centrifuge.	5,000 RPM FDA/ CE Approved .		Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .		Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .		Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	Not quoted	
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	No Deviation	No Deviation
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	temp:chart recorder missing	

14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature – 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	No Deviation	No Deviation
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	No Deviation	No Deviation
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	No Deviation	No Deviation
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	Interface for PC missing	
18	Blood Donation Chair /Venation couches	Standard Size Locally Manufactured	Not Quoted	
19	Computer	P4 FDA/ CE Approved .	Not Quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Not Quoted	
21	Printer	Laser FDA/ CE Approved .	Not Quoted	
22	Generator	30 KV Local/ ISO Approved	Not Quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Not Quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	No Deviation	No Deviation
25	Weighing Machine	Mechanical / Locally Manufactured.	Not Quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Not Quoted	
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	No Deviation	No Deviation
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	No Deviation	No Deviation
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	CE/ FDA Certificate missing	

  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS

Detail of deviation from tender specification / qualification criteria in respect of  
M/S S M Trading company

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500MI) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	Item not quoted	
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers - perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved	Item not quoted	
3	Platelet Agitator Incubator	Digital Temp 20~24 °C Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	Item not quoted	
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	Item not quoted	
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	Item not quoted	
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Item not quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE Certificate missing	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	CE Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE Certificate missing	Authority letter missing
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	Item not quoted	Authority letter missing
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters uprightultra low temperature - 70 °C alarm temperature chart recorder shelves with	Temp:chart recorder missing	Authority letter missing

		air insulated inner doors FDA/ CE Approved .		
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature – 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	Temp:chart recorder missing	Authority letter missing
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Item not quoted	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Item not quoted	
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	Item not quoted	
18	Blood Donation Chair /Venation couches	Standard Size Locally Manufactured	Item not quoted	
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	Item not quoted	
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	Item not quoted	
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	Item not quoted	

  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS

Detail of deviation from tender specification / qualification criteria in respect of  
M/S Platinum

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ml) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	Single bucket 4x500 ml CE certificate missing	
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved	Item not quoted	
3	Platelet Agitator Incubator	Digital Temp 20~24 ° Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	Item not quoted	
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	Item not quoted	
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	Item not quoted	
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Item not quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE Certificate missing	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	CE Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE Certificate missing	Authority letter missing
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	Item not quoted	
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters uprightultra low temperature - 70 °C alarm temperature chart recorder shelves with	Temp:chart recorder missing	

		air insulated inner doors FDA/ CE Approved .		
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature – 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	Temp:chart recorder missing	
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Temp:chart recorder missing	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature – 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Temp:chart recorder missing	
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	Interface PC missing	
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	Item not quoted	
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	
27	Haematology Analyzer	18 Parameters 3 part Diff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	Item not quoted	
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	Item not quoted	
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	Item not quoted	

  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS

Detail of deviation from tender specification / qualification criteria in respect of

M/S Cosmos Enterprises ( Offer #.01)

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ML) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .	CE/ FDA Certificate missing	Authority letter missing
3	Platelet Agitator Incubator	Digital Temp 20~24 °Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	CE/ FDA Certificate missing	Authority letter missing
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	CE/ FDA Certificate missing	Authority letter missing
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	CE/ FDA Certificate missing	Authority letter missing
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	CE/ FDA Certificate missing	Authority letter missing
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
11	Microscope	Binocular FDA/ CE Approved .	CE/ FDA Certificate missing	

12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	Item not quoted	
13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	No Deviation	No Deviation
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature - 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	No Deviation	No Deviation
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	CE/ FDA Certificate missing	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	No Deviation	No Deviation
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	CE/ FDA Certificate missing	Authority letter missing
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	Item not quoted	
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	



27	<b>Haematology Analyzer</b>	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	CE/ FDA Certificate missing	
28	<b>Semi Automated Bio Chemistry Analyzer</b>	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum reagent volume 250ul FDA/ CE Approved .	CE/ FDA Certificate missing	
29	<b>Electrolyte Analyzer</b>	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	CE/ FDA Certificate missing	



**EXECUTIVE DISTRICT OFFICER HEALTH  
DISTRICT GOVERNMENT MIRPURKHAS**

Detail of deviation from tender specification / qualification criteria in respect of  
M/S Global Marketing.

S.#	NAME OF PRODUCT	TENDER SPECIFICATION.	SPEC: DEVIATION	CRITERIA DEVIATION
1	Cryofuge Automated Refrigerated centrifuge floor standing.	Wind shield bucket Rotar 4 Double blood Bag Bucket (8x500ml) with Accessories Temperature range -9~ +40 °C Control Microprocessor, FDA/ CE Approved .	No Deviation	No Deviation
2	Platelet Agitator	24 Bags Capacity Gental Agitator by A Stationery base that moves drawer Assembly Side to side Drawers – perforated non slip Locally Manufactured / ISO Certified or CE/FDA Approved .	No Deviation	No Deviation
3	Platelet Agitator Incubator	Digital Temp 20~24 ° Capacity to hold one unit of platelet agitator Locally Manufactured. / ISO Certified or CE/FDA Approved	No Deviation	No Deviation
4	Blood Bag Expresser	Manual Plasma Extractor Locally Manufactured.	Item not quoted	
5	Tube Sealer	Table top heat sealer FDA/ CE Approved .	No Deviation	No Deviation
6	Weighing Scale	Weighing capacity 10kg Locally Manufactured	Item not quoted	
7	Table Top Centrifuge	10,000 RPM FDA/ CE Approved.	Item not quoted	
8	Table Top Centrifuge	5,000 RPM FDA/ CE Approved .	Item not quoted	
9	Water Bath- Large	18 L-22 L, Digital Over Temperature Protection, Alarm, Timer temp range +5 °C to 99 °C FDA/ CE Approved .	Item not quoted	
10	Water Bath- Small	5L -7 L Digital Over Temperature Protection Alarm Timer temp range + 5 0C to 99 0C FDA/ CE Approved .	Item not quoted	
11	Microscope	Binocular FDA/ CE Approved .	No Deviation	No Deviation
12	Blood Bag Shaker	Rock to mix collected Blood with anticoagulant in the blood bag, LCD panel displays collect total volume preset volume to collect total weight of collected blood and blood bags elapsed time of donation and alarm symbols, audible and visual alarm signals at the end of collection, at abnormal blood flow and at low battery condition automatically takes weight of blood bags to measure net volume of collected blood fully portable with a rechargeable and replaceable battery pack. FDA/ CE Approved .	No Deviation	No Deviation

13	Plasma Freezer with temp:chart recorder	Capacity 450~500 Liters upright ultra low temperature - 70 °C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	Item not quoted	
14	Plasma Freezer	Capacity 250~ 300 liter upright ultra low temperature - 20 0C- 40 0C alarm temperature chart recorder shelves with air insulated inner doors FDA/ CE Approved .	Item not quoted	
15	Blood Bank Refrigerator	Capacity 250~ 300 liter low temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Item not quoted	
16	Blood Bank Refrigerator	Capacity 100~ 150 liter temperature - 2 0C to 6 0C shelves Glass door audio / visual interface for PC, temp alarm door lock with key microprocessor control, temp chart recorder FDA/ CE Approved .	Item not quoted	
17	Pharmacy Refrigerator	Fully automatic defrosting cycle temperature 2~ 0C door lock with key microprocessor contract digital alarm, alarm audio/ visual interface for PC FDA/ CE Approved .	Item not quoted	
18	Blood Donation Chair / Venation couches	Standard Size Locally Manufactured	Item not quoted	
19	Computer	P4 FDA/ CE Approved .	Item not quoted	
20	Printer	Lx 300 Dotmatrix FDA/ CE Approved	Item not quoted	
21	Printer	Laser FDA/ CE Approved .	Item not quoted	
22	Generator	30 KV Local/ ISO Approved	Item not quoted	
23	Split Air conditioner	1 ½ ton ISO Certified	Item not quoted	
24	Gel Cord System	For blood grouping and cross matching FDA/ CE Approved .	Item not quoted	
25	Weighing Machine	Mechanical / Locally Manufactured.	Item not quoted	
26	Stabilizer	UPS 5 kv Local/ ISO Approved	Item not quoted	
27	Haematology Analyzer	18 Parameters 3 part Disff RBC, HGB, Ly #, MO#, GR#, LY%, MO%, HCT, MCY, MCH, RDW, PLT, MPV, PCT, PWD, FDA/ CE Approved .	CE/ FDA Certificate missing	
28	Semi Automated Bio Chemistry Analyzer	Programme test 120 Real time kinetic graphics display graphical LCD with moveable screen keyboard, cuvet system :Minimum regent volume 250ul FDA/ CE Approved .	No deviation	No deviation
29	Electrolyte Analyzer	For Measurement of potassium, sodium, chloride & calcium FDA/ CE Approved .	No deviation	No deviation

  
 EXECUTIVE DISTRICT OFFICER HEALTH  
 DISTRICT GOVERNMENT MIRPURKHAS

**Bill of Quantities / schedule of requirement (tender Equipment Instrument and Machinery for the up gradation  
DHQ Hospital Mirpurkhas and Laboratory Dialysis Center At Civil Hospital Mirpurkhas).**

Tender #	Name of Items	Name of Contrator	Qty:	Approved Rate / unit	Total Amount
1	Cryofuge Automated Refrigerated centrifuge floor standing.	M/S Global Marketing	1	Rs, 2895000.00 Make: Sorval USA	RS,2895000.00
2	Platelet Agitator	M/S Global Marketing	2	Rs, 275000.00 Make:Helmer USA/ BMS Canada	RS.550000.00
3	Platelet Agitator Incubator	M/S Global Marketing	1	Rs.450000.00 Make:Helmer USA	Rs.450000.00
5	Tube Sealer	M/S Global Marketing	2	RS.235000.00 Make: Centron Korea	Rs.470000.00
7	Table Top Centrifuge	M/S Sindh Medical Store	1	RS. 450000.00 Make: Sigma Germany	Rs.450000.00
9	Water Bath- Large	M/S Muslim Medical Services	3	RS.35000.00 Make Haier China	Rs.105000.00
10	Water Bath- Small	M/S Muslim Medical Services	3	RS.25000.00 Make Haier China	Rs.75000.00
11	Microscope	M/S Global Marketing	3	RS. 145000.00 Make Leica Germany	Rs.435000.00
12	Blood Bag Shaker	M/S Global Marketing	3	Rs.235000.00 Make Sentron Korea	Rs.705000.00
14	Plasma Freezer	M/S Sindh Medical Store	2	RS.554913.00 Make Italy	Rs.1109806.00
15	Blood Bank Refrigerator	M/S Sindh Medical Store	1	RS.786609.00 Make Frimed Italy	Rs.786609.00
16	Blood Bank Refrigerator	M/S Sindh Medical Store	2	Rs.585000.00 Make : Helmer USA	Rs.1170000.00

Tender #	Name of Items	Name of Contrator	Qty:	Approved Rate / unit	Total Amount
24	Gel Cord System	M/S Sindh Medical Store	1	Rs.783000.00 Make Diamed Swiss	Rs.783000.00
27	Hacmatology Analyzer	M/S Sindh Medical Store	1	Rs.495000.00 Make: Melet France	Rs.495000.00
28	Semi Automated Bio Chemistry Analyzer	M/S Sindh	1	RS.333000	Rs.333000.00
29	Electrolyte Analyzer	M/S Global Marketing	1	Rs.290000.00 Make Opti Medical USA	Rs.290000.00

  
Dr Shafquat Hussain Daheri  
Executive District Officer Health  
District Mirpurkhas

## FORM OF CONTRACT AGREEMENT

THIS CONTRACT AGREEMENT (hereinafter called the "Agreement") made on the 30th day of June (month) 2011 between E. D. O Health District Muspurkhas (hereafter called the "Procuring Agency") of the one part and M/S Sudd Medical Store (hereafter called the "Contractor") of the other part.


WHEREAS the Procuring Agency is desirous that certain works, viz INSTRUMENTS / EQUIPMENT should be executed by the Contractor and has accepted a bid by the Contractor for the execution and completion of such works and the remedying of any defects therein.

NOW this Agreement witnesseth as follows:


1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract hereinafter referred to.
2. The following documents after incorporating addenda, if any, except those parts relating to Instructions to bidders shall be deemed to form and be read and construed as part of this Agreement, viz:
  - (a) The Contract Agreement;
  - (b) The Letter of Acceptance;
  - (c) The completed Form of Bid;
  - (d) Special Stipulations (Appendix-A to Bid);
  - (e) The Special Conditions of Contract - Part II;
  - (f) The General Conditions - Part I;
  - (g) The priced Bill of Quantities (Appendix-D to Bid);
  - (h) The completed Appendices to Bid (B, C, E to I);
  - (i) The Drawings;
  - (j) The Specifications;
  - (k) \_\_\_\_\_ (any other)
3. In consideration of the payments to be made by the procuring agency to the Contractor as hereinafter mentioned, the Contractor hereby covenants with the procuring agency to execute and complete the works and remedy defects therein in conformity and in all respects with the provisions of the contract.
4. Procuring agency hereby covenants to pay the contractor, in consideration of the execution and completion of the works as per provisions of the contract, the contract Price or such other sum as may become payable under the provisions of the contract at the times and in the manner prescribed by the contract.

IN WITNESS WHEREOF the parties hereto have caused this Agreement to be executed on the day, month and year first before written in accordance with their respective laws.

Signature of the Contactor

(Seal) 

Signature of Procuring Agency

(Seal)   
**EXECUTIVE DISTRICT OFFICER ROADPS  
DISTRICT MIRPURKHAS.**


Signed, Sealed and Delivered in the presence of:

Witness:




(Name, Title and Address)

Witness:



**MUKHTIAR KHAN**  
(Name, Title and Address)

**STORE KEEPER  
INSTRUMENT / EQUIPMENT / MACHINERY**



NEYAZ AHMED STAMP VENDOR  
Lic. No.83, Shop No. 3, Karim Plaza,  
Block-14, Gulshan-e-Iqbal, Karachi

17 MAY 2011

S.No. 64070 Date: 17/5/11  
Issued with Address: 17/5/11  
Through HAIDER ALI KHAN  
Purpose Advocate W/S Wedger H164  
Stamp Vendor's Signature 17/5/11

## Undertaking


On behalf of M/S Sind Medical Stores under signatory state on solemn affirmation that on part of our firm any contravention with prevalent SPP rules and terms & condition of the tender will be tantamount to breach of this contract and make this firm liable to penal action as per rule and procedure.

Assurance is given to supply required items of quality.

Assurance is given to supply required items on approval quoted rates till the end of the current financial year 2010-11.

Assurance is given to supply required items with in the stipulated delivery period on the same quoted rates and provide additional quantity of product is declared substandard, documents submitted by our firm is proved sake/false at any stage of procurement our firm shall be liable to penal action prevalent as per SPP rules.

*[Handwritten signature]*





حکومت پاکستان  
قومی شناختی کارڈ  
42101-207079-0




علی ارشد حکیم  
04/04/1978

نام: محمد  
جنس: مرد  
والد/پرورش کار نامہ: محمد اسلم  
شناختی علامت: بائیں آنکھ کے قریب جوت کانٹان  
تاریخ پیدائش: 04/04/1978

دستخط: محمد اسلم  
دستخط: محمد اسلم

حکومت پاکستان  
قومی شناختی کارڈ  
42000-0567899-3






نام: برحان  
جنس: مرد  
والد/پرورش کار نامہ: محمد اسلم  
شناختی علامت: بائیں آنکھ کے قریب جوت کانٹان  
تاریخ پیدائش: 03/01/1973

دستخط: برحان  
دستخط: برحان

گٹہ کارڈ نمبر برقی ایڈریس میں ڈال دیں۔

حکومت پاکستان  
قومی شناختی کارڈ  
42201-2583256-4

نام: حسین مصطفیٰ  
جنس: مرد  
والد/پرورش کار نامہ: حسین مصطفیٰ  
شناختی علامت: بائیں آنکھ کے قریب جوت کانٹان  
تاریخ پیدائش: 14/07/1990

دستخط: حسین مصطفیٰ  
دستخط: حسین مصطفیٰ

گٹہ کارڈ نمبر برقی ایڈریس میں ڈال دیں۔

## FORM OF CONTRACT AGREEMENT

THIS CONTRACT AGREEMENT (hereinafter called the "Agreement") made on the 30th day of June (month) 2011 between E-D. Health District, Muzhaplas. (hereafter called the "Procuring Agency") of the one part and M/S Muslim Medical Services (hereafter called the "Contractor") of the other part.

WHEREAS the Procuring Agency is desirous that certain works, viz INSTRUMENT/EQUIPMENTS should be executed by the Contractor and has accepted a bid by the Contractor for the execution and completion of such works and the remedying of any defects therein.

NOW this Agreement witnesseth as follows:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract hereinafter referred to.
2. The following documents after incorporating addenda, if any, except those parts relating to Instructions to bidders shall be deemed to form and be read and construed as part of this Agreement, viz:
  - (a) The Contract Agreement;
  - (b) The Letter of Acceptance;
  - (c) The completed Form of Bid;
  - (d) Special Stipulations (Appendix-A to Bid);
  - (e) The Special Conditions of Contract – Part II;
  - (f) The General Conditions – Part I;
  - (g) The priced Bill of Quantities (Appendix-D to Bid);
  - (h) The completed Appendices to Bid (B, C, E to L);
  - (i) The Drawings;
  - (j) The Specifications.
  - (k) \_\_\_\_\_ (any other)
3. In consideration of the payments to be made by the procuring agency to the Contractor as hereinafter mentioned, the Contractor hereby covenants with the procuring agency to execute and complete the works and remedy defects therein in conformity and in all respects with the provisions of the contract.
4. Procuring agency hereby covenants to pay the contractor, in consideration of the execution and completion of the works as per provisions of the contract, the contract Price or such other sum as may become payable under the provisions of the contract at the times and in the manner prescribed by the contract.

IN WITNESS WHEREOF the parties hereto have caused this Agreement to be executed on the day, month and year first before written in accordance with their respective laws.

Signature of the Contactor  
*[Signature]*  
Sales Manager  
**MUSLIM MEDICAL SERVICES**  
4&5 Mezzanine Floor Nobel Heights,  
Opp. Askari Park Main University Road,  
Karachi 70800 Pakistan

Signature of Procuring Agency  
*[Signature]*  
(Seal) **CHIEF DISTRICT OFFICER HEALTH**  
**DISTRICT HIRAPURHAR.**

Signed, sealed and delivered in the presence of:

Witness: *[Signature]*  
**Zubair Hussain**  
Hydrabad - Karachi  
(Name, Title and Address)  
Sales Office  
**MUSLIM MEDICAL SERVICES**  
4&5 Mezzanine Floor Nobel Heights,  
Opp. Askari Park Main University Road,  
Karachi 70800 Pakistan  
Tel: 32-24-4921913, 4131909

Witness: *[Signature]*  
**MUKHTIAR AHMED**  
(Name, Title and Address)  
STORE INCHARGE  
INSTRUMENT/EQUIPMENT/MACHINERY



حکومت پاکستان

قومی شناختی کارڈ

42104-6740513-9

EDD (H) MRS

ONLY

نام:   
 جنس:   
 وید کا نام:   
 شناختی نمبر:   
 تاریخ پیدائش: 25/12/1973

پشاور

حکومت پاکستان

42101-460747 E-5

ONLY For

24/03/1983




(H) MKS

ممبران

کتابت

ممبران



AGIB MEAN STAMP VENDOR  
 Licence No. 04, Shop No. 15 Block-14,  
 Periodic Office, Ghansouhah, Karachi  
 "Date Centre" 13 JUN 2011  
 NO 10551  
 M/s Muslim Medical Services

### Contract Agreement

Contract No. \_\_\_\_\_ Dated 13<sup>th</sup> June 2011

This agreement made this \_\_\_\_\_ day of June 2011 between the Executive district officer health Mirpurkhaas "The Purchaser" acting through District Officer (Procumbent), Health Group of Offices, Mirpurkhaas and M/s Muslim Medical Services here in after referred to as "The Supplier"

Where as the purchaser is Desirous that certain goods should be provided by supplier and where as the supplier has agreed to provide and quoted the rates at discount as per Government Policy, which has accepted by the purchaser for the supply of the following medicines in sum of Rs. 180,000.00 (Rupees One Hundred Eighty Thousand Only. Here in after referred to as "The Contract Price".)

Item #	Description of Store	Unit Price	Qty.	Total Value in Figure
9	Water Bath Large	35,000.00each	3 Units	105,000.00
10	Water Bath Small	25,000.00each	3 Units	75,000.00
Grand Total				180,000.00

The delivery period will be commenced from the date of acceptance of contract by the bidder now this agreement witnessed as follows.

- In this agreement words and expressions shall have the same meanings as are respectively assigned to them in the terms and conditions of Tender enquiry refer to.
- Following documents shall be deemed to from and be read and construed as per this agreement.viz;
  - Contract Agreement
  - Price Schedule

- e. Terms & Conditions of the Contract.
  - f. Schedule of Requirements.
  - g. Bidder Offer including the relevant correspondence with all annexes made before signing the contract agreement.
3. In consideration of the payments to be made by the purchaser to the, we supplier as hereinafter mentioned, the supplier hereby covenants with the purchaser to provide the goods and services to remedy the defects in conformity in all respect with the provision of the Contact.
4. The Purchaser hereby covenant to pay the supplier in consideration of the provision of the goods and services and the remedying of the defects there in, the contract price payable under the provision of the contract at the time and in the manner prescribed in the contract.

In witness there of the parties here to have caused this agreement in accordance with their respective hands and seal, the day, the month and the year first above written.

Supplier  
Signature & Seal *Havoon Rafi*  
Name Havoon Rafi

Purchaser  
Signature & Seal *[Signature]*  
Name \_\_\_\_\_

Address & CNIC No. 352 01-3215029-7  
43-C officers colony  
Lahore

Address & CNIC No. \_\_\_\_\_  
\_\_\_\_\_

Witness (Supplier)  
Signature & Seal *[Signature]*  
Name Amir Mahmood

Witness (Purchaser)  
Signature & Seal *[Signature]*  
Name Muhammad Ahmed

Address & CNIC No. 425, Mezz  
Floor Noble Heights  
Karachi

Address & CNIC No. \_\_\_\_\_  
\_\_\_\_\_



## FORM OF CONTRACT AGREEMENT

THIS CONTRACT AGREEMENT (hereinafter called the "Agreement") made on the 30th day of June (month) 2011 between E.D.O Health Bristol - Mipenselas (hereafter called the "Procuring Agency") of the one part and M/S Global Marketing Services (hereafter called the "Contractor") of the other part.

WHEREAS the Procuring Agency is desirous that certain works, viz \_\_\_\_\_ should be executed by the Contractor and has accepted a bid by the Contractor for the execution and completion of such works and the remedying of any defects therein.

NOW this Agreement witnesseth as follows:

1. In this Agreement words and expressions shall have the same meanings as are respectively assigned to them in the Conditions of Contract hereinafter referred to.
2. The following documents after incorporating addenda, if any, except those parts relating to Instructions to bidders shall be deemed to form and be read and construed as part of this Agreement, viz:
  - (a) The Contract Agreement;
  - (b) The Letter of Acceptance;
  - (c) The completed Form of Bid;
  - (d) Special Stipulations (Appendix-A to Bid);
  - (e) The Special Conditions of Contract - Part II;
  - (f) The General Conditions - Part I;
  - (g) The priced Bill of Quantities (Appendix-D to Bid);
  - (h) The completed Appendices to Bid (B, C, E to L);
  - (i) The Drawings;
  - (j) The Specifications.
  - (k) \_\_\_\_\_ (any other)
3. In consideration of the payments to be made by the procuring agency to the Contractor as hereinafter mentioned, the Contractor hereby covenants with the procuring agency to execute and complete the works and remedy defects therein in conformity and in all respects with the provisions of the contract.
4. Procuring agency hereby covenants to pay the contractor, in consideration of the execution and completion of the works as per provisions of the contract, the contract Price or such other sum as may become payable under the provisions of the contract at the times and in the manner prescribed by the contract.



IN WITNESS WHEREOF the parties hereto have caused this Agreement to be executed on the day, month and year first before written in accordance with their respective laws.

Signature of the Contactor

(Seal)



Signature of Executing Agency

(Seal)

EXECUTIVE DISTRICT OFFICE HEALTH  
DISTRICT EMPORRAG.

Signed, Sealed and Delivered in the presence of:

Witness:

*[Handwritten signature]*


(Name, Title and Address)

Witness:

*[Handwritten signature]*

MUKHTIAR AHMED  
(Name, Title and Address)

STAFF KEEPER  
INSTRUMENT / EQUIPMENT / MACHINERY



## Undertaking

June 10, 2011

On behalf of M/S GLOBAL MARKETING SERVICES under signatory state on solemn affirmation that on part of our firm any contravention with prevalent SPP rules and terms & condition of the tender will be tantamount to breach of this contract and make this firm liable to penal action as per rule and procedure.

Assurance is given to supply required items of quality.

Assurance is given to supply required items on approval quoted rates till the end of the current financial year 2010-11.

Assurance is given to supply required items with in the stipulated delivery period on the same quoted rates and provide additional quantity of product is declared substandard, documents submitted by our firm is proved sake/false at any stage of procurement our firm shall be liable to penal action prevalent as per SPP rules.

Thanking you

Muhammad Ayub  
(Director Operations)  
Global Marketing Services







# Global Marketing Services

■ **Head Office**  
57/1, A-Block, Satellite Town, Rawalpindi  
Ph: 051-4844191-2 Fax: 051-4841861  
UAN: 051-111-145-236  
email: info@gms-world.com  
website: www.gms-world.com

Ref. #: CRT/2011/43

## *To whom it may concern*

*June 10, 2011*

We Messrs. Global Marketing Services located at 57/1, A-Block, St. Town, Rawalpindi authorized Mr Muhammad Ali S/O Mr. Inayat Ali having CNIC No.42301-7659984-5 to sign Executive District Officer (Health) Mirpur Khan Contract on our behalf.

Thanking you

Muhammad Ayub  
(Director Operations)  
Global Marketing Services



■ **LAHORE:** First Floor, Apartment No. 90/C, II,  
Canal Park, (Canal Bank) Gulberg-II,  
Lahore Ph: (042)-5734736

■ **KARACHI:** Rukhsana Building, Plor # E/2, Block 7/B,  
Off. # 10, 3rd Floor, Shaheed-e-Millat Road,  
KCHSI, Karachi Ph: 021-4547985-86

■ **NWFP:** Dabgray Garden,  
Peshawar.  
Ph: 091-2562164

■ **Balochistan:** Office # 15, Shaista Town,  
Satellite Town,  
Quetta Ph: 081-440359

email: info@gms-world.com

website: www.gms-world.com

9290097 FAX 9290098.  
NO. EDOH/ (Procurement)  
Store 7007/09 Dated, 30/06 2011

**OFFICE OF THE  
EXECUTIVE DISTRICT OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.**

The M/S Global Marketing,  
Rukhsana Building Plat-No. E/2 Block 7/8 Office # 10  
3<sup>rd</sup> Floor Shahrah Millat Road Karachi.

Subject:-

**PROCUREMENT OF INSTRUMENTS / EQUIPMENT/ & MACHINERY  
TENDER 2010.11 FOR UP GRADATION OF DHQ HOSPITAL  
MIRPURKHAS LABORATORY & DIALYSIS CENTERS.**

You are requested to please arrange / supply the following products as per terms and Conditions of the above tender and submit the bill / invoice with warranty in triplicate for necessary payment.

S.NO	DESCRIPTION OF STORE	MAKE	QUANTITY REQUIRED	RATE / AU	TOTAL VALUE
1	Cryofuge Automated Refrigerated centrifuge floor Standing	Make Sorval/ Thermo Fisher Germany	1	2895000.	2895000.00
2	Platelet Agitator	Make Helmer USA / BMS Canada	2	275000.	550000.00
3	Platelet Agitator Incubator	Make Helmer USA / BMS Canada	1	450000.	450000.00
4	Tube Sealer	Make Centran Korea	2	235000.	470000.00
5	Microscope	Make Leica germany	3	145000.	435000.00
6	Blood Bag Shaker	Make Centran Korea	3	235000.	705000.00
					5505000.00

TOTAL VALUE IN WORDS :- RS. Fifty five lac five thousand only.

*[Signature]*  
DR. SIYAT QUATRUSSAIN DAHERI  
Executive District Officer Health  
District Government Mirpurkhas.

Cc to:-

The District coordination Officer / District Administrator District Government Mirpurkhas.  
The Secretary to Government of Sindh Health Department Karachi.  
The Account branch EDOH Office Mirpurkhas, with instructions that release of payment may please be subjected to deduction of contract award fee at the rate of 0.025% at the time of payment as given under purchase manual 1991 and submission of security deposit 2.5% on the total amount of work order production of inspection note, warranty and Bank challan of Rs. 1500 in respect of CDL and Rs. 1000 in respect of PDTL as drug testing fee for every batch of supply

*[Signature]*  
DR. SIYAT QUATRUSSAIN DAHERI  
Executive District Officer Health  
District Government Mirpurkhas.

Received by  
M. Ali  
*[Signature]*  
30/06/2011

97

FAX 9290098.

NO. EDOH/ (Procurement)/

7010/ Dated 30/06 2011

12

OFFICE OF THE  
EXECUTIVE DISTRICT OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.

The M/S Global Marketing,  
Rukhsana Building Plat No. E/2 Block 7/8 Office # 10  
3<sup>rd</sup> Floor Shahrah Millat Road Karachi.

Subject:-

**PROCUREMENT OF INSTRUMENTS / EQUIPMENT/ & MACHINERY**  
**TENDER 2010.11 FOR UP GRADATION OF DHQ HOSPITAL**  
**MIRPURKHAS LABORATORY & DIALYSIS CENTERS.**

You are requested to please arrange / supply the following products as per terms and Conditions of the above tender and submit the bill / invoice with warranty in triplicate for necessary payment.

S.NO	DESCRIPTION OF STORE	MAKE	QUANTITY REQUIRED	RATE / AU	TOTAL VALUE
1	Electrolyte Analyzer	Make Opti Medical USA	1	290000.	290000.00
					290000.00

TOTAL VALUE IN WORDS :- RS. Two lac Ninety thousand only.

DR. SHAFQUAT HUSSAIN DAHERI  
Executive District Officer Health  
District Government Mirpurkhas.

Cc to:-

The District coordination Officer / District Administrator District Government Mirpurkhas.  
The Secretary to Government of Sindh Health Department Karachi.  
The Account branch EDOH Office Mirpurkhas, with instructions that release of payment may please be subjected to deduction of contract award fee at the rate of 0.025% at the time of payment as given under purchase manual 1991 and submission of security deposit 2.5% on the total amount of work order production of inspection note, warranty and Bank challan of Rs. 1500 in respect of CDL and Rs, 1000 in respect of PDTL as drug testing fee for every batch of supply

DR. SHAFQUAT HUSSAIN DAHERI  
Executive District Officer Health  
District Government Mirpurkhas.

Received by  
M. S. Ali  
30/06/2011

OFFICE OF THE  
EXECUTIVE DISTRICT OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.

Procurement/ Dated 30/6-2011

To,

The M/S Muslim Medical Services,  
Noble Heights 4&5 Main University Road Karachi.

Subject:-

PROCUREMENT OF INSTRUMENTS / EQUIPMENT/ & MACHINERY  
TENDER 2010.11 FOR UP GRADATION OF DHQ HOSPITAL  
MIRPURKHAS LABORATORY & DIALYSIS CENTERS.

You are requested to please arrange / supply the following products as per terms and Conditions of the above tender and submit the bill / invoice with warranty in triplicate for necessary payment.

S.NO	DESCRIPTION OF STORE	MAKE	QUANTITY REQUIRED	RATE / AU	TOTAL VALUE
1	Water Bath Large	Make Haier China	3	35000.	105000.00
2	Water Bath Small	Make Haier China	3	25000.	75000.00
					180000.00

TOTAL VALUE IN WORDS :- RS. One Lac Eighty thousand only.

DR. SHAFQAT HUSSAIN DAIERI  
Executive District Officer Health  
District Government Mirpurkhas.

Cc to:-

The District coordination Officer / District Administrator District Government Mirpurkhas.  
The Secretary to Government of Sindh Health Department Karachi.  
The Account branch EDOH Office Mirpurkhas, with instructions that release of payment may please be subjected to deduction of contract award fee at the rate of 0.025% at the time o payment as given under purchase manual 1991 and submission of security deposit 2.5% on the total amount of work order production of inspection note, warranty and Bank challan of Rs. 1500 in respect of CDL and Rs, 1000 in respect of PDDL as drug testing fee for every batch of supply

DR. SHAFQAT HUSSAIN DAIERI  
Executive District Officer Health  
District Government Mirpurkhas.

NO. EDOH/ (Procurement) /  
MPS/Store 7017/21 Dated 30/06 2011

OFFICE OF THE  
EXECUTIVE DISTRICT OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.

To,

The M/S Sindh Medical Store,  
13-B Block 06 PECHS Shahrah - e- Faisal Karachi.

Subject:-

PROCUREMENT OF INSTRUMENTS / EQUIPMENT/ & MACHINERY  
TENDER 2010.11 FOR UP GRADATION OF DHQ HOSPITAL  
MIRPURKHAS LABORATORY & DIALYSIS CENTERS.

You are requested to please arrange / supply the following products as per terms and conditions of the above tender and submit the bill / invoice with warranty in triplicate for necessary payment.

S.NO	DESCRIPTION OF STORE	MAKE	QUANTITY REQUIRED	RATE / AU	TOTAL VALUE
1	Haematology Analyzer	Make Melet France	1	495000.	495000.00
2	Sami Automated Bio Chemistry Ananlyzer	Make Awareness technology USA	1	333000.	333000.00
TOTAL VALUE IN WORDS :- RS. Eight Lac twenty eight thousand only.					828000.00

*DR SHAFQUATHUSSAIN DAHERI*  
Executive District Officer Health  
District Government Mirpurkhas.

Cc to:-

The District coordination Officer / District Administrator District Government Mirpurkhas.  
The Secretary to Government of Sindh Health Department Karachi.  
The Account branch EDOH Office Mirpurkhas, with instructions that release of payment may please be subjected to deduction of contract award fee at the rate of 0.025% at the time o payment as given under purchase manual 1991 and submission of security deposit 2.5% on the total amount of work order production of inspection note, warranty and Bank challan of Rs. 1500 in respect of CDL and Rs, 1000 in respect of PDTL as drug testing fee for every batch of supply

*DR SHAFQUATHUSSAIN DAHERI*  
Executive District Officer Health  
District Government Mirpurkhas.



PHONE #. 9290097

FAX 9290098.

MPS/Store 70221 / 24 Dated 2/26 - 2011

**OFFICE OF THE  
EXECUTIVE DISTRICT OFFICER HEALTH,  
DISTRICT GOVERNMENT MIRPURKHAS.**

To,

The M/S Pak Traders,  
214 International Auto Parts Market, Marston Road .  
Karachi

Subject:-

**PROCUREMENT OF INSTRUMENTS / EQUIPMENT/ & FURNITURE  
TENDER 2010.11 FOR USE OF MEDICAL DISPENSARIES.**

You are requested to please arrange / supply the following products as per terms and Conditions of the above tender and submit the bill / invoice with warranty in triplicate for necessary payment.

S.NO	DESCRIPTION OF STORE	MAKE	QUANTITY REQUIRED	RATE / AU	TOTAL VALUE
1	Viginal Speculum	Hilbro Pak Made	3	470.00	1410.0
2	Enemma Can	Hilbro Pak Made	3	1260.00	3780.00
					<b>5190.00</b>

TOTAL VALUE IN WORDS :- RS. Five thousand one hundred & ninety rupees only.

*[Signature]*  
DR SHAFQUAT HUSSAIN DAHERI  
Executive District Officer Health  
District Government Mirpurkhas.

Cc to:-

The District coordination Officer / District Administrator District Government Mirpurkhas.  
The Secretary to Government of Sindh Health Department Karachi.  
The Account branch EDOH Office Mirpurkhas, with instructions that release of payment may please be subjected to deduction of contract award fee at the rate of 0.025% at the time of payment as given under purchase manual 1991 and submission of security deposit 2.5% on the total amount of work order production of inspection note, warranty and Bank challan of Rs. 1500 in respect of CDL and Rs. 1000 in respect of PDDL as drug testing fee for every batch of supply

*[Signature]*  
DR SHAFQUAT HUSSAIN DAHERI  
Executive District Officer Health  
District Government Mirpurkhas.

**SINDH PUBLIC PROCUREMENT REGULATORY AUTHORITY**

**CONTRACT EVALUATION FORM**

**TO BE FILLED IN BY ALL PROCURING AGENCIES FOR PUBLIC CONTRACTS OF WORKS, SERVICES & GOODS WORTH 50 MILLION (PKR) OR ABOVE**

- 1) NAME OF THE ORGANIZATION / DEPTT. E.D.O HEALTH MIRPURKHAS
- 2) PROVINCIAL / LOCAL GOVT / OTHER PROVINCIAL
- 3) TITLE OF CONTRACT TENDER INSTRUMENT/EQUIP / MACHINERY
- 4) TENDER NUMBER 02
- 5) BRIEF DESCRIPTION OF CONTRACT AS ABOVE
- 6) FORUM THAT APPROVED THE SCHEME DISTRICT ADMINISTRATION
- 7) TENDER ESTIMATED VALUE 11.268 MILLION
- 8) ENGINEER'S ESTIMATE  
(For civil works only) \_\_\_\_\_
- 9) ESTIMATED COMPLETION PERIOD (AS PER CONTRACT) CURRENT FINANCIAL YEAR
- 10) TENDER OPENED ON (DATE & TIME) 31.05.2011 @ 01:PM
- 11) NUMBER OF TENDER DOCUMENTS SOLD 12  
(Attach list of buyers)
- 12) NUMBER OF BIDS RECEIVED 12
- 13) NUMBER OF BIDDERS PRESENT AT THE TIME OF OPENING OF BIDS 12
- 14) BID EVALUATION REPORT  
(Enclose a copy) ALREADY SUBMITTED. KARACHI
- 15) NAME AND ADDRESS OF THE SUCCESSFUL BIDDER 1. SINDH MEDICAL STORE. " "  
2. MUSLIM MEDICAL SERVICES " "  
3. GLOBAL MARKETING. " "
- 16) CONTRACT AWARD PRICE AS PER WORK ORDERS.
- 17) RANKING OF SUCCESSFUL BIDDER IN EVALUATION REPORT  
(i.e. 1<sup>st</sup>, 2<sup>nd</sup>, 3<sup>rd</sup> EVALUATION BID). 1<sup>st</sup> - 03  
2<sup>nd</sup> NIL  
3<sup>rd</sup> NIL

18) METHOD OF PROCUREMENT USED : - (Tick one)

- a) SINGLE STAGE - ONE ENVELOPE PROCEDURE \_\_\_\_\_
- b) SINGLE STAGE - TWO ENVELOPE PROCEDURE \_\_\_\_\_
- c) TWO STAGE BIDDING PROCEDURE \_\_\_\_\_
- d) TWO STAGE - TWO ENVELOPE BIDDING PROCEDURE \_\_\_\_\_

PLEASE SPECIFY IF ANY OTHER METHOD OF PROCUREMENT WAS ADOPTED i.e. EMERGENCY, DIRECT CONTRACTING / NEGOTIATION ETC. WITH BRIEF REASONS:

19) APPROVING AUTHORITY FOR AWARD OF CONTRACT DISTRICT ADMINISTRATION

20) WHETHER THE PROCUREMENT WAS INCLUDED IN ANNUAL PROCUREMENT PLAN?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

21) ADVERTISEMENT :

i) SPPRA Website  
(If yes, give date and SPPRA Identification No.)

Yes	<input checked="" type="checkbox"/>	10-4009
No	<input type="checkbox"/>	

ii) News Papers  
(If yes, give names of newspapers and dates)

Yes	<input checked="" type="checkbox"/>	DAILY BUSINESS RECORDER 08-05-2011 DAILY QANAT 10-05-2011
No	<input type="checkbox"/>	DAILY UMMAT 09-05-2011 DAILY IMROZ 08-05-2011

22) NATURE OF CONTRACT

Local	<input checked="" type="checkbox"/>	Int.	<input type="checkbox"/>
-------	-------------------------------------	------	--------------------------

23) WHETHER QUALIFICATION CRITERIA WAS INCLUDED IN BIDDING / TENDER DOCUMENTS?  
(If yes, enclose a copy)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

24) WHETHER BID EVALUATION CRITERIA WAS INCLUDED IN BIDDING / TENDER DOCUMENTS?  
(If yes, enclose a copy)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

25) WHETHER APPROVAL OF COMPETENT AUTHORITY WAS OBTAINED FOR USING A METHOD OTHER THAN OPEN COMPETITIVE BIDDING?

Yes	<input type="checkbox"/>	No	<input checked="" type="checkbox"/>
-----	--------------------------	----	-------------------------------------

26) WAS BID SECURITY OBTAINED FROM ALL THE BIDDERS?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

27) WHETHER THE SUCCESSFUL BID WAS LOWEST EVALUATED BID / BEST EVALUATED BID (in case of Consultancies)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

28) WHETHER THE SUCCESSFUL BIDDER WAS TECHNICALLY COMPLIANT?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

29) WHETHER NAMES OF THE BIDDERS AND THEIR QUOTED PRICES WERE READ OUT AT THE TIME OF OPENING OF BIDS?

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

30) WHETHER EVALUATION REPORT GIVEN TO BIDDERS BEFORE THE AWARD OF CONTRACT?  
(Attach copy of the bid evaluation report)

Yes	<input checked="" type="checkbox"/>	No	<input type="checkbox"/>
-----	-------------------------------------	----	--------------------------

31) ANY COMPLAINTS RECEIVED  
(If yes, result thereof)

Yes	RESERVATIONS CLARIFIED ✓ & CD RETURNED ON HIS WRITTEN REQUEST
No	

32) ANY DEVIATION FROM SPECIFICATIONS GIVEN IN THE TENDER NOTICE / DOCUMENTS  
(If yes, give details)

Yes	✓
No	

33) WAS THE EXTENSION MADE IN RESPONSE TIME?  
(If yes, give reasons)

Yes	.
No	✓ .

34) DEVIATION FROM QUALIFICATION CRITERIA  
(If yes, give details and reasons)

Yes	✓
No	

35) WAS IT ASSURED BY THE PROCURING AGENCY THAT THE SELECTED FIRM IS NOT  
BLACK LISTED?

Yes	✓	No	
-----	---	----	--

36) WAS A VISIT MADE BY ANY OFFICER/OFFICIAL OF THE PROCURING AGENCY TO THE  
SUPPLIER'S PREMISES IN CONNECTION WITH THE PROCUREMENT? IF SO, DETAILS TO  
BE ASCERTAINED REGARDING FINANCING ARRANGEMENT, IF ANY, ROAD:  
(If yes, enclose a copy)

Yes		No	✓
-----	--	----	---

37) WERE PROPER SAFEGUARDS PROVIDED ON MOBILIZATION ADVANCE PAYMENT IN  
THE CONTRACT (BANK GUARANTEE ETC.?)

Yes		No	✓
-----	--	----	---

38) SPECIAL CONDITIONS, IF ANY  
(If yes, give Brief Description)

Yes	
No	✓

Signature & Official Stamp of  
Authorized Officer

  
SPPRA DISTRICT OFFICER HEALTH

FOR OFFICE USE ONLY

SPPRA, Block. No.8, Sindh Secretariat No.4-A, Court Road, Karachi  
Tele: 021-9205356; 021-9205369 & Fax: 021-9206291

**Report evaluation committee regarding technical bids / tender Equipment Instrument and Machinery for the up gradation of Blood Bank DHQ Hospital Mirpurkhas and Laboratory Dialysis Center At Civil Hospital Mirpurkhas.**

Tender #	Name of Items	M/S Muslim Medical Services	M/S Sindh Medical Store	M/S SM Trader	M/S Platinum	M/S Cosmos Enterprises (Offer No.1)	M/S Cosmos Offer No.2	M/S Oasis Medical System	M/S Global Marketing	Ms Lab Link Enterprises	M/S Hospi Care Systems	SK Trading	M/S Shah Enterprises
1	<b>Cryofuge Automated Refrigerated centrifuge floor standing.</b>	Hettich-Germany Against tender specification / qualification criteria	1.Make Helmer USA 2.against tender specification.	NQ	HERMLE-GERMANY/ SIGMA-GERMANY Against tender specification / qualification criteria	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1.Make Ortoalresa: against tender specification/ qualification criteria	Make: Sorval / ThermoFisher Germany as per tender specification / qualification criteria.	Against tender specification / qualification criteria	Make : Hermle against specification	Hanil- Korea Against tender specification	Against tender specification / qualification criteria
2	<b>Platelet Agitator</b>	NQ	1.Make Helmer USA 2.as per tender specification.	NQ	NQ	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1.Make: Helmer. Against qualification criteria	Make: Helmer USA / BMS Canada as per tender specification / qualification criteria	Against tender specification / qualification criteria	Baxter-USA against qualification criteria	Helmer, USA qualification criteria	Against tender specification / qualification criteria
3	<b>Platelet Agitator Incubator</b>	NQ	1.Make Helmer USA 2.as per tender specification.	NQ	NQ	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1.Make: Helmer. Against qualification criteria	Make: Helmer USA / BMS Canada as per tender specification / qualification criteria	NQ	Baxter-USA against qualification criteria	Helmer, USA qualification criteria	Against tender specification / qualification criteria
4	<b>Blood Bag Expresser</b>	NQ	1. Make _____ 2.against tender specification	NQ	NQ	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1.Make: Delcon Germany: against qualification criteria	NQ	Against tender specification / qualification criteria	Baxter-USA Against tender specification / qualification criteria	Qualtek - Italy Against tender specification	Against tender specification / qualification criteria
5	<b>Tube Sealer</b>	NQ	1. Make Dalcone Italy 2.against tender specification	NQ	NQ	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1.Make: Delcon Germany: against qualification criteria	1.Make Centroan korea as per tender specification / qualification criteria	Against tender specification / qualification criteria	Make : Genesis / Buxter against tender specification & qualification criteria	Delcon-Italy against qualification criteria	Against tender specification / qualification criteria
6	<b>Weighing Scale</b>	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	Against tender specification / qualification criteria	1.Make Mr Radwag-Germany against tender specification.	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria
7	<b>Table Top Centrifuge</b>	Daihan-Korea Against tender specification / qualification criteria	1.Make Sigma Germany As Per specification	BOECO=GERMANY Against tender specification / qualification criteria	SIGMA GERMANY Against tender specification / qualification criteria	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make Centurion scientific UK: against tender qualification	NQ	Against tender specification / qualification criteria	Make Sigma against qualification criteria	MPW= Poland Against tender specification	Against tender specification / qualification criteria

8	<b>Table Top Centrifuge</b>	Hetich-Germany Against tender specification / qualification criteria	1.Make Sigma Germany 2.against tender specification.	BOECO=GERMANY Against tender specification / qualification criteria	SIGMA GERMANY Against tender specification / qualification criteria	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make Centurion scientific UK: agaisnt tender qualification	NQ	Against tender specification / qualification criteria	Make Select against qualification criteria	MPW= Poland Against tender specification	Against tender specification / qualification criteria
9	<b>Water Bath-Large</b>	Haier-China as per tender specification / qualification criteria	1.Make GS research koriae 2.against tender specification	MEDLINE- UKAgainst tender specification / qualification criteria	Memmert Germany Against tender specification	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make:Memmert Gemany: agasint qualification criteria.	NQ	Against tender specification / qualification criteria	NQ	Biosan-Litvia Against tender specification / qualification criteria	Against tender specification / qualification criteria
10	<b>Water Bath-Small</b>	Haier-China as per tender specification / qualification criteria	1.Make GS research koriae 2.against tender specification	MEDLINE- UKAgainst tender specification / qualification criteria	Memmert Germany Against tender specification	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Make:Memmert Gemany: agasint qualification criteria.	NQ	Against tender specification / qualification criteria	NQ	Biosan-Litvia Against tender specification / qualification criteria	Against tender specification / qualification criteria
11	<b>Microscope</b>	Motic Against tender specification / qualification criteria	NQ	BOECO=GERMANY Against tender specification / qualification criteria	Meiji Japan Against tender specification	1.Make :GX Optical UK against tender specification	Against tender specification / qualification criteria	Make: Optika Italy: agasint tender specification	Make: Leica Germany as per tender specification / qualification criteria.	Against tender specification / qualification criteria	NQ	Olympus Against tender specification	Against tender specification / qualification criteria
12	<b>Blood Bag Shaker</b>	NQ	1.Hemotech / Dalcone italy as per tender specification	NQ	NQ	NQ	Against tender specification / qualification criteria	1.Make : Delcon Italy: agaisnt tender Qualification criteria	Make:Centron korea as per tender specification/ qualification criteria.	NQ	Genesis Against tender specification / qualification criteria	Delcon-Italy against qualification criteria	Against tender specification / qualification criteria
13	<b>Plasma Freezer with temp chart recorder</b>	Haier-China Against tender specification / qualification criteria	1Make Italy 2against tender specification.	LABOLOGIC-FRANCE Against tender specification / qualification criteria	Sanyo Japan Against tender specification.	1.Make : Fouchetti Italy 2.as per tender specification / qualification critria	Against tender specification / qualification criteria	1Make :Angelantoni Italy: against tender Specification	NQ	Against tender specification / qualification criteria	NQ	Ilshin - Korea Against tender specification / qualification criteria	Against tender specification / qualification criteria
14	<b>Plasma Freezer</b>	Haier-China Against tender specification / qualification criteria	Make Italy as per tender specification / qualifigation criteria	LABOLOGIC-FRANCE Against tender specification / qualification criteria	Sanyo Japan Against tender specification.	1.Make : Fouchetti Italy 2.as per tender specification / qualification critria	Against tender specification / qualification criteria	1Make :Angelantoni Italy: against tender Specification	NQ	Against tender specification / qualification criteria	NQ	Mieling -China against qualification criteria	Against tender specification / qualification criteria
15	<b>Blood Bank Refrigerator</b>	Haier-China Against tender specification / qualification criteria	1.Make Frimed Italy/ BMS Canada as per tender specification / qualification criteria.	NQ	Sanyo Japan Against tender specification.	1.Make :Vission scieitific south korea against tender specification	Against tender specification / qualification criteria	1Make :Angelantoni Italy: against tender Specification	NQ	Against tender specification / qualification criteria	Make Frimed against qualification criteria	Mieling -China against qualification criteria	Against tender specification / qualification criteria

16	<b>Blood Bank Refrigerator</b>	Haier-China Against tender specification / qualification criteria	Make: Helmer USA: / BMS Canada as per tender specification/ qualification.	NQ	Sanyo Japan Against tender specification.	1.Make :Vission scieitific south korea as per tender specification / qualification critirea	Against tender specification / qualification criteria	1Make :Angelantoni Italy: against tender Specification	NQ	Against tender specification / qualification criteria	Make Frimed against qualification criteria	Mieling -China against qualification criteria	Against tender specification / qualification criteria
17	<b>Pharmacy Refrigerator</b>	Haier-China Against tender specification / qualification criteria	1.Make CCI / USA 2.against qualification criteria	NQ	Sanyo Japan Against tender specification.	1.Make :Vission scieitific south korea against tender specification / qualifiction critirea	Against tender specification / qualification criteria	1Make :Angelantoni Italy: against tender Specification	NQ	Against tender specification / qualification criteria	Against tender specification / qualification criteria	Mieling Against tender specification / qualification criteria	Against tender specification / qualification criteria
18	<b>Blood Donation Chair/Venation couches</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	Agaisnt tender specification / qualification cirteria.	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria
19	<b>Computer</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria
20	<b>Printer</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria
21	<b>Printer</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria
22	<b>Generator</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria
23	<b>Split Air conditioner</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	Against tender specification / qualification criteria

24	<b>Gel Cord System</b>	NQ	1.Diamed / Swiss As per tender specification / qualification criteria	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria
25	<b>Weighing Machine</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria
26	<b>Stabilizer</b>	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria	NQ	NQ	NQ	NQ	NQ	Against tender specification / qualification criteria
27	<b>Haematology Analyzer</b>	NQ	1. Make Melet France As per tender specification / qualification criteria	NQ	NQ	1.Make: Nihon Kohden Japan against tender speification	Make BGT Germany:against tender specification / qualification cirteria	1.Make Cell-Tac/ Nihon Japon agaisnt tender specification /qualification cirteris.	NQ	Against tender specification / qualification criteria	Make Diru China against specification & qualification criteria	NQ	Against tender specification / qualification criteria
28	<b>Semi Automated Bio Chemistry Analyzer</b>	NQ	1.Make Awareness technology USA As per tender specification / qualification criteria	NQ	NQ	1.Make: ISE Italy against tender specifiction	Make Clindiag Belgium: against tender specification / qualification cirteria	1.Make:Unilyzer UK agaisnt tender specification /qualification cirteris	Make Vital Scientific, Neatherland as per tender specification / qualification criteria	Against tender specification / qualification criteria	Make Diru China against specification & qualification criteria	Biolab-France Against tender specification	Against tender specification / qualification criteria
29	<b>Electrolyte Analyzer</b>	NQ	1.Make SFR1 France Against tender specification	NQ	NQ	1.Make: Electrical Medical USA against per tender specification	Make: K-Lite Eleetrolyte Aust:against tender specification / qualification cirteris	Make Diamond USA against tender specification /qualification cirteris	Make:Opti Medical USA .as per tender specification/ qualification cirteria.	Against tender specification / qualification criteria	Diamond-USA Against tender specification	Meizhou - China Against tender specification	Against tender specification / qualification criteria

DR M Hanif Baloch  
(DOH) Admin/ Account

Dr Ahmed Khan  
(DDO Procurement/ Logistics)

Mr Aziz Ahmed Daudpoto  
(DDO F&P) Mirpurkhas

Dr Hussain Khan Mari  
Focal Person (WFP)  
District Mirpurkhas

DR Waleed Bin Azhar  
(Chief Operating Officer  
Fatmid Foundation Karachi  
Technical Advisor to Evaluation Committee

Dr Shafquat Hussain Daheri  
Executive District Officer Health  
District Mirpurkhas