

**ADMIN/090/TKC-2013**

May 13, 2013



To,  
Director  
Sindh Public Procurement Regulatory Authority  
Government of Sindh  
Block - 8, Sindh Secretariat No. 4A,  
Court Road  
**KARACHI**  
Phone: 021-9205356 & 69

Subject: **PUBLICATION OF BID EVALUATION REPORT ON SPPRA WEBSITE**


Dear Sir,

Please find enclosed herewith the original Bid Evaluation Reports for procurement of the following equipments:

1. Dialysis Machines
2. Reverse Osmosis System

It is requested that the above reports maybe uploaded on SPPRA website.

For **THE KIDNEY CENTRE**  
Post Graduate Training Institute

  
Pervez Taj  
Manager Procurement

EF-2  
SPPRA INWARD DIARY

No: 10704  
Dated: 15.5

The Kidney Centre Post Graduate Training Institute, 197/9, Rafiqi Shaheed Road, Karachi.75530  
Tel: 35661000-10 Fax: 35661040 & 35661050

E-mail: mail@kidneycentre.com Website: [Http://www.kidneycentre.com](http://www.kidneycentre.com)

## THE KIDNEY CENTRE POST GRADUATE TRAINING INSTITUTE

BID EVALUATION REPORT NIT No. ADMIN/379/TKC-2012 DATED 20.11.2012

### REFERENCE : SPPRA – CLAUSE 45

The tenders were advertised on 19<sup>th</sup> November, 2012 in the following three News Papers:

- |                  |         |
|------------------|---------|
| a. The News      | English |
| b. Daily Express | Urdu    |
| c. Daily Kawish  | Sindhi  |

The bids were received and opened on 6<sup>th</sup> December, 2012. Following quoted:

- Hoorra Pharma
- Universal Enterprises
- Hospital Supply Corporation
- B.Braun Pakistan (Pvt.) Ltd
- Oriental Sales Corporation
- Fresenius Medical Care

The above bids were evaluated and all the above bids, except that of M/s. Hoorra Pharma who did not meet the tender's requirement and also did not submit the earnest money, were accepted.

Thereafter the technical scrutiny was carried out, results of which are given below:

Sr. No.	Bidders Name	TKCI Points	Remarks
<b>DIALYSIS MACHINES</b>			
1	Hospital Supply Corporation	The Machine is not FDA / CE Marked	Not Accepted
2	B. Braun Pakistan (Pvt.) Ltd	Due to kind of disinfectant solution (Citric Acid 50%) used for the Dialysis Machines	Not Accepted
		Due to the changeable roller system for acceptability of the blood tubing lines	
3	Fresenius Medical Care	Required water pressure of the machine is high	Not Accepted
		Due to kind of disinfectant solution (citric acid 25%) used for the dialysis machine	
		Due to our current experience of gel formation in drain lines of the Fresenius machines at The Kidney Centre	
		The system is with bibag system, not bicart	
4	Universal Enterprises	Bicart (Powder Bicarbonate) facility not available with the Nipro Surdial55 machine	Not Accepted
5	Universal Enterprises	The bid for Nipro Surdial55 Plus machine meets all the specifications of The Kidney Center	Accepted

The Kidney Centre Post Graduate Training Institute, 197/9, Rafiqi Shaheed Road, Karachi, 75530  
Tel: 35661000-10 Fax: 35661040 & 35661050

E-mail: mail@kidneycentre.com Website: Http://www.kidneycentre.com

REVERSE OSMOSIS SYSTEMS 1200 & 1600 litres/hours			
1	Fresenius Medical Care	The maintainable pressure by the system is comparatively very less which is also at the first point of the loop only.	Not Accepted
		The rated flow is not given with the rated temperature	
2	Oriental Sales Corporation	The bid for R.O. Systems meets all the specifications of The Kidney Center	Accepted
3	B.Braun Pakistan (Pvt.) Ltd	The bid for R.O. Systems meets all the specifications of The Kidney Center	Accepted

The bidders whose bids were technically not accepted were informed accordingly. The financial bids of only the following bidders whose bids were technically accepted and who were informed in advance, were opened on 18<sup>th</sup> February, 2013. The position is as under:

Sl. No	Bidder Name	TKC Points	Unit Price
<b>DIALYSIS MACHINES</b>			
1	Universal Enterprises	Dialysis Machine Nipro Surdial 55 Plus	USD 12,985.00
<b>REVERSE OSMOSIS SYSTEMS 1200 LITRES/HOUR</b>			
1	B.Braun Pakistan (Pvt.) Ltd	R.O. System 1200 Litres/ Hour	Euro 99,010.00
2	Oriental Sales Corporation	R.O. System 1200 Litres/ Hour	Euro 153,700.00
<b>REVERSE OSMOSIS SYSTEMS 1600 LITRES/HOUR</b>			
1	B.Braun Pakistan (Pvt.) Ltd	R.O. System 1600 Litres/ Hour	Euro 117,510.00
2	Oriental Sales Corporation	R.O. System 1600 Litres/ Hour	Euro 162,400.00

The following bidders whose bids were technically not accepted applied for redressal of grievances. A

Complaint Redressal Committee as per SPPRA clause 31 was constituted.

- Oriental Sales Corporation
- Fresenius Medical Care
- Hospital Supply Corporation
- B.Braun Pakistan (Pvt.) Ltd

The Complaint Redressal Committee, after due deliberations has upheld the decision of the Purchasing Committee. All the complainants and the concerned authorities have been informed accordingly.

The contracts are being awarded as under:

DIALYSIS MACHINES		
1	Dialysis Machines	Universal Enterprises
REVERSE OSMOSIS SYSTEMS 1200 & 1600 litres/hour		
1	B.Braun Pakistan (Pvt.) Ltd	R.O. System 1200 & 1600 Litres/ Hour

*[Signature]*  
14/5/13  
Mr. Shauqat Lillah  
Chairman  
Purchasing Committee

*[Signature]*  
14/5/13  
Muhammad Ahmed  
Member  
Purchasing Committee

*[Signature]*  
14/5/13  
Pervez Taj  
Member  
Purchasing Committee

*[Signature]*  
14/5/13  
Dr. Hasan Nasir Abbas  
Member  
Purchasing Committee

*[Signature]*  
Dr. Abdul Wahab  
Member  
Purchasing Committee  
Representative Govt. of Sindh

*[Signature]*  
14/5/13  
Javali Ahmed  
Member  
Purchasing Committee

The Kidney Centre Post Graduate Training Institute, 197/9, Rafiqi Shaheed Road, Karachi. 75530  
Tel: 5661000-10 Fax: 5661040 & 5661050

E-mail: mail@kidneycentre.com



Website: <http://www.kidneycentre.com>