



No(IFB NO: HD/PRO-eqp/2017 )

**GOVERNMENT OF SINDH  
HEALTH DEPARTMENT**

Karachi dated: 28-12- 2017

To,

The Manager (Assessment),  
Sindh Public Procurement Regulatory Authority,  
Govt. of Sindh, Karachi.

**SUBJECT:- Procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi (SPPRA ID NO. 34700)**

I am directed to submit the following documents for hoisting at SPPRA website.

1. Copy of Notification of Procurement / Technical Committee.
2. Minutes of meeting for opening of Technical Bids.
3. Technical Evaluation report (by Procurement Committee)
4. Technical Evaluation report (by Technical Committee)
5. Attendance Sheet of Technical Bids Opening meeting
6. Minutes of Financial Meeting
7. Financial Evaluation report (by Procurement Committee)
8. Bids Evaluation Reports
9. Attendance Sheet of Financial meeting
10. Intimation to all Bidders
11. Annual Procurement Plan.
12. Certificate of rates as per SPPRA rule 48.
13. Copy of NIT Published in newspapers.
14. Copy of Notification of Complaint Redressal Committee.
15. Copy of minutes of the meeting of pre bid meeting.

  
SECTION OFFICER (PM&I)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

SECTION OFFICER (PM&I)



No(IFB NO: HD/PRO-eqp/2017)

**GOVERNMENT OF SINDH  
HEALTH DEPARTMENT**

Karachi dated: 28-12- 2017

To,

The Manager (Assessment),  
Sindh Public Procurement Regulatory Authority,  
Govt. of Sindh, Karachi.

**SUBJECT:- Procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi (SPPRA ID NO. 34700)**

I am directed to submit the following documents for hoisting at SPPRA website.

1. Copy of Notification of Procurement / Technical Committee.
2. Minutes of meeting for opening of Technical Bids.
3. Technical Evaluation report (by Procurement Committee)
4. Technical Evaluation report (by Technical Committee)
5. Attendance Sheet of Technical Bids Opening meeting
6. Minutes of Financial Meeting
7. Financial Evaluation report (by Procurement Committee)
8. Bids Evaluation Reports
9. Attendance Sheet of Financial meeting
10. Intimation to all Bidders
11. Annual Procurement Plan.
12. Certificate of rates as per SPPRA rule 48.
13. Copy of NIT Published in newspapers.
14. Copy of Notification of Complaint Redressal Committee.
15. Copy of minutes of the meeting of pre bid meeting.

**SECTION OFFICER (PM&I)**

**C.c. to:-**

1. The PS to Secretary, Health Department, Karachi.
2. Office File

  
**SECTION OFFICER (PM&I)**



## NOTIFICATION

No.SO (PM&I) Health./Procurement/2017: In supersession of this department's notification of even number dated 2<sup>nd</sup> October, 2017 and in pursuance of Rule-07 of the Sindh Public Procurement Rules, 2010 (Amended 2017) a Procurement Committee comprising of following Officers for Purchase of Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital Karachi, Health Department, Government of Sindh, for the year 2017-18, is constituted as under;-

- |    |  |                   |
|----|--|-------------------|
| 1- | Additional Secretary (Technical),<br>Health Department, GoS                                      | Chairman          |
| 2- | Section Officer-IV<br>Health Department, GoS   | Member/ Secretary |
| 3- | Professor of Microbiology,<br>Director, Dow Diagnostic & Research Laboratories,<br>DUHS, Karachi | Member            |

### Function & Responsibilities of the Committee

- Preparing bidding documents;
- Carrying out technical as well as financial evaluation of the bids;
- Preparing evaluation reports as provided in Rule-45;
- Making recommendations for the award of contract to the competent authority; and
- Performing any function ancillary and incidental to the above.

- DR. FAZLULLAH PECHUHO-  
SECRETARY TO GOVT. OF SINDH

No.SO (PM&I) Health./Procurement/2017

dated 25<sup>th</sup> October - 2017

Copy forwarded for information & necessary action to:

1. All Members of Committee.
2. The P.S. to Minister, Health Department, Govt. of Sindh, Karachi.
3. The P.S to Secretary, Health Department, Govt. of Sindh, Karachi
4. The Additional Secretary (Dev.), Health Department, Government of Sindh, Karachi
5. Office Order File

SECTION OFFICER (PM&I)



# GOVERNMENT OF SINDH HEALTH DEPARTMENT

## NOTIFICATION

No.SO (PM&I) Health./TC/CHK/2017 :- A Technical Committee is hereby constituted comprising of following members for technical evaluation of bidding documents / specifications for Purchase of Instruments & Equipments for the Central Laboratory of Dr. Ruth K.M Pfau Civil Hospital Karachi, Health Department, Government of Sindh, for the year 2017-18.

01.	Dr. Ghulam Fatima Chief Pathologist, CHK-Central Lab, Dr. Ruth KM Pfau Civil Hospital, Karachi	<b>Chairman</b>
02.	Prof. Dr. Shaheen Sharafat Professor of Microbiology Director, Dow Diagnostic & Research Laboratories DUHS, Karachi	<b>Member</b>
03.	Dr. Suresh Kumar Pathologist CHK-Central Lab, Dr. Ruth KM Pfau Civil Hospital, Karachi	<b>Member</b>


**SECRETARY HEALTH**

No.SO (PM&I) Health./TC/CHK/2017

Karachi dated 4<sup>th</sup> October, 2017

**C.c. to:-**

1. The Chairman and all members of the Committee.
2. The PS to Secretary Health.
3. Office File

  
SECTION OFFICER (PM&I)



**MINUTES OF THE MEETING OF PROCUREMENT COMMITTEE**  
**HELD ON 25-10-2017**

A meeting of Procurement Committee for procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi was held on 25-10-2017 at the office of Health Department 3<sup>rd</sup> Floor, Tughlaq House, Sindh Secretariat Building No.2, Kamal Atta Turk Road, Karachi to evaluate the bids for procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi.

It was informed that the N.I.T. was called on 05-10-2017. A pre-bid conference was held on **16.10.2017** at 2.30 PM at the office of Health Department 3<sup>rd</sup> Floor, Tughlaq House, Sindh Secretariat Building No.2, Kamal Atta Turk Road, Karachi, minutes of pre-bid meeting is attached as (Annexure-A). (07) Tender Documents were issued and accordingly (06) firms participated on the date of opening of tenders. Tenders were opened by the Procurement Committee on 25-10-2017 in presence of the bidders/ representatives of the firms.

Thereafter the bids were technically evaluated (Annex-B) in the light of the criteria mentioned in the bidding documents.

The Procurement Committee unanimously recommended that the technical evaluated bid of **M/S Hospicare Systems, M/S Muslim Medical Services & M/S S. Ejazuddin & Co** may be approved for opening of Financial Evaluation.

Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member

Syed qasim Akber Nimai  
Section Officer (IV)  
Health Department  
Member

Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman



## HEALTH DEPARTMENT GOVERNMENT OF SINDH

**Evaluation for Procurement of Various Medical Instruments & Equipment for Dr. Ruth K.M. Pfau Civil Hospital Karachi (IFB NO: HD/Pro-eqp/2017/) ELIGIBILITY CRITERIA**

N.I.T. NO. (IFB NO: HD/Pro-eqp/2017/)

DATE & TIME OF ISSUE OF TENDERS UPTO	05-10-2017 to 24-10-2017	5.00pm
DATE & TIME OF OPENNING OF BIDS.	10/25/2017	3:30PM
DATE & TIME OF BIDS/TENDERS SUBMISSION.	10/25/2017	3:00PM

SR. #	DETAIL	M/S Global Marketing Services	M/S S. Ejazuddin & Co.	M/S Popular International Pvt Ltd	M/S Hospicare Systems	M/S National Agencies	M/S Muslim Medical Services
1	Original receipt code purchase of tender.	Yes	yes	yes	yes	yes	yes
2	Copy of earnest money without amount in technical bid.	Yes	yes	yes	yes	yes	yes
3	Acceptance of terms and conditions, Tender documents duly signed and stamped by signing authority.	yes	yes	yes	yes	yes	yes
4	Original printed literature / brochure of product showing complete technical details of quoted model and manufacturer.	color copies	yes	yes	yes	yes	yes
5	Copy of GST certificate.	yes	yes	yes	yes	yes	yes
6	Copy of NTN certificate.	yes	yes	yes	yes	yes	yes
7	Copy of last three years paid income tax return.	yes	yes	yes	yes	sales tax given not income tax	yes

Dr. Shaheen Sharafat  
Professor of Microbiology, Director, DOW Diagnostic & Research Laboratories DUHS, Karachi  
Member

Mr. Aslam Pechuho  
Additional Secretary (Tech)  
Health Department, Govt. of Sindh  
Chairman

Syed Qasim Akber Nimal  
Section Officer-IV  
Health Department, Govt. of Sindh  
Member

8	Bank certificate showing financial capability.	<i>not showing</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>
9	Price should not be mentioned on technical bid.	<i>Yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>
10	The bidder must provide list of well reputed organizations / departments in which the bidder has supplied the similar quoted equipment.	<i>yes</i>	<i>yes</i>	<b>no any work order</b>	<i>yes</i>	<b>certificate of general and surgical items given no same assignment</b>	<i>yes</i>
11	complete details of similar assignments/ material supplied and installed during last three years in Sindh.	<i>no</i>	<i>yes</i>	<b>no any work order</b>	<i>yes</i>	<b>certificate of general and surgical items given no same assignment</b>	<i>yes</i>
12	copy of valid professional tax certificate.	<i>no</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>
13	undertaking on judicial paper, stating that the equipment to be supplied is brand new and of latest, free of defects in design material and workmanship and will bind himself for availability of spare for a period of 03 years from the date of supply.	<i>no</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<b>No</b>	<i>yes</i>
14	An affidavit should be provided by supplier that the firm has not been black listed from any Govt./ Autonomous organization.	<i>Yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>	<i>yes</i>
		<b>Not Qualified</b>	<b>Qualified</b>	<b>Not Qualified</b>	<b>Qualified</b>	<b>Not Qualified</b>	<b>Qualified</b>



Dr. Shaheen Sharafat  
Professor of Microbiology, Director, DOW Diagnostic &  
Research Laboratories DUHS, Karachi  
Member



Mr. Aslam Pechuho  
Additional Secretary (Tech)  
Health Department, Govt. of Sindh  
Chairman



Syed Qasim Akber Nimai  
Section Officer-IV  
Health Department, Govt. of Sindh  
Member



## HEALTH DEPARTMENT GOVERNMENT OF SINDH

Evaluation for Procurement of Various Medical Instruments & Equipment for Dr. Ruth  
K.M. Pfau Civil Hospital Karachi (IFB NO: HD/Pro-eqp/2017/) ELIGIBILITY

### CRITERIA

N.I.T. NO. (IFB NO: HD/Pro-eqp/2017/)

DATE & TIME OF ISSUE OF TENDERS UPTO

05-10-2017 to

5.00pm

DATE & TIME OF OPENING OF BIDS.

10/25/2017

3:30PM

DATE & TIME OF BIDS/TENDERS SUBMISSION.

10/25/2017

3:00PM

SR. #	DETAIL	Total Marks	M/S S. Ejazuddin & Co.	M/S Hospicare Systems	M/S Muslim Medical Services
	<b>Company / Firm / Individual Information</b>				
1	No of Years in Business	15	15	15	15
2	Annual Turnover in Millions (last three years)	15	15	15	15
	<b>Technical Proposal</b>				
3	Specifications & Brochures	25	25	25	25
	<b>Financial Capabilities</b>				
4	Income Tax Annual Returns of 5 Years	10	10	10	10
5	Audited Financial Statements/ Bank statement of 5 years	10	10	10	10
6	<b>Relevant Field Experience</b>	25	25	20	25
	<b>Total marks obtained out of 100</b>	<b>100</b>	<b>100</b>	<b>95</b>	<b>100</b>
	<b>Result</b>		<b>Qualified</b>	<b>Qualified</b>	<b>Qualified</b>

Dr Shaheen Sharafat,  
Professor of Microbiology, Director, DOW Diagnostic &  
Research Laboratories DUHS, Karachi  
Member

Syed qasim Akber Nimai  
Section Officer (IV) Health  
Department  
Member

Mr. Aslam Pechuho, Additional Secretary (Technical) Health Department  
Govt of Sindh,  
Chairman



**MINUTES OF MEETING OF TECHNICAL COMMITTEE FOR PROCUREMENT OF VARIOUSE MEDICAL INSTRUMENTS & EQUIPMENTS for Dr. RUTH K.M PFAU CIVIL HOSPITAL, KARACHI FOR EVALUATION OF TECHNICAL BIDS /TECHNICAL PROPOSAL HELD ON 25-10-2017 AT 04.00 PM AT THE OFFICE OF HEALTH DEPARTMENT 3RD FLOOR, TUGHLAQ HOUSE, SINDH SECRETARIAT BUILDING NO 2, KAMAL ATTA TURK ROAD, KARACHI**

S.No	Description/ specification of goods	Specification of the item / good offered by the bidder M/S S. Ejazuddin & Co.	Specification of the item / good offered by the bidder M/S Hospicare Systems	Specification of the item / good offered by the bidder M/S Mulim Medical Services	Qty	Remarks
1	AUTOMATED IDENTIFICATION AND ANTIBIOTIC SUSCEPTIBILITY TESTING INSTRUMENT	Not Quoted	Phoenix M50 System Make BD USA	Not Quoted	1	
	Detect current and emerging resistances accurately		Detect current and emerging resistances accurately			
	Accelerate ID/AST results reporting		Accelerate ID/AST results reporting			
	Capacity: 100 sample panel at one time.		50 panel capacity to stacked 100 panel capacity to grow			
	Random Panel Entry		Random Panel Entry			
	Should have 51 ID wells (45 substrates)		51 ID wells (45 substrates)			
	Should have 85 AST wells (Doubling Dil.)		85 AST wells for true MIC (Doubling Dil.)			
		Not Quoted	Qualified	Not Quoted		



2	<b>Model</b>	Flow cytometer (3 laser & 8 color)	Not Quoted	FACScanto II Make BD USA	Not Quoted	1
	<b>Made</b>	USA				
	<b>Optics</b>	488-sm Solid state, 20 mW output		488-sm Solid state, 20 mW output		
		633-nm HeNe, 17 mW output		633-nm HeNe, 17 mW output		
		Air-cooled 405-nm solid state diode, 30-mW fiber power output		Air-cooled 405-nm solid state diode, 30-mW fiber power output		
		Fixed optical alignment		Fixed optical alignment		
		Spatially separated beams with 9x65 um elliptical spots		Spatially separated beams with 9x65 um elliptical spots		
		Gel Coupled 1.2 NA lens		Gel Coupled 1.2 NA lens		
		Blue Laser Dyes FITC, PE, PerCP or PerCP-Cy <sup>TM</sup> 5.5, PE-Cy <sup>TM</sup> 7 (525, 575, 678 or 695, 785 nm)		Blue Laser Dyes FITC, PE, PerCP or PerCP-Cy <sup>TM</sup> 5.5, PE-Cy <sup>TM</sup> 7 (525, 575, 678 or 695, 785 nm)		
	Red Laser Dyes APC, APC-Cy7 (660, 785 nm)	Red Laser Dyes APC, APC-Cy7 (660, 785 nm)				
	Violet: Pacific Blue <sup>TM</sup> , AmCyan (455, 488 nm)	Violet: Pacific Blue <sup>TM</sup> , AmCyan (455, 488 nm)				
	<b>Fluidics</b>	Automated startup and shutdown and cleaning cycles		Automated startup and shutdown and cleaning cycles		
		Less than or equal to 0.1% carryover		Less than or equal to 0.1% carryover		
Sample flow rate min 10uL/min, max 120uL/min		Sample flow rate min 10uL/min, max 120uL/min				
Sample Acquisition rate 10,000 events/second		Sample Acquisition rate 10,000 events/second				
<b>Parameters</b>	Area (A), Width (W), Height (H) for all channels with upto 2 ratios, and time (T)	Area (A), Width (W), Height (H) for all channels with upto 2 ratios, and time (T)				
<b>Software</b>	FACSDiva <sup>TM</sup> v7.0 or above, FACS clinical v2.4 or above	FACSDiva <sup>TM</sup> v7.0 or above, FACS clinical v2.4 or above				
<b>Operating system</b>	Windows based	Windows based				
		Not Quoted	Qualified	Not Quoted		

*Handwritten signature*

*Handwritten signature*

*Handwritten signature*

3	<b>VERTICAL STEAM STERILIZER (AUTOCLAVE) 85-100 Litres</b>			<b>HVE-85 HIRAYAMA - JAPAN</b>		1	
	<b>Autoclave</b>						
	<b>Features:</b>						
	Capacity	85 – 110 Litres			85 Litres		
	Minimum Chamber size (mm)	300 X 710D					
	Maximum Chamber size (mm)	420 X 795D			420X615D		
	Power source	AC 220, 50/60Hz			AC 220, 50/60Hz		
	Max Power consumption	2.0KW			3.0KW		
	Warming temperature range	45 to 60 °C			45 to 80 °C		
	Agar dissolution temperature range	60 to 100°C			105 to 135°C		
	Chamber material	SUS304 (Stainless steel)			SUS304 (Stainless steel)		
	Sterilization timer	1 to 250 minutes			1 to 250 minutes		
	Auto start timer	1 minutes to 7 days					
Pressure gauge	0 to 0.4 Mpa			0 to 0.4 Mpa			
			<b>Not Quoted</b>	<b>Qualified</b>	<b>Not Quoted</b>		
	<b>HI-TECH FULLY AUTOMATED HEMATOLOGY ANALYZER</b>	<b>XN-1000 06 parts fully Automated Hematology Analyzer of M/S SYSMEX Corporation /Japan</b>				1	
	<b>7 Part Diff with RETICS, Body Fluids with Auto-Sampler</b>	<b>7 Part Diff with RETICS, Body Fluids with Auto-Sampler</b>					
	Can perform CBC with standard NRBC, 7part Diff. Reportable IG% & IG#, Ret-He, dedicated Platelet fluorescence, IPF	Can perform CBC with standard NRBC, 7part Diff. Reportable IG% & IG#, Ret-					
Parameters:	HGB, RBC, HCT, MCV, MCH, MCHC, RDW-SD, RDW CV, MPV,	HGB, RBC, HCT, MCV, MCH, MCHC, RDW-SD, RDW CV, MPV,					
	WBC, NEUT%, LYMPH%, MONO%, EO%, BASO%, NRBC%, IG%.	WBC, NEUT%, LYMPH%, MONO%, EO%, BASO%, NRBC%, IG%.					

*Shreyas Jais* *Prish*

	NEUT#, LYMPH#, MONO#, EO#, BASO#, NRBC#, IG#,	NEUT#, LYMPH#, MONO#, EO#, BASO#, NRBC#, IG#,
	PLT, P-LCR, PDW, PCT, IPF	PLT, P-LCR, PDW, PCT, IPF
	RET%, RET#, LFR, MFR, HFR, IRF, RET-He	RET%, RET#, LFR, MFR, HFR, IRF, RET-He
	WBC-BF, RBC-BF, MN%, MN#, PMN%, PMN#, TC-BF#	WBC-BF, RBC-BF, MN%, MN#, PMN%, PMN#, TC-BF#
Technology:	Fluorescence Flow Cytometry using semi-conductor laser to perform	Fluorescence Flow Cytometry using semi-conductor laser to perform
	WBC, WBC 6 DIFF, NRBC, RET	WBC, WBC 6 DIFF, NRBC, RET
	Hydrodynamic focusing DC Detection Method for RBC, PLT	Hydrodynamic focusing DC Detection Method for RBC, PLT
	PLT-F Specific Separated Chanel Fluorescence Method.	PLT-F Specific Separated Chanel Fluorescence Method.
	Cynide Free SLS method for HGB.	Cynide Free SLS method for HGB.
Throughput:	NRBC by default with each CBC	NRBC by default with each CBC
	100 samples per hour	100 samples per hour
	Min. 7 reportable FDA approved Body Fluids analysis of	Min. 7 reportable FDA approved Body Fluids analysis of
	CSF, CAPD, Synovial & Serous Fluids	CSF, CAPD, Synovial & Serous Fluids
Discrete Modes:	10-12 Discrete Modes with auto dilution mode.	10-12 Discrete Modes with auto dilution mode.
	Specific Body Fluids mode.	Specific Body Fluids mode.
Sample Modes:	Sampler Mode, Manual Mode.	Sampler Mode, Manual Mode.
	Manual Closed Mode, Capillary Mode, Body Fluid Analysis.	Manual Closed Mode, Capillary Mode, Body Fluid Analysis.
Data Storage Capacity:	100,000 samples data storage with histogram/ Scatter grams.	100,000 samples data storage with histogram/ Scatter grams.

4

Not Quoted

Not Quoted

*Shree Jy*

*[Signature]*

Reagent Management:	RFID (Radio Frequency Identification) of Reagent Lot, Expiry Date, Type of Reagent, Number of tests, etc. Color Coded Reagent compartment.	RFID (Radio Frequency Identification) of Reagent Lot, Expiry Date, Type of Reagent, Number of tests, etc. Color Coded Reagent compartment.
Low WBC Mode:	To ensure accurate differential count in marked leukopenic sample (Limited <0.1)	To ensure accurate differential count in marked leukopenic sample (Limited <0.1)
Onboard Rules:	Auto Rerun, Repeat & Reflex Functions	Auto Rerun, Repeat & Reflex Functions
Specific Leukocyte	To meet the need of True Blasts, Blasts Crisis and abnormal lymphocytic	To meet the need of True Blasts, Blasts Crisis and abnormal lymphocytic
Flagging cells.		Flagging cells.
QC Material:	3 levels of controls for 44 reportable parameters.	3 levels of controls for 44 reportable parameters.
QC Control:	20 QC files with 300 data points.	20 QC files with 300 data points.
	1 X bar M file with 300 data points.	1 X bar M file with 300 data points.
International	FDA/ CE Marked.	FDA/ CE Marked.
Approvals:		
Software:	For Strong Flaggings	For Strong Flaggings
Ret He & IRF:	Random access retic analysis with reportable IRF, LFR, MFR.	Random access retic analysis with reportable IRF, LFR, MFR.
Origin:	USA/ Europe/ Japan.	Japan
		Qualified

			Not Quoted	Not Quoted	

*Steacey*  
*Jan*  
*John*

<b>Automated Coagulation Analyzer</b>		<b>Sysmex CS-1600 Japan Automated Coagulation Analyzer</b>				1	
<b>Principles</b>	<b>Multi-Wave Detection System:</b>	Multi-Wave Detection System:					
	Transmitted Light Detection Method	Transmitted Light Detection Method					
	<b>Clotting Assays:</b>	Clotting Assays:					
	405,660,800nm (Percentage Detection Method)	405,660,800nm (Percentage Detection Method)					
	<b>Chromogenic Assays:</b>	Chromogenic Assays:					
	405nm (Rate Method)	405nm (Rate Method)					
	<b>Immunoassays:</b>	Immunoassays:					
	575,800nm (Rate Method, VLI Method)	575,800nm (Rate Method, VLI Method)					
<b>Detector</b>	8channels for Clotting Assays, Chromogenic Assays and Immunoassays	8channels for Clotting Assays, Chromogenic Assays and Immunoassays					
	The transmitted light (or A/D value) is detected every	The transmitted light (or A/D value) is detected every					
	0.1 seconds	0.1 seconds					
	Maximum reading time is 1800seconds	Maximum reading time is 1800seconds					
<b>Parameters</b>	<b>Parameters Clotting Assays:</b>	Parameters Clotting Assays:					
	PT, APTT, Fbg, Dabigatran, Batroxobin Time(Reptilase), TT,	PT, APTT, Fbg, Dabigatran, Batroxobin Time(Reptilase), TT,					
	Extrinsic Factors (II, V, VII, X), Intrinsic Factors (VIII, IX, XI, XII),	Extrinsic Factors (II, V, VII, X), Intrinsic Factors (VIII, IX, XI, XII),					
	PS, PC, LA, ProC Global	PS, PC, LA, ProC Global					
	<b>Chromogenic Assays:</b>	Chromogenic Assays:					
	AT-III, PLG, $\alpha$ 2 -AP, PC, FVIII, Heparin, Rivaroxaban,	AT-III, PLG, $\alpha$ 2 -AP, PC, FVIII, Heparin, Rivaroxaban,					
	C -Inhibitor	C -Inhibitor					
	<b>Immunoassays:</b>	Immunoassays:					
	D-Dimer, vWF:Ag, vWF Ac, Free Protein S, FDP	D-Dimer, vWF:Ag, vWF Ac, Free Protein S, FDP					

*Atsuee Jun*

*[Signature]*


5

Throughput	PT: 120test/h	PT: 120test/h
	PT,APTT: 110 test/h	PT,APTT: 110 test/h
	PT,APTT,Fbg,DD: 70tests/h	PT,APTT,Fbg,DD: 70tests/h
	PT, APTT 50u L	PT, APTT 50u L
Sample	Fbg 10uL PC 15 uL	Fbg 10uL PC 15 uL
Volume	TT 100uL	TT 100uL
	D-Dimer, vWF Ag, vWF Ac 15uL	D-Dimer, vWF Ag, vWF Ac 15uL
	Extrinsic Factors 5uL	Extrinsic Factors 5uL
	Intrinsic Factors 5uL	Intrinsic Factors 5uL
	Rivaroxaban 16u L	Rivaroxaban 16u L
	ATIII, Free protein S, 2α -AP, PLG 21uL	ATIII, Free protein S, 2α -AP, PLG 21uL
	Dabigatran 40 uL	Dabigatran 40 uL
Reagent Holder	28 positions ( 15 °C)	28 positions ( 15 °C)
	5 positions (room temperature)	5 positions (room temperature)
Auto Sampler	Capacity of 50 samples (rack type)	Capacity of 50 samples (rack type)
STAT	Set in sample rack (sample interruption or rack interruption)	Set in sample rack (sample interruption or rack interruption)
Reaction Tubes	Automatically supplied from an internal hopper	Automatically supplied from an internal hopper
	Capacity: 300 tubes	Capacity: 300 tubes
Sample Aspiration	The sample probe, with a liquid surface sensor, quantitatively aspirates plasma from centrifuged whole blood samples	The sample probe, with a liquid surface sensor, quantitatively aspirates plasma from centrifuged whole blood samples
Quality Control	X control, Levy-Jennings control	X control, Levy-Jennings control
	Multi-rule (Westgard Rule) monitoring	Multi-rule (Westgard Rule) monitoring
	300data points x 40 files in memory	300data points x 40 files in memory

Not Quoted

Not Quoted

*Success* *Jur*



	<b>Data Storage</b>	3000sample results with reaction curves	3000sample results with reaction curves			
	<b>Print Out</b>	Graphic or Data Printer (option)	Graphic or Data Printer (option)			
	<b>Dimensions</b>	760 mm x 690mm x 540mm	760 mm x 690mm x 540mm			
	<b>W x D x H (mm)</b>	Approx. 85 (kg)	Approx. 85 (kg)			
	<b>Weight (kg)</b>					
	<b>Power Requirement</b>	100 to 240VAC ( 50 or 60Hz)	100 to 240VAC ( 50 or 60Hz)			
	<b>Power Consumption</b>	680VA max	680VA max			
	<b>Optional Instruments</b>	Graphic Printer, Data Printer	Graphic Printer, Data Printer			
			<b>Qualified</b>	<b>Not Quoted</b>	<b>Not Quoted</b>	
6	<b><u>Pharmaceutical Refrigerator</u></b>		<b>Not Quoted</b>	<b>Not Quoted</b>	<b>HYC -1378</b>	1
	Specifications:				Haier	
	Max Capacity (L)	1400			<b>1378</b>	
	Voltage/Frequency	202/50			202/50	
	Max Power (W)	800			800	
	Temp.range (°C)	2~8			2~8	
	Door	Double-layer glass			Double-layer glass	
	Shelves/Inner door	12/-			12/-	
	Max. Interior Dimensions(mm) (W*D*H):	1320*700*1500			1320*700*1500	
	Max Exterior Dimensions(mm) (W*D*H):	1440*875*1980			1440*875*1980	
Net/gross Weight (kg)	310/360	310/360				
Container Load (sets) 20'/40'/40'H	6/12/12	6/12/12				
			<b>Not Quoted</b>	<b>Not Quoted</b>	<b>Qualified</b>	

*Handwritten signatures and initials in blue ink.*



<u>Ultra Low Freezer</u>	
<b>Features:</b>	
Temperature range: -	40°C to -86°C
GUARD.CONTROL	automatic door mechanism for easy door opening and closing.
Personalized	access control with RFID technology.
VIP.COMPLETE	internal, large surface vacuum insulation panels.
Two-stage cascade refrigeration system with hermetically-sealed compressors.	
Non-flammable refrigerant, CFC-free.	
Heated duplex door gasket prevents icing.	
User friendly microprocessor controller with LED display.	
Alarm test system.	
Built-in visual and audible alarm system: for power failure.	
Impermissible temperature deviation or battery failure.	
Battery operated backup system for full functionality of alarm function in case of power failure for 72 hours.	

7

Not Quoted

<b>ULTRA LOW FREEZER -80</b>	<b>DW -86L628</b>	1
<b>UFV 500 Binder- Germany</b>	<b>Haier China</b>	
40°C to -86°C	40°C to -86°C	
automatic door mechanism for easy door opening and closing.	automatic door mechanism for easy door opening and closing.	
access control with RFID technology.	access control with RFID technology.	
internal, large surface vacuum insulation panels.	internal, large surface vacuum insulation panels.	
Two-stage cascade refrigeration system with hermetically-sealed compressors.	Two-stage cascade refrigeration system with hermetically-sealed compressors.	
Non-flammable refrigerant, CFC-free.	Non-flammable refrigerant, CFC-free.	
Heated duplex door gasket prevents icing.	Heated duplex door gasket prevents icing.	
User friendly microprocessor controller with LED display.	User friendly microprocessor controller with LED display.	
Alarm test system.	Alarm test system.	
Built-in visual and audible alarm system: for power failure.	Built-in visual and audible alarm system: for power failure.	
Impermissible temperature deviation or battery failure.	Impermissible temperature deviation or battery failure.	
Battery operated backup system for full functionality of alarm function in case of power failure for 72 hours.	Battery operated backup system for full functionality of alarm function in case of power failure for 72 hours.	

*Handwritten signatures and initials:*  
 H. S. ...  
 J. ...  
 P. ...

RS 422 interface	
Potential-free contact	
Analog output 4 – 20 mA.	
Inner chamber made of high quality stainless steel.	
Robust, adjustable and lockable casters.	
Easy air filter replacement without tools.	
<b>Specification:</b>	
<b>Max Exterior dimensions</b>	
Width (mm)	900
Height (mm)	1970
Depth (mm)	935
<b>Temperature data</b>	
Setting range (°C)	-90° -- -40°
Average temperature variation (K) <sup>3</sup>	± 2.5

RS 422 interface	RS 422 interface
Potential-free contact	Potential-free contact
Analog output 4 – 20 mA.	Analog output 4 – 20
Inner chamber made of high quality stainless steel.	Inner chamber made of high quality stainless steel.
Robust, adjustable and lockable casters.	Robust, adjustable and lockable casters.
Easy air filter replacement without tools.	Easy air filter replacement without tools.
900	900
1970	1970
935	935
-90° -- -40°	-90° -- -40°
± 2.5	± 2.5

	Not Quoted	Qualified	Qualified		
--	------------	-----------	-----------	--	--

<b><u>Centrifuge Machine with Swing out Rotor</u></b>
Country of Origin: Germany / UK / USA
Features & Specifications:
Capacity: 32-48 tubes
Speed: 5000 – 7000 RPM
RCF: 4000 – 5000 xg
Entry of parameters can be carried out fast and conveniently via the waterproof foil-keypad. Display should be clear and easy to read

<b>ROTOFIX 32A MAKE:</b>
<b>Hettich Germany</b>
Country of Origin: Germany
4x100ml
6000 RPM
4186 xg
easy entry of parameters can be carried out fast and conveniently via the waterproof foil-keypad. Display should be clear and easy to read

8




not Quoted

not Quoted

2

*Shereen* *Jus* *[Signature]*

	Actual values should be displayed during centrifugation			Actual values should be displayed during centrifugation			
	Input in increments of 100 min-1			Input in increments of 100 min-1			
	Time, t/min: Input in minutes.			Time, t/min: Input in minutes.			
	Impulse key: For short centrifuging operations. The rotor spins at the pre-selected RPM as long as the key remains pressed.			Impulse key: For short centrifuging operations. The rotor spins at the pre-selected RPM as long as the key remains pressed.			
	Timer: 1 to 99min or continuous			Timer: 1 to 99min or continuous			
			Not Quoted	Qualified	Not Quoted		
9	<b>VORTEX MIXER</b>		Not Quoted	Not Quoted	RTV -100 Robus (UK2		
	Features:						
	Provide Thorough Mixing.						
	Molded Neoprene Mixing Cup can be easily replaced.						
	Easy & Constant Operation						
	A Cast Aluminum Base Plate provides a low center of gravity, which prevents creeping.						
	Fuse Protection.						
	Source Power: 220V, 50 Hz.				110-220 v 60Hz		
			Not Quoted	Not Quoted	Qualified		

10	<b>Gram Stainer</b>	
	<b>Features:</b>	
	Slide Carousel Capacity: 12 or 30 slides minimum	
	Carousel Rotation Speed: Reagent application 20 RPM, slide drying approx. 950 RPM	
	Cytocentrifuge Rotor Speed: Programmable, 100 to 2000 RPM	
	Display: 4 line by 16 character alpha-numeric LCD dot-matrix	
	Touch Pad Controls: 9 Decolorizer selections, 10 Numeric Programming Keys, Run, Stop, Clean, Program, Cyto-Centrifuge and Volume Test	
	Slides per 500 mL stain pack at medium setting	Maximum 1200 (12 slides/run) Minimum 600 (2 slides/run)
	Reagent Consumption (combined, excluding water): Ranges from 0.7 to 4.0 ml. Medium settings will use 1 to 2 mL per slide	

Not Quoted




<b>Aerospray Gram Series 2</b>
Make: Wescor - USA
1 to 12 or 1 to 30 slides
950 RPM for 45 seconds
100-2000rpm
not mentioned
not mentioned
Maximum 1032 (12 slides/run)
Minimum 996 (12 slides/run)
not mentioned

Not Quoted

1	

*Stacey* *Jim* *[Signature]*

	Operating Time:294 - 475 seconds, depending upon decolorizer setting		Operating Time:294 - 475 seconds, depending upon decolorizer setting		
	Dimensions (W x H x D): 56 cm x 24 cm x 53 cm (22" x 9.5" x 21") maximum		57cm x 25cm x 54cm		
	Electrical Requirements:100V, 120V, 220V or 240V (user-selectable) @ 50 to 60 Hz. 75 watts		100 to 240 V		
	Drain Connection:Nipple on rear panel		Outlet on rear panel		
	Safety: Lid interlock: Lid must be closed, and is locked down during operation		Lid interlock: Lid must be closed, and is locked down during operation		
		Not Quoted	Qualified	Not Quoted	
11	<u>Hematology Stat Slide Stainer/Cytocentrifuge</u>	SYSMEX SP-10 Japan	Aerospray Hematology Series 2		1
	Specifications	Japan	USA		
	Slide Carousel Capacity: 1 to 12 Slides	02 slides per specimen/ 120 samples /hr in real time	1 to 12 or 1 to 30 slides		
	Cytocentrifuge Rotor Speed 100 to 2000 rpm (± 5%), user programmable	Cytocentrifuge Rotor Speed 100 to 2000 rpm (± 5%), user programmable	100 -2000 rpm		
	Electrical Requirements 100 to 240 VAC (±10%) @ 50-60 Hz, 85 watts maximum	Electrical Requirements 100 to 240 VAC (±10%) @ 50-60 Hz, 85 watts maximum	100 to 240 V	Not Quoted	

	Drain Connection Outlet on rear panel		Drain Connection Outlet on rear panel	Drain Connection Outlet on rear panel		
	Safety Lid interlock: Lid must be closed for operation, and is locked down during carousel rotation.		Safety Lid interlock: Lid must be closed for operation, and is locked down during carousel rotation.	Safety Lid interlock: Lid must be closed for operation, and is locked down during carousel rotation.		
	Rotation indicator: LED on front panel lights when the carousel or rotor is rotating.		Rotation indicator: LED on front panel lights when the carousel or rotor is rotating.	Rotation indicator: LED on front panel lights.		
			Qualified	Qualified	NOT Qualified	



Dr Shaheen Sharafat,  
Professor of Microbiology & Director,  
DOW Diagnostic & Research Laboratories,  
DUHS, Karachi

Member


**Prof. Dr. Shaheen Sharafat**

M.B.B.S. (Dow), M.Phil, Ph.D

Chairperson Pathology,

Director

Dow Diagnostic Research and Reference Laboratory (DDRRL)  
Dow University of Health Sciences



Dr Suresh Kumar,  
Pathologist,

CHK-Central Lab,

Dr Ruth KM Pfau civil Hospital , Karachi

Member

**Dr. Suresh Kumar**

M.B.B.S. (SINDE) M. Phil

Consultant Pathologist

Civil Hospital Karachi



Dr Ghulam Fatima,

Chief Pathologist,

CHK-Central Lab,

Dr. Ruth KM Pfau Civil Hospital, Karachi,

Chairperson

**DR. GHULAM FATIMA**

Chief Consultant Pathologist

C.H.K. Central Lab

Civil Hospital, Karachi






**MEETING OF OPENING OF BIDS FOR PROCUREMENT OF  
MEDICAL INSTRUMENTS & EQUIPMENTS FOR  
DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI ON WEDNESDAY  
25<sup>TH</sup> OCTOBER, 2017 @ 04:00 P.M. AT HEALTH DEPARTMENT AT 3RD FLOOR,  
TUGHLAQ HOUSE, SINDH SECRETARIAT BUILDING NO.2, KAMAL ATTA  
TURK ROAD, KARACHI**

**ATTENDANCE SHEET FOR PROCUREMENT COMMITTEE**

S.No.	Name	Designation with Department	Contact Number	Signature
1	Dr. No. Akbar	Addl. Sec Health	0301-2129591	<i>Akbar</i>
2.	SHAHGON SHARAFAT	PROFESSOR OF PATHOLOGY BPS. 21	0300-9294794	<i>Shahgon</i>
3.	Qasim Akbar Nimee	SO-IV H.D.	0333-257-8867	<i>Qasim</i>

**MEETING OF OPENING OF BIDS FOR PROCUREMENT OF  
MEDICAL INSTRUMENTS & EQUIPMENTS FOR  
DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI ON WEDNESDAY  
25<sup>TH</sup> OCTOBER, 2017 @ 04:00 P.M. AT THE OFFICE OF ADDITIONAL  
SECRETARY (PM&I), HEALTH DEPARTMENT AT 3RD FLOOR, TUGHLAQ  
HOUSE, SINDH SECRETARIAT BUILDING NO.2, KAMAL ATTA TURK ROAD,  
KARACHI**

**ATTENDANCE SHEET FOR BIDDERS**

S.No.	Name with designation	Company Name	Contact Number	Signature
1	Popular International (PVI) Ltd M. Shafi	M. Shafi Popular International (PVI) Ltd	03008291683	
2	Khussam Hossain Sales & Application Officer	S. E. Juddin E. W.	03432940365	
3	Muhammad Ali (BUM)	Global Marketing Services	0321 2594512 ali@gms-world.com	
4	Zubair Hussain	Muslim Medical Service	0300550116 zubair.hussain90@gmail.com	
5	Zubair Hussain	Hospicare Systems	0300550116 zubair.hussain90@gmail.com	





**MINUTES OF FINANCIAL BIDS OPENING MEETING HELD ON**  
**18-12-2017**

NIT No.	(IFB NO: HD/PRO-eqp/2017)
SPPRA Sr No.	<b>34700</b>
Method & Procedure of Procurement	<i>Single stage Two envelop procedure</i>

**Procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi**

A meeting of Procurement Committee was held for opening of financial bid on 18-12-2017 at 3.00 pm in respect of subject NIT for procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi. Meeting was attended by all members of procurement committee (PC) and representative of bidders. (Copy of attendance sheet is attached as Annexure A)

The financial bid of technically qualified bidder was opened at 3.00 pm in presences of all participants. The rates quoted by bidders was read aloud and encircled by the Chairman of the PC. All the members of the PC signed each and every page of the financial proposal/bids. The bids don't contain any over writing or cutting. The procurement committee also signed the certificate as per rule 48 of SPPRA where needed. (Attached as annexed B).

The financial evaluation is attached as (Annex-C). The Procurement Committee unanimously approved the financial rates given by M/S Hospicare Systems for item No 01,02,03,08,10&11 , M/S Muslim Medical Services for item No. 06,07&09 M/S S. Ejazuddin & Co for item No. 04 &05.

Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member

Syed Qasim Akber Nimai  
Section Officer (IV)  
Health Department  
Member

Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman



HEALTH DEPARTMENT  
GOVERNMENT OF SINDH

**Financial Evaluation for Procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi.**

NIT No.	(IFB NO: HD/PRO-eqp/2017)
SPPRA Sr No.	<b>34700</b>
Method & Procedure of Procurement	<b>Single stage Two envelop procedure</b>
Technical Bid Opening date	<b>25-10-2017</b>
Financial Bid Opening date	<b>18-12-2017</b>

**Various Medical Instruments & Equipments:**

Item No	Description of items	Required Quantity	Name of Lowest Qualified Bidder	Unit Price in C.I.F.	Total Price in C.I.F.	Unit Price in D.D.P.	Total Price in D.D.P.	Status
1	AUTOMATED IDENTIFICATION AND ANTIBIOTIC SUSCEPTIBILITY TESTING INSTRUMENT	01	M/S Hospicare Systems	\$78,000.00	\$78,000.00	Rs, 12,500,000/=	Rs,12,500,000/=	Only Technically Qualified
2	FLOW CYTOMETER (3 laser & 8 color)	01	M/S Hospicare Systems	\$118,000.00	\$118,000.00	Rs,19,000,000/=	Rs,19,000,000/=	Only Technically Qualified
3	VERTICAL STEAM STERILIZER (AUTOCLAVE) 85-100 Litres	01	M/S Hospicare Systems	-----	-----	Rs,1,175,000/=	Rs,1,175,000/=	Only Technically Qualified

 *reha shah*


4	SEVEN PART AUTOMATED HAEMATOLOGY ANALYZER	01	M/S Ejazuddin &Co.	JPY:5,4 87,800. 00	JPY:5,48 7,800.00	Rs,8,025,908/=	Rs,8,025,908/=	Only Technically Qualified
5	AUTOMATED COAGULATION ANALYZER	01	M/S Ejazuddin &Co.	JPY:4,6 64,400. 00	JPY:4,66 4,400.00	Rs,6,821,685/=	Rs,6,821,685/=	Only Technically Qualified
6	PHARMACEUTICAL REFRIGERATOR IMPORTED TEMP 2-8°C 1340 LITRES	01	M/S Muslim Medical Services	----	----- --	Rs,825,000/=	Rs,825,000/=	Only Technically Qualified
7	ULTRA LOW FREEZER (-80°C)	01	M/S Muslim Medical Services	-	-	Rs,995,000/=	Rs 995,000/=	Lowest
			M/S Hospicare Systems	-	-	Rs,1,195,000/=	Rs1,195,000/=	Highest
8	CENTRIFUGE MACHINE 48 HOLES	02	M/S Hospicare Systems	-	-	Rs,525,000/=	Rs 1,050,000/=	Only Technically Qualified
9	VORTEX MIXER	02	M/S Muslim Medical Services	-	-	Rs,45,000/=	Rs 90,000/=	Only Technically Qualified
10	AUTOMATED GRAM STAINER	01	M/S Hospicare Systems	-	-	Rs,1,450,000/=	Rs 1,450,000/=	Only Technically Qualified

 *reza hussain*


11	AUTOMATED HAEMATOLOGY SLIDE STAINER	01	M/S Hospicare Systems	-	-	Rs,1,450,000/=	Rs 1,450,000/=	Lowest
			M/S Ejazuddin &Co.	JPY:7,9 26,394. 00	JPY:7,92 6,394.00	Rs,11,592,351/=	Rs 11,592,351/=	Highest



Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member



Syed Qasim Akber Nimai  
Section Officer (IV)  
Health Department  
Member



Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman



HEALTH DEPARTMENT  
GOVERNMENT OF SINDH

**Bid Evaluation Report**

1. Name Of Procurement Agency: Health Department, Government of Sindh
2. Tender Reference No: (IFB NO: HD/PRO-eqp/2017)
3. Tender description/Name of work/items: Procurement of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi.
4. Method of Procurement: Single stage Two envelop procedure
5. Tender Published: SPPRA Serial No. 34700, Jang (07-10-2017), The News (07-10-2017), Kawish (10-10-2017)
6. Total Bid documents Sold: (07)
7. Total Bids Received: (06)
8. Technical Bid Opening date: 25-10-2017
9. No. of Bid technically qualified: Three (03)
10. Bid(s) Rejected: 00
11. Financial Bid Opening date: 18-12-2017
12. C.S.& Minutes: Attached
13. Bid Evaluation Report:

S. #	Name of Firm/Bidder	Cost offered by the Bidder	Ranking in term of cost	Comparison with Estimated Cost	Reason for acceptance/rejection	Remarks
1	S. Ejazuddin & Co.	Please see Financial Statement	Please see Financial Statement		Technically Qualified	Items wise rates have been recorded in financial evaluation
2	M/S Hospicare Systems	Please see Financial Statement	Please see Financial Statement		Technically Qualified	Items wise rates have been recorded in financial evaluation
3	M/S Muslim Medical Services	Please see Financial Statement	Please see Financial Statement		Technically Qualified	Items wise rates have been recorded in financial evaluation
4	M/S Global Marketing Services				Technically not Qualified	
5	M/S Popular International Pvt Ltd				Technically not Qualified	

6	<b>M/S National Agencies</b>				Technically not Qualified	
---	------------------------------	--	--	--	---------------------------	--



Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member






Syed Qasim Akber Nimai  
Section Officer (IV)  
Health Department  
Member



Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman

**MEETING FOR FINANCIAL OPENING OF BIDS FOR PURCHASE OF  
VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH  
K.M. PFAU CIVIL HOSPITAL, KARACHI ON MONDAY 18<sup>TH</sup> DECEMBER,  
2017 @ 03:00 P.M. AT HEALTH DEPARTMENT AT 3RD FLOOR, TUGHLAQ  
HOUSE, SINDH SECRETARIAT BUILDING NO.2, KAMAL ATTA TURK ROAD,  
KARACHI**

**ATTENDANCE SHEET FOR PROCUREMENT COMMITTEE**

S.No.	Name	Designation with Department	Contact Number	Signature
1	Dr. M. Akbar Peshawar	Add. Sec Health	0300-2429591	
2.	Dr. Shahzeen Sharafat	Professor BPS-21 Pathology DUHS	0300-9294759	
3.	Qazi Akbar Nizami	SO-IV	0333- 251-8867	

**MEETING FOR FINANCIAL OPENING OF BIDS FOR PURCHASE OF  
VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH  
K.M. PFAU CIVIL HOSPITAL, KARACHI ON MONDAY 18<sup>TH</sup> DECEMBER,  
2017 @ 03:00 P.M. AT HEALTH DEPARTMENT AT 3RD FLOOR, TUGHLAQ  
HOUSE, SINDH SECRETARIAT BUILDING NO.2, KAMAL ATTA TURK ROAD,  
KARACHI**

**ATTENDANCE SHEET FOR BIDDERS**

S.No.	Name with designation	Company Name	Contact Number	Signature
①	Zubair Hussain	Hospicare System	03000550116	Zubair
②	Kashif Hasmat	Muslim Medical Service	03000550024	Kashif
3	Babar Haseeb Khan 14201	Combi Momen Apparatus Spe. S Eja Zubair	03008279480	Babar





NO.HD/PRO-EQP/2017  
GOVERNMENT OF SINDH  
HEALTH DEPARTMENT

Karachi dated 15<sup>th</sup> December, 2017

To,

M/S S. Ejazuddin & Co.  
236/C, Block-II, PECHS  
Shahrah-e-Quideen, Karachi  
021-34537151

**SUBJECT:- MEETING FOR FINANCIAL OPENING OF BIDS FOR PURCHASE OF VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI.**

I am directed to refer to the subject cited above and to inform that a meeting for Financial Opening of Bids submitted for purchase of various medical instruments and equipments for Dr. Ruth K.M. Pfau Civil Hospital, Karachi is scheduled to be held on **Monday 18-12-2017 at 03.00 p.m.** at Health Department, 3<sup>rd</sup> Floor, Tuglaq House, Sindh Secretariat Building No 02, Kamal Atta Turk Road, Karachi.

It is therefore requested to make it convenient to attend the meeting on above mentioned date, time and venue.

  
SECTION OFFICER-IV (H)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

  
SECTION OFFICER-IV (H)



NO.HD/PRO-EQP/2017  
GOVERNMENT OF SINDH  
HEALTH DEPARTMENT

Karachi dated 15<sup>th</sup> December, 2017

To,

M/S Hospicare Systems,  
Mezzanine Floor, Rabbiya Garden,  
Plot No.3, Block-3 MCHS Society,  
Shaheed-e-Millat Road, Karachi.  
021-34921913.

**SUBJECT:- MEETING FOR FINANCIAL OPENING OF BIDS FOR PURCHASE OF VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI.**

I am directed to refer to the subject cited above and to inform that a meeting for Financial Opening of Bids submitted for purchase of various medical instruments and equipments for Dr. Ruth K.M. Pfau Civil Hospital, Karachi is scheduled to be held on **Monday 18-12-2017 at 03.00 p.m.** at Health Department, 3<sup>rd</sup> Floor, Tuglaq House, Sindh Secretariat Building No 02, Kamal Atta Turk Road, Karachi.

It is therefore requested to make it convenient to attend the meeting on above mentioned date, time and venue.

  
SECTION OFFICER-IV (H)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

  
SECTION OFFICER-IV (H)



NO.HD/PRO-EQP/2017  
GOVERNMENT OF SINDH  
HEALTH DEPARTMENT

Karachi dated 15<sup>th</sup> December, 2017

To,

M/S Muslim Medical Services,  
Ground Floor, Rabbiya Garden,  
Plot No.3, Block-3 MCHS Society,  
Shaheed-e-Millat Road, Karachi.  
021-34921913.

**SUBJECT:- MEETING FOR FINANCIAL OPENING OF BIDS FOR PURCHASE  
OF VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS  
FOR DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI.**

I am directed to refer to the subject cited above and to inform that a meeting for Financial Opening of Bids submitted for purchase of various medical instruments and equipments for Dr. Ruth K.M. Pfauf Civil Hospital, Karachi is scheduled to be held on **Monday 18-12-2017 at 03.00 p.m.** at Health Department, 3<sup>rd</sup> Floor, Tuglaq House, Sindh Secretariat Building No 02, Kamal Atta Turk Road, Karachi.

It is therefore requested to make it convenient to attend the meeting on above mentioned date, time and venue.

SECTION OFFICER-IV (H)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

SECTION OFFICER-IV (H)



**NO.HD/PRO-EQP/2017**  
**GOVERNMENT OF SINDH**  
**HEALTH DEPARTMENT**

Karachi dated 15<sup>th</sup> December, 2017

To,

M/S Global Marketing Services,  
House No.227, Street No.12, Block-3,  
Bahadurabad Karachi.  
021-34547985

**SUBJECT:- PURCHASE OF VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI.**

I am directed to inform you that as per the technical evaluation done by the Procurement committee your firm is declared as NOT QUALIFIED.

SECTION OFFICER-IV (H)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

SECTION OFFICER-IV (H)



**NO.HD/PRO-EQP/2017**  
**GOVERNMENT OF SINDH**  
**HEALTH DEPARTMENT**

Karachi dated 15<sup>th</sup> December, 2017

To,

M/S Popular International (Pvt) Ltd,  
141 Justice Inamullah Road, Block-7 & 8,  
KMCHS Near Hill Park, Karachi.  
021-34524827

**SUBJECT:- PURCHASE OF VARIOUS MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH K.M. PFAU CIVIL HOSPITAL, KARACHI.**

I am directed to inform you that as per the technical evaluation done by the Procurement committee your firm is declared as NOT QUALIFIED.

SECTION OFFICER-IV (H)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

SECTION OFFICER-IV (H)



**NO.HD/PRO-EQP/2017  
GOVERNMENT OF SINDH  
HEALTH DEPARTMENT**

Karachi dated 15<sup>th</sup> December, 2017

To,

M/S National Agencies,  
House No.408 Batva Nagar,  
Liaqatabad, Karachi.  
0321-9220484

**SUBJECT:- PURCHASE OF VARIOUS MEDICAL INSTRUMENTS AND  
EQUIPMENTS FOR DR. RUTH K.M. PFAU CIVIL HOSPITAL,  
KARACHI.**

I am directed to inform you that as per the technical evaluation done by the Procurement committee your firm is declared as NOT QUALIFIED.

SECTION OFFICER-IV (H)

C.c. to:-

1. The PS to Secretary, Health Department, Karachi.
2. Office File

SECTION OFFICER-IV (H)



Civil Hospital  
Baba-e-Urdu Road, Karachi  
Dated: July, 2017

### CIVIL HOSPITAL KARACHI PROCUREMENT PLAN FOR THE FINANCIAL YEAR 2017-18

Sr. #	Description of procurement	Funds allocated	Source of fund	Proposed procurement method	Tentative / Timing of procurement				Remarks
					1 <sup>st</sup> Qtr	2 <sup>nd</sup> Qtr	3 <sup>rd</sup> Qtr	4 <sup>th</sup> Qtr	
1	Local Purchase of Drugs / Medicines (24/7 on Emergency Basis)	40.000 (M)	Non-ADP	Single Stage One Envelope	✓				Rule 46 (1)
2	Purchase of Medical Gases	48.825 (M)	Non-ADP	Single Stage One Envelope	✓				Rule 46 (1)
3	Purchase of Drugs / Medicines (Bulk Quantity)	89.600 (M)	Non-ADP	Single Stage Two Envelope	✓	✓	✓	✓	Rule 46 (2)
4	Janitorial Services	23.535 (M)	Non-ADP	Single Stage Two Envelope	✓	✓			Rule 46 (2)
5	Security Services	16.500 (M)	Non-ADP	Single Stage Two Envelope		✓	✓		Rule 46 (2)
6	Uniform & Protective Clothing	7.671 (M)	Non-ADP	Single Stage One Envelope	✓				Rule 46 (1)
7	i) Other Misc. General Store (Sundries)	6.983 (M)	Non-ADP	Single Stage One Envelope	✓	✓	✓	✓	Rule 46 (1)
	ii) Other Misc. General Store (Appliances)								
8	Expenditure on Diet of Patients: Diet-Enteric (Feeding and Food Supplements)	15.000 (M)	Non-ADP	Single Stage One Envelope	✓				Rule 46 (1)
9	Diet for Patients	100.000 (M)	Non-ADP	Single Stage One Envelope		✓	✓		Rule 46 (1)
10	Consumable (Lab items etc)	174.550 (M)	Non-ADP	Single Stage Two Envelope	✓				Rule 46 (2)

*Tauqeer*  
Medical Superintendent  
Civil Hospital, Karachi.

Sr. #	Description of procurement	Funds allocated	Source of fund	Proposed procurement method	Tentative / Timing of procurement				Remarks
					1 <sup>st</sup> Qtr	2 <sup>nd</sup> Qtr	3 <sup>rd</sup> Qtr	4 <sup>th</sup> Qtr	
11	Instruments	44,954 (M)	Non-ADP	Single Stage Two Envelope	✓	✓	✓	✓	Rule 46 (2)
12	Repair of Medical Equipments	22,486 (M)	Non-ADP	Single Stage One Envelope	✓	✓	✓	✓	Rule 46 (1)
13	Repair of Fixture & Furniture	0,953 (M)	Non-ADP	Single Stage One Envelope	✓	✓	✓	✓	- As and when repair work is required - Priority will be given to the Authorized agent of the Medical Equipments - Below One Lac through NIQ - Below One Million through NIT @ SPPRA Website - Above One Million through prominent Newspapers as per guideline of SPPRA
14	Repair of Transports	1,166 (M)	Non-ADP	Single Stage One Envelope	✓	✓	✓	✓	
15	Maintenance & Repair (Office Building)	167,200 (M)	Non-ADP	Single Stage One Envelope	✓	✓	✓	✓	Rule 46 (1)
15	Purchas of Machinery & Equipments	34,900 (M)	SNE (New)	Single Stage Two Envelope			✓	✓	Rule 46 (2)
15	Purchas of Furniture & Fixtures	2,480 (M)	SNE (New)	Single Stage Two Envelope			✓	✓	Rule 46 (2)

*Tawab*  
MEDICAL SUPERINTENDENT  
CIVIL HOSPITAL KARACHI

Medical Superintendent  
Civil Hospital, Karachi





HEALTH DEPARTMENT  
GOVERNMENT OF SINDH

**Certificate as Per Rule 48 of SPPRA Regarding Financial Rates Quoted By  
Bidder for Procurement of Various Medical Instruments & Equipments  
for Dr. Ruth K.M Pfau Civil Hospital, Karachi**

This is certified as per Rule 48 of SPPRA the rates quoted by the only technically qualified bidder M/S S. Ejazuddin & Co for item No. 04 &05 are comparable to the prices or rates of the last awarded contract & the market prices.

Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member

Syed Qasim Akber Nimal  
Section Officer (IV)  
Health Department  
Member

Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman



HEALTH DEPARTMENT  
GOVERNMENT OF SINDH

**Certificate as Per Rule 48 of SPPRA Regarding Financial Rates Quoted By  
Bidder for Procurement of Various Medical Instruments & Equipments  
for Dr. Ruth K.M Pfau Civil Hospital, Karachi**

This is certified as per Rule 48 of SPPRA the rates quoted by the only technically qualified bidder M/S Hospicare Systems for item No 01,02,03,08 &10 are comparable to the prices or rates of the last awarded contract & the market prices.

Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member

Syed Qasim Akber Nimal  
Section Officer (IV)  
Health Department  
Member

Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman



HEALTH DEPARTMENT  
GOVERNMENT OF SINDH

**Certificate as Per Rule 48 of SPPRA Regarding Financial Rates Quoted By  
Bidder for Procurement of Various Medical Instruments & Equipments  
for Dr. Ruth K.M Pfau Civil Hospital, Karachi**

This is certified as per Rule 48 of SPPRA the rates quoted by the only technically qualified bidder M/S Muslim Medical Services for item No. 06&09 are comparable to the prices or rates of the last awarded contract & the market prices.

Dr Shaheen Sharafat,  
Professor of Microbiology,  
Director, DOW Diagnostic & Research  
Laboratories DUHS, Karachi  
Member

Syed Qasim Akber Nimai  
Section Officer (IV)  
Health Department  
Member

Mr. Aslam Pechuho,  
Additional Secretary (Technical)  
Health Department Govt of Sindh,  
Chairman

ABC  
Certified

The Largest Circulated Sindhi Daily of Pakistan

DAILY  
KAWISH

دکٽري وقت ڪراچي، حيدرآباد ۽ سکر مان شايع ٿيندڙ پهرين سنڌي اخبار



# ڪاوش

روزانه

Tuesday 10 October, 2017

(جلد 28) اڱارو 10 آڪٽوبر 2017 ۽ بمطابق 19 محرم الحرام 1439ھ (شمارو 66) قيمت 15 روپيا

## گورنمينٽ آف سنڌ هيلٿ ڊپارٽمينٽ



IFB No:HD/PRO-eqp/2017

Dated: 5th October, 2017

### ٽينڊر نوٽيس

1. هيلٿ ڊپارٽمينٽ، حڪومت سنڌ کي ڊاڪٽريٽ ڪي، ايمر فائو سول اسپتال ڪراچي واسطي مختلف ميڊيڪل انسٽرومينٽس ۽ ايڪيورمينٽس جي لاڳت جي سلسلي ۾ پاڪ رينين ۾ پبلڪ فنڊن مان رقم وصول ٿي آهي. انهيءَ مخصوص ڪيل فنڊ جي هڪ حصي کي هيلٿ ڪاتي واسطي مختلف ميڊيڪل انسٽرومينٽس ۽ ايڪيورمينٽس لاءِ ٽيڪي تحت لعل ادا ڪيڻ تي خرچ ڪرڻ جو خواهشمند آهي.
2. هيلٿ ڊپارٽمينٽ هائي ڊاڪٽريٽ ڪي، ايمر فائو سول اسپتال ڪراچي واسطي مختلف ميڊيڪل انسٽرومينٽس ۽ ايڪيورمينٽس جي پروڪيورمينٽ واسطي سيلز ٽيڪس (SRB) ۽ انڪر ٽيڪس ڪاتي سان رجسٽرڊ ٿيل سٽي ساڪ وارن ۽ اصل ميٽوليڪچررز يا سندن مجاز ڊيلرن/ڊسٽريبيوٽرز کان ايس بي بي رول (2) 46 ”سنگل اسٽيج-ٽو اينويبل طريقيڪار“ سان مطابقت ۾ مهينو ڊاڪ گهرائي ٿو.
3. وضاحتون، شرطن ۽ ضابطن سميت ٻين تفصيلن تي ٻڌل واک ڪاغذ ”سيڪشن آفيسر (PM&I)، ٿرڊ فلور، تعلق هائوس، سنڌ سيڪريٽريٽ بلڊنگ نمبر 2، ڪمال انٽارڪ روڊ ڪراچي جي آفيس مان اشاعت جي تاريخ کان 2017-10-24 تائين سرڪاري ڪم ڪار واري وقت دوران 3000 رينين جي ٽينڊر تي ٻي آرڊر/ڊمانڊ جرافٽ جي صورت ۾ بحق سيڪريٽري، هيلٿ ڊپارٽمينٽ گورنمينٽ آف سنڌ ڪرڻ تي حاصل ڪري سگهجن ٿا.
4. واک ڏيندڙن کي سندن مهينو ٽينڊر ڪاغذ/واڪن (ٽيڪنيڪل ۽ فنانشل) الڳ لفافن ۾ 2017-10-25 تي ٽيپري 3:30 وڳي يا ان کان اڳ اڀريشنل سيڪريٽري (PM&I) هيلٿ ڊپارٽمينٽ، ٿرڊ فلور تعلق هائوس، سنڌ سيڪريٽريٽ بلڊنگ نمبر 2، ڪمال انٽارڪ روڊ ڪراچي، جي آفيس ۾ امانت لاءِ گهرائي آهن.
5. فقط ٽيڪنيڪل واک/پروپوزلس 2017-10-25 تي صابر 4:00 وڳي پروڪيورمينٽ ڪميٽي آڏو واک ڏيندڙن/فرمن جي سندن مجاز نمائندن جي موجودگيءَ ۾ مٿي ڄاڻايل آفيس ۾ کوليا ويندا.
6. سمورن دلچسپي رکندڙ واک ڏيندڙن لاءِ هڪ پري-بڊ ڪانفرنس/ميٽنگ 2017-10-16 تي 2:30 وڳي ايڊيشنل سيڪريٽري (PM&I) هيلٿ ڊپارٽمينٽ، ٿرڊ فلور تعلق هائوس، سنڌ سيڪريٽريٽ بلڊنگ نمبر 02، ڪمال انٽارڪ روڊ ڪراچي، جي آفيس ۾ منعقد ڪئي ويندي.
7. آگه پاڪ رينين ۾ ڄاڻايل گهرجن ۽ جي ايس ٽي، ايس آر بي (جتي لاڳو ڪرڻ جوڳو هجي) سميت انڪر ٽيڪس وغيره شامل هجي، جي ايس ٽي، مقرر ٿيل آگهن تي ايٽ سورس ڪالسيو ويندو. وڌيڪ تفصيل واک ڪاغذن ۾ شامل آهن.
8. واکن سان گڏ ڄاڻايل قيمت جو 3% واک سيڪيورٽي واک امانت واري تاريخ تي ٻي آرڊر/ڪال ڊيپازٽ جي صورت ۾ ”سيڪريٽري هيلٿ ڊپارٽمينٽ، گورنمينٽ آف سنڌ“ جي نالي ۾ امانت گهرجي، ايس بي بي آر تي روزانو 2010 (ترميم ٿيل 2017) موجب واک جي ڪارگر هٿن وارو ملو (90) ڏينهن آهي.
9. عام موڪل جي صورت ۾ واک ورنڊو ڪرڻ واري ڏينهن تي ٻڌل پروگرام موجب امانتيا ۽ کوليا ويندا.
10. واک ڪاغذ، ايس بي بي آر تي ويب سائيٽ ([www.spprasindh.gov.pk](http://www.spprasindh.gov.pk)) تان پڻ ڏانهن لوڊ ڪري سگهجن ٿا پر واک ڏيندڙ کي ٽيڪنيڪل واک سان ڄاڻايل طريقي ۾ ٽينڊر تي لازمي امانت گهرجي.
11. پروڪيورمينٽ ڪميٽي ايس بي بي آر تي روزانو 2010 (ترميم ٿيل 2017) جي واسطيدار تشريحات موجب ٽيڪي جي ڏيڻ/قبوليت کان اڳ ڪنهن به ٻي سببن واکن کي منظور يا رد ڪرڻ جو حق محفوظ رکي ٿي.

سيڪشن آفيسر (PM&I)  
هيلٿ ڊپارٽمينٽ، GoS

INF/KRY.No.4188/2017

SAY NO TO CORRUPTION

استان دستگري جي خلاف متحد آهيون

EXIT



سنڌ بر تعليم جي بهتري لاءِ علمي ۽ پنهنجن نفاذ لاءِ 8 3 9 8 تي ايس ايم ايس ڪريو

IFB NO: HD/PRO-eqp/2017/

## حکومت سندھ... محکمہ صحت

کراچی مورخہ 15 اکتوبر 2017

### ٹینڈر نوٹس

1۔ محکمہ صحت، حکومت سندھ کو ڈاکٹر روڈ کے ایم فائو سول ہسپتال کراچی کیلئے مختلف میڈیکل انسٹرومنٹس اور ایکویپمنٹ کیلئے پبلک ٹنڈر سے پاکستانی روپے میں مختص کردہ رقم موصول ہوئی ہے۔ اس مختص کردہ فنڈز کی پابندی کے تحت کچھ نئے کوکچر صحت کیلئے مختلف میڈیکل انسٹرومنٹس اور ایکویپمنٹ کیلئے کئے جانے والے کنٹریکٹ کے تحت کی جانے والی اہل ادا کنندگیوں کیلئے استعمال کرنے کا ارادہ ہے۔

2۔ محکمہ صحت سب SPP رول (2) 46 کے "ایک مرحلہ۔ دو لٹنڈر طریقہ کار" کی مطابقت میں سیکشن (SRB) اور ایگمنٹس ڈیپارٹمنٹ میں رجسٹرڈ معروف اور اورینٹل میڈیکل سٹورز یا ان کے مجاز کردہ ڈیلرز/ڈسٹری بیوٹرز سے ڈاکٹر روڈ کے ایم فائو سول ہسپتال کراچی کیلئے مختلف میڈیکل انسٹرومنٹس اور ایکویپمنٹس کی پروکیورمنٹ کیلئے سربراہ پیشکشیں طلب کرتا ہے۔

3۔ تصدیقات، شرائط و ضوابط بشمول دیگر تفصیلات پر مشتمل بڈنگ دستاویزات دفتر سیکشن آفیسر (PM&I)، تیسری منزل، تعلق ہاؤس، سندھ سیکریٹریٹ بلڈنگ نمبر 02، کمال اتارک روڈ کراچی سے تاریخ اشاعت پلا سے 24.10.2017 تک دوران اوقات کار سب سے 3000 روپے کی ٹینڈر فیس بشکل بے آرڈر ڈیمانڈ ڈرافٹ بنام "سیکریٹری، محکمہ صحت، حکومت سندھ" کے عوض حاصل کی جاسکتی ہیں۔

4۔ پیشکش دہندگان کو اپنی سربراہ ٹینڈر دستاویزات/پیشکشیں (ٹیکنیکل اور فنانس) جداگانہ لفافوں میں 25.10.2017 سے پہلے 3:30 بجے تک دفتر ایڈیشنل سیکریٹری، (PM&I)، محکمہ صحت، تیسری منزل، تعلق ہاؤس، سندھ سیکریٹریٹ بلڈنگ نمبر 02، کمال اتارک روڈ کراچی میں جمع کرا دینی جائیں۔

5۔ صرف ٹیکنیکل پیشکشیں/پروپوزلز 25.10.2017 سے پہلے 4:00 بجے پروکیورمنٹ کمیٹی کے سامنے پیشکش دہندگان/ان کے مجاز کردہ فرم کے نمائندوں کی موجودگی میں مذکورہ بالا دفتر میں کھولی جائیں گی۔

6۔ دلچسپی رکھنے والے تمام پیشکش دہندگان کیلئے پری بڈنگ/مینگ کا انعقاد 16.10.2017 سے پہلے 2:30 بجے دفتر ایڈیشنل سیکریٹری (PM&I)، محکمہ صحت، تیسری منزل، تعلق ہاؤس، سندھ سیکریٹریٹ بلڈنگ نمبر 02، کمال اتارک روڈ کراچی میں کیا جائیگا۔

7۔ نرخ پاکستانی روپے میں پیش کئے جائیں، جن میں SRB، GST (جہاں اطلاق ہوتا ہو)، ایگمنٹس وغیرہ شامل ہونے چاہئیں۔ GST مقررہ شرح سے بڑھ کر نہیں کیا جائیگا۔ مزید تفصیلات بڈنگ دستاویزات کے ساتھ منسلک ہیں۔

8۔ لازم ہے کہ پیشکش کی کل لاگت کی 3% بڈ سکورٹی، پیشکش جمع کرانے کی تاریخ پر پیشکش کے ساتھ منسلک کیلئے آرڈر/کال ڈیپازٹ بنام "سیکریٹری، محکمہ صحت، حکومت سندھ" جمع کرائی جائے۔ پیشکش موثر رہنے کی مدت SPPRA رولز 2010 (ترمیم شدہ 2017) کی رو سے (90) دن ہے۔

9۔ عام تعطیل ہونے کی صورت میں پیشکشیں دیے گئے شیڈول کے مطابق اگلے یوم کارپرجع کرائی جائیں گی۔

10۔ بڈنگ دستاویزات SPPRA ویب سائٹ (www.pprasindh.gov.pk) سے ڈاؤن لوڈ کی جاسکتی ہیں تاہم اس صورت میں پیشکش دہندہ پر مقررہ صورت میں ٹینڈر فیس جمع کرنا لازم ہے۔

11۔ پروکیورمنٹ کمیٹی کنٹریکٹ تفویض/قبول کے جانے سے قبل SPPRA رولز 2010 (ترمیم شدہ 2017) کے متعلقہ مندرجات سے مشروط کوئی یا تمام پیشکشیں مسترد کر دینے کا حق محفوظ رکھتی ہے۔

دستخط... سیکشن آفیسر (PM&I)  
محکمہ صحت، حکومت سندھ

INF-KRY-4188/17  
محکمہ صحت کراچی کے خلاف

SAY NO TO  
CORRUPTION

8398

# NATIONAL 5A



## GOVERNMENT OF SINDH HEALTH DEPARTMENT

### TENDER NOTICE

1. The Health Department, Government of Sindh has received an allocation from the Public Funds in Pak rupees towards the cost of Various Medical Instruments & Equipments for Dr. Ruth K.M Pfau Civil Hospital, Karachi. It is intended that part of the proceeds of this allocated fund will be applied to eligible payments under the contract for Various Medical Instruments & Equipments for Health Department.
2. The Health Department now invites sealed bids in accordance with SPP Rule 46(2) Single Stage - Two Envelope procedures from well reputed and original manufacturers or their authorized dealers / distributors registered with Sales Tax (SRB) and Income Tax Department for procurement of Various Medical Instruments and Equipments for Dr. Ruth K.M Pfau Civil Hospital Karachi.
3. The Bidding Document containing specification, terms and conditions including other details can be obtained from the office of "Section Officer (PM&I), 3rd Floor, Tughlaq House, Sindh Secretariat Building No. 02, Kamal Atta Turk Road, Karachi" from Date of Publication up to 24-10-2017 during official working hours, against the tender fee amount of Rs. 3000/- in the shape of pay order/Demand Draft in the name of Secretary, Health Department, Government of Sindh.
4. The bidders are required to submit their sealed tender documents / Bids (technical & financial) in separate envelopes on or before 23-10-2017 at 3:30 p.m. at the office of Additional Secretary (PM&I), Health Department, 3rd Floor, Tughlaq House, Sindh Secretariat Building No. 02, Kamal Atta Turk Road, Karachi.
5. Only Technical bids/proposals will be opened on 25-10-2017 at 4:00 p.m. before the procurement committee in the presence of bidders/their representative of the firms at the above mentioned office.
6. A pre-bid conference/meeting will be held for all interested bidders on 16-10-2017 at 2:30 p.m. at the office of Additional Secretary (PM&I), Health Department, 3rd Floor, Tughlaq House, Sindh Secretariat Building No. 02, Kamal Atta Turk Road, Karachi.
7. The rates shall be quoted in Pak rupees and should be inclusive of GST, SRB (where applicable), Income Tax etc. GST will be deducted at source at the prescribed rates. The further details are enclosed in the bidding documents.
8. The Bid Security of 3% of the quoted price must be submitted along with bids on bid submission date in the shape of pay order/Call deposit in the name of "Secretary, Health Department, Government of Sindh". The bid validity period is (90) Days as per SPPRA Rules 2010 (Amended 2017).
9. In case of public Holiday, the bids shall be submitted and opened as per given schedule on the next working day.
10. The bidding documents can also be download from SPPRA website ([www.spprasindh.gov.pk](http://www.spprasindh.gov.pk)) but the bidder must submit the tender fee in prescribed manner along with technical bid.
11. Procurement Committee reserves the right to accept or reject any or all bids prior to award/acceptance of contract as per relevant provisions of SPPRA Rules 2010 (Amended 2017).

Sd/-  
SECTION OFFICER (PM&I)  
HEALTH DEPARTMENT, GOS

Say No to Corruption  
INF/KRY-4188/17

8398



NO.HD(P&E) 3-2 (427)/2014  
GOVERNMENT OF SINDH  
HEALTH DEPARTMENT  
(Procurement, Monitoring and Inspection Cell)

NOTIFICATION

No. HD(P&E)3-2/(427)/2014: In supersession to this department's notification of even number dated: 10-04-2017 and in pursuance of Rule 31 of the Sindh Public Procurement Rules, 2010, the Govt. of Sindh, Health Department, re-constituted Complaint Redressal Committee (CRC), comprising of the following officers for scrutinizing the complaints of aggrieved bidders against tender invited by Health Institutions / Hospitals / Programs / Projects in Sindh.

01	Secretary Health, Govt. of Sindh.	Chairman
02	Representative from Accountant General Sindh	Member
03	Independent expert from relevant field concerning (to be nominated by the Head of Procuring Agency)	Member
04	Deputy Secretary (PM&I)	Member/ Secretary
05	Deputy Secretary (General)	Member

TORs

- To scrutinize the complaints from the aggrieved bidders and decide the cases strictly in accordance with SPP Rules 2010.

SECRETARY HEALTH

No. HD(P&E)3-2/(427)/2014:

Karachi, dated 31st May, 2017

C.C to:

1. The Director General Health Services Sindh, Hyderabad.
2. The District Health Officers (All)
3. The Medical Superintendents (All)
4. The P.S. to Chief Secretary Sindh, Karachi.
5. The Managing Director, Sindh Public Procurement Regulatory Authority, Karachi.
6. The Special Secretary/Adl. Secretary (Admn/Development/Public Health) Health Department.
7. The Chairman & all members of the Committee.
8. The P.S. to Secretary Health.

  
(NAVEED AHMED SOOMRO)  
SECTION OFFICER (PM&I)

**MINUTES OF PRE-BID MEETING REGARDING PURCHASE OF VARIOUS  
MEDICAL INSTRUMENTS AND EQUIPMENTS FOR DR. RUTH K.M. PFAU  
CIVIL HOSPITAL, KARACHI HELD ON 16<sup>TH</sup> OCTOBER, 2017 @ 02:30 P.M.  
IN THE PM&I WING, HEALTH DEPARTMENT, GOVT. OF SINDH**

(List of participant is enclosed herewith)


A Pre bid meeting for procurement of various medical instruments and equipments for Dr. Ruth K.M Pfauf Civil Hospital, Karachi was held on 16<sup>th</sup> October, 2017 @ 02:30 p.m.

The following company / vendor attended the meeting :-

**1. M/s S. Ejazuddin & Co.**

During the meeting no such major objection regarding the bid was raised by the bidder except that the representative of S. Ejazuddin & Co informed that in the procurement of goods no Sindh Revenue Board (SRB) registration is required and no sales tax on services is deducted from the goods which are procured. It is also very much clear in the rule 46(1)(a)(iii) of SPPRA Rules that registration with the Sindh Revenue Board (SRB) is applicable for procurement of works & services only. The committee agreed to it and deleted the SRB registration requirement from the mandatory clause of the bidding document.

The meeting ended with the vote of thanks to the above participant.

  
**Section Officer (PM&I)  
Health Department  
Govt. of Sindh**

C.c. to:-

**1. Participants.**



