

# **BID EVALUATION REPORT**

**SINDH GOVERNMENT HOSPITAL KORANIG-05,  
KARACHI**

**TENDER FOR PURCHASE OF MACHINERY  
EQUIPMENT, INSTRUMENT, MISC: ITEMS,  
UNDER ADP  
SCHEME NO. 394 FOR THE YEAR 2015-2016.**

**TENDER NOTICE**

The Medical Superintendent, Sindh Government Hospital, Korangi No 5, Karachi, hereby invites sealed bids from various manufacturers / sole agents / authorized distributors for the supply of item mentioned below as per bidding process given under Clause 46(2) single stage two envelope procedure of Sindh Public Procurement Rules 2010 (Amended 2013).

A complete set of tender enquiry may be purchased from office of the Medical Superintendent Sindh Govt Hospital Korangi 5, Karachi, and PM&I Cell, Health Department 3rd Floor Tughlaque House, Sindh Secretariat Karachi from April 26, 2016 on submission of written application upon cash payment of non-refundable fee mentioned below. The bids must be submitted on opening date upto 12:00 noon which will be opened publicly in presence of the bidders or their authorized representatives who choose to attend at 01:00 p.m. at the office of the Medical Superintendent Sindh Govt Hospital Korangi 5. All bids must be accompanied by a bid security @ 2.5% of the total bid cost in shape of pay order.

The purchaser reserves the right to accept / reject any / all bids under the relevant provision of SPPRA Rules-2010 (Amended 2013).

1	Description of Store	Machinery Equipment, Instrument, Misc for Sindh Government Hospital, Korangi No. 5, Karachi	Cost of Tender Rs. 2000/-
2	Description of Store	Furniture, Fixture and Linen for Sindh Government Hospital, Korangi No. 5, Karachi	Cost of Tender Rs. 1000/-
3	Description of Store	Machinery Equipment, Instrument, Misc for 03 Urban Health Centres at Karachi	Cost of Tender Rs. 2000/-
4	Description of Store	Furniture, Fixture and Linen for 03 Urban Health Centres at Karachi	Cost of Tender Rs. 1000/-
5	Date of Sale of Tender Document	26th April, 2016	
6	Last Date of Sale of Tender	12-05-2016	
7	Date of Submission of Tender	12-05-2016 upto 12:00 Noon	
8	Date of Opening	12-05-2016 at 1:00 pm	
9	Tender Opening Venue	Seminar Hall Sindh Government Hospital, Korangi No. 5, Karachi	

**N.B:**

1. In case Government announces any public holiday then the tenders will be opened on next working day
2. All quotations shall include Govt. taxes including GST, if applicable
3. Information regarding this NIT may also be downloaded from SPPRA website: [www.pprasindh.gov.pk](http://www.pprasindh.gov.pk) & Sindh Govt. website [www.sindh.gov.pk](http://www.sindh.gov.pk)

INF-KRY No. 1685/16

ہم شہری کے خلاف تہمتیں

Say No to Corruption

MEDICAL SUPERINTENDENT  
SINDH GOVT. HOSPITAL, KORANGI-5, KARACHI**CORRIGENDUM**

This is with reference to the Tender Notice published in daily Dawn, Express dated 25th April, 2016 pertaining to the:

1. Machinery Equipment, Instrument, Misc for Sindh Government Hospital, Korangi No. 5, Karachi.
  2. Furniture, Fixture and Linen for Sindh Government Hospital, Korangi No. 5, Karachi.
  3. Machinery Equipment, Instrument, Misc. for 03 Urban Health Centre at Karachi
  4. Furniture, Fixture and Linen for 03 Urban Health Centre at Karachi.
- This is to notify that the venue for tender opening is at E.O.C Hall, I.I. Depot Cantt, Karachi.

MEDICAL SUPERINTENDENT  
SINDH GOVT. HOSPITAL KORANGI, KARACHI

INF-KRY No. 1967/16

Say No to Corruption

ہم شہری کے خلاف تہمتیں



SPPRA

## Sindh Public Procurement Regulatory Authority Government of Sindh

[Home](#) [SPPRA Notifications](#) [SPPRA Board](#) [SPP Act 2009](#) [SPP Rules 2010](#) [Amendments](#) [SPPRA Team](#) [Contact Us](#)

[Tenders](#)

[Evaluation Reports](#)

[Black Listed Firms](#)

[Committee](#)

[Sample Documents](#)

[Policy Instructions](#)

[Review Panel](#)

[Capacity Building](#)

[Downloads](#)

[Gallery](#)

[Notifications](#)

[Procurement Plans](#)

[Site Statistics](#)

[Login](#)

Search Tenders Archive By City Names click here

- ☒ Serial No.
- ☐ City (e.g sukkur)
- ☐ SPPRA ID
- ☐ Tender Title
- ☐ Advertise Date
- ☐ Month/Year

Advance Search

submit

1

S. No .	SPPRA Tender ID	Tender Title	Download	Upload On	Advertise Date	Closing Date	Corrigendum /SBD/ Notices	Remarks	Evaluation Reports
27853	2147483847SPPRA ID WITHHELD	Sindh Government Hospital KORANGI KARACHI NIT WITH SBD		28-04-2016	28.Apr.2016	12.May.2016	13 May 2016	BER, CS, TER, AOC Required	





**OFFICE OF THE MEDICAL SUPERINTENDENT  
SINDH GOVT. HOSPITAL, KORANGI-5, KARACHI  
BID EVALUATION REPORT**


- |   |  |
|---|--|
| 1. Name of Procuring Agency             | : Medical Superintendent, Sindh Govt. Hospital, Korangi-5, Karachi.  |
| 2. Tender Reference No                  | : SGHK/2612/16   |
| 3. Tender Description/Name of work/item | : <u>TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS, UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.</u> |
| 4. Method of Procurement                | : <u>Single Stage Two Envelope bidding procedure</u>   |
| 5. Tender Published                     | : Daily DAWN and Daily Jang<br>Electronic Media SPPRA ID No. 27853   |
| 6. Total Bid documents Sold             | : 35 nos.  |
| 7. Total Bids Received                  | : 24 nos.  |
| 8. Technical Bid Opening date           | : 12-5-2016  |
| 9. No. of Bid technically qualified     | : 22 nos.  |
| 10. Bid(s) Rejected                     | : 2 no.  |
| 11. Financial Bid Opening date          | : 20-5-2016  |
| 12. Bid Evaluation Report:              |  |


S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
COMPARATIVE STATEMENT ATTACHED						

The Committee has recommended to award the contract to the successful bidders


  
**Additional Deputy Commissioner-I  
District Korangi Karachi**  
Representative of Deputy Commissioner  
Korangi - Karachi (Member)

  
**Asstt. Accountant General**  
Representative of Accountant General Sindh  
Karachi (Member)

  
Dr. Shakoor Abbasi  
DHO-Malir, Karachi (Member)

  
Dr. Tahir Aziz  
DHO-(West) -Karachi (Member)  
**DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST**

**DR. ABDUL SHAKOOR ABBASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.**

  
Dr. Muhammad Tofique  
Medical Superintendent/Chairman  
Sindh Govt. Hospital-Korangi- Karachi  
**Dr. Muhammad Tofique  
Medical Superintendent  
SINDH GOVT. HOSPITAL  
Korangi No. 5, Karachi**



## MINUTES OF MEETING OF TECHNICAL BID OPENING

MINUTES OF TENDER OPENING MEETING REGARDING TENDER NO. SGHK/-2612/16 FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS FOR SINDH GOVT. HOSPITAL KORANGI-5, KARACHI UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016 HELD ON MAY 12TH, 2016.

Tender opening meeting started with the recitation of the Holy Quran at I.I Depot, Rafique Shaheed Road -Karachi. Medical Superintendent/Chairman purchase committee chaired the meeting for the purchase of machinery equipment, instrument, Misc. Items to be installed at Sindh Govt. Hospital, Korangi – 5, Karachi. Chairman procurement Committee briefed all the participants regarding the method and schedule of procurement. The Chairman Procurement committee has also briefed that NIT was published in the leading Newspapers mentioned below:


- i. Daily DAWN
- ii. Daily Jang

As the method of procurement was single stage two envelop procedure therefore first technical proposals of all companies/firms/bidders were opened in the presence of Technical Committee and members of procurement committee. The attendance sheet of the participants and bidders are also attached herewith (photocopy).

However, the submitted bids were shown to all, no late bid was entertained. Attendance of the bidders was sought. Total Thirty Five (35) tender documents were sold and Twenty Four (24) bids were participated. The documents were handed over to Technical Committee for further evaluation.

The Chairman and all the members of Technical Committee signed the bidding documents of all the bidders. The financial proposal were kept in sealed box in lock and key before the bidders and all the participants. The bidding documents were further proceeded for the Technical Evaluation before the respective Technical Committee. The meeting ended with thanks.


The list of participants is attached herewith.

  
\_\_\_\_\_  
Medical Superintendent  
Sindh Govt. Hospital, Korangi-5-Karachi  
(Chairman)  
**Dr. Muhammad Tofique**  
**Medical Superintendent**  
**SINDH GOVT. HOSPITAL**  
**Korangi No. 5, Karachi**



**LIST OF PARTICIPANTS**

01	Medical Superintendent, Sindh Govt. Hospital, Korangi-5-Karachi (Chairman)
02	Prof. Tariq Mahmood, Head of the Department of Radiology, JPMC-Karachi (Member)
03	Dr. Umer Daraz, Chief Pathologist, CHK(Member)
04	Dr. Robina Bashir, Add. Medical Superintendent, SGH- Korangi – 5 (Member)
05	Dr. Riazuddin Shaikh, Cardiologist , SGH- Korangi – 5 (Member)
06	Dr. Abdul Hameed Shaikh, Pathologist , SGH- Korangi – 5 (Member)
07	Dr. Abdul Razzaq Siddiqui , Nephrologist , SGH- Korangi – 5 (Member)
08	Dr. M. Murtaza Raza, Radiologist , SGH- Korangi – 5 (Member)
09	Dr. Abdul Jabbar, Sr. Dental Surgeon , SGH- Korangi – 5 (Member)
10	Dr. Shehneela , Pediatrician (Member)

  
Medical Superintendent  
Sindh Govt. Hospital, Korangi-5-Karachi  
(Chairman)  
**Dr. Muhammad Tofique**  
Medical Superintendent  
SINDH GOVT. HOSPITAL  
Korangi No. 5, Karachi



## ATTENDENT SHEET FOR TECHNICAL BIDS OF FINANCIAL YEAR 2015-16 FOR PURCHASING OF

Machinery Equipment, Instrument, Misc for Sindh Government Hospital, Korangi No. 5, Karachi

DUE ON 12.05.2016

SR. NO.	NAME OF FIRM	SIGNATURE	CONTACT NO.
✓ 1	Universal Enterprises	[Signature]	0333-0374410
✓ 2	BIOS	[Signature]	0333-3831988
✓ 3	MEDURG (Pvt) Ltd	[Signature]	0300-2842124
✓ 4	Standard Supplier	[Signature]	0300-2566277
✓ 5	Medcare Pvt Ltd	[Signature]	0300-2120607
✓ 6	T.K Medical Instruments	[Signature]	0300-9203101
✓ 7	Hospicare System	[Signature]	0300-550116
✓ 8	Sindh Medical Stores <del>M. H. Qureshi</del>	[Signature]	03012956227
✓ 9	Hospital Solutions	[Signature]	03332235157
✓ 10	Elite C- (Pvt) Ltd	[Signature]	0311-0209044
✓ 11	Fresenius Medical Care	[Signature]	0305-4441521 0305-4440562
✓ 12	Lucky Linn International	[Signature]	0314-2015730 0345-2628269
✓ 13	Hoora Pharmacy	[Signature]	0300-3331507
✓ 14	Zeal Pharmaceutical	[Signature]	0300-2544007
✓ 15	ZAM TRADING Co	[Signature]	03329236312
✓ 16	Hospital Supply Corporation	[Signature]	0921-2629715
✓ 17	GULF MARKET. International M. REHAN Qureshi	[Signature]	0333-3960655
✓ 18	LAB Medical Solutions	[Signature]	03355377002
✓ 19	M/S Popular International	[Signature]	0321-8755107
✓ 20	M/S Medicomst Co	[Signature]	0336-3745522



21.	M/S Seetajudhara	Khuram	03432940365
22.	M/S H & Enterprise	Zuber	0311-1076235
23.	R.Z. Services.	Haji	0345-2717832
24.	Bashon Pharma	(Signature)	03000300098
(25.)	Bio Sciences International	(Signature)	0300-3331507
(26.)	Radiant Medical	(Signature)	03132055651
27.	<hr/>		
28.			
29.			
30.			
31.			
32.			
33.			
34.			
35.			
36.			
37.			
38.			
39.			
40.			
41.			
42.			
43.			























12/5/16

12/5/16

Takrif 12/5/16  
(DHO Lahore)



DUE ON 20.05.2016

SR. NO.	NAME OF FIRM	SIGNATURE	CONTACT NO.
01	Standard Suppliers		0300-256277
02	Mediquips. SMC		0333-3899841
03	Hospital Supply Corporation		0321-2629715
04	Hospicare System		0300530116
05	Lab Medikal Solutions		0335-5370002
06	Medi Medlitec <sup>Co</sup> <del>Medi</del> Medlitec		0336-5765322
07	Fresenius Medical		0305-4441521 0305-4441100
08	Lucky Lab <sup>Lab</sup> <del>Lab</del> <sup>Lab</sup>		0314-2015700 0395-2628217
09	Hospital Solution		0321-2092500
10	Zeal Pharmaceutical <sup>Services</sup>		0300-2544007
11	HOBAN Pharma.		0300-2287365
12	Bios		0333-3836988
13	M/s Popular <sup>Pvt. Ltd</sup> international		0321-8755104
14	Quip Medical <sup>Co</sup> <del>Co</del> <sup>Co</sup>		0333-2766695
15	C.M.T.C. (C) PVT LTD		0300-3380117
16	ZAM TRADING Co		0332-9236312
17	R.Z. Services		0345-2717831
18	Boshon Pharma <sup>Co</sup> <del>Co</del> <sup>Co</sup>		030030090
19	MEDI URGE (Pvt) Ltd		0300-2842124
20	S. E. Fajuddin <sup>Co</sup> <del>Co</del> <sup>Co</sup>		03432940365



21


21	UNIVERSAL BOOK PAPER
----	----------------------

[illegible]

0333, 0322, 141

03008246132





## **FINANCIAL COMPARATIVE STATEMENT**

**SINDH GOVERNMENT HOSPITAL , KORANIG-05, KARACHI**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT,**

**MISC: ITEMS, UNDERADP**

**SCHEME NO. 394 FOR THE YEAR 2015-2016.**

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

S.No.	Name of Equipment	Name of technically Responsive bidder	Unit Price on US\$/Euro/GBP and others	Exchange rate	Price Per Unit in Pak Rupees.	Remarks
1.	Emergency resuscitation trolley	Bashson Pharma	1,100Euro.	117.06	128,766Pkr	M/s. ZAM Trading approved as 1 <sup>st</sup> lowest
		ZAM Trading	1,129US\$.	104.50	117,980Pkr	
2.	6 Ch ECG Machine	Medequips	2,399US\$.	104.50	250,695Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		Zeal Pharm. Se.	1,960US\$.	104.50	204,820 Pkr	
		BIOS	1,119Euro.	117.06	130,990Pkr	
		Bashon Pharma	1,390US\$.	104.50	145,255 Pkr	
3.	Multi Para monitor	TK Inst.	1,149Euro.	117.06	134,502Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		M/s. Medequip	6,428US\$.	104.50	671,726Pkr	
		M/s. BIOS	2,730US\$.	104.50	285,285Pkr	

DR. TAHIR AZIZ  
DISTRICT HEALTH OFFICER  
KARACHI WEST

DR. ABDUL SHAKOOR ABBASI  
DISTRICT HEALTH OFFICER  
KARACHI WEST

DR. TAHIR AZIZ  
DISTRICT HEALTH OFFICER  
KARACHI WEST



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

4.	Defibrillator	Medequips Option:	5,335US\$, 750US\$, 1,200US\$, 7,285US\$,	104.50	557,507Pkr 78,375Pkr 125,400Pkr 761,282Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		Total:				
		Elate Cc Options:	3,195Euro. 800Euro. 800Euro. 4,795Euro.	117.06	374,007Pkr 93,648Pkr 93,648Pkr 561,303Pkr	
		Total:				
5.	Suction machine heavy duty	BIOS Option:	3,690Euro. 450Euro. 545Euro. 4,685Euro.	117.06	431,951Pkr 52,677Pkr 63,798Pkr 548,426Pkr	M/s. TK Instrument approved as 1 <sup>st</sup> lowest
		Universal Enterprise	1,216.80Euro.	117.06	142,439Pkr	
		BIOS	1,050Euro.	117.06	122,913Pkr	
		Zeal Pharm. Se.	566Euro.	117.06	66,256Pkr	
		Standard Suppliers	980GBP 695Euro.	152.52 117.06	149,470Pkr 81,357Pkr	

*Copy*  
**DR. TAHIR AZIZ SHAIKH**  
DISTRICT HEALTH OFFICER  
KARACHI WEST

**DR. ABDUL SHAKOOR ABBASI**  
DISTRICT HEALTH OFFICER  
KORANGI-5, KARACHI

**Dr. Muhammad Tofigue**  
Medical Superintendent  
SINDH GOVT. HOSPITAL,  
Korangi No. 5, Karachi

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR

THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

6.	Central Patient Monitor	Medequip	4,800US\$.	104.50	501,600Pkr	As per tender, Central Monitor was required compatible with item no. 3, from same Manufacturer.  Since offer of M/s. BIOS declared 1 <sup>st</sup> lowest in item no. 3, hence offer of M/s. BIOS is approved for this item.
		Zeal Pharm. Se.	3265US\$.	104.50	341,192Pkr	
		BIOS	5,250US\$.	104.50	548,625Pkr	
7.	Infusion pump	Universal Enterprise	128,500Pkr	-	128,500Pkr	M/s. TK Med. approved as 1 <sup>st</sup> lowest
		Mediurge	1215Euro.	117.06	142,228Pkr	
		Hospicare	1225US\$.	104.50	128,012Pkr	
		TK Medical..	848Euro.	117.06	99,267Pkr	
8.	Portable lights	M/s. BIOS	110,000Pkr	-	110,000Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		Zeal Pharm. Se.	6,530US\$.	104.50	682,385Pkr	

*Chhigir*  
**DR. TAHIR AZIZ SHANIKH**  
DISTRICT HEALTH OFFICER  
KARACHI WEST

**DR. ABDUL SHAKOOR ADEAST**  
DISTRICT HEALTH OFFICER  
DISTRICT HEALTH OFFICER

**DR. Muhammad Fajir**  
Medical Superintendent  
SINDH GOVT. HOSPITAL  
Korangi No. 6, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

9.	OT Table (Local)	HAQ Enterprise	99,786Pkr	-		99,786Pkr	M/s. RZ service approved as 1 <sup>st</sup> lowest
		RZ Service	95,000Pkr	-		95,000Pkr	
10.	Ventilator	Hospicare	18,000US\$.	104.50		1,881,000Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		BIOS	14,544US\$.	104.50		1,519,848Pkr	
		Elate Cc	13,300Euro.	117.06		1,556,898Pkr	
11.	Portable blood pressure monitoring apparatus with pediatric cuff						No bid received hence item may be retender.
12.	Baby Weighing scale	Zeal Pharm. Se.	40US\$.	104.50		4,180Pkr	M/s. Zeal Pharma approved as 1 <sup>st</sup> lowest
13.	Glucometer	RZ Service	1,830Pkr	-		1,830Pkr	M/s. Popular International approved as 1 <sup>st</sup> lowest
		Popular International	1,080Pkr			1,080Pkr	

*[Signature]*  
DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*[Signature]*  
DR. ABDUL SHAKOOR AHMAD  
DISTRICT HEALTH OFFICER  
DISTRICT MAJLIE KARACHI.

*[Signature]*  
Dr. Muhammad Tofigue  
Medical Superintendent  
SINDH GOVT. HOSPITAL,  
Korangi No. 5, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR

THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

14.	Baby incubator	Zeal Pharm. Se.	4,898US\$.	104.50	511,841Pkr	M/s. Bashson pharma approved as 1 <sup>st</sup> lowest
		Bashson Pharma	4,000US\$.	104.50	418,000Pkr	
		Mediurge	6,660Euro.	117.06	779,620Pkr	
		TK Inst.	5,380US\$.	104.50	562,210Pkr	
15.	Phototherapy unit	Zeal Pharm. Se.	1,176US\$.	104.50	122,892Pkr	M/s. TK Medical approved as 1 <sup>st</sup> lowest
		Mediurge	2,590Euro.	117.06	303,185Pkr	
		TK Medical	564US\$.	104.50	58,938Pkr	
16.	Nebulizer	Zeal Pharm. Se.	33US\$.	104.50	3,448Pkr	M/s. Zeal Pharma approved as 1 <sup>st</sup> lowest
17.	Infant ambu bag					No bid received for this item
18.	Laryngoscope pediatric/infant	HAQ Enterprise	19,786Pkr	-	19,786Pkr	M/s. HAQ Enterprise approved as 1 <sup>st</sup> lowest

*DR. TAHIR AZIZ SHAIKH*  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*DR. ABDUL SHAKOOR ABRASI*  
DISTRICT HEALTH OFFICER  
KARACHI

*Dr. Muhammad Tofigue*  
Medical Superintendent  
SINDH GOVT HOSPITAL  
Korangi No. 5, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

19.	Over head radiant warmer	ZAM Trading	1,965US\$.	104.50	205,342Pkr	M/s. ZAM Trading approved as 1 <sup>st</sup> lowest
		TK Inst.	2,040US\$.	104.50	213,180Pkr	
20.	EEG Machine	Medequips	23,000US\$.	104.50	2,403,500Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		BIOS	15,555Euro.	117.06	1,820,868Pkr	
21.	Evoked Potential EMG Machine	Medequips	25,900US\$.	104.50	2,706,550Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		BIOS	17,335Euro.	117.06	2,029,235Pkr	
22.	Bronchofiberscope with complete set of accessories	Bashson Pharma	1,588,000Pkr	-	1,588,000Pkr	M/s. Bashson Pharma approved as 1 <sup>st</sup> lowest

*[Signature]*  
DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*[Signature]*  
DR. ABDUL SHAKOOR ABBASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

23.	Spirometer	Zeal Pharm. Se.	1,960US\$.	104.50	204,820Pkr	M/s. Zeal Pharma approved as 1 <sup>st</sup> lowest
		BIOS	485,000Pkr	-	485,000Pkr	
		ZAM Trading	4,047US\$.	104.50	422,911Pkr	
24.	Pulse oximeter	Zeal Pharm. Se.	810US\$.	104.50	84,645Pkr	M/s. Zeal Pharma approved as 1 <sup>st</sup> lowest
		BIOS	119,000Pkr	-	119,000Pkr	
25.	Abraham punch biopsy needles					No bid received for this item
26.	Dialysis machine	Universal Enterprise	8,920US\$.	104.50	932,140Pkr	M/s. Universal Enterprise approved as 1 <sup>st</sup> lowest
		Presenius Medical	8,949US\$.	104.50	935,170Pkr	

*Dr. Tahir Aziz Shaikh*  
DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*Dr. Abdul Shakoor Abbasi*  
DR. ABDUL SHAKOOR ABBASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

*Dr. Muhammad Tofigue*  
Dr. Muhammad Tofigue  
Medical Superintendent  
SINDH GOVT HOSPITAL  
Korangi No. 5, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

27.	RO Plant with complete fitting for 10 nos. Dialysis machine	Universal Enterprise	645,000Pkr.	-	645,000Pkr	M/s. Universal Enterprise approved as 1 <sup>st</sup> lowest
		Fresenius Medical	41,500US\$.	104.50	4,336,750Pkr	
28.	Dialyzer Re-processing system	Universal Enterprise	8,875US\$.	104.50	927,437Pkr	M/s. Universal Enterprise approved as 1 <sup>st</sup> lowest
		Fresenius Medical	10,500US\$.	104.50	1,097,250Pkr	
29.	500mA X-ray unit	Medequips	43,975US\$.	104.50	4,595,387Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		BIOS	28,400Euro.	117.06	3,324,504Pkr	
30.	300mA X-ray unit for OPD	Medequips	34,000US\$.	104.50	3,553,000Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
		BIOS	26,900Euro.	117.06	3,148,914Pkr	
31.	X-Ray Auto processor	Mediurge	10,280Euro.	117.06	1,203,377Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
			599,000Pkr	-	599,000Pkr	

*Calvin*  
DR. TAHIR AZIZ SHAKKI  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*Shafiq*  
DR. SHAFIQ SHAFIQ ARBASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

*Dr. Muhammad Tofigue*  
Dr. Muhammad Tofigue  
Medical Superintendent  
SINDH GOVT. HOSPITAL  
Korangi No. 5, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

32.	Digital color Ultrasound machine	Medequips	35,500US\$.	104.50	3,709,750Pkr	M/s. Medequips approved as 1 <sup>st</sup> lowest
33.	Digital Black & White Ultrasound system	ZAM Trading	10,080US\$.	104.50	1,053,360Pkr	M/s. ZAM Trading approved as 1 <sup>st</sup> lowest
34.	100mA Mobile Radiography unit	BIOS	14,990Euro.	117.06	1,754,729Pkr	M/s. BIOS approved as 1 <sup>st</sup> lowest
35.	Video endoscope complete system (both upper & lower GI)	Hospital Solution	9,138,000Pkr	-	9,138,000Pkr	Quoted rate is very high from the approved rate of PC-1, hence may be retender.
36.	Chemistry Analyzer	ZAM Trading	9,127US\$.	104.50	953,771Pkr	M/s. Gulf Marketing approved as 1 <sup>st</sup> lowest
		Sindh Medical Store	3,496US\$.	104.50	365,332Pkr	
		Hooria Pharma	4,500Euro.	117.06	526,770Pkr	
		Lab Medikal Solution	3,750US\$.	104.50	391,875Pkr	
		Gulf Marketing	250,000Pkr	-	250,000Pkr	

*DR. TAHIR AZIZ SHAIKH*  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*DR. ABDUL SHAKOOR ABBASI*  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*Dr. Muhammad Tufique*  
Medical Superintendent  
Korangi Hospital




# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

		Popular International	2940US\$.	104.50	307,230Pkr	
		Lucky Lab Int.	354,000Pkr	-	354,000Pkr	
37.	Electrolyte Analyzer	Bashon Pharma	3,263US\$.	104.50	340,983Pkr	M/s. Popular International approved as 1 <sup>st</sup> lowest
		ZAM Trading	4,683US\$.	104.50	489,373Pkr	
		Sindh Medical Store	3,074US\$.	104.50	321,233Pkr	
		Lab Medikal Solution	4,430US\$.	104.50	462,935Pkr	
		Gulf Marketing	250,000Pkr	-	250,000Pkr	
		Popular International	2205.6US\$.	104.50	230,485Pkr	
38.	Incubator for lab	ZAM Trading	2538US\$.	104.50	265,221Pkr	M/s. Popular International approved as 1 <sup>st</sup> lowest
		Popular International	892.8US\$.	104.50	93,297Pkr	

Signature:   
AZIZ SHAIKH  
OFFICER

Signature: 

*DR. TAHIR AZIZ SHAIKH*  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*DR. ABDUL SHAKOOR ABBASI*  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

*Dr. Mahmood Tofigue*  
Medical Superintendent  
SINDH GOVT HOSPITAL,  
Korangi No. 6, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

39.	Hot air oven	ZAM Trading	2,224US\$.	104.50	232,408Pkr	M/s. ZAM Trading approved as 1 <sup>st</sup> lowest
40.	Centrifuge machine (24 heads)	ZAM Trading	2,738US\$.	104.50	286,121Pkr	M/s. ZAM Trading approved as 1 <sup>st</sup> lowest
		Popular International	2,834.4US\$.	104.50	296,195Pkr	
41.	Ultrasonic water bath	Popular International	585.25US\$.	104.50	61,158Pkr	M/s. Popular International approved as 1 <sup>st</sup> lowest
42.	Electronic balance for lab	ZAM Trading	1,825US\$.	104.50	190,712Pkr	M/s. ZAM Trading approved as 1 <sup>st</sup> lowest
43.	Sterilizer autoclave vertical type					No Bid received hence may be retendered.

SEMAI HIGWAY  
DISTRICT HEALTH OFFICER  
KORANGI  
KARACHI WEST

TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

DR. ABDUL SYAFQUR ARNASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

44.	Binocular Microscope	Popular International	960US\$.	104.50	100,320Pkr	M/s. Popular International approved as 1 <sup>st</sup> lowest
45.	Juster (fixed)	ZAM Trading	127US\$.	104.50	132,271Pkr	M/s. Popular International approved as 1 <sup>st</sup> lowest
		Lab Medikal Solution	90US\$.	104.50	9,405Pkr	
		Popular International	47.4US\$.	104.50	4,953Pkr	
46.	Refrigerator rotaray					No bid received for this item
47.	Dental Chair	Bashson Pharma	385,000Pkr	-	385,000Pkr	M/s. Bashson Pharma approved as 1 <sup>st</sup> lowest

MR. AZIZ SHAIKH  
HEALTH OFFICER

*Tahir Aziz*  
DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

*Abdul Shakoor Abbasi*  
DR. ABDUL SHAKOOR ABBASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

48.	Ultrasonic Scaler	Bashson Pharma	55,000Pkr	-		55,000Pkr	M/s. Bashson Pharma approved as 1 <sup>st</sup> lowest
49.	Autoclave dental						No bid received for this item
50.	Dry head sterilizer dental						No bid received for this item
51.	Dental instrument sets	RZ Service	5,900Pkr	-		5,900Pkr	M/s. RZ Service approved as 1 <sup>st</sup> lowest
		HAQ Enterprise	12,786Pkr	-		12,786Pkr	
52.	Flow meter	Standard Suppliers	5,500Pkr	-		5,500Pkr	
		ZAM Trading	80US\$.	104.50		8360Pkr	M/s. Standard Suppliers approved as 1 <sup>st</sup> lowest
		HAQ Enterprise	12,786Pkr	-		12,786Pkr	

DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

DR. ABDUL SHAKOOR ARNASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

Page 13 of 17



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

53.	Stabilizer 5 KVA		No bid received for this item																
54.	Oxygen cylinder 240 liters	<table> <tr> <td>ZAM Trading</td><td>33,350Pkr</td><td>-</td><td>33,350Pkr</td></tr> <tr> <td>HAQ Enterprise</td><td>29,786Pkr</td><td>-</td><td>29,786Pkr</td></tr> <tr> <td>RZ Service</td><td>34,250Pkr</td><td>-</td><td>34,250Pkr</td></tr> <tr> <td>TK Inst.</td><td>46,000Pkr</td><td>-</td><td>46,000Pkr</td></tr> </table>	ZAM Trading	33,350Pkr	-	33,350Pkr	HAQ Enterprise	29,786Pkr	-	29,786Pkr	RZ Service	34,250Pkr	-	34,250Pkr	TK Inst.	46,000Pkr	-	46,000Pkr	M/s. HAQ Enterprise approved as 1 <sup>st</sup> lowest
ZAM Trading	33,350Pkr	-	33,350Pkr																
HAQ Enterprise	29,786Pkr	-	29,786Pkr																
RZ Service	34,250Pkr	-	34,250Pkr																
TK Inst.	46,000Pkr	-	46,000Pkr																
55.	Oxygen cylinder 48/c liter	<table> <tr> <td>ZAM Trading</td><td>23,350Pkr</td><td>-</td><td>23,350Pkr</td></tr> <tr> <td>HAQ Enterprise</td><td>17,786Pkr</td><td>-</td><td>17,786Pkr</td></tr> <tr> <td>RZ Service</td><td>17,150Pkr</td><td>-</td><td>17,150Pkr</td></tr> <tr> <td>TK Inst.</td><td>31,000Pkr</td><td>-</td><td>31,000Pkr</td></tr> </table>	ZAM Trading	23,350Pkr	-	23,350Pkr	HAQ Enterprise	17,786Pkr	-	17,786Pkr	RZ Service	17,150Pkr	-	17,150Pkr	TK Inst.	31,000Pkr	-	31,000Pkr	M/s. RZ Service approved as 1 <sup>st</sup> lowest
ZAM Trading	23,350Pkr	-	23,350Pkr																
HAQ Enterprise	17,786Pkr	-	17,786Pkr																
RZ Service	17,150Pkr	-	17,150Pkr																
TK Inst.	31,000Pkr	-	31,000Pkr																

*Copy*  
DR. TAHIR AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

DR. ABDUL SHANOOZ ABRASI  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.


Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

56.	Computer with monitor , laser printer and other complete accessories	HAQ Enterprise	139,786Pkr	-		139,786Pkr	M/s. HAQ Enterprise approved as 1 <sup>st</sup> lowest
57.	Refrigerator	HAQ Enterprise	45,786Pkr	-		45,786Pkr	M/s. HAQ Enterprise approved as 1 <sup>st</sup> lowest
		TK Inst.	65,000Pkr	-		65,000Pkr	
58.	IT Net working						No bid received for this item
59.	Cold storage facility	Standard suppliers	18,000Euro.	117.06		2,107,080Pkr	M/s. Standard Suppliers approved as 1 <sup>st</sup> lowest
60.	Color TV						No bid received for this item

  
DR. TAHIRA AZIZ SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

  
DR. ARSHAD SHAHOOD ARSHADI  
DISTRICT HEALTH OFFICER  
DISTRICT HALL KARACHI.

  
DR. MUHAMMAD TAHIR  
Medical Superintendent  
SINDH GOVT. HOSPITAL,  
Korangi No. 5, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR

THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

61.	A/c 1 ½ tons	HAQ Enterprise	83,786Pkr	-		M/s. TK Medical approved as 1 <sup>st</sup> lowest
		TK Medical	62,500Pkr	-		
62.	A/c Split 2 tons	TK Medical	103,000Pkr	-		M/s. TK Medical approved as 1 <sup>st</sup> lowest
63.	Photocopier					No bid received for this item
64.	Fax Machine					No bid received for this item
65.	Multi Media					No bid received for this item
66.	Over Head Project with screen set	Bashson Pharma	105,700Pkr	-		M/s. Bashson Pharma approved as 1 <sup>st</sup> lowest

DR. GHANIM RAZI SHAIKH  
DISTRICT HEALTH OFFICER  
KARACHI WEST

DR. ABUL SHAKOOR ABBASI  
DISTRICT HEALTH OFFICER  
DISTRICT MAIN KARACHI.

DR. JAWAD TAYYUB  
DISTRICT HEALTH OFFICER  
KORANGI-5, KARACHI

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Financial bid opening : May 20<sup>th</sup>, 2016

## COMPARATIVE STATEMENT

67	Central oxygen supply system	Universal Enterprise	127,500GBP.	152.52	19,446,300PKr	M/s. Mediurge approved as 1 <sup>st</sup> lowest
		ZAM Trading	119,240US\$.	104.50	12,460,580PKr	
		Standafrd suppliers	10,416,500PKr	-	10,416,500PKr	
		Mediurge	75,175Euro.	117.06	8,799,985PKr	

Representative of Deputy Commissioner-1  
Korangi - Karachi (Member)

**District Korangi Karachi**

Dr. Shakoor Abbasi  
DHO-Malir, Karachi (Member)

**DR. ABDUL SHAKOOR ABBASI**  
DISTRICT HEALTH OFFICER  
DISTRICT MALIR KARACHI.

Representative of Accountant General Sindh,  
Karachi (Member)

**Accountant General Sindh**  
Office of the  
Karachi

Dr. Tahir Aziz  
DHO-(West) -Karachi (Member)

**DR. TAHIR AZIZ SHAIKH**  
DISTRICT HEALTH OFFICER  
KARACHI WEST

Dr. Muhammad Tofigue  
Medical Superintendent/Chairman  
Sindh Govt. Hospital-Korangi- Karachi

**Dr. Muhammad Tofigue**  
Medical Superintendent

**SINDH GOVT. HOSPITAL**  
Korangi No. 5, Karachi





# **TECHNICAL EVALUATION**

**SINDH GOVERNMENT HOSPITAL , KORANIG-05, KARACHI**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT,**

**MISC: ITEMS, UNDER ADP**

**SCHEME NO. 394 FOR THE YEAR 2015-2016.**

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	01				
Name of Items	Emergency resuscitation trolley				
Quantity	18 Nos.				
Emergency resuscitation trolley					
Approved in Bid Evaluation	Zeal Pharma	Mediarge	Bashson Pharma	ZAM Trading	
Model	Yes	No.	Yes	Yes	
Make	MX37EMG	R-8002	PCL	238	
	Baldia	Hersil	Franch Hospital	GIMA	
Light weight aluminium or polypropylene structure with 4 castors and 5 or more drawers.	Country of origin is not as per tender specification	Not approved in Bid evaluation.	4 drawers	5 drawers	
Clear and smooth design for easy cleaning			Yes	Yes	
Complete with all accessories such as Cardiac board.			Yes	Yes	
IV Pole		Attached Agency letter is not valid.	Yes	Yes	
Oxygen regulator			Yes	Yes	
Flow meter with humidifier			Yes	Yes	
Top tray			Yes	Yes	
Waste collection bin.			Yes	Yes	
Resuscitator kit for adult and peds.			Yes	Yes	
USA/UK/Europe/Japan	Taiwan		Italy	Italy	
Remarks	Non Responsive	Non responsive	Responsive	Responsive	



SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

Item No	02					
Name of Items	6 Ch ECG Machine					
Quantity	06 Nos.					
<div>✓</div> <div>✓</div> <div>✓</div> <div>✓</div>						
6 Ch ECG Machine	Medequips	Zeal Pharm. Se.	BIOS	Bashon Pharma	TK Inst.	Hospicare
Approved in Bid Evaluation	Yes	Yes	Yes	Yes	Yes	Yes
Model	ECG 1250K	FT-1	MT plus	KES 601T	Ascard Grey	ECG100S
Make	Nihon Kohden	Schillar	BTL	Kalamed	Aspel	Cardioline
Resting 6 Chanel ECG Machine with alphanumeric key board, LCD display of 5.5" or more. Pace maker detection.	5.7"	5.7"	5.7"	5.7"	5.5"	4.3"
Indicators for loose leads	Yes	Yes	Yes	Yes	Yes	Yes
Auto and manual modes.	Yes	Yes	Yes	Yes	Yes	Yes
Printer Paper speed (cm/s) : 1, 2.5, 5.	Yes	Yes	Yes	Yes	Yes	Yes
Sensitivity (cm/mV): 0.25, 0.5, 1.2	Yes	Yes	Yes	Yes	Yes	Yes
Diagnostic analysis and interpretation as standard	Yes	Yes	Yes	Yes	Yes	Not mentioned
Memory : 100 or more records, 10 sec long strip	Yes	Yes	Yes	Yes	Yes	Yes
Paper type: roll/A-4	Yes	Yes	Yes	Yes	Yes	Yes
Battery backup for 75 minutes for monitoring and 30 min for printing in both automatic and manual mode.	Yes	Yes	Yes	Yes	Yes	Yes
Optional PC connection software	Not quoted	Not quoted	Yes	Yes	Yes	Yes
Complete with: Set of electrodes, patient cable, power cord, gel, paper roll 58mm / 25m and dust cover.	Yes	No	Yes	Yes	Yes	Yes
USA/UK/ Europe/Japan	Japan	Switzerland	Bulgaria	Germany	Poland	Italy
Remarks	Responsive	Responsive	Responsive	Responsive	Responsiv <sup>e</sup>	Non responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	03				
Name of Items	Multi Para monitor				
Quantity	24 Nos.				
<b>Multi Para monitor</b>		<b>Medequips</b>	<b>Zeal Pharm. Se.</b>	<b>Mediurge</b>	<b>BIOS</b>
Approved in Bid Evaluation	Yes	Yes	Yes	No.	Yes
Model	BSM3763K	Truescope –classic	Storm 5800	OMNI III	
Make	Nihon Kohden	Schiller	Dixon	Infinitum Medical	
15" or more color TFT touch screen	15"	Yes	The attached brochure do not have Manufacturers' address & other details in order to confirm country of origin.	15"	Yes
Wave form display up to 8 or more	Yes	Yes		Yes	Yes
ECG 5 leads, NIBP, SPO2 & TEMP.	Yes	Yes		Yes	Yes
Trend recording up to 64 hours or more	72 hours	96 hour		72 hours	Yes
HR Display Range: 30-300bpm or more	Yes	10-300bpm		Yes	Yes
Hard wire/wires connectivity facility for central monitoring station.	Yes	Yes		Yes	Yes
Network connectivity HL7 based to transfer data to the hospital information system.	Yes	Not available		Yes	Yes
Battery backup up to 3 hours or more.	Yes	No, 120 min		Yes	Yes
Complete with set of ECG electrodes, NIBP cuffs (Adult, Pediatric & Neonate) SPO2 probes 2 sets. (Adult, Pediatric, Neonate) Temp probe, local trolley with basket	Yes	Not mentioned	User list of quoted items not attached. HL 7	Yes	Yes
USA/UK/Europe/Japan	Japan	Switzerland	connectivity not available	USA	
Remarks	Responsive	Non Responsive	Non Responsive	Responsive	



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	04				
Name of Items	Defibrillator				
Quantity	02 Nos.				
Defibrillator					
Approved in Bid Evaluation	Yes	Medequips	Elate Cc	BIOS	Zeal Pharm. Se.
Model	Tec-5621K	Yes	Yes	Yes	Yes
Make	Nilton Koh.	Primedie	Osatu	Defiguard 4000	Schilliar
Defibrillator with monitor & recorder	Yes	Yes	Yes	Yes	Yes
Biphasic & truncated wave form	Yes	Yes	Yes	Yes	Yes
5" or more screen.	6.5"	5.7"	5.5"	5.9"	5.9"
Synchronized output with ECG	Yes	Yes	Yes	Yes	Yes
Control of energy charges / delivering on main panel and apex paddle.	Yes	Yes	Yes	Yes	Not mentioned
Charging indicators	Yes	Yes	Yes	Yes	Yes
Manual and AED mode energy up to 200J or more.	270J	360J	200J	AED not quoted	Not mentioned
Charging time for full energy should be less than 6 sec.	Yes	Yes	Yes	Yes	Yes
Display of heart rate, ECG through paddles and lead I, II & III patient cables.	Yes	Yes	Yes	Yes	Yes
Built-in recorder for printing of full summary of standard 50mm paper.	Yes	Yes	Yes	Yes	Yes
Alarm for high and low heart rate, low battery warning.	Yes	Yes	Yes	Yes	Yes
Built in rechargeable battery with charges for more than 100 shocks	Yes	Yes	Yes	Yes	Yes
Auto test/self-check	Yes	Yes	Yes	Yes	Yes
External pediatric and adult paddles.	Yes	Yes	Yes	Yes	Not

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

ECG cable	Yes	Yes	Yes	mentioned
AC 220V/50Hz operated	Yes	Yes	Yes	Yes
SpO2 and external Pace maker should be quoted as option.	Yes	Yes	Yes	Not quoted
USA/UK/Europe/Japan	Japan	Germany	Spain	Switzerland
Remarks	Responsive	Responsive	Responsive	Non responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	05					
Name of Items	Suction machine heavy duty					
Quantity	12 Nos.					
✓						
Suction machine heavy duty	Universal Enterprise	BIOS	Zeal Pharm. Se.	Standard Suppliers	Mediurge	TK Inst.
Approved in Bid Evaluation	Yes	Yes	Yes	Yes	No.	Yes.
Model	Novela	H-40	Maxispeed 6.4p	SAM 35	Eurovac H-40	Hospive 350
Make	Uzumcu	Hersil	3A	MG Medical	Hersil	CAMI
Noise less suction machine with a noise level of not more than 47db for smooth and calm operation.	Yes	Yes	Yes	Yes	Not approved in Bid evaluation.	Yes
Should have 4 extra strong wheels (1 of them antistatic and 2 of them with brakes)	Yes	Yes	Yes	Yes		Yes
Should have suction capacity of 40l/min	Yes	Yes	Yes	Yes		Yes
Vacuum pressure 650 -700 mmhg	Yes	Yes	Yes	Yes		Yes
ON/OFF switch	Yes	Yes	Yes	Yes		Yes
Should have Knob for Suction Control	Yes	Yes	Yes	Yes	Attached Agency letter is not valid.	Yes
Should be piston type, oil free , maintenance free	Yes	Yes	Yes	Yes		Yes
Should have 2 Polycarbonate collection jar of 4 liters with safety valve to prevent overfilling	Yes	Yes	Yes	Yes		Yes
Should be Complete With	Yes	Yes	Yes	Yes		Yes
Suction tube of 3-4 m	Yes	Yes	Yes	Yes		Yes
2 No Collection jar of 4 liters	Yes	Yes	Yes	Yes		Yes
With foot switch	Yes	Yes	Yes	Yes		Yes

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Power cord	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Anti-Bacterial filter	Yes	Yes	Yes	Yes	Yes	Yes	Yes	Yes
USA/UK/Europe/Japan	Turkey	Spain	Italy	UK	Spain	Italy	Italy	Italy
Remarks	Responsive	Responsive	Responsive	Responsive	Responsive	Non Responsive	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	06
Name of Items	Central Patient Monitor
Quantity	1 No.

Central Patient Monitor	Medequip	Zeal Pharm. Se.	BIOS
Approved in Bid Evaluation	Yes	Yes	Yes
Model	CNS-9101K	Schiller	OMNI View
Make	Nihon. Koh.	Schiller	Infinium Med.
Must be compatible with item no 3 " Multiparameter Monitor" and also from the same Manufacturer.	Yes	Yes	Yes
Multi-Channel 21-Inches colored Central Monitor with recording facility On Laser Printer	21"	Not mentioned	21"
Digital HIR with high and low alarms on all parameters selected on bedside monitors.	Yes	Yes	Yes
Other parameters to be monitored as selected in the bedside monitors with automatic recording on alarm	Yes	Yes	Yes
should have facility for 2 <sup>nd</sup> displays	Yes	No	Yes
monitor which can operate separately different information to monitor many patient data.	Yes	Yes	Yes
Hard wire system with capabilities of Tele Monitoring as a future upgrade.	Yes	Yes	Yes
Capability to store 72-hours trend data and	Yes	Yes	Yes
up to 250 arrhythmia episodes of 8-seconds each	Yes	Yes	Yes
and up to 4000 ST events for each patient	Yes	Yes	Yes
Capability to store 72-hours	Yes	Yes	Yes
full disclosure ECG as well as 5 other waveforms for each patient.	Yes	Yes	Yes
Built-in automatic Arrhythmia analysis and ST analysis	Yes	Yes	Yes
USA/UK/Europe/Japan	Japan	Switzerland	USA
Remarks	Responsive	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	07								
Name of Items	Infusion pump								
Quantity	20 Nos.								
Infusion pump		Zeal Pharm. Se.	Universal Enterprise	Baston Pharma	Mediurge	Hospicare	TK Inst.		
Approved in Bid Evaluation		Yes	Yes	Yes	Yes	Yes	Yes	Yes	
Model		ST-1000	TE-112	DF 12M	Infusia VP7	Infusomat P	CA-2000		
Make		Smatronic	Terumo	Moog	Fresenius Kabi	B Braun	Canafusion		
Volumetric peristaltic infusion pump.		Country of origin is not as per tender specification	Yes	Yes	Yes	Yes	Yes	Yes	
Flow rate 0.1 to 99.9 ml/h.			Yes	Yes	Yes	Yes	Yes	Yes	
Unit compatible with most IV set.			Yes	Yes	Yes	Yes	Yes	Yes	
Memory function incase of power failure to retain the most recent flow rate & total volume.			Yes	Yes	Yes	Yes	Yes	Yes	
Comprehensive alarm for Air in line, occlusion, door open, low battery, infusion complete, empty container.			Yes	Yes	Yes	Yes	Yes	Yes	
Built in battery charger.			Yes	Yes	Yes	Yes	Yes	Yes	
Unit complete with IV sets.			Yes	Yes	Yes	Yes	Yes	Yes	
USA/UK/Europe/Japan		Brazil	Japan	Luthania	Germany	Germany	Canada		
Remarks		Non Responsive	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive	



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	08
Name of Items	Portable lights
Quantity	02 Nos.

Portable lights	Zeal Pharm.	BIOS
Approved in Bid Evaluation	Se. Yes	Yes
Model	Lgnis 60	E-100
Make	Famed	Dentis
LED light. 60,000 LUX intensity	Yes	Yes
color rendering Ra > 95	Yes	Yes
4500K Color temperatures	Yes	Yes
No of LEDs: 8 nos or more.	Yes	Yes
Mobile stand with flexible arm and four castors with from same Manufacturer.	Yes	Yes
<b>FDA approved</b>	Yes	Yes
Korea/Taiwan/Malaysia	Poland	Korea
Remarks	Responsive	Responsive

# **SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.**

**394 FOR THE YEAR 2015-2016.**

**Tender Opened on May 11<sup>th</sup>, 2016**

## **TECHNICAL EVALUATION**

Item No	09
Name of Items	OT Table (Local
Quantity	02 Nos.

OT Table (Local	HAQ Enterprise	RZ Service
Good quality local	Yes	Yes
Remarks	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	10
Name of Items	Ventilator
Quantity	04 Nos.

Ventilator	Elate Ce	Hospicare	Mediurge	BIOS
Approved in Bid Evaluation	Yes <i>Date of issue and validity is not mentioned in the attached Agency certificate.</i>	Yes	No. <i>User list not attached.</i>	Yes
Model	4000ICU	PB 840	EXTEND	evolution 3e
Make	Siare	Purrete & Bennet	Air liquid	E-vent
Adult, Peads and Neonate	Yes	Yes	Yes	Yes
Touch screen LCD 12" or more display with turbine	12" TFT	Yes	Not turbine type	12.1" LCD
ACMV, SIMV, Spontaneous Ventilation	Yes	Yes	Spontaneous mode is not mentioned	Yes
Volume based : V-CMV; V-SIMV	Yes	Yes	Yes	Yes
Spontaneous based: CPAP + Pressure support, CPAP+ Volume support	Yes	Yes	Yes	Yes
Volume controlled pressure target ventilation.	Not mentioned	Yes	Not mentioned	Yes
Tidal volume : 5 - 2000 ml	2-3000ml	Yes	20-2000ml	5-200ml



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## **TECHNICAL EVALUATION**

Pcontrol3 - 70 cmH2O	Not mentioned	Yes	Not mentioned	Yes
Psupport 0 - 80 cmH2O	2-80cmH2O	Yes	Yes	Yes
Peak flow (mandatory) up to 115lpm	Not mentioned	Yes	Yes	Yes
I-Times 0.1 - 10.0 sec	9.6sec	Yes	Yes	Yes
Oxygen concentration 21 - 100%	Yes	Yes	Yes	Yes
Breath Types	Yes	Yes	Yes	Yes
Breath Triggering	Not mentioned	Yes	Yes	Yes
Flow triggering 0.1 - 20lpm	0.3-15lpm	Yes	Not mentioned	Yes
Monitoring display	Yes	Yes	Yes	Yes
Delivered oxygen concentration 21-100%	Yes	Yes	Yes	Yes
Trend Data (70 hours or more)	Yes	Yes	Yes	Yes
Pressure: Ppeak, Pmean, PEEP/CPAP, Pplateau,	CPAP Not available	Yes	Yes	Yes
Flow: Inspiratory flow & Expiratory flow	Yes	Yes	Yes	Yes
Minute volume	Yes	Yes	Yes	Yes
Spontaneous tidal volume	Not mentioned	Yes	Yes	Yes
Time: Inspiratory & Expiratory time	Yes	Yes	Yes	Yes
Lung mech: Cstat, Auto PEEP, RSB1, PMax	Not mentioned	Yes	Not mentioned	Yes
POI/PiMax	Not mentioned	Yes	Not mentioned	Yes
Loops : flow & pressure volume loop, pressure & time loop , flow & time loop	Pressure & time loop not mentioned	Yes	Yes	Yes
Alarm: complete set of alarm for gas, pressure, volume, power cut off, patient disconnected etc.	Yes	Yes	Yes	Yes
Battery backup of 2 hours or more.	Yes	Yes	Not	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

				mentioned	Yes
Built in trolley from same Manufacturer	Yes	Yes	Yes	Yes	Yes
Patient circuit support arm	Yes	Yes	Yes	Yes	Yes
10 x disposable Patient circuit	Yes	Yes	Yes	Yes	Yes
USA/UK/Europe/Japan	Italy	USA	France	USA	USA
Remarks	Non responsive	Responsive	Non Responsive	Responsive	

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	11	No bid received for this item
Name of Items	Portable blood pressure monitoring apparatus with pediatric cuff	
Quantity	04 Nos.	

Item No	12
Name of Items	Baby Weighing scale
Quantity	1 No.

✓

Baby Weighing scale	Zeal Pharm. Se.
Imported	Yes
Remarks	Responsive



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Item No	13
Name of Items	Glucometer
Quantity	02 Nos.

Glucometer	RZ. Service	Popular International
Imported	Yes	Yes
Remarks	Responsive	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 11<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	14
Name of Items	Baby incubator
Quantity	04 Nos.

Baby incubator	Zeal Pharm. Se.	Bashon Pharma	Mediurge	TK Inst.
Approved in Bid Evaluation	Yes	Yes	Yes	Yes
Model	BT-500	Line 4	Cristina Self Control	BB-3000
Make	Bistos	Olidelf	Cobarns	Dison
Intensive care incubator with air and skin temperature control.	Yes	Yes	Yes	Yes
Control module with micro processed servo controlled system .	Yes	Yes	Yes	Yes
Skin and air temperature probes.	Yes	Yes	Yes	Yes
Seven alarms and foldable front door.	Yes	Yes	Yes	Yes
Double wall and passive humidity control .	Yes	Yes	Yes	Yes
Configurable to support and mount a wide range of accessories .	Yes	Yes	Yes	Yes
Tilt mechanism (+/- 100) Cabinet stand with central column with two side double drawers , lateral protection, four wheels and two brakes.	Yes	Yes	Yes	Yes
Unit should complete with four automatic ports , two lateral iris ports and six ports for tubing, sliding mattress tray with trendelenburg and fowler positing x-rays transparent.	Yes	Yes	Yes	Yes
Taiwan/China /Malaysia	USA	Brazil	Italy	China
Remarks	Responsive	Responsive	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	15
Name of Items	Phototherapy unit
Quantity	02 Nos.

✓ ✓ ✓

Phototherapy unit	Zeal Pharm. Se.	Mediurge	TK Inst.
Approved in bid Evaluation	Yes	Yes	Yes
Model	BT-400	05A	BL-50
Make	Bistos	Cobarns	Dison
Input power : ≤ 200VA . Light source: blue light tube .	Yes	Yes	Yes
Tube quantity: 5 . Light source effective life : 2000hours .	Yes	Yes	Yes
Wave length: 425nm~475nm .	Yes	Yes	Yes
Light radiancy : ≤ 550uW/cm².	Yes	Yes	Yes
Time display range: 999h 59min.	Yes	Yes	Yes
Display time accuracy : 2 h +/- 1 min . Adjust angle 30° , 60° , 90°	Yes	Yes	Yes
Taiwan/China/Malaysia	USA	Italy	China
Remarks	Responsive	Responsive	Responsive

22.

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	16
Name of Items	Nebulizer
Quantity	10 Nos.

Nebulizer	Zeal Pharm. Sc.
Approved in Bid Evaluation	Yes
Model	Automizer
Make	3A
Compressor: Piston.Fuse: T1,6 A/250V.	Yes
Operating Pressure: 0.90 bar. Air flow: 14l/min.	Yes
Noise Level: 50db. Temporary Use: 30/30 min.	Yes
ON/OFF. Class of Risk: IIa.	Yes
Accessories: Adult face mask 01.	Yes
Pediatric Face Mask. Faster jet Bulb.	Yes
Mouthpiece. Nasal prong. Connection tubing.	Yes
USA/UK/Europe/Japan	Italy
Remarks	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 11<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	17	No bid received for this item
Name of Items	Infant ambu bag	
Quantity	02 Nos.	

Item No	18
Name of Items	Laryngoscope pediatric/infant
Quantity	02 Nos.

Laryngoscope pediatric/infant	HAQ Enterprise
Imported	Yes
Remarks	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

Item No	19
Name of Items	Over head radiant warmer
Quantity	02 Nos.

Over head radiant warmer	ZAM Trading	TK Inst.
Approved in Bid Evaluation	Yes	Yes
Model	Not mentioned	BN-100
Make	Beijing Jolonsangyo	Dison
Operational condition : +18°C ~ + 30°C.	Yes	Yes
Environment temperature: 30% ~ 75%.	Yes	Yes
Environment relative humidity : <0.3m/s.	Yes	Yes
Temperature control mode : Pre-heat, auto and manual .	Yes	Yes
Temperature control range: 25°C ~ 37°C.	Yes	Yes
Temperature override mode to : 38°C.	Yes	Yes
Temperature sensor accuracy: <=0.3°C.	Yes	Yes
Mattress temperature infirmity : <=2.0°C.	Yes	Yes
Radiant box horizon angle adjust range : 0~+/-90° step-less .	Yes	Yes
Mattress tray size : 68cm x 56cm .	Yes	Yes
Mattress tray max load : 10 kg.	Yes	Yes
Transfusion pole max load : 2 kg.	Yes	Yes
Shelf max load: 2 kg. R-232 connector .	Yes	Yes
Mattress to radiant box : 81 cm .	Yes	Yes
Power failure alarm .	Yes	Yes
Sensor failure alarm .	Yes	Yes
Taiwan/China/Malaysia	China	China
Remarks	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 17<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	20
Name of Items	EEG Machine
Quantity	1 No.

EEG Machine	Medequips	BIOS
Approved in Bid Evaluation	Yes	Yes
Model	EEG 1200K	Neurowerk EEG 36
Make	Nihon Kohden	Sigma
32 Channel EEG, digital amplifier	32 ch	36 Ch.
Should have at-least 26 EEG channel including referential (EEG) &4+2 polygraphicchannel	Yes	Yes
Optical isolated network cable for PC connection.	Yes	Yes
24 bit A/D converter.	Yes	Yes
32 channel Amplifier with LAN connection.	Yes	Yes
Photic stimulator	Yes	Yes
EEG recording and review software	Yes	Yes
Isolation transformer 400VA or more	Not mentioned	Yes
Desktop PC station with power adaptor, CPU Intel core, 4 GB RAM, 500GB hard disk, DVDR, Windows 7 Professional. Display 22" or more.	Yes	Yes

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 11<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Device software should be preinstalled.	Yes	Yes
original trolley from same Manufacturer	Yes	Yes
complete with following accessories:	Yes	Yes
set of EEG Electrodes for full channel	Yes	Yes
Electrode cables	Yes	Yes
EEG head cap	Yes	Yes
Wrist and ear electrode	Yes	Yes
Laser printer	Yes	Yes
USA/UK/Europe/Japan	Japan	Germany
Remarks	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	21		
Name of Items	Evoked Potential EMG Machine		
Quantity	1 No.		
Evoked Potential EMG Machine		Medequips	BIOS
Approved in bid evaluation		Yes	Yes
Model		MEB 9400K	Neurowerk EMG 2 Ch
Make		Nihon Kohden	Sigma
2 Channel EMG System		Yes	Yes
EMG/NCV/EP		Yes	Yes
Should have Capability to Measure & Analysis of EMG (Spontaneous activity, voluntary and maximum activity		Yes	Yes
Should have Motor and sensory nerve conduction velocity, mixed nerve studies, F wave, H reflex, Blink reflex, Repetitive stimulation		Yes	Yes
Should have Middle and late latency		Yes	Yes
Sympathetic skin response		Yes	Yes
Should have Preconfigured SEP programs for median, tibial and trigeminal nerves		Yes	Yes
ADC should be 24 bit or more		Yes	Yes
Software should have neuro-anatomic guide for electrode placement		Yes	Yes
Should have capability to average with artifact suppression		Yes	Yes
Should have facility of Split screen for on line, averaged and stored signals, presentation of results		Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Should have defined report forms	Yes	Yes
User defined number of measuring traces	Yes	Yes
Automatic marking of pre-definable markers	Yes	Yes
Sound presentation of the EMG	Yes	Yes
Should have facility to export report into PDF or any other computer read format	Yes	Yes
Personal computer: Intel Pentium 2 GHz or better	Yes	Yes
2 GB RAM or better	Yes	Yes
500 GB HD or more	Yes	Yes
Network Card/Modem should be builtin	Yes	Yes
CD/RW or DVD/RW	Yes	Yes
Sampling rate should not be less than 50 kHz/channel	Yes	Yes
Audio output (for additional multimedia speakers)	Yes	Yes
Should include LAN Ethernet for EMG processor adapter	Yes	Yes
LAN Ethernet for network adapter & PC	Yes	Yes
Should have 2x external Trigger In/Out	Yes	Yes
Input impedance > 110 MOhm or better	Yes	Yes
Lower cutoff frequency 0.02 Hz - 1000 Hz or better	Yes	Yes
Should be complete with	Yes	Yes
23" or more TFT Monitor	Yes	Yes
Keyboard, mouse	Yes	Yes
Current stimulator	Yes	Yes
Foot switch	Yes	Yes
Operating System should be Windows 7 or Windows 8	Yes	Yes
Original Trolley from same manufacturer with integrated isolation transformer	Yes	Yes
Head box stand	Yes	Yes



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.**

**394 FOR THE YEAR 2015-2016.**

**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Compatible laser printer	Yes	Yes
(USA/UK/Europe/Japan)	Japan	Germany
Remarks	Responsive	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

Item No	22
Name of Items	Bronchofiberscope with complete set of accessories
Quantity	1 No.

Bronchofiberscope with complete set of accessories	Bashson Pharma
Approved in bid evaluation	Yes
Model	Karl Storz Adult
Make	Karl Storz
Field of view - in excess of 110 degree.	Yes
depth of field - from 3 mm – 50 mm.	Yes
distal end outer diameter – less than 5 mm.	Yes
insertion tube outer diameter – maximum 5 mm	Yes
working length – 600 mm	Yes
channel inner diameter - 2.2 mm or more.	Yes
minimum visible distance from end – 5mm.	Yes
angulations achieved – up – 180 degree, down – 130 degree.	Yes
total length – 900 mm.	Yes
compatible light source, suction pump, leak tester and trolley in which entire assembly may be mounted. bronchoscope and accessories should be easily disinfected and sterilized with commonly available methods	Yes
USA/UK/Europe/Japan	Germany
Remarks	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	23
Name of Items	Spirometer
Quantity	1 Nos.

Spirometer	Zeal Pharm. Se.	BIOS	ZAM Trading
Approved in bid evaluation	Yes	Yes	Yes
Model	SP-1	Datospir 120D	Vitalograph
Make	Schiller	Sibelmed	Vitalograph
TURBINE type transducer.	Yes	Yes	Yes
FVC, VC, MW tests in PRE and POST bronchodilatation mode.	Yes	Yes	Yes
Graphics superposition.	Yes	Yes	Yes
Check-up programme.	Yes	Yes	Yes
Graphic incentive for children.	Yes	Yes	Yes
Configuration programme.	Yes	Yes	Yes
Internal data base for more than 150 tests with graphics.	Yes	Yes	Yes
112 mm. (4 in.) internal printer.	Yes	Yes	Yes
Data transfer via INTERNET with PC of Patients tests.	Yes	Yes	Yes
Device check-up data.	Yes	Yes	Yes
Software update.	Yes	Yes	Yes
Tests exportation to other management systems.	Yes	Yes	Yes
Calibration programme for quality control or test, according to standards	Yes	Yes	Yes
USA/UK/Europe/Japan	Switzerland	Spain	UK
Remarks	Responsive	Responsive	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	24
Name of Items	Pulse oximeter
Quantity	02 Nos.

Pulse oximeter	Zeal Pharm. Se.	BIOS
Approved in bid evaluation	Yes	Yes
Model	Argus OXM plus	CLBO
Make	Schiller	Infinium Medical
Portable, mounted on trolley.	Yes	Yes
5: or more" touch screen TFT display	Yes	Yes
2 trace waveform	Yes	Yes
SpO2% Range: 0 ~ 100%	Yes	Yes
Pulse Rate Range: 30-250 bpm	Yes	Yes
Alarm	Yes	Yes
Red light LED wavelength SpO2 Probe	Yes	Yes
Complete with leads and infant accessories	Yes	Yes
USA/UK/Europe/Japan	Switzerland	USA
Remarks	Responsive	Responsive



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 11<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Item No	25	No bid received for this item
Name of Items	Abrahm punch biopsy needles	
Quantity	02 Nos.	

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION



Item No	26
Name of Items	Dialysis machine
Quantity	10 Nos.

Dialysis machine	Universal Enterprise	Hospital Supply Cor.	Fresenius Medical
Approved in bid evaluation	Yes	Yes	Yes
Model	Surdial 55	TQS-88	4008S
Make	NIPRO	Toray	Fresenius
Complete with battery backup	Yes	Yes	Yes
The machine saves Operation History, Alarm History, Dialysis Record and Parts.	Yes	Yes	Yes
Life Records automatically and necessary data can be shown on LCD.	Yes	Yes	Yes
Dialysis Machine Capable of Acetate Dialysis, Bicarbonate Dialysis and Isolated Ultra-filtration.	Yes	Yes	Yes
Easy to see 5.7 inch LCD Touch Panel and easy to handle control panel.	Yes	Yes	Yes
Dialysis treatment parameters such as UF Goal, UF Rate, Blood Flow and Syringe Flow are easily settable on the control panel.	Yes	Yes	Yes
The machine saves Operation History, Alarm History, Dialysis Record and Parts Life Records automatically with LCD display.	Yes	Yes	Yes
Profile	Yes	Yes	Yes
UF Rate and Dialysis Conductivity are selectable from 4 Patterns	Yes	Yes	Yes
Selectable Concentrate	Yes	Yes	Yes
3 Types of concentrate are pre-settable	Yes	Yes	Yes
Function to Lighten weight of waste	Yes	Yes	Yes
After Dialysis Treatment,	Yes	Yes	Yes
The Saline water will be removed from inside of blood line and dialyzer to the Drain water line.	Yes	Yes	Yes
Battery Backup	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

Automatic Power Backup System	Yes	Yes	Yes
Safety Design	Yes	Yes	Yes
Double Chamber System to precise UF Volumetric control	Yes	Yes	Yes
3 CPU System for Control, Driving and Monitoring Functions	Yes	Yes	Yes
3 Sensors for Syringe Check for Syringe fitting and speed	Yes	Yes	Yes
The machine performs self check of the closed circuit	Yes	Yes	Yes
The machine monitors reverse rotation of blood pump for preventing an accident.	Yes	Yes	Yes
Supply Water Flow 700 ml/min	Yes	Yes	Yes
Supply Water Temperature 10 – 25 °C	Yes	Yes	Yes
Relative Humidity 30-85 % Rh (Except dew stat)	Yes	Yes	Yes
Atmospheric pressure 0.070-0.106Mpa	Yes	Yes	Yes
Dialysis Flow 400-600 ml / min	Yes	Yes	Yes
UF rate 0.00 – 3.00 L / h	Yes	Yes	Yes
UF volume 0.00 – 40.0L, Dialysate type Bicarbonate / Acetate	Yes	Yes	Yes
Dialysis Conductivity 12.0 – 16.0mS/cm	Yes	Yes	Yes
Dialysate step 1 – 6 ( 0.3Ms/ cm per one step )	Yes	Yes	Yes
Blood Pump Flow Rate	Yes	Yes	Yes
Tube Diameter 8.0mm 30 - 600ml / min	Yes	Yes	Yes
Tube Diameter 6.5mm 20 - 400 ml / min	Yes	Yes	Yes
Heparin infusion rate	Yes	Yes	Yes
Syringe size 10ml 0-9.9 ml / h	Yes	Yes	Yes
Syringe size 20ml 0-9.9 ml / h	Yes	Yes	Yes
Air bubble detector Ultrasonic detection system	Yes	Yes	Yes
Blood leak detector Photoelectric detection system	Yes	Yes	Yes
Monitor Section	Yes	Yes	Yes
Dialysate temperature indicator 0-99.9 °C	Yes	Yes	Yes
Dialysate flow indicator 400 – 600 ml / min	Yes	Yes	Yes
Dialysate conductivity indicator 0.0 – 18.0mS / cm	Yes	Yes	Yes
Dialysate pressure indicator -400 – 400 mmHg	Yes	Yes	Yes
Arterial pressure indicator -200 – 400 mmHg	Yes	Yes	Yes

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Venous pressure indicator -200 – 400 mmHg	Yes	Yes	Yes
Transmembrane pressure indicator -400 – 400 mmHg	Yes	Yes	Yes
UF volume indicator 0.00 – 40.0L	Yes	Yes	Yes
Blood pump flow indicator 0 – 600 mL / min	Yes	Yes	Yes
Heparin pump flow indicator 0 – 9.9 mL / h	Yes	Yes	Yes
Disinfection and Rinsing	Yes	Yes	Yes
Chemicals Sodium hypochlorite 5%	Yes	Yes	Yes
USA/UK/Europe/Japan	Japan	Japan	Germany
Remarks	Responsive	Responsive	Responsive

27



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	27
Name of Items	RO Plant with complete fitting for 10 nos. Dialysis machine
Quantity	1 Nos.

✓  
✓  
Revised tender

RO Plant with complete fitting for 10 nos. Dialysis machine	Universal Enterprise	Hospital Supply Cor.	Fresenius Medical
Approved in bid evaluation	Yes	Yes	Yes
Model	3000GPD	EPRO-3000	Aquab Plus
Make	Local	Local	Fresenius Medical
Designed on Feed Water	Yes	Yes	Germany
Remarks	Responsive	Responsive	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	28		
Name of Items	Dialyzer Re-processing system		
Quantity	1 Nos.		
Dialyzer Re-processing system	Universal Enterprise	Hospital Supply Cor.	Fresenius Medical
Approved in bid evaluation	Yes	Yes	Yes
Model	-	Compact	ADR-88
Make	Medivator	Mediopl	Fresenius Medical
Remarks	Responsive	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	29
Name of Items	500mA X-ray unit
Quantity	1 No.

500mA X-ray unit	Medequips	Medequips Alternate	BIOS
Approved in bid evaluation	Yes	Yes	Yes
Model	KSO32S	KSO50S	N-40HF
Make	Toshiba	Toshiba	Stephanix
40 kW or more High frequency X-Ray Generator	32 kW	50 kW	40 kW
4 ways floating top table with fixed height	Yes	Yes	Yes
floor mounted tube stand	Yes	Yes	Yes
manual collimator	Yes	Yes	Yes
140KHU X-ray tube	Yes	Yes	Yes
wall bucky stand	Yes	Yes	Yes
pair of HT cables 10m or more.	Yes	Yes	Yes
Radiographic kVP 40~125kV or more in one kV step.	Yes	Yes	Yes
Radiographic mA 10~500mA or more	Yes	Yes	Yes
mAs range: 0.4~500mAs or more.	Yes	Yes	Yes
Exposure time: 1 ms ~ 6s.	2ms-3.2sec	1ms-9sec	1ms-10sec
2 point kVP and mAs selection. 3 point kVP, mA & time selection.	Yes	Yes	Yes
300 or more Anatomical program.	Yes	Yes	Yes
Electromagnetic brakes Longitudinal and lateral travel movement for table bucky,	Yes	Yes	Yes

X W te

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Longitudinal bucky travel.	Yes	Yes	Yes
Cassette size : 18 x 24 to 35x43cm.	Yes	Yes	Yes
Moving grid 8:1.	Yes	Yes	Yes
Foot switch for electro magnetic brake control.	Yes	Yes	Yes
Tube support balancing System with counterweights and Electromagnetic brakes.	Yes	Yes	Yes
Control panel with push button to control vertical, longitudinal and transverse travel.	Yes	Yes	Yes
Column rotation +/-90deg or better.	Yes	Yes	Yes
Manual collimator.	Yes	Yes	Yes
LED light indicator for alignment with bucky.	Yes	Yes	Yes
Chest bucky stand with the cassettes formats from 18X24 to 35X43 on either orientation.	Yes	Yes	Yes
Moving Potter Bucky super speed.	Yes	Yes	Yes
Vertical travel of bucky not less than 150cm.	Yes	Yes	Yes
Moving grid: 8:1 ratio.	Yes	Yes	Yes
Dual Focal X-ray tube : Small focus 1.0 mm and Large focus 2.0 mm.	0.6/1.2mm	0.6/1.2mm	1/2mm
Maximum Tube voltage 125KVp and Anode Heat capacity 140 KHU.	140KHU	300KHU	140KHU
USA/UK/Europe/Japan	Japan	Japan	France
Remarks	Non-Responsive	Responsive	Responsive

X

W

W



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	30
Name of Items	300mA X-ray unit for OPD
Quantity	1 No.

300mA X-ray unit for OPD	Medequip	Medequip Alternate	BIOS
Approved in bid evaluation	Yes	Yes	Yes
Model	KSO-25S	KSO-32S	N-32HF
Make	Toshiba	Toshiba	Stephanix
30 kW or more High frequency X-Ray Generator	25kW	32 kW	32kW
4 ways floating top table with fixed height	Yes	Yes	Yes
floor mounted tube stand	Yes	Yes	Yes
manual collimator	Yes	Yes	Yes
140KHU X-ray tube	Yes	Yes	Yes
wall bucky stand	Yes	Yes	Yes
pair of HT cables 10m or more.	Yes	Yes	Yes
Radiographic kVP 40~125kV or more in one kV step.	Yes	Yes	Yes
Radiographic mA 10~300Ma or better.	Yes	Yes	Yes
mAs range: 0.4~500mAs or more.	Yes	Yes	Yes
Exposure time: 1 ms ~ 6s.	2ms-3.2sec	2ms-3.2sec	1ms-10sec
2 point kVP and mAs selection.	Yes	Yes	Yes
3 point kVP, mA & time selection.	Yes	Yes	Yes
300 or more Anatomical program.	Yes	Yes	Yes
Electromagnetic brakes Longitudinal and lateral travel movement for table bucky,	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Longitudinal bucky travel.	Yes	Yes	Yes
Cassette size : 18 x 24 to 35x43cm.	Yes	Yes	Yes
Moving grid 8:1 .	Yes	Yes	Yes
Foot switch for electro megnatic brake control.	Yes	Yes	Yes
Tube support balancing System with counterweights and Electromagnetic brakes.	Yes	Yes	Yes
Control panel with push button to control vertical, longitudinal and transverse travel.	Yes	Yes	Yes
Column rotation +/-90deg or better.	Yes	Yes	Yes
Manual collimator.	Yes	Yes	Yes
LED light indicator for alignment with bucky.	Yes	Yes	Yes
Chest bucky stand with the cassettes formats from 18X24 to 35X43 on either orientation.	Yes	Yes	Yes
Moving Potter Bucky super speed.	Yes	Yes	Yes
Vertical travel of bucky not less than 150cm. Moving grid: 8:1 ratio.	Yes	Yes	Yes
Dual Focal X-ray tube : Small focus 1.0 mm and Large focus 2.0 mm.	1/2mm	0.6/1.2mm	1/2mm
Maximum Tube voltage 125KVp and Anode Heat capacity 140 KHU.	140KHU	140KHU	140KHU
USA/UK/Europe/Japan	Japan	Japan	France
Remarks	Non-Responsive	Responsive	Responsive

X

W

W



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	31
Name of Items	X-Ray Auto processor
Quantity	2 Nos.

X-Ray Auto processor	Mediurge	BIOS
Approved in bid evaluation	Yes	Yes
Model	Compac 2	MP-8800
Make	Protoc	Okamoto
Table top film processor to process commercially available medical x-ray films	Yes	Yes
With roller transport system.	Yes	Yes
Processing capacity of 80 film/hour or more.	Yes	Yes
With temperature control system	Yes	Yes
Automatic replenishment system.	Yes	Yes
Complete with tank for developer (3.5L or more), fixer (2.4L or more) and washer (2L or more).	Yes	Yes
Compatible with all common types of films and chemicals.	Yes	Yes
or equivalent	Yes	Yes
USA/UK/Europe/Japan	Germany	Japan
Remarks	Responsive	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	32
Name of Items	Digital color Ultrasound machine
Quantity	1 Nos.

Digital color Ultrasound machine	Medequips	ZAM Trading
Approved in bid evaluation	Yes	Yes
Model	Xario-200	MyLab six
Make	Toshiba	Esate
High frequency Multi beam former Color ultrasound Machine.	Yes	Yes
18" or more Anti flickering rotatable LED display with	19"	19"
8" Touch screen	8.4"	8.9"
4 active probe connectors	3 nos.	3 nos.
Complete with original trolley	Yes	Yes
Modes: 2D mode, M mode, CW/PW imaging, Color M mode, Anatomical M mode, 3D/4D, THI, Pulse inversion harmonic imaging, Trapezoidal Imaging, 2D automatic optimization, color automatic optimization, Speckle Reduction, auto calculation etc.	Yes	Yes
2D and color panoramic, color compare mode, Stress Echo, LGT/TGC etc	Yes	Yes
Frame rate: 800 frames or more	Yes	Yes
Cine loop: 1000 or more	Yes	Yes
Depth : 32 cm or more	Yes	Yes
200° scanning angle	Yes	Yes
1000% zooming facility	Yes	Yes
DICOM Compatible	Yes	Yes
Storage/Printer/Worklist/	Yes	Yes
PPS/Structured Report	Yes	Yes
VGA, DVD and USB port	Yes	Yes

Bid evaluation  
 criteria  
 h, i, j on  
 not fulfilled



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

500GB Hard Disk	Yes	Yes
Complete with :	Yes	Yes
2-6 MHz Multi frequency Convex probe	Yes	Yes
4-16 MHz Multi frequency linear probe	Yes	Yes
2-6MHz Multi frequency phased array probe.	Yes	Yes
Printer and UPS	Yes	Yes
CE and FDA 510K Approved	Yes	Yes
USA/UK/Europe/Japan	Japan	Italy
Remarks	Responsive	Responsive

*✓*  
*Bid evaluation criteria is not fulfilled*

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	33
Name of Items	Digital Black & White Ultrasound system
Quantity	1 Nos.

Digital Black & White Ultrasound system	Medequip	ZAM Trading
Approved in bid evaluation	Yes	Yes
Model	DP-20	HS-2200
Make	Mindray	Honda
High Resolution Portable B/W Ultrasound Machine having Digital Beam Forming System for Abdominal/Obstetrical/Gynae/Urology/small parts with the following Composition: 12" or more LCD Monitor having wide viewing angle with Tilt, Swivel & up/down.	Country of origin is not attached	Yes
Two or more probe connectors		12.1"
Depth: 24 cm or more. STC gain.		Yes
B/M scan, B(Single, Dual), B/M, M		Yes
3~5 MHz Multi Frequency Convex Probe.		Yes
Linear probe 5~10MHz (small parts) .		Yes
comprehensive measurement package/functions with six or more different methods for obs/Gynae. Twin pregnancy measurement package etc		Yes
CD/DVD writer or USB port . Digital B/W Ultrasound printer.		UPS
Voltage Stabilizer.		Yes
Mobile Trolley		Yes
CE and FDA 510K Approved	China	Japan
USA/UK/Europe/Japan		
Remarks	Non-Responsive	Responsive

X

Tw

check & eval



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.**

**394 FOR THE YEAR 2015-2016.**

**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

Item No	34
Name of Items	100mA Mobile Radiography unit
Quantity	1 Nos.
100mA Mobile Radiography unit	BIOS
Approved in bid evaluation	Yes
Model	Movix 4 E+
Make	Stephanix
High frequency X-Ray inverter type Generator	Yes
Having output power of 4kW or more	4kW
X-Ray tube protections	Yes
KVP range: 40-110kV or more	Yes
MA Range: 10-100mA or better	Yes
mAS range: 1.0-200mAs or better	Yes
Centering light indicator	Yes
Standard power supply cable	Yes
standard x-ray exposure switch cable	Yes
vertical movement of the Arm to lower or raise the tube collimator assembly	Yes
Rotation of the Collimator with reference to the power module (+/- 90deg) or better.	Yes
USA/UK/Europe/Japan	France
Remarks	Responsive

*W*

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	35
Name of Items	Video endoscope complete system (both upper & lower GI)
Quantity	1 Nos.

Video endoscope complete system (both upper & lower GI)	Hospital Solution
Approved in bid evaluation	Yes
Model	BF-PE2
Make	Olympus
HIGH DEFINITION GASTROINTESTINAL VIDEO SCOPE	Yes
Waterproof One-touch Connector, High Performance HDTV Compatible	Yes
Close Focus, Slim Design, (NBI System Facility Compatible	Yes
With standard set of accessories	Yes
Field of view: 140deg	Yes
Direction of View: Forward Viewing	Yes
Depth of field: 2 to 100 mm	Yes
Outer diameter: 9.2 mm	Yes
Working length: 1030mm	Yes
Total Working length: 1350mm	Yes
Angulation Range: Up 210°, Down 90° Right 100° Left 100deg	Yes
Inner Diameter: 2.8mm	Yes
Minimum visible distance: 3.0mm	Yes
HIGH DEFINITION COLONO VIDEO SCOPE	Yes
Waterproof One-touch Connector, High Performance HDTV Compatible	Yes
Close Focus, * Water Jet, NBI System Facility Compatible	Yes
VARAIBLE STIFFNESS with standard set of accessories	Yes
Field of view: 140 deg	Yes
Direction of View: Forward Viewing	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Depth of field:	2 - 100mm	Yes
Outer diameter:	12.8 mm	Yes
Working length:	1680mm	Yes
Flushing Pump for Colono Videoscope		Yes
VIDEO SCOPE SYSTEM		Yes
Combination of HDTV and NBI is Now Available with an LED Light		Yes
Frequency	50 / 60 Hz	Yes
Rated input	200VA	Yes
Accep. Ratio	16:9 and 16:10 Output	Yes
Digital signal output	HD-SDI (SMPTE 292M), SD-SDI (SMPTE 259M), DV (IEEE 1394), DVI (WUXGA, 1080p or SXGA) can be selected	Yes
White balance adjustment	White balance adjustment is possible using the white balance button on the front panel	Yes
Standard color chart output	The "Color bar" or the "50% white" screen can be display	Yes
DVR, Video Printer	Image flying system • Flushing pump • Endoscopic co2 regulation unit	Yes
Patient data	The following data can be displayed on the monitor using the keyboard	Yes
Patient ID • Patient name • Sex • Age • Date of birth • Date of recording (time, stopwatch) •		Yes
Comments		Yes
Displaying the record state	The recording state of the following ancillary equipment can be displayed on the monitor	Yes
Portable memory and internal buffer • DVR • Video printer • Image filing system		Yes
Memorization selected setting • Color tone • Iris mode • Enhancement • Color enhancement mode • Contrast • AGC • Color mode • White balance • Brightness adjustment method • Brightness • Air feeding		Yes
High Definition LCD Medical Grade Monitor 21" (SONY		Yes
HDTV Compatible		Yes
Local Video Trolley with anti castor and keyboard drawer		Yes
USA/UK/Europe/Japan		Japan
Remarks		Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	36
Name of Items	Chemistry Analyzer
Quantity	1 Nos.

Chemistry Analyzer	ZAM Trading	Sindh Medical Store	Hoora Pharma	Lab Medikal Solution	Gulf Marketing	Popular International	Lucky Lab Int.
Approved in bid evaluation	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Model	RX Monza	Starfax 4500	Star Dust MC 15	-	Tecno 786 plus	Humalyzer 3500	BTS - 350
Make	Randox	Awareness technology	Diasys Diagnostic	Metrolab	Audit Diagnostic	Human	Biosystems
Semi automatic Chemistry Analyzer.	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Light Source: Quartz halogen lamp 12V-20W.	Yes	Yes	Yes	Yes	Not mentioned	Yes	Yes
Wave Length range: Automatic by 12 positions filter wheel.	Not mentioned	Yes	Yes	Yes	Yes	Yes	Yes
6 standard interference filter : 340,405,505,546,578,& 620nm.	Yes	Yes	No	Yes	Yes	Yes	Yes
Photometric range: -0.1 to 2.3 Absorbance.	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Blanking : Automatic zero setting.	Not mentioned	Yes	Yes	Yes	Yes	Yes	Yes
Operator interface : Membrane key board, for direct function and alpha-numeric entry.	Yes	Yes	Yes	Yes	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 17<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Measurement procedure: kinetic with linearity check.							
Kinetic, with linearity check and sample slope blank.	Yes	Yes	Yes	Yes	Yes	Yes	Not mentioned
Two point kinetic with or without reagent blank; end point with or without reagent blank.	Yes	Yes	Not mentioned	Yes	Yes	Yes	Yes
Bi-chromatic end point, with or without reagent blank. End point, with sample blank	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Multiple Testing : up to nine replicates: means, SD & CV.	Yes	Yes	Yes	Yes	Yes	Not mentioned	Yes
For end point fixed at 2 seconds.	No	Yes	Yes	Yes	Yes	Yes	Yes
Delay time: Programmable , 0 to 999 seconds.	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Calibration: factor, one-point , two point and multipoint.	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Flow Cell: Metal with quartz window measuring volume 25ul.	Yes	Yes	Yes	Yes	Yes	Yes	Yes
Temperature control: by means of peltier elements. Fixed temperature at 37°C	Yes	Yes	Yes	Yes	Yes	Yes	Yes
USA/UK/Europe/Japan	UK	USA	Germany	Argentina	Ireland	German Y	Spain
Remarks	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive

As per

Signature

Non

Not mentioned



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	37					
Name of Items	Electrolyte Analyzer					
Quantity	1 Nos.					
Electrolyte Analyzer	Bashon Pharma	ZAM Trading	Sindh Medical Store	Lab Medical Solution	Gulf Marketing	Popular International
Approved in bid evaluation	Yes	Yes	Yes	Yes	Yes	Yes
Model	Smartlyte	Prolyte	Selectalyte	SR-1520	E Lyte plus	Humalyte plus 3
Make	Diamond Diagnostic	Diamond Diagnostic	Awareness technology	Arkray	Electra Medical	Human
Electrolyte Analyzer for Whole Blood, Serum, Plasma, or Urine.	Yes	Yes	Yes	Yes	Yes	Yes
100ul Whole Blood, Serum or Plasma. 60ul Capillary Sample	Yes	Yes	Yes	Yes	Yes	Yes
Detection Range	Yes	Yes	Yes	Not mentioned	Yes	Yes
Blood Na+:20-200mmol/L, K+:0.2-40mmol/L	Yes	Yes	Yes	Yes	Yes	Yes
Cl:-25-200mmol/L, Urine Na+:25-1000mmol/L (undiluted), K+:1.0-500mmol/L (undiluted)	Yes	Yes	Yes	Yes	Yes	Yes
Cl:-25-500mmol/L (undiluted)	Yes	Yes	Yes	Yes	Yes	Yes
Analysis Time 60 sec or better	Yes	Yes	Yes	Yes	Yes	Yes
Data Storage facility . Automatic calibration.	Yes	Yes	Yes	Yes	Yes	Yes
Alphanumeric display.	Yes	Yes	Yes	Yes	Yes	Yes
40 column printer RS-232 communication port	Yes	Yes	Yes	Yes	Yes	Yes
USA/UK/Europe/Japan	USA	USA	USA	Japan	USA	Germany
Remarks	Responsive	Responsive	Responsive	Responsive	Responsive	Responsive

*JS spaces*



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	38
Name of Items	Incubator for lab
Quantity	1 Nos.

Incubator for lab	ZAM Trading	Popular International
Approved in bid evaluation	Yes	Yes
Model	RI53	Huma therm
Make	Binder	Human
Microprocessor Control.	Yes	Yes
PID temperature control.	Yes	Yes
LED digital display to show actual PV and preset chamber.	Yes	Yes
Over temperature safety device.	Yes	Yes
99 hours electronic digital timer.	Yes	Yes
Temperature 300°C. resolution 1°C.	Yes	Yes
LED 3 digit display .	Not mentioned	Yes
0.1°C stability Stainless steel interior and powder coated exterior.	Yes	Yes
Capacity 36 Liters .	Yes	Yes
chamber size : 330x330x330. 2 x shelves	Yes	Yes
USA/UK/Europe/Japan	Germany	Germany
Remarks	Responsive	Responsive

As per specs  
 JCEW


# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	39
Name of Items	Hot air oven
Quantity	1 Nos.

Hot air oven	ZAM Trading
Approved in bid evaluation	Yes
Model	RE53
Make	Binder
Capacity : 34L. Microprocessor Control system.	Yes
Temp control: PID.	Yes
Digital Temp setting .	Yes
Digital Temp readout.	Yes
Air convection: Forced convection by fan.	Yes
Temp uniformity $\pm 2.0^{\circ}\text{C}$ at $55^{\circ}\text{C}$ .	Not mentioned
Temp stability $\pm 0.5^{\circ}\text{C}$ at $37^{\circ}\text{C}$ .	Yes
Over-temp alarm & shutout : Yes, $4^{\circ}\text{C}$ to the set point.	Not mentioned
Timer : Digital, 99 hr. 59min. + HOLD.	Yes
Temperature range : Digital, 99 hr. 59min.	Yes
HOLD Standard accessory : 2 S/S mesh shelves, adjustable	Yes
USA/UK/Europe/Japan	Germany
Remarks	Responsive

As per specs  




# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	40
Name of Items	Centrifuge machine (24 heads)
Quantity	1 Nos.

Centrifuge machine (24 heads)	ZAM Trading	Zeal Pharm. Se.	Popular International
Approved in bid evaluation	Yes	Yes	Yes
Model	Mikro 120	PLC-012H	Humax 14K
Make	Hettich	Germany	Humax
Complete with 24-place angle rotor.	Yes	Country of origin is not as per tender specification	Yes
Microprocessor-controlled.	Yes		Yes
Dimension 305X355X320 mm	Not mentioned		Yes
Head : 24 nos.	Yes		Yes
Motor 80 W	Yes		Yes
Tachometer LED digital Dial tachometer	Yes		Yes
Timer LED digital, 99 min LED	Yes		No
Brake	No		Yes
Locking latch	Yes		Yes
Safety shutout	Yes		
USA/UK/Europe/Japan	German y	Taiwan	German y
Remarks	Responsive	Not responsive	Responsive

As per spec

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	41
Name of Items	Ultrasonic water bath
Quantity	2 Nos.

Ultrasonic water bath	Popular International
Approved in bid evaluation	Yes
Model	Huma aqua 5
Make	Human
Volume: 22L. Uniformity: .5 °C. Heater: 1 KW	Yes
Readout: Thermometer.	Yes
Temperature Controller: PID, +/-0.1 °C.	Yes
Stirrer: Built-in	Yes
USA/UK/Europe/Japan	Germany
Remarks	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	42
Name of Items	Electronic balance for lab
Quantity	1 Nos.

Electronic balance for lab	ZAM Trading
Approved in bid evaluation	Yes
Model	PA 214C
Make	Ohaus
Capacity & Resolution: 400g x 0.01g	Yes
Percentage: Resolution: 0.01 / 0.1 / 1%	No
Stabilization Time: 2 sec, Typically	Yes
Sensitivity drift: (10 ppm/°c (10°c-30°c)	Yes
Linearity: +/-0.02 g	Yes
Repeatability/st D.Dev:0.01g	Not mentioned ?
Display Refresh: 5 or 10 times	Yes
Operating temperature: +5°c-40 th less than 85% no condensation	Yes
USA/UK/Europe/Japan	USA
Remarks	Responsive

As per spec

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	43
Name of Items	Sterilizer autoclave vertical type
Quantity	1 Nos.

Sterilizer autoclave vertical type	No Bid received for this item
Approved in bid evaluation	
Model	
Make	
Tabletop autoclave sterilizer.	
25 liter chamber capacity.	
ry, re dry function.	
Auto exhaust function.	
Using Pressure: Unwrapped /wrapped liquid.	
And Temperature 115-134°/0.9-2.1kg/cm².	
Designed temperature : 140°.	
Safety device : Overheat (Low water ) cut off switch safety valve and release valve emergency exhaust.	
valve door safety holder pressure control switch	
or equivalent	
USA/UK/Europe/Japan	
Remarks	X



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	44
Name of Items	Binocular Microscope
Quantity	1 Nos.

Binocular Microscope	Popular International
Approved in bid evaluation	Yes
Model	Humascope premium LED
Make	Human
Slide type Binocular head with 45° inclined and 360° rotation.	Yes
Wide-field eyepiece 18 and 13mm paired.	Yes
Quadruple nose piece with built in click stopper.	Yes
DIN achromatic objectives.	Yes
Built-in mechanical stage 140mm x 140mm, moving range 50mm x 75mm, with specimen holder controlled by coaxial knobs at low position.	Yes
Focusable condenser and diaphragm iris in filters and filter holder.	Yes
Built-in electrical illuminator with halogen bulb.	Yes
Coaxial focusing system with built in tension adjustment and focus stopper.	Germany
USA/UK/Europe/Japan	Responsive
Remarks	

✓  
 AS per specs  
 JAE

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	45
Name of Items	Juster (fixed)
Quantity	4 Nos.

Juster (fixed)	ZAM Trading	Lab Medikal Solution	Popular International
Approved in bid evaluation	Yes	Yes	Yes
Model	-	-	Humapette Fix
Make	Biohit	Capp	Human
USA/UK/Europe/Japan	Finland	Denmark	Germany
Remarks	Responsive	Responsive	Responsive

As per spaces  
 JCEV



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**



Item No	46	No bid received for this item
Name of Items	Refrigerator rotaray	
Quantity	1 Nos.	

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 17<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	47	
Name of Items	Dental Chair	
Quantity	4 Nos.	
<b>Dental Unit consisting of :</b>		
Approved in bid evaluation		Bashson Pharma
Model		Yes
Make		N-3
Dental Chair		Yes
Doctor's Tray		Yes
Cuspidor unit		Yes
Junction box		Yes
Dental light		Yes
<b>Dental Unit complete</b>		
Dental Units fully Automatic Set unit Chair /Doctor School / Air Compressor / Triple Syringe / Hand Piece / Unit Splitter		Yes
Dental Treatment Centre		Yes
Patient Chair		Yes
With electrical chair movement, double articulated headrest and backrest.		Yes
2 preset chair positions: start and rinsing position		Yes
3 programmable chair positions		Yes
Last memory position the chair returns to the previous treatment position after using the preset rinsing lowest chair position can reach 380mm, with emergency treatment position.		Yes

*As per form*





# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Chair movement is controllable by dentist and assistant touch panels, as well as the multifunctional foot everything is integrated into the chair base, no cables on the floor, hygienic and clean.	Yes
One main switch to control air, water and power.	Yes
Chair instrument lock enables safe operation, the chair is locked while an instrument is working. With safety system, backrest and seat movement can be stopped once it meets obstacle.	Yes
Dentist element	Yes
Hanging hoses / Height of dentist element is adjustable / Instrument holder / Control Panel can be used for operating functions, such as chair movement preset and programmable cup filler, bowl rinsing operating light on / off and intensity adjustment / 1 three way syringe.	Yes
2 Ultra Push Air Turbine Autoclaveable 400,000 rpm, ( mini / standard ). 1 Air motor with contra and straight angle hand pieces. Contra and straight angle hand pieces. Contra and straight hand pieces are autoclave. / 1 Ultrasonic Scaler Piezon with four tips built in type.	Yes
1/- Foot Controller	Yes
1/- air pressure gauge.	Yes
1/- silicon mat which can be sterilized	Yes
Assistant element	Yes
With an angle adjustable support arm.	Yes
With suction filter system for easy cleaning.	Yes
1 three way syringe	Yes
1 HVE	Yes
1 saliva ejector	Yes
1 LED Curing light	Yes

*Dr. Farhan*


*[Signature]*

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016,  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Water unit	Yes
Mounted to the chair / The cuspidor can be swiveled and removable for easy cleaning / Fresh water bottle 1.5L / A switch to choose city water or fresh water bottle / Programmable cup filler and bowl rinsing systems.	Yes
Brightness is 25000 Lux / 12V, 50 watts	Yes
Light on / off and intensity can be controlled through the touch panel / The light turns off automatically during chair movement, and turns back to the previous intensity when the chair stops. Working stool	Yes
Mobile on 5 castors / Height of seat and backrest is adjustable / Backrest angle is adjustable and lockable.	Yes
Oil free Dental Air Compressor	Yes
Motor (3/4-1 H.P) / Output: 100 L / min or above / Working Pressure:-5-8 bar / Tank Capacity : 24 litre / Power 220 Volt, 50 Hz, Complete set.	Yes
USA/UK/Europe/Japan	China
Remarks	Responsive

at 12/5/16  




# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	48
Name of Items	Ultrasonic Scaler ✓
Quantity	4 Nos.

Ultrasonic Scaler	Bashson Pharma
Approved in bid evaluation	Yes
Model	Wood packer
Make	-
LCD display	Yes
Input/Inlet pressure: 1-5Mpa	Yes
Output power : 3W-20W	Yes
Semi-offset output power: <2N	Yes
Inlet Pressure: 0.01Mpa-0.5Mpa	Yes
Operating Frequency: 24-32 KHZ	Yes
Complete with 2 x bottles and 6 x tips	Yes
USA/UK/Europe/Japan	China
Remarks	Responsive

*Handwritten signatures and initials at the bottom of the page.*

**SINDHI GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.**

**394 FOR THE YEAR 2015-2016.**

**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Item No	49	No bid received for this item
Name of Items	Autoclave dental	
Quantity	1 Nos.	

Item No	50	No bid received for this item
Name of Items	Dry head sterilizer dental	
Quantity	1 Nos.	






**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Item No	51
Name of Items	Dental instrument sets
Quantity	20 Nos.

Dental instrument sets	HAQ Enterprise	RZ Service
Approved in bid evaluation	Yes	Yes
Local, export quality	Yes	Yes
Remarks	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	52
Name of Items	Flow meter
Quantity	20 Nos.

Flow meter	Standard Suppliers	ZAM Trading	HAQ Enterprise
Approved in bid evaluation	Yes	Yes	Yes
USA/UK/Europe/Japan	UK	UK	Japan
Remarks	Responsive	Responsive	Responsive

Item No	53	No bid received for this item
Name of Items	Stabilizer 5 KVA	
Quantity	4 Nos.	



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## **TECHNICAL EVALUATION**

Item No	54
Name of Items	Oxygen cylinder 240 liters
Quantity	20 Nos.

Oxygen cylinder 240 liters	ZAM Trading	HAQ Enterprise	RZ Service	TK Inst.
Approved in bid evaluation	Yes	Yes	Yes	Yes
BOC tested	Yes	Yes	Yes	Yes
Remarks	Responsive	Responsive	Responsive	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	55
Name of Items	Oxygen cylinder 48/c liter
Quantity	20 Nos.

Oxygen cylinder 48/ liter	ZAM Trading	HAQ Enterprise	RZ Service	TK Inst.
Approved in bid evaluation	Yes	Yes	Yes	Yes
BOC tested	Yes	Yes	Yes	Yes
Remarks	Responsive	Responsive	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## **TECHNICAL EVALUATION**

Item No	56
Name of Items	Computer with monitor , laser printer and other complete accessories
Quantity	10 Nos.

Computer with monitor , laser printer and other complete accessories	HAQ Enterprise
Approved in bid evaluation	Yes
Make	Dell
Core 5i	Yes
4.5 Ghz optical super combo	Yes
500 GB HDD	Yes
4 GB DDR 3 RAM	Yes
17" LCD Monitor	Yes
Key Board & Mouse	Yes
Laser Printer	Yes
Remarks	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	57
Name of Items	Refrigerator
Quantity	25 Nos.

Refrigerator	HAQ Enterprise	TK Inst.
Approved in bid evaluation	Yes	Yes
Dawlance/Sharp/Haier	Dawlance	Dawlance
Remarks	Responsive	Responsive

Item No	58	No bid received for this item
Name of Items	IT Net working	
Quantity	Nos.	



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI - SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	59
Name of Items	Cold storage facility ✓
Quantity	1 Nos.

Cold storage facility	Standard suppliers
Approved in bid evaluation	Yes
Model	MMC 2+W2
Make	Evermed
2 body storage with modular construction.	Yes
Complete with bay and tier made of 304 grade stain less steel storage cabinet	Yes
80mm high density rigid insulation polyurethane insulation chiller	Yes
stain less steel height adjustable racking & rolling system	Yes
304 grade stain less steel body storage trays	Yes
10°C to -5°C cooling system with digital control panel	Yes
body transport trolley.	Yes
USA/UK/Europe/Japan	Italy
Remarks	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	60	
Name of Items	Color TV	
Quantity	2 Nos.	No bid received for this item ✓

Item No	61	
Name of Items	A/c 1 ½ tons	
Quantity	40 Nos.	

A/c 1 ½ tons	HAQ Enterprise	TK Inst.
Approved in bid evaluation	Yes	Yes
Dawlance/Sharp/Haier	Haier	Dawlance
Remarks	Responsive	Responsive



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Item No	62
Name of Items	A/c Split 2 tons
Quantity	10 Nos.

A/c Split 2 tons	TK Inst.
Approved in bid evaluation	Yes
Dawlanee/Sharp/Haier	Dawlanee
Remarks	Responsive

**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Item No	63	No bid received for this item ✓
Name of Items	Photocopier	
Quantity	1 Nos.	

Item No	64	No bid received for this item ✓
Name of Items	Fax Machine	
Quantity	1 Nos.	

Item No	65	No bid received for this item ✓
Name of Items	Multi Media	
Quantity	1 Nos.	



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

**TECHNICAL EVALUATION**

Item No	66
Name of Items	Over Head Project with screen set
Quantity	1 Nos.

Over Head project with screen	Bashson Pharma
Approved in bid evaluation	Yes
Remarks	Responsive

# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 12<sup>th</sup>, 2016

## TECHNICAL EVALUATION

*CE E May 11*

Item No	67
Name of Items	Central oxygen supply system
Quantity	1 Nos.

Central oxygen supply system	Universal Enterprise	ZAM Trading	Standard suppliers	Mediurge
Approved in bid Evaluation	Yes	Yes	Yes	Yes
Model	Manifold O2	Manifold	-	160GAR
Make:	CPX	-	Penlon	G.Samaras
AUTOMATIC MANIFOLD FOR OXYGEN FOR 2 X 10 CYLINDERS WITH CONNECTION TO VIE TANK.	Yes	Yes	Yes	Yes
The Control Panel should handle three sources i.e. Two sided cylinders and one VIE Tank.	Yes	Yes	Yes	Yes
The Manifold shall be a fully automatic type & shall switch from Tank / Generation System to Cylinders	Yes	Yes	Yes	Yes
bank without fluctuation in delivery supply line pressure	Yes	Yes	Yes	Yes
The system should be able to maintain continuous supply in case of power failure.	Yes	Yes	Yes	Yes
Capacity: 70-90, M3/h.	Yes	Yes	Yes	Yes
Fully automatic self-contained shuttle-valve with no electrical power required for switching, Microprocessor based control panel with display.	Yes	Yes	Yes	Yes
Units of measure (psi/Pa/Bar).	Yes	Yes	Yes	Yes
AUTOMATIC MANIFOLD FOR NITROUS OXIDE (2 X 5)	Yes	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

The Manifold shall be a fully automatic type & shall switch from "Bank in Use" to "Reserve" bank with out fluctuation in delivery supply line pressure&with out the need for external power after the switch-over	Yes	Yes	Yes	Yes
the "Reserve" bank shall then become the "Bank in Use" and the bank in use shall become the reserve bank. The System should be able maintain continuous supply in case of Power failure.	Yes	Yes	Yes	Yes
Capacity 10 - 50 M3/h.	Yes	Yes	Yes	Yes
Fully automatic self-contained shuttle-valve with no electrical power required for switching, Microprocessor based control panel Service and maintenance works can be carried out without stopping operation.	Yes	Yes	Yes	Yes
<b>WALL OUTLET.</b> Outlet should have double lock quick connect and gas specific. The latch / locking release mechanism should be incorporated on the gas outlet with geometrical key indexing and should have a parking facility so that in the event of release of probe by pressing the spring loaded center, the probe / flowmeter does not fall down.	Yes	Yes	Yes	Yes
Gas outlets wall mounted, Complete in Box/casing gas Specific, Self sealing valve,indexed to eliminate interchangeability/erroneous tapping of gas and cover plate.	Yes	Yes	Yes	Yes
Oxygen Qty:92 Nos.	Yes	Yes	Yes	Yes
Nitrous Oxide Qty:6 Nos.	Yes	Yes	Yes	Yes
Compressed Air 4/5 Bar Qty:92 Nos.	Yes	Yes	Yes	Yes
<b>Air Venturi System with Gauge</b> The injector should be connected to wall source of pressure gas using direct probe or rail mounting system. Max. Vacuum 650 to 700 mbar. System should be supplied with 02liter Un Breakable &Sterilizable collection jar.	Yes	Yes	Yes	Yes
<b>ITEM 5(a) Multiple Zone Valve Box with Monitoring Gauge for three gases</b>	Yes	Yes	Yes	Yes
Microprocessor controlled with digital display of pressure.	Yes	Yes	Yes	Yes
Each wall type zone valve box shall consist of the following Components.	Yes	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

A steel valve box which can house two to four shut-off-ball valve with tube extensions.	Yes	Yes	Yes	Yes	Yes
<b>ITEM 5(b)</b>	Yes	Yes	Yes	Yes	Yes
<b>Multiple Zone Valve Box with Monitoring Gauge for two gases</b>	Yes	Yes	Yes	Yes	Yes
Microprocessor controlled with digital display of pressure. Each wall type zone valve box shall consist of the following Components. A steel valve box which can house two to four shut-off-ball valve with tube extensions.	Yes	Yes	Yes	Yes	Yes
<b>ITEM 6</b>	Yes	Yes	Yes	Yes	Yes
<b>CEILING PENDANTS:</b> 06 pendants	Yes	Yes	Yes	Yes	Yes
Each pendant comprises 02 Oxygen, 2 Air 4 Bar & 1 Nitrous Oxide Point	Yes	Yes	Yes	Yes	Yes
Flexible Ceiling Type Outlets:	Yes	Yes	Yes	Yes	Yes
<b>ITEM 7</b>	Yes	Yes	Yes	Yes	Yes
<b>OXYGEN FLOW METER WITH HUMIDIFIER &amp; PROBES:</b>	Yes	Yes	Yes	Yes	Yes
With autoclave able and unbreakable humidifier bottle and probe.	Yes	Yes	Yes	Yes	Yes
<b>ITEM 8 COPPER PIPING (IMPORTED FROM UK, EUROPE, USA):</b>	Yes	Yes	Yes	Yes	Yes
Supply and installation of seamless copper pipe, deoxidized and for use in piped medical gas systems, therefore it should be guaranteed that the tubes are leak-tight, free of pores, blisters etc, Tube should be drawn seamless, according to HMT 2022 (HTM 02-01) Or equivalent.	Yes	Yes	Yes	Yes	Yes
<b>Material :-</b>	Yes	Yes	Yes	Yes	Yes
Copper tubes, including fixing clips for ceiling and walls, sleeves, reducers, tee's bends and soldering material.	Yes	Yes	Yes	Yes	Yes
1350 Meter Size 15 x 1.	Yes	Yes	Yes	Yes	Yes
1500 Meter Size 12 x 1.	Yes	Yes	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.**

**394 FOR THE YEAR 2015-2016.**

**Tender Opened on May 12<sup>th</sup>, 2016**

## TECHNICAL EVALUATION

<b>ANESTHESIA SCAVENGING SYSTEM</b>	Yes	Yes	Yes	Yes
Anesthetic gas scavenging systems designed to remove exhaled anesthetic gas mixtures from operating theatres driven by Air/Oxygen at 05 bar or through vacuum plant, anesthetic rooms and recovery area.	Yes	Yes	Yes	Yes
<b>Oil Free Duplex Air Compressor System</b>	Yes	Yes	Yes	Yes
<b>01 DUPLEX SYSTEM</b>	Yes	Yes	Yes	Yes
100% oil free 10 HP duplex Medical Air Compressor Plan with air dryer, filters, Bypass system, Electrical Panel	Yes	Yes	Yes	Yes
Air Compressor System x 2	Yes	Yes	Yes	Yes
2 x Model: 7.5OP-9.5V/6A	Yes	Yes	Yes	Yes
Motor Power (KW / HP): 3.7 / 5	Yes	Yes	Yes	Yes
Piston Displacement: 541 lit/min	Yes	Yes	Yes	Yes
Free Air Flow: 440 lit/min	Yes	Yes	Yes	Yes
Working Pressure: 9.5 KG/CM2G (135 PSI)	Yes	Yes	Yes	Yes
Power Source: 380V / 3PH / 50HZ	Yes	Yes	Yes	Yes
Noise Level: 74 dB(A) at 1.5 Meter	Yes	Yes	Yes	Yes
Tank: 125 Liters	Yes	Yes	Yes	Yes
<b>Air Dryers</b>	Yes	Yes	Yes	Yes
2 x Dryer Model: AMH-12	Yes	Yes	Yes	Yes
Volume Flow Rate @ Inlet: 1200 lit/min (@80 °C	Yes	Yes	Yes	Yes
Max. Operating Pressure: 14 barg.	Yes	Yes	Yes	Yes
Max. Ambient Temperature: 45 °C	Yes	Yes	Yes	Yes
Pressure Dew Point: +3 °C	Yes	Yes	Yes	Yes
Refrigerant: R134 A	Yes	Yes	Yes	Yes
Power Source: 230V / 1PH / 50HZ	Yes	Yes	Yes	Yes
<b>Filtration</b>	Yes	Yes	Yes	Yes
2 x Pre-Filter: FT-012P – 1200 lit/min (3 Micron) ½"	Yes	Yes	Yes	Yes
4 x After-Filter: FT-012X – 1200 lit/min (0.01 Micron) ½"	Yes	Yes	Yes	Yes
1.9) By Pass System (Local with Imported Fittings):	Yes	Yes	Yes	Yes
Isolation Valves, Adaptors, Regulators & Pressure Switches.	Yes	Yes	Yes	Yes



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO.

394 FOR THE YEAR 2015-2016.

Tender Opened on May 17<sup>th</sup>, 2016

## TECHNICAL EVALUATION

Electrical Panel	Yes	Yes	Yes	Yes
Simple Controlled Electrical Panel, Pressure based Automatic Switching between Compressors, Contractors, Overload, Power Indication Lights and Circuit Breakers.	Yes	Yes	Yes	Yes
UK/USA/BEC/Japan.	UK	UK	UK	Greece
Remarks	Responsive	Responsive	Responsive	Responsive

  
Prof. Tariq Mahmood  
HOD, Radiology, JPMC (Member)

  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Razauddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Majid Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Al. Nuraza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehzeela  
Pediatrician (Member)

  
Dr. Muhammad Raza  
Medical Superintendent  
Sindh Govt. Medical Superintendent  
Dr. Abdul Razzaq Siddiqui  
Medical Superintendent  
Sindh GOVT. HOSPITAL  
Korangi No. 5, Karachi





# **BID EVALUATION**

**SINDH GOVERNMENT HOSPITAL  
KORANIG-05, KARACHI**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT,  
INSTRUMENT, MISC: ITEMS, UNDER ADP  
SCHEME NO. 394 FOR THE YEAR 2015-2016.**

**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder:** M/s. Mediequip


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	Yes
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	Yes
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hanif Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. Hospital Solution**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	No
Satisfactory report from the user, with for 2 years satisfactory report.	No
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Khazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hamid Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 11<sup>th</sup>, 2016**

**BID EVALUATION**


**Name of Bidder: M/s. TK Instrument**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	No
One year business history with the Manufacturer	No
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	No
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq/Mehmood  
HOD, Radiology, IPMC (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. M. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Raheed Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shahnazeela  
Pediatrician (Member)



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


## TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.


Tender Opened on May 12<sup>th</sup>, 2016


### BID EVALUATION


Name of Bidder: M/s. Bashon Pharma


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	No
Satisfactory report from the user, with for 2 years satisfactory report.	No
Proof for the availability of workshop & trained engineers	No
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehnaz  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


Name of Bidder: M/s. Standard Suppliers


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. M. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Naeem Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehnaz  
Pediatrician (Member)



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


## TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.


Tender Opened on May 12<sup>th</sup>, 2016


### BID EVALUATION


Name of Bidder: M/s. S.M.S Diagnostic


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Kiazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehnaza  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 11<sup>th</sup>, 2016**

**BID EVALUATION**


**Name of Bidder: M/s. Lab Medical Solutions (Pvt) Limited**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	No
Satisfactory report from the user, with for 2 years satisfactory report.	No
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mahmood  
HOD, Radiology, JPMC (Member)


  
Dr. Razauddin Shaikh  
Cardiologist (Member)


  
Dr. M. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Raheed Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehneela  
Pediatrician (Member)



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


Name of Bidder: M/s. BIOS


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	Yes
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology / JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shemeela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi




**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


Name of Bidder: M/s. Universal Enterprises


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Raheem Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehnaz  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi




# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


## TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016. Tender Opened on May 12<sup>th</sup>, 2016


### BID EVALUATION


Name of Bidder: M/s. Elate c.c Pvt Ltd


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	
Item no. 04	Yes
Item no. 10	No
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Razauddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hamid Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 11<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. Gulf Marketing international**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	No
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Kiazuddin Shaikh  
Cardiologist (Member)


  
Dr. M. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Razzaq Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehneela  
Pediatrician (Member)



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 11<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. RZ SERVICES**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	N/A
Sole agency/distributorship certificate.	N/A
3 year validity of Sole agency certificate	N/A
One year business history with the Manufacturer	N/A
Proof of past import from Manufacturer, copy of AWB/BL/GD	N/A
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	N/A
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	N/A
Satisfactory report from the user, with for 2 years satisfactory report.	N/A
Proof for the availability of workshop & trained engineers	N/A
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Riazuddin Shaikh,  
Cardiologist (Member)


  
Dr. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehnaz  
Pediatrician (Member)



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. Zeal Pharmaceutical Services**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	No
Copy of valid Professional Tax certificate	No
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Muneeb Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehneela  
Pediatrician (Member)



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. Hooru Pharma**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology/JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi




# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


## TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016. Tender Opened on May 12<sup>th</sup>, 2016


### BID EVALUATION


Name of Bidder: M/s. Medi Urge (Pvt) Ltd.

Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
2,10,14,15	No
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
Item no. 1 & 5	Expired
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	No
Proof for the availability of workshop & trained engineers	Yes
REMARKS	Responsive


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Mufazza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


## TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.


Tender Opened on May 12<sup>th</sup>, 2016


### BID EVALUATION


Name of Bidder: M/s. Hospital Supply Corporation


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. M. Mufaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehnecia  
Pediatrician (Member)




**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder:** M/s. Fresenius Medical


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 11<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. HAQ Enterprises**


<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	N/A
Sole agency/distributorship certificate.	N/A
3 year validity of Sole agency certificate	N/A
One year business history with the Manufacturer	N/A
Proof of past import from Manufacturer, copy of AWB/BL/GD	N/A
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	N/A
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	N/A
Satisfactory report from the user, with for 2 years satisfactory report.	N/A
Proof for the availability of workshop & trained engineers	N/A
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology/JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Muneed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehneela  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**


**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder: M/s. ZAM Trading**


	<b>Documents required as per tender terms and condition</b>	<b>Remarks</b>
a	Photo Copy of Bid security without amount.	Yes
b	Original tender purchase receipt	Yes
c	Receipts for alternate offer	N/A
d	Detail Technical data sheet of quoted items.	Yes
e	Original catalogue of quoted items.	Yes
f	Copy of quality certificates i.e. CE/FDA	Yes
g	Sole agency/distributorship certificate.	Yes
h	3 year validity of Sole agency certificate	No
i	One year business history with the Manufacturer	No
j	Proof of past import from Manufacturer, copy of AWB/BL/GD	No
k	certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
l	Copy of valid Professional Tax certificate	Yes
m	Copy of Income Tax Certificate	Yes
n	Copy GST Registration Certificate.	Yes
o	copy of valid PNRA certificate	N/A
p	List of users in Sindh and all over Pakistan.	Yes
q	Satisfactory report from the user, with for 2 years satisfactory report.	Yes
r	Proof for the availability of workshop & trained engineers	Yes
	<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology/JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Jameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehnaz  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.


## TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.


Tender Opened on May 12<sup>th</sup>, 2016


### BID EVALUATION


Name of Bidder: M/s. Lucky Lab International


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	No
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. M. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdulameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehzeela  
Pediatrician (Member)




**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**


**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**


**BID EVALUATION**


**Name of Bidder:** M/s. Hospicare System


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	No
Copy of valid Professional Tax certificate	No
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
<b>REMARKS</b>	<b>Responsive</b>


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. M. Murtaza Raza  
Radiologist (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi

  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. Shehneela  
Pediatrician (Member)



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI -  
SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.  
Tender Opened on May 11<sup>th</sup>, 2016**

**BID EVALUATION**

Name of Bidder: M/s. Meditech Instrument

Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	<del>Yes</del> Amount is mentioned
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
REMARKS The bidder has mentioned the amount of earnest money in the technical bid by hence bid is disqualified.	Non Responsive

Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)

Dr. Umer Daraz  
Chief Pathologist, CHK (Member)

Dr. Robina Bashir  
Add. Medical Superintendent (Member)

Dr. Riazuddin Shaikh  
Cardiologist (Member)

Dr. Abdul Hameed Shaikh  
Pathologist (Member)

Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

Dr. M. Murtaza Raza  
Radiologist (Member)

Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

Dr. Shehnaz  
Pediatrician (Member)

Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi



**SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI –  
SINDH.**

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS  
UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**

**Tender Opened on May 12<sup>th</sup>, 2016**

**BID EVALUATION**

Name of Bidder: M/s. S.Ejaz u ddin & Co

Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Amount mentioned in Pay order
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	Yes
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	Yes
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	Yes
copy of valid PNRA certificate	N/A
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Yes
REMARKS Bidder has mentioned the amount of	Non Responsive

Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)

Dr. Umer Daraz  
Chief Pathologist, CHK (Member)

Dr. Robina Bashir  
Add. Medical Superintendent (Member)

Dr. Riazuddin Shaikh  
Cardiologist (Member)

Dr. Abdul Hamid Shaikh  
Pathologist (Member)

Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

Dr. M. Murtaza Raza  
Radiologist (Member)

Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

Dr. Shehnaz  
Pediatrician (Member)

Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi




# SINDH GOVERNMENT HOSPITAL, KORANGI-5, KARACHI – SINDH.

**TENDER FOR PURCHASE OF MACHINERY EQUIPMENT, INSTRUMENT, MISC. ITEMS**  
**UNDER ADP SCHEME NO. 394 FOR THE YEAR 2015-2016.**  
**Tender Opened on May 12<sup>th</sup>, 2016**

## BID EVALUATION


Name of Bidder: M/s. Popular International


Documents required as per tender terms and condition	Remarks
Photo Copy of Bid security without amount.	Yes
Original tender purchase receipt	Yes
Receipts for alternate offer	N/A
Detail Technical data sheet of quoted items.	Yes
Original catalogue of quoted items.	Yes
Copy of quality certificates i.e. CE/FDA	Yes
Sole agency/distributorship certificate.	Yes
3 year validity of Sole agency certificate	No
One year business history with the Manufacturer	Yes
Proof of past import from Manufacturer, copy of AWB/BL/GD	No
certificate from the manufacturer that the offered machine / equipment being used in the country of origin	Yes
Copy of valid Professional Tax certificate	Yes
Copy of Income Tax Certificate	Yes
Copy GST Registration Certificate.	N/A
copy of valid PNRA certificate	Yes
List of users in Sindh and all over Pakistan.	Yes
Satisfactory report from the user, with for 2 years satisfactory report.	Yes
Proof for the availability of workshop & trained engineers	Responsive


  
Prof. Tariq Mehmood  
HOD, Radiology, JPMC (Member)


  
Dr. Umer Daraz  
Chief Pathologist, CHK (Member)


  
Dr. Robina Bashir  
Add. Medical Superintendent (Member)


  
Dr. Riazuddin Shaikh  
Cardiologist (Member)


  
Dr. Abdul Hameed Shaikh  
Pathologist (Member)

  
Dr. Abdul Razzaq Siddiqui  
Nephrologist (Member)

  
Dr. M. Murtaza Raza  
Radiologist (Member)

  
Dr. Abdul Jabbar  
Sr. Dental Surgeon (Member)

  
Dr. Shehnaz  
Pediatrician (Member)

  
Dr. Muhammad Tofique  
Medical Superintendent  
Sindh Govt. Hospital-Korangi, Karachi