

Bid Evaluation Report

1. Name of Procuring Agency: International Center for Chemical and Biological Sciences, University of Karachi, Karachi-75270.
2. Tender Reference No: Re-Tender Notice No. ICCBS-HEJ-EQPT-030315 (5TH TIME)
3. Tender Description/Name of work/item: Purchase/Import of Table Top Centrifuge
4. Method of Procurement: Single Stage – Single Envelope
5. Tender Published: Serial No. 22713; SPPRA website
(SPPRA ID No. & News papers names with dates)
6. Total Bid documents Sold; Three
7. Total Bids Received: Three
8. Technical Bid Opening date: (if applicable) N.A. (Provide details in separate form)
9. No. of Bid technically qualified (if applicable): N.A.
10. Bid(s) Rejected: Nil (2 rejected in technical evaluation)
11. Financial Bid Opening date: 03-03-2015 (5th time)

PURCHASE COMMITTEE, ICCBS INSTITUTIONS

[Signature]
1. PROF. DR. M. SHAIQ ALI
Chairman (I.C.C.B.S)

[Signature]
2. Dr. Farzana Shaheen
Member (I.C.C.B.S)

[Signature]
3. Dr. Sonia Siddiqui
Member (I.C.C.B.S)

Members from Institution

[Signature]
1. Syed Jehanzeb
Member (I.B.A)


[Signature]
2. Dr. Shaista Perveen
Member (NED University)


Members outside the Institution


12. Bid Evaluation Report:

S No	Name of Firm or Bidder	Cost offered by the Bidder	Ranking in terms of cost	Comparison with Estimated cost	Reasons for acceptance/rejection	Remarks
0	1	2	3	4	5	6
1.	Eco Biosciences International	Rs. 687,930/-	3 rd lowest (evaluated)	1.75% less	According to specs and within the estimated cost range	Recommended for placement of order
2.	Techno Medical Systems (Pvt.) Ltd.	Rs. 658,300/-	Invalid	Invalid	Invalid	Rejected
3.	Al-Rehman Enterprises	Rs. 610,000/-	Invalid	Invalid	Invalid	Rejected

PURCHASE COMMITTEE, ICCBS INSTITUTIONS



1. PROF. DR. M. SHAIQ ALI
Chairman (I.C.C.B.S)


2. Dr. Farsana Shaheen
Member (I.C.C.B.S)


3. Dr. Sonia Siddiqui
Member (I.C.C.B.S)

Members from Institution


1. Dr. Syed Shanzel
Member (I.C.C.B.S)


2. Dr. Shaista Perveen
Member (NED University)

Members outside the Institution

Signatures of the Members of the Committee


TECHNICAL COMMITTEE MINUTES OF TENDER NO ICCBS-HEJ-EQPT-030315
FOR PURCHASE OF CONSUMABLES

Technical committee meeting for Re-Tender No: ICCBS-HEJ-EQPT-030315 (5nd time) was held on 13-3-2015. Following recommendations were made:


TABLE TOP CENTRIFUGE

3 Bids were received from M/s ECO BIOSCIENCES, M/s AL-RAHMAN ENTERPRISES and M/s TECHNO MEDICAL SYSTEM.

Bids offered by M/s AL-RAHMAN ENTERPRISES and M/s TECHNO MEDICAL SYSTEM were rejected since the centrifuge offered by M/s AL-RAHMAN ENTERPRISES didn't fulfill the specification in terms of capacity (2-7 mL), the required capacity was 15 and 50 mL. The bid offered by M/s TECHNO MEDICAL SYSTEM was also rejected because the centrifuge offered had less speed range (10,115xg), whereas the requirement was much higher (20,913xg). After careful evaluation, committee has recommended that the tender may be awarded to M/s ECO BIOSCIENCES, Model # 5810, Eppendorf, Germany, as they are the only valid bidder among all other bidders and the quoted item is as per our specification.


Prof. Dr. Bina Shahven Siddiqui
(Convenor)


Dr. Atia-tul-Wahab
(End-User)


Mr. Sohail Ahmed
(Member)


Mr. Haris Qureshi
(Member)