




BID EVALUATION REPORT

SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI



BID EVALUATION REPORT

- | | |
|--|--|
| 1 Name of Procuring Agency | Civil Hospital Karachi - Health Department |
| 2 Tender Reference No. | PD-Trauma Centre-CHK/042 (Rev) 2015-16 |
| 3 Tender Description / Name of Work | Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period) of Medical Equipment of the project: Shaheed Mohtarma Benazir Bhutto Accident Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi |
| 4 Tender Item Number | 1 |
| 5 Name of Item | Diagnostic & Therapeutic Angiography System |
| 6 Quantity | 01 Unit |
| 7 Method of Procurement | Single Stage - Two Envelope Procedure |
| 8 Tender Published | Daily Dawn, Daily Jang, Daily Ibrat and Website of SPPRA & Civil Hospital Karachi |
| 9 Total Bid Documents Sold | Nine |
| 10 Total Bid Received | Two |
| 11 Technical Bid opening date (if applicable) | May 20, 2016 |
| 12 Number of Bid technically qualified (if applicable) | Two |
| 13 Number of Bid technically disqualified | Nil |
| 14 Financial Bid opening date | May 30, 2016 |
| 15 Bid Evaluation Report | |

S.No.	Name of Firm or Bidder	Cost Offered by Bidder	Ranking in Terms of Cost	Comparison with Estimated cost	Reasons for Acceptance	Remarks
1	2	3	4	5	6	7
1	M/S Medequips SMC	Rs. 64,794,135.00 each on CIF Basis & rates on Rs. 48,000,000/- on DDP basis	1st Lowest	Within overall cost of PC-1	On lowest basis	The Procurement Committee approved and accepted the offer of M/S Medequips SMC at the rate of Rs. 64,794,135.00 each on CIF basis. Four optional items at the total rate of Rs. 14,654,477.-.
2	M/S Siemens Healthcare (Pvt.) Ltd.	Rs. 67,306,347.00 each on CIF basis & Rs. 70,500,000.00 each on DDP basis	2nd Lowest			

Reasons of acceptance / rejection have been recorded in "Technical Evaluation Report".


Dr. Jamil Ahmed Siddiqui
 Medical Superintendent
 Civil Hospital Karachi


Associate Prof. (Dr.) Ehtesham Al-Ibrani
 Head of Department of Burns
 Civil Hospital Karachi


Engr. Hayat Kemal
 Project Director, SMBB Accident, Emergency &
 Trauma Centre and Ancillary Services Complex at
 Civil Hospital Karachi


Waris Khalil
 Divisional Accounts Officer
 A.G. Sindh Office


Abdul Karim Memon
 Mukhtarkar (Revenue), Garden Representative of
 Deputy Commissioner (South), Karachi

BID EVALUATION REPORT

- | | |
|--|--|
| 1 Name of Procuring Agency | Civil Hospital Karachi - Health Department |
| 2 Tender Reference No. | PD-Trauma Centre-CHK/042 (Rev) 2015-16 |
| 3 Tender Description / Name of Work | Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period) of Medical Equipment of the project Shaheed Mohtama Benazir Bhutto Accident Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi |
| 4 Tender Item Number | 2 |
| 5 Name of Item | Hand Held Fogging Machine with Motor Blower Nozzle |
| 6 Quantity | 10 Units |
| 7 Method of Procurement | Single Stage - Two Envelope Procedure |
| 8 Tender Published | Daily Dawn, Daily Jang, Daily Ibrat and Website of SPPRA & Civil Hospital Karachi |
| 9 Total Bid Documents Sold | Ninth |
| 10 Total Bid Received | One |
| 11 Technical Bid opening date (if applicable) | May 20, 2016 |
| 12 Number of Bid technically qualified (if applicable) | One |
| 13 Number of Bid technically disqualified | Nil |
| 14 Financial Bid opening date | May 30, 2016 |
| 15 Bid Evaluation Report | |

S.No.	Name of Firm or Bidder	Cost Offered by Bidder	Ranking in Terms of Cost	Comparison with Estimated cost	Reasons for Acceptance	Remarks
1	2	3	4	5	6	7
1	M/S Karachi Medical Company	Rs. 119,330.00 each on CIF Basis & rates on DDP basis were not offered	Single quoted bidder	Within overall cost of PC-1	On lowest basis	The Procurement Committee approved and accepted the offer of M/S Karachi Medical Company at the rate of Rs. 119,330.00 each on CIF basis.

Reasons of acceptance / rejection have been recorded in "Technical Evaluation Report".


Dr. Jamil Ahmed Siddiqui
 Medical Superintendent
 Civil Hospital Karachi


Associate Prof. (Dr.) Ehmer Al-Ibran
 Head of Department of Burns
 Civil Hospital Karachi


Engr. Hayat Kemal
 Project Director, SMBB Accident, Emergency &
 Trauma Centre and Ancillary Services Complex at
 Civil Hospital Karachi


Wazir Khalil
 Divisional Accounts Officer
 A.G. Sindhi Office


Abilul Karim Memon
 Mukhtiar (Revenue), Garden Representative of
 Deputy Commissioner (South), Karachi

BID EVALUATION REPORT

1	Name of Procuring Agency	Civil Hospital Karachi - Health Department
2	Tender Reference No.	PD-Trauma Centre-CHK/042 (Rev) 2015-16
3	Tender Description / Name of Work	Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period) of Medical Equipment of the project Shaheed Mohtarma Benazir Bhutto Accident Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi
4	Tender Item Number	3
5	Name of Item	Skin Graft Mesher
6	Quantity	02 Units
7	Method of Procurement	Single Stage - Two Envelope Procedure
8	Tender Published	Daily Dawn, Daily Jang, Daily Ibrat and Website of SPPRA & Civil Hospital Karachi
9	Total Bid Documents Sold	Ninth
10	Total Bid Received	One
11	Technical Bid opening date (if applicable)	May 20, 2016
12	Number of Bid technically qualified (if applicable)	One
13	Number of Bid technically disqualified	Nil
14	Financial Bid opening date	May 30, 2016
15	Bid Evaluation Report	

S.No.	Name of Firm or Bidder	Cost Offered by Bidder	Ranking in Terms of Cost	Comparison with Estimated cost	Reasons for Acceptance	Remarks
1	2	3	4	5	6	7
1	M/S Karachi Medical Company	Rs. 1,020,596.00 each on CIF Basis & rates on DDP basis were not offered	Single quoted bidder	Within overall cost of PC-1	On lowest basis	The Procurement Committee approved and accepted the offer of MS Karachi Medical Company at the rate of Rs. 1,020,596.00 each on CIF basis.

Reasons of acceptance / rejection have been recorded in "Technical Evaluation Report".


Dr. Jamil Ahmed Siddiqui
 Medical Superintendent
 Civil Hospital Karachi


Associate Prof. (Dr.) Ehmer Al-Ibran
 Head of Department of Burns
 Civil Hospital Karachi


Engr. Hayat Kemal
 Project Director, SM/BB Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi


Waris Khalil
 Divisional Accounts Officer
 A.G. Sindh Office


Abdul Karim Memon
 Mukhtiar (Revenue), Garden Representative of Deputy Commissioner (South), Karachi

BID EVALUATION REPORT

- | | |
|--|---|
| 1 Name of Procuring Agency | Civil Hospital Karachi - Health Department |
| 2 Tender Reference No. | PD-Trauma Centre-CHK/042 (Rev) 2015-16 |
| 3 Tender Description / Name of Work | Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period) of Medical Equipment of the project Shaheed Mohtarma Benazir Bhutto Accident Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi |
| 4 Tender Item Number | 4 |
| 5 Name of Item | Dermatome Battery Operated |
| 6 Quantity | 03 Units |
| 7 Method of Procurement | Single Stage - Two Envelope Procedure |
| 8 Tender Published | Daily Dawn, Daily Jang, Daily Ibrai and Website of SPPRA & Civil Hospital Karachi |
| 9 Total Bid Documents Sold | Nine |
| 10 Total Bid Received | Three |
| 11 Technical Bid opening date (if applicable) | 20-May-16 |
| 12 Number of Bid technically qualified (if applicable) | Two |
| 13 Number of Bid technically disqualified | One |
| 14 Financial Bid opening date | May 30, 2016 |
| 15 Bid Evaluation Report | |

S.No.	Name of Firm or Bidder	Cost Offered by Bidder	Ranking in Terms of Cost	Comparison with Estimated cost	Reasons for Acceptance	Remarks
1	2	3	4	5	6	7
1	M/S Karachi Medical Company (Original Offer / Humeca)	Rs. 1,665,980.00 each on CIF Basis & Rs. 1,667,000.00 each on DDP Basis	Single quoted bid	Within overall cost of 2C-1	On lowest basis	The Procurement Committee approved and accepted the offer of M/S Karachi Medical Company at the rate of Rs. 1,665,980.00 each on CIF basis. It has also decided that Two units of Original Offer (Humeca) & One unit of Alternate Offer (B Braun) to be procured.
2	M/S Karachi Medical Company (Alternate Offer / B. Braun)					

Reasons of acceptance / rejection have been recored in "Technical Evaluation Report".



Dr. Jamil Ahmed Siddiqui
Medical Superintendent
Civil Hospital Karachi



Associate Prof. (Dr.) Ehmer Al-Ibran
Head of Department of Burns
Civil Hospital Karachi



Engr. Hayat Kemal
Project Director, SMBB Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi



Waris Khalid
Divisional Accounts Officer
A.G. Sindh Office



Abdul Karim Memon
Mukhtiar (Revenue), Garden Representative of Deputy Commissioner (South), Karachi

MINUTES OF THE PROCUREMENT COMMITTEE

SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

Meeting of the Procurement Committee of Trauma Centre Civil Hospital Karachi was held on **Monday May 30, 2016** at 11:30 a.m. in the Board Room of Civil Hospital Karachi. Dr. Jamil Ahmed Siddiqui, Medical Superintendent Civil Hospital Karachi chaired the meeting. The financial bids were opened after finalization of technical bids by the Procurement Committee for procurement of Machinery/Equipment under Tender for Supplying, Installation, Commissioning and Maintenance (During Maintenance Period) of **Medical Equipment & Ancillaries** of the project Shaheed Mohtarma Benazir Bhutto, Accident, Emergency & Trauma Center and Ancillary Services Complex at Civil Hospital Karachi for financial year **2015-2016**. The technical bids of tender enquiry number **PD-Trauma Centre-CHK/042 (Rev) 2015-16** were opened on Friday May 20, **2016 at 11:30 a.m.** and subsequently approved by the Procurement Committee for procurement of the required items.

- The following members attended the meeting:

01	Dr. Jamil Ahmed Siddiqui Medical Superintendent, Civil Hospital Karachi	Chairperson
02	Associate Prof. (Dr.) Ehmer Al-Ibran Head of Department of Burns, Civil Hospital Karachi	Member
03	Engr. Hayat Kemal Project Director, SMBB Trauma Centre, Civil Hospital Karachi	Member
04	Mr. Waris Khalil Divisional Accounts Officer, A.G. Sindh Office	Member
05	Mr. Abdul Karim Memon Mukhtiarkar (Revenue) Garden / Representative of Deputy Commissioner Karachi (South)	Member

EVALUATION OF FINANCIAL PROPOSAL:

Initiating the discussions, this office presented the evaluated Technical Data Sheets for perusal of the Procurement Committee (Copy of Technical Evaluation Report is enclosed at A) and subsequently the Financial Bids of Technically Qualified Bidders were opened as under:

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
Item No. : 1
Name of Item : **Diagnostic & Therapeutic Angiography System**
Quantity Required : 01 Unit
Estimated Price : Rs: 56,144,185/- each

Name of Firms	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
Make	Toshiba Medical Systems Corporation / Fysiceon	Siemens AG
Model	Infx 8000C (Infinix I Series)	Artis Zee (Ceiling Mounted)

The offers of M/S Medequips SMC and M/S Siemens Healthcare (Pvt.) Ltd. meet all the required specification as per tender requisite. The Procurement Committee decided to proceed with the opening of financial proposals of qualified firms.

M/S Medequip SMC offered the rate US\$ 619,000 /- equivalent to PKR 64,794,135/- on C&F basis and Rs: 68,000,000/- on DDP basis. Moreover, following optional items were also quoted;

S. No	Item	Price in C&F	PKR Equivalence	DDP
1	Advance dose tracking system kit XIDF-DTS802	US\$35,000	3,663,643	3,850,000
2.	STEPPING DSA	US\$30,000	3,140,265	3,300,000
3.	Advanced Spot Fluoroscopy	US\$35,000	3,663,643	3,850,000
4.	Additional Workstation	US\$35,000	3,663,643	3,850,000
5.	Multimodality Roadmap for CT and MRI Integration	US\$35,000	3,663,643	3,850,000

M/S Medequip SMC offered following Post warranty service Contract:

S. No	Item	Amount	PKR Equivalence
1.	Without Parts	US\$ 15,000 per year (Payable quarterly in equivalent PKR at the time of billing)	1,570,133
2.	With All Parts Except Tube	US\$ 60,000 per year (Payable quarterly in equivalent PKR at the time of billing)	6,280,530



3.	With all parts including Tube	US\$ 80,000 per year (Payable quarterly in equivalent PKR at the time of billing)	8,374,040
4.	Cost of X-Ray Tube	US\$ 115,000 with standard Toshiba's warranty	12,037,683

M/S Siemens Healthcare (Pvt.) Ltd. offered the rate US\$ 643,000/- equivalent to PKR 67,306,347/- on C&F basis and Rs: 70,500,000/- on DDP basis.

The Procurement Committee **approved and accepted** the offer of M/S Medequip SMC at the rate of Rs. 64,794,135/- each on C&F basis for the basic unit. Moreover, following optional items were also approved on C & F basis at below mentioned PKR equivalence.

S. No	Items	Price in C&F	PKR Equivalence
1.	Advance dose tracking system kit XIDF-DTS802	US\$35,000	3,663,643.00
2.	Advanced Spot Fluoroscopy	US\$35,000	3,663,643.00
3.	Additional Workstation	US\$35,000	3,663,643.00
4.	Multimodality Roadmap for CT and MRI Integration	US\$35,000	3,663,643.00
TOTAL OPTIONAL ITEMS			14,654,572.00

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
Item No. : 2
Name of Item : **Hand Held Fogging Machine with Motor Blower Nozzle Technology**
Quantity Required : 10 Units
Estimated Price : Rs: 125,000/- each

Name of Firms	M/S Karachi Medical Company
Make	Fogmaster Corporation
Model	NOZ-L-JET 7807

The offer of M/S Karachi Medical Company meets all the required specification as per tender requisite. The Procurement Committee decided to proceed with the opening of financial proposals of qualified firm.

M/S Karachi Medical Company offered the rate US\$ 1,140/- equivalent to PKR 119,330/- on C&F basis while rate on DDP basis was not quoted.

The Procurement Committee **approved and accepted** the offer of M/S Karachi Medical Company at the rate of Rs. 119,330/- each on C&F basis.

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. : 3
 Name of Item : **Skin Graft Mesher**
 Quantity Required : 02 Units
 Estimated Cost : Rs: 1,800,000/- each

Name of Firms	M/S Karachi Medical Company
Make	Brennen Medical (Now) Molnlycke Healthcare
Model	Brenin Skin Graft Mesher

The offer of M/S Karachi Medical Company meets all the required specification as per tender requisite. The Procurement Committee decided to proceed with the opening of financial proposals of qualified firm.

M/S Karachi Medical Company offered the rate US\$ 9,750/- each equivalent to PKR 1,020,596/- on C&F basis while rate on DDP basis was not quoted.

The Procurement Committee **approved and accepted** the offer of M/S Karachi Medical Company at the rate of Rs. 1,020,596/- each on C&F basis.

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. : 4
 Name of Item : **Dermatome Battery Operated**
 Quantity Required : 03 Units
 Estimated Price : Rs: 1,300,000/-

Name of Firms	M/S Oriental Sales Corporation	M/S Karachi Medical Company (Original Offer)	M/S Karachi Medical Company (Alternate Offer)
Make	Integra LifeSciences Corporation	Humecca BV	Aesculap / B.Braun
Model	Integra Padgett Dermatomes Model S	D80	Acculan 3Ti

The offers of M/S Karachi Medical Company (Original Offer) and M/S Karachi Medical Company (Alternate Offer) meet all the required specification as per tender requisite. The offer of M/S Oriental Sales Corporation does not meet the required basic specification as per tender requisite hence declared rejected. The Procurement Committee decided to proceed with the opening of financial proposals of qualified firms.

M/S Karachi Medical Company offered the rate US\$ 15,915.66/- each equivalent to PKR 1,665,980/- on C&F basis and Rs: 1,667,000/- each on DDP basis for both offers.

The Procurement Committee **approved and accepted** the offer of M/S Karachi Medical Company at the rate of Rs. 1,665,980/- each on C&F basis. It has also decided that two units of original offer and one unit of alternate offer to be procured.

Dr. Jamil Ahmed Siddiqui
Medical Superintendent
Civil Hospital Karachi

Associate Prof. (Dr.) Ehsan Al-Ibran
Head of Department of Burns
Civil Hospital Karachi

Engr. Hayat Kemal
Project Director

Waris Khalil
Divisional Accounts Officer
A.G. Sindh Office

Abdul Karim Memon
Mukhtiarkar (Revenue), Garden / Representative
of Deputy Commissioner (South), Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

FINANCIAL COMPARATIVE STATEMENT

Tender No. : PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Date of Opening of Financial Proposals : May 30, 2016
 Tender Item No. : 1
 Name of Item : Diagnostic & Therapeutic Angiography System
 Quantity : 01 Unit
 Estimated Cost / PC-1 Cost : Rs.56,144,185.00

S.No.	Name of Qualified Firm	Quoted Amount				Earnest Money
		CIF Price		DDP Price		
		Unit Price	Total Price	Unit Price	Total Price	
1	M/S Medequips SMC	USD 619,000.00	USD 619,000.00	68,000,000.00	68,000,000.00	Pay Order No. 02314495 dated 27-04-2016 amounting to Rs.950,000.00 of Meezan Bank (0104), Gulshan-e-Iqbal Branch, Karachi
2	M/S Siemens Healthcare (Pvt.) Ltd.	USD 643,000.00	USD 643,000.00	70,500,000.00	70,500,000.00	Pay Order No. 02314495 dated 27-04-2016 amounting to Rs.950,000.00 of Meezan Bank (0104), Gulshan-e-Iqbal Branch, Karachi

Handwritten signatures and initials in blue and green ink.

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

FINANCIAL COMPARATIVE STATEMENT

Tender No. : PD-Trauma Centre-CHK-042 (Rev) 2015-16
 Date of Opening of Financial Proposals : 30-May-16
 Tender Item No. : 2
 Name of Item : Hand Held Fogging Machine with Motor Blower Nozzle
 Quantity : 10 Units
 Estimated Cost / PC-I Cost : Rs. 125,000.00

S.No.	Name of Qualified Firm	Quoted Amount				Earnest Money
		CIF Price		DDP Price		
		Unit Price	Total Price	Unit Price	Total Price	
1	M/S Karachi Medical Company	USD 1,140.00	USD 11,400.00	Not Quoted	Not Quoted	Pay Order No. 02314495 dated 27-04-2016 amounting to Rs.950,000.00 of Meezan Bank (0104), Gulshan-e-Iqbal Branch, Karachi

(Handwritten signatures in green and blue ink)

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

FINANCIAL COMPARATIVE STATEMENT

Tender No. : PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Date of Opening of Financial Proposals : 30-May-16
 Tender Item No. : 3
 Name of Item : Skin Graft Mesher
 Quantity : 02 Units
 Estimated Cost / PC-I Cos: : Rs. 1,800,000.00

S.No.	Name of Qualified Firm	Quoted Amount				Earnest Money
		CIF Price		DDP Price		
		Unit Price	Total Price	Unit Price	Total Price	
1	M/S Karachi Medical Company	USD 9,750.00	USD 19,500.00	Not Quoted	Not Quoted	Pay Order No. 02314495 dated 27-04-2016 amounting to Rs.950,000.00 of Meezan Bank (0104), Gulshan-e-Iqbal Branch, Karachi

(Handwritten signatures in blue and green ink)

**SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF
MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY &
TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI**

FINANCIAL COMPARATIVE STATEMENT

Tender No. : PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Date of Opening of Financial Proposals : 30-May-16
 Tender Item No. : 4
 Name of Item : **Dermatome Battery Operated**
 Quantity : 03 Units
 Estimated Cost / PC-I Cost : Rs. 1,300,000.00


S.No.	Name of Qualified Firm	Quoted Amount				Earnest Money
		CIF Price		DDP Price		
		Unit Price	Total Price	Unit Price	Total Price	
1	M/S Karachi Medical Company (Original Offer / Humecca)	USD 15,916.66	USD 47,746.98	Rs. 1,667,000.00	Rs: 5,001,000.00	Pay Order No. 02314495 dated 27-04-2016 amounting to Rs:950,000.00 of Meezan Bank (0104), Gulshan-e-Iqbal Branch, Karachi
2	M/S Karachi Medical Company (Alternate Offer / B. Braun)	USD 15,916.66	USD 47,746.98	Rs. 1,667,000.00	Rs: 5,001,000.00	

(Handwritten signatures in green and blue ink)

TECHNICAL EVALUATION REPORT

NAME OF WORK: SUPPLYING, INSTALLATION, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO, ACCIDENT, EMERGENCY & TRAUMA CENTER AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

- The tenders for the above work were called on the basis of Single Stage Two Envelope System through press and appeared in Daily Dawn dated 02-05-2016, Daily Jang dated 02-05-2016, Daily Ibrat dated 01-05-2016 (copies enclosed).
- The advertisement also appeared on SPPRA website at serial No. 28147 dated 05-05-2016 and Website of Civil Hospital Karachi at Serial No. 01 dated 04-05-2016 (copies enclosed).
- In response of above, following Eight prospective bidders purchased tender document till closing date.
 1. M/S Medequips SMC
 2. M/S Siemens Healthcare (Pvt.) Ltd.
 3. M/S Oriental Sales Corporation
 4. M/S Karachi Medical Company
 5. M/S Faraz Associates
 6. M/S DWP Technologies (Pvt.) Ltd.
 7. M/S Shirazi Trading
 8. M/S Karachi Medical Company (Alternate Offer)
- M/S Saad Sales Services uploaded the tender documents from the website and submit the pay order of tender fee with Technical Proposals on 20-05-2016.
- The Technical Proposals were opened by the Procurement committee on 20-05-2016 (Copies of Attendance Sheets are enclosed). The following six bidders submitted their bids:
 1. M/S Medequips S.M.C
 2. M/S Siemens Healthcare (Pvt.) Ltd.



3. M/S Oriental Sales Corporation
4. M/S Karachi Medical Company
5. M/S Karachi Medical Company (Alternate Offer)
6. M/S Saad Sales Services

- The Technical Offers examined and evaluated as per the Technical Evaluation Criteria provided in the bidding documents and verified the documents and bid security submitted by the bidders (Technical Evaluation Sheets are enclosed).

Following is the position of Technical Evaluation:-

1. M/S Medequips S.M.C secured 85 out of 100 marks.
2. M/S Siemens Healthcare (Pvt.) Ltd. secured 70 out of 100 marks.
3. M/S Oriental Sales Corporation secured 70 out of 100 marks.
4. M/S Karachi Medical Company secured 90 out of 100 marks.
5. M/S Karachi Medical Company (Alternate Offer) secured 90 out of 100 marks.
6. M/S Saad Sales Services secured 15 out of 100 marks.

- **The passing marks are 70 out of 100.**

- Following firms were **qualified** in the Technical Evaluation:

1. M/S Medequips S.M.C
2. M/S Siemens Healthcare (Pvt.) Ltd.
3. M/S Oriental Sales Corporation
4. M/S Karachi Medical Company
5. M/S Karachi Medical Company (Alternate Offer)

- Following Firms were **disqualified** in the Technical Evaluation:

1. M/S Saad Sales Services

- **The Specifications Compliance / Bids Evaluations of Technically Qualified Firms** were evaluated as per Bid Evaluation Criteria provided in the Bidding Documents (Bid Evaluation Sheets are enclosed). Following is the position of Bids:-

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
Item No. : 1
Name of Item : **Diagnostic & Therapeutic Angiography System**
Quantity Required : 01 Unit



Name of Firms	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
Make	Toshiba Medical Systems Corporation / Fysicon	Siemens AG
Model	Infx 8000C (Infinix I Series)	Artis Zee (Ceiling Mounted)

- The offers of M/S Medequips SMC and M/S Siemens Healthcare (Pvt.) Ltd. meet all the required specification as per tender requisite.
- **The Procurement Committee decided to proceed with the opening of financial proposals of qualified firms.**

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
Item No. : 2
Name of Item : **Hand Held Fogging Machine with Motor Blower Nozzle Technology**
Quantity Required : 10 Units

Name of Firms	M/S Karachi Medical Company
Make	Fogmaster Corporation
Model	NOZ-L-JET 7807

- The offer of M/S Karachi Medical Company meets all the required specification as per tender requisite.
- **The Procurement Committee decided to proceed with the opening of financial proposals of qualified firm.**

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
Item No. : 3
Name of Item : **Skin Graft Mesher**
Quantity Required : 02 Units


Name of Firms	M/S Karachi Medical Company
Make	Brennen Medical (Now) Molnlycke Healthcare
Model	Brenin Skin Graft Mesher


- The offer of M/S Karachi Medical Company meets all the required specification as per tender requisite.
- **The Procurement Committee decided to proceed with the opening of financial proposals of qualified firm.**

Tender Enquiry No : PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. : 4
 Name of Item : **Dermatome Battery Operated**
 Quantity Required : 03 Units


Name of Firms	M/S Oriental Sales Corporation	M/S Karachi Medical Company (Original Offer)	M/S Karachi Medical Company (Alternate Offer)
Make	Integra LifeSciences Corporation	Humeca BV	Aesculap / B.Braun
Model	Integra Padgett Dermatomes Model S	D80	Acculan 3Ti

- The offers of M/S Karachi Medical Company (Original Offer) and M/S Karachi Medical Company (Alternate Offer) meet all the required specification as per tender requisite.
- The offer of M/S Oriental Sales Corporation does not meet all the required specification as per tender requisite hence declared rejected.
- **The Procurement Committee decided to proceed with the opening of financial proposals of qualified firms.**


Dr. Jamil Ahmed Siddiqui
 Medical Superintendent
 Civil Hospital Karachi


Associate Prof. (Dr.) Ehmer Al-Ibran
 Head of Department of Burns
 Civil Hospital Karachi


Engr. Hayat Kemal
 Project Director


Waris Khalil
 Divisional Accounts Officer
 A.G. Sindh Office


Abdul Karim Memon
 Mukhtiarkar (Revenue), Garden / Representative
 of Deputy Commissioner (South), Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

LIST OF QUALIFIED FIRMS
(As per Technical & Bid Evaluation Criteria given in Bidding Documents)


S. NO.	DESCRIPTION OF ITEM	BIDDERS	TECHNICALLY QUALIFIED BIDDERS	REMARKS
1	Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period) of Medical Equipment for the project Shaheed Mohtarma Benazir Bhutto, Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi	1) M/S Medequips S.M.C 2) M/S Siemens Healthcare (Pvt.) Ltd. 3) M/S Oriental Sales Corporation 4) M/S Karachi Medical Company 5) M/S Karachi Medical Company (Alternate Offer) 6) M/S Saad Sales Services	1) M/S Medequips S.M.C 2) M/S Siemens Healthcare (Pvt.) Ltd. 3) M/S Karachi Medical Company 4) M/S Karachi Medical Company (Alternate Offer)	


Dr. Jamil Ahmed Siddiqui
Medical Superintendent
Civil Hospital Karachi


Warris Khalil
Divisional Accounts Officer
A.G. Sindh Office


Associate Prof. Ehnar Al-Ibrani
Head of Department of Burns
Civil Hospital Karachi


Engr. Hayat Kemal
Project Director, SMIBB Trauma Centre
Civil Hospital Karachi


Abdul Karim Memon
Mukhtiarkar (Revenue) Garden / Representative of
Deputy Commissioner (South) Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

LIST OF TECHNICALLY QUALIFIED FIRMS

(As per Technical Evaluation Criteria given in Bidding Documents)

S. NO.	DESCRIPTION OF ITEM	BIDDERS	TECHNICALLY QUALIFIED BIDDERS	REMARKS
1	Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period) of Medical Equipment for the project Shaheed Mohtarma Benazir Bhutto, Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi	1) M/S Medequips S.M.C 2) M/S Siemens Healthcare (Pvt.) Ltd. 3) M/S Oriental Sales Corporation 4) M/S Karachi Medical Company 5) M/S Karachi Medical Company (Alternate Offer) 6) M/S Saad Sales Services	1) M/S Medequips S.M.C 2) M/S Siemens Healthcare (Pvt.) Ltd. 3) M/S Oriental Sales Corporation 4) M/S Karachi Medical Company 5) M/S Karachi Medical Company (Alternate Offer)	

Mirza

Dr. Jamil Ahmed Siddiqui
Medical Superintendent
Civil Hospital Karachi

Waris Khalil

Waris Khalil
Divisional Accounts Officer
A.G. Sindh Office

Engr. Hayat Kemal

Associate Prof. Enmar Al-Ibram
Head of Department of Burns
Civil Hospital Karachi

Abdul Karim Memon

Abdul Karim Memon
Mukhtiar (Revenue) Garden / Representative of
Deputy Commissioner (South) Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF
 MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND
 ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI
 TECHNICAL EVALUATION SHEET

S. No.	Essential Requirements / Evaluation Criteria	Qualifying Marks	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.	M/S Oriental Sales Corporation
MANDATORY REQUIREMENTS:					
1	Certificates of Registration with Income Tax & Sales Tax Departments		Attached	Attached	Attached
2	Affidavit regarding Blacklisting & Litigation		Attached	Attached	Attached on Letterhead
TECHNICAL EVALUATION:					
1	Establishment of Firms since a. More than 3 years - 05 marks b. More than 2 years - 10 marks As per attached Proforma "A"	10	Registered with FBR on 16-12-2005 as Company 10 More than 23 Institutes as per work orders attached but Proforma not attached 10	Registered with FBR on 26-02-2007 as Company 10 6 Institutions 10	Registered with FBR on 16-10-2002 for Non specialized wholesale trade 10 3 Institution 5
2	Major Institution Served during last year a. Less than 5 Institutions - 05 marks b. Up to or more than 5 Institutions - 10 marks As per attached Proforma "B"	10	32 work orders since 2010 but Proforma not attached 30	6 works since 2009 30	6 works since 2014 20
3	Previous Performance of similar works a. More than 3 years - 30 marks b. Up to 3 years - 20 marks As per attached Proforma "C"	30	10 of 11 employees including 4 Engineers with company profile	List of 26 Technical Staff and list of tools attached	List of 25 employees & 5 engineers with undertaking for workshop attached
4	Man Power Requirement 1. Must have proper well established workshop (Procurement Committee may visit the same) 2. Must have certified technical staff (Certificate must be attached) with other staff on ledger (List of persons be attached)	10 (5 marks for each)	5	10	10
5	Bank Certificate Bank Certificate that the bidder can perform business upto Rs. 25.00 million or above	10	Not attached 0	Not attached 0	Not attached 0






SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF
MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND
ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI
TECHNICAL EVALUATION SHEET

S. No.	Essential Requirements / Evaluation Criteria	Qualifying Marks	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.	M/S Oriental Sales Corporation
6	Financial Soundness of the Firm The bidder should submit following: a. Bank Statement for last 3 years with Bank Letter b. Audited Balance Sheet & Income Tax Returns for last 3 years	20 (10 marks for each)	1 Attached 20	2 Bank Statement of March 2016 is attached with Financial Statement of 3 years on their own letter head 0	3 2 years audited balance sheet & 3 years bank statement attached 15
7	Questionnaire / Proforma (Mandatory) Page 10 & 11 of the Bidding Documents Volume I	10 70	Attached 10	Attached 10	Attached 10
QUALIFYING MARKS			85	70	70
TOTAL MARKS		100			

NOTE:

- 1) M/S Medequips SMC secured 85 marks and stand qualified
- 2) M/S Siemens Healthcare (Pvt.) Ltd. secured 70 marks and stand qualified.
- 3) M/S Oriental Sales Corporation secured 70 marks and stand qualified.



Dr. Jamil Ahmed Siddiqui
Medical Superintendent
Civil Hospital Karachi


Warris Khalid

Divisional Accounts Officer
A.G. Sindh Office



Associate Prof. Eihmar Al Ibran
Head of Department of Burns
Civil Hospital Karachi



Engr. Hayat Kemal
Project Director, SMBB Trauma Centre Civil
Hospital Karachi


Abdul Karim Memon
Mukhtarkar (Revenue) Garden / Representative of
Deputy Commissioner (South) Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF
 MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND
 ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI
 TECHNICAL EVALUATION SHEET

S. No.	Essential Requirements/ Evaluation Criteria	Qualifying Marks	M/S Karachi Medical Company (Original & Alternate Offer)	M/S Saad Sales Services
			4	5
MANDATORY REQUIREMENTS:				
1	Certificates of Registration with Income Tax & Sales Tax Departments		Attached	Not attached
2	Affidavit regarding Blacklisting & Litigation		Attached	Affidavit for Litigation is not attached
TECHNICAL EVALUATION:				
1	Establishment of Firms since a. More than 3 years - 05 marks b. More than 5 years - 10 marks As per attached Proforma "A"	10	Registered with FBR on 22-11-2001 as Non-Specialized wholesale trade Distributor, Importer	Not attached
2	Major Institution Served during last year a. Less than 5 institutions - 05 marks b. Up to or more than 5 institutions - 10 marks As per attached Proforma "B"	10	8 Institute	Not attached
3	Previous Performance of similar works a. More than 3 years - 30 marks b. Up to 3 years - 20 marks As per attached Proforma "C"	30	More than 43 works since 2011	Not attached
4	Man Power Requirement 1. Must have proper well established workshop (Procurement Committee may visit the same) 2. Must have certified technical staff (Certificate must be attached) with other staff on ledger (List of persons be attached)	10 (5 marks for each)	List of 30 employees and tool kit attached	Not attached
5	Bank Certificate Bank Certificate that the bidder can perform business upto Rs. 25.00 million or above.	10	Not attached	Bank Letter attached
			0	10



**SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF
MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND
ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI**

TECHNICAL EVALUATION SHEET

S. No.	Essential Requirements/ Evaluation Criteria	Qualifying Marks	M/S Karachi Medical Company (Original & Alternate Offer)	M/S Saad Sales Services
6	Financial Soundness of the Firm The bidder should submit following: a. Bank Statement for last 3 years with Bank Letter b. Audited Balance Sheet & Income Tax Returns for last 3 years.	20 (10 marks for each)	4 Attached 20 Attached	5 Income Tax Returns from 2012 to 2014 attached 5 Blank Proforma attached
7	Questionnaire / Proforma (Mandatory) Page 10 & 11 of the Bidding Documents Volume I	10	10 Attached	0 Blank Proforma attached
QUALIFYING MARKS		70	90	15
TOTAL MARKS		100		

NOTE:

4) M/S Karachi Medical Company Secured 90 marks they submit receipt of original and alternate offer and stand qualified.

5) M/S Saad Sales Services secured 15 marks and stand disqualified.



Dr. Jamil Ahmed Siddiqui
Medical Superintendent
Civil Hospital Karachi



Waris Khatib
Divisional Accounts Officer
A.G. Sindh Office



Associate Prof. Ehsan Al-Ibraan
Head of Department of Burns
Civil Hospital Karachi



Engr. Hayat Kemal
Project Director, SMBB Trauma Centre
Civil Hospital Karachi

Abdul Karim Memon
Mukhtarkar (Revenue) Garden / Representative of
Deputy Commissioner (South) Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAIN TENANCE PERIOD) OF MEDICAL EQUIPMENT & ANCILLARIES FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

LIST OF QUALIFIED FIRMS



(As per Bid Evaluation Criteria given in Bidding Documents)

S. NO.	DESCRIPTION OF ITEM	BIDDERS	TECHNICALLY QUALIFIED BIDDERS	REMARKS
1	Diagnostic & Therapeutic Angiography System	1) M/S Medequips S.M.C 2) M/S Siemens Healthcare (Pvt.) Ltd.	1) M/S Medequips S.M.C 2) M/S Siemens Healthcare (Pvt.) Ltd.	Both Bidders Qualified
2	Hand Held Fogging Machine with Motor Blower Nozzle Technology	1) M/S Karachi Medical Company	1) M/S Karachi Medical Company	Single Bidder Quoted & Qualified
3	Skin Graft Mesher	1) M/S Karachi Medical Company	1) M/S Karachi Medical Company	Single Bidder Quoted & Qualified
4	Dermatome Battery Operated	1) M/S Oriental Sales Corporation 2) M/S Karachi Medical Company 3) M/S Karachi Medical Company (Alternate Offer)	1) M/S Karachi Medical Company 2) M/S Karachi Medical Company 2) M/S Karachi Medical Company (Alternate Offer)	Both offers of single bidder Qualified


Dr. Jamil Ahmed Siddiqui
 Medical Superintendent
 Civil Hospital Karachi

 Divisional Accounts Officer
 A.C. Sindh Office


Dr. Ehter Al-Ibran
 Head of Department of Burns
 Civil Hospital Karachi



Engr. Hayat Kenal
 Project Director, SMH Trauma Centre
 Civil Hospital Karachi

Abdul Karim Afton
 Mukhtariakar (Revenue) Garden / Representative of
 Deputy Commissioner (South) Karachi

SUPPLY, INSTALLATION, TESTING, COMMISSIONING AND MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT & ANCILLARIES FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL, KARACHI

BID EVALUATION SHEET

(As per Bid Evaluation Criteria given in Bidding Documents)


S. No.	Bid Evaluation Criteria	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd	M/S Oriental Sales Corporation	M/S Karachi Medical Company (original offer)	M/S Karachi Medical Company (Alternate offer)
a.	Specifications Compliance	100%	100%	0%	95%	80%
b.	After sale service facility (the vendors should submit an undertaking on Non-Judicial Stamp Paper of Rs. 50/- alongwith Technical Proposals for the same).	Yes	Yes	Yes	Yes	Yes
c.	Authority letter from Manufacturer where applicable	Yes	Yes	Yes	Yes	Yes
d.	Guarantee of availability of parts of the supply articles for 10 years	Yes	Yes	Yes	Yes	Yes
e.	Compliance of terms & Conditions and Additional Terms & Conditions of the Bid Form	Yes	Yes	Yes	Yes	Yes


Dr. Jamil Ahmed Siddiqui
 Medical Superintendent
 Civil Hospital Karachi


Associate Prof. Emar Al-Ibran
 Head of Department of Burns
 Civil Hospital Karachi


Engr. Hayat Kemal
 Project Director, SMIBB Trauma Centre
 Civil Hospital Karachi


Waris Khan
 Divisional Accounts Officer
 A.C. Sindh Office


Abdul Karim Memon
 Mukhtiar (Revenue) Garden / Representative of
 Deputy Commissioner (South) Karachi

Supply, Installation, Testing, Commissioning & Maintenance (during maintenance period) of **Medical Equipment** for the project **Shahheed Mohitarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi**

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No.
 Item No.
 Name of Item
 Quantity Required

PD-Trauma Centre-CHK/042 (Rev) 2015-16
 I
Diagnostic & Therapeutic Angiography System
 01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
Make	Toshiba Medical Systems Corporation / Fysicon	Siemens AG
Model	Infx 8000C (Infinix I Series)	Artis Zee (Ceiling Mounted)
Diagnostic & Therapeutic Angiography System		
Single Plane General Angiography / Vascular Interventional System for Trauma work:		
<ul style="list-style-type: none"> A fully digital flat panel single plane General angiography system dedicated for diagnostic & interventional and peripheral procedures with the following minimum specifications and application softwares. The quoted model must be latest and bidder must mention the number of systems installed worldwide and in Pakistan. The local vendor must have sold angiography equipment in the province and submit letter of satisfaction of similar general / vascular angio-systems installed by them regarding satisfactory after sales service performance from three reputable public / private centers of Sindh. This is a knockout clause. 	Yes Yes 75-Pakistan 25-Karachi	Yes Yes 42-Pakistan 18-Karachi
S.No.	Item details - model and specifications	
I POSITIONING ARM:		
Ceiling suspended arm		
1.2	The system should be ceiling mounted with capability to do head to toe patient imaging without moving the patient	
1.3	Real time display of rotation angulations.	
	Yes	Yes
	Yes	Yes

(Handwritten signatures and initials in blue and green ink)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No.

Item No.

Name of Item

Quantity Required

PD-Trauma Centre-CHK/042 (Rev) 2015-16

1

Diagnostic & Therapeutic Angiography System

01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
1.4 Geometry: C-arm / G-arm	Yes	Yes
1.5 RAO / LAO +/- 120 ° or more.	C-Arm	C-Arm
1.6 Cranial / Caudal: Min. +/-45° in all detector positions	Yes	Yes
1.7 Rotation Speed: 40°/sec or more in LAO / RAO.	Yes	Yes
1.8 Isocentric Height: Variable / Fixed.	Yes	Yes
1.9 Auto Positioning: Programmable auto positioning of selected.	Yes	Yes
1.10 Angulations, (50 or more) programmable positions.)	Yes	Yes
1.11 The control panel can be mounted at any side of the patient table.	Yes	Yes
1.12 All the rotational / angles should be digitally displayed	Yes	Yes
1.13 Motorized / manual parking / rotation of the positioning arm.	Yes	Yes
2 Digital Flat Panel Detector 12 Inch x 16 Inch	Yes	Yes
2.1 Single plane C-Arm / G-Arm	Yes	Yes
2.2 Image matrix of 1024 x 1024 x 14 bits or more.	Yes	Yes
2.3 30x40cm large field of view - Rectangular Panel is required	Yes	Yes
2.4 Built in temperature stabilizer.	Yes	Yes
2.5 Integrated collision protection feature.	Yes	Yes
2.6 All other standard accessories according to this digital flat panel.	Yes	Yes
2.7 Dose management with floro filters	Yes	Yes
2.8 Pixel Size of 200 um or better.	194	184
3 Patient Support/ Table:		
3.1 Catheterization Table	Yes	Yes
3.2 Floor mounted with up/down, longitudinal and transverse movements	Yes	Yes
3.3 Original table from manufacturer	Yes	Yes
3.4 Longitudinal Stroke: 1000-1500 mm.	Yes	Yes
3.5 Lateral stroke: 100-300 mm	Yes	Yes
3.6 CPR in any table position.	Yes	Yes

Handwritten signatures and initials in blue and green ink.

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16

1

Name of Item Diagnostic & Therapeutic Angiography System

Quantity Required 01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
3.7 All table side movement controls.	Yes	Yes
3.8 Table top should be of such construction in material and durability to accept patient's weight of not less than 150 kg plus 100 kg for resuscitation.	220kg	390 kg
3.9 Table dimensions should be able to accommodate patients of all ages.	Yes	Yes
3.10 Table top should have large metal free over hang for un-obstructed imaging coverage.	Yes	Yes
3.11 Complete accessories should be provided including arm holder.	Yes	Yes
3.12 Hand grip, arm support and arm rest and positioning aids.	Yes	Yes
3.13 Left / right table pivoting: + / - 90 degree	Yes	Yes
4 X-Ray Tube:	Yes	Yes
4.1 Triple focus with at least 3.0 MHU or better anode heat storage.	Yes	Yes
4.2 Large Capacity to enable continuous heat dissipation during serial exposure. Tube power should be same as generator power i.e. 100kw.	Yes	Yes
4.3 Rotating anode bearing less liquid metal technology tube.	Yes	Yes
4.4 Focus 0.3, 0.6 and 1.0 mm or similar	Yes	Yes
4.5 Dose management with auto adjustment fluoro-filters.	Yes	Yes
5 X-Ray Generator:		
5.1 Microprocessor based high frequency using fiber optic for data communication between each imaging system.	Yes	Yes
5.2 Dedicated X-Ray generator of 100KW.	Yes	Yes
5.3 Radiographic rating minimum 1000mA.	Yes	Yes
5.4 Serial filming exposures with shortest exposure of 1ms, with automatic kV and mA control for optimum image quality.	Yes	Yes
5.5 The system should have capability of digital radiography and fluoroscopy.	Yes	Yes
5.6 Should have capability of doing digital pulsed fluoroscopy at 10/2.5/15 frames and 25/30frames per second or similar.	Yes	Yes

(Handwritten signatures and initials in blue and green ink)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

PD-Trauma Centre-CHK/042 (Rev) 2015-16

Tender Enquiry No.
Item No.
Name of Item
Quantity Required

1
Diagnostic & Therapeutic Angiography System
01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
5.7 Automatic kV, mA & pulse width regulations.	Yes	Yes
6 Digital Imaging & Acquisition / Fluoroscopy		
6.1 Digital system.	Yes	Yes
6.2 Acquisition, storage and display in 1024 x 1024 x 12 bits or more at 12.5 / 15 and 25/30 FPS and 512 x 512 at 60 FPS for pediatric work.	Yes	Yes
6.3 Parallel processing capability / multitasking facility.	Yes	Yes
6.4 Real time filtering and road map function.	Yes	Yes
6.5 Magnetic Disk Capacity for Storage of 64,000 images or greater in 1024 x 1024 on the magnetic disk of main console.	Yes	Yes
6.6 Minimum scene length to be 10 seconds in 1024 matrix.	Yes	Yes
6.7 Digital pulsed fluoroscopy with 12.5 / 15 and 25 / 30 FPS in 1024 x 1024 x 10 bits or more.	Yes	Yes
6.8 Images to be stored on and retrieved from archival disks for possible manipulation and quantification using available software packages.	Yes	Yes
6.9 Rotational DSA package with 40 degrees/second or more minimum head end cam rotation speed	Yes	Yes
7 Monitoring System:		
7.3 Two medical grade monitors for live images and road mapping in the control room 18 inch or larger LCD/LED and 2 in cath room	Yes	Yes
7.4 All the monitors will be of Medical Graded, complied with international standards for medical monitors.	Yes	Yes
8 Recording / Archiving & Communication System:		
8.1 Recording / archiving system should be DICOM-3 compatible.	Yes	Yes
8.2 The digital images should be stored as backup on CD/DVD.	Yes	Yes
8.3 DICOM (send/storage, commitment, retrieve/query)	Yes	Yes
8.4 Ethernet connection to connect with other terminals.	Yes	Yes

(Handwritten signatures and initials)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16

Item No. 1
 Name of Item Diagnostic & Therapeutic Angiography System
 Quantity Required 01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
8.3 DICOM (send/storage, commitment, retrieve/query)	Yes	Yes
8.4 Ethernet connection to connect with other terminals.	Yes	Yes
8.5 Integrated Intercom system.	Yes	Yes
9 Software / Hardware Packages:		
9.1 Complete analysis package for the following vascular applications.	Yes	Yes
9.2 Dynamic pre and post intervention comparison with one image live and other reference.	Yes	Yes
9.3 Automatic loop replay after acquisition or fluoroscopy.	Yes	Yes
9.4 Dynamic real time pan / zoom.	Yes	Yes
9.5 Dynamic real time digital image processing like edge enhancement or gamma correction, noise reduction (spatial filtration.)	Yes	Yes
9.6 Roadmap facility	Yes	Yes
9.7 Simultaneous display of fluoroscopy and reference images, not only as static images but as dynamic loop.	Yes	Yes
9.8 Multitasking capability of recording CD in the background while acquiring new patients fluoro images	Yes	Yes
9.9 Online image density (gray scale) correction.	Yes	Yes
9.10 Facility to review previous studies in the examination room from the patients old CD.	Yes	Yes
9.11 Automatic positioning of the C-Arm corresponding to reference image.	Yes	Yes
9.12 Store fluoro facility to store fluoroscopy.	Yes	Yes
9.13 Stent visualization system latest version by the company.	SNRF	CLEAR STENT
9.14 Table should have stepping top for peripheral dsa / Bolus Chase for 5 or more steps program	Yes	C-Arm stepping
10 Surgical Shadow Less Light:		
10.1 Ceiling suspended LED arm, for Angiographic and surgical procedures.	Yes	Yes

(Handwritten signatures and initials in green and blue ink)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

PD-Trauma Centre-CHK/042 (Rev) 2015-16

1
Diagnostic & Therapeutic Angiography System
 01 Unit

Tender Enquiry No.

Item No.

Name of Item

Quantity Required

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
11 Radiation Protection:		
11.1 Ceiling suspended tiltable lead glass for radiation protection of operators head & neck regions and lower body parts.	Yes	Yes
11.2 Collision tolerant.	Yes	Yes
11.3 Lower body radiation protection flap.	INFAB - US Model	TRUMPF / MAQUET
12 Accessories of Angiography System:		
12.1 PHYSIOLOGICAL HEMODYNAMIC MONITORING SYSTEM: (from same manufacturer of angio system or its joint venture partner) integration should be FDA/CE/MHLW approved. Multichannel for 2 channels IBP, Cardiac output with thermo dilution method. 12 lead ECG, and SpO2	QMAPP (Fysicon) CE approved	SENSIS HEMOLITE (Manufacturer) FDA approved
12.2 Digital display of all the parameters like IBP, Heart rate, cardiac output parameters.	Yes	Yes
12.3 Pulse oximeter probe	Yes	Yes
12.4 Two invasive blood pressure transducer with disposable transducer domes.	Yes	Yes
12.5 Interface with angio system	Yes	Yes
12.6 One Complete 12 lead ECG Cable set.	Yes	Yes
12.7 Cardiac Output (01 No)	Yes	Yes
12.8 Fully programmable latest model contrast medium injectors with volume and flow control and facility of contrast medium.	MEDRAD ARTERION MARK - 7	MEDTRON INJECTOR
13 Other Accessories for Angiography Suite		
13.1 UPS -120 KVA or more sine wave online with battery bank to support whole system and accessories with atleast 10 minutes back up	REFILLO (Italian made) Yes	EMERSON Yes
13.2 Lead aprons (06)	Yes - Infab	Yes - Mavig
13.3 Lead goggles (06)	Yes - Infab	Yes - Mavig

(Handwritten signatures and initials in blue and green ink)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. 1
 Name of Item Diagnostic & Therapeutic Angiography System
 Quantity Required 01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
13.4 Thyroid protectors (06)	Yes - Infab	Yes - Mavig
13.5 Lead gloves (06) pairs.	Yes - Infab	Yes - Mavig
13.6 Lead glass window of size 5' x 2.5 feet	Yes	Yes
14 Quality & Safety Standards / Radiation Protection:		
Knockout clause : Main angiography system should have FDA 510K approval and MDD compliance and country of manufacturer of this unit should be USA, Europe, Japan.	FDA, CE and MHL W	FDA & CE
15 Training		
15.1 Training of Operator for one week at site	As per T.E	Yes
15.2 Application training on system to doctors	As per T.E	Yes
16 Installation & Commissioning		
16.1 The contracting firm will install complete system including all accessories within six months starting from date of establishing LC.	As per T.E	Yes
16.2 Manufacturer and supplier shall be responsible for rectifying with all possible speed at their own expense and manage any defect or fault in the system developing at any time during installation and commissioning period.	As per T.E	Yes
17 Warranty & Post Warranty Agreement & Maintenance Contract:		
Warranty period shall be of 3 years starting from date of commissioning, covering services and all parts including X-ray tube (unlimited exposures in 3 years), Detector, software, purchased non proprietor & optional items, machine accessories like UPS, Injector, etc. Warranty should come from the manufacturer; local agent warranty is not acceptable and vendor should submit proof of manufacturer warranty. This is a knockout clause	As per T.E	Yes

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16

Item No. 1
 Name of Item Diagnostic & Therapeutic Angiography System
 Quantity Required 01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
17.2 A certificate shall be produced by the manufacturer guaranteeing availability of a certified and trained engineer atleast in Karachi for atleast TEN YEARS to be immediately available on site, should any need arise.	As per T.E	Yes
17.3 Maintenance of the machine / equipment during the whole period of warranty shall be responsibility of the vendor, failing which, the vendor will have to pay down time penalty as indicated in terms and condition.	As per T.E	100-95% No penalty
17.4 An annual optimal uptime of 95% is considered as acceptable level of performance. Uptime shall be defined as time available to user for doing data acquisition and processing during working hours of six days a week throughout the year.	As per T.E	Yes
17.5 Compatible ultrasound/Doppler with linear probe for identification and collaboration will be included.	FAMIO 30 Colour Doppler 10L5PL W-704M	Colour Doppler Portable Acuson P-500
17.6 Manufacturer / supplier will guarantee the availability of spare parts and accessories for the system for at least ten years.	As per T.E	Yes
17.7 Manufacturer / supplier shall check system performance during and after every two months. Planned Preventive repair and maintenance time will not be considered as downtime.	As per T.E	Yes
17.8 Vendor will mention the rate of service charges with and without parts for the complete system, after the expiry of 3 years warranty period on yearly basis along with the bid in USD	Mentioned in Financial Offer	Yes
17.9 Manufacturer / supplier should also confirm that in event of the services costs falling or reducing in line with the manufacturer policy worldwide due to technical innovations or other reviews, they will reduce the maintenance cost accordingly	As per T.E	Yes
17.10 Up-gradation of computing system / FMIS should be carried out as available during warranty period as desired by end user for console and workstation and after warranty within the maintenance agreement cost	As per T.E	Yes

(Handwritten signatures and initials)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16
Item No. 1
Name of Item Diagnostic & Therapeutic Angiography System
Quantity Required 01 Unit

Name of Firms:	M/S Medequips SMC	M/S Siemens Healthcare (Pvt.) Ltd.
17.11 Up-gradation of computing system / FMIS should be carried out as available during warranty period as desired by end user for console and workstation and after warranty within the maintenance agreement cost	As per T.E	Yes
18 Optional Items:		
18.1 System Up gradation to higher version should be indicated if provided by manufacturer	As per T.E	Yes
18.2 Any additional option available with each manufacturer may be offered as optional	1. Advance dose tracking system kit XIDF-DTS802 2. STEPPING DSA 3. Advanced spot Fluoroscopy 4. Additional Workstation 5. Multimodality Roadmap for CT and MRI Integration	Not quoted
19 Cost		
19.1 Cost on C&F basis	Yes	Yes
Complete Job: Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period)		
Country of Manufacturer: USA, EEC, Japan or Equivalent	Japan / Netherland	Germany
Delivery Period:	As per T.E	As per T.E
Warranty:	As per T.E	As per T.E
Catalogue / Brochure should be attached duly flagged.	Attached	Attached

Both qualified.

M S C

DR. SHAM MOHAPATRA
Professor of Orthodontics
Dental College & Hospital, Raipur

Supply, Installation, Testings, Commissioning & Maintenance (during maintenance period) of **Medical Equipment** for the project Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. 2
 Name of Item **Hand Held Fogging Machine with Motor Blower Nozzle Technology**
 Quantity Required 10 Units

Name of Firms:	M/S Karachi Medical Company ✓
Make	Fogmaster Corporation
Model	NOZ-L-JET 7807
Hand Held Fogging Machine with Motor Blower Nozzle Technology	
• Chemicals: Liquid Capacity Output Rate Materials	Yes
• 1 Hp, 120VAC, 7	Yes
• amp; 240VAC, 3.5 amp	Yes
• Balanced fan, 20,000 rpm	Yes
• Counter-rotating vortex design	Yes
• High turbulence in nozzle shears should feed liquid into fog-sized droplets (7-30 microns VMD). Nozzle should have no small orifices and resistant to plugging.	Yes
• Nozzle should be able to atomize both oil-based and water-based liquids.	Yes
• Machine should be used for controlling odors, sanitizing, dis- infecting, controlling flies, mosquitoes and other insects, carpet mildew control, dust suppression, humidification, flood and fire damage restoration and other innovative applications. Particle size distribution should vary with liquid viscosity, surface tension, density and output rate. 1 gallon [4 L] Up to 10 oz [300 ml] per minute	Yes
• Power head, wand, tank - aluminum Tank gasket - Buna N Hose - vinyl Tubing - fuel and oil resistant vinyl Fittings should be of brass	Yes
• The power head can be adjusted to up or down angles as required.	Yes
• Made of durable aluminum, it's light-weight, but tough.	Yes
Complete Job: Supply, Installation, Testings, Commissioning and Maintenance (during maintenance period)	
Country of Origin: UK, USA, Japan or Equivalent	USA
Delivery Period:	As per T.E.
Warranty:	As per T.E.
Catalogue / Brochure should be attached duly flagged.	Attached

Qualified

Signature

Dr. Yasmeen Kharal
 Additional medical Superintendent /DD/Shaheed Mohtarma Benazir Bhutto Accident Emergency Centre and Ancillary Services Complex @ Civil Hospital Karachi

Signature

Signature

Supply, Installation, Testing, Commissioning & Maintenance (during maintenance period) of **Medical Equipment** for the project Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. 3
 Name of Item **Skin Graft Mesher**
 Quantity Required 02 Units

Name of Firms:	M/S Karachi Medical Company
Make	Brennen Medical (Now) Mohlycke Healthcare
Model	Brenin Skin Graft Mesher
Skin Graft Mesher	
<ul style="list-style-type: none"> Heavy Duty Skin Graft Mesher complete with all accessories. 	
Specifications/ Features:	
<ul style="list-style-type: none"> Heavy Duty made of Medical Grade Stainless Steel. 	Not Mentioned
<ul style="list-style-type: none"> Totally a self-contained instrument to expand the skin graft. 	Yes
<ul style="list-style-type: none"> Usable with any length piece of skin. 	Yes
Non-Crushing Non-Carrier based Instrument	Yes
<ul style="list-style-type: none"> Meshes by pinching NOT by cutting the graft. 	Yes
<ul style="list-style-type: none"> System shall allow to mesh any graft length without using any plastic carriers. 	Yes
<ul style="list-style-type: none"> Capable of creating a uniform pattern and maintaining graft integrity. 	Yes
<ul style="list-style-type: none"> Ratchet handle shall operate smoothly to provide effortless meshing. 	Yes
<ul style="list-style-type: none"> Heavy weight base to provide stability and holding the mesher firmly in place for unassisted operations. 	Yes
<ul style="list-style-type: none"> It shall not contain blades to eliminate the need for sharpening and reduces risk of injury. 	Yes
<ul style="list-style-type: none"> Fitted with Hand grip/Crossbar giving an extra element of steadiness during meshing and making unit easily transportable. 	Yes
<ul style="list-style-type: none"> Platform width: 4(1/2)" or 6 (1/2)" 	4(1/2)"
<ul style="list-style-type: none"> Variety of Ratios:- 1:1, 2:1, 3:1, 4:1 & 6:1 	Yes
Complete Job: Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period)	
Country of Manufacturer: USA, EEC, Japan or Equivalent	USA

(Handwritten signatures and initials in blue and green ink)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. 3
 Name of Item Skin Graft Mesher
 Quantity Required 02 Units

Name of Firms:	M/S Karachi Medical Company
Delivery Period:	As per T.E
Warranty:	As per T.E
Catalogue / Brochure should be attached duly flagged.	Attached

as

[Signature]

[Signature]

[Signature]

[Signature]

Qualified

DR. EHMER
 ASSISTANT PROFESSOR
 BURNS CENTRE
 CIVIL HOSPITAL AND
 QIBV UNIVERSITY OF HEALTH SCIENCES

Supply, Installation, Testing, Commissioning & Maintenance (during maintenance period) of **Medical Equipment** for the project Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

PD-Trauma Centre-CHK/042 (Rev) 2015-16

Tender Enquiry No.

Item No.

Name of Item

Quantity Required

4

Dermatome Battery Operated

03 Units

Name of Firms:	M/S Oriental Sales Corporation	M/S Karachi Medical Company (Original Offer)	M/S Karachi Medical Company (Alternate Offer)
Make	Integra LifeSciences Corporation	Humecca BV	Aesculap / B.Braun
Model	Integra Padgett Dermatomes Model S Electrical Operated	D80	Acculian 3Ti
Dermatome Battery Operated	Yes	Yes	Yes
Specifications		Yes	Yes
Battery operated		Yes	1100 mAh
Voltage: 2400 mAh		Yes	NIMH
Battery Chemistry: Li-Ion(no Memory effect) Charger for Dermatome		Yes	240v/50HZ
Power supply primary: 100-240v/50HZ		Yes	9.6v:1.05
Charge current: 1200mA		Yes	One hour
Time to charge battery: 2-3 hours approx		Yes	Depends on charge state and capacity
Weight Without Battery and Motor cartridge: 700-800g		Yes	1396g
Total weight: Not more than 1500g		1.33g	1700g
Length: up to 12 Inches		11.6"	290mm

(Handwritten signatures and initials)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No. PD: Trauma Centre-CHK/042 (Rev) 2015-16
 Item No. 4

Name of Item Dermatome Battery Operated

Quantity Required 03 Units

Name of Firms:	M/S Oriental Sales Corporation	M/S Karachi Medical Company (Original Offer)	M/S Karachi Medical Company (Alternate Offer)
Width Head Type : 4.09 Inches		Yes	110mm
Height : 1.5 to 2 inches (Approximately)		Less than 2.05 inches	90mm
No Vibration and shock		Not Mentioned	Not Mentioned
Cutting speed 7000 strokes/Minutes		Yes	0-8500stroke/min
Cordless, battery operated and light weight		Yes	Yes
Precise thickness of the graft from 0.0 to 1.2 mm (0.000 - 0.048") in 0.1 mm (0.008") increments.		Yes	0.2-1.2 in 0.1mm increments
Width-reducing clamps on the dermatome head for cutting of smaller graft widths.		Yes	1/10mm steps
Thickness adjustment should become fixed as required		Yes	Yes
Should have powerful Li-Ion batteries		Yes	NIMH
Should have autoclave case		Yes	Yes
Dermatome Complete Set should Include:			
1) Dermatome		Yes	Yes
2) Box of 1X10 Blade		Yes	Yes
3) Dermatome Motor		Yes	Yes
4) Dermatome Battery		Yes	Yes
5) Width Reducing Plates (3)		Yes	Yes
6) Charger For Battery		Yes	Yes
7) Three Options with Charger		Yes	Yes
1) Lubricating Oil		Yes	Yes
2) Safety Funnel		Yes	Yes
3) Safety Hooke		Yes	Yes

(Handwritten signatures and initials in blue and green ink)

TECHNICAL DATA SHEET (SPECIFICATION COMPLIANCE)

Tender Enquiry No.: PD-Trauma Centre-CHK/042 (Rev) 2015-16
 Item No.: 4

Name of Item: Dermatome Battery Operated
 Quantity Required: 03 Units

Name of Firms:	M/S Oriental Sales Corporation	M/S Karachi Medical Company (Original Offer)	M/S Karachi Medical Company (Alternate Offer)
4) Operating Manual		Will be supplied with item	Will be supplied with item
5) Instruction Manual		Will be supplied with item	Will be supplied with item
6) Service Manual		Will be supplied with item	Will be supplied with item
7) DVD of Maintenance and after care Videos		Will be supplied with item	Will be supplied with item
8) Screw Driver		Yes	Not Needed
Complete Job: Supply, Installation, Testing, Commissioning and Maintenance (during maintenance period)			
Country of Manufacturer: USA, EEC, Japan or Equivalent	USA	Netherland	Germany
Delivery Period:	As per I.T	As per T.E	As per T.E
Warranty:	As per I.T	As per T.E	As per T.E
Catalogue / Brochure should be attached duly flagged.	Attached	Attached	Attached

*Original Quot Alternates
 Offers of M/S Kmc are
 approved & qualified*

AS
MS
my

DR. EHMER
 ASSISTANT PROFESSOR
 BURNS CENTRE
 CIVIL HOSPITAL AND
 DOW UNIVERSITY OF HEALTH SCIENCES

SHAHEED MOHTARMA BENAZIR BHUTTO Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi						
TENDER NOTICE						
Civil Hospital Karachi invites sealed bids on Single Stage — Two Envelope Procedure as per Sindh Public Procurement Regulatory Authority Rules from the manufacturers / importers / sole agents / contractors for the following works in Civil Hospital Karachi.						
S. No.	Name of Work	Bidding Document Cost (Rs.)	Bid / Tender Security	Tender Purchasing Date	Date of Submission and Opening	Time of Completion of Works
1.	Supply, Installation, Testing, Commissioning & Maintenance (during maintenance period) of following items of the Medical Equipment for the project Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi.	1000.00	1% of the Bid Amount	25-04-2016 to 19-05-2016	20-05-2016 at 11:00 a.m. & 20-05-2016 at 11:30 a.m., respectively	Three (3) Months
	1. Diagnostic & Therapeutic Angiography System (01 Unit) 2. Decontamination Fogging Machine (10 Units) 3. Skin Graft Mesher (02 Units) 4. Dermatome Battery Operated (03 Units)					
<ol style="list-style-type: none"> A complete set of tender documents containing BOQ, specification and terms & conditions can be purchased from the Office of the Project Director Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi through pay order of Rs. 1000.00 each in favour of Project Director Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi on submission of application along with copy of NTN Certificate by self or through authorized representative. Tender documents can be downloaded from SPPRA website www.pprasindh.gov.pk The bids must be submitted on the above date by 11:00 am in Board Room of the Administration Block of Civil Hospital Karachi situated on Second Floor. These bids will be opened publicly at 11:30 a.m. on the same day in presence of bidders or their authorized representatives who choose to attend on the place of submission. Technical and Financial Proposals should be submitted in separate envelopes and both should be sealed in third envelope. All NIT(s) shall include Government Taxes etc. (if applicable). The deductions in the bills will be done as per government rules. Bid / Tender Security @ 1% of total value of the quoted items by firms in shape of Pay Order / Demand Draft must be attached with "Financial Bid" in favour of the Project Director Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi. A photocopy of the bid / tender security in shape of Pay Order / Demand Draft without showing the amount figure must be submitted with Technical Bid, failing which the bid will be ignored. In case of announcement of public holiday or any unfavorable circumstances on the day of opening, the bids shall be submitted and opened as per given schedule on the next working day. 						
INF-KRY No. 1873/16		ہم رشوری کا خلاف نہیں۔		Say No to Corruption		PROJECT DIRECTOR

MEDICAL EQUIPMENT



آجر پهرين مئي 2016 ع

شھيد محترم بيٽظير ڀٽو ايڪسيڊنٽ، ايمرجنسي اينڊ ٽراما سينٽر اينڊ اينسيلري سروس ڪمپليڪس واقع سول اسپتال، ڪراچي

ٽينڊر نوٽيس

سول اسپتال ڪراچي جي هيٺ ڄاڻايل ڪمن لاءِ سول اسپتال ڪراچي پاران سنڌ پبلڪ پروسڪيورمينٽ ريگيوليٽري اٿارٽي رولس تحت سنگل اسٽيج ٽو اينڊ ٽي ميٽوڊيڪل/ امپورٽرز/ سول ايجنس / نيڪيڊان ڪان مهيند واک گهريل آهن

ايس نمبر	ڪم جو نالو	واڪ دستاويز (واڪه (روپيا)	واڪه/ٽينڊر سيڪيورٽي	ٽينڊر خريد ڪرڻ جي تاريخ	جمع ڪرائڻ جي تاريخ	مدر
1	شھيد محترم بيٽظير ڀٽو ايڪسيڊنٽ، ايمرجنسي اينڊ ٽراما سينٽر اينڊ اينسيلري سروسز ڪمپليڪس پروجيڪٽ واقع سول اسپتال ڪراچي لائينڊيڪل ايڪيويٽمينٽ لاءِ هيٺ ڄاڻايل اسمن جي (مينٽيننس مادي دوران) سهلائي اسٽاليشن ۽ مينٽيننس	1000.00	واڪه رقم جو 1%	25-04-16 کان 19-05-16	20-05-16 صبح 11 وڳي ۽ 19-05-16 صبح 11:30 وڳي	03 مهينا
1	ڊائيجوسٽڪ اينڊ ٽيراپيوٽڪ اينڊيوگرافي سسٽم (1 يونٽ)					
2	ڊيڪوريشن فائنگ مشين (1 يونٽ)					
3	ڊرماٽو بئري آريئڊ (03 يونٽ)					

- ٽينڊر دستاويزن جو مڪمل سيٽ بي او ڪيو، تصديقات ۽ شرط ۽ شروط پروجيڪٽ ڊائريڪٽر، شھيد محترم بيٽظير ڀٽو ايڪسيڊنٽ، ايمرجنسي اينڊ ٽراما سينٽر اينڊ اينسيلري سروسز ڪمپليڪس واقع سول اسپتال ڪراچي مان 1000 روپيا هر هڪ لاءِ پروجيڪٽ ڊائريڪٽر شھيد محترم بيٽظير ڀٽو ايڪسيڊنٽ ايمرجنسي ۽ ٽراما سينٽر ۽ اينسيلري سروسز ڪمپليڪس واقع سول اسپتال ڪراچي جي نالي ادا ڪري خريد ڪري سگهجن ٿا. درخواست سان پنهنجي يا بااختيار نمائندي جز اين ٿي اين سرٽيفيڪٽ جمع ڪرائڻو پوندو.
- ٽينڊر دستاويز ايش بي بي آر اي جي ويب سائٽ: www.pprasindh.gov.pk تان به وائون لوڊ ڪري سگهجن ٿا.
- واڪه مٿي ڄاڻايل تاريخ تي 11:00 وڳي صبح جو سول اسپتال ڪراچي جي سيڪنڊ فلور تي ڦاٽر ايڊمنسٽريشن بلاڪ جي بورڊ روم ۾ جمع ڪرائڻا پوندا.
- واڪه سڀني جي سامهون ساڳئي ڏينهن 11:30 وڳي صبح جو ان موقعي تي موجود رهڻ جي خواهشمند واکه ڏيندڙن يا سندن بااختيار نمائندن جي وڃي وڃي وڃي ڪوٺيا ويندا.
- ٽيڪنيڪل ۽ فائنل پروپوزل ڦار ڦارڻن و بند ڪري نئين لڦاڻي ۾ پئي وڃي مهيند ڪري ڏيڻا پوندا.
- سورا اين ائي تي (ايس، سان حڪومتي ٽيڪس (لاڳو هوندا تہ) شامل ڪرڻا پوندا.
- بلن مان حڪومت جي ڦاٽن موجب ڪٽڙي ڪئي ويندي.
- واڪه / ٽينڊر سيڪيورٽي اسمن جي ڪمپل ڪل ويلو جو هڪ سيڪڙو بي آرڊر/ ڊمانڊ درافت جي صورت ۾ "فائنل واکه" سان پروجيڪٽ ڊائريڪٽر شھيد محترم بيٽظير ڀٽو ايڪسيڊنٽ، ايمرجنسي اينڊ ٽراما سينٽر اينڊ اينسيلري سروسز ڪمپليڪس واقع سول اسپتال ڪراچي جي نالي شامل ڪرڻي پوندي.
- واڪه/ ٽينڊر سيڪيورٽي جي فوٽر ڪاهي بي آرڊر/ ڊمانڊ درافت سان لازمي انگن ۾ ٽيڪنيڪل واکه سان شامل ڪرڻي پوندي، ڪوٺاهي جي صورت ۾ واکه تي غور ٿي ڪيو ويندو.
- جيڪڏهن عام مرڪل جو اعلان ٿيو يا ڪا اڻ ٿيڻي صورتحال تي د واکه پئي ڪم واري ڏينهن تي جمع ڪرائي ڪوٺيا ويندا.

پروجيڪٽ ڊائريڪٽر

INF-KRY.NO.1873/2016

SAY NO TO CORRUPTION

ڪرپشن جو خاتمو آڻيو

8398

پتو: پتيل جي چوڪي، علاء علي، ڄامشورو، بيجر وٽي، ڄامشورو، سنڌ

تلفون نمبر: 8398

پتو: ايس ايس ايس ايس ڪرپشن

Home Public Tenders SPPRA Notifications SPPRA Board SPP Act 2009 SPP Rules 2010 Ammenements SPPRA Team Contact Us

Tenders * Serial No.
 Text
 Committee Tender Title Advance Search
 Sample Documents SPPRA ID submit
 Advertise Date
 Policy Instructions
 Review Panel City

Black Listed Firms 1 Page 1 Documents Viewed 935315

Capacity Building	S.No.	SPPRA ID	Tender Title	Download	Uploaded On	Advertised On	Closing On	Corr.gendum	Comments
Complaints			Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre at Civil Hospital KARACHI NI WITH SBJ		05-05-2016	04.May.2016	20.May.2016		BER, CS TER AOX Request

Download
 Statistics

Downloads

Evaluation Reports

FAQ's

Feed Back

Gallery

Notifications

Procurement Plans

Site Statistics

Useful Links

Login

Users Online 42

15-Jan-16

Update

Privacy Policy



<http://chk.gov.pk>

search

search

Civil Hospital Karachi

- [Blog \(http://chk.gov.pk/blog-2/\)](http://chk.gov.pk/blog-2/)
- [Disaster Management \(http://chk.gov.pk/disaster-management/\)](http://chk.gov.pk/disaster-management/)
- [Developmental Projects \(http://chk.gov.pk/developmental-projects/\)](http://chk.gov.pk/developmental-projects/)
- [Events ▶\(http://chk.gov.pk/category/events/\)](http://chk.gov.pk/category/events/)
- [Public private partnership ▶\(http://chk.gov.pk/public-private-partnership/\)](http://chk.gov.pk/public-private-partnership/)
- [Trainings ▶](#)
- [Tenders ▶](#)
- [Awards \(http://chk.gov.pk/category/award/\)](http://chk.gov.pk/category/award/)

Shaheed Mohtarma Benazir Bhutto Accident, Emergency & Trauma Centre Active Tenders

Search:

#	Tender Title	Advertised On	Closing On	View
1	Bidding Document of Medical Equipment	04/05/2016	04/06/2016	View more (http://chk.gov.pk/tender-detail/43)

Showing 1 to 1 of 1 entries

Previous 1 Next

- [RSS \(http://troikasolution.tk/civil/#\)](http://troikasolution.tk/civil/#)
- [Back to Top](#)

Copyright © 2016 Civil Hospital. All rights reserved.

Powered By Troika Solution (<http://www.troikasolution.com>)

ATTENDANCE SHEET

NAME OF WORK: SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI







DATE OF OPENING: MAY 20, 2016 (TECHNICAL PROPOSALS)

S.NO.	MEMBERS OF PROCUREMENT COMMITTEE	SIGNATURE	PHONE NUMBER
1.	Medical Superintendent Civil Hospital Karachi.		
2.	Head of Department of Burns Civil Hospital Karachi		
3.	Project Director SMBB Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi		
4.	Divisional Accounts Officer A.G. Sindh Office		
5.	Representative of Deputy Commissioner (South) Karachi		

ATTENDANCE SHEET

NAME OF WORK: SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

DATE OF OPENING: MAY 20, 2016 (TECHNICAL PROPOSALS)

S.NO.	BIDDERS	NAME & CONTACT NUMBER	SIGNATURE
1.	M/S Medequips SMC	Adnan Hussain Khan 0333-3899842	
2.	M/S Siemens Healthcare Pvt. Ltd.	YAWAR ABBAS 0322 2224369	
3.	M/S Oriental Sales Corporation	M. RAZAUDDIN 03212428259	
4.	M/S Karachi Medical Company	Tariq Mirza 0300-9281707 0300-9281707	
5.	M/S Faraz Associates	Absent	—
6.	M/S DWP Technologies (Pvt.) Ltd.	Absent	—
7.	M/S Shirazi Trading	Absent	—
8.	M/S Karachi Medical Company	Tariq Mirza 0300-9281707	
9.	M/S SAAD Sales Services	A. waka's 0321-9224352	











ATTENDANCE SHEET

NAME OF WORK: SUPPLY, INSTALLATION, TESTING, COMMISSIONING & MAINTENANCE (DURING MAINTENANCE PERIOD) OF MEDICAL EQUIPMENT FOR THE PROJECT SHAHEED MOHTARMA BENAZIR BHUTTO ACCIDENT, EMERGENCY & TRAUMA CENTRE AND ANCILLARY SERVICES COMPLEX AT CIVIL HOSPITAL KARACHI

DATE OF OPENING: MAY 30, 2016 (FINANCIAL PROPOSALS)

S.NO.	MEMBERS OF PROCUREMENT COMMITTEE	SIGNATURE	PHONE NUMBER
1.	Medical Superintendent Civil Hospital Karachi.		03002514001
2.	Head of Department of Burns Civil Hospital Karachi		
3.	Project Director SMBB Accident, Emergency & Trauma Centre and Ancillary Services Complex at Civil Hospital Karachi		
4.	Divisional Accounts Officer A.G. Sindh Office		
5.	Representative of Deputy Commissioner (South) Karachi		

ATTENDANCE SHEET






NAME OF WORK:

SUPPLY, INSTALLATION, TESTING, COMMISSIONING
& MAINTENANCE (DURING MAINTENANCE PERIOD)
OF MEDICAL EQUIPMENT FOR THE PROJECT
SHAHEED MOHTARMA BENAZIR BHUTTO
ACCIDENT, EMERGENCY & TRAUMA CENTRE AND
ANCILLARY SERVICES COMPLEX AT CIVIL
HOSPITAL KARACHI

DATE OF OPENING:

MAY 30, 2016

(FINANCIAL PROPOSALS)

S.NO.	BIDDERS	NAME & CONTACT NUMBER	SIGNATURE
1.	M/S Medequips SMC	Adnan 0333-3899842	
2.	M/S Siemens Healthcare Pvt. Ltd.	YAWAR ABBAS 0322 2224369	
3.	M/S Oriental Sales Corporation	M. Raza UDDIN 0321 2428259	
4.	M/S Karachi Medical Company (Original Offer)	Tasir Miza 0300 9281707	
5.	M/S Karachi Medical Company (Alternate Offer)	Tasir Miza 0300 9281707	
6.	M/S Saad Sales Services	(A)	

