Bioassay Research (CBSBR) DR. PANJWANI CENTER FOR MOLECULAR MEDICINE AND DRUG RESEARCH

INTERNATIONAL CENTER FOR CHEMICAL AND BIOLOGICAL SCIENCES UNIVERSITY OF KARACHI, KARACHI-75270

RE TENDER NOTICE NO. PCMD-BIOEQUIV-9. 2011/12

Tender: 3rd Time

01) Defibrillator

Sealed tenders are invited only from the sales tax registered with Sindh Revenue Board and income tax registered suppliers for the Purchase/Import on C & F basis of Scientific Equipments required for the project entitled, "Center for Bioequivalence Studies and Bioassay Research at Dr. Panjwani Center for Molecular Medicine and Drug Research".

The detail and specifications of the item can be collected from the date of notification of the advertisement on the websites, from Purchase Office of PCMD, on any working day between 9.00 a.m. to 12.30 p.m., (except Friday), on payment of Rs. 300/- (non-refundable), in shape of a pay order (Demand Draft by the out of Karachi bidders), in favor of the Director ICCBS., or downloaded from the websites www.iccs.edu and www.pprasindh.gov.pk The last date of issuing tender is 16-6-2012. The tenders may be submitted with 2% of the bid value as bid security, in shape of a pay order in favor of the Director ICCBS., latest by 2.30 p.m. on 18-06-2012. The tenders will be opened in Purchase Office PCMD of the Center at 03.00 p.m. on the same day in presence of the bidders or their representatives. The Procuring Agency may reject all or any bid subject to the relevant provision of SPP rules.

CO-DIRECTOR PCMD

Quantity: 01set

01. Defibrillator:

Specification:

Waveform:

Damped sinusoidal

Output

2,3,5,7,7,10,20,30,50,70,100,150,200,300, and 360 joules

Energy:

Charge Push-button on apex paddle and on front panel Control

Charge Time:

Less than 10 seconds to 360 joules when powered by a fully charged battery

Armed Indicator

Charge done tone, and charge done lamp on apex paddle

Paddle Contact Indicator: 3-color LED bar graph array on STERNUM paddle indicates quality of debrillator paddle

Paddles:

Standard paddles are anterior/anterior, adult, and pediatric. Adult electrodes (83 cm) slide off to expose pediatric electrodes (21 cm)

Synchronizer:

"SYNC" message appears on the monitor and is annoted periodically on the recorder while in synchronous mode. An audible beep sounds with each detected R-wave, while a marker pulse on the monitor and syn designator on the recorder strip indicate discharge

point. Upon activation, discharge occurs within 30ms of marker pulse.

Monitor:

Input:

ECG activity may be viewed through the paddles or patient cable. Lead I, II, or III

selectable in LEADS position. The monitor and recorder indicate selected ECG source

Heart Rate Display:

Digital readout on monitor from 20 to 300 bpm

Heart Rate Alarms:

Adjustable limits from 20 to 280 bpm

NOTE:

IQ/OQ/PQ to be conducted by the vendors engineers at site.

SUMMARY SHEET

TENDER NOTICE NO. PCMD-BIOEQUIV-9. 2011/12 (3rd Time) The tender will liable to be rejected if this form will not accompan the tender bid / quote

SERIAL No.	BID	FOREIGN	CONNERGION	
	VALUE	CURRENCY	CONVERSION RATE.	PRICE IN PKR.
	TOTAL BID VA	LUE IN DICE		
	Earnest Money			
	Payorder / Bank Draft No:		Date:	
Signature		Seal:		
		Seal.		