

PHONE: 0243-9280151-52

OFFICE OF THE DISTRICT HEALTH OFFICER KHAIRPUR

NO: DHO KHP/ACCT/- 7197/98 DATED 27-3-2017.

To,

The Director,
Information (Advertisement),
Publication Department,
95-96 Pakistan Secretariat,
Karachi.

SUBJECT: RELEASE OF GOVERNMENT ADVERTISEMENT IN LEADING NEWS PAPERS

The tender notice along (seven copies) for supply of instruments, equipments and furniture for health institutions of district Khairpur for the year 2016-2017 are send herewith for publication in the tree leading news papers.

It is requested the attached tender notice may kindly be published in three news papers i.e daily Dawn, daily Jung and daily Kawish only under intimation to this office at your earliest.

The bill in duplicate may please be sent to this office payment.

**DISTRICT HEALTH OFFICER
KHAIRPUR**

C.C for information:-

1. The Additional Secretary M&I Cell health department Sindh @ Karachi
2. The Director General health services Sindh @ Hyderabad
3. The Deputy Commissioner Khairpur.
4. The Managing Director SPPRA Sindh Karachi.

Accepted
**DISTRICT HEALTH OFFICER
KHAIRPUR**

PHONE: 0243-9280151-52

OFFICE OF THE DISTRICT HEALTH OFFICER KHAIRPUR

NO: DHO KHP/ACCT/- 7197/99 DATED 27-3-2017.

TENDER NOTICE

The District Health Officer Khairpur hereby invites sealed bid two envelop technical Bid/ financial bid from various manufacturers / importers /distributors for the supply of items mentioned below. A complete set of tender enquiry may be purchased from the office of the undersigned, PM&I Cell, health department, Sindh secretariat 3rd floor tughluk house Karachi and the Director General Health Services Sindh, Hyderabad on submission of written application upon cash payment of non-refundable fee mentioned in each tender enquiry. The bids must be delivered before opening date ~~2017~~ ³⁻³⁻²⁰¹⁷ and 12:00 Noon, which will be opened publicly in the presence of the bidders or their authorized representative who wish to attend at ~~2017~~ ³⁻⁵⁻²⁰¹⁷ at 1:00 pm. The purchase committee / procurement agency reserve the right to accept / reject any/ all bids subject to the relevant provisions of SPPRA Rules 2010.

Descriptions	Date of opening	Cost of Tender
Instruments/ equipments and furniture	2017 ⁰³⁻⁰⁵⁻²⁰¹⁷	Rs. 2000

N.B.

1. The tenders will be sold from date of publication to ~~20-04-2017~~ ¹⁸⁻⁴⁻²⁰¹⁷
2. Bids must be quoted in Pak Rupees.
3. In case Government announces any public holiday then tender will be opened on next working day.
4. All rates should include Government taxes including GST if applicable.
5. This NIT can be download from website www.sindh.gov.pk or www.sppra.gov.pk



**DISTRICT HEALTH OFFICER
KHAIRPUR**

PHONE: 0243-9280151-52

OFFICE OF THE DISTRICT HEALTH OFFICER KHAIRPUR

NO: DHO KHP/STORE/- 7115/17 DATED 24-3-2017.

To.

The Secretary,
Government of Sindh
Health Department Karachi.

SUBJECT: PROCUREMENT COMMITTEE FOR THE YEAR 2016-2017

It is submitted that the release order of District ADP Revenue component for the current financial year 2016-2017 has been released by the Deputy Commissioner Khairpur for the purchase of Instruments' Equipment' Furniture for the development scheme of District Khairpur.

The proposal for procurement committee is given below:-

- | | | |
|---|--|-----------------|
| • | A.D.H.O District Khairpur | Chairmen |
| • | Taluka Health officer Khairpur | Member |
| • | Representative of Deputy Commissioner Khairpur | Member |

**DISTRICT HEALTH OFFICER
KHAIRPUR**

C.C for information:-

1. The Director General health services Sindh @ Hyderabad
2. The Deputy Commissioner Khairpur.



**DISTRICT HEALTH OFFICER
KHAIRPUR**



GOVERNMENT OF SINDH
HEALTH DEPARTMENT

NOTIFICATION

No. HD/P&F-3-2(427)-2014. In supersession to this department's notification no. 22/S.O/M&F-2-1-2015(CRC) dated 22-09-2015 and in pursuance of Rule 3 and 52 of the Sindh Public Procurement Rules, 2010, the Government of Sindh Health Department hereby constitute a Medical Committee of Reference comprising the following members for scrutinizing the complaints of aggrieved bidders against tender invited by Health Department through Central Procurement Committee

01	Special Secretary Health Department	Chairman
02	Representative of Sindh Medical Association	Member
03	Independent expert from relevant field concerning (to be nominated by the Head of Procuring Agency)	Member
04	Deputy Secretary, PMA&F	Member Secretary
05	Deputy Secretary, General	Member

FOR:

- To scrutinize the complaints from the aggrieved bidders and decide the issue of award/contract in accordance with SPP Rules 2010.


SECRETARY HEALTH

No. HD/P&F-3-2(427)-2014

Revised dated 09/01/2015

C.C. to:

- The Director General Health Services Sindh, Hyderabad
- The P.S. to Chief Secretary Sindh, Karachi
- The Managing Director, Sindh Public Procurement Regulatory Authority, Karachi
- The Special Secretary/Additional Secretary (Admin Development Public Health) Health Department
- The Chairman & all members of the Committee
- The P.S. to Secretary Health


NAWAF AHMED SOOMRO
SECTION OFFICER PMA&F



GOVERNMENT OF SINDH
HEALTH DEPARTMENT

NOTIFICATION

No. DD(Dev)HD/KHP/Pe): A Procurement Committee under Rules-7 of Sindh Public Procurement Rules 2010 is hereby constituted comprising the following for purchase of Instruments / Equipments / Furniture for the Dev. Program Scheme of District Khairpur financial year 2016-17.

01. A.D.H.O District Khairpur. Chairman.
02. Tahsil Chief Medical Officer, Khairpur. Member.
03. Representative of Deputy Commissioner, Sanghar. Member.

The DORS Functions / Responsibilities of the Procurement Committee in accordance with Rule-8 of SPP Rules 2010 shall be as under:

- Preparing bidding documents;
- Carrying out technical as well as financial evaluation of the bids;
- Preparing evaluation report as provided in Rule-15;
- Making recommendations for the award of contract to the competent authority;
- Perform any other technical, ancillary and incidental to the above.


SECRETARY HEALTH

No. DD(Dev)HD/KHP/Pe)

Karachi, dated: 02nd April 2016

C.C to:

- The Managing Director, Sindh Public Procurement Regulatory Authority, Karachi
- The DHO, Sanghar
- The Chairman & all members of the Committee.
- The P & A Minister / Secretary Health


(NAVEED AHMED SOOMRO)
SECTION OFFICER (PM&I)

DAWN TUESDAY APRIL 18, 2017

**OFFICE OF THE DISTRICT
HEALTH OFFICER KHAIRPUR**

PHONE: 0243-9280151-52

NO: DHO KHP/ACCT/7197/99

DATED: 27-03-2017

TENDER NOTICE

The District Health Officer, Khairpur District, is invited to receive tenders for the purchase of the following items for the District Health Office, Khairpur. The tenders should be submitted in a sealed envelope addressed to the District Health Officer, Khairpur, and should be accompanied by a copy of the tender document. The tenders should be submitted by 12:00 noon on 30th March 2017. The lowest and acceptable tender will be accepted. The District Health Officer reserves the right to reject any or all tenders without assigning any reason therefor. The District Health Officer is not bound to accept the lowest tender.

Descriptions Instruments, equipment and furniture	Date of Opening	Cost of Tender
---	-----------------	----------------

N.B.

1. The tenderer should submit a copy of the tender document to the District Health Officer, Khairpur, by 12:00 noon on 27th March 2017.
2. The tenderer should submit a copy of the tender document to the District Health Officer, Khairpur, by 12:00 noon on 27th March 2017.
3. The tenderer should submit a copy of the tender document to the District Health Officer, Khairpur, by 12:00 noon on 27th March 2017.
4. The tenderer should submit a copy of the tender document to the District Health Officer, Khairpur, by 12:00 noon on 27th March 2017.

DISTRICT HEALTH OFFICER
KHAIRPUR

Signature of District Health Officer, Khairpur
Date: 27/03/2017



دفتر ڈسٹرکٹ ہیلتھ آفیسر خیرپور

فون نمبر: 0260151-521

No. DHQHP/ACCT/57/99

Date: 27-03-2017

تعمیراتی کاموں کی اطلاع

ڈسٹرکٹ ہیلتھ آفیسر خیرپور و صدر ڈسٹرکٹ ہیلتھ آفیسر کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

سرگرمیوں کی تفصیلات: ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

1. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

2. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

3. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

4. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

5. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

قیمت: Rs. 2000
 تاریخ: 03-05-2017
 دفتر: ڈسٹرکٹ ہیلتھ آفیسر خیرپور

نوٹ:

1. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔
2. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔
3. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔
4. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔
5. ڈسٹرکٹ ہیلتھ آفیسر خیرپور کی ذمہ داری کے تحت درج ذیل کاموں کی اطلاع دی جا رہی ہے۔

ڈسٹرکٹ ہیلتھ آفیسر خیرپور

INF-KRY: No. 1999/17

www.sppra.gov.pk www.sindh.gov.pk

OFFICE OF THE DISTRICT HEALTH OFFICER, KHAIRPUR
GOVERNMENT OF SINDH



TENDER FOR THE PURCHASE OF
MACHINERY / EQUIPMENT / FURNITURE / FIXTURES
DISTRICT HEALTH OFFICER
KHAIRPUR
FOR THE YEAR 2016-2017

Cost of Tender is Rs. 2000/= non refundable.

1	Tender No.	DHOKHP/ACCT/-7197/99		
2	Tender selling date from	19-04-2017	TO	03-05-2017 12:00 PM
3	Tender submission date	03-05-2017	AT	12:00 PM
4	Tender Opening Date	03-05-2017	AT	01:00 PM



OFFICE OF THE DISTRICT HEALTH OFFICER KHAIRPUR

TERMS AND CONDITIONS

Single Stage – Two Envelopes Bidding Procedure.

1. Single stage - two envelopes procedure shall be applied:
2. The Tender for supply of Machinery / Equipment shall comprise a single stage containing two separate envelopes. Each envelope shall contain separately the technical proposal and the financial proposal;
3. The envelopes shall be marked as "TECHNICAL PROPOSAL" and "FINANCIAL PROPOSAL" in bold and legible letters to avoid confusion;
4. The committee shall evaluate the technical proposal, without reference to the financial proposal and reject any proposal which does not conform to the specified requirements;
5. Financial proposal for the supply of Machinery / Equipment failing to disqualify in the technical evaluation will be returned to the bidders unopened.
6. The bidder quoting the final lowest price and qualifying in the technical evaluation shall be declared Successful.
7. The Tenders shall be submitted with all documents in sealed envelope with sealing Wax. Tenders must be filled in with Blue or Black Ink in the columns provided / on separate letter Head duly signed.
8. The Tender must be free from erasing, cutting and over writing. In case of erasing, cutting and over writing, authorized person should initial it.
9. Store is required immediately, the tenders may how ever give their short guarantee delivery period by which the supply will be complete positively. No extension will be granted / accorded for the supply of initial quantity.
10. The successful Tenderer will have to deposit with purchaser security deposit in shape of pay order / call deposit, 2.5% value of contract.
11. The purchase committee has reserved the right to purchase all or part of stores under this Tender as per Budget provision.
12. The sales Tax and Income Tax will be deducted as per Government Rules.



OFFICE OF THE DISTRICT HEALTH OFFICER KHAIRPUR

BID EVALUATION CRITERIAL

The technical proposals will be evaluated on merits of the, followings:

S.No.	Evaluation Criteria	Yes	No.
1	Conforming technical specification		
2	Proof of financial certificate Rs.50 (M)		
3	Bank statement last 3 years		
4	Proof of General Sales Tax paid during the last three years		
5	Proof of Income Tax paid during the last three years		
6	Proven Track Record (previous record last 03years)		
7	Copy of Professional Tax Certificate current year		
8	Copy of Income Tax Certificate		
9	Copy of Sales Tax Certificate		
10	The Bidder will submit an affidavit on legal stamp paper of Rs.100/-to the effect that their firm has not been blacklisted in the past on any ground by any Government.		
12	Original Tender purchase Receipt		
13	Registration with FBR		
14	Chamber Of Commerce Certificate		

1. The Bidder shall furnish, as part of its financial proposal/bid, refundable Bid Security/Earnest Money in Pak Rupees @ 2% of quoted amount in the shape of Bank Draft / Pay Order / Call Deposit in the name of the District Health Officer KHAIRPUR. The financial bid found deficient of the Bid Security will be rejected. No personal cheque in lieu thereof will be acceptable at any cost. The previous Bid Security, if any, will not be considered or carried forward. However, the Bid Security of the successful Bidder will be returned upon submission of Performance Guarantee equal to 2% of the Contract amount that will remain with the DHO KHAIRPUR till satisfactory completion of the Contract period. In case of unsuccessful bidders, the Bid Security will be returned as soon as possible.

2. The entire quantity of the ordered supplies shall be delivered within 30 days or earlier from the date of issuance of supply order / contract award without any penalty.

3. The payment will be made after submitting the Bill and Inspection.



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

4. The concerned firms will pay all Government Taxes as per Government rules.
5. The firm concerned shall submit the rates for (2016- 2017) till June 2017.
6. The tender rates should be inclusive of all taxes income tax and sale tax etc. Payable to Federal and Provincial Government or local bodies and no claim on the account shall be entertained.
7. You must submit sample along with tender on the day of tender opening of failed to submit samples before the tender opening we have discretion to approve or reject the tender.

CERTIFICATE

We guarantee to supply the stores exactly in accordance with requirements specified in the invitation to tender.

1. Signature of Tenderer _____
2. Designation _____
3. Address _____
4. Stamp _____

Signature of the Tenderer
with seal & address

DISTRICT HEALTH OFFICER
KHAIRPUR



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

TECHNICAL PROPOSAL

TENDER FOR THE PURCHASE OF
MACHINERY / EQUIPMENT / FURNITURE / FIXTURES FOR DISTRICT HEALTH OFFICER
KHAIRPUR
FOR THE YEAR 2016-2017

Cost of Tender is Rs. 2000/= non refundable.

- 1 Tender selling date from 19-04-2017 TO 03-05-2017 12:00 PM
- 2 Tender submission date 03-05-2017 AT 12:00 PM
- 3 Tender Opening Date 03-05-2017 01:00 PM

The Tenders shall quote on free delivery to consignee's end Office of the District Health Officer, KHAIRPUR Price's Figures and words including all Government Taxes (i.e. Income Tax, Sale Tax and other Government Taxes), failing which the offer will be ignored.

S.NO	Particulars	Qty	Make/ Model	Country Of Origin
Machinery / Equipment / Instrument				
1	Suction Machine Make: UK/USA/JAPAN/GERMANY OR EQUIVALENT Frequency(Hz): 50/60Hz Electrical Power (Kw): 0.4Kw Suction motor rotation speed: 3440 rpm Main voltage: 230VAC Noise level: 50db	55		
2	Ultrasound Machine Make: UK/USA/JAPAN/GERMANY OR EQUIVALENT Scanning Method Electronic Convex :Electronic Linear B-Mode specifications; M-mode specifications: M-mode display : Scrolling with 4 speeds DISPLAY MODES: B/M Scan mode: B(Single / Dual) b)B/M C)M 12.1" Digital TFT LCD Color Monitor. Selectable transmittal focus,	07		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	<p>Reception: Continuous Dynamic Focusing. Gray Scale: 256 Biopsy Lines System Software. Upgradeable through CD Drive. Input /Output Composite Out pur Black & White , Rs-232C Ethernet Gain: 62-100bd, STC Switches 0.24cm, 8-steps. Image Processing, Dynamic Range, Edge Enhancement & Post Process. B-Mode measurements: Distance 4-Channel, Angle 2-Channels Area 4-Channel, volume, 4-Channes. Auto annotation facility. Measurement Functions: Distance, Angel, Ratio, Histogram, Area/ Circumference, B-Mode profile, sop Ratio Time, Speed, Biometric Tables for fetal growth measurements with trend graphs.</p>			
3	<p>X-ray Machine 100 MA Make: UK/USA/JAPAN/GERMANY OR EQUVALINET SPECIFICATION: Out put: 24KW/30KW . Max mA out put : 300mA. KVP (Radiography) : 35-125KVP. KVP (Fluoroscopy): 45-100KVP. Timer (Radio.): 0.01 to 5 sec. Timer (Fluoro) : Cumm. Timer of 5 min. Technic selection : 50s, 100s, 200L, 300L . Fluoro mA : 0 3mA. Floor to ceiling column stand Vertical Column assembly. Vertical carriage assembly. Base assembly, ceiling assembly Floor rail. Balancing System with counter weight. Ceiling Height Approx. from 248(228) to 299(279) cm. Above floor level. Ceiling Rail Approx. 300cm. Floor Rail Approx. Vertical Travel Approx. 38.5(48) to 189(176)cm floor to focal spot. Transverse Travel Approx. 96(76) to 134(100)cm. Longitudinal Travel Approx. 240 cm. Rotating of Vertical Column +/- 180 deg with electromechanical. Detents at each 90 degposition. Rotation of Horizontal X-Ray Beam +/- 180 deg with manual lock. Vertical lock release.</p>	3		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	<p>Transverse lock Release. Longitudinal lock release. Rotation lock release. Table Four - way movable table top. The material of the tabletop should be high - impact plastic Laminate. Smooth movement .Electromagnetic locks Longitudinal should be $\pm 255\text{mm}$. 8:1, 85 Lines/inch X-ray Tube Large Focal Spot 2mm.Small Focal Spot 1mm.Rotation Anode 3600 RPM. Heat Unit 140 KHU Manual X-Ray beam Collimator. Pair of H.T Cable. Transverse should be $\pm 170\text{mm}$.VerticalBucky Travel should be 540mm.Grid Vertical chest Bucky stand Balancing System with counter weight. Vertical Travel Approx: 120cm film centre to floor distance from 50 to 185cm. Front Plate to Film Distance Approx 3 cm. Electromagnetic Lock. Film Formats From 5" x 7" to 14" x 17" Cassettes. Cassette Insertion From left to right. Smooth & Quick Grid Movement Reciprocating motion. Grid Ratio & Density : r = 8 : 1, d = 85 lines/inch. Inherent Filtration of Front Plate Less than 0.5mm aluminum at 100 kvp. Focusing Distance to = from 100 to 180cm.</p>			
4	<p>Auto Clave (Electric Udrimues Capacity) Make: UK/USA/JAPAN/GERMANY OR EQUVALINET</p> <p>Chamber Size: 16 liter design with high performance wide range use easy to maintain Intelligence design sterilization technology.</p>	19		



**OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR**

5	<p>B.P Operator Make: UK/USA/JAPAN/GERMANY OR EQUVALINET</p> <p>Casing is of a high profile Al-Zn alloy produced by Pressure Die Casting.</p> <p>Bellows are of proven material for long lasting and better repeatability Sector mechanism is specially designed for maximum repeat accuracy Printed Dial has 20 to 300 mm of Hg scale with 2 mm of Hg resolution Crystal clear cover glass for best visibility All plastic and rubber parts are of first grade virgin material Metal parts are treated with Nickel and Chromium plating to prevent corrosion Best quality painting is adapted to whole body of the instrument Special cuff is provided for getting excellent grip Soft carrying case for ease in handling</p> <p>NOTE: Sample required with technical bid</p>	175		
6	<p>Stethoscope</p> <p>Make: UK/USA/JAPAN/GERMANY OR EQUVALINET</p> <p>Length: 27" (69 cm) Weight: .4079 lbs (185 g) Binaural Construction: Single lumen Chest piece Finish: Chrome Chest piece Technology: Single sided Diaphragm Diameter: 2.0" (51 cm) Diaphragm Material: Polyurethane-Coated Silicone Diaphragm Type: Digital Electronic Filtering - Ambient Noise Reduction (ANR) Ear tip Type: Soft Sealing Extra Ear tips: No Headset Material: Wide diameter aerospace alloy / Anodized aluminum Patient: Adult, Infant, Pediatric Performance: 10+ Special Adaptors: No Applications: Cardiology/High Performance</p> <p>NOTE: Sample required with technical bid</p>	164		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

7	<p>Nebulizer Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Ultrasonic energy for uniform and highly dense 1-5 microns. ♣ More than 96% of 0.3 micron or larger air borne dust particles is effectively shut out with the air filter to provide purified air for aerosol nebulization. Medication cup with replaceable diaphragm. ♣ Easily detachable fan cover and pneumo clean (Air filter). ♣ Made of highly resistant sterilize able resin. ♣ Stand with solution bottle for safety. ♣ Nebulizing rate: 4 ml/min or greater. ♣ Mist particle size: Approx 1-5 microns. ♣ Nebulizing time setting: 1-30 min& continuous ♣ Medication cup capacity: 150 ml. ♣ Accessories: ♣ Tray set for nebulizer with tray track and pole mount fitting.</p>	81		
8	<p>Baby Weighing Scale Make: UK/USA/JAPAN/GERMANY OR EQUVALINET The generous tray on this baby scale leaves room for a soft blanket. The tray is non-tip and the softly rounded walls are high to prevent lively babies from rolling off during measurement. The LCD and digits are large and clear and a built in spirit level ensures level positioning. The tray is detachable for easy cleaning. A measuring rod can be added so that weight and length can be measured at the same time.. Functions TARE HOLD key for lively babies ZERO Auto OFF Features Detachable tray LCD display with 3 cm high digits Battery or power supply with AC adapter</p>	03		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	D & C Kit			
	Hegar Dilator Set of 8	1		
	Curette D / E	4		
	Vulsellum forceps	1		
	Sherman Curette	1		
	Ramply sponge holding forceps	4		
	Towel Clip	6		
	Cusco Specillum	1		
	Sims specillum Set of 3	1		
	Sims Uterine Sound	1		
9	Metal Cathetar	1	50	
	Brown Tenacullum forceps	1		
	Flashing Curette	1		
	Plain forceps 8"	1		
	Bowl 4"	2		
	Bowl 8"	1		
	Instruments Tray	1		
	Total Pieces	28		
	NOTE: Sample required with technical bid			
	CE/FDA USA/ISO 13485-2003			
	Cutler/OTINS/Hilbro			
	Delivery Kit			
	Plain forceps 7"	1		
	Tooth forceps 7"	1		
	Cord Scissor USA Pattern	1		
10	Mayo Scissor Str	1	50	
	Mosquito Artery forceps str	6		
	Artery forceps Str 6"	6		
	Needle Holder 7"	1		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	Sims Speculum Med	1		
	Gali pot	2		
	Large Bowl 8"	1		
	Kidney tray 10"	1		
	Instruments tray 10"x12"	1		
	Total Pieces	23		
	NOTE: Sample required with technical bid			
	CE/FDA USA/ISO 13485-2003			
	Cutler/OTINS/Hilbro			
11	Delivery Table Make: UK/USA/JAPAN/GERMANY OR EQUIVALINET - Lateral tilt left/ right 20°/ 20° - Trendelenburg reverse trendelenburg .. 30°/ 30° - Back section up/ down tilt angle 45° / 0° - Head section up/ down tilt angle 30°/ 90° - Leg section up/ down tilt angle 0° / 90° Standard accessories: 1. Head rest 2. Shoulder supports 3. Body supports 4. Arm rest 5. Screen frame 6. Knee crutch 7. Mattress 8. Built-in kidney lift, removable 9. Foot pedal 10. Crank handle 11. Locking screw 12. Rotation control 13. Trendelenburg movement control 14. Lateral tilt control 15. Head & back section control 16. Eccentric lever 17. Kidney lift control 18. Oil seal with cap 19. Table base 50mm thickness		10	



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	Optional accessories: 1. Restrain strap 2. Hand grips (pair) 3. U-shaped headrest 4. Raised armrest 5. S/S funnel & waste receptacle & corrugated tubing			
12	Dressing Drum Round Stainless Steel NOTE: Sample required with technical bid	128		
13	Cutler/Platinum/Hilbro/OTINS Drip Stand NOTE: Sample required with technical bid	97		
14	Cutler/Platinum/Hilbro/OTINS SMR Set Freer elevator and separators 2 Killian nasal Gouge V shaped Bayonet 2 Killian nasal Gouge V shaped Str 1 Lucae nasal Chisel 2 Hammar S.S 1 Frazier suction tube No 8 - 10 - 12 3 Telly nasal dressing forceps 1 Luc turbinate forceps small 1 Luc turbinate forceps Medium 1 Hartmann Crocodil forceps 1 Killian nasal speculum Set of 4 1 Heyman turbinactomy scissor 1 Thudiam nasal speculum set of 5 1 Trocar Cannula 1 Jenson middleton bone nibler D/A 1 Walsham septum forceps right 1 Walsham septum forceps left 1 Instruments tray 1 Total Pieces 23	40		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	<p>NOTE: Sample required with technical bid</p> <p>CE/FDA USA/ISO 13485-2003 Cutler/OTINS/Hilbro</p>		
15	<p>Examination Coach Main Frame made by 1 ¼ x 1 ¼ M.S Pipe 18 gauge with head side adjustable in synthetic leather of best quality with foam 10 years warranty and the frame is in powder coating</p> <p>Cutler/OTINS/Hilbro/Platinum</p>	119	
16	<p>General Surgical Instrument</p> <p>B.P Handle No 3 2</p> <p>B.P Handle No 4 2</p> <p>B.P Handle No 7 2</p> <p>Sponge holding forcps 6</p> <p>Dissecting forceps Plain / Tooth 6" 1+1 2</p> <p>Dissecting forceps Plain / Tooth 8" 1+1 2</p> <p>Adson forceps Plain & Tooth 2+2 4</p> <p>McIndoe forceps 6" 1</p> <p>Metzenbaum Scissor 6" Str & Cvd 1+1 2</p> <p>Metzenbaum Scissor 7" Str & Cvd 1+1 2</p> <p>Metzenbaum Scissor 8" Str & Cvd 1+1 2</p> <p>Mayo Scissor 6 3/4" Str & Cvd 1+1 2</p> <p>Crile artery forceps Str 6" 12</p> <p>Crile artery forceps Cvd 6" 24</p> <p>Mosquito artery forceps Str 12</p> <p>Mosquito artery forceps Cvd 18</p> <p>Artery forceps 8" Str & Cvd 4+4 8</p> <p>Artery forceps 9" Cvd fine 2</p> <p>Kocher Artery forceps 8" Str & Cvd 4+4 8</p> <p>Babcock intestinal forceps 6" 2</p> <p>Babcock intestinal forceps 8" 2</p> <p>Allis tissue forceps 6" 4</p> <p>Allis tissue forceps 8" 4</p> <p>Backhaus Towel Clip 6</p> <p>Moynihan Forceps 8" 2</p>	60	



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	Langenback Retractor Small 2 Langenback Retractor Med 2 Morris Retractor Medium 2 Balfour Selfretaining Retractor 1 Sinus forceps 1 Volkman Curette D/E 2 Aneurysm Needle 1 Watson Chyne Dissector 1 Yaunker suction nozzle 1 Mayo Hegar Needle holder 6" 2 Mayo Hegar Needle holder 7" 2 Mayo Hegar Needle holder 8" 2 Mayo Hegar Needle holder 9" 2 Kidney pedicle Clamp 4 Kidney Tray 10" 2 Bowl 4" 2 Instruments tray large without lid 1 <p style="text-align: right;">Total Pieces 163</p>			
	<p>NOTE: Sample required with technical bid</p> <p>CE/FDA USA/ISO 13485-2003 Cutler/OTINS/Hilbro</p>			
17	<p>Air Condition Split 1-1.5 Ton Cooling Capacity (BTU/Watts) 18000/5300 Rated current input (A) 08 EER 03 Power Supply (V/Hz) 220-240/50 Cooling Power input (max)W 2500 Cooling Current input (Max)A 11.6 Electric Shock Protection CLASS I Indoor airflow (m³/H) 800 Rated Power Input (W) 1800 Indoor/Outdoor noise dB 37-44/54 Refrigerant R22/1300g Gross weight (IDU/ODU) 15-42 (Kg) Dimension indoor mm 1055*370*300 Dimension outdoor mm 920*335-595</p>	25		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	<p>Ideal Room size (sq. ft) 151-250 Evaporator: Hydrophilic fins with inner grove copper tube Condenser: Golden fins with inner grove copper tube</p> <p>Make: Dawlance/PEL/Gree</p>			
18	<p>Printer for Ultra Sound Machine</p>	7		
19	<p>Refrigerator PRA-160 At Ambient Temp 55C Capacity 420 Liters Nominal Tension 220 Volts Frequency 50Hz Compressor Wattage 217W Total Current 1.8 A Freezing capacity (24hrs) 9.5 Kg Defrosting AUTO Adjustable Thermostat YES CFC Gas (R-34a) 0.18Kg, 1845 H(mm) Packing Dimension W(mm) 730, D(mm) 780 Net weight 85Kg Gross weight 99Kg</p> <p>Make: Dawlance/ PEL</p>	20		
20	<p>Stabilizer A-20 SP Special for very low voltage 3 Relay 2 Meter Input V 100 Watts 2000</p> <p>Make: Universal/Platinum/Golden</p>	62		
21	<p>Electric Water Cooler Tank: 45 Liters Body: Megan, Steel Water Tank: Non megan, Steel Voltage Rating: Operate on 220-240. 50 cycles A.C supply single phase Water Temperature Controllable from 9C to 15C</p>	25		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	<p>Refrigeration System: Fitter with high quality world famous quick Compressor, Condenser and fan motor to ensure long lasting Performance, All cooling system hermitically sealed with RE-12 R-22 as refrigerant.</p> <p>Water System: Main Cooling system is made of Copper and Brass. Water Storage tank is made of high quality copper sheet tinned from inside and tested on maximum pressure according to (PSI-142, 10KG). Fitted with high quality chrome plated brass taps</p> <p>Make: Canon/Super Asia/China</p>			
22	<p>UPS with all accessories</p> <p>10 Tube Lights or 25 Energy Savers & 11 Fans</p> <p>Key Features</p> <p>Built-in intelligent 3-stage charger control for efficient charging and to prevent overcharging Light weighted, wall mounted, rack design User friendly LCD & LED Multiple protections: low battery alarm, low battery shut down, overload protection, over heating protection, short circuit protection Highly efficient DC-to-AC conversion, minimizing energy loss Free AC & DC cables Innovative charging system especially designed to suit local type batteries and local weather conditions</p> <p>Load Range 1500 Watts 2000VA/24VDC/220VAC With Battery Make: Osaka/Exide/</p>	21		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

Furniture / Fixture				
23	<p>Office Chair Made with talhi/shesham wood seat & back netted with Spirit Polish Standard Size Best Quality NOTE: Sample required with technical bid</p>	420		
24	<p>Office Table Made by sheesham wood with glass top and side rack handles, channels locks are best quality finish by spirit polish Size: (L) 4' x (W) 2 ½' x (H) 2 ½' Best Quality NOTE: Sample required with technical bid</p>	24		
25	<p>Revolving Chair Main frame made with talhi/shesham wood seat & back plastic netted plastic based Stand with spirit polish. Best Quality NOTE: Sample required with technical bid</p>	156		
26	<p>Wheel Chairs Wheel Chair Extra Wide Heavy Duty Model 951 B Powder Coating Mild Steel Frame, 22" Seat Width, Flip-Up-Arm Est, Detachable Footrest, Rubber Foam Castor, Pneumatic Mag Wheel. Weight Capacity : 150KG. NOTE: Sample required with technical bid Make: China/Pakistan</p>	33		
27	<p>Meeting Hall Chair Revolving Revolving Chairs - Adjustable lumbar support and instant seat height adjustment. Heavy duty plastic 5-prong base Height/width adjustable arms with soft, durable urethane pads. Minimum Dimensions: Overall: 30-1/4"Wx28-3/4"Dx42" to 44-1/2"H. NOTE: Sample required with technical bid</p>	96		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

28	Meeting Hall Table Training Hall Table (20' X 4'-6" X 3' H) : Made of high density laminations sheet with pvc edge banding on edges Front side of table should be covered with lamination sheet) . The Colour of Lamination sheet will be specified by Engineer in charge.	1		
29	Patients Trolley Made by stainless steel square pipe in 16 gauge with stretcher made in 1 ¼ Round Pipe in stainless Steel with Curve sheet 6 inch imported Casters Locally Made Export Quality NOTE: Sample required with technical bid	36		
30	Steel Fix Chair 3 in each Imported 3 seater metallic bench Non resisting stuff Full metallic structure 100% imported All Stuff available in pin pack cotton NOTE: Sample required with technical bid Make: China/Platinum/Deskon	169		
31	Bed Side Lockers Made by 22 gauge sheet with one space and a cabinet with 2 inch wheel finished by epoxy powder coated top is made by fiber material is very strong and water proof Draw panel and door is made by fiber material door is hang by stainless steel hanger finished by scratch and chip resistance powder coating Locally mad Export Quality. NOTE: Sample required with technical bid	30		



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

32	Fowler Bed Single crank operation Removable head and foot panels Rotary buffers for protection Mobile of 5" diagonally locking casters Dimension (78.5*36.5*24) Inches Main Frame made by 1 ¼ x 2 ½ x 16 gauge upper frame NOTE: Sample required with technical bid	30		
33	Instrument Trolley 18 X 24 With square pipe top S.S powder coated NOTE: Sample required with technical bid	6		
34	Instrument Tray Full Stainless steel NOTE: Sample required with technical bid	34		
35	Cup Board Iron Steel cupboards (6.5 feet x 3 feet) for record keeping Steel Almira of the size 78"x36"x18" with 2 steel doors for the books/library having four shelves making five compartments, 20 gauge over all with locking arrangement and complete in all respect NOTE: Sample required with technical bid	57		
36	Iron Rack Four Shelf NOTE: Sample required with technical bid	63		
37	Maters Single 6.5'*3.5'*4" With leather/ regisine cover NOTE: Sample required with technical bid Make: Master/Celeste	54		

DISTRICT HEALTH OFFICER
KHAIRPUR



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

FINANCIAL PROPOSAL

TENDER FOR THE PURCHASE OF
MACHINERY / EQUIPMENT / FURNITURE / FIXTURES DISTRICT HEALTH OFFICER
KHAIRPUR
FOR THE YEAR 2016-2017

Cost of Tender is Rs. 2000/= non refundable.

- 1 Tender selling date from 18-04-2017 TO 03-05-2017 12:00 PM
- 2 Tender submission date 03-05-2017 AT 12:00 PM
- 3 Tender Opening Date 03-05-2017 01:00 PM

The Tenders shall quote on free delivery to consignee's end Office of the District Health Officer, KHAIRPUR Price should be quoted in Figures and words including all Government Taxes (i.e Income Tax, Sale Tax and other Government Taxes), failing which the offer will be ignored.

S.NO	Particulars	Qty	Make/ Model	Country Of Origin	Rate	Amount
Machinery / Equipment / Instrument						
1	Suction Machine Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Frequency(Hz): 50/60Hz Electrical Power (Kw): 0.4Kw Suction motor rotation speed: 3440 rpm Main voltage: 230VAC Noise level: 50db	55				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

<p>2</p> <p>Ultrasound Machine Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Scanning Method Electronic Convex :Electronic Linear B-Mode specifications; M-mode specifications: M-mode display : Scrolling with 4 speeds DISPLAY MODES: B/M Scan mode: B(Single / Dual) b)B/M C)M 12.1" Digital TFT LCD Color Monitor. Selectable transmittal focus, Reception: Continuous Dynamic Focusing. Gray Scale: 256 Biopsy Lines System Software. Upgradeable through CD Drive. Input/Output Composite Output Black & White, RS- 232C Ethernet Gain: 62-100db, STC Switches 0.24 cm, 8-steps. Image Processing, Dynamic Range, Edge Enhancement & Post Process. B-Mode measurements: Distance 4- Channel, Angle 2 Channels Area 4-Channel, volume, 4-Channels. Auto annotation facility. Measurement Functions: Distance, Angel, Ratio, Histogram, Area/Circumference, B-Mode profile, sop Ratio Time, Speed, Biometric Tables for fetal growth measurements with trend graphs.</p>	<p>07</p>				
--	-----------	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

3	<p>X-ray Machine 100 MA Make: UK/USA/JAPAN/GERMANY OR EQUVALINET SPECIFICATION: Output: 24KW/30KW. Max mA output: 300mA.KVP(Radiography):35-125KVP.KVP (Fluoroscopy): 45- 100KVP. Timer (Radio.):0.01 to 5sec.Timer (Fluoro): Cumm. Timer of 5min.Technic selection: 50s, 100s,200L, 300L. Fluoro mA :0-3mA. Floor to ceiling column stand Vertical Column assembly. Vertical carriage assembly. Base assembly, ceiling assembly Floor rail. Balancing System with counter weight. Ceiling Height Approx. from 248(228) to 299(279) cm. Above floor level.Ceiling Rail Approx. 300 cm. Floor Rai lApprox.Vertical Travel Approx. 38.5(48) to 189(176) cm floor to focal spot Transverse Travel Approx. 96(76) to 134(100) cm. Longitudinal Travel Approx. 240 cm.Rotating of Vertical Column +/- 180 deg with electromechanical Detents at each 90 degposition. Rotation of Horizontal X Ray Beam +/- 180 deg with manual lock. Vertical lock release.</p>	03				
---	---	----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	<p>Transverse lock Release, Longitudinal lock release. Rotation lock release. Table Four-way movable table top. The material of the table top should be high-impact plastic Laminate. Smooth movement. Electromagnetic locks Longitudinal should be ± 255mm. 8:1,85 Lines/inch X-ray Tube Large Focal Spot 2mm. Small Focal Spot 1mm. Rotation Anode 3600 RPM. Heat Unit 140 KHU Manual X-Ray beam Collimator. Pair of H.T Cable. Transverse should be ± 170mm. Vertical Bucky Travel should be 540mm. Grid Vertical chest Bucky stand Balancing System with counter weight. Vertical Travel Approx: 120cm film centre to floor distance from 50 to 185cm. Front Plate to Film Distance Appro x 3cm. Electro magnetic Lock. Film Formats From 5"x7" to 14"x17" Cassettes. Cassette Insertion From left to right. Smooth & Quick Grid Movement Reciprocating motion. Grid Ratio & Density: r= 8:1, d=85 lines/inch, Inherent Filtration of Front Plate Less than 0.5mm aluminum at 100 kvp. Focusing Distance to= from 100 to 180 cm.</p>					
1	<p>Auto Clave (Electric Udrimues Capacity) Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Chamber Size; 16liter design with high performance wide range use easy to maintain Intelligence design sterilization technology.</p>	19				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

5	<p>B.P Operator Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Casing is of a high profile Ai-Zn alloy produced by Pressure Die Casting. Bellows are of proven material for long lasting and better repeatability Sector mechanism is specially designed for maximum repeat accuracy Printed Dial has 20 to 300mm of Hg scale with 2mm of Hg resolution Crystal clear cover glass for best visibility All plastic and rubber parts are of first grade virgin material Metal parts are treated with Nickel and Chromium plating to prevent corrosion Best quality painting is adapted to whole body of the instrument Special cuff is provided for getting excellent grip Soft carrying case for ease in handling</p> <p>NOTE: Sample required with technical bid</p>	175				
---	--	-----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

5	<p>B.P Operator Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Casing is of a high profile Al-Zn alloy produced by Pressure Die Casting. Bellows are of proven material for long lasting and better repeatability Sector mechanism is specially designed for maximum repeat accuracy Printed Dial has 20 to 300mm of Hg scale with 2mm of Hg resolution Crystal clear cover glass for best visibility All plastic and rubber parts are of first grade virgin material Metal parts are treated with Nickel and Chromium plating to prevent corrosion Best quality painting is adapted to whole body of the instrument Special cuff is provided for getting excellent grip Soft carrying case for ease in handling</p> <p>NOTE:Sample required with technical bid</p>	175				
---	---	-----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

6	<p>Stethoscope Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Length: 27" (69cm) Weight: 4079 lbs (185g) Binaural Construction: Single lumen Chest piece Finish: Chrome Chestpiece Technology: Single sided Diaphragm Diameter: 2.0" (51cm) Diaphragm Material: Polyurethane-Coated Silicone Diaphragm Type: Digital Electronic Filtering Ambient Noise Reduction (ANR) Ear tip Type: Soft Sealing Extra Ear tips: No Headset Material: Wide diameter aerospacealloy/ Anodized aluminum Patient: Adult, Infant, Pediatric Performance: 10 + Special Adaptors: No Applications: Cardiology / High Performance</p> <p>NOTE: Sample required with technical bid</p>	164				
---	--	-----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

7	<p>Nebulizer Make: UK/USA/JAPAN/GERMANY OR EQUVALINET Ultra sonic energy for uniform and highly dense 1-5 microns. A More than 96% of 0.3 micron or larger air borne dust particles is effectively shut out with the air filter to provide purified air for aerosol nebulization. Medication cup with replaceable diaphragm. *Easily detachable fan cover and pneumoclean (Airfilter). * Made of highly resistant sterilize able resin. *Stand with solution bottle for safety. *Nebulizingrate: 4 ml /min or greater. *Mist particle size:Approx 1-5 microns. *Nebulizing time setting: 1-30min & continuous Medication cup capacity: 150ml. * Accessories: *Tray setf or nebulizer with tray track and pole mount fitting.</p>	81				
---	--	----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

8	<p>Baby Weighing Scale Make:UK/USA/JAPAN/GERMANY OR EQUVALINET The generous tray on this baby scale leaves room for a soft blanket. The tray is non-tip and the softly rounded walls are high to prevent lively babies from rolling off during measurement. The LCD and digits are large and clear and a built in spirit level ensures level positioning. The tray is detachable for easy cleaning. A measuring rod can be added so that weight and length can be measured at the same time.. Functions TARE HOLD key for lively babies ZERO Auto OFF Features Detachable tray LCD display with 3cm high digits Battery or power supply with AC adapter</p>	03																						
	<table border="1"> <tr> <td data-bbox="331 1203 1052 1239">D & C Kit</td> <td data-bbox="1052 1203 1131 1239"></td> </tr> <tr> <td data-bbox="331 1239 1052 1274">Hegar Dilator Set of 8</td> <td data-bbox="1052 1239 1131 1274">1</td> </tr> <tr> <td data-bbox="331 1274 1052 1310">Curette D / E</td> <td data-bbox="1052 1274 1131 1310">4</td> </tr> <tr> <td data-bbox="331 1310 1052 1346">Vulsellum forceps</td> <td data-bbox="1052 1310 1131 1346">1</td> </tr> <tr> <td data-bbox="331 1346 1052 1382">Sherman Curette</td> <td data-bbox="1052 1346 1131 1382">1</td> </tr> <tr> <td data-bbox="331 1382 1052 1417">Ramply sponge holding forceps</td> <td data-bbox="1052 1382 1131 1417">4</td> </tr> <tr> <td data-bbox="331 1417 1052 1453">Towel Clip</td> <td data-bbox="1052 1417 1131 1453">6</td> </tr> <tr> <td data-bbox="331 1453 1052 1489">Cusco Speculum</td> <td data-bbox="1052 1453 1131 1489">1</td> </tr> <tr> <td data-bbox="331 1489 1052 1524">Sims speculum, Set of 3</td> <td data-bbox="1052 1489 1131 1524">1</td> </tr> </table>	D & C Kit		Hegar Dilator Set of 8	1	Curette D / E	4	Vulsellum forceps	1	Sherman Curette	1	Ramply sponge holding forceps	4	Towel Clip	6	Cusco Speculum	1	Sims speculum, Set of 3	1					
D & C Kit																								
Hegar Dilator Set of 8	1																							
Curette D / E	4																							
Vulsellum forceps	1																							
Sherman Curette	1																							
Ramply sponge holding forceps	4																							
Towel Clip	6																							
Cusco Speculum	1																							
Sims speculum, Set of 3	1																							



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

9	Sims Uterine Sound	1	50				
	Metal Cathetar	1					
	Brown Tenaculum forceps	1					
	Flashing Curette	1					
	Plain forceps 8"	1					
	Bowl 4"	2					
	Bowl 8"	1					
	Instruments Tray	1					
	Total Pieces	28					
NOTE: Sample required with technical bid		CE/FDA USA/ISO					
13485-2003		Cutler/OTINS/Hiibro					
10	Delivery Kit		50				
	Plain forceps 7"	1					
	Tooth forceps 7"	1					
	Cord Scissor USA pattern	1					
	Mayo Scissor Str	1					
	Mosquito Artery forceps str	6					
	Artery forceps Str 6"	6					
	Needle Holder 7'	1					
	Sims Speculum Med	1					
	Gali Pot	2					
	Large Bowl 8"	1					
	Kidney Tray 10"	1					
	Instruments tray 10" x 12"	1					
	Total Pieces	23					
NOTE: Sample required with technical bid		CE/FDA USA/ISO					
13485-2003		Cutler/OTINS/Hiibro					



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

11	<p>Delivery Table Make: UK/USA/JAPAN/GERMANY OR EQUVALINET</p> <ul style="list-style-type: none">• Lateral tilt left/right....20°/20°■ Trend elenburg/reverse trend elenburg..30°/30°■ Back section up/down tilt angle....45°/0°■ Head section up/down tilt angle....30°/90°• Leg section up/down tilt angle....0°/90° <p>Standardaccessories:</p> <ol style="list-style-type: none">1. Head rest2. Shoulder supports3. Body supports4. Arm rest5. Screen frame6. Knee crutch7. Mattress8. Built-in kidney lift, removable9. Foot pedal10. Crank handle11. Locking screw12. Rotation control13. Trend eleburg movement control14. Lateral tilt control15. Head & back section control16. Eccentric lever17. Kidney lift control18. Oil seal with cap19. Table base 50mm thicknass	10				
----	---	----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

	Optional accessories: 1.Restrain strap 2.Hand grips (pair) 3.U-shaped head rest 4.Raised arm rest 5. S/S funnel & waste receptacle & corrugated tubing					
12	Dressing Drum Round Stainless Steel Sample required with technical bid Cutler/Platinum/Hilbro/OTINS	NOTE:	128			
13	Drip Stand Rounded stainless steel Sample required with technical bid Cutler/Platinum/Hilbro/OTINS	NOTE:	97			



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

14	SMR Set						
	Freer elevator and separators	2					
	Killian nasal Gouge V shaped Bayonet	2					
	Killian nasal Gouge V shaped Str	1					
	Lucae nasal Chisel	2					
	Hammar S.S	1					
	Frazier suction tube No 8-10-12	3					
	Telly nasal dressing forceps	1					
	Luc turbinate forceps small	1					
	Luc turbinate forceps Medium	1					
	Hartmann Crocodil forceps	1					
	Killian nasal speculum Set of 4	1					
	Heyman turbinactomy scissor	1	40				
	Thudiam nasal speculum set of 5	1					
	Trocar Cannula	1					
	Jenson middleton bone nibler D/A	1					
	Walsham septum forceps right	1					
	Walsham septum forceps left	1					
	Instruments tray	1					
		Total Pieces	23				
			Sample required with				
		technical bid					
		CE/FDA USA/ISO 13485-2003					
	Cutler/OHNS/Hibro						



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

15	<p>Examination Coach Main Frame made by 1 1/4 x 1 1/4 M.S Pipe 18 gauge with head side adjustable in synthetic leather of best quality with foam 10 years warranty and the frame is in powder coating Cutler/OTINS/Hilbro/Platinum</p>	119					
	<p>General Surgical Instrument</p> <p>B.P Handle No 3</p> <p>B.P Handle No 4</p> <p>B.P Handle No 7</p> <p>Sponge holding forceps</p> <p>Dissecting forceps Plain / Tooth 6" 1+1</p> <p>Dissecting forceps Plain / Tooth 8" 1+1</p> <p>Adson forceps Plain & Tooth 2+2</p> <p>McIndoe forceps 6"</p> <p>Metzenbaum Scissor 6" Str & Cvd 1+1</p> <p>Metzenbaum Scissor 7" Str & Cvd 1+1</p> <p>Metzenbaum Scissor 8" Str & Cvd 1+1</p> <p>Mayo Scissor 6 3/4" Str & Cvd 1+1</p> <p>Crile artery forceps Str 6"</p> <p>Crile artery forceps Cvd 6"</p> <p>Mosquito artery forceps Str</p> <p>Mosquito artery forceps Cvd</p> <p>Artery forceps 8" Str & Cvd 4+4</p> <p>Artery forceps 9" Cvd fine</p> <p>Keener Artery forceps 8" Str & Cvd</p> <p>Babcock intestinal forceps 6"</p> <p>Babcock intestinal forceps 8"</p>	<p>2</p> <p>2</p> <p>2</p> <p>6</p> <p>2</p> <p>2</p> <p>4</p> <p>1</p> <p>2</p> <p>2</p> <p>2</p> <p>2</p> <p>12</p> <p>24</p> <p>12</p> <p>18</p> <p>8</p> <p>2</p> <p>8</p> <p>2</p> <p>2</p>					



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

16

Allis tissue forceps	4
Allis tissue forceps	4
Backhaus Towel Clip	6
Moynihan Forceps 8"	2
Langenback Retractor Small	2
Langenback Retractor Med	2
Morris Retractor Medium	2
Balfour Selfretaining Retractor	1
Sinus forceps	1
Volkman Curette D/E	2
Aneurysm Needle	1
Watson Chyne Dissector	1
Yaunker suction nozal	1
Mayo Hegar Needle holder 6"	2
Mayo Hegar Needle holder 7"	2
Mayo Hegar Needle holder 8"	2
Mayo Hegar Needle holder 9"	2
Kidney pedicel Clamp	4
Kidney Tray 10"	2
Bowl 4"	2
Instruments tray large without lid	1
Total Pieces	163
NOTE: Sample required with technical bid	
CE/FDA USA/ISO 13485-2003	
Cutler/OTINS/Hilbro	

60



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

17	Air Condition Spilit 1-1.5 Ton Cooling Capacity (BTU/Watts) 18000/5300 Rated current input (A)08 EER 03 Power Supply (V/HZ) 220-240/50 Cooling Power input (max)W 2500 Cooling Current input (Max)A 11.6 Electric Shock Protection CLASS 1 Indoor airflow (m3/H) 800 Rated Power Input (W) 1800 Indoor/Outdoor noise dB 37-44/54 Refrigerant R22/1300g Gross weight (IDU/ODU) 15-42 (Kg) Dimension indoor mm1055*370*300 Dimension outdoor mm920*335-595 Ideal Room size (sq.ft) 151-250 Evaporator: Hydrophilic fins with inner grove copper tube Condenser: Golden fins with inner grove copper tube Make: Dawlance/PEL/Gree	25				
18	Printer for Ultra Sound Machine	7				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

19	Refrigerator PRA-160 At Ambient Temp 55C Capacity 420 Liters Nominal Tension 220 Volts Frequency 50Hz Compressor Wattage 217W Total Current 1.8A Freezing capacity (24hrs) 9.5Kg Defrosting AUTO Adjustable Thermostat YES CFC Gas (R-34a) 0.18Kg, 1845 H(mm) Packing Dimension W (mm) 730, D(mm) 780 Net weight 85Kg Gross weight 99Kg Make: Dawlance/PEL	20				
20	Stabilizer A-20 SP Special for very low voltage 3 Relay 2 Meter Input V 100 Watts 2000 Make: Universal/Platinum/Golden	62				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

21	<p>Electric Water Cooler Tank: 45 Liters Body: Megan, Steel Water Tank: Nonmegan, Steel Voltage Rating: Operateon 220-240. 50 cycles A.C supply single phase Water Temperature Controllable from 9C to 15C</p> <p>Refrigeration System: Fitter with high quality world famous quick Compressor, Condenser and fan motor to ensure long lasting Performance, All cooling system hermitically sealed with RE-12R-22 as refrigerant.</p> <p>WaterSystem: Main Cooling system is made of Copper and Brass. Water Storage tank is made of high quality copper sheet tinned from inside and tested on maximum pressure according to (PSI-142,10KG). Fitted with high quality chrome plated brass taps</p> <p>Make: Canon/SuperAsia/China</p>	25				
----	---	----	--	--	--	--



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

22	UPS with all accessories 10 Tube Lights or 25 Energy Savers & 11 Fans Key Features Built-in intelligent 3-stage charger control for efficient charging and to prevent overcharging Light weighted, wall mounted, rack design User friendly LCD & LED Multiple protections: low battery alarm, low battery shutdown, over load protection, over heating protection, short circuit protection Highly efficient DC-to-AC conversion, minimizing energy loss Free AC & DC cables Innovative charging system especially designed to suit local type batteries and local weather conditions Load Range 1500 Watts 2000VA/24VDC/220VAC With Battery Make:Osaka/Exide	21				
Furniture / Fixture						
23	Office Chair Made with talhi/shesham wood seat & back netted with Spirit Polish Standard Size Best Quality NOTE: Sample required with technical bid	420				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

24	<p>Office Table Made by sheesham wood with glass top and side rack handles, channel stocks are best quality finish by spirit polish Size: (L) 4x(W)2 1/2 (H) 1/2 Best Quality NOTE: Sample required with technical bid</p>	24				
25	<p>Revolving Chair Main frame made with talhi/shesham wood seat & back plastic netted plastic based Stand with spirit polish. Best Quality NOTE: Sample required with technical bid</p>	156				
26	<p>Wheel Chairs Wheel Chair Extra Wide Heavy Duty Model 951B Powder Coating Mild Steel Frame, 22" Seat Width, Flip-Up-Arm Est, Detachable Footrest, Rubber Foam Castor, Pneumatic Mag Wheel. Weight Capacity: 150KG. NOTE: Sample required with technical bid Make: China/Pakistan</p>	33				
27	<p>Meeting Hall Chair Revolving Revolving Chairs-Adjustable lumbar support and instant seat height adjustment. Heavy duty plastic 5-prong base Height/width adjustable arms with soft, durable urethane pads. Minimum Dimensions: Overall: 30-1/4"Wx28-3/4"Dx42"to44-1/2"H. NOTE: Sample required with technical bid</p>	96				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

28	Meeting Hall Table Training Hall Table (20'X4'-6"X3'H): Made of high density laminations sheet with pvc edge banding on edges Front side of table should be covered with lamination sheet). The Colour of Lamination sheet will be specified by Engineer in charge.	1				
29	Patients Trolley Made by stainless steel square pipe in 16 gauge with stretcher made in 1 1/4 Round Pipe in stainless Steel with Curve sheet 6 inch imported Casters Locally Made ExportQuality NOTE: Sample required with technical bid	36				
30	Steel Fix Chairs in each Imported 3 seater metallic bench Non resisting stuff Full metallic structure 100% imported All Stuff available in pin pack cotton NOTE: Sample required with technical bid Make: China/Platinum/Deskon	169				
31	Bed Side Lockers Made by 22 gauge sheet with one space and a cabinet with 2 inch wheel finished by epoxy powder coated top is made by fiber material is very strong and water proof Draw panel and door is made by fiber material door is hang by stainless steel hanger finished by scratch and chip resistance powder coating Locally made Export Quality. NOTE: Sample required with technical bid	30				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

32	Fowler Bed Single crank operation Removable head and foot panels Rotary buffers for protection Mobile of 5" diagonally locking casters Dimension (78.5*36.5*24) Inches Main Frame made by 1 1/4 x 2 1/2 x 16 gauge upper frame NOTE: Sample required with technical bid	30				
33	Instrument Trolley 18 X 24 With square pipe top 5.5 powder coated NOTE: Sample required with technical bid	6				
34	Instrument Tray Full Stainless steel NOTE: Sample required with technical bid	34				
35	Cup Board Iron cup boards (6.5 feet x 3 feet) for record keeping Steel Almirah of the size 78"x36"x18" with 2 steel doors for the books/library having four shelves making five compartments, 20 gauge over all with locking arrangement and complete in all respect NOTE: Sample required with technical bid	57	Steel			
36	Iron Rack Four Shelf NOTE: Sample required with technical bid	63				



OFFICE OF THE DISTRICT HEALTH OFFICER
KHAIRPUR

37	Maters Single 6.5'*3.5'*4" With leather / regisine cover NOTE:Sample required with technical bid Make: Master/Celeste	54				
----	---	----	--	--	--	--

~~DISTRICT HEALTH OFFICER
KHAIRPUR~~